



U.S. Department of the Interior  
U.S. Geological Survey

**RESULTS OF THE U.S. GEOLOGICAL SURVEY'S ANALYTICAL  
EVALUATION PROGRAM FOR STANDARD REFERENCE SAMPLES  
DISTRIBUTED IN SEPTEMBER 2002**

---

Open-File Report 02-481

**Results of the U.S. Geological Survey's Analytical  
Evaluation Program for Standard Reference Samples  
Distributed in September 2002**

**By Mark T. Woodworth and Brooke F. Connor**

---

---

**U.S. GEOLOGICAL SURVEY**

**Open-File Report 02-481**

**Lakewood, Colorado  
2003**

**U.S. DEPARTMENT OF THE INTERIOR**

**Gayle A. Norton, Secretary**

**U.S. GEOLOGICAL SURVEY**

**Charles G. Groat, Director**

---

For additional information  
write to:

Chief, Branch of Quality Systems  
U.S. Geological Survey  
Water Resources Division  
Box 25046, Mail Stop 401  
Denver Federal Center  
Denver, Colorado 80225-0046

Copies of this report can be  
purchased from:

U.S. Geological Survey  
Branch of Information Services  
DFC, Bldg. 810, Box 25286  
Denver, Colorado 80225-0286  
888-ASK-USGS

## CONTENTS

	Page
Definition of analytical methods, abbreviations, and symbols.....	iv
Abstract.....	1
Introduction.....	1
Preparation of standard reference samples.....	6
Laboratory analyses.....	7
Statistical presentation of data.....	10
Laboratory performance ratings.....	10
Reference.....	11

## FIGURE

Figure 1. Statistical parameters shown on data graphs in tables 11-16.....	11
--	----

## TABLES

1. USGS used laboratories that participated in the analyses of standard reference samples distributed in September 2002.....	3
2. Other laboratory participants in the analyses of standard reference samples distributed in September 2002.....	4
3. Analytes determined in standard reference samples distributed in September 2002.....	8
4. Overall laboratory performance ratings for standard reference samples distributed in September 2002.....	12
5. Laboratory performance ratings for standard reference sample T-171 (trace constituents).....	14
6. Laboratory performance ratings for standard reference sample M-164 (major constituents).....	26
7. Laboratory performance ratings for standard reference sample N-75 (nutrient constituents).....	33
8. Laboratory performance ratings for standard reference sample N-76 (nutrient constituents).....	35
9. Laboratory performance ratings for standard reference sample P-39 (low ionic-strength constituents).....	37
10. Laboratory performance ratings for standard reference sample Hg-35 (mercury).....	40
11. Statistical summary of reported data for standard reference sample T-171 (trace constituents).....	41
12. Statistical summary of reported data for standard reference sample M-164 (major constituents).....	69
13. Statistical summary of reported data for standard reference sample N-75 (nutrient constituents).....	85
14. Statistical summary of reported data for standard reference sample N-76 (nutrient constituents).....	90
15. Statistical summary of reported data for standard reference sample P-39 (low ionic-strength constituents).....	95
16. Statistical summary of reported data for standard reference sample Hg-35 (mercury).....	106
17. Most probable values for constituents and properties in standard reference samples distributed in September 2002.....	107

## Definition of analytical methods, abbreviations, and symbols

### Abbreviations and figure symbols

C = Celsius
$F\sigma$ = F-pseudosigma - nonparametric statistic for deviation
HCl = hydrochloric acid
Hg = mercury sample
$HNO_3$ = nitric acid
Lh = lower hinge value
L = liter
Lab = laboratory
mg/L = milligrams per liter
mL = milliliter
M - = major ion sample
MPV = most probable value (center line on graphs)
n = number of analyses
N = Normality
N - = nutrient sample
NR = not rated, less than values reported or insufficient data
OLR = overall laboratory rating for each sample type
OWR = overall weighted rating for all sample types
P - = precipitation sample (low ionic-strength, typically <50 $\mu$ S/cm)
ppm = parts per million
SRS = USGS standard reference sample
T - = trace metal sample
Uh = upper hinge value
USGS = United States Geological Survey
V = number of rated analyses
Z-value = number of F-pseudosigmas from the MPV
$\mu$ g/L = micrograms per liter
$\mu$ m = micrometer
$\mu$ S/cm = microsiemens per centimeter at 25° Celsius
< = less than
-- = not reported

### Analytical methods and codes

Code	Method
0	Other
1	Atomic absorption: direct, air
2	Atomic absorption: direct, nitrous oxide
3	Atomic absorption: graphite furnace
4	Inductively coupled plasma
5	Direct current plasma
6	Inductively coupled plasma/mass spectrometry
7	Ion chromatography
8	Atomic absorption: cold vapor
9	Atomic fluorescence
10	Atomic absorption: extraction
11	Atomic absorption: hydride
12	Flame emission
20	Titration: colorimetric
21	Titration: electrometric
22	Colorimetric
40	Ion selective electrode
41	Electrometric [pH and specific conductance]
50	Gravimetric
51	Turbidimetric

### Formulas

MPV = median value (excluding less than values)

F-pseudosigma ( $F\sigma$ ) =  $(U_h - L_h)/1.349$

$U_h$  = median of the upper half of the reported values (excluding less than values)

$L_h$  = median of the lower half of the reported values (excluding less than values)

Z-value =  $(\text{reported value} - \text{MPV})/F\text{-pseudosigma}$

OLR = mean of all rated analytes for sample type

OWR = 
$$\frac{(\text{OLR} \cdot V_1) + (\text{OLR} \cdot V_2) \dots (\text{OLR} \cdot V_n)}{(V_1 + V_2 + \dots + V_n)}$$
 for each SRS type

### Ratings

Rating	Absolute Z-value
4 (Excellent)	0.00 to 0.50
3 (Good)	0.51 to 1.00
2 (Satisfactory)	1.01 to 1.50
1 (Marginal)	1.51 to 2.00
0 (Unsatisfactory)	Greater than 2.00

# RESULTS OF THE U.S. GEOLOGICAL SURVEY'S ANALYTICAL EVALUATION PROGRAM FOR STANDARD REFERENCE SAMPLES DISTRIBUTED IN SEPTEMBER 2002

By Mark T. Woodworth and Brooke F. Connor

## ABSTRACT

This report presents the results of the U.S. Geological Survey's analytical evaluation program for six standard reference samples -- T-171 (trace constituents), M-164 (major constituents), N-75 (nutrient constituents), N-76 (nutrient constituents), P-39 (low ionic-strength constituents), and Hg-35 (mercury) -- that were distributed in September 2002 to laboratories enrolled in the U.S. Geological Survey sponsored interlaboratory testing program. Analytical data received from 102 laboratories were evaluated with respect to overall laboratory performance and relative laboratory performance for each analyte in the six reference samples. Results of these evaluations are presented in tabular form. Also presented are tables and graphs summarizing the analytical data provided by each laboratory for each analyte in the six standard reference samples. The most probable value for each analyte was determined using nonparametric statistics.

## INTRODUCTION

The U.S. Geological Survey (USGS) conducts an interlaboratory analytical evaluation program semiannually. This program provides a variety of standard reference samples (SRSs) to accomplish quality assurance testing of laboratories and to provide an adequate supply of samples that contribute to quality control programs of participating laboratories. Natural-matrix reference materials are preferred for use in this interlaboratory evaluation program. A series of samples are prepared and distributed each spring and fall. The program began in 1962 with a single sample containing major constituents that was prepared from distilled water and reagent-grade chemicals. Twenty-three USGS laboratories participated in the first analytical evaluation program. Since that time, objectives of the program have been to:

- (1) evaluate and improve the performance of USGS and other participating laboratories;
- (2) provide a library of carefully prepared, homogeneous, stable, reference materials for use in the quality control programs of laboratories;
- (3) identify analytical problem areas;
- (4) identify quality assurance needs with respect to environmental analyses and develop new reference materials to meet these needs; and
- (5) evaluate the accuracy and precision of analytical methods.

Over 275 USGS and non-USGS laboratories are enrolled in the program, which can currently provide 5 different types of SRSs:

1. Trace constituents.
2. Major constituents.
3. Nutrient constituents.
4. Low ionic-strength constituents.
5. Mercury.

Though this is not a laboratory certification program, participation in this continuing quality assurance program is mandatory for all laboratories providing water-quality data for USGS sponsored reports or storage in the USGS national databases. The results from this study can be used to alert participating laboratories of possible deficiencies in their analytical operations and provide reference materials for laboratory quality-control programs. Laboratories that provide data for the USGS are identified with a laboratory identification number while all other laboratories are kept confidential.

A supply of SRSs from previous evaluations, is available. USGS offices and participating laboratories can purchase these SRSs for further testing, continuing quality assurance, and quality-control programs by contacting:

U.S. Geological Survey  
Branch of Quality Systems  
SRS Purchasing  
Denver Federal Center, Bldg. 53  
P. O. Box 25046, MS 401  
Denver, Colorado 80225-0046  
(303) 236-1875

This report summarizes the analytical results submitted by 102 laboratories for the September 2002 evaluation (table 1 and table 2). Analytical results for the following are presented in this report:

T-171	Trace constituents	N-76	Nutrient constituents
M-164	Major constituents	P-39	Low ionic-strength constituents
N-75	Nutrient constituents	Hg-35	Mercury

Laboratories that are providing analytical services to USGS offices are requested to analyze the appropriate SRSs for the same analytes requested by the USGS offices. All laboratories are requested to include the analytical methods used to determine the concentration of each analyte. When analytical method information was provided, it has been included in tables 11-16.

Not all SRSs are requested or necessarily analyzed by all the laboratories; nor do all laboratories enrolled in the program participate in each evaluation.

**Table 1. USGS used laboratories that participated in the analyses of standard reference samples distributed in September 2002**

<b>Lab</b>	<b>Participating Laboratory</b>	<b>City</b>	<b>State</b>
1	U.S. Geological Survey - National Water Quality Laboratory	Denver	CO
4	U.S. Geological Survey - Utah District Laboratory	Salt Lake City	UT
5	U.S. Bureau of Reclamation	Denver	CO
10	U.S. Bureau of Reclamation	Boise	ID
12	Metro Wastewater Reclamation District	Denver	CO
16	Oklahoma Department of Environmental Quality	Oklahoma City	OK
21	UC Davis - Department of Environmental Science & Policy	Davis	CA
23	City of Fort Collins - Water Quality Laboratory	Ft. Collins	CO
25	Kentucky Geological Survey	Lexington	KY
33	U.S. Geological Survey - Panola Mountain Watershed Project	Atlanta	GA
46	Wisconsin State Laboratory of Hygiene	Madison	WI
59	Division of Consolidated Laboratory Services	Richmond	VA
70	University of Iowa Hygienic Laboratory	Des Moines	IA
72	New Jersey Department of Health	Trenton	NJ
91	Georgia Department of Natural Resources	Atlanta	GA
102	Heidelberg College	Tiffin	OH
105	Pennsylvania Department of Environmental Protection	Harrisburg	PA
109	North Dakota State Water Commission Laboratory	Bismarck	ND
110	U.S. Geological Survey - New York District Laboratory	Troy	NY
118	Virginia Tech - Occoquan Watershed Monitoring Laboratory	Manassas	VA
121	University of Hawaii - Department of Oceanography	Honolulu	HI
134	Ocala Water Quality and Research Laboratory	Ocala	FL
142	North Dakota Department of Health	Bismarck	ND
147	U.S. Geological Survey - Surface Water Quality Research	Boulder	CO
180	Clean Water Services	Hillsboro	OR
193	Vermont Department of Environmental Conservation Laboratory	Waterbury	VT
205	Olsen's Agriculture Laboratory	McCook	NE
208	U.S. Geological Survey - Water Resources Division	San Diego	CA
212	Sewer Trent Laboratories	Arvada	CO
219	U.S. Geological Survey - Geologic Division	Denver	CO
220	U.S. Bureau of Reclamation	Bismarck	ND
224	University of Arkansas - Water Quality Laboratory	Fayetteville	AR
254	U.S. Geological Survey, WRD, NRP	Menlo Park	CA
255	Colorado Springs Utilities - Water Resource Department	Colorado Springs	CO
319	Fairfax County Environmental Services	Lorton	VA
330	Kennecott Environmental Laboratory	Magna	UT
333	U.S. Geological Survey - Colorado District WEBB Laboratory	Denver	CO
341	Michigan Department of Environmental Quality	Lansing	MI
349	South Dakota Department of Health	Pierre	SD
356	Washington State Department of Ecology - Manchester Laboratory	Port Orchard	WA
366	TriMatrix Laboratory	Grand Rapids	MI
369	New Hampshire Department of Environmental Services	Concord	NH
373	City of Tulsa - Quality Assurance Laboratory	Tulsa	OK
374	U.S. Geological Survey - Miami Subdistrict Laboratory	Miami	FL
378	Minnesota Department of Health	Minneapolis	MN
379	Mississippi Department of Environmental Quality Laboratory	Pearl	MS
380	New Mexico Health Department	Albuquerque	NM
381	North Atlantic Coastal Laboratory	Wellfleet	MA
383	University of Vermont - Department of Plant and Soil Science	Burlington	VT
384	SUNY - College of Environmental Science and Forestry	Syracuse	NY
385	Virginia Tech Soil Testing Laboratory	Charleston	WV
386	Johnson County Environmental Laboratory	Mission	KS
387	Virginia Tech Hydro-Geochemistry Laboratory	Charleston	WV



**Table 2. Other laboratory participants in the analyses of standard reference samples distributed in September 2002**

<b>Participating Laboratory</b>	<b>City</b>	<b>State</b>
Albion Environmental	College Station	TX
Aqua Tech Environmental Laboratory (ATEL)	Marion	OH
Chemical Solutions Ltd	Mechanicsburg	PA
City of Northglenn - Water Treatment Facility	Northglenn	CO
City of Albuquerque - Water Quality Laboratory	Albuquerque	NM
City of Pueblo - Wastewater Treatment Plant	Pueblo	CO
City of Tallahassee - Water Quality Laboratory	Tallahassee	FL
Columbia Analytical Services	Rochester	NY
Denver Water Department	Denver	CO
Desert Research Institute	Reno	NV
Florida Department of Environmental Protection	Tallahassee	FL
High Sierra Water Laboratory	Truckee	CA
Huffman Laboratories	Golden	CO
Institute of Ecosystem Studies	Millbrook	NY
Kansas Geological Survey	Lawrence	KS
Lower Colorado River Authority - Environmental Laboratory Services	Austin	TX
Madison Public Health Laboratory	Madison	WI
Metropolitan Water District of Southern California	La Verne	CA
Montana Bureau of Mines & Geology	Butte	MT
Old Dominion University - Applied Marine Research Laboratory	Norfolk	VA
Severn Trent Laboratories	Tallahassee	FL
South Florida Water Management District	West Palm Beach	FL
Southwest Florida Water Management District	Brooksville	FL
Suffolk County Water Authority Laboratory	Hauppauge	NY
Texas A&M University - Trace Element Research Laboratory	College Station	TX
U.S. Bureau of Reclamation	Alamosa	CO
U.S. Department of Agriculture - Cooperative Chemical Analytical Laboratory	Corvallis	OR
U.S. Forest Service	Ft. Collins	CO
U.S. Geological Survey Geologic Division - Coal Analytical Laboratory	Denver	CO
University of Georgia - Soil, Plant, & Water Laboratory	Athens	GA
University of Maryland - Chesapeake Biological Laboratory	Solomons	MD
University of Maryland - Horn Point Laboratory	Cambridge	MD
University of Montana - Murdock Environmental Biogeochemistry Laboratory	Missoula	MT
University of Tennessee - Department of Civil & Environmental Engineering	Knoxville	TN
UZ HydroChemistry Laboratory, YMPB	Denver	CO
West Coast Analytical Service, Inc.	Santa Fe Springs	CA
WMRC - Hazardous Waste Research Center	Champaign	IL
Wyoming Department of Agriculture	Laramie	WY

**Table 2. Other laboratory participants in the analyses of standard reference samples distributed in September 2002 -- continued**

<b>Middle East Participating Laboratory</b>	<b>Location</b>	
Center for Environmental & Occupational Health Sciences	Ramallah	West Bank via Israel
Environmental Research Centre - Jubeiha	Amman	Jordan
Geological Survey of Israel Laboratory	Jerusalem	Israel
Israeli Hydrologic Service Laboratory	Beit-Dagan	Israel
Mekorot Laboratory	Ashqelon	Israel
Mekorot Water Co. Ltd. Central Lab	Nazareth Illit	Israel
Palestinian Water Authority	Al-Beireh	West Bank via Israel
Public Health Laboratory	Tel Aviv	Israel
Public Health Laboratory - Ministry of Health	Beer Sheva	Israel
Water & Soil Environmental Research Unit	Bethlehem	West Bank via Israel
Water Authority of Jordan	Amman	Jordan

## PREPARATION OF STANDARD REFERENCE SAMPLES

All of the SRSs used in this evaluation were prepared by USGS personnel located in Lakewood, Colorado, and were analyzed for analyte concentrations and physical property values before mailing. A supply of these SRSs is maintained and are available to purchase by participating laboratories and USGS offices for use in their quality-control programs.

Trace constituents sample T-171 was prepared using water collected from Big Thompson River west of Drake, Colorado. The water was pumped through a 0.2- and 0.1-micrometer ( $\mu\text{m}$ ) filter into a 1325-liter (L) polypropylene drum. The water was continuously circulated and passed through a 0.1- $\mu\text{m}$  filter and ultraviolet sterilizer for 24 hours. The water was then acidified to a pH <2 with nitric acid ( $\text{HNO}_3$ ) and chlorinated to 5 parts per million (ppm) free chlorine with sodium hypochlorite. The trace constituent concentrations were adjusted by adding reagent-grade chemicals. The sample was circulated through a 0.1- $\mu\text{m}$  filter and an ultraviolet sterilizer for an additional 24 hours prior and during bottling. The 500-milliliter (mL) polypropylene bottles and caps were acid leached with 0.16N  $\text{HNO}_3$ , deionized-water rinsed, and autoclave sterilized.

Major constituents sample M-164 was prepared using water collected from Bear Creek east of Kittridge, Colorado. The water was pumped through a 0.2- and 0.1- $\mu\text{m}$  filter into a 1325-L polypropylene drum. The water was continuously circulated and passed through a 0.1- $\mu\text{m}$  filter and ultraviolet sterilizer for 24 hours. The water was then chlorinated to 5-ppm free chlorine with sodium hypochlorite. The major constituent concentrations were adjusted by adding reagent-grade chemicals. The sample was circulated an additional 24 hours, then allowed to sit for 48 hours. During bottling, the sample was pumped through an ultraviolet sterilizer and a 0.1- $\mu\text{m}$  filter. The 500-mL polypropylene bottles and caps were acid leached with 0.16N  $\text{HNO}_3$ , deionized-water rinsed, and autoclave sterilized.

Nutrient constituents sample N-75 was prepared in a 50-L polypropylene carboy using deionized water. This SRS was prepared the week prior to sample distribution. The water was circulated through a 0.1- $\mu\text{m}$  filter and kept chilled with ice (12 degrees Celsius) during the entire preparation procedure. Ultraviolet sterilization was performed up until the addition of reagent-grade chemicals. The 60-mL amber glass vials and teflon-faced rubber-lined caps were acid leached with 0.1N hydrochloric acid (HCl), deionized-water rinsed, and autoclave sterilized.

Nutrient constituents sample N-76 was prepared in a 190-L polypropylene drum using water collected from Chicago Creek south of Idaho Springs, Colorado. This SRS was prepared the week prior to sample distribution. The water was circulated through a 0.1- $\mu\text{m}$  filter and kept chilled with ice (12 degrees Celsius) during the entire preparation procedure. Ultraviolet sterilization was performed up until the addition of reagent-grade chemicals. The 250-mL polyethylene bottles were acid leached with 0.1N HCl, deionized-water rinsed, and autoclave sterilized.

Low ionic-strength constituents sample P-39 was prepared in a 600-L polypropylene drum using water collected from West Fork Clear Creek east of Berthoud Falls, Colorado. The sample was diluted to obtain lower concentrations. Prior and during bottling, the sample was circulated through a 0.1- $\mu\text{m}$  filter and an ultraviolet sterilizer. The 500-mL polypropylene bottles and caps were acid leached with 0.16N  $\text{HNO}_3$ , deionized-water rinsed, and autoclave sterilized.

Mercury sample Hg-35 was prepared using deionized water. The sample was prepared in a 45-L glass carboy. It was preserved with 5 mL/L of 12 N HCl. The desired mercury concentration was obtained by adding a mercury standard solution. The 250-mL borosilicate glass bottles and teflon-lined caps were new, acid leached, and deionized-water rinsed.

#### LABORATORY ANALYSES

The participating laboratories were asked to determine constituents that are summarized in table 3. The number of analytes range from 1 in Hg-35 (mercury) to 28 in T-171 (trace constituents).

**Table 3. Analytes determined in standard reference samples distributed in September 2002**

[mg/L, milligrams per liter; µg/L, micrograms per liter; □µS/cm, microsiemens per centimeter at 25 degrees Celsius]

<b>Constituent or Property</b>		<b>Units</b>	<b>T-171</b>	<b>M-164</b>	<b>N-75</b>	<b>N-76</b>	<b>P-39</b>	<b>Hg-35</b>
Acidity	Acidity as CaCO <sub>3</sub>	mg/L					X	
Alk	Alkalinity as CaCO <sub>3</sub>	mg/L		X				
Ag	Silver	µg/L	X					
Al	Aluminum	µg/L	X					
As	Arsenic	µg/L	X					
B	Boron	µg/L	X	X				
Ba	Barium	µg/L	X					
Be	Beryllium	µg/L	X					
Ca	Calcium	mg/L	X	X			X	
Cd	Cadmium	µg/L	X					
Cl	Chloride	mg/L		X			X	
Co	Cobalt	µg/L	X					
Cr	Chromium	µg/L	X					
Cu	Copper	µg/L	X					
ROE	Dissolved Solids	mg/L		X				
F	Fluoride	mg/L		X			X	
Fe	Iron	µg/L	X					
Hg	Mercury	µg/L						X
K	Potassium	mg/L	X	X			X	
Li	Lithium	µg/L	X					
Mg	Magnesium	mg/L	X	X			X	
Mn	Manganese	µg/L	X					
Mo	Molybdenum	µg/L	X					
Na	Sodium	mg/L	X	X			X	
NH <sub>3</sub> as N	Ammonia	mg/L			X	X		
NH <sub>3</sub> + Org N as N	Ammonia + Organic N	mg/L			X	X		
Ni	Nickel	µg/L	X					
NO <sub>3</sub> as N	Nitrate	mg/L			X	X		
Pb	Lead	µg/L	X					
pH	pH	unit		X			X	
PO <sub>4</sub> as P	Orthophosphate	mg/L			X	X	X	
total P as P	Phosphorus	mg/L		X	X	X		
Sb	Antimony	µg/L	X					
Se	Selenium	µg/L	X					
SiO <sub>2</sub>	Silica	mg/L	X	X				
SO <sub>4</sub>	Sulfate	mg/L		X			X	
Sp Cond	Specific Conductance	µS/cm		X			X	
Sr	Strontium	µg/L	X	X				
Tl	Thallium	µg/L	X					
U	Uranium	µg/L	X					
V	Vanadium	µg/L	X	X				
Zn	Zinc	µg/L	X					

Laboratories were requested to identify the method used for each constituent according to analytical method codes in the list of definitions, abbreviations, and symbols (page iv).

Participating laboratories were also asked to identify the method used, such as those references listed next, to further define the methods.

1. American Public Health Association, American Water Works Association, and Water Environment Federation, 1995, Standard methods for the examination of water and wastewater (19th ed.): Washington, D.C., American Public Health Association, variable pagination.
2. American Society for Testing and Materials, 1995, Annual book of ASTM standards: Philadelphia, v. 11.0, and v. 11.02.3.
3. Kopp, J.F., and McKee, G.F., 1979, Methods for chemical analysis of water and wastes: Cincinnati, U.S. Environmental Protection Agency, EPA 600/4-79-020, rev. 1983,460 p.
4. Fishman, M.J., and Friedman, L.C., eds., 1989. Methods for determination of inorganic substances in water and fluvial sediments (3rd ed.): U.S. Geological Survey Techniques of Water-Resources Investigations, Book 5, Chapter A1, 545 p.
5. Miscellaneous manufacturer's instrument manuals or references.

## STATISTICAL PRESENTATION OF DATA

Data in this report are evaluated using nonparametric statistics as described by Hoaglin and others (1983). This statistical approach is a resistant statistic because outliers have less influence on the median, than on the mean in traditional parametric statistics. Analytical data for each analyte are presented in tabular and graphical forms in tables 11 - 16. Tabulated data for each analyte include the laboratory identification number; reported values; analytical method; most probable value (MPV); number of reported analyses, excluding less than values, (n); data range; the F-pseudostandard deviation; and the Z-value. The Z-value is equivalent to the Z-score of traditional statistics. The F-pseudostandard deviation approximates the standard deviation ( $\sigma$ ) of traditional statistics when the data has a Gaussian distribution.

The median value, calculated from the reported results, is the MPV. The F-pseudostandard deviation is calculated by dividing the fourth-spread (analogous to interquartile range) by 1.349; therefore the smaller the F-pseudostandard deviation the more precise the determinations. The 1.349 value is derived from the number of standard deviations that encompasses 50% of the data. Statistical tables show that 25% of the area under a normal curve lies  $0.6745\sigma$  from the mean, so 50% lies  $1.349\sigma$ . The MPV and F-pseudostandard deviation are replaced with the term "inadequate data" when the overall number of analyses is less than seven or the calculated F-pseudostandard deviation is greater than the MPV. However, if an analyte has at least five analyses by a given method, the median and F-pseudostandard deviation are reported in the block of data listed for each method. Based on an assessment of analyte data (Keith Long, Branch of Quality Systems, verbal comm., 1998), when the F-pseudostandard deviation is less than 5 percent of the MPV, the rating criterion is set to 5 percent of the MPV. When applicable, rating criterion is shown in tables 5 - 17.


The graphical plot of the reported data is shown in figure 1. The upper and lower boundaries of the graphical plots are +3 and -3 F-pseudostandard deviation deviations from the median. Reported values are grouped by analytical method in ascending order of value.

## LABORATORY PERFORMANCE RATINGS

To facilitate laboratory intercomparison, laboratory performance ratings are included in tables 4 - 16 in this report. For each SRS, averages of all the analyte ratings and the number of rated analyses are given for each participating laboratory. The actual reported values by all the laboratories were used to calculate the statistical results and performance ratings presented in this report. Laboratory determination of each analyte is rated on a scale 4 to 0, based on the absolute Z-value. The listing of ratings and Z-values are presented in the list of analytical methods, abbreviations, and symbols given on page iv and in tables 5 - 10.

Laboratories reporting "less than" values are not performance rated unless their less than value is less than the MPV and has a Z-value greater than 2. In this case, the laboratory would receive a rating of 0 for that analyte.

Ratings are based on the relative performance of laboratories on specific samples and should be reviewed and evaluated on a case-by-case basis for each laboratory considering such factors as methods used and data needs of specific USGS projects using the laboratory data.



NOTE: vertical scale is the concentration value of the individual analyte in appropriate units (see table 3). Horizontal scale is the laboratory reported values separated by method (different symbols) and plotted by increasing values. Numbers next to each symbol at the bottom of the figure are analytical method codes as described on page iv. Laboratory-reported results outside  $\pm 3$  F-pseudostigma from the median are not shown on the graphs.

**Figure 1.** Statistical parameters shown on data graphs in tables 11-16

#### REFERENCE

Hoaglin, D.C., Mosteller, F., and Tukey, J.W., Eds. 1983, Understanding robust and exploratory data analysis: New York, NY, John Wiley, Inc., p. 38-41.



**Table 4. Overall laboratory performance ratings for standard reference samples distributed September 2002**

[SRS, standard reference sample; Lab, laboratory identification number; OWR, overall weighted rating for all sample types; OLR, overall laboratory rating for sample type; V/66, number of rated analyses out of 66 from all sample types; V/28, V/16, V/5, V/5, V/11, and V/1 are number of rated analyses for each sample type (T-171, M-164, N-75, N-76, P-39, HG-35) respectively; NR, not rated; --, not reported.]

Lab	SRS =		T-171		M-164		N-75		N-76		P-39		HG-35	
	OWR	V/66	OLR	V/28	OLR	V/16	OLR	V/5	OLR	V/5	OLR	V/11	OLR	V/1
1	3.3	63	3.6	28	2.9	15	3.2	5	3.4	5	3.0	9	4.0	1
2	3.3	9	--	--	--	--	--	--	--	--	3.3	9	--	--
4	1.3	3	--	--	1.3	3	--	--	--	--	--	--	--	--
5	2.4	54	2.2	19	2.6	16	3.0	5	1.8	5	2.7	9	--	--
7	2.7	24	2.7	24	--	--	--	--	--	--	--	--	--	--
8	3.0	62	3.0	27	3.0	16	3.8	5	2.2	5	3.3	9	NR	0
10	3.4	28	3.0	7	3.5	11	3.8	5	3.6	5	--	--	--	--
12	2.2	16	1.8	4	2.1	7	--	--	2.6	5	--	--	--	--
16	2.3	51	2.0	26	2.9	15	2.4	5	2.0	5	--	--	--	--
21	3.3	6	4.0	1	--	--	3.2	5	--	--	--	--	--	--
23	2.9	40	2.9	9	3.2	13	1.8	4	2.6	5	2.9	9	NR	0
24	2.9	23	2.8	10	3.0	13	--	--	--	--	--	--	--	--
25	1.5	55	0.9	21	1.9	15	2.0	5	2.6	5	1.6	9	--	--
26	2.5	20	2.3	6	2.8	11	--	--	1.7	3	--	--	--	--
30	4.0	2	--	--	--	--	--	--	4.0	2	--	--	--	--
31	3.2	5	--	--	--	--	3.2	5	--	--	--	--	--	--
32	3.0	45	3.1	28	2.9	16	--	--	--	--	--	--	0.0	1
33	2.4	34	2.4	7	2.2	12	0.0	3	2.7	3	3.3	9	--	--
38	3.1	26	--	--	3.6	10	1.8	5	3.4	5	3.0	6	--	--
42	2.5	45	2.5	24	2.3	15	3.0	3	3.0	3	--	--	--	--
45	3.1	56	3.3	26	3.1	15	1.3	3	1.3	3	3.4	8	3.0	1
46	3.3	29	2.6	8	3.8	11	3.3	4	3.2	5	--	--	4.0	1
59	3.0	34	3.2	20	2.8	13	--	--	--	--	--	--	0.0	1
64	3.5	31	4.0	5	3.1	10	3.8	4	3.5	4	3.4	8	--	--
70	2.0	41	1.3	19	3.1	12	2.2	5	1.6	5	--	--	--	--
72	2.1	10	--	--	--	--	1.6	5	2.6	5	--	--	--	--
76	3.8	26	3.8	12	3.9	10	4.0	2	3.5	2	--	--	--	--
85	3.5	26	--	--	2.9	11	4.0	5	3.8	5	3.8	5	--	--
86	3.3	48	3.4	20	3.5	13	2.0	3	4.0	3	3.2	9	--	--
91	2.9	13	--	--	2.2	5	3.0	4	3.8	4	--	--	--	--
97	2.7	45	2.4	22	2.9	14	2.5	4	3.4	5	--	--	--	--
100	2.6	35	2.6	20	2.7	15	--	--	--	--	--	--	--	--
102	2.1	20	--	--	2.1	10	1.4	5	2.8	5	--	--	--	--
105	2.8	49	3.3	17	3.1	14	1.3	4	2.8	5	2.2	9	NR	0
109	2.5	17	3.0	6	2.3	11	--	--	--	--	--	--	--	--
110	3.3	12	4.0	4	--	--	3.0	2	--	--	3.0	6	--	--
113	3.4	47	3.4	16	3.4	13	2.6	5	4.0	5	3.6	8	--	--
118	2.8	15	--	--	2.6	5	2.8	5	3.0	5	--	--	--	--
121	2.1	15	2.8	9	1.2	6	--	--	--	--	--	--	--	--
134	3.6	61	3.6	27	3.6	16	3.0	4	3.4	5	3.8	9	NR	0
138	3.4	60	3.3	24	3.4	16	3.8	5	3.8	5	3.4	9	4.0	1
142	3.1	34	3.5	15	2.8	9	2.8	5	2.6	5	--	--	--	--
146	1.9	38	2.3	15	2.3	13	0.6	5	1.0	5	--	--	--	--
147	3.2	9	3.6	8	--	--	--	--	--	--	--	--	0.0	1
149	3.1	31	3.2	24	2.7	7	--	--	--	--	--	--	--	--
151	3.0	30	2.5	14	3.8	12	2.5	4	--	--	--	--	--	--
180	3.2	51	3.5	21	3.3	11	3.0	5	2.6	5	2.7	9	NR	0
183	2.3	12	--	--	2.6	7	1.5	2	2.3	3	--	--	--	--
190	3.4	49	2.9	16	3.6	14	3.6	5	3.6	5	3.7	9	--	--
193	2.8	35	2.7	12	2.6	8	3.7	3	2.5	4	3.1	7	3.0	1

**Table 4. Overall laboratory performance ratings for standard reference samples distributed September 2002**

**-- continued**

[SRS, standard reference sample; Lab, laboratory identification number; OWR, overall weighted rating for all sample types; OLR, overall laboratory rating for sample type; V/66, number of rated analyses out of 66 from all sample types; V/28, V/16, V/5, V/5, V/11, and V/1 are number of rated analyses for each sample type (T-171, M-164, N-75, N-76, P-39, HG-35) respectively; NR, not rated; --, not reported.]

Lab	SRS =		T-171		M-164		N-75		N-76		P-39		HG-35	
	OWR	V/66	OLR	V/28	OLR	V/16	OLR	V/5	OLR	V/5	OLR	V/11	OLR	V/1
205	1.5	2	--	--	--	--	--	--	1.5	2	--	--	--	--
208	3.2	6	--	--	3.5	2	--	--	3.0	2	3.0	2	--	--
212	1.9	45	1.7	23	2.3	16	--	--	1.4	5	--	--	2.0	1
219	2.0	38	2.3	26	1.3	12	--	--	--	--	--	--	--	--
220	2.2	23	2.0	14	2.4	9	--	--	--	--	--	--	--	--
224	2.2	18	--	--	2.5	13	--	--	1.4	5	--	--	--	--
227	2.6	19	2.9	7	2.3	7	--	--	2.6	5	--	--	--	--
228	2.1	8	--	--	--	--	--	--	--	--	2.1	8	--	--
230	3.2	37	3.3	24	3.2	13	--	--	--	--	--	--	--	--
235	3.2	26	3.2	25	--	--	--	--	--	--	--	--	3.0	1
246	2.6	15	2.6	14	--	--	--	--	--	--	--	--	3.0	1
247	2.7	43	2.8	9	2.9	15	2.0	5	1.6	5	3.2	9	NR	0
254	3.2	25	3.1	16	3.6	9	--	--	--	--	--	--	--	--
255	2.2	15	1.8	11	3.3	4	--	--	--	--	--	--	--	--
256	3.0	29	2.8	17	3.3	12	--	--	--	--	--	--	--	--
257	1.5	17	0.7	7	2.0	10	--	--	--	--	--	--	--	--
259	3.7	31	3.7	18	3.7	13	--	--	--	--	--	--	--	--
265	3.3	45	3.5	28	3.1	11	--	--	--	--	3.2	6	--	--
266	3.1	12	--	--	3.1	12	--	--	--	--	--	--	--	--
273	1.2	36	0.8	12	1.5	15	--	--	--	--	1.1	9	--	--
274	1.0	26	0.0	5	1.1	12	--	--	--	--	1.6	9	--	--
276	3.0	9	--	--	3.0	9	--	--	--	--	--	--	--	--
277	1.0	1	--	--	1.0	1	--	--	--	--	--	--	--	--
279	1.7	12	1.5	4	1.8	4	--	--	--	--	1.8	4	--	--
296	2.1	33	2.2	26	1.9	7	--	--	--	--	--	--	--	--
304	3.6	20	3.5	19	--	--	--	--	--	--	--	--	4.0	1
307	2.5	15	1.1	7	3.6	5	--	--	4.0	3	--	--	NR	0
313	3.2	10	--	--	--	--	3.2	5	3.2	5	--	--	--	--
316	3.6	5	--	--	--	--	3.6	5	--	--	--	--	--	--
318	3.6	5	--	--	--	--	3.6	5	--	--	--	--	--	--
319	2.0	2	--	--	2.0	2	--	--	--	--	--	--	--	--
320	3.0	8	--	--	--	--	2.8	5	3.3	3	--	--	--	--
326	2.9	34	3.0	16	3.7	10	--	--	--	--	1.9	8	--	--
328	2.5	59	2.6	25	2.6	15	1.6	5	2.8	5	2.6	9	--	--
330	3.0	25	3.0	20	2.8	5	--	--	--	--	--	--	--	--
333	2.9	9	--	--	3.0	3	--	--	--	--	2.8	6	--	--
341	2.8	42	2.5	19	3.1	13	3.2	5	3.0	5	--	--	--	--
349	2.6	35	2.8	24	2.6	7	--	--	1.3	4	--	--	--	--
356	3.2	34	3.3	23	3.5	6	--	--	3.0	4	--	--	0.0	1
366	2.2	21	--	--	2.6	11	1.0	5	2.4	5	--	--	--	--
369	3.6	5	--	--	--	--	3.6	5	--	--	--	--	--	--
373	2.6	10	--	--	--	--	3.0	5	2.2	5	--	--	--	--
374	3.0	1	--	--	3.0	1	--	--	--	--	--	--	--	--
378	2.8	10	--	--	--	--	2.6	5	3.0	5	--	--	--	--
379	2.3	43	1.5	12	2.3	12	2.4	5	3.0	5	2.9	9	NR	0
380	2.7	10	--	--	--	--	2.8	5	2.6	5	--	--	--	--
381	1.7	9	1.0	5	--	--	2.5	4	--	--	--	--	--	--
383	2.0	3	--	--	--	--	--	--	2.0	3	--	--	--	--
384	1.7	29	1.7	9	1.8	9	0.5	2	0.0	2	2.4	7	--	--
385	2.3	4	--	--	2.3	4	--	--	--	--	--	--	--	--
386	3.1	20	3.0	7	3.1	13	--	--	--	--	--	--	--	--
387	0.0	2	--	--	0.0	2	--	--	--	--	--	--	--	--

**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents)**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Silver		Aluminum		Arsenic		Boron		Barium	
	OLR	V/28	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
	MPV =	F-pseudosigma =										
			2.44 µg/L		19.4 µg/L		3.50 µg/L		23.5 µg/L		13.6 µg/L	
			0.119 (0.122)		1.63		0.274		3.04		0.46 (0.68)	
1	3.6	28	2.35	3	18.923	4	3.435	4	23.16	4	13.563	4
5	2.2	19	<4.00	NR	<30.0	NR	3.73	3	25.9	3	11.9	0
7	2.7	24	--	--	48	0	3.4	4	<30	NR	13.5	4
8	3.0	27	2.5	4	19.8	4	3.6	4	23.9	4	14	3
10	3.0	7	--	--	--	--	3.5	4	--	--	--	--
12	1.8	4	1.8	0	--	--	4	1	--	--	--	--
16	2.0	26	2.7	0	17	2	3.2	2	15.5	0	13	3
21	4.0	1	--	--	--	--	--	--	--	--	--	--
23	2.9	9	2.26	2	--	--	<10	NR	--	--	--	--
24	2.8	10	--	--	--	--	--	--	24	4	14	3
25	0.9	21	<17	NR	36	0	3.7	3	20.2	2	10.7	0
26	2.3	6	2.13	0	--	--	--	--	--	--	--	--
32	3.1	28	2.5	4	19.4	4	3.38	4	24.2	4	13.7	4
33	2.4	7	--	--	<50	NR	--	--	--	--	14.49	2
42	2.5	24	<1.0	0	<30.0	NR	3.52	4	<30.0	NR	13.8	4
45	3.3	26	2.93	0	18.9	4	3.69	3	25.4	3	13.1	3
46	2.6	8	--	--	<100	NR	3.92	1	--	--	14.3	2
59	3.2	20	2.45	4	--	--	3.74	3	--	--	14	3
64	4.0	5	--	--	--	--	--	--	--	--	--	--
70	1.3	19	<10	NR	24	0	4.2	0	<100	NR	14.3	2
76	3.8	12	--	--	--	--	3.366	4	--	--	<20.0	NR
86	3.4	20	1.38	0	20.8	3	--	--	14	0	13.4	4
97	2.4	22	2.36	3	19.2	4	3.42	4	--	--	14.4	2
100	2.6	20	2.4	4	18.8	4	3.04	1	<40.0	NR	13.9	4
105	3.3	17	2.4	4	20.9	3	<4.0	NR	<200	NR	13	3
109	3.0	6	--	--	--	--	--	--	--	--	--	--
110	4.0	4	--	--	18.63	4	--	--	--	--	--	--
113	3.4	16	2.34	3	18.8	4	--	--	--	--	13.3	4
121	2.8	9	--	--	--	--	--	--	--	--	15	0
134	3.6	27	2.4	4	21	3	3.35	3	24	4	13.8	4
138	3.3	24	2.42	4	20.5	3	3.33	3	18	1	13.8	4
142	3.5	15	2.16	0	--	--	3.5	4	--	--	13.4	4
146	2.3	15	2.16	0	73.6	0	5.99	0	--	--	13.6	4
147	3.6	8	--	--	--	--	3.36	3	--	--	13.8	4
149	3.2	24	2.36	3	19.3	4	3.28	3	--	--	13.36	4
151	2.5	14	<10	NR	19	4	3.5	4	0.026	0	14	3
180	3.5	21	2.35	3	18.7	4	3.26	3	--	--	13.8	4
190	2.9	16	2	0	17.7	2	3.08	1	--	--	--	--
193	2.7	12	2.65	1	18.9	4	<5.0	NR	--	--	13	3
212	1.7	23	2.44	4	31.4	0	<15	NR	22.9	4	16.3	0
219	2.3	26	--	--	19.3	4	3	1	20	2	12.9	2
220	2.0	14	<5	NR	<20	NR	4.3	0	25	4	14.4	2
227	2.9	7	--	--	23.7	0	--	--	--	--	--	--
230	3.3	24	--	--	21	3	3.6	4	--	--	13.9	4
235	3.2	25	2.48	4	17.2	2	3.65	3	21.4	3	13.6	4

**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte =		Silver		Aluminum		Arsenic		Boron		Barium		
MPV =		2.44 µg/L		19.4 µg/L		3.50 µg/L		23.5 µg/L		13.6 µg/L		
F-pseudosigma =		0.119 (0.122)		1.63		0.274		3.04		0.46 (0.68)		
Lab	OLR	V/28	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
246	2.6	14	--	--	<30	NR	<5	NR	--	--	12.6	2
247	2.8	9	<10	NR	<80	NR	<40	NR	<50	NR	<10	0
254	3.1	16	--	--	<50	NR	<10	NR	24.1	4	13.6	4
255	1.8	11	2.7	0	--	--	4	1	25.8	3	--	--
256	2.8	17	2.5	4	18.3	3	6.34	0	--	--	--	--
257	0.7	7	--	--	--	--	3.108	2	--	--	--	--
259	3.7	18	2.49	4	22.4	1	3.26	3	23.5	4	13.6	4
265	3.5	28	2.3	2	19.5	4	3.5	4	23	4	13.5	4
273	0.8	12	--	--	--	--	--	--	--	--	16.5	0
274	0.0	5	--	--	--	--	--	--	--	--	--	--
279	1.5	4	--	--	--	--	--	--	--	--	--	--
296	2.2	26	3.3	0	20.1	4	3.78	2	--	--	13.8	4
304	3.5	19	2.49	4	18.8	4	3.62	4	--	--	13.5	4
307	1.1	7	3.3	0	--	--	2.55	0	--	--	--	--
326	3.0	16	--	--	--	--	3.6	4	24.2	4	13.8	4
328	2.6	25	2.8	0	<40	NR	3.4	4	19	2	13.4	4
330	3.0	20	2.6	2	18.6	4	3.8	2	--	--	13.9	4
341	2.5	19	2.5	4	<50	NR	3.1	2	--	--	13	3
349	2.8	24	2.1	0	21	3	3.6	4	24.5	4	13.2	3
356	3.3	23	2.5	4	49	0	3.62	4	--	--	--	--
379	1.5	12	3.5	0	15.1	0	3.3	3	--	--	11.9	0
381	1.0	5	--	--	--	--	--	--	--	--	--	--
384	1.7	9	--	--	25.7	0	--	--	--	--	--	--
386	3.0	7	--	--	--	--	--	--	--	--	--	--

**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Beryllium		Calcium		Cadmium		Cobalt		Chromium	
	MPV =	F-pseudosigma =	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
			1.18 µg/L		6.75 mg/L		4.92 µg/L		4.71 µg/L		3.10 µg/L	
			0.107		0.263 (0.337)		0.226 (0.246)		0.259		0.274	
1			1.212	4	6.775	4	4.9	4	4.76	4	2.763	2
5			1.18	4	6.53	3	5.35	1	<3.00	0	4.78	0
7			1	1	6.58	4	5	4	4.6	4	<2	0
8			1.2	4	6.9	4	5	4	5.3	0	3.3	3
10			--	--	--	--	4.8	4	--	--	3.1	4
12			--	--	--	--	--	--	--	--	<20	NR
16			1.1	3	6.3	2	4.8	4	3.8	0	2.8	2
21			--	--	--	--	--	--	--	--	--	--
23			--	--	6.45	3	4.63	2	--	--	<4.00	NR
24			--	--	6.25	2	--	--	--	--	--	--
25			1.1	3	6.07	0	8.8	0	6.3	0	<15	NR
26			--	--	6.84	4	4.62	2	--	--	--	--
32			1.25	3	6.55	3	5.58	0	4.58	4	3.04	4
33			--	--	7.22	2	--	--	--	--	--	--
42			1.29	2	6.21	1	4.13	0	4.37	2	3	4
45			1.27	3	6.63	4	4.89	4	4.72	4	3.28	3
46			<2	NR	6.85	4	4.64	2	--	--	3.75	0
59			1.14	4	6.75	4	5	4	4.3	1	3.08	4
64			--	--	6.67	4	--	--	--	--	--	--
70			1.4	0	6.82	4	6	0	5.7	0	<5.0	NR
76			1.114	3	--	--	4.936	4	4.668	4	<5.0	NR
86			1.18	4	6.82	4	4.67	3	4.79	4	2.81	2
97			--	--	7.04	3	4.46	1	5.35	0	<2.1	0
100			1.38	1	7.02	3	5.08	3	<5.0	NR	3.14	4
105			1.22	4	6.61	4	5	4	<50	NR	<4	NR
109			--	--	6.57	3	--	--	--	--	--	--
110			--	--	6.82	4	--	--	--	--	--	--
113			1.15	4	6.72	4	4.76	3	--	--	3.18	4
121			--	--	6.85	4	5.1	3	--	--	--	--
134			1.2	4	6.92	3	5	4	4.8	4	3.3	3
138			1.18	4	6.98	3	5.23	2	4.73	4	2.86	3
142			1.25	3	--	--	4.88	4	4.7	4	2.74	2
146			<4.00	NR	6.85	4	4.99	4	4.96	3	2.74	2
147			--	--	--	--	4.88	4	--	--	--	--
149			1.09	3	6.4	2	4.87	4	4.67	4	3.24	3
151			1.2	4	--	--	5.3	1	--	--	2.8	2
180			1.12	3	6.75	4	4.9	4	4.45	3	3.08	4
190			--	--	7.15	2	4.88	4	--	--	3.27	3
193			1.21	4	6.94	3	4.97	4	5.1	2	<5	NR
212			0.73	0	6.43	3	4.06	0	4.73	4	3.46	2
219			1.1	3	7	3	4.8	4	4.4	2	2.9	3
220			1.1	3	6.67	4	5.5	0	<10	NR	<10	NR
227			--	--	6.59	4	4.93	4	--	--	--	--
230			1.3	2	7	3	5.1	3	4.8	4	2.7	2
235			1.05	2	7.08	3	5.24	2	4.71	4	2.96	3

**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Beryllium		Calcium		Cadmium		Cobalt		Chromium	
	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
	MPV = 1.18 µg/L		6.75 mg/L		4.92 µg/L		4.71 µg/L		3.10 µg/L	
	F-pseudosigma = 0.107		0.263 (0.337)		0.226 (0.246)		0.259		0.274	
246	1	1	6.74	4	4.5	1	4.3	1	<5	NR
247	<10	NR	6.54	3	<10	NR	<10	NR	<10	NR
254	--	--	6.86	4	4.6	2	4.5	3	<5	NR
255	--	--	--	--	5.6	0	--	--	3.3	3
256	1.24	3	--	--	4.73	3	4.41	2	3.18	4
257	--	--	--	--	3.29	0	--	--	--	--
259	--	--	6.65	4	4.87	4	4.73	4	3.26	3
265	1.2	4	6.9	4	4.9	4	4.5	3	2.7	2
273	1.4	0	6.64	4	9.3	0	--	--	--	--
274	--	--	4.82	0	--	--	--	--	--	--
279	--	--	7.44	0	--	--	--	--	--	--
296	1.16	4	6.7	4	5.41	0	5.01	2	3.44	2
304	1.1	3	--	--	5.05	3	4.6	4	3.2	4
307	--	--	--	--	4.56	2	--	--	4	0
326	--	--	6.76	4	4.8	4	5.3	0	10.5	0
328	1.1	3	6.2	1	5.1	3	<8	NR	3.4	2
330	1.2	4	6.7	4	5.2	2	--	--	3.1	4
341	1.1	3	6.1	1	4.8	4	<15	NR	3	4
349	1.3	2	7	3	4.8	4	4.4	2	3.1	4
356	1.16	4	6.8	4	5.11	3	4.65	4	2.99	4
379	1.3	2	--	--	5.5	0	4.9	3	3.7	0
381	--	--	4.13	0	--	--	--	--	--	--
384	--	--	6.9567	3	--	--	--	--	--	--
386	--	--	6.5	3	--	--	--	--	--	--

**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Copper		Iron		Potassium		Lithium		Magnesium	
	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
	MPV = 1.00 µg/L F-pseudosigma = 0.172		53.4 µg/L 4.82		2.80 mg/L 0.093 (0.140)		6.30 µg/L 0.599		2.78 mg/L 0.111 (0.139)	
1	0.968	4	54.21	4	2.85	4	6.033	4	2.78	4
5	<4.00	NR	48.9	3	3.12	0	8.1	0	2.68	3
7	0.96	4	45.5	1	2.69	3	<20	NR	2.71	4
8	1.2	2	<100	NR	2.9	3	6.9	3	2.9	3
10	<2	NR	57	3	--	--	--	--	--	--
12	1.1	3	<50	NR	--	--	--	--	--	--
16	0.6	0	51.6	4	2.9	3	--	--	2.7	3
21	--	--	55.2	4	--	--	--	--	--	--
23	<5.00	NR	--	--	2.84	4	--	--	2.72	4
24	--	--	46.4	2	2.9	3	--	--	2.61	2
25	13.4	0	61.7	1	2.06	0	25.8	0	2.769	4
26	--	--	--	--	2.52	1	--	--	2.81	4
32	1.05	4	48	2	2.84	4	6.75	3	2.89	3
33	--	--	<100	NR	2.79	4	--	--	3.092	0
42	<2.0	NR	56.2	3	2.66	3	5.58	2	2.65	3
45	0.971	4	62.2	1	2.77	4	--	--	2.78	4
46	<3	NR	<300	NR	<3	NR	--	--	2.78	4
59	0.97	4	--	--	2.77	4	--	--	2.65	3
64	--	--	--	--	2.81	4	--	--	2.71	4
70	<5.0	NR	42.9	0	2.84	4	--	--	2.87	3
76	<20.0	NR	--	--	2.782	4	6.115	4	2.797	4
86	--	--	54.6	4	2.83	4	6.18	4	2.73	4
97	--	--	42.5	0	2.81	4	--	--	2.74	4
100	<5.0	NR	54.9	4	2.87	4	<50	NR	2.93	2
105	<10	NR	59	2	2.84	4	<25	NR	2.78	4
109	--	--	53	4	2.71	3	--	--	2.77	4
110	--	--	--	--	--	--	--	--	2.8	4
113	--	--	54.3	4	2.8	4	--	--	2.75	4
121	--	--	--	--	--	--	--	--	2.82	4
134	1	4	53	4	2.75	4	6.5	4	2.8	4
138	0.949	4	47.1	2	2.8	4	--	--	2.86	3
142	<1	NR	--	--	--	--	--	--	--	--
146	<20.0	NR	57.2	3	2.81	4	--	--	2.88	3
147	0.96	4	--	--	--	--	--	--	--	--
149	1.41	0	52	4	2.79	4	--	--	2.82	4
151	<10	NR	54	4	--	--	<10	NR	--	--
180	0.886	3	55.9	3	2.97	2	--	--	2.77	4
190	<2.3	NR	50.8	3	2.68	3	--	--	2.89	3
193	<12.5	NR	<100	NR	2.82	4	--	--	2.62	2
212	<10	NR	57.7	3	2.95	2	5.65	2	2.65	3
219	0.85	3	50	3	2.9	3	5.5	2	2.9	3
220	1.2	2	40.2	0	2.37	0	<10	NR	2.75	4
227	<1.46	NR	54	4	--	--	--	--	2.76	4
230	<5	NR	50	3	2.9	3	6.3	4	2.8	4
235	1	4	48.8	3	--	--	--	--	2.89	3

**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Copper		Iron		Potassium		Lithium		Magnesium	
	MPV = 1.00 µg/L		53.4 µg/L		2.80 mg/L		6.30 µg/L		2.78 mg/L	
	F-pseudosigma = 0.172		4.82		0.093 (0.140)		0.599		0.111 (0.139)	
	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
246	<3	NR	55	4	2.74	4	6	4	2.68	3
247	<10	NR	<50	NR	2.67	3	<10	NR	2.77	4
254	<5	NR	53.4	4	2.78	4	6.3	4	2.92	3
255	1.2	2	57	3	--	--	--	--	--	--
256	<10.00	NR	55.05	4	--	--	6.7	3	--	--
257	--	--	--	--	--	--	--	--	--	--
259	--	--	52.3	4	2.73	4	--	--	2.76	4
265	1.95	0	53	4	2.8	4	6	4	2.85	4
273	--	--	47.5	2	2.47	0	9.8	0	2.92	3
274	--	--	--	--	2.5	0	--	--	3.86	0
279	--	--	--	--	3.08	1	--	--	2.95	2
296	1.26	1	44	1	2.8	4	8.5	0	2.7	3
304	1	4	--	--	--	--	--	--	--	--
307	<1.46	NR	--	--	--	--	--	--	--	--
326	--	--	58.1	3	2.96	2	--	--	2.85	4
328	1.2	2	56	3	2.57	1	6.05	4	2.56	1
330	1	4	--	--	2.78	4	--	--	2.78	4
341	1.5	0	54	4	2.8	4	<10	NR	3	1
349	1.1	3	67	0	2.8	4	--	--	2.8	4
356	0.98	4	53	4	2.8	4	--	--	2.85	4
379	<1.5	NR	48.1	2	--	--	--	--	--	--
381	--	--	--	--	2.73	4	--	--	2.99	1
384	--	--	53.3	4	3.01	2	--	--	3.05	1
386	--	--	56.5	3	2.95	2	--	--	2.7	3



**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Manganese		Molybdenum		Sodium		Nickel		Lead	
	MPV =		12.4 µg/L		1.97 µg/L		8.60 mg/L		5.67 µg/L		1.09 µg/L	
	F-pseudosigma =		0.44 (0.62)		0.133		0.282 (0.430)		0.256 (0.283)		0.148	
	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	12.4	4	2.938	0	8.49	4	5.796	4	1.024	4		
5	12.6	4	<10.0	NR	8.77	4	<10.0	NR	1.37	1		
7	12.4	4	<3	NR	8.3	3	5.7	4	0.95	3		
8	14	0	1.9	3	8.8	4	5.5	3	1.09	4		
10	11.6	2	--	--	--	--	--	--	<2	NR		
12	<20	NR	--	--	--	--	<20	NR	1	3		
16	12	3	2.5	0	8.1	2	4.3	0	1.2	3		
21	--	--	--	--	--	--	--	--	--	--		
23	12.13	4	--	--	7.7	0	5.63	4	<5.00	NR		
24	11.7	2	--	--	8.17	3	--	--	--	--		
25	11.2	1	--	--	7.151	0	8.2	0	5.5	0		
26	--	--	--	--	8.34	3	--	--	--	--		
32	12	3	1.97	4	9.15	2	5.4	3	1.16	4		
33	<100	NR	--	--	8.58	4	--	--	--	--		
42	11.8	3	<10	NR	8.33	3	5.25	2	1.1	4		
45	12.5	4	1.98	4	8.57	4	5.8	4	1.11	4		
46	12.2	4	--	--	8.73	4	<50	NR	<3	NR		
59	12.6	4	1.79	2	8.65	4	5.66	4	1.09	4		
64	--	--	--	--	8.61	4	--	--	--	--		
70	14.2	0	<5.0	NR	8.9	3	7.2	0	1.3	2		
76	--	--	<5.0	NR	8.594	4	5.577	4	1.058	4		
86	12.2	4	2.08	3	8.71	4	5.75	4	--	--		
97	11.3	1	1.94	4	8.86	3	5.8	4	1.27	2		
100	12.3	4	16.5	0	8.83	3	<20	NR	1.23	3		
105	13	2	<4	NR	8.76	4	<50	NR	1.1	4		
109	14.99	0	--	--	8.53	4	--	--	--	--		
110	--	--	--	--	--	--	--	--	--	--		
113	12.4	4	2.54	0	8.45	4	5.77	4	--	--		
121	13	2	--	--	8.55	4	--	--	--	--		
134	12.9	3	1.95	4	8.42	4	6	2	1	3		
138	12.4	4	<2.0	NR	8.96	3	5.54	4	0.937	3		
142	--	--	1.94	4	--	--	5.67	4	<1	NR		
146	12.2	4	<10.0	NR	8.95	3	4.9	0	<5.00	NR		
147	--	--	--	--	--	--	--	--	1.05	4		
149	12.47	4	1.84	3	8.6	4	5.47	3	1.03	4		
151	12	3	<2	NR	--	--	6	2	<1	NR		
180	12.1	4	2.01	4	8.51	4	5.38	3	1	3		
190	12.2	4	--	--	8.77	4	5.44	3	1.08	4		
193	<50	NR	<5	NR	9.26	1	<12.5	NR	<5	NR		
212	13.4	1	3.31	0	8.44	4	6.86	0	2.33	0		
219	12	3	<2	NR	8.7	4	5.4	3	0.85	1		
220	11.44	2	<10	NR	8.23	3	<10	NR	10.6	0		
227	--	--	--	--	--	--	--	--	1.65	0		
230	13	2	2	4	8.8	4	5.7	4	1.05	4		
235	12.5	4	1.98	4	8.8	4	5.68	4	1.08	4		

**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Manganese		Molybdenum		Sodium		Nickel		Lead	
	MPV =		12.4 µg/L		1.97 µg/L		8.60 mg/L		5.67 µg/L		1.09 µg/L	
	F-pseudosigma =		0.44 (0.62)		0.133		0.282 (0.430)		0.256 (0.283)		0.148	
	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
246	11.3	1	<3	NR	8.1	2	5.6	4	<10	NR		
247	10	0	1.97	4	8.52	4	<50	NR	<40	NR		
254	11.8	3	--	--	8.42	4	6.4	0	<15	NR		
255	12.3	4	--	--	--	--	7.5	0	1.2	3		
256	12.5	4	1.8	2	--	--	5.51	3	<3.00	NR		
257	14.682	0	--	--	--	--	5.403	3	2.5	0		
259	12.5	4	--	--	8.65	4	5.9	3	--	--		
265	12.3	4	1.8	2	8.5	4	5.4	3	1	3		
273	16.5	0	--	--	12.4	0	--	--	--	--		
274	--	--	--	--	6.15	0	--	--	--	--		
279	--	--	--	--	8.89	3	--	--	--	--		
296	5	0	1.94	4	8.9	3	5.72	4	1.07	4		
304	12.3	4	1.89	3	--	--	5.65	4	1.06	4		
307	--	--	--	--	--	--	4.93	0	0.94	3		
326	12.6	4	--	--	8.46	4	5.88	3	0.6	0		
328	13	2	1.9	3	7.88	1	5.5	3	1	3		
330	12.9	3	1.7	0	8.78	4	6	2	1.1	4		
341	12	3	<25	NR	8.6	4	5.7	4	1	3		
349	<20	NR	2.2	1	8.03	2	5.7	4	1.2	3		
356	12	3	--	--	9.03	3	5.65	4	1.11	4		
379	12.6	4	--	--	--	--	5.1	1	--	--		
381	14.6	0	--	--	14.62	0	--	--	--	--		
384	10.9333	0	--	--	7.91	1	--	--	--	--		
386	12.5	4	--	--	8.67	4	--	--	--	--		

**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Antimony		Selenium		Silica		Strontium		Thallium	
	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
	MPV = 2.73 µg/L		1.39 µg/L		3.50 mg/L		84.0 µg/L		8.35 µg/L	
	F-pseudosigma = 0.234		0.182		0.163 (0.175)		4.04 (4.20)		0.519	
1	3.142	1	1.385	4	3.515	4	82.937	4	8.304	4
5	<20.0	NR	2.24	0	3.53	4	83	4	--	--
7	2.6	3	<1	0	3.38	3	83.5	4	8.1	4
8	2.79	4	1.37	4	3.7	2	78	2	8.5	4
10	--	--	1.2	2	--	--	--	--	--	--
12	--	--	<2	NR	--	--	--	--	--	--
16	2.5	3	0.6	0	--	--	80	3	8	3
21	--	--	--	--	--	--	--	--	--	--
23	--	--	--	--	--	--	--	--	--	--
24	--	--	--	--	3.64	3	84.2	4	--	--
25	<50	NR	<16	NR	2.895	0	88.7	2	<7.10	0
26	--	--	--	--	--	--	--	--	--	--
32	2.75	4	1.29	3	3.58	4	83	4	8.68	3
33	--	--	--	--	3.72	2	81.11	3	--	--
42	2.65	4	1.8	0	3.14	0	87.8	3	8.13	4
45	2.7	4	1.46	4	1.66	0	--	--	8.11	4
46	--	--	--	--	--	--	--	--	--	--
59	3.13	1	<5	NR	--	--	80.8	3	9.13	2
64	--	--	--	--	3.5	4	--	--	--	--
70	5.3	0	<5.0	NR	3.27	2	--	--	8.3	4
76	--	--	<2.0	NR	--	--	--	--	8.304	4
86	--	--	--	--	--	--	82.3	4	--	--
97	2.91	3	1.08	1	3.64	3	86.2	3	8.7	3
100	3.22	0	1.4	4	4.05	0	88.3	2	9.29	1
105	2.9	3	<7	NR	3.445	4	87	3	8.6	4
109	--	--	--	--	--	--	--	--	--	--
110	--	--	--	--	3.49	4	--	--	--	--
113	--	--	--	--	--	--	82.8	4	7.06	0
121	--	--	--	--	3.53	4	83	4	--	--
134	2.65	4	1.15	2	3.41	3	84	4	8.35	4
138	2.73	4	1.4	4	--	--	85.8	4	8.24	4
142	2.7	4	1.42	4	--	--	--	--	8.11	4
146	<20.0	NR	<10.0	NR	--	--	--	--	15.7	0
147	--	--	--	--	--	--	81.5	3	--	--
149	2.52	3	1.21	3	--	--	--	--	6.25	0
151	3	2	<5	NR	--	--	81	3	7.8	2
180	2.56	3	1.32	4	--	--	--	--	--	--
190	--	--	1.19	2	3.5	4	84.8	4	--	--
193	<12.5	NR	<5	NR	--	--	94.3	0	8.38	4
212	<10	NR	<15	NR	3.44	4	85.6	4	14.2	0
219	2.6	3	1.1	1	3.7	2	91	1	7.1	0
220	<10	NR	1.4	4	--	--	--	--	<30	NR
227	--	--	--	--	--	--	--	--	--	--
230	--	--	1.4	4	3.6	3	80	3	8.4	4
235	2.82	4	1.23	3	3.76	2	83.6	4	9.81	0

**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Antimony		Selenium		Silica		Strontium		Thallium	
	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
	MPV = 2.73 µg/L		1.39 µg/L		3.50 mg/L		84.0 µg/L		8.35 µg/L	
	F-pseudosigma = 0.234		0.182		0.163 (0.175)		4.04 (4.20)		0.519	
246	--	--	--	--	--	--	79.8	3	--	--
247	2.75	4	<100	NR	--	--	80	3	<50	NR
254	--	--	--	--	3.33	3	88.4	2	<50	NR
255	--	--	1.7	1	--	--	--	--	--	--
256	4.17	0	1.29	3	3.4	3	90	2	--	--
257	1.434	0	2.093	0	--	--	--	--	--	--
259	--	--	--	--	3.45	4	84.5	4	--	--
265	2.5	3	1.3	4	3.5	4	85	4	8	3
273	--	--	--	--	3.2	1	95.5	0	--	--
274	--	--	--	--	1.38	0	--	--	--	--
279	--	--	--	--	--	--	--	--	--	--
296	2.91	3	1.74	1	--	--	113	0	8.8	3
304	2.73	4	1.62	2	--	--	84.1	4	9.75	0
307	--	--	1.29	3	--	--	--	--	--	--
326	--	--	--	--	--	--	83.9	4	--	--
328	2.7	4	1.5	3	3.56	4	82	4	8	3
330	3	2	1.7	1	--	--	--	--	8.7	3
341	2.3	1	<1.0	0	--	--	--	--	7.9	3
349	2.6	3	1.3	4	--	--	--	--	8.8	3
356	2.92	3	1.51	3	3.6	3	97.2	0	8.86	3
379	--	--	--	--	--	--	--	--	--	--
381	--	--	--	--	--	--	--	--	--	--
384	--	--	--	--	1.71	0	81.9	4	--	--
386	--	--	--	--	3.74	2	--	--	--	--

**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Uranium		Vanadium		Zinc	
	RV	Rating	RV	Rating	RV	Rating
	Analyte =		Analyte =		Analyte =	
	MPV = 1.23 µg/L		MPV = 2.39 µg/L		MPV = 6.21 µg/L	
	F-pseudosigma = 0.096		F-pseudosigma = 0.148		F-pseudosigma = 1.108	
1	1.23	4	2.281	3	6.628	4
5	--	--	<4.00	NR	6.64	4
7	1.2	4	0.99	0	3.7	0
8	1.3	3	2.6	2	7.5	2
10	--	--	--	--	5	2
12	--	--	--	--	<20	NR
16	1.2	4	2.3	3	1.5	0
21	--	--	--	--	--	--
23	--	--	--	--	5.17	3
24	--	--	--	--	--	--
25	--	--	<19	NR	7.6	2
26	--	--	--	--	--	--
32	1.45	0	2.24	2	7.53	2
33	--	--	--	--	--	--
42	1.33	2	2.33	4	6.65	4
45	1.23	4	2.48	3	6.04	4
46	--	--	--	--	<50	NR
59	--	--	2.21	2	<10	NR
64	--	--	--	--	--	--
70	1.5	0	<5.0	NR	7.9	1
76	<5.0	NR	2.313	3	--	--
86	--	--	2.39	4	6.11	4
97	--	--	<2.3	NR	<3.2	0
100	--	--	<5.0	NR	<5.0	NR
105	--	--	<20	NR	11	0
109	--	--	--	--	--	--
110	--	--	--	--	--	--
113	--	--	--	--	--	--
121	--	--	5.6	0	--	--
134	--	--	2.5	3	6.5	4
138	--	--	2.62	1	6.05	4
142	1.19	4	2.26	3	5.96	4
146	--	--	<10.0	NR	<20.0	NR
147	1.28	3	--	--	6.1	4
149	1.29	3	2.43	4	6.1	4
151	--	--	2.1	1	<20	NR
180	--	--	--	--	5.95	4
190	--	--	--	--	<12	NR
193	--	--	<25	NR	<25	NR
212	<60	NR	3.16	0	13.7	0
219	1	0	2.7	0	5.5	3
220	--	--	<10	NR	<20	NR
227	--	--	--	--	5.84	4
230	1.19	4	2.5	3	8	1
235	--	--	2.45	4	6.99	3

**Table 5. Laboratory performance ratings for standard reference sample T-171 (trace constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; µg/L, micrograms per liter; mg/L, milligrams per liter; V/28, number of rated analyses out of 28 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte = Uranium		Vanadium		Zinc	
	RV	Rating	RV	Rating	RV	Rating
	MPV = 1.23 µg/L		2.39 µg/L		6.21 µg/L	
	F-pseudostigma = 0.096		0.148		1.108	
246	--	--	<5	NR	7.4	2
247	--	--	<10	NR	<40	NR
254	1.33	2	<10	NR	5.2	3
255	--	--	--	--	<9.0	NR
256	--	--	2.34	4	<10.00	NR
257	--	--	--	--	--	--
259	--	--	--	--	5.9	4
265	1.2	4	2.4	4	6	4
273	--	--	--	--	--	--
274	--	--	--	--	--	--
279	--	--	--	--	--	--
296	1.35	2	3.85	0	7.71	2
304	--	--	2.37	4	6.59	4
307	--	--	--	--	--	--
326	--	--	--	--	6.2	4
328	0.7	0	<8	NR	6	4
330	--	--	--	--	6.8	3
341	--	--	<2.0	0	<10	NR
349	1.2	4	2.3	3	8.5	0
356	--	--	2.38	4	6.22	4
379	--	--	2.5	3	--	--
381	--	--	--	--	--	--
384	--	--	--	--	--	--
386	--	--	--	--	--	--

**Table 6. Laboratory performance ratings for standard reference sample M-164 (major constituents)**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter; µg/L, micrograms per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius; V/16, number of rated analyses out of 16 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Alkalinity		Boron		Calcium		Chloride		Fluoride	
	OLR	V/16	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
	MPV =	F-pseudosigma =	11.4 mg/L	0.96	48.2 µg/L	3.63	12.7 mg/L	0.55 (0.64)	23.2 mg/L	0.89 (1.16)	0.604 mg/L	0.0497
1	2.9	15	11.9	3	48.06	4	13.08	3	22.4	3	0.635	3
4	1.3	3	20.3	0	--	--	--	--	23.6	4	--	--
5	2.6	16	11.2	4	44.6	3	12.2	3	22.5	3	0.56	3
8	3.0	16	10.2	2	48.3	4	12.5	4	22.6	3	0.68	1
10	3.5	11	11.5	4	--	--	12.8	4	22.8	4	0.47	0
12	2.1	7	11	4	--	--	--	--	22	2	--	--
16	2.9	15	11.4	4	31	0	12	2	18.6	0	0.63	3
23	3.2	13	10	2	--	--	12.4	4	23.4	4	0.583	4
24	3.0	13	11.4	4	48	4	12.1	3	23.2	4	0.593	4
25	1.9	15	11	4	42.1	1	10.82	0	21.3	1	0.58	4
26	2.8	11	11.5	4	--	--	13.06	3	23.56	4	0.484	0
32	2.9	16	11.7	4	47	4	13.5	2	25.2	1	0.57	3
33	2.2	12	10.04	2	--	--	13.8	1	23.8	3	0.563	3
38	3.6	10	11.61	4	--	--	12.56	4	--	--	--	--
42	2.3	15	14.9	0	35.8	0	11.8	2	23.2	4	0.604	4
45	3.1	15	12	3	50	4	12.5	4	23.2	4	0.535	2
46	3.8	11	11.3	4	--	--	12.5	4	23.7	4	0.591	4
59	2.8	13	9.34	0	--	--	13.4	2	22.4	3	0.52	1
64	3.1	10	--	--	--	--	12.94	4	23.9	3	--	--
70	3.1	12	11	4	<100	NR	12.7	4	24	3	0.61	4
76	3.9	10	--	--	48.6	4	--	--	23.2	4	--	--
85	2.9	11	14.5	0	49	4	--	--	23.9	3	0.62	4
86	3.5	13	--	--	47.7	4	12.6	4	23.8	3	0.637	3
91	2.2	5	9.5	1	--	--	--	--	23	4	--	--
97	2.9	14	12	3	--	--	13.1	3	23.5	4	0.656	2
100	2.7	15	12.6	2	<40	0	13.4	2	23.4	4	0.634	3
102	2.1	10	--	--	--	--	14	0	23.6	4	0.56	3
105	3.1	14	11.6	4	<200	NR	13.1	3	23.5	4	0.69	1
109	2.3	11	13.52	0	--	--	12.38	4	26.55	0	0.63	3
113	3.4	13	7.84	0	--	--	12.7	4	23.8	3	0.63	3
118	2.6	5	17	0	--	--	--	--	--	--	--	--
121	1.2	6	--	--	--	--	12.1	3	--	--	--	--
134	3.6	16	12.9	1	48.2	4	13.31	3	23.4	4	0.61	4
138	3.4	16	9.95	2	41.3	1	13.1	3	23.3	4	0.561	3
142	2.8	9	11	4	--	--	--	--	23.2	4	0.604	4
146	2.3	13	8.08	0	--	--	12.7	4	22.5	3	0.513	1
149	2.7	7	11	4	--	--	--	--	24	3	0.6	4
151	3.8	12	11	4	--	--	12.8	4	23	4	0.628	4
180	3.3	11	11.5	4	--	--	12.8	4	25.2	1	0.556	3
183	2.6	7	10	2	--	--	--	--	26.2	0	0.634	3
190	3.6	14	11.3	4	--	--	12.9	4	23.5	4	0.6	4
193	2.6	8	9.1	0	--	--	13.2	3	--	--	--	--
208	3.5	2	--	--	--	--	--	--	22.77	4	--	--
212	2.3	16	9.48	1	46.2	3	11.9	2	22.9	4	0.628	4
219	1.3	12	--	--	56	0	15.7	0	21	1	0.6	4

**Table 6. Laboratory performance ratings for standard reference sample M-164 (major constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter; µg/L, micrograms per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius; V/16, number of rated analyses out of 16 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Alkalinity		Boron		Calcium		Chloride		Fluoride	
	OLR	V/16	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
	MPV =	F-pseudosigma =	11.4 mg/L	0.96	48.2 µg/L	3.63	12.7 mg/L	0.55 (0.64)	23.2 mg/L	0.89 (1.16)	0.604 mg/L	0.0497
220	2.4	9	13.35	0	51.2	3	12.39	4	24.88	2	0.61	4
224	2.5	13	10	2	--	--	12.22	3	22.41	3	0.78	0
227	2.3	7	12.8	2	--	--	12.8	4	22.1	3	--	--
230	3.2	13	12.5	2	--	--	13.2	3	23	4	0.64	3
247	2.9	15	11	4	70	0	12.5	4	23.4	4	0.55	2
254	3.6	9	--	--	48.3	4	12.9	4	23	4	--	--
255	3.3	4	--	--	--	--	12.8	4	--	--	0.52	1
256	3.3	12	12.06	3	--	--	12.49	4	22.78	4	0.51	1
257	2.0	10	--	--	--	--	12	2	27	0	0.55	2
259	3.7	13	10.7	3	50.4	3	12.6	4	22.7	4	0.63	3
265	3.1	11	--	--	46	3	12.6	4	22.1	3	0.63	3
266	3.1	12	12	3	--	--	13.4	2	23.3	4	0.58	4
273	1.5	15	11.2	4	144	0	13.9	1	22.94	4	3.19	0
274	1.1	12	10.92	4	--	--	10.44	0	17.96	0	0.85	0
276	3.0	9	11.72	4	--	--	12.7	4	28	0	--	--
277	1.0	1	--	--	--	--	--	--	--	--	--	--
279	1.8	4	--	--	--	--	12.33	3	--	--	--	--
296	1.9	7	--	--	58	0	13.5	2	--	--	--	--
307	3.6	5	--	--	--	--	--	--	22	2	--	--
319	2.0	2	--	--	54	1	--	--	24	3	--	--
326	3.7	10	--	--	45.8	3	12.64	4	22.45	3	--	--
328	2.6	15	12	3	40	0	12.2	3	22	2	0.59	4
330	2.8	5	10.7	3	--	--	--	--	22.6	3	0.63	3
333	3.0	3	9.81	1	--	--	--	--	--	--	--	--
341	3.1	13	<20	NR	52	2	12.2	3	23	4	--	--
349	2.6	7	12	3	--	--	--	--	25	1	0.59	4
356	3.5	6	9.7	1	--	--	--	--	22.8	4	0.62	4
366	2.6	11	12.3	3	--	--	12.3	3	23.3	4	--	--
374	3.0	1	--	--	--	--	--	--	24	3	--	--
379	2.3	12	11.4	4	--	--	12.2	3	25	1	0.69	1
384	1.8	9	--	--	--	--	13.2667	3	20.4524	0	0.566	3
385	2.3	4	--	--	--	--	13.7	1	--	--	--	--
386	3.1	13	12	3	--	--	12.4	4	22.9	4	0.605	4
387	0.0	2	--	--	--	--	--	--	4.5	0	--	--



**Table 6. Laboratory performance ratings for standard reference sample M-164 (major constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter; µg/L, micrograms per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius; V/16, number of rated analyses out of 16 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Potassium		Magnesium		Sodium		pH		Residue on Evaporation	
	MPV =		3.03 mg/L		2.78 mg/L		37.0 mg/L		7.06		198 mg/L	
	F-pseudosigma =		0.126 (0.152)		0.104 (0.139)		1.67 (1.85)		0.297 (0.353)		16.3	
	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	2.96	4	2.817	4	37.31	4	6.17	0	198	4		
4	--	--	--	--	--	--	--	--	--	--		
5	3.26	1	2.6	2	37.5	4	6.57	2	171	1		
8	3	4	2.7	3	36.7	4	7.45	2	203	4		
10	3	4	2.8	4	37.2	4	--	--	188	3		
12	--	--	--	--	--	--	6.8	3	162	0		
16	3.1	4	2.7	3	36	3	7.1	4	203	4		
23	3.07	4	2.96	2	6.18	0	7.3	3	188	3		
24	3	4	2.5	0	36.3	4	6.6	2	--	--		
25	2.402	0	2.806	4	32.01	0	6.78	3	186	3		
26	2.91	3	2.8	4	37.04	4	5.4	0	182	3		
32	3.2	2	2.82	4	40.5	1	6.94	4	220	2		
33	3.14	3	3.11	0	34.03	1	7.13	4	--	--		
38	3.12	3	2.966	2	36.47	4	7.1	4	192	4		
42	2.92	3	2.66	3	35.2	3	6.66	2	--	--		
45	2.93	3	2.71	4	35.7	3	7.1	4	210	3		
46	2.9	3	2.7	3	37.6	4	7	4	204	4		
59	2.98	4	2.5	0	37.4	4	7.2	4	204	4		
64	3.03	4	2.69	3	38.2	3	7.28	3	--	--		
70	3.03	4	2.83	4	38.3	3	--	--	2200	0		
76	2.973	4	2.734	4	37.06	4	--	--	--	--		
85	--	--	--	--	--	--	6.56	2	186	3		
86	3.11	3	2.78	4	38.1	3	6.83	3	--	--		
91	--	--	--	--	--	--	6.9	4	--	--		
97	3.03	4	2.73	4	38.1	3	7.13	4	198	4		
100	3.06	4	2.89	3	37.8	4	7.22	4	195	4		
102	2.5	0	3.1	0	30	0	--	--	--	--		
105	3.01	4	2.81	4	38.3	3	6.5	1	186	3		
109	3.31	1	2.8	4	36.99	4	6.7	3	212	3		
113	3.03	4	2.73	4	36.9	4	--	--	193	4		
118	--	--	--	--	--	--	6.8	3	220	2		
121	--	--	2.6	2	33.6	1	--	--	--	--		
134	2.96	4	2.78	4	36.53	4	7.232	4	197.5	4		
138	3	4	2.81	4	38.5	3	7.23	4	195	4		
142	--	--	--	--	--	--	6.27	0	220	2		
146	3.12	3	2.78	4	38.3	3	7.13	4	210	3		
149	--	--	--	--	--	--	7.2	4	185	3		
151	3.1	4	2.8	4	37	4	7.05	4	191	4		
180	3.07	4	2.72	4	36.7	4	7.07	4	--	--		
183	--	--	--	--	--	--	--	--	212	3		
190	2.96	4	2.85	4	37.2	4	6.86	3	203	4		
193	3.18	3	2.58	2	35.9	3	7.06	4	--	--		
208	--	--	--	--	--	--	--	--	--	--		
212	2.84	2	2.55	1	35.6	3	7.02	4	167	1		
219	3.7	0	3.4	0	43.6	0	--	--	--	--		

**Table 6. Laboratory performance ratings for standard reference sample M-164 (major constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter; µg/L, micrograms per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius; V/16, number of rated analyses out of 16 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Potassium		Magnesium		Sodium		pH		Residue on Evaporation	
	MPV =		3.03 mg/L		2.78 mg/L		37.0 mg/L		7.06		198 mg/L	
	F-pseudsigma =		0.126 (0.152)		0.104 (0.139)		1.67 (1.85)		0.297 (0.353)		16.3	
	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
220	2.75	1	2.71	4	36.16	4	--	--	--	--	--	--
224	2.89	3	2.65	3	34.42	2	6.72	3	200	4		
227	--	--	2.72	4	--	--	--	--	214	3		
230	3.1	4	2.7	3	39	2	6.7	3	--	--		
247	2.89	3	2.77	4	37	4	7.25	3	180	2		
254	3	4	2.86	3	37.3	4	--	--	--	--		
255	--	--	2.78	4	--	--	--	--	--	--		
256	3.07	4	--	--	34.7	2	6.95	4	--	--		
257	2.9	3	--	--	34.5	2	7.06	4	180	2		
259	3.06	4	2.73	4	37.6	4	--	--	--	--		
265	2.9	3	2.7	3	36.5	4	--	--	--	--		
266	3.26	1	2.4	0	37.5	4	7.16	4	196	4		
273	3.29	1	30.1	0	40.6	1	7.03	4	157	0		
274	2.8	1	3.38	0	31.97	0	7.05	4	--	--		
276	2.95	3	2.95	2	38.3	3	7.24	3	198	4		
277	--	--	--	--	--	--	--	--	226.7	1		
279	3.62	0	3.01	1	35.5	3	--	--	--	--		
296	3.1	4	2.8	4	38.8	3	--	--	--	--		
307	--	--	--	--	--	--	7.08	4	--	--		
319	--	--	--	--	--	--	--	--	--	--		
326	3.06	4	2.75	4	36.61	4	7.17	4	--	--		
328	2.74	1	2.62	2	35	2	6.87	3	214	3		
330	--	--	--	--	--	--	7.31	3	--	--		
333	--	--	--	--	--	--	7.2	4	--	--		
341	3	4	3	1	38	3	7.06	4	200	4		
349	--	--	--	--	--	--	6.71	3	190	4		
356	--	--	--	--	--	--	--	--	201	4		
366	2.96	4	2.6	2	34.4	2	7.33	3	210	3		
374	--	--	--	--	--	--	--	--	--	--		
379	3.6	0	2.7	3	39.5	2	7.2	4	204	4		
384	3.37	0	2.9867	2	36.02	3	--	--	--	--		
385	3.2	2	2.9	3	38.5	3	--	--	--	--		
386	3.25	2	2.7	3	37.1	4	6.87	3	232	0		
387	--	--	--	--	--	--	--	--	--	--		

**Table 6. Laboratory performance ratings for standard reference sample M-164 (major constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter; µg/L, micrograms per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius; V/16, number of rated analyses out of 16 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Silica		Sulfate		Specific Conductance		Strontium		Total Phosphorus as P	
	MPV =	F-pseudsigma =	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
			8.10 mg/L		11.2 mg/L		301 µS/cm		89.6 µg/L		0.249 mg/L	
			0.371 (0.405)		0.55 (0.56)		6.7 (15.0)		4.67		0.0215	
1			11.9	0	10.8	3	282	2	87.04	3	--	--
4			--	--	9.8	0	--	--	--	--	--	--
5			7.85	3	9.99	0	302	4	86.5	3	0.255	4
8			7.8	3	11.1	4	300	4	87	3	0.2	0
10			7.9	4	11.5	3	300	4	--	--	--	--
12			--	--	<3	0	298	4	--	--	0.22	2
16			--	--	11.2	4	302	4	84	2	0.22	2
23			8.15	4	11.4	4	295	4	--	--	0.238	3
24			8.42	3	9.8	0	290	3	89.9	4	--	--
25			6.478	0	10	0	304	4	95.2	2	0.26	3
26			--	--	10.61	2	300	4	--	--	--	--
32			8.1	4	11.7	3	307	4	90.5	4	0.24	4
33			8.62	2	11.5	3	270.6	1	84.89	3	--	--
38			7.93	4	--	--	294.9	4	--	--	0.261	3
42			7.28	0	12.4	0	302	4	94.7	2	0.255	4
45			8.01	4	10.3	1	298	4	--	--	0.287	1
46			--	--	--	--	299	4	--	--	0.25	4
59			--	--	11	4	302	4	86.3	3	--	--
64			8.1	4	11.5	3	247	0	--	--	0.257	4
70			7.92	4	11.8	2	302	4	--	--	0.216	1
76			--	--	11.37	4	300	4	87.22	4	0.2463	4
85			8.6	2	11.9	2	295	4	88	4	0.25	4
86			--	--	11.5	3	308	4	87.5	4	0.259	4
91			--	--	13	0	285	2	--	--	--	--
97			12	0	7.58	0	299	4	91.4	4	0.22	2
100			9.45	0	11.4	4	300	4	93.1	3	--	--
102			7.8	3	11.4	4	313	3	--	--	0.252	4
105			8.153	4	11.5	3	293	4	95	2	0.242	4
109			--	--	12.75	0	310	3	--	--	--	--
113			7.76	3	11	4	305	4	86.8	3	0.248	4
118			7.92	4	--	--	302	4	--	--	--	--
121			7.1	0	--	--	--	--	81	1	--	--
134			7.79	3	10.75	3	297	4	89.94	4	0.259	4
138			8.29	4	11.2	4	290	3	92.2	3	0.253	4
142			--	--	10.9	3	305	4	--	--	0.228	3
146			--	--	10.4	2	280	2	--	--	0.282	1
149			--	--	10	0	325	1	--	--	--	--
151			7.85	3	10.4	2	296	4	--	--	--	--
180			--	--	11	4	324	1	--	--	0.261	3
183			--	--	11.5	3	313.5	3	--	--	0.242	4
190			8.2	4	11.2	4	298	4	<0.25	0	0.234	3
193			8.14	4	--	--	284	2	--	--	--	--
208			--	--	10.77	3	--	--	--	--	--	--
212			7.68	2	10.5	2	257	0	89.7	4	0.303	0
219			9.8	0	11	4	--	--	114	0	0.26	3

**Table 6. Laboratory performance ratings for standard reference sample M-164 (major constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter; µg/L, micrograms per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius; V/16, number of rated analyses out of 16 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Silica		Sulfate		Specific Conductance		Strontium		Total Phosphorus as P	
	MPV =		8.10 mg/L		11.2 mg/L		301 µS/cm		89.6 µg/L		0.249 mg/L	
	F-pseudosigma =		0.371 (0.405)		0.55 (0.56)		6.7 (15.0)		4.67		0.0215	
	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
220	--	--	12.64	0	--	--	--	--	--	--	--	--
224	8.96	0	11.35	4	309	3	--	--	--	--	0.23	3
227	--	--	--	--	845	0	--	--	--	--	0.115	0
230	8.3	4	11	4	299	4	85.2	3	--	--	--	--
247	9.06	0	11.2	4	301	4	90	4	0.21	1	--	--
254	7.76	3	11	4	--	--	94.5	2	--	--	--	--
255	--	--	11	4	--	--	--	--	--	--	--	--
256	8.21	4	11.17	4	302	4	93.3	3	--	--	--	--
257	--	--	11.5	3	317	2	--	--	0.32	0	--	--
259	8.1	4	11.4	4	301	4	89.5	4	0.231	3	--	--
265	7.4	1	12	2	--	--	88	4	--	--	--	--
266	7.95	4	11	4	311	3	--	--	--	--	--	--
273	9.31	0	11.8	2	306	4	96.4	2	0.3	0	--	--
274	3.16	0	9.27	0	293	4	--	--	0.0359	0	--	--
276	--	--	--	--	301	4	--	--	--	--	--	--
277	--	--	--	--	--	--	--	--	--	--	--	--
279	--	--	--	--	--	--	--	--	--	--	--	--
296	--	--	--	--	--	--	128	0	--	--	--	--
307	--	--	11.45	4	300	4	--	--	0.244	4	--	--
319	--	--	--	--	--	--	--	--	--	--	--	--
326	--	--	11.16	4	302	4	85.6	3	--	--	--	--
328	8.25	4	12	2	299	4	86	3	0.26	3	--	--
330	--	--	10.6	2	--	--	--	--	--	--	--	--
333	--	--	--	--	303	4	--	--	--	--	--	--
341	--	--	9	0	303	4	90	4	0.254	4	--	--
349	--	--	<10	0	312	3	--	--	--	--	--	--
356	--	--	11.1	4	303	4	--	--	--	--	--	--
366	--	--	12.2	1	275	1	--	--	0.232	3	--	--
374	--	--	--	--	--	--	--	--	--	--	--	--
379	--	--	11.7	3	278	2	--	--	0.21	1	--	--
384	3.9333	0	11.7708	2	--	--	87.1333	3	--	--	--	--
385	--	--	--	--	--	--	--	--	--	--	--	--
386	8.13	4	12.1	1	301	4	--	--	0.251	4	--	--
387	--	--	5.4	0	--	--	--	--	--	--	--	--

**Table 6. Laboratory performance ratings for standard reference sample M-164 (major constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter; µg/L, micrograms per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius; V/16, number of rated analyses out of 16 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Vanadium	
MPV =	4.91 µg/L
F-pseudosigma =	0.262
Lab	RV Rating
1	4.726 3
4	-- --
5	5.24 2
8	4.7 3
10	-- --
12	-- --
16	4.8 4
23	-- --
24	-- --
25	<19 NR
26	-- --
32	5.7 0
33	-- --
38	-- --
42	5.13 3
45	5.06 3
46	-- --
59	4.97 4
64	-- --
70	-- --
76	5.102 3
85	-- --
86	4.82 4
91	-- --
97	<4.90 NR
100	5.53 0
102	-- --
105	<20 NR
109	-- --
113	-- --
118	-- --
121	3.8 0
134	4.91 4
138	4.88 4
142	4.42 1
146	3.92 0
149	-- --
151	-- --
180	-- --
183	-- --
190	-- --
193	-- --
208	-- --
212	4.92 4
219	4.8 4

Analyte = Vanadium	
MPV =	4.91 µg/L
F-pseudosigma =	0.262
Lab	RV Rating
220	<10 NR
224	-- --
227	-- --
230	5.2 2
247	<10 NR
254	-- --
255	-- --
256	5.09 3
257	-- --
259	-- --
265	4.8 4
266	-- --
273	-- --
274	-- --
276	-- --
277	-- --
279	-- --
296	6.18 0
307	-- --
319	-- --
326	-- --
328	<8 NR
330	-- --
333	-- --
341	4.7 3
349	-- --
356	-- --
366	-- --
374	-- --
379	-- --
384	-- --
385	-- --
386	-- --
387	-- --

**Table 7. Laboratory performance ratings for standard reference sample N-75 (nutrient constituents)**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter; V/5, number of rated analyses out of 5 possible; RV, reported value; <, less than; NR, not rated; --, not reported.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Ammonia +		Ammonia +		Nitrate as N		Total		Orthophosphate	
	OLR	V/5	Ammonia as N	Organic N as N	Organic N as N	Nitrate as N	Nitrate as N	Phosphorus as P	Phosphorus as P	as P	as P	
	MPV =	F-pseudosigma =	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	3.2	5	0.043	0	0.103	4	0.09	4	0.131	4	0.094	4
5	3.0	5	0.071	3	0.08	3	0.098	2	0.125	4	0.09	3
8	3.8	5	0.077	4	0.084	3	0.093	4	0.129	4	0.095	4
10	3.8	5	0.07	3	0.12	4	0.09	4	0.129	4	0.097	4
16	2.4	5	0.08	4	0.12	4	0.06	0	0.115	1	0.089	3
21	3.2	5	0.077	4	0.094	3	0.083	1	0.126	4	0.094	4
23	1.8	4	<0.10	NR	0.5	0	0.09	4	0.12	3	0.11	0
25	2.0	5	0.09	2	0.12	4	0.113	0	0.13	4	0.169	0
31	3.2	5	0.077	4	0.09	3	0.083	1	0.126	4	0.094	4
33	0.0	3	0.132	0	--	--	0.128	0	--	--	0.13	0
38	1.8	5	0.083	3	0.09	3	0.088	3	0.151	0	0.115	0
42	3.0	3	--	--	--	--	0.096	3	0.128	4	0.103	2
45	1.3	3	--	--	--	--	0.111	0	0.143	1	0.101	3
46	3.3	4	0.072	4	<0.14	NR	0.086	2	0.123	3	0.098	4
64	3.8	4	0.08	4	--	--	0.09	4	0.132	3	0.093	4
70	2.2	5	0.07	3	0.156	2	0.091	4	0.111	0	0.104	2
72	1.6	5	0.056	1	0.069	2	0.074	0	0.119	2	0.101	3
76	4.0	2	0.077	4	--	--	--	--	0.1255	4	--	--
85	4.0	5	0.075	4	0.097	4	0.091	4	0.13	4	0.096	4
86	2.0	3	0.055	1	--	--	0.084	1	0.13	4	--	--
91	3.0	4	0.0641	2	0.104	4	0.095	3	0.135	3	--	--
97	2.5	4	0.073	4	<0.12	NR	0.091	4	0.1	0	0.086	2
102	1.4	5	0.12	0	0.05	1	0.12	0	0.137	2	0.093	4
105	1.3	4	0.13	0	<1.00	NR	0.1	1	0.125	4	0.058	0
110	3.0	2	0.09	2	--	--	0.09	4	--	--	--	--
113	2.6	5	0.045	0	0.09	3	0.089	3	0.125	4	0.099	3
118	2.8	5	0.091	2	0.1	4	0.097	3	0.119	2	0.101	3
134	3.0	4	0.077	4	<0.2	NR	0.087	3	0.143	1	0.092	4
138	3.8	5	0.075	4	0.129	4	0.097	3	0.131	4	0.094	4
142	2.8	5	0.109	0	0.112	4	0.096	3	0.131	4	0.09	3
146	0.6	5	0.103	0	0.374	0	0.16	0	0.162	0	0.089	3
151	2.5	4	0.08	4	--	--	0.099	2	0.142	1	0.091	3
180	3.0	5	0.077	4	0.12	4	0.119	0	0.121	3	0.098	4
183	1.5	2	--	--	--	--	--	--	0.133	3	0.229	0
190	3.6	5	0.079	4	0.143	3	0.092	4	0.12	3	0.093	4
193	3.7	3	0.07	3	<0.5	NR	0.09	4	0.131	4	--	--
247	2.0	5	0.14	0	0.2	0	0.09	4	0.12	3	0.09	3
313	3.2	5	0.0709	3	0.0761	2	0.0936	4	0.121	3	0.0967	4
316	3.6	5	0.0774	4	0.135	3	0.0904	4	0.131	4	0.0998	3
318	3.6	5	0.08	4	0.113	4	0.0966	3	0.12	3	0.0968	4
320	2.8	5	0.0767	4	0.1518	2	0.0911	4	0.1275	4	0.918	0
328	1.6	5	0.13	0	0.105	4	0.08	0	0.13	4	0.4	0
341	3.2	5	0.072	4	0.11	4	0.09	4	0.129	4	0.079	0
366	1.0	5	0.104	0	0.03	0	0.1	1	0.118	2	0.103	2
369	3.6	5	0.075	4	0.125	4	0.096	3	0.13	4	0.091	3

**Table 7. Laboratory performance ratings for standard reference sample N-75 (nutrient constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter; V/5, number of rated analyses out of 5 possible; RV, reported value; <, less than; NR, not rated; --, not reported.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	OLR	V/5	Ammonia + Organic N as N				Nitrate as N		Total Phosphorus as P		Orthophosphate as P	
			Ammonia as N		Organic N as N		Nitrate as N		Total Phosphorus as P		Orthophosphate as P	
			MPV =	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV
			0.077 mg/L		0.113 mg/L		0.092 mg/L		0.128 mg/L		0.095 mg/L	
			0.0114		0.0348		0.0052		0.0082		0.0072	
<b>373</b>	3.0	5	0.065	2	0.181	1	0.094	4	0.127	4	0.095	4
<b>378</b>	2.6	5	0.0707	3	0.223	0	0.0943	4	0.134	3	0.0913	3
<b>379</b>	2.4	5	0.06	2	0.152	2	0.093	4	0.084	0	0.092	4
<b>380</b>	2.8	5	0.0863	3	0.126	4	0.086	2	0.118	2	0.09	3
<b>381</b>	2.5	4	0.076	4	--	--	0.098	2	0.106	0	0.094	4
<b>384</b>	0.5	2	--	--	--	--	0.1003	1	--	--	0.2599	0

**Table 8. Laboratory performance ratings for standard reference sample N-76 (nutrient constituents)**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter; V/5, number of rated analyses out of 5 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Ammonia +		Ammonia +		Nitrate as N		Total		Orthophosphate	
	OLR	V/5	Ammonia as N	Organic N as N	Organic N as N	Nitrate as N	Nitrate as N	Phosphorus as P	Phosphorus as P	as P	as P	
	MPV =	F-pseudosigma =	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
			0.885 mg/L		0.950 mg/L		0.891 mg/L		0.958 mg/L		0.895 mg/L	
			0.0482		0.0615		0.0460		0.0389 (0.0479)		0.0623	
1	3.4	5	0.915	3	0.94	4	0.884	4	0.954	4	0.973	2
5	1.8	5	0.833	2	0.81	0	0.94	2	1.004	3	0.986	2
8	2.2	5	0.84	3	0.84	1	0.93	3	0.99	3	0.99	1
10	3.6	5	0.9	4	0.97	4	0.91	4	0.974	4	0.972	2
12	2.6	5	0.86	3	1.11	0	0.92	3	0.926	3	0.886	4
16	2.0	5	0.91	3	1.09	0	0.69	0	0.985	3	0.895	4
23	2.6	5	0.895	4	0.916	3	0.858	3	0.92	3	0.703	0
25	2.6	5	0.9	4	0.98	4	0.814	1	0.94	4	0.754	0
26	1.7	3	0.843	3	--	--	0.822	2	--	--	0.323	0
30	4.0	2	--	--	--	--	0.872	4	--	--	0.876	4
33	2.7	3	0.951	2	--	--	0.891	4	--	--	0.812	2
38	3.4	5	0.979	1	0.94	4	0.881	4	0.966	4	0.868	4
42	3.0	3	--	--	--	--	0.948	2	0.943	4	0.956	3
45	1.3	3	--	--	--	--	0.816	1	1.036	1	0.817	2
46	3.2	5	0.845	3	1.05	1	0.913	4	0.966	4	0.883	4
64	3.5	4	0.9	4	--	--	0.92	3	0.97	4	0.95	3
70	1.6	5	0.759	0	1.96	0	0.884	4	1.53	0	0.92	4
72	2.6	5	0.814	2	0.981	4	0.93	3	1.026	2	0.964	2
76	3.5	2	0.8796	4	--	--	--	--	0.9905	3	--	--
85	3.8	5	0.88	4	0.92	4	0.92	3	0.972	4	0.91	4
86	4.0	3	0.861	4	--	--	0.9	4	0.954	4	--	--
91	3.8	4	0.841	3	0.934	4	0.9	4	0.98	4	--	--
97	3.4	5	0.915	3	0.98	4	0.92	3	0.99	3	0.897	4
102	2.8	5	1.04	0	0.94	4	0.87	4	0.958	4	0.98	2
105	2.8	5	0.84	3	1.24	0	0.91	4	0.963	4	0.837	3
113	4.0	5	0.885	4	0.93	4	0.898	4	0.944	4	0.926	4
118	3.0	5	0.976	1	0.947	4	0.923	3	0.963	4	0.948	3
134	3.4	5	0.898	4	0.917	3	0.926	3	0.975	4	0.946	3
138	3.8	5	0.845	3	0.939	4	0.89	4	0.964	4	0.875	4
142	2.6	5	0.922	3	1.172	0	0.92	3	0.992	3	0.871	4
146	1.0	5	0.764	0	1.08	0	0.828	2	1.13	0	0.936	3
180	2.6	5	0.892	4	0.886	2	0.909	4	0.886	2	0.99	1
183	2.3	3	--	--	--	--	1.15	0	0.917	3	0.902	4
190	3.6	5	0.876	4	0.993	3	0.883	4	0.933	3	0.91	4
193	2.5	4	0.88	4	1.07	1	0.93	3	0.904	2	--	--
205	1.5	2	0.92	3	--	--	1.01	0	--	--	--	--
208	3.0	2	--	--	--	--	0.922	3	--	--	0.84	3
212	1.4	5	0.85	3	0.95	4	0.76	0	0.62	0	0.62	0
224	1.4	5	0.94	2	0.62	0	0.871	4	0.84	0	0.8	1
227	2.6	5	0.928	3	0.818	0	0.847	3	0.933	3	0.892	4
247	1.6	5	1.07	0	1.05	1	0.78	0	0.92	3	0.87	4
307	4.0	3	0.882	4	--	--	0.89	4	0.949	4	--	--
313	3.2	5	0.903	4	0.849	1	0.889	4	0.984	3	0.888	4
320	3.3	3	0.9263	3	0.9893	3	0.8969	4	--	--	--	--
328	2.8	5	0.86	3	0.95	4	0.86	3	0.94	4	2.88	0



**Table 8. Laboratory performance ratings for standard reference sample N-76 (nutrient constituents) -- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter; V/5, number of rated analyses out of 5 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Ammonia as N		Ammonia + Organic N as N		Nitrate as N		Total Phosphorus as P		Orthophosphate as P	
	OLR	V/5	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
	MPV =	F-pseudostigma =	0.885 mg/L	0.0482	0.950 mg/L	0.0615	0.891 mg/L	0.0460	0.958 mg/L	0.0389 (0.0479)	0.895 mg/L	0.0623
<b>341</b>	3.0	5	0.88	4	0.95	4	0.879	4	0.983	3	0.7	0
<b>349</b>	1.3	4	0.75	0	0.84	1	0.82	1	0.931	3	--	--
<b>356</b>	3.0	4	0.798	1	--	--	0.905	4	0.931	3	0.87	4
<b>366</b>	2.4	5	0.907	4	0.85	1	0.845	3	0.883	1	0.929	3
<b>373</b>	2.2	5	0.888	4	0.936	4	0.85	3	1.107	0	1.063	0
<b>378</b>	3.0	5	0.852	3	1.01	3	0.963	1	0.955	4	0.865	4
<b>379</b>	3.0	5	0.86	3	0.98	4	0.85	3	0.87	1	0.91	4
<b>380</b>	2.6	5	0.964	1	0.973	4	0.86	3	0.866	1	0.865	4
<b>383</b>	2.0	3	0.91	3	--	--	0.93	3	--	--	0.72	0
<b>384</b>	0.0	2	--	--	--	--	0.449	0	--	--	2.934	0

**Table 9. Laboratory performance ratings for standard reference sample P-39 (low ionic-strength constituents)**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter;  $\mu$ S/cm, microsiemens per centimeter at 25 degrees Celsius; V/11, number of rated analyses out of 11 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Acidity		Calcium		Chloride		Fluoride		Potassium	
	OLR	V/11	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
	MPV =	F-pseudosigma =	inadequate data		8.65 mg/L 0.274 (0.433)		2.07 mg/L 0.112		0.650 mg/L 0.0852		1.57 mg/L 0.073 (0.079)	
1	3.0	9	--	--	8.629	4	2.36	0	0.687	4	1.57	4
2	3.3	9	--	--	9.013	3	2.196	2	0.63	4	1.609	4
5	2.7	9	--	--	8.3	3	1.87	1	1.08	0	1.56	4
8	3.3	9	--	--	8.5	4	2.1	4	0.74	2	1.6	4
23	2.9	9	--	--	8.13	2	2.01	3	0.61	4	1.54	4
25	1.6	9	21	NR	7.659	0	1.9	1	1.1	0	1.111	0
33	3.3	9	--	--	8.63	4	2.11	4	0.625	4	1.46	2
38	3.0	6	--	--	8.65	4	--	--	--	--	1.62	3
45	3.4	8	--	--	8.63	4	1.87	1	0.587	3	1.57	4
64	3.4	8	--	--	8.59	4	2.07	4	--	--	1.57	4
85	3.8	5	--	--	--	--	2.12	4	0.64	4	--	--
86	3.2	9	--	--	8.88	3	2.03	4	0.797	1	1.6	4
105	2.2	9	21.4	NR	8.83	4	2.09	4	0.83	0	1.69	1
110	3.0	6	--	--	8.46	4	1.98	3	--	--	--	--
113	3.6	8	--	--	8.72	4	2.07	4	0.68	4	1.54	4
134	3.8	9	--	--	8.96	3	2.05	4	0.66	4	1.53	4
138	3.4	9	--	--	9	3	2.05	4	0.611	4	1.57	4
180	2.7	9	--	--	8.72	4	1.87	1	0.605	3	1.84	0
190	3.7	9	--	--	8.68	4	2.04	4	0.64	4	1.46	2
193	3.1	7	--	--	8.83	4	--	--	--	--	1.63	3
208	3.0	2	--	--	--	--	2.09	4	--	--	--	--
228	2.1	8	--	--	8.12	2	1.84	0	--	--	1.57	4
247	3.2	9	51	NR	8.48	4	1.91	2	0.63	4	1.43	1
265	3.2	6	--	--	8.7	4	1.7	0	--	--	1.6	4
273	1.1	9	15.3	NR	9.63	0	2.34	0	3.31	0	1.79	0
274	1.6	9	14.92	NR	4.82	0	15.96	0	0.74	2	1.4	0
279	1.8	4	--	--	9.9	0	--	--	--	--	1.82	0
326	1.9	8	--	--	8.73	4	2.17	3	--	--	1.44	1
328	2.6	9	21	NR	8.44	4	4.15	0	0.63	4	1.47	2
333	2.8	6	--	--	--	--	2.09	4	--	--	1.51	3
379	2.9	9	--	--	7.4	0	2.16	3	0.73	3	1.6	4
384	2.4	7	--	--	9.1233	2	2.1316	3	0.6128	4	1.68	2

**Table 9. Laboratory performance ratings for standard reference sample P-39 (low ionic-strength constituents)**

-- continued

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter;  $\mu$ S/cm, microsiemens per centimeter at 25 degrees Celsius; V/11, number of rated analyses out of 11 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte =		Magnesium		Sodium		pH		Orthophosphate as P		Sulfate	
	MPV =		0.812 mg/L		6.19 mg/L		3.66		inadequate data		29.3 mg/L	
	F-pseudosigma =		0.0393 (0.0406)		0.297 (0.310)		0.039 (0.183)				1.08 (1.47)	
	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	0.817	4	5.957	3	3.66	4	--	--	34.2	0		
2	0.843	3	6.391	3	3.628	4	--	--	29.961	4		
5	0.753	2	6.22	4	3.41	2	0.001	NR	29.2	4		
8	0.81	4	6.1	4	3.64	4	<0.3	NR	32.1	1		
23	0.89	1	37.8	0	3.66	4	<0.01	NR	29.1	4		
25	0.819	4	5.044	0	3.8	3	<0.003	NR	27.7	2		
33	0.829	4	5.96	3	3.65	4	0.181	NR	29.8	4		
38	0.675	0	6.09	4	3.7	4	--	--	--	--		
45	0.818	4	6.28	4	--	--	<0.1	NR	28.5	3		
64	0.78	3	6.18	4	3.67	4	<0.002	NR	29.5	4		
85	--	--	--	--	3.65	4	--	--	28.6	4		
86	0.833	3	6.36	3	3.62	4	--	--	29.5	4		
105	0.851	3	6.66	1	3.7	4	<0.01	NR	46	0		
110	0.758	2	--	--	3.68	4	--	--	28.15	3		
113	0.78	3	6.23	4	--	--	<0.004	NR	28.5	3		
134	0.791	3	6.09	4	3.68	4	<0.01	NR	29.43	4		
138	0.812	4	6.46	3	3.68	4	<0.004	NR	28.3	3		
180	0.799	4	6.09	4	3.59	4	0.015	NR	29.7	4		
190	0.82	4	6.26	4	3.51	3	<0.01	NR	29.4	4		
193	0.75	1	5.96	3	3.65	4	--	--	28.7	4		
208	--	--	--	--	--	--	--	--	27.8	2		
228	0.74	1	7.92	0	3.65	4	--	--	30.53	3		
247	0.79	3	6.19	4	3.66	4	<0.01	NR	29.3	4		
265	0.83	4	6.3	4	--	--	--	--	28.5	3		
273	0.89	1	7.03	0	3.6	4	0.01	NR	31.88	1		
274	3.38	0	4.92	0	3.59	4	0.0033	NR	29.03	4		
279	0.84	3	6.33	4	--	--	--	--	--	--		
326	0.672	0	5.73	2	3.82	3	--	--	27.74	2		
328	0.76	2	5.96	3	3.66	4	0.043	NR	34	0		
333	--	--	7.07	0	3.68	4	<0.01	NR	31.1	2		
379	0.8	4	6	3	3.85	2	0.046	NR	29.3	4		
384	0.8803	1	5.63	1	--	--	--	--	28.9206	4		

**Table 9. Laboratory performance ratings for standard reference sample P-39 (low ionic-strength constituents)  
-- continued**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses; mg/L, milligrams per liter;  $\mu\text{S}/\text{cm}$ , microsiemens per centimeter at 25 degrees Celsius; V/11, number of rated analyses out of 11 possible; RV, reported value; <, less than; NR, not rated; --, not reported; ( ), rating criterion.]

<u>Rating</u>	<u>Absolute Z-value</u>	<u>Rating</u>	<u>Absolute Z-value</u>
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

		<b>Specific</b>	
<b>Analyte =</b>		<b>Conductance</b>	
MPV =		193 $\mu\text{S}/\text{cm}$	
F-pseudosigma =		8.9 (9.6)	
<b>Lab</b>		<b>RV</b>	<b>Rating</b>
1		190	4
2		186.7	3
5		195	4
8		200	3
23		189	4
25		194	4
33		176.8	1
38		198.7	3
45		195	4
64		159	0
85		185	3
86		201	3
105		183.5	3
110		182.8	2
113		198	3
134		196	4
138		180	2
180		245	0
190		190	4
193		186	3
208		--	--
228		200.7	3
247		201	3
265		--	--
273		196	4
274		193	4
279		--	--
326		154.6	0
328		195	4
333		192.7	4
379		185	3
384		--	--

**Table 10. Laboratory performance ratings for standard reference sample HG-35 (Mercury)**

[MPV, most probable value; Lab, laboratory identification number; OLR, overall laboratory rating for all rated analyses;  $\mu\text{g/L}$ , micrograms per liter; V/1, number of rated analyses out of 1 possible; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	


**Analyte = Mercury**

MPV = 0.026  $\mu\text{g/L}$

F-pseudosigma = 0.0024

Lab	OLR	V/1	RV	Rating
1	4.0	1	0.0264	4
8	NR	0	<0.1	NR
23	NR	0	<0.10	NR
32	0.0	1	0.021	0
45	3.0	1	0.0284	3
46	4.0	1	0.0262	4
59	0.0	1	0.0185	0
105	NR	0	<0.2	NR
134	NR	0	<0.1	NR
138	4.0	1	0.0268	4
147	0.0	1	0.033	0
180	NR	0	<0.050	NR
193	3.0	1	0.0246	3
212	2.0	1	0.0238	2
235	3.0	1	0.025	3
246	3.0	1	0.0279	3
247	NR	0	<0.2	NR
304	4.0	1	0.0268	4
307	NR	0	<0.0955	NR
356	0.0	1	0.032	0
379	NR	0	<0.1	NR


**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents)**



SUMMARY	Methods			Method Codes	Statistics	
	3	4	6			
n =	10	10	17	03 Atomic absorption: graphite furnace	<b>MPV = 2.44 µg/L</b>	
Minimum =	1.8	1.38	2.1	04 Inductively coupled plasma	F-pseudosigma = 0.119	
Maximum =	3.3	3.5	3.3	06 Inductively coupled plasma/mass spectrometry	Rating criterion = 0.122	
Median =	2.35	2.47	2.49		n = 37	
F-pseudosigma =	0.200	0.222	0.104		Uh = 2.50	
					Lh = 2.34	

Lab	Rating	Z-value	Method Codes			Lab	Rating	Z-value	Method Codes		
			3	4	6				3	4	6
1	3	-0.74	2.35	--	--	341	4	0.49	--	--	2.5
5	NR	--	--	<4.00	--	349	0	-2.79	--	--	2.1
8	4	0.49	--	--	2.5	356	4	0.49	--	--	2.5
12	0	-5.25	1.8	--	--	379	0	8.69	--	3.5	--
16	0	2.13	--	2.7	--						
23	2	-1.48	2.26	--	--						
25	NR	--	--	<17	--						
26	0	-2.54	2.13	--	--						
32	4	0.49	--	--	2.5						
42	0	-11.80	--	--	<1.0						
45	0	4.02	--	--	2.93						
59	4	0.08	--	--	2.45						
70	NR	--	<10	--	--						
86	0	-8.69	--	1.38	--						
97	3	-0.66	2.36	--	--						
100	4	-0.33	2.4	--	--						
105	4	-0.33	--	--	2.4						
113	3	-0.82	2.34	--	--						
134	4	-0.33	--	2.4	--						
138	4	-0.16	--	2.42	--						
142	0	-2.30	--	--	2.16						
146	0	-2.30	--	2.16	--						
149	3	-0.66	--	--	2.36						
151	NR	--	--	--	<10						
180	3	-0.74	--	--	2.35						
190	0	-3.61	2	--	--						
193	1	1.72	2.65	--	--						
212	4	0.00	--	2.44	--						
220	NR	--	--	<5	--						
235	4	0.33	--	--	2.48						
247	NR	--	--	<10	--						
255	0	2.13	--	--	2.7						
256	4	0.49	--	2.5	--						
259	4	0.41	--	2.49	--						
265	2	-1.15	--	--	2.3						
296	0	7.05	--	--	3.3						
304	4	0.41	--	--	2.49						
307	0	7.05	3.3	--	--						
328	0	2.95	--	2.8	--						
330	2	1.31	--	--	2.6						


**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued**



SUMMARY	Methods				Method Codes		Statistics	
	3	4	5	6				
n =	3	15	0	19	03 Atomic absorption: graphite furnace		MPV = 19.4 µg/L	
Minimum =	17.7	15.1	0	17	04 Inductively coupled plasma		F-pseudosigma = 1.63	
Maximum =	19.2	73.6		48	05 Direct current plasma		n = 37	
Median =		21.0		19.3	06 Inductively coupled plasma/mass spectrometry		Uh = 21.0	
F-pseudosigma =		7.23		1.22			Lh = 18.8	

Lab	Rating	Z-value	Method Codes				Lab	Rating	Z-value	Method Codes			
			3	4	5	6				3	4	5	6
1	4	-0.29	--	--	--	18.92	328	NR	--	--	<40	--	--
5	NR	--	--	<30.0	--	--	330	4	-0.49	--	--	--	18.6
7	0	17.54	--	--	--	48	341	NR	--	--	--	--	<50
8	4	0.25	--	--	--	19.8	349	3	0.98	--	--	--	21
16	2	-1.47	--	--	--	17	356	0	18.15	--	--	49	--
25	0	10.18	--	36	--	--	379	0	-2.64	--	15.1	--	--
32	4	0.00	--	--	--	19.4	384	0	3.86	--	25.7	--	--
33	NR	--	--	--	<50	--							
42	NR	--	--	--	<30.0	--							
45	4	-0.31	--	--	--	18.9							
46	NR	--	--	<100	--	--							
70	0	2.82	--	--	--	24							
86	3	0.86	--	20.8	--	--							
97	4	-0.12	19.2	--	--	--							
100	4	-0.37	--	18.8	--	--							
105	3	0.92	--	--	--	20.9							
110	4	-0.47	--	18.63	--	--							
113	4	-0.37	--	18.8	--	--							
134	3	0.98	--	21	--	--							
138	3	0.67	--	20.5	--	--							
146	0	33.23	--	73.6	--	--							
149	4	-0.06	--	--	--	19.3							
151	4	-0.25	--	--	--	19							
180	4	-0.43	--	--	--	18.7							
190	2	-1.04	17.7	--	--	--							
193	4	-0.31	18.9	--	--	--							
212	0	7.36	--	31.4	--	--							
219	4	-0.06	--	--	--	19.3							
220	NR	--	--	<20	--	--							
227	0	2.64	--	23.7	--	--							
230	3	0.98	--	--	--	21							
235	2	-1.35	--	--	--	17.2							
246	NR	--	--	<30	--	--							
247	NR	--	--	<80	--	--							
254	NR	--	--	<50	--	--							
256	3	-0.67	--	18.3	--	--							
259	1	1.84	--	22.4	--	--							
265	4	0.06	--	--	--	19.5							
296	4	0.43	--	--	--	20.1							
304	4	-0.37	--	--	--	18.8							

Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued




SUMMARY	Methods					Statistics	
	3	4	6	10	11	Method Codes	
n =	10	6	27	1	1	03 Atomic absorption: graphite furnace	MPV = 3.50 µg/L
Minimum =	2.55	3.26	3	6.34	3.5	04 Inductively coupled plasma	F-pseudosigma = 0.274
Maximum =	4.3	5.99	4.2			06 Inductively coupled plasma/mass spectrometry	n = 45
Median =	3.39	3.47	3.50			10 Atomic absorption: extraction	Uh = 3.70
F-pseudosigma =	0.623	0.297	0.194			11 Atomic absorption: hydride	Lh = 3.33

Lab	Rating	Z-value	Method Codes					Lab	Rating	Z-value	Method Codes				
			3	4	6	10	11				3	4	6	10	11
1	4	-0.24	--	--	3.435	--	259	3	-0.88	--	3.26	--	--	--	
5	3	0.84	3.73	--	--	--	265	4	0.00	--	--	3.5	--	--	
7	4	-0.36	--	--	3.4	--	296	2	1.02	--	--	3.78	--	--	
8	4	0.36	--	--	3.6	--	304	4	0.44	--	--	3.62	--	--	
10	4	0.00	--	--	--	3.5	307	0	-3.46	2.55	--	--	--	--	
12	1	1.82	4	--	--	--	326	4	0.36	--	3.6	--	--	--	
16	2	-1.09	--	--	3.2	--	328	4	-0.36	--	--	3.4	--	--	
23	NR	--	<10	--	--	--	330	2	1.09	--	--	3.8	--	--	
25	3	0.73	--	3.7	--	--	341	2	-1.46	--	--	3.1	--	--	
32	4	-0.44	--	--	3.38	--	349	4	0.36	--	--	3.6	--	--	
42	4	0.07	--	--	3.52	--	356	4	0.44	--	--	3.62	--	--	
45	3	0.69	--	--	3.69	--	379	3	-0.73	--	3.3	--	--	--	
46	1	1.53	3.92	--	--	--									
59	3	0.88	--	--	3.74	--									
70	0	2.55	--	--	4.2	--									
76	4	-0.49	--	--	3.366	--									
97	4	-0.29	3.42	--	--	--									
100	1	-1.68	3.04	--	--	--									
105	NR	--	--	--	<4.0	--									
134	3	-0.55	3.35	--	--	--									
138	3	-0.62	--	3.33	--	--									
142	4	0.00	--	--	3.5	--									
146	0	9.08	--	5.99	--	--									
147	3	-0.51	--	--	3.36	--									
149	3	-0.80	--	--	3.28	--									
151	4	0.00	--	--	3.5	--									
180	3	-0.88	--	--	3.26	--									
190	1	-1.53	3.08	--	--	--									
193	NR	--	<5.0	--	--	--									
212	NR	--	--	<15	--	--									
219	1	-1.82	--	--	3	--									
220	0	2.92	4.3	--	--	--									
230	4	0.36	--	--	3.6	--									
235	3	0.55	--	--	3.65	--									
246	NR	--	--	<5	--	--									
247	NR	--	--	<40	--	--									
254	NR	--	--	<10	--	--									
255	1	1.82	--	--	4	--									
256	0	10.35	--	--	6.34	--									
257	2	-1.43	3.108	--	--	--									




**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued**



SUMMARY	Methods		Statistics						
	4	6							
n =	16	7	<b>MPV = 23.5 µg/L</b>						
Minimum =	14	0.026	F-pseudosigma = 3.04						
Maximum =	25.9	25.4	n = 23						
Median =	23.3	23.9	Uh = 24.2						
F-pseudosigma =	3.45	1.59	Lh = 20.1						
	<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th colspan="2">Method Codes</th> </tr> <tr> <th>04</th> <th>06</th> </tr> </thead> <tbody> <tr> <td>Inductively coupled plasma</td> <td>Inductively coupled plasma/mass spectrometry</td> </tr> </tbody> </table>		Method Codes		04	06	Inductively coupled plasma	Inductively coupled plasma/mass spectrometry	
Method Codes									
04	06								
Inductively coupled plasma	Inductively coupled plasma/mass spectrometry								

Method Codes				
Lab	Rating	Z-value	4	6
1	4	-0.11	23.16	--
5	3	0.79	25.9	--
7	NR	--	<30	--
8	4	0.13	--	23.9
16	0	-2.63	15.5	--
24	4	0.16	24	--
25	2	-1.09	20.2	--
32	4	0.23	--	24.2
42	NR	--	--	<30.0
45	3	0.63	--	25.4
70	NR	--	<100	--
86	0	-3.13	14	--
100	NR	--	<40.0	--
105	NR	--	--	<200
134	4	0.16	24	--
138	1	-1.81	18	--
151	0	-7.72	--	0.026
212	4	-0.20	22.9	--
219	2	-1.15	20	--
220	4	0.49	25	--
235	3	-0.69	--	21.4
247	NR	--	<50	--
254	4	0.20	24.1	--
255	3	0.76	25.8	--
259	4	0.00	23.5	--
265	4	-0.16	--	23
326	4	0.23	24.2	--
328	2	-1.48	19	--
349	4	0.33	--	24.5


**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued**



SUMMARY	Methods				Method Codes	Statistics	
	3	4	5	6			
n =	2	22	1	22	03 Atomic absorption: graphite furnace	<b>MPV = 13.6 µg/L</b>	
Minimum =	13	10.7	14.49	12.9	04 Inductively coupled plasma	F-pseudosigma = 0.46	
Maximum =	14.4	16.5		14.3	05 Direct current plasma	Rating criterion = 0.68	
Median =		13.6		13.7	06 Inductively coupled plasma/mass spectrometry	n = 47	
F-pseudosigma =		0.52		0.40		Uh = 14.0	
						Lh = 13.3	

Lab	Rating	Z-value	Method Codes				Lab	Rating	Z-value	Method Codes			
			3	4	5	6				3	4	5	6
1	4	-0.05	--	13.56	--	--	273	0	4.26	--	16.5	--	--
5	0	-2.50	--	11.9	--	--	296	4	0.29	--	--	--	13.8
7	4	-0.15	--	13.5	--	--	304	4	-0.15	--	--	--	13.5
8	3	0.59	--	--	--	14	326	4	0.29	--	13.8	--	--
16	3	-0.88	--	13	--	--	328	4	-0.29	--	--	--	13.4
24	3	0.59	--	14	--	--	330	4	0.44	--	--	--	13.9
25	0	-4.26	--	10.7	--	--	341	3	-0.88	--	--	--	13
32	4	0.15	--	--	--	13.7	349	3	-0.59	--	--	--	13.2
33	2	1.31	--	--	14.49	--	379	0	-2.50	--	11.9	--	--
42	4	0.29	--	--	--	13.8							
45	3	-0.74	--	--	--	13.1							
46	2	1.03	--	14.3	--	--							
59	3	0.59	--	--	--	14							
70	2	1.03	--	--	--	14.3							
76	NR	--	--	--	--	<20.0							
86	4	-0.29	--	13.4	--	--							
97	2	1.18	14.4	--	--	--							
100	4	0.44	--	13.9	--	--							
105	3	-0.88	--	--	--	13							
113	4	-0.44	--	13.3	--	--							
121	0	2.06	--	15	--	--							
134	4	0.29	--	13.8	--	--							
138	4	0.29	--	13.8	--	--							
142	4	-0.29	--	--	--	13.4							
146	4	0.00	--	13.6	--	--							
147	4	0.29	--	--	--	13.8							
149	4	-0.35	--	--	--	13.36							
151	3	0.59	--	--	--	14							
180	4	0.29	--	--	--	13.8							
193	3	-0.88	13	--	--	--							
212	0	3.97	--	16.3	--	--							
219	2	-1.03	--	--	--	12.9							
220	2	1.18	--	14.4	--	--							
230	4	0.44	--	--	--	13.9							
235	4	0.00	--	--	--	13.6							
246	2	-1.47	--	12.6	--	--							
247	0	-5.29	--	<10	--	--							
254	4	0.00	--	13.6	--	--							
259	4	0.00	--	13.6	--	--							
265	4	-0.15	--	--	--	13.5							

Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued




SUMMARY	Methods			Method Codes		Statistics	
	3	4	6				
n =	2	13	25	03 Atomic absorption: graphite furnace		<b>MPV = 1.18 µg/L</b>	
Minimum =	1.1	0.73	1	04 Inductively coupled plasma		F-pseudosigma = 0.107	
Maximum =	1.21	1.4	1.4	06 Inductively coupled plasma/mass spectrometry		n = 40	
Median =	1.18		1.20			Uh = 1.25	
F-pseudosigma =	0.104		0.111			Lh = 1.10	

Lab	Rating	Z-value	Method Codes			Lab	Rating	Z-value	Method Codes		
			3	4	6				3	4	6
1	4	0.30	--	--	1.212	349	2	1.12	--	--	1.3
5	4	0.00	--	1.18	--	356	4	-0.19	--	--	1.16
7	1	-1.67	--	--	1	379	2	1.12	--	1.3	--
8	4	0.19	--	--	1.2						
16	3	-0.74	--	1.1	--						
25	3	-0.74	--	1.1	--						
32	3	0.65	--	--	1.25						
42	2	1.02	--	--	1.29						
45	3	0.84	--	--	1.27						
46	NR	--	--	<2	--						
59	4	-0.37	--	--	1.14						
70	0	2.05	--	--	1.4						
76	3	-0.61	--	--	1.114						
86	4	0.00	--	1.18	--						
100	1	1.86	--	1.38	--						
105	4	0.37	--	--	1.22						
113	4	-0.28	--	1.15	--						
134	4	0.19	--	1.2	--						
138	4	0.00	--	1.18	--						
142	3	0.65	--	--	1.25						
146	NR	--	--	<4.00	--						
149	3	-0.84	--	--	1.09						
151	4	0.19	--	--	1.2						
180	3	-0.56	--	--	1.12						
193	4	0.28	1.21	--	--						
212	0	-4.19	--	0.73	--						
219	3	-0.74	--	--	1.1						
220	3	-0.74	1.1	--	--						
230	2	1.12	--	--	1.3						
235	2	-1.21	--	--	1.05						
246	1	-1.67	--	1	--						
247	NR	--	--	<10	--						
256	3	0.56	--	1.24	--						
265	4	0.19	--	--	1.2						
273	0	2.05	--	1.4	--						
296	4	-0.19	--	--	1.16						
304	3	-0.74	--	--	1.1						
328	3	-0.74	--	--	1.1						
330	4	0.19	--	--	1.2						
341	3	-0.74	--	--	1.1						

Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued


T-171 CALCIUM (Ca) in mg/L



SUMMARY	Methods							Method Codes		Statistics	
	0	1	2	4	5	6	20				
n =	2	7	1	39	1	5	1	00 Other		MPV = 6.75 mg/L	
Minimum =	6.5	4.13	7.15	6.07	7.22	6.4	4.82	01 Atomic absorption: direct, air		F-pseudosigma = 0.263	
Maximum =	7.44	7		7.08		7		02 Atomic absorption: direct, nitrous oxide		Rating criterion = 0.337	
Median =		6.57		6.76		6.63		04 Inductively coupled plasma		n = 56	
F-pseudosigma =		0.423		0.189		0.111		05 Direct current plasma		Uh = 6.90	
								06 Inductively coupled plasma/mass spectrometry		Lh = 6.55	
								20 Titration: colorimetric			

Lab	Rating	Z-value	Method Codes							Lab	Rating	Z-value	Method Codes						
			0	1	2	4	5	6	20				0	1	2	4	5	6	20
1	4	0.09	--	--	--	6.775	--	--	--	254	4	0.34	--	--	--	6.86	--	--	--
5	3	-0.64	--	--	--	6.53	--	--	--	259	4	-0.28	--	--	--	6.65	--	--	--
7	4	-0.49	--	--	--	6.58	--	--	--	265	4	0.46	--	--	--	6.9	--	--	--
8	4	0.46	--	--	--	6.9	--	--	--	273	4	-0.31	--	--	--	6.64	--	--	--
16	2	-1.32	--	--	--	6.3	--	--	--	274	0	-5.71	--	--	--	--	--	--	4.82
23	3	-0.87	--	6.45	--	--	--	--	--	279	0	2.06	7.44	--	--	--	--	--	--
24	2	-1.47	--	--	--	6.25	--	--	--	296	4	-0.13	--	--	--	--	--	6.7	--
25	0	-2.00	--	--	--	6.07	--	--	--	326	4	0.04	--	--	--	6.76	--	--	--
26	4	0.28	--	--	--	6.84	--	--	--	328	1	-1.62	--	--	--	6.2	--	--	--
32	3	-0.58	--	--	--	--	--	6.55	--	330	4	-0.13	--	--	--	6.7	--	--	--
33	2	1.41	--	--	--	--	7.22	--	--	341	1	-1.91	--	6.1	--	--	--	--	--
42	1	-1.59	--	--	--	6.21	--	--	--	349	3	0.76	--	7	--	--	--	--	--
45	4	-0.34	--	--	--	--	--	6.63	--	356	4	0.16	--	--	--	6.8	--	--	--
46	4	0.31	--	--	--	--	6.85	--	--	381	0	-7.75	--	4.13	--	--	--	--	--
59	4	0.01	--	6.75	--	--	--	--	--	384	3	0.63	--	--	--	6.957	--	--	--
64	4	-0.22	--	--	--	6.67	--	--	--	386	3	-0.73	6.5	--	--	--	--	--	--
70	4	0.22	--	--	--	6.82	--	--	--										
86	4	0.22	--	--	--	6.82	--	--	--										
97	3	0.87	--	--	--	7.04	--	--	--										
100	3	0.82	--	--	--	7.02	--	--	--										
105	4	-0.40	--	--	--	6.61	--	--	--										
109	3	-0.52	--	6.57	--	--	--	--	--										
110	4	0.22	--	--	--	6.82	--	--	--										
113	4	-0.07	--	--	--	6.72	--	--	--										
121	4	0.31	--	--	--	6.85	--	--	--										
134	3	0.52	--	--	--	6.92	--	--	--										
138	3	0.70	--	--	--	6.98	--	--	--										
146	4	0.31	--	--	--	6.85	--	--	--										
149	2	-1.02	--	--	--	--	--	6.4	--										
180	4	0.01	--	--	--	6.75	--	--	--										
190	2	1.20	--	--	7.15	--	--	--	--										
193	3	0.58	--	6.94	--	--	--	--	--										
212	3	-0.93	--	--	--	6.43	--	--	--										
219	3	0.76	--	--	--	7	--	--	--										
220	4	-0.22	--	--	--	6.67	--	--	--										
227	4	-0.46	--	--	--	6.59	--	--	--										
230	3	0.76	--	--	--	--	--	7	--										
235	3	0.99	--	--	--	7.08	--	--	--										
246	4	-0.01	--	--	--	6.74	--	--	--										
247	3	-0.61	--	--	--	6.54	--	--	--										


**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued**



SUMMARY	Methods			Method Codes			Statistics	
	3	4	6	03 Atomic absorption: graphite furnace	04 Inductively coupled plasma	06 Inductively coupled plasma/mass spectrometry	MPV = 4.92 µg/L	F-pseudosigma = 0.226
n =	12	19	25				Rating criterion = 0.246	n = 56
Minimum =	3.29	4.06	4.8				Uh = 5.11	Lh = 4.80
Maximum =	5.5	9.3	6					
Median =	4.84	4.80	5.00					
F-pseudosigma =	0.322	0.297	0.230					

Lab	Rating	Z-value	Method Codes			Lab	Rating	Z-value	Method Codes		
			3	4	6				3	4	6
1	4	-0.06	4.9	--	--	254	2	-1.28	--	4.6	--
5	1	1.77	5.35	--	--	255	0	2.79	--	--	5.6
7	4	0.35	--	--	5	256	3	-0.75	--	4.73	--
8	4	0.35	--	--	5	257	0	-6.61	3.29	--	--
10	4	-0.47	4.8	--	--	259	4	-0.18	--	4.87	--
16	4	-0.47	--	4.8	--	265	4	-0.06	--	--	4.9
23	2	-1.16	--	4.63	--	273	0	17.84	--	9.3	--
25	0	15.81	--	8.8	--	296	0	2.01	--	--	5.41
26	2	-1.20	4.62	--	--	304	3	0.55	--	--	5.05
32	0	2.71	--	--	5.58	307	2	-1.44	4.56	--	--
42	0	-3.19	--	4.13	--	326	4	-0.47	--	4.8	--
45	4	-0.10	--	--	4.89	328	3	0.75	--	--	5.1
46	2	-1.12	4.64	--	--	330	2	1.16	--	--	5.2
59	4	0.35	--	--	5	341	4	-0.47	--	--	4.8
70	0	4.42	--	--	6	349	4	-0.47	--	--	4.8
76	4	0.09	--	--	4.936	356	3	0.79	--	--	5.11
86	3	-1.00	--	4.67	--	379	0	2.38	--	5.5	--
97	1	-1.85	4.46	--	--						
100	3	0.67	5.08	--	--						
105	4	0.35	--	--	5						
113	3	-0.63	--	4.76	--						
121	3	0.75	--	5.1	--						
134	4	0.35	--	5	--						
138	2	1.28	--	5.23	--						
142	4	-0.14	--	--	4.88						
146	4	0.31	--	4.99	--						
147	4	-0.14	--	--	4.88						
149	4	-0.18	--	--	4.87						
151	1	1.57	--	--	5.3						
180	4	-0.06	--	--	4.9						
190	4	-0.14	4.88	--	--						
193	4	0.22	4.97	--	--						
212	0	-3.48	--	4.06	--						
219	4	-0.47	--	--	4.8						
220	0	2.38	5.5	--	--						
227	4	0.06	--	4.93	--						
230	3	0.75	--	--	5.1						
235	2	1.32	--	--	5.24						
246	1	-1.69	--	4.5	--						
247	NR	--	--	<10	--						


Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods			Method Codes	Statistics	
	3	4	6			
n =	2	13	20	03 Atomic absorption: graphite furnace	MPV = 4.71 µg/L	
Minimum =	5.1	3.8	4.3	04 Inductively coupled plasma	F-pseudostigma = 0.259	
Maximum =	5.35	6.3	5.7	06 Inductively coupled plasma/mass spectrometry	n = 35	
Median =	4.73 4.66				Uh = 4.85	
F-pseudostigma =	0.297 0.196				Lh = 4.50	

Lab	Rating	Z-value	Method Codes			Lab	Rating	Z-value	Method Codes		
			3	4	6				3	4	6
1	4	0.19	--	--	4.76	356	4	-0.23	--	--	4.65
5	0	-6.59	--	<3.00	--	379	3	0.73	--	4.9	--
7	4	-0.42	--	--	4.6						
8	0	2.27	--	--	5.3						
16	0	-3.51	--	3.8	--						
25	0	6.13	--	6.3	--						
32	4	-0.50	--	--	4.58						
42	2	-1.31	--	--	4.37						
45	4	0.04	--	--	4.72						
59	1	-1.58	--	--	4.3						
70	0	3.82	--	--	5.7						
76	4	-0.16	--	--	4.668						
86	4	0.31	--	4.79	--						
97	0	2.47	5.35	--	--						
100	NR	--	--	<5.0	--						
105	NR	--	--	--	<50						
134	4	0.35	--	4.8	--						
138	4	0.08	--	4.73	--						
142	4	-0.04	--	--	4.7						
146	3	0.96	--	4.96	--						
149	4	-0.15	--	--	4.67						
180	3	-1.00	--	--	4.45						
193	2	1.50	5.1	--	--						
212	4	0.08	--	4.73	--						
219	2	-1.19	--	--	4.4						
220	NR	--	--	<10	--						
230	4	0.35	--	--	4.8						
235	4	0.00	--	--	4.71						
246	1	-1.58	--	4.3	--						
247	NR	--	--	<10	--						
254	3	-0.81	--	4.5	--						
256	2	-1.16	--	4.41	--						
259	4	0.08	--	4.73	--						
265	3	-0.81	--	--	4.5						
296	2	1.16	--	--	5.01						
304	4	-0.42	--	--	4.6						
326	0	2.27	--	5.3	--						
328	NR	--	--	<8	--						
341	NR	--	--	<15	--						
349	2	-1.19	--	--	4.4						


**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued**



SUMMARY	Methods			Method Codes			Statistics		
	3	4	6	03 Atomic absorption: graphite furnace	04 Inductively coupled plasma	06 Inductively coupled plasma/mass spectrometry	MPV = 3.10 µg/L	F-pseudosigma = 0.274	n = 39
n =	7	12	20						
Minimum =	2.763	2.74	2.7						
Maximum =	4	10.5	3.44						
Median =	3.14	3.22	3.08						Uh = 3.30
F-pseudosigma =	0.341	0.552	0.245						Lh = 2.93

Lab	Rating	Z-value	Method Codes			Lab	Rating	Z-value	Method Codes		
			3	4	6				3	4	6
1	2	-1.23	2.763	--	--	259	3	0.58	--	3.26	--
5	0	6.13	--	4.78	--	265	2	-1.46	--	--	2.7
7	0	-4.01	--	--	<2	296	2	1.24	--	--	3.44
8	3	0.73	--	--	3.3	304	4	0.36	--	--	3.2
10	4	0.00	3.1	--	--	307	0	3.28	4	--	--
12	NR	--	--	<20	--	326	0	26.98	--	10.5	--
16	2	-1.09	--	2.8	--	328	2	1.09	--	--	3.4
23	NR	--	--	<4.00	--	330	4	0.00	--	--	3.1
25	NR	--	--	<15	--	341	4	-0.36	3	--	--
32	4	-0.22	--	--	3.04	349	4	0.00	--	--	3.1
42	4	-0.36	--	--	3	356	4	-0.40	--	--	2.99
45	3	0.66	--	--	3.28	379	0	2.19	--	3.7	--
46	0	2.37	3.75	--	--						
59	4	-0.07	--	--	3.08						
70	NR	--	--	--	<5.0						
76	NR	--	--	--	<5.0						
86	2	-1.06	--	2.81	--						
97	0	-3.65	--	<2.1	--						
100	4	0.15	3.14	--	--						
105	NR	--	--	--	<4						
113	4	0.29	--	3.18	--						
134	3	0.73	--	3.3	--						
138	3	-0.88	--	2.86	--						
142	2	-1.31	--	--	2.74						
146	2	-1.31	--	2.74	--						
149	3	0.51	--	--	3.24						
151	2	-1.09	--	--	2.8						
180	4	-0.07	--	--	3.08						
190	3	0.62	3.27	--	--						
193	NR	--	--	--	<5						
212	2	1.31	--	3.46	--						
219	3	-0.73	--	--	2.9						
220	NR	--	--	<10	--						
230	2	-1.46	--	--	2.7						
235	3	-0.51	--	--	2.96						
246	NR	--	--	<5	--						
247	NR	--	--	<10	--						
254	NR	--	--	<5	--						
255	3	0.73	--	--	3.3						
256	4	0.29	--	3.18	--						

**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued**




SUMMARY	Methods			Method Codes	Statistics	
	3	4	6			
n =	2	3	21	03 Atomic absorption: graphite furnace	<b>MPV = 1.00 µg/L</b>	
Minimum =	1.1	0.949	0.6	04 Inductively coupled plasma	F-pseudosigma = 0.172	
Maximum =	1.2	13.4	1.95	06 Inductively coupled plasma/mass spectrometry	n = 26	
Median =			1		Uh = 1.20	
F-pseudosigma =			0.172		Lh = 0.968	

Lab	Rating	Z-value	Method Codes			Lab	Rating	Z-value	Method Codes		
			3	4	6				3	4	6
1	4	-0.19	--	--	0.968	296	1	1.51	--	--	1.26
5	NR	--	--	<4.00	--	304	4	0.00	--	--	1
7	4	-0.23	--	--	0.96	307	NR	--	<1.46	--	--
8	2	1.16	--	--	1.2	328	2	1.16	--	--	1.2
10	NR	--	<2	--	--	330	4	0.00	--	--	1
12	3	0.58	1.1	--	--	341	0	2.91	--	--	1.5
16	0	-2.33	--	--	0.6	349	3	0.58	--	--	1.1
23	NR	--	--	<5.00	--	356	4	-0.12	--	--	0.98
25	0	72.10	--	13.4	--	379	NR	--	--	<1.5	--
32	4	0.29	--	--	1.05						
42	NR	--	--	--	<2.0						
45	4	-0.17	--	--	0.971						
46	NR	--	<3	--	--						
59	4	-0.17	--	--	0.97						
70	NR	--	--	--	<5.0						
76	NR	--	--	--	<20.0						
100	NR	--	--	<5.0	--						
105	NR	--	--	--	<10						
134	4	0.00	--	1	--						
138	4	-0.30	--	0.949	--						
142	NR	--	--	--	<1						
146	NR	--	--	<20.0	--						
147	4	-0.23	--	--	0.96						
149	0	2.38	--	--	1.41						
151	NR	--	--	--	<10						
180	3	-0.66	--	--	0.886						
190	NR	--	<2.3	--	--						
193	NR	--	<12.5	--	--						
212	NR	--	--	<10	--						
219	3	-0.87	--	--	0.85						
220	2	1.16	1.2	--	--						
227	NR	--	--	<1.46	--						
230	NR	--	--	--	<5						
235	4	0.00	--	--	1						
246	NR	--	--	<3	--						
247	NR	--	--	<10	--						
254	NR	--	--	<5	--						
255	2	1.16	--	--	1.2						
256	NR	--	--	<10.00	--						
265	0	5.52	--	--	1.95						




Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods							Method Codes	Statistics	
	0	1	3	4	5	6	22			
n =	1	5	0	35	0	5	1	00 Other	MPV =	53.4 µg/L
Minimum =	56.5	50.8	0	40.2	0	44	55.2	01 Atomic absorption: direct, air	F-pseudsigma =	4.82
Maximum =		67		61.7		62.2		03 Atomic absorption: graphite furnace	n =	47
Median =		54.0		53.4		50.0		04 Inductively coupled plasma	Uh =	56.0
F-pseudsigma =		2.97		4.91		2.97		05 Direct current plasma	Lh =	49.5
								06 Inductively coupled plasma/mass spectrometry		
								22 Colorimetric		

Lab	Rating	Z-value	Method Codes							Lab	Rating	Z-value	Method Codes						
			0	1	3	4	5	6	22				0	1	3	4	5	6	22
1	4	0.17	--	--	--	54.21	--	--	--	256	4	0.34	--	--	--	55.05	--	--	--
5	3	-0.93	--	--	--	48.9	--	--	--	259	4	-0.23	--	--	--	52.3	--	--	--
7	1	-1.64	--	--	--	45.5	--	--	--	265	4	-0.08	--	--	--	53	--	--	--
8	NR	--	--	--	--	<100	--	--	--	273	2	-1.22	--	--	--	47.5	--	--	--
10	3	0.75	--	57	--	--	--	--	--	296	1	-1.95	--	--	--	--	--	44	--
12	NR	--	--	--	--	<50	--	--	--	326	3	0.98	--	--	--	58.1	--	--	--
16	4	-0.37	--	--	--	51.6	--	--	--	328	3	0.54	--	--	--	56	--	--	--
21	4	0.37	--	--	--	--	--	--	55.2	341	4	0.12	--	--	--	54	--	--	--
24	2	-1.45	--	--	--	46.4	--	--	--	349	0	2.82	--	67	--	--	--	--	--
25	1	1.72	--	--	--	61.7	--	--	--	356	4	-0.08	--	--	--	53	--	--	--
32	2	-1.12	--	--	--	--	--	48	--	379	2	-1.10	--	--	--	48.1	--	--	--
33	NR	--	--	--	--	<100	--	--	--	384	4	-0.02	--	--	--	53.3	--	--	--
42	3	0.58	--	--	--	56.2	--	--	--	386	3	0.64	56.5	--	--	--	--	--	--
45	1	1.83	--	--	--	--	--	62.2	--										
46	NR	--	--	--	--	<300	--	--	--										
70	0	-2.18	--	--	--	42.9	--	--	--										
86	4	0.25	--	--	--	54.6	--	--	--										
97	0	-2.26	--	--	--	42.5	--	--	--										
100	4	0.31	--	--	--	54.9	--	--	--										
105	2	1.16	--	--	--	59	--	--	--										
109	4	-0.08	--	53	--	--	--	--	--										
113	4	0.19	--	--	--	54.3	--	--	--										
134	4	-0.08	--	--	--	53	--	--	--										
138	2	-1.31	--	--	--	47.1	--	--	--										
146	3	0.79	--	--	--	57.2	--	--	--										
149	4	-0.29	--	--	--	--	--	52	--										
151	4	0.12	--	54	--	--	--	--	--										
180	3	0.52	--	--	--	55.9	--	--	--										
190	3	-0.54	--	50.8	--	--	--	--	--										
193	NR	--	--	--	<100	--	--	--	--										
212	3	0.89	--	--	--	57.7	--	--	--										
219	3	-0.71	--	--	--	50	--	--	--										
220	0	-2.74	--	--	--	40.2	--	--	--										
227	4	0.12	--	--	--	54	--	--	--										
230	3	-0.71	--	--	--	--	--	50	--										
235	3	-0.95	--	--	--	48.8	--	--	--										
246	4	0.33	--	--	--	55	--	--	--										
247	NR	--	--	--	--	<50	--	--	--										
254	4	0.00	--	--	--	53.4	--	--	--										
255	3	0.75	--	--	--	57	--	--	--										


Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods						Method Codes		Statistics	
	0	1	4	5	6	12				
n =	2	13	29	1	6	1	00 Other		MPV = 2.80 mg/L	
Minimum =	2.95	2.68	2.06	2.79	2.77	2.5	01 Atomic absorption: direct, air		F-pseudsigma = 0.093	
Maximum =	3.08	2.96	3.12		2.9		04 Inductively coupled plasma		Rating criterion = 0.140	
Median =	2.80	2.80	2.80		2.80		05 Direct current plasma		n = 52	
F-pseudsigma =		0.052	0.156		0.043		06 Inductively coupled plasma/mass spectrometry		Uh = 2.86	
							12 Flame emission		Lh = 2.74	

Lab	Rating	Z-value	Method Codes							Lab	Rating	Z-value	Method Codes						
			0	1	4	5	6	12	0				1	4	5	6	12		
1	4	0.36	--	2.85	--	--	--	--	273	0	-2.36	--	--	2.47	--	--	--		
5	0	2.29	--	--	3.12	--	--	--	274	0	-2.14	--	--	--	--	--	2.5		
7	3	-0.79	--	--	2.69	--	--	--	279	1	2.00	3.08	--	--	--	--	--		
8	3	0.71	--	--	2.9	--	--	--	296	4	0.00	--	--	--	--	2.8	--		
16	3	0.71	--	--	2.9	--	--	--	326	2	1.14	--	2.96	--	--	--	--		
23	4	0.29	--	2.84	--	--	--	--	328	1	-1.64	--	--	2.57	--	--	--		
24	3	0.71	--	--	2.9	--	--	--	330	4	-0.14	--	--	2.78	--	--	--		
25	0	-5.29	--	--	2.06	--	--	--	341	4	0.00	--	2.8	--	--	--	--		
26	1	-2.00	--	--	2.52	--	--	--	349	4	0.00	--	2.8	--	--	--	--		
32	4	0.29	--	--	--	--	2.84	--	356	4	0.00	--	--	2.8	--	--	--		
33	4	-0.07	--	--	--	2.79	--	--	381	4	-0.50	--	2.73	--	--	--	--		
42	3	-1.00	--	--	2.66	--	--	--	384	2	1.50	--	--	3.01	--	--	--		
45	4	-0.21	--	--	--	--	2.77	--	386	2	1.07	2.95	--	--	--	--	--		
46	NR	--	--	--	<3	--	--	--											
59	4	-0.21	--	2.77	--	--	--	--											
64	4	0.07	--	2.81	--	--	--	--											
70	4	0.29	--	--	2.84	--	--	--											
76	4	-0.13	--	--	--	--	2.782	--											
86	4	0.21	--	--	2.83	--	--	--											
97	4	0.07	--	2.81	--	--	--	--											
100	4	0.50	--	--	2.87	--	--	--											
105	4	0.29	--	--	2.84	--	--	--											
109	3	-0.64	--	2.71	--	--	--	--											
113	4	0.00	--	--	2.8	--	--	--											
134	4	-0.36	--	2.75	--	--	--	--											
138	4	0.00	--	--	2.8	--	--	--											
146	4	0.07	--	--	2.81	--	--	--											
149	4	-0.07	--	--	--	--	2.79	--											
180	2	1.21	--	--	2.97	--	--	--											
190	3	-0.86	--	2.68	--	--	--	--											
193	4	0.14	--	2.82	--	--	--	--											
212	2	1.07	--	--	2.95	--	--	--											
219	3	0.71	--	--	2.9	--	--	--											
220	0	-3.07	--	--	2.37	--	--	--											
230	3	0.71	--	--	--	--	2.9	--											
246	4	-0.43	--	--	2.74	--	--	--											
247	3	-0.93	--	--	2.67	--	--	--											
254	4	-0.14	--	--	2.78	--	--	--											
259	4	-0.50	--	--	2.73	--	--	--											
265	4	0.00	--	--	2.8	--	--	--											


**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued**



SUMMARY	Methods			Method Codes	Statistics	
	0	4	6			
n =	1	9	9	00 Other	MPV =	6.30 µg/L
Minimum =	6.18	5.58	5.5	04 Inductively coupled plasma	F-pseudosigma =	0.599
Maximum =		25.8	8.5	06 Inductively coupled plasma/mass spectrometry	n =	19
Median =		6.50	6.12		Uh =	6.83
F-pseudosigma =		1.557	0.532		Lh =	6.02

Lab	Rating	Z-value	Method Codes		
			0	4	6
1	4	-0.45	--	--	6.033
5	0	3.00	--	8.1	--
7	NR	--	--	--	<20
8	3	1.00	--	--	6.9
25	0	32.54	--	25.8	--
32	3	0.75	--	--	6.75
42	2	-1.20	--	5.58	--
76	4	-0.31	--	--	6.115
86	4	-0.20	6.18	--	--
100	NR	--	--	--	<50
105	NR	--	--	<25	--
134	4	0.33	--	6.5	--
151	NR	--	--	--	<10
212	2	-1.08	--	5.65	--
219	2	-1.33	--	--	5.5
220	NR	--	--	<10	--
230	4	0.00	--	--	6.3
246	4	-0.50	--	6	--
247	NR	--	--	<10	--
254	4	0.00	--	6.3	--
256	3	0.67	--	6.7	--
265	4	-0.50	--	--	6
273	0	5.84	--	9.8	--
296	0	3.67	--	--	8.5
328	4	-0.42	--	--	6.05
341	NR	--	--	<10	--


Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods						Method Codes		Statistics	
	0	1	4	5	6	20				
n =	3	8	38	1	6	1	00 Other		MPV = 2.78 mg/L	
Minimum =	2.7	2.62	2.56	3.092	2.7	3.86	01 Atomic absorption: direct, air		F-pseudosigma = 0.111	
Maximum =	2.95	3	3.05		2.89		04 Inductively coupled plasma		Rating criterion = 0.139	
Median =	2.79	2.78			2.80		05 Direct current plasma		n = 57	
F-pseudosigma =	0.189	0.089			0.030		06 Inductively coupled plasma/mass spectrometry		Uh = 2.88	
							20 Titration: colorimetric		Lh = 2.73	

Lab	Rating	Z-value	Method Codes							Lab	Rating	Z-value	Method Codes						
			0	1	4	5	6	20	0				1	4	5	6	20		
1	4	0.00	--	--	2.78	--	--	--	247	4	-0.07	--	--	2.77	--	--	--		
5	3	-0.72	--	--	2.68	--	--	--	254	3	1.01	--	--	2.92	--	--	--		
7	4	-0.50	--	--	2.71	--	--	--	259	4	-0.14	--	--	2.76	--	--	--		
8	3	0.86	--	--	2.9	--	--	--	265	4	0.50	--	--	2.85	--	--	--		
16	3	-0.58	--	--	2.7	--	--	--	273	3	1.01	--	--	2.92	--	--	--		
23	4	-0.43	--	2.72	--	--	--	--	274	0	7.77	--	--	--	--	--	3.86		
24	2	-1.22	--	--	2.61	--	--	--	279	2	1.22	2.95	--	--	--	--	--		
25	4	-0.08	--	--	2.769	--	--	--	296	3	-0.58	--	--	--	--	2.7	--		
26	4	0.22	--	--	2.81	--	--	--	326	4	0.50	--	--	2.85	--	--	--		
32	3	0.79	--	--	--	--	2.89	--	328	1	-1.58	--	--	2.56	--	--	--		
33	0	2.24	--	--	--	3.092	--	--	330	4	0.00	--	--	2.78	--	--	--		
42	3	-0.94	--	--	2.65	--	--	--	341	1	1.58	--	3	--	--	--	--		
45	4	0.00	--	--	--	--	2.78	--	349	4	0.14	--	2.8	--	--	--	--		
46	4	0.00	--	--	2.78	--	--	--	356	4	0.50	--	--	2.85	--	--	--		
59	3	-0.94	--	2.65	--	--	--	--	381	1	1.51	--	2.99	--	--	--	--		
64	4	-0.50	2.71	--	--	--	--	--	384	1	1.94	--	--	3.05	--	--	--		
70	3	0.65	--	--	2.87	--	--	--	386	3	-0.58	2.7	--	--	--	--	--		
76	4	0.12	--	--	--	--	2.797	--											
86	4	-0.36	--	--	2.73	--	--	--											
97	4	-0.29	--	--	2.74	--	--	--											
100	2	1.08	--	--	2.93	--	--	--											
105	4	0.00	--	--	2.78	--	--	--											
109	4	-0.07	--	2.77	--	--	--	--											
110	4	0.14	--	--	2.8	--	--	--											
113	4	-0.22	--	--	2.75	--	--	--											
121	4	0.29	--	--	2.82	--	--	--											
134	4	0.14	--	--	2.8	--	--	--											
138	3	0.58	--	--	2.86	--	--	--											
146	3	0.72	--	--	2.88	--	--	--											
149	4	0.29	--	--	--	--	2.82	--											
180	4	-0.07	--	--	2.77	--	--	--											
190	3	0.79	--	2.89	--	--	--	--											
193	2	-1.15	--	2.62	--	--	--	--											
212	3	-0.94	--	--	2.65	--	--	--											
219	3	0.86	--	--	2.9	--	--	--											
220	4	-0.22	--	--	2.75	--	--	--											
227	4	-0.14	--	--	2.76	--	--	--											
230	4	0.14	--	--	--	--	2.8	--											
235	3	0.79	--	--	2.89	--	--	--											
246	3	-0.72	--	--	2.68	--	--	--											


**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued**



SUMMARY	Methods						Statistics	
	0	1	3	4	5	6	Method Codes	
n =	1	3	2	27	0	21	00 Other	MPV = 12.4 µg/L
Minimum =	12.5	11.6	12.2	10	0	5	01 Atomic absorption: direct, air	F-pseudostigma = 0.44
Maximum =	14.99	14.68	16.5	14.2			03 Atomic absorption: graphite furnace	Rating criterion = 0.62
Median =			12.3	12.4			04 Inductively coupled plasma	n = 54
F-pseudostigma =			0.63	0.44			05 Direct current plasma	Uh = 12.6
							06 Inductively coupled plasma/mass spectrometry	Lh = 12.0

Lab	Rating	Z-value	Method Codes						Lab	Rating	Z-value	Method Codes					
			0	1	3	4	5	6				0	1	3	4	5	6
1	4	0.08	--	--	--	--	12.4	255	4	-0.08	--	--	--	12.3	--	--	
5	4	0.40	--	--	--	12.6	--	256	4	0.24	--	--	--	12.5	--	--	
7	4	0.08	--	--	--	--	12.4	257	0	3.78	--	--	14.68	--	--		
8	0	2.67	--	--	--	--	14	259	4	0.24	--	--	--	12.5	--	--	
10	2	-1.21	--	11.6	--	--	--	265	4	-0.08	--	--	--	--	--	12.3	
12	NR	--	--	--	--	<20	--	273	0	6.72	--	--	--	16.5	--	--	
16	3	-0.57	--	--	--	12	--	296	0	-11.90	--	--	--	--	--	5	
23	4	-0.36	--	--	--	12.13	--	304	4	-0.08	--	--	--	--	--	12.3	
24	2	-1.05	--	--	--	11.7	--	326	4	0.40	--	--	--	12.6	--	--	
25	1	-1.86	--	--	--	11.2	--	328	2	1.05	--	--	--	13	--	--	
32	3	-0.57	--	--	--	--	12	330	3	0.89	--	--	--	--	--	12.9	
33	NR	--	--	--	--	<100	--	341	3	-0.57	--	--	--	--	--	12	
42	3	-0.89	--	--	--	--	11.8	349	NR	--	--	--	--	--	--	<20	
45	4	0.24	--	--	--	--	12.5	356	3	-0.57	--	--	--	--	--	12	
46	4	-0.24	--	--	--	12.2	--	379	4	0.40	--	--	--	12.6	--	--	
59	4	0.40	--	--	--	--	12.6	381	0	3.64	--	14.6	--	--	--	--	
70	0	3.00	--	--	--	--	14.2	384	0	-2.29	--	--	--	10.93	--	--	
86	4	-0.24	--	--	--	12.2	--	386	4	0.24	12.5	--	--	--	--	--	
97	1	-1.70	--	--	--	11.3	--										
100	4	-0.08	--	--	--	12.3	--										
105	2	1.05	--	--	--	--	13										
109	0	4.28	--	14.99	--	--	--										
113	4	0.08	--	--	--	12.4	--										
121	2	1.05	--	--	--	13	--										
134	3	0.89	--	--	--	12.9	--										
138	4	0.08	--	--	--	12.4	--										
146	4	-0.24	--	--	--	12.2	--										
149	4	0.19	--	--	--	--	12.47										
151	3	-0.57	--	--	--	--	12										
180	4	-0.40	--	--	--	--	12.1										
190	4	-0.24	--	--	12.2	--	--										
193	NR	--	--	--	<50	--	--										
212	1	1.70	--	--	--	13.4	--										
219	3	-0.57	--	--	--	--	12										
220	2	-1.47	--	--	--	11.44	--										
230	2	1.05	--	--	--	--	13										
235	4	0.24	--	--	--	--	12.5										
246	1	-1.70	--	--	--	11.3	--										
247	0	-3.81	--	--	--	10	--										
254	3	-0.89	--	--	--	11.8	--										


Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods				Method Codes		Statistics	
	1	3	4	6				
n =	1	1	8	15	01 Atomic absorption: direct, air		<b>MPV = 1.97 µg/L</b>	
Minimum =	2.2	1.94	1.8	1.7	03 Atomic absorption: graphite furnace		F-pseudostigma = 0.133	
Maximum =			16.5	2.01	04 Inductively coupled plasma		n = 25	
Median =			2.52	1.94	06 Inductively coupled plasma/mass spectrometry		Uh = 2.08	
F-pseudostigma =			0.822	0.082			Lh = 1.90	

Lab	Rating	Z-value	Method Codes			
			1	3	4	6
1	0	7.25	--	--	2.938	--
5	NR	--	--	--	<10.0	--
7	NR	--	--	--	--	<3
8	3	-0.52	--	--	--	1.9
16	0	3.97	--	--	2.5	--
32	4	0.00	--	--	--	1.97
42	NR	--	--	--	--	<10
45	4	0.07	--	--	--	1.98
59	2	-1.35	--	--	--	1.79
70	NR	--	--	--	--	<5.0
76	NR	--	--	--	--	<5.0
86	3	0.82	--	--	2.08	--
97	4	-0.22	--	1.94	--	--
100	0	108.89	--	--	16.5	--
105	NR	--	--	--	--	<4
113	0	4.27	--	--	2.54	--
134	4	-0.15	--	--	1.95	--
138	NR	--	--	--	<2.0	--
142	4	-0.22	--	--	--	1.94
146	NR	--	--	--	<10.0	--
149	3	-0.97	--	--	--	1.84
151	NR	--	--	--	--	<2
180	4	0.30	--	--	--	2.01
193	NR	--	--	<5	--	--
212	0	10.04	--	--	3.31	--
219	NR	--	--	--	--	<2
220	NR	--	--	--	<10	--
230	4	0.22	--	--	--	2
235	4	0.07	--	--	--	1.98
246	NR	--	--	--	<3	--
247	4	0.00	--	--	--	1.97
256	2	-1.27	--	--	1.8	--
265	2	-1.27	--	--	--	1.8
296	4	-0.22	--	--	--	1.94
304	3	-0.60	--	--	--	1.89
328	3	-0.52	--	--	--	1.9
330	0	-2.02	--	--	--	1.7
341	NR	--	--	--	--	<25
349	1	1.72	2.2	--	--	--


Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods						Method Codes		Statistics	
	0	1	4	5	6	12				
n =	2	11	34	1	6	1	00 Other		MPV =	8.60 mg/L
Minimum =	8.67	7.7	7.151	8.58	8.57	6.15	01 Atomic absorption: direct, air		F-pseudostigma =	0.282
Maximum =	8.89	14.62	12.4		9.15		04 Inductively coupled plasma		Rating criterion =	0.430
Median =		8.61	8.52		8.70		05 Direct current plasma		n =	55
F-pseudostigma =		0.252	0.334		0.227		06 Inductively coupled plasma/mass spectrometry		Uh =	8.80
							12 Flame emission		Lh =	8.42

Lab	Rating	Z-value	Method Codes							Lab	Rating	Z-value	Method Codes						
			0	1	4	5	6	12				0	1	4	5	6	12		
1	4	-0.26	--	--	8.49	--	--	--	259	4	0.12	--	--	8.65	--	--	--		
5	4	0.40	--	--	8.77	--	--	--	265	4	-0.23	--	--	8.5	--	--	--		
7	3	-0.70	--	--	8.3	--	--	--	273	0	8.84	--	--	12.4	--	--	--		
8	4	0.47	--	--	8.8	--	--	--	274	0	-5.70	--	--	--	--	--	6.15		
16	2	-1.16	--	--	8.1	--	--	--	279	3	0.67	8.89	--	--	--	--	--		
23	0	-2.09	--	7.7	--	--	--	--	296	3	0.70	--	--	--	--	8.9	--		
24	3	-1.00	--	--	8.17	--	--	--	326	4	-0.33	--	--	8.46	--	--	--		
25	0	-3.37	--	--	7.151	--	--	--	328	1	-1.67	--	--	7.88	--	--	--		
26	3	-0.60	--	--	8.34	--	--	--	330	4	0.42	--	--	8.78	--	--	--		
32	2	1.28	--	--	--	--	9.15	--	341	4	0.00	--	8.6	--	--	--	--		
33	4	-0.05	--	--	--	8.58	--	--	349	2	-1.33	--	8.03	--	--	--	--		
42	3	-0.63	--	--	8.33	--	--	--	356	3	1.00	--	--	9.03	--	--	--		
45	4	-0.07	--	--	--	--	8.57	--	381	0	14.00	--	14.62	--	--	--	--		
46	4	0.30	--	--	8.73	--	--	--	384	1	-1.60	--	--	7.91	--	--	--		
59	4	0.12	--	8.65	--	--	--	--	386	4	0.16	8.67	--	--	--	--	--		
64	4	0.02	--	8.61	--	--	--	--											
70	3	0.70	--	--	8.9	--	--	--											
76	4	-0.01	--	--	--	--	8.594	--											
86	4	0.26	--	--	8.71	--	--	--											
97	3	0.60	--	8.86	--	--	--	--											
100	3	0.53	--	--	8.83	--	--	--											
105	4	0.37	--	--	8.76	--	--	--											
109	4	-0.16	--	8.53	--	--	--	--											
113	4	-0.35	--	--	8.45	--	--	--											
121	4	-0.12	--	--	8.55	--	--	--											
134	4	-0.42	--	8.42	--	--	--	--											
138	3	0.84	--	--	8.96	--	--	--											
146	3	0.81	--	--	8.95	--	--	--											
149	4	0.00	--	--	--	--	8.6	--											
180	4	-0.21	--	--	8.51	--	--	--											
190	4	0.40	--	8.77	--	--	--	--											
193	1	1.53	--	9.26	--	--	--	--											
212	4	-0.37	--	--	8.44	--	--	--											
219	4	0.23	--	--	8.7	--	--	--											
220	3	-0.86	--	--	8.23	--	--	--											
230	4	0.47	--	--	--	--	8.8	--											
235	4	0.47	--	--	8.8	--	--	--											
246	2	-1.16	--	--	8.1	--	--	--											
247	4	-0.19	--	--	8.52	--	--	--											
254	4	-0.42	--	--	8.42	--	--	--											

**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued**




SUMMARY	Methods				Method Codes	Statistics	
	0	3	4	6			
n =	2	4	14	24	00 Other	MPV = 5.67 µg/L	
Minimum =	5.65	4.93	4.3	5.25	03 Atomic absorption: graphite furnace	F-pseudostigma = 0.256	
Maximum =	5.88	5.8	8.2	7.5	04 Inductively coupled plasma	Rating criterion = 0.283	
Median =			5.69	5.68	06 Inductively coupled plasma/mass spectrometry	n = 44	
F-pseudostigma =			0.363	0.202		Uh = 5.80	
						Lh = 5.46	

Lab	Rating	Z-value	Method Codes				Lab	Rating	Z-value	Method Codes			
			0	3	4	6				0	3	4	6
1	4	0.46	--	--	--	5.796	259	3	0.83	--	--	5.9	--
5	NR	--	--	--	<10.0	--	265	3	-0.94	--	--	--	5.4
7	4	0.12	--	--	--	5.7	296	4	0.19	--	--	--	5.72
8	3	-0.58	--	--	--	5.5	304	4	-0.05	--	--	--	5.65
12	NR	--	--	--	<20	--	307	0	-2.59	--	4.93	--	--
16	0	-4.82	--	--	4.3	--	326	3	0.76	5.88	--	--	--
23	4	-0.12	--	--	5.63	--	328	3	-0.58	--	--	--	5.5
25	0	8.95	--	--	8.2	--	330	2	1.18	--	--	--	6
32	3	-0.94	--	--	--	5.4	341	4	0.12	--	--	--	5.7
42	2	-1.47	--	--	--	5.25	349	4	0.12	--	--	--	5.7
45	4	0.48	--	--	--	5.8	356	4	-0.05	5.65	--	--	--
46	NR	--	--	--	<50	--	379	1	-1.99	--	--	5.1	--
59	4	-0.02	--	--	--	5.66							
70	0	5.42	--	--	--	7.2							
76	4	-0.31	--	--	--	5.577							
86	4	0.30	--	--	5.75	--							
97	4	0.48	--	5.8	--	--							
100	NR	--	--	--	<20	--							
105	NR	--	--	--	--	<50							
113	4	0.37	--	--	5.77	--							
134	2	1.18	--	--	6	--							
138	4	-0.44	--	--	5.54	--							
142	4	0.02	--	--	--	5.67							
146	0	-2.70	--	--	4.9	--							
149	3	-0.69	--	--	--	5.47							
151	2	1.18	--	--	--	6							
180	3	-1.01	--	--	--	5.38							
190	3	-0.79	--	5.44	--	--							
193	NR	--	--	<12.5	--	--							
212	0	4.22	--	--	6.86	--							
219	3	-0.94	--	--	--	5.4							
220	NR	--	--	--	<10	--							
230	4	0.12	--	--	--	5.7							
235	4	0.05	--	--	--	5.68							
246	4	-0.23	--	--	5.6	--							
247	NR	--	--	--	<50	--							
254	0	2.59	--	--	6.4	--							
255	0	6.48	--	--	--	7.5							
256	3	-0.55	--	--	5.51	--							
257	3	-0.92	--	5.403	--	--							




Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods			Method Codes	Statistics	
	3	4	6			
n =	8	6	26	03 Atomic absorption: graphite furnace	MPV = 1.09 µg/L	
Minimum =	0.94	0.6	0.85	04 Inductively coupled plasma	F-pseudostigma = 0.148	
Maximum =	10.6	5.5	1.3	06 Inductively coupled plasma/mass spectrometry	n = 40	
Median =	1.25	1.33	1.08		Uh = 1.20	
F-pseudostigma =	0.663	1.033	0.064		Lh = 1	

Lab	Rating	Z-value	Method Codes			Lab	Rating	Z-value	Method Codes		
			3	4	6				3	4	6
1	4	-0.41	--	--	1.024	257	0	9.54	2.5	--	--
5	1	1.92	1.37	--	--	265	3	-0.57	--	--	1
7	3	-0.91	--	--	0.95	296	4	-0.10	--	--	1.07
8	4	0.03	--	--	1.09	304	4	-0.17	--	--	1.06
10	NR	--	<2	--	--	307	3	-0.98	0.94	--	--
12	3	-0.57	1	--	--	326	0	-3.27	--	0.6	--
16	3	0.78	--	--	1.2	328	3	-0.57	--	--	1
23	NR	--	--	<5.00	--	330	4	0.10	--	--	1.1
25	0	29.78	--	5.5	--	341	3	-0.57	--	--	1
32	4	0.51	--	--	1.16	349	3	0.78	--	--	1.2
42	4	0.10	--	--	1.1	356	4	0.17	--	--	1.11
45	4	0.17	--	--	1.11						
46	NR	--	<3	--	--						
59	4	0.03	--	--	1.09						
70	2	1.45	--	--	1.3						
76	4	-0.18	--	--	1.058						
97	2	1.25	1.27	--	--						
100	3	0.98	1.23	--	--						
105	4	0.10	--	--	1.1						
134	3	-0.57	--	1	--						
138	3	-1.00	--	0.937	--						
142	NR	--	--	--	<1						
146	NR	--	--	<5.00	--						
147	4	-0.24	--	--	1.05						
149	4	-0.37	--	--	1.03						
151	NR	--	--	--	<1						
180	3	-0.57	--	--	1						
190	4	-0.03	1.08	--	--						
193	NR	--	<5	--	--						
212	0	8.40	--	2.33	--						
219	1	-1.59	--	--	0.85						
220	0	64.18	10.6	--	--						
227	0	3.81	--	1.65	--						
230	4	-0.24	--	--	1.05						
235	4	-0.03	--	--	1.08						
246	NR	--	--	<10	--						
247	NR	--	--	<40	--						
254	NR	--	--	<15	--						
255	3	0.78	--	--	1.2						
256	NR	--	--	<3.00	--						


Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods			Method Codes	Statistics	
	3	4	6			
n =	3	3	25			MPV = 2.73 µg/L
Minimum =	1.434	2.65	2.3	03 Atomic absorption: graphite furnace		F-pseudosigma = 0.234
Maximum =	3.22	4.17	5.3	04 Inductively coupled plasma		n = 31
Median =			2.73	06 Inductively coupled plasma/mass spectrometry		Uh = 2.92
F-pseudosigma =			0.230			Lh = 2.60

Lab	Rating	Z-value	Method Codes		
			3	4	6
1	1	1.76	--	--	3.142
5	NR	--	--	<20.0	--
7	3	-0.56	--	--	2.6
8	4	0.26	--	--	2.79
16	3	-0.98	--	--	2.5
25	NR	--	--	<50	--
32	4	0.09	--	--	2.75
42	4	-0.34	--	--	2.65
45	4	-0.13	--	--	2.7
59	1	1.71	--	--	3.13
70	0	11.01	--	--	5.3
97	3	0.77	2.91	--	--
100	0	2.10	3.22	--	--
105	3	0.73	--	--	2.9
134	4	-0.34	--	2.65	--
138	4	0.00	--	2.73	--
142	4	-0.13	--	--	2.7
146	NR	--	--	<20.0	--
149	3	-0.90	--	--	2.52
151	2	1.16	--	--	3
180	3	-0.73	--	--	2.56
193	NR	--	<12.5	--	--
212	NR	--	--	<10	--
219	3	-0.56	--	--	2.6
220	NR	--	--	<10	--
235	4	0.39	--	--	2.82
247	4	0.09	--	--	2.75
256	0	6.17	--	4.17	--
257	0	-5.55	1.434	--	--
265	3	-0.98	--	--	2.5
296	3	0.77	--	--	2.91
304	4	0.00	--	--	2.73
328	4	-0.13	--	--	2.7
330	2	1.16	--	--	3
341	1	-1.84	--	--	2.3
349	3	-0.56	--	--	2.6
356	3	0.81	--	--	2.92


Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods						Statistics	
	3	4	6	9	10	11	Method Codes	
n =	8	1	19	1	1	1	03 Atomic absorption: graphite furnace	MPV = 1.39 µg/L
Minimum =	1.08	1.4	0.6	1.23	1.29	1.2	04 Inductively coupled plasma	F-pseudosigma = 0.182
Maximum =	2.24		1.8				06 Inductively coupled plasma/mass spectrometry	n = 31
Median =	1.35		1.40				09 Atomic fluorescence	Uh = 1.51
F-pseudosigma =	0.427		0.196				10 Atomic absorption: extraction	Lh = 1.26
							11 Atomic absorption: hydride	

Lab	Rating	Z-value	Method Codes						
			3	4	6	9	10	11	
1	4	0.00	--	--	1.385	--	--	--	
5	0	4.71	2.24	--	--	--	--	--	
7	0	-2.12	--	--	<1	--	--	--	
8	4	-0.08	--	--	1.37	--	--	--	
10	2	-1.02	--	--	--	--	--	1.2	
12	NR	--	--	<2	--	--	--	--	
16	0	-4.32	--	--	0.6	--	--	--	
25	NR	--	--	<16	--	--	--	--	
32	3	-0.52	--	--	1.29	--	--	--	
42	0	2.29	--	--	1.8	--	--	--	
45	4	0.41	--	--	1.46	--	--	--	
59	NR	--	--	--	<5	--	--	--	
70	NR	--	--	--	<5.0	--	--	--	
76	NR	--	--	--	<2.0	--	--	--	
97	1	-1.68	1.08	--	--	--	--	--	
100	4	0.08	1.4	--	--	--	--	--	
105	NR	--	--	--	<7	--	--	--	
134	2	-1.29	1.15	--	--	--	--	--	
138	4	0.08	--	1.4	--	--	--	--	
142	4	0.19	--	--	1.42	--	--	--	
146	NR	--	--	<10.0	--	--	--	--	
149	3	-0.96	--	--	1.21	--	--	--	
151	NR	--	--	--	<5	--	--	--	
180	4	-0.36	--	--	1.32	--	--	--	
190	2	-1.07	1.19	--	--	--	--	--	
193	NR	--	--	<5	--	--	--	--	
212	NR	--	--	<15	--	--	--	--	
219	1	-1.57	--	--	1.1	--	--	--	
220	4	0.08	1.4	--	--	--	--	--	
230	4	0.08	--	--	1.4	--	--	--	
235	3	-0.85	--	--	--	1.23	--	--	
247	NR	--	--	<100	--	--	--	--	
255	1	1.73	--	--	1.7	--	--	--	
256	3	-0.52	--	--	--	--	1.29	--	
257	0	3.90	2.093	--	--	--	--	--	
265	4	-0.47	--	--	1.3	--	--	--	
296	1	1.95	--	--	1.74	--	--	--	
304	2	1.29	--	--	1.62	--	--	--	
307	3	-0.52	1.29	--	--	--	--	--	
328	3	0.63	--	--	1.5	--	--	--	


Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods					Method Codes	Statistics	
	0	4	5	6	22			
n =	1	24	1	3	4	00 Other	MPV =	3.50 mg/L
Minimum =	3.74	1.71	3.72	1.66	1.38	04 Inductively coupled plasma	F-pseudosigma =	0.163
Maximum =		4.05		3.6	3.64	05 Direct current plasma	Rating criterion =	0.175
Median =		3.50				06 Inductively coupled plasma/mass spectrometry	n =	33
F-pseudosigma =		0.141				22 Colorimetric	Uh =	3.60
							Lh =	3.38

Lab	Rating	Z-value	Method Codes				
			0	4	5	6	22
1	4	0.09	--	3.515	--	--	--
5	4	0.17	--	3.53	--	--	--
7	3	-0.69	--	3.38	--	--	--
8	2	1.14	--	3.7	--	--	--
24	3	0.80	--	3.64	--	--	--
25	0	-3.46	--	2.895	--	--	--
32	4	0.46	--	--	--	3.58	--
33	2	1.26	--	--	3.72	--	--
42	0	-2.06	--	3.14	--	--	--
45	0	-10.51	--	--	--	1.66	--
64	4	0.00	--	3.5	--	--	--
70	2	-1.31	--	--	--	--	3.27
97	3	0.80	--	--	--	--	3.64
100	0	3.14	--	4.05	--	--	--
105	4	-0.31	--	3.445	--	--	--
110	4	-0.06	--	3.49	--	--	--
121	4	0.17	--	3.53	--	--	--
134	3	-0.51	--	3.41	--	--	--
190	4	0.00	--	--	--	--	3.5
212	4	-0.34	--	3.44	--	--	--
219	2	1.14	--	3.7	--	--	--
230	3	0.57	--	--	--	3.6	--
235	2	1.49	--	3.76	--	--	--
254	3	-0.97	--	3.33	--	--	--
256	3	-0.57	--	3.4	--	--	--
259	4	-0.29	--	3.45	--	--	--
265	4	0.00	--	3.5	--	--	--
273	1	-1.71	--	3.2	--	--	--
274	0	-12.11	--	--	--	--	1.38
328	4	0.34	--	3.56	--	--	--
356	3	0.57	--	3.6	--	--	--
384	0	-10.23	--	1.71	--	--	--
386	2	1.37	3.74	--	--	--	--


Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods					Statistics	
	1	3	4	5	6	Method Codes	
n =	1	1	28	1	9	01 Atomic absorption: direct, air	<b>MPV = 84.0 µg/L</b>
Minimum =	94.3	84.8	79.8	81.11	78	03 Atomic absorption: graphite furnace	F-pseudosigma = 4.04
Maximum =			97.2		113	04 Inductively coupled plasma	Rating criterion = 4.20
Median =			84.4		81.5	05 Direct current plasma	n = 40
F-pseudosigma =			3.84		2.08	06 Inductively coupled plasma/mass spectrometry	Uh = 87.4
							Lh = 82.0

Lab	Rating	Z-value	Method Codes				
			1	3	4	5	6
1	4	-0.24	--	--	82.94	--	--
5	4	-0.23	--	--	83	--	--
7	4	-0.11	--	--	83.5	--	--
8	2	-1.42	--	--	--	--	78
16	3	-0.94	--	--	80	--	--
24	4	0.06	--	--	84.2	--	--
25	2	1.13	--	--	88.7	--	--
32	4	-0.23	--	--	--	--	83
33	3	-0.68	--	--	--	81.11	--
42	3	0.92	--	--	87.8	--	--
59	3	-0.75	--	--	--	--	80.8
86	4	-0.39	--	--	82.3	--	--
97	3	0.54	--	--	86.2	--	--
100	2	1.04	--	--	88.3	--	--
105	3	0.73	--	--	87	--	--
113	4	-0.27	--	--	82.8	--	--
121	4	-0.23	--	--	83	--	--
134	4	0.01	--	--	84	--	--
138	4	0.44	--	--	85.8	--	--
147	3	-0.58	--	--	--	--	81.5
151	3	-0.70	--	--	--	--	81
190	4	0.20	--	84.8	--	--	--
193	0	2.47	94.3	--	--	--	--
212	4	0.39	--	--	85.6	--	--
219	1	1.68	--	--	91	--	--
230	3	-0.94	--	--	--	--	80
235	4	-0.08	--	--	--	--	83.6
246	3	-0.99	--	--	79.8	--	--
247	3	-0.94	--	--	80	--	--
254	2	1.06	--	--	88.4	--	--
256	2	1.44	--	--	90	--	--
259	4	0.13	--	--	84.5	--	--
265	4	0.25	--	--	85	--	--
273	0	2.75	--	--	95.5	--	--
296	0	6.92	--	--	--	--	113
304	4	0.04	--	--	--	--	84.1
326	4	-0.01	--	--	83.9	--	--
328	4	-0.46	--	--	82	--	--
356	0	3.16	--	--	97.2	--	--
384	4	-0.49	--	--	81.9	--	--


**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued**



SUMMARY	Methods				Method Codes	Statistics	
	0	3	4	6			
n =	1	5	3	24	00 Other	MPV =	8.35 µg/L
Minimum =	8.304	7.06	8.24	6.25	03 Atomic absorption: graphite furnace	F-pseudosigma =	0.519
Maximum =		9.29	15.7	9.81	04 Inductively coupled plasma	n =	33
Median =		8.38		8.30	06 Inductively coupled plasma/mass spectrometry	Uh =	8.80
F-pseudosigma =		0.259		0.556		Lh =	8.10

Lab	Rating	Z-value	Method Codes			
			0	3	4	6
1	4	-0.09	--	--	--	8.304
7	4	-0.48	--	--	--	8.1
8	4	0.29	--	--	--	8.5
16	3	-0.67	--	--	--	8
25	0	-2.41	--	--	<7.10	--
32	3	0.64	--	--	--	8.68
42	4	-0.42	--	--	--	8.13
45	4	-0.46	--	--	--	8.11
59	2	1.50	--	--	--	9.13
70	4	-0.10	--	--	--	8.3
76	4	-0.09	8.304	--	--	--
97	3	0.67	--	8.7	--	--
100	1	1.81	--	9.29	--	--
105	4	0.48	--	--	--	8.6
113	0	-2.49	--	7.06	--	--
134	4	0.00	--	8.35	--	--
138	4	-0.21	--	--	8.24	--
142	4	-0.46	--	--	--	8.11
146	0	14.16	--	--	15.7	--
149	0	-4.05	--	--	--	6.25
151	2	-1.06	--	--	--	7.8
193	4	0.06	--	8.38	--	--
212	0	11.27	--	--	14.2	--
219	0	-2.41	--	--	--	7.1
220	NR	--	--	--	<30	--
230	4	0.10	--	--	--	8.4
235	0	2.81	--	--	--	9.81
247	NR	--	--	--	<50	--
254	NR	--	--	--	<50	--
265	3	-0.67	--	--	--	8
296	3	0.87	--	--	--	8.8
304	0	2.70	--	--	--	9.75
328	3	-0.67	--	--	--	8
330	3	0.67	--	--	--	8.7
341	3	-0.87	--	--	--	7.9
349	3	0.87	--	--	--	8.8
356	3	0.98	--	--	--	8.86


**Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued**



SUMMARY	Methods			Method Codes	Statistics	
	0	4	6			
n =	1	0	17	00 Other	<b>MPV = 1.23 µg/L</b>	
Minimum =	1.33	0	0.7	04 Inductively coupled plasma	F-pseudosigma = 0.096	
Maximum =			1.5	06 Inductively coupled plasma/mass spectrometry	n = 18	
Median =			1.23		Uh = 1.33	
F-pseudosigma =			0.074		Lh = 1.20	

Lab	Rating	Z-value	Method Codes		
			0	4	6
1	4	0.00	--	--	1.23
7	4	-0.31	--	--	1.2
8	3	0.73	--	--	1.3
16	4	-0.31	--	--	1.2
32	0	2.28	--	--	1.45
42	2	1.04	--	--	1.33
45	4	0.00	--	--	1.23
70	0	2.80	--	--	1.5
76	NR	--	--	--	<5.0
142	4	-0.42	--	--	1.19
147	3	0.52	--	--	1.28
149	3	0.62	--	--	1.29
212	NR	--	--	<60	--
219	0	-2.39	--	--	1
230	4	-0.42	--	--	1.19
254	2	1.04	1.33	--	--
265	4	-0.31	--	--	1.2
296	2	1.25	--	--	1.35
328	0	-5.50	--	--	0.7
349	4	-0.31	--	--	1.2

Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued




SUMMARY	Methods			Method Codes	Statistics	
	3	4	6			
n =	0	9	18	03 Atomic absorption: graphite furnace	MPV = 2.39 µg/L	
Minimum =	0	2.3	0.99	04 Inductively coupled plasma	F-pseudosigma = 0.148	
Maximum =		5.6	3.85	06 Inductively coupled plasma/mass spectrometry	n = 27	
Median =		2.50	2.35		Uh = 2.50	
F-pseudosigma =		0.230	0.141		Lh = 2.30	

Lab	Rating	Z-value	Method Codes			Lab	Rating	Z-value	Method Codes		
			3	4	6				3	4	6
1	3	-0.74	--	--	2.281	379	3	0.74	--	2.5	--
5	NR	--	--	<4.00	--						
7	0	-9.44	--	--	0.99						
8	2	1.42	--	--	2.6						
16	3	-0.61	--	2.3	--						
25	NR	--	--	<19	--						
32	2	-1.01	--	--	2.24						
42	4	-0.40	--	--	2.33						
45	3	0.61	--	--	2.48						
59	2	-1.21	--	--	2.21						
70	NR	--	--	--	<5.0						
76	3	-0.52	--	--	2.313						
86	4	0.00	--	2.39	--						
97	NR	--	--	<2.3	--						
100	NR	--	--	<5.0	--						
105	NR	--	--	--	<20						
121	0	21.65	--	5.6	--						
134	3	0.74	--	2.5	--						
138	1	1.55	--	2.62	--						
142	3	-0.88	--	--	2.26						
146	NR	--	--	<10.0	--						
149	4	0.27	--	--	2.43						
151	1	-1.96	--	--	2.1						
193	NR	--	--	<25	--						
212	0	5.19	--	3.16	--						
219	0	2.09	--	2.7	--						
220	NR	--	--	<10	--						
230	3	0.74	--	--	2.5						
235	4	0.40	--	--	2.45						
246	NR	--	--	<5	--						
247	NR	--	--	<10	--						
254	NR	--	--	<10	--						
256	4	-0.34	--	2.34	--						
265	4	0.07	--	--	2.4						
296	0	9.85	--	--	3.85						
304	4	-0.13	--	--	2.37						
328	NR	--	--	<8	--						
341	0	-2.63	--	--	<2.0						
349	3	-0.61	--	--	2.3						
356	4	-0.07	--	--	2.38						




Table 11. Statistical summary of reported data for standard reference sample T-171 (trace constituents) -- continued



SUMMARY	Methods			Method Codes	Statistics	
	1	4	6			
n =	1	14	21	01 Atomic absorption: direct, air	MPV = 6.21 µg/L	
Minimum =	5	1.5	3.7	04 Inductively coupled plasma	F-pseudosigma = 1.108	
Maximum =		13.7	11	06 Inductively coupled plasma/mass spectrometry	n = 36	
Median =		6.08	6.63		Uh = 7.45	
F-pseudosigma =		0.593	1.105		Lh = 5.96	

Lab	Rating	Z-value	Method Codes			Lab	Rating	Z-value	Method Codes		
			1	4	6				1	4	6
1	4	0.38	--	--	6.628	259	4	-0.28	--	5.9	--
5	4	0.39	--	6.64	--	265	4	-0.19	--	--	6
7	0	-2.26	--	--	3.7	296	2	1.35	--	--	7.71
8	2	1.16	--	--	7.5	304	4	0.34	--	--	6.59
10	2	-1.09	5	--	--	326	4	-0.01	--	6.2	--
12	NR	--	--	<20	--	328	4	-0.19	--	6	--
16	0	-4.25	--	1.5	--	330	3	0.53	--	--	6.8
23	3	-0.94	--	5.17	--	341	NR	--	--	--	<10
25	2	1.25	--	7.6	--	349	0	2.07	--	--	8.5
32	2	1.19	--	--	7.53	356	4	0.01	--	--	6.22
42	4	0.40	--	--	6.65						
45	4	-0.15	--	--	6.04						
46	NR	--	--	<50	--						
59	NR	--	--	--	<10						
70	1	1.52	--	--	7.9						
86	4	-0.09	--	6.11	--						
97	0	-2.72	--	<3.2	--						
100	NR	--	--	<5.0	--						
105	0	4.32	--	--	11						
134	4	0.26	--	6.5	--						
138	4	-0.14	--	6.05	--						
142	4	-0.23	--	--	5.96						
146	NR	--	--	<20.0	--						
147	4	-0.10	--	--	6.1						
149	4	-0.10	--	--	6.1						
151	NR	--	--	--	<20						
180	4	-0.23	--	--	5.95						
190	NR	--	<12	--	--						
193	NR	--	<25	--	--						
212	0	6.76	--	13.7	--						
219	3	-0.64	--	--	5.5						
220	NR	--	--	<20	--						
227	4	-0.33	--	5.84	--						
230	1	1.62	--	--	8						
235	3	0.70	--	--	6.99						
246	2	1.07	--	7.4	--						
247	NR	--	--	<40	--						
254	3	-0.91	--	5.2	--						
255	NR	--	--	<9.0	--						
256	NR	--	--	<10.00	--						


Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents)



SUMMARY	Methods						Statistics	
	0	20	21	22	40	41	Method Codes	
n =	1	9	43	2	1	1	00 Other	MPV = 11.4 mg/L
Minimum =	12	9.5	7.84	11.4	11.5	11.61	20 Titration: colorimetric	F-pseudostigma = 0.96
Maximum =		17	20.3	12.8			21 Titration: electrometric	n = 57
Median =		11.3	11.2				22 Colorimetric	Uh = 12.0
F-pseudostigma =		1.53	0.96				40 Ion selective electrode	Lh = 10.7
							41 Electrometric	

Method Codes								Method Codes									
Lab	Rating	Z-value	0	20	21	22	40	41	Lab	Rating	Z-value	0	20	21	22	40	41
1	3	0.52	--	--	11.9	--	--	--	227	2	1.45	--	--	--	12.8	--	--
4	0	9.24	--	--	20.3	--	--	--	230	2	1.14	--	--	12.5	--	--	--
5	4	-0.21	--	--	11.2	--	--	--	247	4	-0.42	--	--	11	--	--	--
8	2	-1.25	--	--	10.2	--	--	--	256	3	0.68	--	12.06	--	--	--	--
10	4	0.10	--	--	11.5	--	--	--	259	3	-0.73	--	--	10.7	--	--	--
12	4	-0.42	--	--	11	--	--	--	266	3	0.62	--	--	12	--	--	--
16	4	0.00	--	--	--	11.4	--	--	273	4	-0.21	--	--	11.2	--	--	--
23	2	-1.45	--	10	--	--	--	--	274	4	-0.50	--	--	10.92	--	--	--
24	4	0.00	--	--	11.4	--	--	--	276	4	0.33	--	--	11.72	--	--	--
25	4	-0.42	--	--	11	--	--	--	328	3	0.62	--	12	--	--	--	--
26	4	0.10	--	--	11.5	--	--	--	330	3	-0.73	--	--	10.7	--	--	--
32	4	0.31	--	--	11.7	--	--	--	333	1	-1.65	--	--	9.81	--	--	--
33	2	-1.41	--	--	10.04	--	--	--	341	NR	--	--	--	<20	--	--	--
38	4	0.22	--	--	--	--	--	11.61	349	3	0.62	--	--	12	--	--	--
42	0	3.63	--	--	14.9	--	--	--	356	1	-1.76	--	9.7	--	--	--	--
45	3	0.62	--	--	12	--	--	--	366	3	0.93	--	12.3	--	--	--	--
46	4	-0.10	--	11.3	--	--	--	--	379	4	0.00	--	--	11.4	--	--	--
59	0	-2.14	--	--	9.34	--	--	--	386	3	0.62	12	--	--	--	--	--
70	4	-0.42	--	--	11	--	--	--									
85	0	3.22	--	--	14.5	--	--	--									
91	1	-1.97	--	9.5	--	--	--	--									
97	3	0.62	--	--	12	--	--	--									
100	2	1.25	--	--	12.6	--	--	--									
105	4	0.21	--	--	11.6	--	--	--									
109	0	2.20	--	--	13.52	--	--	--									
113	0	-3.69	--	--	7.84	--	--	--									
118	0	5.81	--	17	--	--	--	--									
134	1	1.56	--	--	12.9	--	--	--									
138	2	-1.50	--	--	9.95	--	--	--									
142	4	-0.42	--	--	11	--	--	--									
146	0	-3.45	--	--	8.08	--	--	--									
149	4	-0.42	--	--	11	--	--	--									
151	4	-0.42	--	--	11	--	--	--									
180	4	0.10	--	--	--	--	11.5	--									
183	2	-1.45	--	10	--	--	--	--									
190	4	-0.10	--	--	11.3	--	--	--									
193	0	-2.39	--	--	9.1	--	--	--									
212	1	-1.99	--	--	9.48	--	--	--									
220	0	2.02	--	--	13.35	--	--	--									
224	2	-1.45	--	--	10	--	--	--									


**Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued**



SUMMARY	Methods		Statistics
	4	6	
n =	20	7	<b>MPV = 48.2 µg/L</b>
Minimum =	31	46	F-pseudosigma = 3.63
Maximum =	144	58	n = 27
Median =	48.1	48.3	Uh = 50.8
F-pseudosigma =	6.12	1.31	Lh = 45.9

Lab	Rating	Z-value	Method Codes	
			4	6
1	4	-0.04	--	48.06
5	3	-0.99	44.6	--
8	4	0.03	--	48.3
16	0	-4.74	31	--
24	4	-0.06	48	--
25	1	-1.68	42.1	--
32	4	-0.33	--	47
42	0	-3.41	35.8	--
45	4	0.50	--	50
70	NR	--	<100	--
76	4	0.11	--	48.6
85	4	0.22	49	--
86	4	-0.14	47.7	--
100	0	-2.26	<40	--
105	NR	--	<200	--
134	4	0.00	48.2	--
138	1	-1.90	41.3	--
212	3	-0.55	46.2	--
219	0	2.15	56	--
220	3	0.83	51.2	--
247	0	6.00	70	--
254	4	0.03	48.3	--
259	3	0.61	50.4	--
265	3	-0.61	--	46
273	0	26.37	144	--
296	0	2.70	--	58
319	1	1.60	54	--
326	3	-0.66	45.8	--
328	0	-2.26	40	--
341	2	1.05	52	--


Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued



SUMMARY	Methods								Method Codes		Statistics	
	0	1	2	4	5	6	7	20	01	02	MPV =	12.7 mg/L
n =	3	7	3	38	1	4	1	3 00	Other		F-pseudsigma =	0.55
Minimum =	12.33	12.2	12.56	10.82	13.8	12.5	12.49	10.44	01 Atomic absorption: direct, air		Rating criterion =	0.64
Maximum =	13.4	13.4	12.9	15.7				12.2	02 Atomic absorption: direct, nitrous oxide		n =	60
Median =		12.8		12.7					04 Inductively coupled plasma		Uh =	13.1
F-pseudsigma =		0.45		0.59					05 Direct current plasma		Lh =	12.4
									06 Inductively coupled plasma/mass spectrometry			
									07 Ion chromatography			
									20 Titration: colorimetric			

Lab	Rating	Z-value	Method Codes								Lab	Rating	Z-value	Method Codes							
			0	1	2	4	5	6	7	20				0	1	2	4	5	6	7	20
1	3	0.60	--	--	--	13.08	--	--	--	--	254	4	0.31	--	--	--	12.9	--	--	--	--
5	3	-0.79	--	--	--	12.2	--	--	--	--	255	4	0.16	--	--	--	12.8	--	--	--	--
8	4	-0.31	--	--	--	12.5	--	--	--	--	256	4	-0.33	--	--	--	--	--	--	12.49	--
10	4	0.16	--	12.8	--	--	--	--	--	--	257	2	-1.10	--	--	--	--	--	--	--	12
16	2	-1.10	--	--	--	12	--	--	--	--	259	4	-0.16	--	--	--	12.6	--	--	--	--
23	4	-0.47	--	12.4	--	--	--	--	--	--	265	4	-0.16	--	--	--	12.6	--	--	--	--
24	3	-0.94	--	--	--	12.1	--	--	--	--	266	2	1.27	13.4	--	--	--	--	--	--	--
25	0	-2.96	--	--	--	10.82	--	--	--	--	273	1	1.89	--	--	--	13.9	--	--	--	--
26	3	0.57	--	--	--	13.06	--	--	--	--	274	0	-3.56	--	--	--	--	--	--	--	10.44
32	2	1.26	--	--	--	--	--	13.5	--	--	276	4	0.00	--	--	12.7	--	--	--	--	--
33	1	1.73	--	--	--	--	13.8	--	--	--	279	3	-0.58	12.33	--	--	--	--	--	--	--
38	4	-0.22	--	--	12.56	--	--	--	--	--	296	2	1.26	--	--	--	--	13.5	--	--	--
42	2	-1.42	--	--	--	11.8	--	--	--	--	326	4	-0.09	--	--	--	12.64	--	--	--	--
45	4	-0.31	--	--	--	--	--	12.5	--	--	328	3	-0.79	--	--	--	12.2	--	--	--	--
46	4	-0.31	--	--	--	12.5	--	--	--	--	341	3	-0.79	--	12.2	--	--	--	--	--	--
59	2	1.10	--	13.4	--	--	--	--	--	--	366	3	-0.63	--	--	--	12.3	--	--	--	--
64	4	0.38	--	--	--	12.94	--	--	--	--	379	3	-0.79	--	--	--	--	--	--	--	12.2
70	4	0.00	--	--	--	12.7	--	--	--	--	384	3	0.89	--	--	--	13.27	--	--	--	--
86	4	-0.16	--	--	--	12.6	--	--	--	--	385	1	1.57	--	--	--	13.7	--	--	--	--
97	3	0.63	--	--	--	13.1	--	--	--	--	386	4	-0.47	12.4	--	--	--	--	--	--	--
100	2	1.10	--	--	--	13.4	--	--	--	--											
102	0	2.05	--	--	--	14	--	--	--	--											
105	3	0.63	--	--	--	13.1	--	--	--	--											
109	4	-0.50	--	12.38	--	--	--	--	--	--											
113	4	0.00	--	--	--	12.7	--	--	--	--											
121	3	-0.94	--	--	--	12.1	--	--	--	--											
134	3	0.96	--	--	--	13.31	--	--	--	--											
138	3	0.63	--	--	--	13.1	--	--	--	--											
146	4	0.00	--	--	--	12.7	--	--	--	--											
151	4	0.16	--	12.8	--	--	--	--	--	--											
180	4	0.16	--	--	--	12.8	--	--	--	--											
190	4	0.31	--	--	12.9	--	--	--	--	--											
193	3	0.79	--	13.2	--	--	--	--	--	--											
212	2	-1.26	--	--	--	11.9	--	--	--	--											
219	0	4.72	--	--	--	15.7	--	--	--	--											
220	4	-0.49	--	--	--	12.39	--	--	--	--											
224	3	-0.76	--	--	--	12.22	--	--	--	--											
227	4	0.16	--	--	--	12.8	--	--	--	--											
230	3	0.79	--	--	--	--	--	13.2	--	--											
247	4	-0.31	--	--	--	12.5	--	--	--	--											


Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued



SUMMARY	Methods					Statistics	
	0	7	20	21	22	Method Codes	
n =	3	46	7	5	8	00 Other	MPV = 23.2 mg/L
Minimum =	22.9	4.5	17.96	22.45	18.6	07 Ion chromatography	F-pseudosigma = 0.89
Maximum =	24	25.2	28	27	25	20 Titration: colorimetric	Rating criterion = 1.16
Median =	23.2	24.0	23.3	23.3		21 Titration: electrometric	n = 69
F-pseudosigma =	0.82	2.45	2.85	0.82		22 Colorimetric	Uh = 23.8
							Lh = 22.6

Method Codes						Method Codes					
Lab	Rating	Z-value	0	7	22	Lab	Rating	Z-value	0	7	22
1	3	-0.69	--	22.4	--	219	1	-1.90	--	21	--
4	4	0.34	--	23.6	--	220	2	1.45	--	24.88	--
5	3	-0.60	--	22.5	--	224	3	-0.68	--	22.41	--
8	3	-0.52	--	22.6	--	227	3	-0.95	--	22.1	--
10	4	-0.34	--	22.8	--	230	4	-0.17	--	23	--
12	2	-1.03	--	--	22	247	4	0.17	--	23.4	--
16	0	-3.97	--	--	18.6	254	4	-0.17	--	23	--
23	4	0.17	--	23.4	--	256	4	-0.36	--	22.78	--
24	4	0.00	--	--	23.2	257	0	3.28	--	--	27
25	1	-1.64	--	21.3	--	259	4	-0.43	--	--	22.7
26	4	0.31	--	23.56	--	265	3	-0.95	--	22.1	--
32	1	1.72	--	25.2	--	266	4	0.09	--	--	23.3
33	3	0.52	--	23.8	--	273	4	-0.22	22.94	--	--
42	4	0.00	--	23.2	--	274	0	-4.52	--	17.96	--
45	4	0.00	--	23.2	--	276	0	4.14	--	28	--
46	4	0.43	--	--	23.7	307	2	-1.03	--	22	--
59	3	-0.69	--	22.4	--	319	3	0.69	24	--	--
64	3	0.60	--	23.9	--	326	3	-0.65	--	--	22.45
70	3	0.69	--	24	--	328	2	-1.03	--	22	--
76	4	0.00	--	23.2	--	330	3	-0.52	--	22.6	--
85	3	0.60	--	23.9	--	341	4	-0.17	--	--	23
86	3	0.52	--	23.8	--	349	1	1.55	--	25	--
91	4	-0.17	--	23	--	356	4	-0.34	--	22.8	--
97	4	0.26	--	--	23.5	366	4	0.09	--	--	23.3
100	4	0.17	--	23.4	--	374	3	0.69	--	24	--
102	4	0.34	--	23.6	--	379	1	1.55	--	--	25
105	4	0.26	--	23.5	--	384	0	-2.37	--	20.45	--
109	0	2.89	--	--	26.55	386	4	-0.26	22.9	--	--
113	3	0.52	--	23.8	--	387	0	-16.12	--	4.5	--
134	4	0.17	--	23.4	--						
138	4	0.09	--	23.3	--						
142	4	0.00	--	23.2	--						
146	3	-0.60	--	22.5	--						
149	3	0.69	--	24	--						
151	4	-0.17	--	23	--						
180	1	1.72	--	25.2	--						
183	0	2.59	--	--	26.2						
190	4	0.26	--	23.5	--						
208	4	-0.37	--	22.77	--						
212	4	-0.26	--	22.9	--						


**Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued**



SUMMARY	Methods						Statistics	
	0	7	20	22	40	41	Method Codes	
n =	1	27	1	2	22	1	00 Other	<b>MPV = 0.604 mg/L</b>
Minimum =	0.605	0.484	0.63	0.591	0.47	0.69	07 Ion chromatography	F-pseudosigma = 0.0497
Maximum =		0.78		0.85	3.19		20 Titration: colorimetric	n = 54
Median =		0.600			0.607		22 Colorimetric	Uh = 0.630
F-pseudosigma =		0.052			0.037		40 Ion selective electrode	Lh = 0.563
							41 Electrometric	

Lab	Rating	Z-value	Method Codes						Lab	Rating	Z-value	Method Codes					
			0	7	20	22	40	41				0	7	20	22	40	41
1	3	0.62	--	--	--	--	0.635	--	256	1	-1.89	--	0.51	--	--	--	--
5	3	-0.89	--	0.56	--	--	--	--	257	2	-1.09	--	--	--	--	0.55	--
8	1	1.53	--	0.68	--	--	--	--	259	3	0.52	--	--	--	--	0.63	--
10	0	-2.70	--	--	--	--	0.47	--	265	3	0.52	--	0.63	--	--	--	--
16	3	0.52	--	0.63	--	--	--	--	266	4	-0.48	--	--	--	--	0.58	--
23	4	-0.42	--	0.583	--	--	--	--	273	0	52.07	--	--	--	--	3.19	--
24	4	-0.22	--	--	--	--	0.593	--	274	0	4.95	--	--	--	0.85	--	--
25	4	-0.48	--	0.58	--	--	--	--	328	4	-0.28	--	--	--	--	0.59	--
26	0	-2.42	--	0.484	--	--	--	--	330	3	0.52	--	--	0.63	--	--	--
32	3	-0.68	--	0.57	--	--	--	--	349	4	-0.28	--	--	--	--	0.59	--
33	3	-0.83	--	0.563	--	--	--	--	356	4	0.32	--	0.62	--	--	--	--
42	4	0.00	--	0.604	--	--	--	--	379	1	1.73	--	--	--	--	--	0.69
45	2	-1.39	--	0.535	--	--	--	--	384	3	-0.77	--	0.566	--	--	--	--
46	4	-0.26	--	--	--	0.591	--	--	386	4	0.02	0.605	--	--	--	--	--
59	1	-1.69	--	--	--	--	0.52	--									
70	4	0.12	--	--	--	--	0.61	--									
85	4	0.32	--	--	--	--	0.62	--									
86	3	0.66	--	0.637	--	--	--	--									
97	2	1.05	--	--	--	--	0.656	--									
100	3	0.60	--	--	--	--	0.634	--									
102	3	-0.89	--	0.56	--	--	--	--									
105	1	1.73	--	0.69	--	--	--	--									
109	3	0.52	--	--	--	--	0.63	--									
113	3	0.52	--	0.63	--	--	--	--									
134	4	0.12	--	--	--	--	0.61	--									
138	3	-0.87	--	--	--	--	0.561	--									
142	4	0.00	--	--	--	--	0.604	--									
146	1	-1.83	--	0.513	--	--	--	--									
149	4	-0.08	--	0.6	--	--	--	--									
151	4	0.48	--	0.628	--	--	--	--									
180	3	-0.97	--	0.556	--	--	--	--									
183	3	0.60	--	--	--	--	0.634	--									
190	4	-0.08	--	--	--	--	0.6	--									
212	4	0.48	--	--	--	--	0.628	--									
219	4	-0.08	--	0.6	--	--	--	--									
220	4	0.12	--	0.61	--	--	--	--									
224	0	3.54	--	0.78	--	--	--	--									
230	3	0.72	--	0.64	--	--	--	--									
247	2	-1.09	--	0.55	--	--	--	--									
255	1	-1.69	--	--	--	--	0.52	--									


Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued



SUMMARY	Methods								Statistics	
	0	1	4	5	6	7	12	Method Codes		
n =	2	15	31	1	5	1	3	00 Other	MPV =	3.03 mg/L
Minimum =	3.25	2.95	2.402	3.14	2.93	3.07	2.8	01 Atomic absorption: direct, air	F-pseudostigma =	0.126
Maximum =	3.62	3.31	3.7		3.2		3.26	04 Inductively coupled plasma	Rating criterion =	0.152
Median =		3.03	3.00		3.10			05 Direct current plasma	n =	58
F-pseudostigma =		0.085	0.152		0.094			06 Inductively coupled plasma/mass spectrometry	Uh =	3.12
								07 Ion chromatography	Lh =	2.95
								12 Flame emission		

Lab	Rating	Z-value	Method Codes								Lab	Rating	Z-value	Method Codes							
			0	1	4	5	6	7	12	0				1	4	5	6	7	12		
1	4	-0.46	--	2.96	--	--	--	--	--	256	4	0.26	--	--	--	--	--	3.07	--		
5	1	1.52	--	--	3.26	--	--	--	--	257	3	-0.86	--	--	--	--	--	--	2.9		
8	4	-0.20	--	--	3	--	--	--	--	259	4	0.20	--	--	3.06	--	--	--	--		
10	4	-0.20	--	3	--	--	--	--	--	265	3	-0.86	--	--	2.9	--	--	--	--		
16	4	0.46	--	--	3.1	--	--	--	--	266	1	1.52	--	--	--	--	--	--	3.26		
23	4	0.26	--	3.07	--	--	--	--	--	273	1	1.72	--	--	3.29	--	--	--	--		
24	4	-0.20	--	--	3	--	--	--	--	274	1	-1.52	--	--	--	--	--	--	2.8		
25	0	-4.15	--	--	2.402	--	--	--	--	276	3	-0.53	--	2.95	--	--	--	--	--		
26	3	-0.79	--	--	2.91	--	--	--	--	279	0	3.89	3.62	--	--	--	--	--	--		
32	2	1.12	--	--	--	--	3.2	--	--	296	4	0.46	--	--	--	--	3.1	--	--		
33	3	0.73	--	--	--	3.14	--	--	--	326	4	0.20	--	3.06	--	--	--	--	--		
38	3	0.59	--	3.12	--	--	--	--	--	328	1	-1.91	--	--	2.74	--	--	--	--		
42	3	-0.73	--	--	2.92	--	--	--	--	341	4	-0.20	--	3	--	--	--	--	--		
45	3	-0.66	--	--	--	--	2.93	--	--	366	4	-0.46	--	--	2.96	--	--	--	--		
46	3	-0.86	--	--	2.9	--	--	--	--	379	0	3.76	--	--	3.6	--	--	--	--		
59	4	-0.33	--	2.98	--	--	--	--	--	384	0	2.24	--	--	3.37	--	--	--	--		
64	4	0.00	--	3.03	--	--	--	--	--	385	2	1.12	--	--	3.2	--	--	--	--		
70	4	0.00	--	--	3.03	--	--	--	--	386	2	1.45	3.25	--	--	--	--	--	--		
76	4	-0.38	--	--	--	--	2.973	--	--												
86	3	0.53	--	--	3.11	--	--	--	--												
97	4	0.00	--	3.03	--	--	--	--	--												
100	4	0.20	--	--	3.06	--	--	--	--												
102	0	-3.50	--	--	2.5	--	--	--	--												
105	4	-0.13	--	--	3.01	--	--	--	--												
109	1	1.85	--	3.31	--	--	--	--	--												
113	4	0.00	--	--	3.03	--	--	--	--												
134	4	-0.46	--	2.96	--	--	--	--	--												
138	4	-0.20	--	--	3	--	--	--	--												
146	3	0.59	--	--	3.12	--	--	--	--												
151	4	0.46	--	3.1	--	--	--	--	--												
180	4	0.26	--	--	3.07	--	--	--	--												
190	4	-0.46	--	2.96	--	--	--	--	--												
193	3	0.99	--	3.18	--	--	--	--	--												
212	2	-1.25	--	--	2.84	--	--	--	--												
219	0	4.42	--	--	3.7	--	--	--	--												
220	1	-1.85	--	--	2.75	--	--	--	--												
224	3	-0.92	--	--	2.89	--	--	--	--												
230	4	0.46	--	--	--	--	3.1	--	--												
247	3	-0.92	--	--	2.89	--	--	--	--												
254	4	-0.20	--	--	3	--	--	--	--												

Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued




SUMMARY	Methods							Statistics	
	0	1	2	4	5	6	20	Method Codes	
n =	3	9	1	39	1	5	1	00 Other	MPV = 2.78 mg/L
Minimum =	2.4	2.5	2.95	2.5	3.11	2.7	3.38	01 Atomic absorption: direct, air	F-pseudosigma = 0.104
Maximum =	3.01	3		30.1		2.82		02 Atomic absorption: direct, nitrous oxide	Rating criterion = 0.139
Median =		2.80		2.73		2.73		04 Inductively coupled plasma	n = 59
F-pseudosigma =		0.119		0.082		0.067		05 Direct current plasma	Uh = 2.84
								06 Inductively coupled plasma/mass spectrometry	Lh = 2.70
								20 Titration: colorimetric	

Lab	Rating	Z-value	Method Codes							Lab	Rating	Z-value	Method Codes						
			0	1	2	4	5	6	20				0	1	2	4	5	6	20
1	4	0.27	--	--	--	2.817	--	--	--	247	4	-0.07	--	--	--	2.77	--	--	--
5	2	-1.29	--	--	--	2.6	--	--	--	254	3	0.58	--	--	--	2.86	--	--	--
8	3	-0.58	--	--	--	2.7	--	--	--	255	4	0.00	--	--	--	2.78	--	--	--
10	4	0.14	--	2.8	--	--	--	--	--	259	4	-0.36	--	--	--	2.73	--	--	--
16	3	-0.58	--	--	--	2.7	--	--	--	265	3	-0.58	--	--	--	2.7	--	--	--
23	2	1.29	--	2.96	--	--	--	--	--	266	0	-2.73	2.4	--	--	--	--	--	--
24	0	-2.01	--	--	--	2.5	--	--	--	273	0	196.55	--	--	--	30.1	--	--	--
25	4	0.19	--	--	--	2.806	--	--	--	274	0	4.32	--	--	--	--	--	--	3.38
26	4	0.14	--	--	--	2.8	--	--	--	276	2	1.22	--	--	2.95	--	--	--	--
32	4	0.29	--	--	--	--	--	2.82	--	279	1	1.65	3.01	--	--	--	--	--	--
33	0	2.37	--	--	--	--	3.11	--	--	296	4	0.14	--	--	--	--	--	2.8	--
38	2	1.34	--	2.966	--	--	--	--	--	326	4	-0.22	--	--	--	2.75	--	--	--
42	3	-0.86	--	--	--	2.66	--	--	--	328	2	-1.15	--	--	--	2.62	--	--	--
45	4	-0.50	--	--	--	--	--	2.71	--	341	1	1.58	--	3	--	--	--	--	--
46	3	-0.58	--	--	--	2.7	--	--	--	366	2	-1.29	--	--	--	2.6	--	--	--
59	0	-2.01	--	2.5	--	--	--	--	--	379	3	-0.58	--	--	--	2.7	--	--	--
64	3	-0.65	--	--	--	2.69	--	--	--	384	2	1.49	--	--	--	2.987	--	--	--
70	4	0.36	--	--	--	2.83	--	--	--	385	3	0.86	--	--	--	2.9	--	--	--
76	4	-0.33	--	--	--	--	--	2.734	--	386	3	-0.58	2.7	--	--	--	--	--	--
86	4	0.00	--	--	--	2.78	--	--	--										
97	4	-0.36	--	--	--	2.73	--	--	--										
100	3	0.79	--	--	--	2.89	--	--	--										
102	0	2.30	--	--	--	3.1	--	--	--										
105	4	0.22	--	--	--	2.81	--	--	--										
109	4	0.14	--	2.8	--	--	--	--	--										
113	4	-0.36	--	--	--	2.73	--	--	--										
121	2	-1.29	--	--	--	2.6	--	--	--										
134	4	0.00	--	--	--	2.78	--	--	--										
138	4	0.22	--	--	--	2.81	--	--	--										
146	4	0.00	--	--	--	2.78	--	--	--										
151	4	0.14	--	2.8	--	--	--	--	--										
180	4	-0.43	--	--	--	2.72	--	--	--										
190	4	0.50	--	2.85	--	--	--	--	--										
193	2	-1.44	--	2.58	--	--	--	--	--										
212	1	-1.65	--	--	--	2.55	--	--	--										
219	0	4.46	--	--	--	3.4	--	--	--										
220	4	-0.50	--	--	--	2.71	--	--	--										
224	3	-0.94	--	--	--	2.65	--	--	--										
227	4	-0.43	--	--	--	2.72	--	--	--										
230	3	-0.58	--	--	--	--	--	2.7	--										




Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued



SUMMARY	Methods						Statistics	
	0	1	4	5	6	12	Method Codes	
n =	2	13	34	1	5	4	00 Other	MPV = 37.0 mg/L
Minimum =	35.5	6.18	30	34.03	35.7	31.97	01 Atomic absorption: direct, air	F-pseudostigma = 1.67
Maximum =	37.1	38.3	43.6		40.5	37.5	04 Inductively coupled plasma	Rating criterion = 1.85
Median =		37.2	37.0		38.8		05 Direct current plasma	n = 59
F-pseudostigma =		1.09	1.56		1.44		06 Inductively coupled plasma/mass spectrometry	Uh = 38.1
							12 Flame emission	Lh = 35.8

Lab	Rating	Z-value	Method Codes							Lab	Rating	Z-value	Method Codes						
			0	1	4	5	6	12	0				1	4	5	6	12		
1	4	0.17	--	--	37.31	--	--	--	254	4	0.16	--	--	37.3	--	--	--		
5	4	0.27	--	--	37.5	--	--	--	256	2	-1.24	--	--	--	--	--	34.7		
8	4	-0.16	--	--	36.7	--	--	--	257	2	-1.35	--	--	--	--	--	34.5		
10	4	0.11	--	37.2	--	--	--	--	259	4	0.32	--	--	37.6	--	--	--		
16	3	-0.54	--	--	36	--	--	--	265	4	-0.27	--	--	36.5	--	--	--		
23	0	-16.66	--	6.18	--	--	--	--	266	4	0.27	--	--	--	--	--	37.5		
24	4	-0.38	--	--	36.3	--	--	--	273	1	1.95	--	--	40.6	--	--	--		
25	0	-2.70	--	--	32.01	--	--	--	274	0	-2.72	--	--	--	--	--	31.97		
26	4	0.02	--	--	37.04	--	--	--	276	3	0.70	--	38.3	--	--	--	--		
32	1	1.89	--	--	--	--	40.5	--	279	3	-0.81	35.5	--	--	--	--	--		
33	1	-1.61	--	--	--	34.03	--	--	296	3	0.97	--	--	--	--	38.8	--		
38	4	-0.29	--	36.47	--	--	--	--	326	4	-0.21	--	--	36.61	--	--	--		
42	3	-0.97	--	--	35.2	--	--	--	328	2	-1.08	--	--	35	--	--	--		
45	3	-0.70	--	--	--	--	35.7	--	341	3	0.54	--	38	--	--	--	--		
46	4	0.32	--	--	37.6	--	--	--	366	2	-1.41	--	--	34.4	--	--	--		
59	4	0.22	--	37.4	--	--	--	--	379	2	1.35	--	--	39.5	--	--	--		
64	3	0.65	--	38.2	--	--	--	--	384	3	-0.53	--	--	36.02	--	--	--		
70	3	0.70	--	--	38.3	--	--	--	385	3	0.81	--	--	38.5	--	--	--		
76	4	0.03	--	--	--	--	37.06	--	386	4	0.05	37.1	--	--	--	--	--		
86	3	0.59	--	--	38.1	--	--	--											
97	3	0.59	--	38.1	--	--	--	--											
100	4	0.43	--	--	37.8	--	--	--											
102	0	-3.78	--	--	30	--	--	--											
105	3	0.70	--	--	38.3	--	--	--											
109	4	-0.01	--	36.99	--	--	--	--											
113	4	-0.05	--	--	36.9	--	--	--											
121	1	-1.84	--	--	33.6	--	--	--											
134	4	-0.25	--	36.53	--	--	--	--											
138	3	0.81	--	--	38.5	--	--	--											
146	3	0.70	--	--	38.3	--	--	--											
151	4	0.00	--	37	--	--	--	--											
180	4	-0.16	--	--	36.7	--	--	--											
190	4	0.11	--	37.2	--	--	--	--											
193	3	-0.59	--	35.9	--	--	--	--											
212	3	-0.76	--	--	35.6	--	--	--											
219	0	3.57	--	--	43.6	--	--	--											
220	4	-0.45	--	--	36.16	--	--	--											
224	2	-1.39	--	--	34.42	--	--	--											
230	2	1.08	--	--	--	--	39	--											
247	4	0.00	--	--	37	--	--	--											


Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued



SUMMARY	Methods			Method Codes		Statistics	
	0	40	41				
n =	2	11	41	00 Other		<b>MPV = 7.06</b>	
Minimum =	6.87	5.4	6.17	40 Ion selective electrode		F-pseudostigma = 0.297	
Maximum =	7.06	7.31	7.45	41 Electrometric		Rating criterion = 0.353	
Median =		7.06	7.05			n = 54	
F-pseudostigma =		0.245	0.297			Uh = 7.20	
						Lh = 6.80	

Method Codes						Method Codes					
Lab	Rating	Z-value	0	40	41	Lab	Rating	Z-value	0	40	41
1	0	-2.51	--	--	6.17	266	4	0.30	--	--	7.16
5	2	-1.37	--	6.57	--	273	4	-0.07	--	--	7.03
8	2	1.12	--	--	7.45	274	4	-0.01	--	--	7.05
12	3	-0.72	--	--	6.8	276	3	0.52	--	--	7.24
16	4	0.13	--	--	7.1	307	4	0.07	--	--	7.08
23	3	0.69	--	--	7.3	326	4	0.33	--	--	7.17
24	2	-1.29	--	--	6.6	328	3	-0.52	--	--	6.87
25	3	-0.78	--	--	6.78	330	3	0.72	--	7.31	--
26	0	-4.69	--	5.4	--	333	4	0.41	--	--	7.2
32	4	-0.33	--	--	6.94	341	4	0.01	--	7.06	--
33	4	0.21	--	--	7.13	349	3	-0.98	--	6.71	--
38	4	0.13	--	7.1	--	366	3	0.78	--	--	7.33
42	2	-1.12	--	--	6.66	379	4	0.41	--	--	7.2
45	4	0.13	--	--	7.1	386	3	-0.52	6.87	--	--
46	4	-0.16	--	--	7						
59	4	0.41	--	--	7.2						
64	3	0.64	--	--	7.28						
85	2	-1.40	--	--	6.56						
86	3	-0.64	--	--	6.83						
91	4	-0.44	--	--	6.9						
97	4	0.21	--	7.13	--						
100	4	0.47	--	--	7.22						
105	1	-1.57	--	--	6.5						
109	3	-1.01	--	--	6.7						
118	3	-0.72	--	--	6.8						
134	4	0.50	--	7.232	--						
138	4	0.50	--	--	7.23						
142	0	-2.23	--	--	6.27						
146	4	0.21	--	--	7.13						
149	4	0.41	--	--	7.2						
151	4	-0.01	--	--	7.05						
180	4	0.04	--	7.07	--						
190	3	-0.55	--	6.86	--						
193	4	0.01	--	7.06	--						
212	4	-0.10	--	--	7.02						
224	3	-0.95	--	--	6.72						
230	3	-1.01	--	--	6.7						
247	3	0.55	--	--	7.25						
256	4	-0.30	--	--	6.95						
257	4	0.01	7.06	--	--						


**Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued**



SUMMARY	Methods				Statistics	
	0	4	40	50	Method Codes	
n =	4	1	1	40	00 Other	MPV = 198 mg/L
Minimum =	157	180	212	162	04 Inductively coupled plasma	F-pseudosigma = 16.3
Maximum =	232			2200	40 Ion selective electrode	n = 46
Median =				199	50 Gravimetric	Uh = 210
F-pseudosigma =				15.6		Lh = 188

Lab	Rating	Z-value	Method Codes				Lab	Rating	Z-value	Method Codes			
			0	4	40	50				0	4	40	50
1	4	0.00	--	--	--	198	341	4	0.12	--	--	--	200
5	1	-1.66	--	--	--	171	349	4	-0.49	--	--	--	190
8	4	0.31	--	--	--	203	356	4	0.18	--	--	--	201
10	3	-0.61	--	--	--	188	366	3	0.74	--	--	--	210
12	0	-2.21	--	--	--	162	379	4	0.37	--	--	--	204
16	4	0.31	--	--	--	203	386	0	2.08	232	--	--	--
23	3	-0.61	--	--	--	188							
25	3	-0.74	--	--	--	186							
26	3	-0.98	--	--	--	182							
32	2	1.35	--	--	--	220							
38	4	-0.37	--	--	--	192							
45	3	0.74	--	--	--	210							
46	4	0.37	--	--	--	204							
59	4	0.37	--	--	--	204							
70	0	122.76	--	--	--	2200							
85	3	-0.74	--	--	--	186							
97	4	0.00	--	--	--	198							
100	4	-0.18	--	--	--	195							
105	3	-0.74	--	--	--	186							
109	3	0.86	--	--	--	212							
113	4	-0.31	--	--	--	193							
118	2	1.35	--	--	--	220							
134	4	-0.03	--	--	--	197.5							
138	4	-0.18	--	--	--	195							
142	2	1.35	--	--	--	220							
146	3	0.74	--	--	--	210							
149	3	-0.80	--	--	--	185							
151	4	-0.43	--	--	--	191							
183	3	0.86	--	--	212	--							
190	4	0.31	--	--	--	203							
212	1	-1.90	--	--	--	167							
224	4	0.12	--	--	--	200							
227	3	0.98	--	--	--	214							
247	2	-1.10	--	180	--	--							
257	2	-1.10	180	--	--	--							
266	4	-0.12	196	--	--	--							
273	0	-2.51	157	--	--	--							
276	4	0.00	--	--	--	198							
277	1	1.76	--	--	--	226.7							
328	3	0.98	--	--	--	214							


Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued



SUMMARY	Methods					Statistics	
	0	4	5	6	22	Method Codes	
n =	1	19	1	3	18	00 Other	MPV = 8.10 mg/L
Minimum =	8.13	3.933	8.62	8.01	3.16	04 Inductively coupled plasma	F-pseudosigma = 0.371
Maximum =		9.8		8.3	12	05 Direct current plasma	Rating criterion = 0.405
Median =		7.85			8.05	06 Inductively coupled plasma/mass spectrometry	n = 42
F-pseudosigma =		0.511			0.519	22 Colorimetric	Uh = 8.30
							Lh = 7.80

Lab	Rating	Z-value	Method Codes					Lab	Rating	Z-value	Method Codes				
			0	4	5	6	22				0	4	5	6	22
1	0	9.38	--	--	--	--	11.9	384	0	-10.29	--	3.933	--	--	--
5	3	-0.62	--	7.85	--	--	--	386	4	0.07	8.13	--	--	--	
8	3	-0.74	--	7.8	--	--	--								
10	4	-0.49	--	--	--	--	7.9								
23	4	0.12	--	--	--	--	8.15								
24	3	0.79	--	8.42	--	--	--								
25	0	-4.00	--	6.478	--	--	--								
32	4	0.00	--	--	--	8.1	--								
33	2	1.28	--	--	8.62	--	--								
38	4	-0.42	--	--	--	--	7.93								
42	0	-2.02	--	7.28	--	--	--								
45	4	-0.22	--	--	--	8.01	--								
64	4	0.00	--	8.1	--	--	--								
70	4	-0.44	--	--	--	--	7.92								
85	2	1.23	--	--	--	--	8.6								
97	0	9.63	--	--	--	--	12								
100	0	3.33	--	9.45	--	--	--								
102	3	-0.74	--	--	--	--	7.8								
105	4	0.13	--	8.153	--	--	--								
113	3	-0.84	--	--	--	--	7.76								
118	4	-0.44	--	--	--	--	7.92								
121	0	-2.47	--	7.1	--	--	--								
134	3	-0.77	--	7.79	--	--	--								
138	4	0.47	--	--	--	--	8.29								
151	3	-0.62	--	--	--	--	7.85								
190	4	0.25	--	--	--	--	8.2								
193	4	0.10	--	--	--	--	8.14								
212	2	-1.04	--	7.68	--	--	--								
219	0	4.20	--	9.8	--	--	--								
224	0	2.12	--	--	--	--	8.96								
230	4	0.49	--	--	--	8.3	--								
247	0	2.37	--	--	--	--	9.06								
254	3	-0.84	--	7.76	--	--	--								
256	4	0.27	--	8.21	--	--	--								
259	4	0.00	--	8.1	--	--	--								
265	1	-1.73	--	7.4	--	--	--								
266	4	-0.37	--	--	--	--	7.95								
273	0	2.99	--	9.31	--	--	--								
274	0	-12.20	--	--	--	--	3.16								
328	4	0.37	--	8.25	--	--	--								


Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued



SUMMARY	Methods						Statistics	
	0	6	7	22	50	51	Method Codes	
n =	2	1	47	7	1	5	00 Other	MPV = 11.2 mg/L
Minimum =	11.8	11.7	5.4	7.58	12.75	9.27	06 Inductively coupled plasma/mass spectrometry	F-pseudosigma = 0.55
Maximum =	12.1		13	12.2		11.7	07 Ion chromatography	Rating criterion = 0.56
Median =			11.2	10.6		11.2	22 Colorimetric	n = 63
F-pseudosigma =			0.53	1.37		0.33	50 Gravimetric	Uh = 11.5
							51 Turbidimetric	Lh = 10.8

Lab	Rating	Z-value	Method Codes						Lab	Rating	Z-value	Method Codes					
			0	6	7	22	50	51				0	6	7	22	50	51
1	3	-0.71	--	--	10.8	--	--	--	220	0	2.57	--	--	12.64	--	--	--
4	0	-2.50	--	--	9.8	--	--	--	224	4	0.27	--	--	11.35	--	--	--
5	0	-2.16	--	--	9.99	--	--	--	230	4	-0.36	--	--	11	--	--	--
8	4	-0.18	--	--	11.1	--	--	--	247	4	0.00	--	--	11.2	--	--	--
10	3	0.54	--	--	11.5	--	--	--	254	4	-0.36	--	--	11	--	--	--
12	0	-14.64	--	--	--	<3	--	--	255	4	-0.36	--	--	11	--	--	--
16	4	0.00	--	--	--	--	--	11.2	256	4	-0.05	--	--	11.17	--	--	--
23	4	0.36	--	--	11.4	--	--	--	257	3	0.54	--	--	11.5	--	--	--
24	0	-2.50	--	--	--	9.8	--	--	259	4	0.36	--	--	11.4	--	--	--
25	0	-2.14	--	--	10	--	--	--	265	2	1.43	--	--	12	--	--	--
26	2	-1.05	--	--	10.61	--	--	--	266	4	-0.36	--	--	--	--	--	11
32	3	0.89	--	11.7	--	--	--	--	273	2	1.07	11.8	--	--	--	--	--
33	3	0.54	--	--	11.5	--	--	--	274	0	-3.45	--	--	--	--	--	9.27
42	0	2.14	--	--	12.4	--	--	--	307	4	0.45	--	--	--	--	--	11.45
45	1	-1.61	--	--	10.3	--	--	--	326	4	-0.07	--	--	11.16	--	--	--
59	4	-0.36	--	--	11	--	--	--	328	2	1.43	--	--	12	--	--	--
64	3	0.54	--	--	11.5	--	--	--	330	2	-1.07	--	--	10.6	--	--	--
70	2	1.07	--	--	11.8	--	--	--	341	0	-3.93	--	--	9	--	--	--
76	4	0.30	--	--	11.37	--	--	--	349	0	-2.14	--	--	<10	--	--	--
85	2	1.25	--	--	11.9	--	--	--	356	4	-0.18	--	--	11.1	--	--	--
86	3	0.54	--	--	11.5	--	--	--	366	1	1.79	--	--	12.2	--	--	--
91	0	3.21	--	--	13	--	--	--	379	3	0.89	--	--	--	--	--	11.7
97	0	-6.46	--	--	--	7.58	--	--	384	2	1.02	--	--	11.77	--	--	--
100	4	0.36	--	--	11.4	--	--	--	386	1	1.61	12.1	--	--	--	--	--
102	4	0.36	--	--	11.4	--	--	--	387	0	-10.36	--	--	5.4	--	--	--
105	3	0.54	--	--	11.5	--	--	--									
109	0	2.77	--	--	--	--	12.75	--									
113	4	-0.36	--	--	11	--	--	--									
134	3	-0.80	--	--	10.75	--	--	--									
138	4	0.00	--	--	11.2	--	--	--									
142	3	-0.54	--	--	10.9	--	--	--									
146	2	-1.43	--	--	10.4	--	--	--									
149	0	-2.14	--	--	10	--	--	--									
151	2	-1.43	--	--	10.4	--	--	--									
180	4	-0.36	--	--	11	--	--	--									
183	3	0.54	--	--	--	11.5	--	--									
190	4	0.00	--	--	11.2	--	--	--									
208	3	-0.77	--	--	10.77	--	--	--									
212	2	-1.25	--	--	10.5	--	--	--									
219	4	-0.36	--	--	11	--	--	--									


Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued



SUMMARY	Methods			Method Codes		Statistics	
	0	40	41				
n =	2	6	54	00 Other		<b>MPV =</b>	<b>301 µS/cm</b>
Minimum =	301	278	247	40 Ion selective electrode		F-pseudosigma =	6.7
Maximum =	312	313.5	845	41 Electrometric		Rating criterion =	15.0
Median =		299	301			n =	62
F-pseudosigma =		3.7	6.7			Uh =	304
						Lh =	295

Lab	Rating	Z-value	Method Codes			Lab	Rating	Z-value	Method Codes		
			0	40	41				0	40	41
1	2	-1.23	--	--	282	212	0	-2.90	--	--	257
5	4	0.10	--	--	302	224	3	0.57	--	--	309
8	4	-0.03	--	--	300	227	0	36.24	--	--	845
10	4	-0.03	--	--	300	230	4	-0.10	--	--	299
12	4	-0.17	--	--	298	247	4	0.03	--	--	301
16	4	0.10	--	302	--	256	4	0.10	--	--	302
23	4	-0.37	--	--	295	257	2	1.10	--	--	317
24	3	-0.70	--	--	290	259	4	0.03	--	--	301
25	4	0.23	--	--	304	266	3	0.70	--	--	311
26	4	-0.03	--	--	300	273	4	0.37	--	--	306
32	4	0.43	--	--	307	274	4	-0.50	--	--	293
33	1	-1.99	--	--	270.6	276	4	0.03	--	--	301
38	4	-0.37	--	--	294.9	307	4	-0.03	--	--	300
42	4	0.10	--	--	302	326	4	0.10	--	--	302
45	4	-0.17	--	--	298	328	4	-0.10	--	--	299
46	4	-0.10	--	299	--	333	4	0.17	--	--	303
59	4	0.10	--	--	302	341	4	0.17	--	--	303
64	0	-3.56	--	--	247	349	3	0.77	312	--	--
70	4	0.10	--	--	302	356	4	0.17	--	--	303
76	4	-0.03	--	--	300	366	1	-1.70	--	--	275
85	4	-0.37	--	--	295	379	2	-1.50	--	278	--
86	4	0.50	--	--	308	386	4	0.03	301	--	--
91	2	-1.03	--	--	285						
97	4	-0.10	--	--	299						
100	4	-0.03	--	--	300						
102	3	0.83	--	--	313						
105	4	-0.50	--	--	293						
109	3	0.63	--	--	310						
113	4	0.30	--	--	305						
118	4	0.10	--	--	302						
134	4	-0.23	--	297	--						
138	3	-0.70	--	--	290						
142	4	0.30	--	--	305						
146	2	-1.36	--	--	280						
149	1	1.63	--	--	325						
151	4	-0.30	--	--	296						
180	1	1.56	--	--	324						
183	3	0.87	--	313.5	--						
190	4	-0.17	--	298	--						
193	2	-1.10	--	--	284						


Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued



SUMMARY	Methods							Statistics	
	0	3	4	5	6	7	Method Codes		
n =	2	0	22	1	8	1	00 Other	MPV =	89.6 µg/L
Minimum =	85.6	0	81	84.89	85.2	87.13	03 Atomic absorption: graphite furnace	F-pseudosigma =	4.67
Maximum =	94.5		114		128		04 Inductively coupled plasma	n =	34
Median =			89.9		87.1		05 Direct current plasma	Uh =	93.1
F-pseudosigma =			4.30		2.67		06 Inductively coupled plasma/mass spectrometry	Lh =	86.8
							07 Ion chromatography		

Lab	Rating	Z-value	Method Codes						
			0	3	4	5	6	7	
1	3	-0.55	--	--	--	--	87.04	--	
5	3	-0.66	--	--	86.5	--	--	--	
8	3	-0.56	--	--	--	--	87	--	
16	2	-1.20	--	--	84	--	--	--	
24	4	0.06	--	--	89.9	--	--	--	
25	2	1.20	--	--	95.2	--	--	--	
32	4	0.19	--	--	--	--	90.5	--	
33	3	-1.01	--	--	--	84.89	--	--	
42	2	1.09	--	--	94.7	--	--	--	
59	3	-0.71	--	--	--	--	86.3	--	
76	4	-0.51	--	--	--	--	87.22	--	
85	4	-0.34	--	--	88	--	--	--	
86	4	-0.45	--	--	87.5	--	--	--	
97	4	0.39	--	--	91.4	--	--	--	
100	3	0.75	--	--	93.1	--	--	--	
105	2	1.16	--	--	95	--	--	--	
113	3	-0.60	--	--	86.8	--	--	--	
121	1	-1.84	--	--	81	--	--	--	
134	4	0.07	--	--	89.94	--	--	--	
138	3	0.56	--	--	92.2	--	--	--	
190	0	-19.13	--	<0.25	--	--	--	--	
212	4	0.02	--	--	89.7	--	--	--	
219	0	5.22	--	--	114	--	--	--	
230	3	-0.94	--	--	--	--	85.2	--	
247	4	0.09	--	--	90	--	--	--	
254	2	1.05	94.5	--	--	--	--	--	
256	3	0.79	--	--	93.3	--	--	--	
259	4	-0.02	--	--	89.5	--	--	--	
265	4	-0.34	--	--	88	--	--	--	
273	2	1.46	--	--	96.4	--	--	--	
296	0	8.22	--	--	--	--	128	--	
326	3	-0.86	85.6	--	--	--	--	--	
328	3	-0.77	--	--	86	--	--	--	
341	4	0.09	--	--	--	--	90	--	
384	3	-0.53	--	--	--	--	87.13	--	

Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued




SUMMARY	Methods					Statistics	
	0	4	6	20	22	Method Codes	
n =	1	5	1	1	34	00 Other	MPV = 0.249 mg/L
Minimum =	0.251	0.115	0.246	0.25	0.036	04 Inductively coupled plasma	F-pseudostigma = 0.0215
Maximum =		0.26			0.32	06 Inductively coupled plasma/mass spectrometry	n = 42
Median =		0.255			0.246	20 Titration: colorimetric	Uh = 0.259
F-pseudostigma =		0.044			0.022	22 Colorimetric	Lh = 0.230

Lab	Rating	Z-value	Method Codes					Lab	Rating	Z-value	Method Codes				
			0	4	6	20	22				0	4	6	20	22
5	4	0.28	--	--	--	--	0.255	379	1	-1.81	--	--	--	--	0.21
8	0	-2.28	--	0.2	--	--	--	386	4	0.09	0.251	--	--	--	--
12	2	-1.35	--	--	--	--	0.22								
16	2	-1.35	--	--	--	--	0.22								
23	3	-0.51	--	--	--	--	0.238								
25	3	0.51	--	--	--	--	0.26								
32	4	-0.42	--	--	--	--	0.24								
38	3	0.56	--	--	--	--	0.261								
42	4	0.28	--	0.255	--	--	--								
45	1	1.77	--	--	--	--	0.287								
46	4	0.05	--	--	--	0.25	--								
64	4	0.37	--	--	--	--	0.257								
70	1	-1.54	--	--	--	--	0.216								
76	4	-0.13	--	--	0.246	--	--								
85	4	0.05	--	--	--	--	0.25								
86	4	0.47	--	0.259	--	--	--								
97	2	-1.35	--	--	--	--	0.22								
102	4	0.14	--	--	--	--	0.252								
105	4	-0.33	--	--	--	--	0.242								
113	4	-0.05	--	--	--	--	0.248								
134	4	0.47	--	--	--	--	0.259								
138	4	0.19	--	--	--	--	0.253								
142	3	-0.98	--	--	--	--	0.228								
146	1	1.54	--	--	--	--	0.282								
180	3	0.56	--	--	--	--	0.261								
183	4	-0.33	--	--	--	--	0.242								
190	3	-0.70	--	--	--	--	0.234								
212	0	2.51	--	--	--	--	0.303								
219	3	0.51	--	0.26	--	--	--								
224	3	-0.88	--	--	--	--	0.23								
227	0	-6.23	--	0.115	--	--	--								
247	1	-1.81	--	--	--	--	0.21								
257	0	3.30	--	--	--	--	0.32								
259	3	-0.84	--	--	--	--	0.231								
273	0	2.37	--	--	--	--	0.3								
274	0	-9.91	--	--	--	--	0.036								
307	4	-0.23	--	--	--	--	0.244								
328	3	0.51	--	--	--	--	0.26								
341	4	0.23	--	--	--	--	0.254								
366	3	-0.79	--	--	--	--	0.232								




Table 12. Statistical summary of reported data for standard reference sample M-164 (major constituents) -- continued



SUMMARY	Methods		Statistics
	4	6	
n =	11	12	MPV = 4.91 µg/L
Minimum =	3.8	4.42	F-pseudostigma = 0.262
Maximum =	5.53	6.18	n = 23
Median =	4.88	5.02	Uh = 5.12
F-pseudostigma =	0.152	0.335	Lh = 4.76

Lab	Rating	Z-value	Method Codes	
			4	6
1	3	-0.70	--	4.726
5	2	1.26	5.24	--
8	3	-0.80	--	4.7
16	4	-0.42	4.8	--
25	NR	--	<19	--
32	0	3.02	--	5.7
42	3	0.84	--	5.13
45	3	0.57	--	5.06
59	4	0.23	--	4.97
76	3	0.73	--	5.102
86	4	-0.34	4.82	--
97	NR	--	<4.90	--
100	0	2.37	5.53	--
105	NR	--	<20	--
121	0	-4.24	3.8	--
134	4	0.00	4.91	--
138	4	-0.11	4.88	--
142	1	-1.87	--	4.42
146	0	-3.78	3.92	--
212	4	0.04	4.92	--
219	4	-0.42	4.8	--
220	NR	--	<10	--
230	2	1.11	--	5.2
247	NR	--	<10	--
256	3	0.69	5.09	--
265	4	-0.42	--	4.8
296	0	4.85	--	6.18
328	NR	--	<8	--
341	3	-0.80	--	4.7


**Table 13. Statistical summary of reported data for standard reference sample N-75 (nutrient constituents)**



SUMMARY	Methods					Statistics	
	0	20	21	22	40	Method Codes	
n =	1	1	1	41	2	00 Other	<b>MPV = 0.077 mg/L</b>
Minimum =	0.076	0.077	0.064	0.043	0.07	20 Titration: colorimetric	F-pseudosigma = 0.0114
Maximum =				0.14	0.08	21 Titration: electrometric	n = 46
Median =				0.077		22 Colorimetric	Uh = 0.086
F-pseudosigma =			#####			40 Ion selective electrode	Lh = 0.071

Lab	Rating	Z-value	Method Codes					Lab	Rating	Z-value	Method Codes				
			0	20	21	22	40				0	20	21	22	40
1	0	-2.98	--	--	--	0.043	--	366	0	2.37	--	--	--	0.104	--
5	3	-0.53	--	--	--	0.071	--	369	4	-0.18	--	--	--	0.075	--
8	4	0.00	--	0.077	--	--	--	373	2	-1.05	--	--	--	0.065	--
10	3	-0.61	--	--	--	--	0.07	378	3	-0.55	--	--	--	0.071	--
16	4	0.26	--	--	--	0.08	--	379	2	-1.49	--	--	--	0.06	--
21	4	0.00	--	--	--	0.077	--	380	3	0.81	--	--	--	0.086	--
23	NR	--	--	--	--	<0.10	--	381	4	-0.09	0.076	--	--	--	--
25	2	1.14	--	--	--	0.09	--								
31	4	0.00	--	--	--	0.077	--								
33	0	4.82	--	--	--	0.132	--								
38	3	0.53	--	--	--	0.083	--								
46	4	-0.44	--	--	--	0.072	--								
64	4	0.26	--	--	--	0.08	--								
70	3	-0.61	--	--	--	0.07	--								
72	1	-1.84	--	--	--	0.056	--								
76	4	0.00	--	--	--	0.077	--								
85	4	-0.18	--	--	--	0.075	--								
86	1	-1.93	--	--	--	0.055	--								
91	2	-1.13	--	--	0.064	--	--								
97	4	-0.35	--	--	--	0.073	--								
102	0	3.77	--	--	--	0.12	--								
105	0	4.64	--	--	--	0.13	--								
110	2	1.14	--	--	--	0.09	--								
113	0	-2.80	--	--	--	0.045	--								
118	2	1.23	--	--	--	0.091	--								
134	4	0.00	--	--	--	0.077	--								
138	4	-0.18	--	--	--	0.075	--								
142	0	2.80	--	--	--	0.109	--								
146	0	2.28	--	--	--	0.103	--								
151	4	0.26	--	--	--	--	0.08								
180	4	0.00	--	--	--	0.077	--								
190	4	0.18	--	--	--	0.079	--								
193	3	-0.61	--	--	--	0.07	--								
247	0	5.52	--	--	--	0.14	--								
313	3	-0.53	--	--	--	0.071	--								
316	4	0.04	--	--	--	0.077	--								
318	4	0.26	--	--	--	0.08	--								
320	4	-0.03	--	--	--	0.077	--								
328	0	4.64	--	--	--	0.13	--								
341	4	-0.44	--	--	--	0.072	--								


**Table 13. Statistical summary of reported data for standard reference sample N-75 (nutrient constituents) -- continued**



SUMMARY	Methods			Method Codes	Statistics	
	0	20	22			
n =	3	1	31	00 Other	<b>MPV = 0.113 mg/L</b>	
Minimum =	0.05	0.084	0.03	20 Titration: colorimetric	F-pseudostigma = 0.0348	
Maximum =	0.12		0.5	22 Colorimetric	n = 35	
Median =			0.120		Uh = 0.139	
F-pseudostigma =			0.038		Lh = 0.092	

Lab	Rating	Z-value	Method Codes		
			0	20	22
1	4	-0.29	--	--	0.103
5	3	-0.95	--	--	0.08
8	3	-0.83	--	0.084	--
10	4	0.20	--	--	0.12
16	4	0.20	--	--	0.12
21	3	-0.55	--	--	0.094
23	0	11.11	--	--	0.5
25	4	0.20	0.12	--	--
31	3	-0.66	--	--	0.09
38	3	-0.66	--	--	0.09
46	NR	--	--	--	<0.14
70	2	1.23	--	--	0.156
72	2	-1.26	--	--	0.069
85	4	-0.46	--	--	0.097
91	4	-0.26	--	--	0.104
97	NR	--	--	--	<0.12
102	1	-1.81	0.05	--	--
105	NR	--	--	--	<1.00
113	3	-0.66	--	--	0.09
118	4	-0.37	--	--	0.1
134	NR	--	--	--	<0.2
138	4	0.46	--	--	0.129
142	4	-0.03	0.112	--	--
146	0	7.49	--	--	0.374
180	4	0.20	--	--	0.12
190	3	0.86	--	--	0.143
193	NR	--	--	--	<0.5
247	0	2.50	--	--	0.2
313	2	-1.06	--	--	0.076
316	3	0.63	--	--	0.135
318	4	0.00	--	--	0.113
320	2	1.11	--	--	0.152
328	4	-0.23	--	--	0.105
341	4	-0.09	--	--	0.11
366	0	-2.38	--	--	0.03
369	4	0.34	--	--	0.125
373	1	1.95	--	--	0.181
378	0	3.16	--	--	0.223
379	2	1.12	--	--	0.152
380	4	0.37	--	--	0.126


**Table 13. Statistical summary of reported data for standard reference sample N-75 (nutrient constituents) -- continued**



SUMMARY	Methods			Method Codes	Statistics	
	0	7	22			
n =	1	11	37	00 Other	<b>MPV = 0.092 mg/L</b>	
Minimum =	0.098	0.09	0.06	07 Ion chromatography	F-pseudosigma = 0.0052	
Maximum =		0.128	0.16	22 Colorimetric	n = 49	
Median =		0.099	0.091		Uh = 0.097	
F-pseudosigma =		0.013	0.005		Lh = 0.090	

Method Codes						Method Codes					
Lab	Rating	Z-value	0	7	22	Lab	Rating	Z-value	0	7	22
1	4	-0.39	--	--	0.09	341	4	-0.39	--	--	0.09
5	2	1.16	--	--	0.098	366	1	1.54	--	--	0.1
8	4	0.19	--	0.093	--	369	3	0.77	--	--	0.096
10	4	-0.39	--	--	0.09	373	4	0.39	--	--	0.094
16	0	-6.17	--	--	0.06	378	4	0.44	--	--	0.094
21	1	-1.73	--	--	0.083	379	4	0.19	--	--	0.093
23	4	-0.39	--	--	0.09	380	2	-1.16	--	--	0.086
25	0	4.05	--	0.113	--	381	2	1.16	0.098	--	--
31	1	-1.73	--	--	0.083	384	1	1.60	--	0.1	--
33	0	6.94	--	--	0.128						
38	3	-0.77	--	--	0.088						
42	3	0.77	--	0.096	--						
45	0	3.66	--	0.111	--						
46	2	-1.16	--	--	0.086						
64	4	-0.39	--	--	0.09						
70	4	-0.19	--	--	0.091						
72	0	-3.47	--	--	0.074						
85	4	-0.19	--	--	0.091						
86	1	-1.54	--	--	0.084						
91	3	0.58	--	--	0.095						
97	4	-0.19	--	--	0.091						
102	0	5.40	--	0.12	--						
105	1	1.54	--	--	0.1						
110	4	-0.39	--	0.09	--						
113	3	-0.58	--	--	0.089						
118	3	0.96	--	--	0.097						
134	3	-0.96	--	--	0.087						
138	3	0.96	--	0.097	--						
142	3	0.77	--	--	0.096						
146	0	13.10	--	--	0.16						
151	2	1.35	--	0.099	--						
180	0	5.20	--	--	0.119						
190	4	0.00	--	--	0.092						
193	4	-0.39	--	--	0.09						
247	4	-0.39	--	0.09	--						
313	4	0.31	--	--	0.094						
316	4	-0.31	--	--	0.09						
318	3	0.89	--	--	0.097						
320	4	-0.17	--	--	0.091						
328	0	-2.31	--	--	0.08						


**Table 13. Statistical summary of reported data for standard reference sample N-75 (nutrient constituents) -- continued**



SUMMARY	Methods				Method Codes		Statistics	
	0	4	6	22				
n =	1	3	1	43	00 Other		<b>MPV = 0.128 mg/L</b>	
Minimum =	0.106	0.128	0.126	0.084	04 Inductively coupled plasma		F-pseudosigma = 0.0082	
Maximum =	0.13		0.162		06 Inductively coupled plasma/mass spectrometry		n = 48	
Median =			0.128		22 Colorimetric		Uh = 0.131	
F-pseudosigma =			0.008				Lh = 0.120	

Lab	Rating	Z-value	Method Codes				Lab	Rating	Z-value	Method Codes			
			0	4	6	22				0	4	6	22
1	4	0.40	--	--	--	0.131	341	4	0.15	--	--	--	0.129
5	4	-0.34	--	--	--	0.125	366	2	-1.20	--	--	--	0.118
8	4	0.15	--	0.129	--	--	369	4	0.28	--	--	--	0.13
10	4	0.15	--	--	--	0.129	373	4	-0.09	--	--	--	0.127
16	1	-1.56	--	--	--	0.115	378	3	0.77	--	--	--	0.134
21	4	-0.21	--	--	--	0.126	379	0	-5.37	--	--	--	0.084
23	3	-0.95	--	--	--	0.12	380	2	-1.20	--	--	--	0.118
25	4	0.28	--	--	--	0.13	381	0	-2.67	0.106	--	--	--
31	4	-0.21	--	--	--	0.126							
38	0	2.85	--	--	--	0.151							
42	4	0.03	--	0.128	--	--							
45	1	1.87	--	--	--	0.143							
46	3	-0.58	--	--	--	0.123							
64	3	0.52	--	--	--	0.132							
70	0	-2.05	--	--	--	0.111							
72	2	-1.07	--	--	--	0.119							
76	4	-0.28	--	--	0.126	--							
85	4	0.28	--	--	--	0.13							
86	4	0.28	--	0.13	--	--							
91	3	0.89	--	--	--	0.135							
97	0	-3.40	--	--	--	0.1							
102	2	1.13	--	--	--	0.137							
105	4	-0.34	--	--	--	0.125							
113	4	-0.34	--	--	--	0.125							
118	2	-1.07	--	--	--	0.119							
134	1	1.87	--	--	--	0.143							
138	4	0.40	--	--	--	0.131							
142	4	0.40	--	--	--	0.131							
146	0	4.20	--	--	--	0.162							
151	1	1.75	--	--	--	0.142							
180	3	-0.83	--	--	--	0.121							
183	3	0.64	--	--	--	0.133							
190	3	-0.95	--	--	--	0.12							
193	4	0.40	--	--	--	0.131							
247	3	-0.95	--	--	--	0.12							
313	3	-0.83	--	--	--	0.121							
316	4	0.40	--	--	--	0.131							
318	3	-0.95	--	--	--	0.12							
320	4	-0.03	--	--	--	0.128							
328	4	0.28	--	--	--	0.13							


**Table 13. Statistical summary of reported data for standard reference sample N-75 (nutrient constituents) -- continued**



SUMMARY	Methods			Method Codes	Statistics	
	0	7	22			
n =	1	7	38	00 Other	<b>MPV = 0.095 mg/L</b>	
Minimum =	0.094	0.09	0.058	07 Ion chromatography	F-pseudosigma = 0.0072	
Maximum =		0.26	0.918	22 Colorimetric	n = 46	
Median =		0.101	0.095		Uh = 0.101	
F-pseudosigma =		0.017	0.007		Lh = 0.091	

Lab	Rating	Z-value	Method Codes			Lab	Rating	Z-value	Method Codes		
			0	7	22				0	7	22
1	4	-0.14	--	--	0.094	373	4	0.00	--	--	0.095
5	3	-0.70	--	--	0.09	378	3	-0.51	--	--	0.091
8	4	0.00	--	0.095	--	379	4	-0.42	--	--	0.092
10	4	0.28	--	--	0.097	380	3	-0.70	--	--	0.09
16	3	-0.83	--	--	0.089	381	4	-0.14	0.094	--	--
21	4	-0.14	--	--	0.094	384	0	22.93	--	0.26	--
23	0	2.09	--	--	0.11						
25	0	10.29	--	--	0.169						
31	4	-0.14	--	--	0.094						
33	0	4.87	--	0.13	--						
38	0	2.78	--	--	0.115						
42	2	1.11	--	0.103	--						
45	3	0.83	--	0.101	--						
46	4	0.42	--	--	0.098						
64	4	-0.28	--	--	0.093						
70	2	1.25	--	--	0.104						
72	3	0.83	--	--	0.101						
85	4	0.14	--	--	0.096						
97	2	-1.25	--	--	0.086						
102	4	-0.28	--	--	0.093						
105	0	-5.15	--	--	0.058						
113	3	0.56	--	--	0.099						
118	3	0.83	--	--	0.101						
134	4	-0.42	--	--	0.092						
138	4	-0.14	--	--	0.094						
142	3	-0.70	--	--	0.09						
146	3	-0.83	--	--	0.089						
151	3	-0.56	--	0.091	--						
180	4	0.42	--	--	0.098						
183	0	18.64	--	--	0.229						
190	4	-0.28	--	--	0.093						
247	3	-0.70	--	0.09	--						
313	4	0.24	--	--	0.097						
316	3	0.67	--	--	0.1						
318	4	0.25	--	--	0.097						
320	0	114.46	--	--	0.918						
328	0	42.42	--	--	0.4						
341	0	-2.23	--	--	0.079						
366	2	1.11	--	--	0.103						
369	3	-0.56	--	--	0.091						


**Table 14. Statistical summary of reported data for standard reference sample N-76 (nutrient constituents)**



SUMMARY	Methods					Statistics	
	7	20	21	22	40	Method Codes	
n =	1	1	1	44	2	07 Ion chromatography	<b>MPV = 0.885 mg/L</b>
Minimum =	0.843	0.84	0.841	0.75	0.882	20 Titration: colorimetric	F-pseudosigma = 0.0482
Maximum =				1.07	0.9	21 Titration: electrometric	n = 49
Median =				0.890		22 Colorimetric	Uh = 0.915
F-pseudosigma =				0.046		40 Ion selective electrode	Lh = 0.850

Lab	Rating	Z-value	Method Codes					Lab	Rating	Z-value	Method Codes				
			7	20	21	22	40				7	20	21	22	40
1	3	0.62	--	--	--	0.915	--	341	4	-0.10	--	--	--	0.88	--
5	2	-1.08	--	--	--	0.833	--	349	0	-2.80	--	--	--	0.75	--
8	3	-0.93	--	0.84	--	--	--	356	1	-1.81	--	--	--	0.798	--
10	4	0.31	--	--	--	--	0.9	366	4	0.46	--	--	--	0.907	--
12	3	-0.52	--	--	--	0.86	--	373	4	0.06	--	--	--	0.888	--
16	3	0.52	--	--	--	0.91	--	378	3	-0.68	--	--	--	0.852	--
23	4	0.21	--	--	--	0.895	--	379	3	-0.52	--	--	--	0.86	--
25	4	0.31	--	--	--	0.9	--	380	1	1.64	--	--	--	0.964	--
26	3	-0.87	0.843	--	--	--	--	383	3	0.52	--	--	--	0.91	--
33	2	1.37	--	--	--	0.951	--								
38	1	1.95	--	--	--	0.979	--								
46	3	-0.83	--	--	--	0.845	--								
64	4	0.31	--	--	--	0.9	--								
70	0	-2.61	--	--	--	0.759	--								
72	2	-1.47	--	--	--	0.814	--								
76	4	-0.11	--	--	--	0.88	--								
85	4	-0.10	--	--	--	0.88	--								
86	4	-0.50	--	--	--	0.861	--								
91	3	-0.91	--	--	0.841	--	--								
97	3	0.62	--	--	--	0.915	--								
102	0	3.22	--	--	--	1.04	--								
105	3	-0.93	--	--	--	0.84	--								
113	4	0.00	--	--	--	0.885	--								
118	1	1.89	--	--	--	0.976	--								
134	4	0.27	--	--	--	0.898	--								
138	3	-0.83	--	--	--	0.845	--								
142	3	0.77	--	--	--	0.922	--								
146	0	-2.51	--	--	--	0.764	--								
180	4	0.15	--	--	--	0.892	--								
190	4	-0.19	--	--	--	0.876	--								
193	4	-0.10	--	--	--	0.88	--								
205	3	0.73	--	--	--	0.92	--								
212	3	-0.73	--	--	--	0.85	--								
224	2	1.14	--	--	--	0.94	--								
227	3	0.89	--	--	--	0.928	--								
247	0	3.84	--	--	--	1.07	--								
307	4	-0.06	--	--	--	0.882	--								
313	4	0.37	--	--	--	0.903	--								
320	3	0.86	--	--	--	0.926	--								
328	3	-0.52	--	--	--	0.86	--								

Table 14. Statistical summary of reported data for standard reference sample N-76 (nutrient constituents) -- continued




SUMMARY	Methods			Method Codes	Statistics	
	0	20	22			
n =	3	1	36	00 Other	<b>MPV = 0.950 mg/L</b>	
Minimum =	0.947	0.84	0.62	20 Titration: colorimetric	F-pseudosigma = 0.0615	
Maximum =	1.172		1.96	22 Colorimetric	n = 40	
Median =		0.950			Uh = 1.00	
F-pseudosigma =		0.062			Lh = 0.919	

Lab	Rating	Z-value	Method Codes		
			0	20	22
1	4	-0.16	--	--	0.94
5	0	-2.28	--	--	0.81
8	1	-1.79	--	0.84	--
10	4	0.33	--	--	0.97
12	0	2.60	--	--	1.11
16	0	2.28	--	--	1.09
23	3	-0.55	--	--	0.916
25	4	0.49	0.98	--	--
38	4	-0.16	--	--	0.94
46	1	1.63	--	--	1.05
70	0	16.42	--	--	1.96
72	4	0.50	--	--	0.981
85	4	-0.49	--	--	0.92
91	4	-0.26	--	--	0.934
97	4	0.49	--	--	0.98
102	4	-0.16	--	--	0.94
105	0	4.71	--	--	1.24
113	4	-0.33	--	--	0.93
118	4	-0.05	0.947	--	--
134	3	-0.54	--	--	0.917
138	4	-0.18	--	--	0.939
142	0	3.61	1.172	--	--
146	0	2.11	--	--	1.08
180	2	-1.04	--	--	0.886
190	3	0.70	--	--	0.993
193	1	1.95	--	--	1.07
212	4	0.00	--	--	0.95
224	0	-5.36	--	--	0.62
227	0	-2.15	--	--	0.818
247	1	1.63	--	--	1.05
313	1	-1.64	--	--	0.849
320	3	0.64	--	--	0.989
328	4	0.00	--	--	0.95
341	4	0.00	--	--	0.95
349	1	-1.79	--	--	0.84
366	1	-1.63	--	--	0.85
373	4	-0.23	--	--	0.936
378	3	0.98	--	--	1.01
379	4	0.49	--	--	0.98
380	4	0.37	--	--	0.973




**Table 14. Statistical summary of reported data for standard reference sample N-76 (nutrient constituents) -- continued**



SUMMARY	Methods					Method Codes	Statistics	
	0	7	20	22	40			
n =	2	15	1	35	1	00 Other	<b>MPV = 0.891 mg/L</b>	
Minimum =	0.922	0.449	0.82	0.69	1.15	07 Ion chromatography	F-pseudosigma = 0.0460	
Maximum =	0.923	0.948		1.01		20 Titration: colorimetric	n = 54	
Median =		0.870		0.900		22 Colorimetric	Uh = 0.920	
F-pseudosigma =		0.053		0.030		40 Ion selective electrode	Lh = 0.858	

Lab	Rating	Z-value	Method Codes					Lab	Rating	Z-value	Method Codes				
			0	7	20	22	40				0	7	20	22	40
1	4	-0.14	--	--	--	0.884	--	307	4	-0.01	--	--	--	0.89	--
5	2	1.08	--	--	--	0.94	--	313	4	-0.03	--	--	--	0.889	--
8	3	0.86	--	0.93	--	--	--	320	4	0.14	--	--	--	0.897	--
10	4	0.42	--	--	--	0.91	--	328	3	-0.66	--	--	--	0.86	--
12	3	0.64	--	--	--	0.92	--	341	4	-0.25	--	--	--	0.879	--
16	0	-4.36	--	--	--	0.69	--	349	1	-1.53	--	--	0.82	--	--
23	3	-0.71	--	0.858	--	--	--	356	4	0.32	--	--	--	0.905	--
25	1	-1.66	--	0.814	--	--	--	366	3	-0.99	--	--	--	0.845	--
26	2	-1.49	--	0.822	--	--	--	373	3	-0.88	--	--	--	0.85	--
30	4	-0.40	--	0.872	--	--	--	378	1	1.58	--	--	--	0.963	--
33	4	0.01	--	0.891	--	--	--	379	3	-0.88	--	--	--	0.85	--
38	4	-0.21	--	--	--	0.881	--	380	3	-0.66	--	--	--	0.86	--
42	2	1.25	--	0.948	--	--	--	383	3	0.86	--	0.93	--	--	--
45	1	-1.62	--	0.816	--	--	--	384	0	-9.61	--	0.449	--	--	--
46	4	0.49	--	--	--	0.913	--								
64	3	0.64	--	--	--	0.92	--								
70	4	-0.14	--	--	--	0.884	--								
72	3	0.86	--	--	--	0.93	--								
85	3	0.64	--	--	--	0.92	--								
86	4	0.21	--	--	--	0.9	--								
91	4	0.21	--	--	--	0.9	--								
97	3	0.64	--	--	--	0.92	--								
102	4	-0.45	--	0.87	--	--	--								
105	4	0.42	--	--	--	0.91	--								
113	4	0.16	--	--	--	0.898	--								
118	3	0.71	0.923	--	--	--	--								
134	3	0.77	--	--	--	0.926	--								
138	4	-0.01	--	0.89	--	--	--								
142	3	0.64	--	--	--	0.92	--								
146	2	-1.36	--	--	--	0.828	--								
180	4	0.40	--	--	--	0.909	--								
183	0	5.65	--	--	--	--	1.15								
190	4	-0.16	--	--	--	0.883	--								
193	3	0.86	--	--	--	0.93	--								
205	0	2.60	--	--	--	1.01	--								
208	3	0.69	0.922	--	--	--	--								
212	0	-2.84	--	--	--	0.76	--								
224	4	-0.42	--	0.871	--	--	--								
227	3	-0.95	--	0.847	--	--	--								
247	0	-2.40	--	0.78	--	--	--								


Table 14. Statistical summary of reported data for standard reference sample N-76 (nutrient constituents) -- continued



SUMMARY	Methods				Method Codes		Statistics	
	0	4	6	22				
n =	1	4	1	41	00 Other		<b>MPV = 0.958 mg/L</b>	
Minimum =	0.963	0.933	0.991	0.62	04 Inductively coupled plasma		F-pseudosigma = 0.0389	
Maximum =	0.99		1.53		06 Inductively coupled plasma/mass spectrometry		Rating criterion = 0.0479	
Median =				0.958	22 Colorimetric		n = 47	
F-pseudosigma =				0.042			Uh = 0.984	
							Lh = 0.931	

Lab	Rating	Z-value	Method Codes				Lab	Rating	Z-value	Method Codes			
			0	4	6	22				0	4	6	22
1	4	-0.08	--	--	--	0.954	349	3	-0.56	--	--	--	0.931
5	3	0.96	--	--	--	1.004	356	3	-0.56	--	--	--	0.931
8	3	0.67	--	0.99	--	--	366	1	-1.57	--	--	--	0.883
10	4	0.33	--	--	--	0.974	373	0	3.11	--	--	--	1.107
12	3	-0.67	--	--	--	0.926	378	4	-0.06	--	--	--	0.955
16	3	0.56	--	--	--	0.985	379	1	-1.84	--	--	--	0.87
23	3	-0.79	--	--	--	0.92	380	1	-1.92	--	--	--	0.866
25	4	-0.38	--	--	--	0.94							
38	4	0.17	--	--	--	0.966							
42	4	-0.31	--	0.943	--	--							
45	1	1.63	--	--	--	1.036							
46	4	0.17	--	--	--	0.966							
64	4	0.25	--	--	--	0.97							
70	0	11.94	--	--	--	1.53							
72	2	1.42	--	--	--	1.026							
76	3	0.68	--	--	0.991	--							
85	4	0.29	--	--	--	0.972							
86	4	-0.08	--	0.954	--	--							
91	4	0.46	--	--	--	0.98							
97	3	0.67	--	--	--	0.99							
102	4	0.00	--	--	--	0.958							
105	4	0.10	--	--	--	0.963							
113	4	-0.29	--	--	--	0.944							
118	4	0.10	0.963	--	--	--							
134	4	0.35	--	--	--	0.975							
138	4	0.13	--	--	--	0.964							
142	3	0.71	--	--	--	0.992							
146	0	3.59	--	--	--	1.13							
180	2	-1.50	--	--	--	0.886							
183	3	-0.86	--	--	--	0.917							
190	3	-0.52	--	--	--	0.933							
193	2	-1.13	--	--	--	0.904							
212	0	-7.06	--	--	--	0.62							
224	0	-2.46	--	--	--	0.84							
227	3	-0.52	--	0.933	--	--							
247	3	-0.79	--	--	--	0.92							
307	4	-0.19	--	--	--	0.949							
313	3	0.54	--	--	--	0.984							
328	4	-0.38	--	--	--	0.94							
341	3	0.52	--	--	--	0.983							

Table 14. Statistical summary of reported data for standard reference sample N-76 (nutrient constituents) -- continued



SUMMARY	Methods			Method Codes	Statistics	
	0	7	22			
n =	2	10	35	00 Other	<b>MPV = 0.895 mg/L</b>	
Minimum =	0.948	0.323	0.62	07 Ion chromatography	F-pseudosigma = 0.0623	
Maximum =	0.98	2.934	2.88	22 Colorimetric	n = 47	
Median =		0.855	0.897		Uh = 0.949	
F-pseudosigma =		0.107	0.053		Lh = 0.865	

				Method Codes							Method Codes		
Lab	Rating	Z-value		0	7	22	Lab	Rating	Z-value	0	7	22	
1	2	1.25	--	--	--	0.973	366	3	0.55	--	--	0.929	
5	2	1.46	--	--	--	0.986	373	0	2.70	--	--	1.063	
8	1	1.53	--	0.99	--	--	378	4	-0.48	--	--	0.865	
10	2	1.24	--	--	--	0.972	379	4	0.24	--	--	0.91	
12	4	-0.14	--	--	--	0.886	380	4	-0.48	--	--	0.865	
16	4	0.00	--	--	--	0.895	383	0	-2.81	--	0.72	--	
23	0	-3.08	--	--	--	0.703	384	0	32.75	--	2.934	--	
25	0	-2.26	--	--	--	0.754							
26	0	-9.19	--	0.323	--	--							
30	4	-0.31	--	--	--	0.876							
33	2	-1.33	--	0.812	--	--							
38	4	-0.43	--	--	--	0.868							
42	3	0.98	--	0.956	--	--							
45	2	-1.25	--	0.817	--	--							
46	4	-0.19	--	--	--	0.883							
64	3	0.88	--	--	--	0.95							
70	4	0.40	--	--	--	0.92							
72	2	1.11	--	--	--	0.964							
85	4	0.24	--	--	--	0.91							
97	4	0.03	--	--	--	0.897							
102	2	1.37	0.98	--	--	--							
105	3	-0.93	--	--	--	0.837							
113	4	0.50	--	--	--	0.926							
118	3	0.85	0.948	--	--	--							
134	3	0.82	--	--	--	0.946							
138	4	-0.32	--	--	--	0.875							
142	4	-0.39	--	--	--	0.871							
146	3	0.66	--	--	--	0.936							
180	1	1.53	--	--	--	0.99							
183	4	0.11	--	--	--	0.902							
190	4	0.24	--	--	--	0.91							
208	3	-0.88	--	0.84	--	--							
212	0	-4.42	--	--	--	0.62							
224	1	-1.53	--	--	--	0.8							
227	4	-0.05	--	--	--	0.892							
247	4	-0.40	--	0.87	--	--							
313	4	-0.11	--	--	--	0.888							
328	0	31.88	--	--	--	2.88							
341	0	-3.13	--	--	--	0.7							
356	4	-0.40	--	--	--	0.87							

**Table 15. Statistical summary of reported data for standard reference sample P-39 (low ionic strength constituents)**


**P-39 ACIDITY (as CaCO<sub>3</sub>) in mg/L**

SUMMARY	Methods		Method Codes	Statistics
	20	21		
n =	1	5	20 Titration: colorimetric	inadequate data (n < 7)
Minimum =	21	14.92	21 Titration: electrometric	
Maximum =		51		
Median =		21.0		
F-pseudostandard deviation =		4.52		

Method Codes

Lab	Rating	Z-value	20	21
25	NR	0.00	--	21
105	NR	0.09	--	21.4
247	NR	6.63	--	51
273	NR	-1.26	--	15.3
274	NR	-1.34	--	14.92
328	NR	0.00	21	--


**Table 15. Statistical summary of reported data for standard reference sample P-39 (low ionic strength constituents)**  
 -- continued



SUMMARY	Methods								Method Codes		Statistics	
	0	1	2	4	5	6	7	20				
n =	2	4	2	16	1	1	1	2	00	Other	MPV = 8.65 mg/L	
Minimum =	8.629	8.12	8.65	7.659	8.63	8.63	9.013	4.82	01	Atomic absorption: direct, air	F-pseudostigma = 0.274	
Maximum =	9.9	8.83	8.68	9.63				7.4	02	Atomic absorption: direct, nitrous oxide	Rating criterion = 0.433	
Median =				8.72						04	Inductively coupled plasma	n = 29
F-pseudostigma =				0.334						05	Direct current plasma	Uh = 8.83
										06	Inductively coupled plasma/mass spectrometry	Lh = 8.46
										07	Ion chromatography	
										20	Titration: colorimetric	

Lab	Rating	Z-value	Method Codes							
			0	1	2	4	5	6	7	20
1	4	-0.05	8.629	--	--	--	--	--	--	--
2	3	0.84	--	--	--	--	--	--	9.013	--
5	3	-0.81	--	--	--	8.3	--	--	--	--
8	4	-0.35	--	--	--	8.5	--	--	--	--
23	2	-1.20	--	8.13	--	--	--	--	--	--
25	0	-2.29	--	--	--	7.659	--	--	--	--
33	4	-0.05	--	--	--	--	8.63	--	--	--
38	4	0.00	--	--	8.65	--	--	--	--	--
45	4	-0.05	--	--	--	--	--	8.63	--	--
64	4	-0.14	--	8.59	--	--	--	--	--	--
86	3	0.53	--	--	--	8.88	--	--	--	--
105	4	0.42	--	--	--	8.83	--	--	--	--
110	4	-0.44	--	--	--	8.46	--	--	--	--
113	4	0.16	--	--	--	8.72	--	--	--	--
134	3	0.72	--	--	--	8.96	--	--	--	--
138	3	0.81	--	--	--	9	--	--	--	--
180	4	0.16	--	--	--	8.72	--	--	--	--
190	4	0.07	--	--	8.68	--	--	--	--	--
193	4	0.42	--	8.83	--	--	--	--	--	--
228	2	-1.23	--	8.12	--	--	--	--	--	--
247	4	-0.39	--	--	--	8.48	--	--	--	--
265	4	0.12	--	--	--	8.7	--	--	--	--
273	0	2.27	--	--	--	9.63	--	--	--	--
274	0	-8.86	--	--	--	--	--	--	--	4.82
279	0	2.89	9.9	--	--	--	--	--	--	--
326	4	0.18	--	--	--	8.73	--	--	--	--
328	4	-0.49	--	--	--	8.44	--	--	--	--
379	0	-2.89	--	--	--	--	--	--	--	7.4
384	2	1.09	--	--	--	9.123	--	--	--	--


**Table 15. Statistical summary of reported data for standard reference sample P-39 (low ionic strength constituents)**  
 -- continued



SUMMARY	Methods				Method Codes	Statistics	
	0	7	20	22			
n =	1	26	1	1	00 Other	<b>MPV = 2.07 mg/L</b>	
Minimum =	2.34	1.7	15.96	2.16	07 Ion chromatography	F-pseudosigma = 0.112	
Maximum =		4.15			20 Titration: colorimetric	n = 29	
Median =		2.06			22 Colorimetric	Uh = 2.13	
F-pseudosigma =		0.148				Lh = 1.98	

Lab	Rating	Z-value	Method Codes			
			0	7	20	22
1	0	2.58	--	2.36	--	--
2	2	1.12	--	2.196	--	--
5	1	-1.78	--	1.87	--	--
8	4	0.27	--	2.1	--	--
23	3	-0.53	--	2.01	--	--
25	1	-1.51	--	1.9	--	--
33	4	0.36	--	2.11	--	--
45	1	-1.78	--	1.87	--	--
64	4	0.00	--	2.07	--	--
85	4	0.44	--	2.12	--	--
86	4	-0.36	--	2.03	--	--
105	4	0.18	--	2.09	--	--
110	3	-0.80	--	1.98	--	--
113	4	0.00	--	2.07	--	--
134	4	-0.18	--	2.05	--	--
138	4	-0.18	--	2.05	--	--
180	1	-1.78	--	1.87	--	--
190	4	-0.27	--	2.04	--	--
208	4	0.18	--	2.09	--	--
228	0	-2.05	--	1.84	--	--
247	2	-1.42	--	1.91	--	--
265	0	-3.29	--	1.7	--	--
273	0	2.40	2.34	--	--	--
274	0	123.60	--	--	15.96	--
326	3	0.89	--	2.17	--	--
328	0	18.51	--	4.15	--	--
333	4	0.18	--	2.09	--	--
379	3	0.80	--	--	--	2.16
384	3	0.55	--	2.132	--	--


**Table 15. Statistical summary of reported data for standard reference sample P-39 (low ionic strength constituents)**  
 -- continued



SUMMARY	Methods				Method Codes	Statistics
	7	22	40	41		
n =	13	1	7	1	07 Ion chromatography	<b>MPV = 0.650 mg/L</b>
Minimum =	0.587	0.74	0.611	0.73	22 Colorimetric	F-pseudostigma = 0.0852
Maximum =	1.1		3.31		40 Ion selective electrode	n = 22
Median =	0.630		0.640		41 Electrometric	Uh = 0.740
F-pseudostigma =	0.137		0.029			Lh = 0.625

Lab	Rating	Z-value	Method Codes			
			7	22	40	41
1	4	0.43	--	--	0.687	--
2	4	-0.23	0.63	--	--	--
5	0	5.04	1.08	--	--	--
8	2	1.06	0.74	--	--	--
23	4	-0.47	0.61	--	--	--
25	0	5.28	1.1	--	--	--
33	4	-0.29	0.625	--	--	--
45	3	-0.74	0.587	--	--	--
85	4	-0.12	--	--	0.64	--
86	1	1.72	0.797	--	--	--
105	0	2.11	0.83	--	--	--
113	4	0.35	0.68	--	--	--
134	4	0.12	--	--	0.66	--
138	4	-0.46	--	--	0.611	--
180	3	-0.53	0.605	--	--	--
190	4	-0.12	--	--	0.64	--
247	4	-0.23	0.63	--	--	--
273	0	31.20	--	--	3.31	--
274	2	1.06	--	0.74	--	--
328	4	-0.23	--	--	0.63	--
379	3	0.94	--	--	--	0.73
384	4	-0.44	0.613	--	--	--

**Table 15. Statistical summary of reported data for standard reference sample P-39 (low ionic strength constituents)**  
**-- continued**




SUMMARY	Methods							Method Codes		Statistics	
	0	1	4	5	6	7	12				
n =	1	8	15	1	1	2	1	00 Other		<b>MPV = 1.57 mg/L</b>	
Minimum =	1.82	1.44	1.111	1.46	1.57	1.57	1.4	01 Atomic absorption: direct, air		F-pseudosigma = 0.073	
Maximum =		1.63	1.84				1.609	04 Inductively coupled plasma		Rating criterion = 0.079	
Median =		1.56	1.60					05 Direct current plasma		n = 29	
F-pseudosigma =		0.074	0.085					06 Inductively coupled plasma/mass spectrometry		Uh = 1.61	
								07 Ion chromatography		Lh = 1.51	
								12 Flame emission			

Lab	Rating	Z-value	Method Codes						
			0	1	4	5	6	7	12
1	4	0.00	--	1.57	--	--	--	--	--
2	4	0.50	--	--	--	--	1.609	--	--
5	4	-0.13	--	--	1.56	--	--	--	--
8	4	0.38	--	--	1.6	--	--	--	--
23	4	-0.38	--	1.54	--	--	--	--	--
25	0	-5.85	--	--	1.111	--	--	--	--
33	2	-1.40	--	--	--	1.46	--	--	--
38	3	0.64	--	1.62	--	--	--	--	--
45	4	0.00	--	--	--	--	1.57	--	--
64	4	0.00	--	1.57	--	--	--	--	--
86	4	0.38	--	--	1.6	--	--	--	--
105	1	1.53	--	--	1.69	--	--	--	--
113	4	-0.38	--	--	1.54	--	--	--	--
134	4	-0.51	--	1.53	--	--	--	--	--
138	4	0.00	--	--	1.57	--	--	--	--
180	0	3.44	--	--	1.84	--	--	--	--
190	2	-1.40	--	1.46	--	--	--	--	--
193	3	0.76	--	1.63	--	--	--	--	--
228	4	0.00	--	--	--	--	1.57	--	--
247	1	-1.78	--	--	1.43	--	--	--	--
265	4	0.38	--	--	1.6	--	--	--	--
273	0	2.80	--	--	1.79	--	--	--	--
274	0	-2.17	--	--	--	--	--	1.4	--
279	0	3.18	1.82	--	--	--	--	--	--
326	1	-1.66	--	1.44	--	--	--	--	--
328	2	-1.27	--	--	1.47	--	--	--	--
333	3	-0.76	--	--	1.51	--	--	--	--
379	4	0.38	--	--	1.6	--	--	--	--
384	2	1.40	--	--	1.68	--	--	--	--




**Table 15. Statistical summary of reported data for standard reference sample P-39 (low ionic strength constituents)**  
 -- continued



SUMMARY	Methods							Method Codes		Statistics	
	0	1	4	5	6	7	20				
n =	1	7	17	1	1	1	1	00	Other	<b>MPV = 0.812 mg/L</b>	
Minimum =	0.84	0.672	0.753	0.829	0.818	0.843	3.38	01	Atomic absorption: direct, air	F-pseudosigma = 0.0393	
Maximum =		0.89	0.89					04	Inductively coupled plasma	Rating criterion = 0.0406	
Median =		0.750	0.810					05	Direct current plasma	n = 29	
F-pseudosigma =		0.069	0.030					06	Inductively coupled plasma/mass spectrometry	Uh = 0.833	
								07	Ion chromatography	Lh = 0.780	
								20	Titration: colorimetric		

Lab	Rating	Z-value	Method Codes						
			0	1	4	5	6	7	20
1	4	0.12	--	--	0.817	--	--	--	--
2	3	0.76	--	--	--	--	--	0.843	--
5	2	-1.45	--	--	0.753	--	--	--	--
8	4	-0.05	--	--	0.81	--	--	--	--
23	1	1.92	--	0.89	--	--	--	--	--
25	4	0.17	--	--	0.819	--	--	--	--
33	4	0.42	--	--	--	0.829	--	--	--
38	0	-3.37	--	0.675	--	--	--	--	--
45	4	0.15	--	--	--	0.818	--	--	--
64	3	-0.79	--	0.78	--	--	--	--	--
86	3	0.52	--	--	0.833	--	--	--	--
105	3	0.96	--	--	0.851	--	--	--	--
110	2	-1.33	--	--	0.758	--	--	--	--
113	3	-0.79	--	--	0.78	--	--	--	--
134	3	-0.52	--	--	0.791	--	--	--	--
138	4	0.00	--	--	0.812	--	--	--	--
180	4	-0.32	--	--	0.799	--	--	--	--
190	4	0.20	--	0.82	--	--	--	--	--
193	1	-1.53	--	0.75	--	--	--	--	--
228	1	-1.77	--	0.74	--	--	--	--	--
247	3	-0.54	--	--	0.79	--	--	--	--
265	4	0.44	--	--	0.83	--	--	--	--
273	1	1.92	--	--	0.89	--	--	--	--
274	0	63.25	--	--	--	--	--	3.38	--
279	3	0.69	0.84	--	--	--	--	--	--
326	0	-3.45	--	0.672	--	--	--	--	--
328	2	-1.28	--	--	0.76	--	--	--	--
379	4	-0.30	--	--	0.8	--	--	--	--
384	1	1.68	--	--	0.88	--	--	--	--


**Table 15. Statistical summary of reported data for standard reference sample P-39 (low ionic strength constituents)**  
 -- continued



SUMMARY	Methods							Method Codes		Statistics	
	0	1	4	5	6	7	12				
n =	1	6	17	1	1	2	1	00 Other		<b>MPV = 6.19 mg/L</b>	
Minimum =	6.33	5.96	5.044	5.96	6.28	6.391	4.92	01 Atomic absorption: direct, air		F-pseudostigma = 0.297	
Maximum =		37.8	7.07			7.92		04 Inductively coupled plasma		Rating criterion = 0.310	
Median =		6.14	6.19					05 Direct current plasma		n = 29	
F-pseudostigma =		0.126	0.297					06 Inductively coupled plasma/mass spectrometry		Uh = 6.36	
								07 Ion chromatography		Lh = 5.96	
								12 Flame emission			

Lab	Rating	Z-value	Method Codes						
			0	1	4	5	6	7	12
1	3	-0.75	--	--	5.957	--	--	--	--
2	3	0.65	--	--	--	--	--	6.391	--
5	4	0.10	--	--	6.22	--	--	--	--
8	4	-0.29	--	--	6.1	--	--	--	--
23	0	102.13	--	37.8	--	--	--	--	--
25	0	-3.70	--	--	5.044	--	--	--	--
33	3	-0.74	--	--	--	5.96	--	--	--
38	4	-0.32	--	6.09	--	--	--	--	--
45	4	0.29	--	--	--	--	6.28	--	--
64	4	-0.03	--	6.18	--	--	--	--	--
86	3	0.55	--	--	6.36	--	--	--	--
105	1	1.52	--	--	6.66	--	--	--	--
113	4	0.13	--	--	6.23	--	--	--	--
134	4	-0.32	--	6.09	--	--	--	--	--
138	3	0.87	--	--	6.46	--	--	--	--
180	4	-0.32	--	--	6.09	--	--	--	--
190	4	0.23	--	6.26	--	--	--	--	--
193	3	-0.74	--	5.96	--	--	--	--	--
228	0	5.59	--	--	--	--	7.92	--	--
247	4	0.00	--	--	6.19	--	--	--	--
265	4	0.36	--	--	6.3	--	--	--	--
273	0	2.71	--	--	7.03	--	--	--	--
274	0	-4.10	--	--	--	--	--	4.92	--
279	4	0.45	6.33	--	--	--	--	--	--
326	2	-1.49	--	--	5.73	--	--	--	--
328	3	-0.74	--	--	5.96	--	--	--	--
333	0	2.84	--	--	7.07	--	--	--	--
379	3	-0.61	--	--	6	--	--	--	--
384	1	-1.81	--	--	5.63	--	--	--	--

**Table 15. Statistical summary of reported data for standard reference sample P-39 (low ionic strength constituents)**  
 -- continued



SUMMARY	Methods			Method Codes	Statistics	
	21	40	41			
n =	1	7	18	21 Titration: electrometric	<b>MPV = 3.66</b>	
Minimum =	3.65	3.41	3.59	40 Ion selective electrode	F-pseudostigma = 0.039	
Maximum =		3.85	3.82	41 Electrometric	Rating criterion = 0.183	
Median =		3.68	3.66		n = 26	
F-pseudostigma =		0.104	0.030		Uh = 3.68	
					Lh = 3.63	

Lab	Rating	Z-value	Method Codes		
			21	40	41
1	4	0.00	--	--	3.66
2	4	-0.17	--	--	3.628
5	2	-1.37	--	3.41	--
8	4	-0.11	--	--	3.64
23	4	0.00	--	--	3.66
25	3	0.77	--	--	3.8
33	4	-0.05	--	--	3.65
38	4	0.22	--	3.7	--
64	4	0.05	--	--	3.67
85	4	-0.05	3.65	--	--
86	4	-0.22	--	--	3.62
105	4	0.22	--	--	3.7
110	4	0.11	--	3.68	--
134	4	0.11	--	3.68	--
138	4	0.11	--	--	3.68
180	4	-0.38	--	3.59	--
190	3	-0.82	--	3.51	--
193	4	-0.05	--	--	3.65
228	4	-0.05	--	--	3.65
247	4	0.00	--	--	3.66
273	4	-0.33	--	--	3.6
274	4	-0.38	--	--	3.59
326	3	0.87	--	--	3.82
328	4	0.00	--	--	3.66
333	4	0.11	--	--	3.68
379	2	1.04	--	3.85	--

**Table 15. Statistical summary of reported data for standard reference sample P-39 (low ionic strength constituents)  
-- continued**


**P-39 ORTHOPHOSPHATE as ORTHOPHOSPHORUS (PO<sub>4</sub> as P) in mg/L**

SUMMARY	Methods		Statistics
	7	22	
n =	1	6	inadequate data (F $\sigma$ > MPV)
Minimum =	0.181	0.001	
Maximum =		0.046	
Median =		0.013	
F-pseudosigma =		0.029	
	Method Codes		
		07 Ion chromatography	
		22 Colorimetric	

Method Codes

Lab	Rating	Z-value	7	22
5	NR	-0.50	--	0.001
8	NR	--	<0.3	--
23	NR	--	--	<0.01
25	NR	--	--	<0.003
33	NR	5.92	0.181	--
45	NR	--	<0.1	--
64	NR	--	--	<0.002
105	NR	--	--	<0.01
113	NR	--	--	<0.004
134	NR	--	--	<0.01
138	NR	--	--	<0.004
180	NR	0.00	--	0.015
190	NR	--	--	<0.01
247	NR	--	<0.01	--
273	NR	-0.18	--	0.01
274	NR	-0.42	--	0.003
328	NR	1.00	--	0.043
333	NR	--	<0.01	--
379	NR	1.10	--	0.046


**Table 15. Statistical summary of reported data for standard reference sample P-39 (low ionic strength constituents)**  
**-- continued**



SUMMARY	Methods			Method Codes	Statistics	
	0	7	51			
n =	1	27	2	00 Other	<b>MPV = 29.3 mg/L</b>	
Minimum =	31.88	27.7	29.03	07 Ion chromatography	F-pseudosigma = 1.08	
Maximum =		46	29.3	51 Turbidimetric	Rating criterion = 1.47	
Median =		29.3			n = 30	
F-pseudosigma =		1.02			Uh = 30.0	
					Lh = 28.5	

Lab	Rating	Z-value	Method Codes		
			0	7	51
1	0	3.34	--	34.2	--
2	4	0.45	--	29.96	--
5	4	-0.07	--	29.2	--
8	1	1.91	--	32.1	--
23	4	-0.14	--	29.1	--
25	2	-1.09	--	27.7	--
33	4	0.34	--	29.8	--
45	3	-0.55	--	28.5	--
64	4	0.14	--	29.5	--
85	4	-0.48	--	28.6	--
86	4	0.14	--	29.5	--
105	0	11.40	--	46	--
110	3	-0.78	--	28.15	--
113	3	-0.55	--	28.5	--
134	4	0.09	--	29.43	--
138	3	-0.68	--	28.3	--
180	4	0.27	--	29.7	--
190	4	0.07	--	29.4	--
193	4	-0.41	--	28.7	--
208	2	-1.02	--	27.8	--
228	3	0.84	--	30.53	--
247	4	0.00	--	29.3	--
265	3	-0.55	--	28.5	--
273	1	1.76	31.88	--	--
274	4	-0.18	--	29.03	--
326	2	-1.06	--	27.74	--
328	0	3.21	--	34	--
333	2	1.23	--	31.1	--
379	4	0.00	--	29.3	--
384	4	-0.26	--	28.92	--


**Table 15. Statistical summary of reported data for standard reference sample P-39 (low ionic strength constituents)**  
 -- continued



SUMMARY	Methods		Method Codes	Statistics	
	40	41			
n =	3	25	40 Ion selective electrode	<b>MPV =</b>	<b>193 <math>\mu\text{S}/\text{cm}</math></b>
Minimum =	185	154.6	41 Electrometric	F-pseudostigma =	8.9
Maximum =	196	245		Rating criterion =	9.6
Median =		193		n =	28
F-pseudostigma =		9.6		Uh =	197
				Lh =	185

Lab	Rating	Z-value	Method Codes	
			40	41
1	4	-0.30	--	190
2	3	-0.64	--	186.7
5	4	0.22	--	195
8	3	0.74	--	200
23	4	-0.40	--	189
25	4	0.12	--	194
33	1	-1.66	--	176.8
38	3	0.61	--	198.7
45	4	0.22	--	195
64	0	-3.51	--	159
85	3	-0.81	--	185
86	3	0.85	--	201
105	3	-0.97	--	183.5
110	2	-1.04	--	182.8
113	3	0.53	--	198
134	4	0.33	196	--
138	2	-1.33	--	180
180	0	5.41	--	245
190	4	-0.30	190	--
193	3	-0.71	--	186
228	3	0.81	--	200.7
247	3	0.85	--	201
273	4	0.33	--	196
274	4	0.02	--	193
326	0	-3.97	--	154.6
328	4	0.22	--	195
333	4	-0.02	--	192.7
379	3	-0.81	185	--

**Table 16. Statistical summary of reported data for standard reference sample HG-35 (mercury)**



SUMMARY	Methods				Method Codes		Statistics	
	6	8	9	11	06	08	09	11
n =	2	4	7	0	06 Inductively coupled plasma/mass spectrometry			
Minimum =	0.021	0.025	0.019	0	08 Atomic absorption: cold vapor			<b>MPV = 0.026 µg/L</b>
Maximum =	0.028	0.032	0.035		09 Atomic fluorescence			F-pseudosigma = 0.0024
Median =			0.027		11 Atomic absorption: hydride			n = 13
F-pseudosigma =			0.002					Uh = 0.028
								Lh = 0.025

Lab	Rating	Z-value	Method Codes			
			6	8	9	11
1	4	0.00	--	0.026	--	--
8	NR	--	--	<0.1	--	--
23	NR	--	--	--	--	<0.10
32	0	-2.21	0.021	--	--	--
45	3	0.82	0.028	--	--	--
46	4	-0.08	--	--	0.026	--
59	0	-3.23	--	--	0.019	--
105	NR	--	--	<0.2	--	--
134	NR	--	--	<0.1	--	--
138	4	0.16	--	--	0.027	--
147	0	2.75	--	--	0.033	--
180	NR	--	--	--	<0.050	--
193	3	-0.74	--	0.025	--	--
212	2	-1.06	--	--	0.024	--
235	3	-0.57	--	0.025	--	--
246	3	0.61	--	--	0.028	--
247	NR	--	--	<0.2	--	--
304	4	0.16	--	--	0.027	--
307	NR	--	--	--	--	-0.0955
356	0	2.29	--	0.032	--	--
379	NR	--	--	<0.1	--	--

**Table 17. Most probable values for constituents and properties in standard reference samples distributed in September 2002**

[MPV, most probable value; n, number of samples; µg/L, microgram per liter; mg/L, milligram per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius; ( ) rating criterion.]

**T-171**

<b>Analyte =</b>	<b>Silver</b>	<b>Aluminum</b>	<b>Arsenic</b>	<b>Boron</b>	<b>Barium</b>
MPV =	2.44 µg/L	19.4 µg/L	3.50 µg/L	23.5 µg/L	13.6 µg/L
F-pseudosigma =	0.119 (0.122)	1.63	0.274	3.04	0.46 (0.68)
n =	37	37	45	23	47
<b>Analyte =</b>	<b>Beryllium</b>	<b>Calcium</b>	<b>Cadmium</b>	<b>Cobalt</b>	<b>Chromium</b>
MPV =	1.18 µg/L	6.75 mg/L	4.92 µg/L	4.71 µg/L	3.10 µg/L
F-pseudosigma =	0.107	0.263 (0.337)	0.226 (0.246)	0.259	0.274
n =	40	56	56	35	39
<b>Analyte =</b>	<b>Copper</b>	<b>Iron</b>	<b>Potassium</b>	<b>Lithium</b>	<b>Magnesium</b>
MPV =	1.00 µg/L	53.4 µg/L	2.80 mg/L	6.30 µg/L	2.78 mg/L
F-pseudosigma =	0.172	4.82	0.093 (0.140)	0.599	0.111 (0.139)
n =	26	47	52	19	57
<b>Analyte =</b>	<b>Manganese</b>	<b>Molybdenum</b>	<b>Sodium</b>	<b>Nickel</b>	<b>Lead</b>
MPV =	12.4 µg/L	1.97 µg/L	8.60 mg/L	5.67 µg/L	1.09 µg/L
F-pseudosigma =	0.44 (0.62)	0.133	0.282 (0.430)	0.256 (0.283)	0.148
n =	54	25	55	44	40
<b>Analyte =</b>	<b>Antimony</b>	<b>Selenium</b>	<b>Silica</b>	<b>Strontium</b>	<b>Thallium</b>
MPV =	2.73 µg/L	1.39 µg/L	3.50 mg/L	84.0 µg/L	8.35 µg/L
F-pseudosigma =	0.234	0.182	0.163 (0.175)	4.04 (4.20)	0.519
n =	31	31	33	40	33
<b>Analyte =</b>	<b>Uranium</b>	<b>Vanadium</b>	<b>Zinc</b>		
MPV =	1.23 µg/L	2.39 µg/L	6.21 µg/L		
F-pseudosigma =	0.096	0.148	1.108		
n =	18	27	36		

**M-164**

<b>Analyte =</b>	<b>Alkalinity</b>	<b>Boron</b>	<b>Calcium</b>	<b>Chloride</b>	<b>Fluoride</b>
MPV =	11.4 mg/L	48.2 µg/L	12.7 mg/L	23.2 mg/L	0.604 mg/L
F-pseudosigma =	0.96	3.63	0.55 (0.64)	0.89 (1.16)	0.0497
n =	57	27	60	69	54
<b>Analyte =</b>	<b>Potassium</b>	<b>Magnesium</b>	<b>Sodium</b>	<b>pH</b>	<b>Residue on Evaporation</b>
MPV =	3.03 mg/L	2.78 mg/L	37.0 mg/L	7.06	198 mg/L
F-pseudosigma =	0.126 (0.152)	0.104 (0.139)	1.67 (1.85)	0.297 (0.353)	16.3
n =	58	59	59	54	46
<b>Analyte =</b>	<b>Silica</b>	<b>Sulfate</b>	<b>Specific Conductance</b>	<b>Strontium</b>	<b>Total Phosphorus as P</b>
MPV =	8.10 mg/L	11.2 mg/L	301 µS/cm	89.6 µg/L	0.249 mg/L
F-pseudosigma =	0.371 (0.405)	0.55 (0.56)	6.7 (15.0)	4.67	0.0215
n =	42	63	62	34	42
<b>Analyte =</b>	<b>Vanadium</b>				
MPV =	4.91 µg/L				
F-pseudosigma =	0.262				
n =	23				



**Table 17. Most probable values for constituents and properties in standard reference samples distributed in September 2002 -- continued**

[MPV, most probable value; n, number of samples; µg/L, microgram per liter; mg/L, milligram per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius; ( ) rating criterion.]

<b>N-75</b>	<b>Analyte =</b>	<b>Ammonia as N</b>	<b>Ammonia + Organic N as N</b>	<b>Nitrate as N</b>	<b>Total Phosphorus as P</b>	<b>Orthophosphate as P</b>
	MPV =	0.077 mg/L	0.113 mg/L	0.092 mg/L	0.128 mg/L	0.095 mg/L
	F-pseudosigma =	0.0114	0.0348	0.0052	0.0082	0.0072
	n =	46	35	49	48	46

<b>N-76</b>	<b>Analyte =</b>	<b>Ammonia as N</b>	<b>Ammonia + Organic N as N</b>	<b>Nitrate as N</b>	<b>Total Phosphorus as P</b>	<b>Orthophosphate as P</b>
	MPV =	0.885 mg/L	0.950 mg/L	0.891 mg/L	0.958 mg/L	0.895 mg/L
	F-pseudosigma =	0.0482	0.0615	0.0460	0.0389 (0.0479)	0.0623
	n =	49	40	54	47	47

<b>P-39</b>	<b>Analyte =</b>	<b>Acidity</b>	<b>Calcium</b>	<b>Chloride</b>	<b>Fluoride</b>	<b>Potassium</b>
	MPV =	inadequate data	8.65 mg/L	2.07 mg/L	0.650 mg/L	1.57 mg/L
	F-pseudosigma =		0.274 (0.433)	0.112	0.0852	0.073 (0.079)
	n =		29	29	22	29

	<b>Analyte =</b>	<b>Magnesium</b>	<b>Sodium</b>	<b>pH</b>	<b>Orthophosphate as P</b>	<b>Sulfate</b>
	MPV =	0.812 mg/L	6.19 mg/L	3.66	inadequate data	29.3 mg/L
	F-pseudosigma =	0.0393 (0.0406)	0.297 (0.310)	0.039 (0.183)		1.08 (1.47)
	n =	29	29	26		30

	<b>Analyte =</b>	<b>Specific Conductance</b>
	MPV =	193 µS/cm
	F-pseudosigma =	8.9 (9.6)
	n =	28

<b>HG-35</b>	<b>Analyte =</b>	<b>Mercury</b>
	MPV =	0.026 ug/L
	F-pseudosigma =	0.0024
	n =	13