

UNITED STATES
DEPARTMENT OF THE INTERIOR
GEOLOGICAL SURVEY

REPORT OF THE U.S. GEOLOGICAL SURVEY'S ANALYTICAL EVALUATION
PROGRAM--STANDARD REFERENCE WATER SAMPLES M4 (MAJOR CONSTITUENTS),
T93 (TRACE CONSTITUENTS), N15 (NUTRIENTS), P9 (PRECIPITATION SNOWMELT),
AND AMW2 (ACID MINE WATER).

Denver, Colorado

March 1986

CONTENTS

	<u>Page</u>
Abstract	1
Introduction	1
Purpose and plan	1
Preparation of samples	2
Determinations	3
Statistical evaluation	6
Laboratory performance and reported values	6
References.	7
Participating laboratories	8

TABLES

Table 1.	Explanation of abbreviations and symbols used in computer-printout tables	12
2.	Overall laboratory performance; Standard Reference Water Sample M4 (major constituents)	13
3.	Overall laboratory performance; Standard Reference Water Sample T93 (trace constituents)	19
4.	Overall laboratory performance; Standard Reference Water Sample N15 (nutrients)	25
5.	Overall laboratory performance; Standard Reference Water Sample P9 (precipitation snowmelt)	27
6.	Overall laboratory performance; Standard Reference Water Sample AMW2 (acid mine water)	30
7.	Analytical data; Standard Reference Water Sample M4 (major constituents)	33
8.	Statistics by method for Standard Reference Water Sample M4 (major constituents)	66
9.	Analytical data; Standard Reference Water Sample T93 (trace constituents)	69
10.	Statistics by method for Standard Reference Water Sample T93 (trace constituents)	98
11.	Analytical data; Standard Reference Water Sample N15 (nutrients)	101
12.	Statistics by method for Standard Reference Water Sample N15 (nutrients)	112
13.	Analytical data; Standard Reference Water Sample P9 (precipitation snowmelt)	113
14.	Statistics by method for Standard Reference Water Sample P9 (precipitation snowmelt)	132
15.	Analytical data; Standard Reference Water Sample AMW2 (acid mine water)	134
16.	Statistics by method for Standard Reference Water Sample AMW2 (acid mine water)	153

Comments, suggestions, or questions regarding these samples or this program may be made by calling (303) 236-3612 (FTS 776-3612), or by writing to Victor J. Janzer, SRWS Program, U.S. Geological Survey, 5293 Ward Road, Arvada, CO 80002.

ABSTRACT

The U.S. Geological Survey began an interlaboratory testing program of reference water samples in 1962. Program objectives then were, and now are, to provide a means for participating water laboratories to: (1) Identify analytical problem areas; (2) ascertain the accuracy and precision of common water analyses and analytical methods; and (3) provide reference samples for quality-assurance testing. Participation in this continuing quality-assurance program is mandatory for all domestic laboratories providing water-analysis data for U.S. Geological Survey use.

This report presents analytical data submitted by the laboratories that analyzed the reference samples distributed in **December 1985**. Relative performance ratings achieved by the laboratories for each determination, statistical evaluation of the data, and data summaries are given in 16 tables.

INTRODUCTION

The U.S. Geological Survey began an interlaboratory testing program of reference water samples in 1962 with a single major-constituent reference sample prepared from distilled water and reagent grade chemicals. Principal objectives of this continuing program are to provide a means for participating water laboratories to: (1) Identify analytical problem areas; (2) ascertain the accuracy and precision of analytical methods for determining various constituents and physical properties of water; and (3) provide reference samples for continuing quality-assurance testing of U.S. Geological Survey and various cooperator and contract laboratories. Twenty-three U.S. Geological Survey laboratories participated in the 1962 effort to determine 6 constituents in a single major-constituent Standard Reference Water Sample (SRWS). Today, more than 120 domestic laboratories, both Survey and non-Survey, participate in the program, which currently uses up to ten SRS types: (1) major constituents; (2) trace constituents; (3) nutrients; (4) herbicides; (5) insecticides; (6) water and suspended-sediment mixture for trace metals; (7) precipitation snowmelt; (8) priority pollutants; (9) acid mine drainage; and (10) a sediment (bed material) for trace metals.

Participation in this continuing quality-assurance program is mandatory for all laboratories providing water-analysis data for U.S. Geological Survey use. Major constituent, trace-constituent, and nutrient SRWS are prepared and distributed to participating laboratories twice each year. One or more of the other SRS types also may be included. This report presents analytical data submitted by the laboratories that analyzed the reference samples distributed to them in December 1985. Samples were analyzed during December and January, and data were requested to be submitted by January 24. Data received through February 10, 1986, have been included in this report. Relative performance ratings achieved by the laboratories for each determination, and statistical evaluations of the data are given in 16 tables.

PURPOSE AND PLAN

Periodic participation in reference sample analysis provides the means to alert participating laboratories to deficiencies in their analytical operations, and provides reference solutions for continuing quality assurance testing. Standard Reference Water Samples are prepared and distributed approximately every 6 months for analysis by U.S. Geological Survey and other cooperating laboratories. These analyses provide independent and objective evaluations of water-quality data provided by some of these laboratories for Survey use and publication. Non-U.S. Geological Survey laboratories participating in these studies are identified only by a confidential code number whereas participating U.S. Geological Survey laboratories are identified by location, name and code number.

This report summarizes the analytical results submitted by 101 of the 116 laboratories that requested and were shipped test samples in December 1985. **The original date of January 24, 1986, that was specified as a deadline for data return was extended several times.**

Samples which were distributed during December 1985 included SRWS M4 (major constituents), SRWS T93 (trace constituents), SRWS N15 (nutrients), P9 (precipitation snowmelt), and AMW2 (acid mine water). Not all samples are requested nor necessarily analyzed by all laboratories, nor do all laboratories enrolled in the program, participate in each round of analyses. Each participating laboratory was asked to perform at least those determinations that it makes routinely on the respective sample type, and to indicate the analytical methods used for each constituent. When method information was provided, it has been included in the respective data tables.

PREPARATION OF SAMPLES

SRWS M4 (major constituents) and AMW2 (acid mine water) were prepared from natural surface waters. SRWS T93 (trace constituents) and N15 (nutrients) were prepared from cold tap water. SRWS P9 (precipitation) was prepared from melted snow.

Sample M4 was prepared by allowing the suspended sediment to settle for several days in the collection drums. The partly clarified sample was then filtered sequentially through a 10 μm (micrometer) nominal size prefilter, then a 5 μm nominal size intermediate filter and finally a 0.45 μm membrane filter, into a large clean polyethylene drum.

The cold tap water used to prepare both T93 and N15 was filtered into separate clean polyethylene drums using only the 0.45 μm filters. Thymol [1.25 mg/L (milligrams per liter)] was added to both M4 and T93 to inhibit fungal and bacterial growth. No other additions were made to M4. Natural trace metal abundances in T93, however, were supplemented by the addition of selected constituents as solutions of reagent grade chemicals. These included aluminum, iron, manganese, antimony, arsenic, barium, beryllium, cadmium, chromium, cobalt, copper, lead, lithium, molybdenum, nickel, selenium, silver, thallium and zinc. Solutions for M4 and T93 were mixed overnight with a motor driven Teflon^{1/} coated stirrer, after which they were again filtered through 0.45 μm membrane filters, then passed through a flow-through ultraviolet [254-nm (nanometer)] sterilizer and packaged under ultraviolet radiation, in dry-heat sterilized 1-L Teflon or autoclaved polypropylene bottles. A 0.2 μm "final membrane filter" was also used when preparing T93.

Natural nutrient concentration levels in SRWS N15 were supplemented by the addition of ammonium, nitrate, nitrite, and orthophosphate "ions", and organic nitrogen and phosphorous as dissolved reagent grade chemicals. It was preserved by the addition of mercuric chloride (40 mg/L). Sodium chloride (450 mg/L) was also added. This is equivalent to the U.S. Geological Survey technique for field preservation of nutrient samples, using mercuric chloride-sodium chloride tablets. The sample was then mixed overnight with a motor-driven, Teflon-coated stirrer, packaged in polyethylene bottles, without sterilization, and stored in the dark at 4 °C (Celsius), until needed. The samples for this round-robin testing were all shipped at ambient temperatures.

^{1/} The use of the trade name in this report is for identification purposes only and does not constitute endorsement by the U.S. Geological Survey.

DETERMINATIONS

Abbreviations or symbols are listed below for each determination made on the various SRWS. These abbreviations and symbols are used in tables 2-16. Additional abbreviations and symbols used in tables 7-16 are explained in table 1.

Standard Reference Water Samples in this listing include: M4 (major constituents), T93 (trace constituents), N15 (nutrients), P9 (precipitation snowmelt), and AMW2 (acid-mine water).

Abbreviation/Symbol	M4 (mg/L) <u>1/</u>	T93 (µg/L) <u>2/</u>	N15 (mg/L)	P9 (mg/L) <u>3/</u>	AMW2 (µg/L) <u>4/</u>
ALK(CACO3) = Alkalinity (as CaCO ₃)	x				
ACID@CACO3 = Acidity (as CaCO ₃)		x		x	x
AG = Silver		x		x	x
AL = Aluminum		x			x
AS = Arsenic		x			x
B = Boron	x				
BA = Barium		x			x
BE = Beryllium		x			x
BR = Bromide	x				
CA = Calcium	x			x	
CD = Cadmium		x		x	x
CL = Chloride	x			x	x
CO = Cobalt		x		x	x
CR TOT = Chromium, total		x		x	x
CU = Copper		x		x	x
DSRD 180 = Dissolved solids, 180°C	x				

Standard Reference Water Samples in this listing include: M4 (major constituents), T93 (trace constituents), N15 (nutrients), P9 (precipitation snowmelt), and AMW2 (acid-mine water), (continued).

Abbreviation/Symbol			M4 (mg/L) <u>1/</u>	T93 (µg/L) <u>2/</u>	N15 (mg/L)	P9 (mg/L) <u>3/</u>	AMW2 (µg/L) <u>4/</u>
F	=	Fluoride	x			x	x
FE	=	Iron		x		x	x
HG	=	Mercury		x			x
I	=	Iodide	x				
K	=	Potassium	x			x	
LI	=	Lithium		x			x
MG	=	Magnesium	x			x	
MN	=	Manganese		x		x	x
MO	=	Molybdenum		x			x
NA	=	Sodium	x			x	
NH3-N	=	Ammonia as nitrogen			x	x	
NI	=	Nickel		x			x
NO2-N	=	Nitrite as nitrogen	x		x		
NO3-N	=	Nitrate as nitrogen	x		x		x
ORG-N	=	Organic nitrogen as nitrogen			x		
PB	=	Lead		x		x	x
PH	=	pH	x			x	x
PO4-P	=	Orthophosphate as phosphorus			x		
P, TOTAL	=	Phosphorus, Total as phosphorus	x		x		
SB	=	Antimony		x			
SE	=	Selenium		x			x

Standard Reference Water Samples in this listing include: M4 (major constituents), T93 (trace constituents), N15 (nutrients), P9 (precipitation snowmelt), and AMW2 (acid-mine water), (continued).

Abbreviation/Symbol		M4 (mg/L) <u>1/</u>	T93 (µg/L) <u>2/</u>	N15 (mg/L)	P9 (mg/L) <u>3/</u>	AMW2 (µg/L) <u>4/</u>
SIO2	= Silica	x				x
SO4	= Sulfate	x			x	x
SP.COND.	= Specific conductance	x			x	x
SR	= Strontium	x	x			x
TL	= Thallium		x		x	
V	= Vanadium	x				
ZN	= Zinc		x		x	x

1/ Results in mg/L except specific conductance (microsiemens or micromhos per centimeter at 25 °C); pH (units); boron, bromide, iodide, strontium, and vanadium (micrograms per liter).

2/ Results in µg/L except acidity (milligrams per liter).

3/ Results in mg/l except pH (units) and specific conductance (microsiemens or micromhos per centimeter at 25 °C); cadmium, chromium, cobalt, copper, iron, lead, manganese, silver, thallium, and zinc (micrograms per liter).

4/ Results in µg/L except specific conductance (microsiemens per centimeter at 25 °C); pH (units); acidity, aluminum, chloride, fluoride, iron, manganese, nitrate, silica, sulfate and zinc (milligrams per liter).

STATISTICAL EVALUATION

Statistical evaluations of the data were made to estimate the most probable value (MPV) for each of the constituents determined. Reported values of "less than" were considered as "not determined" and were not used (ignored) in the computation of the means, standard deviations, and so forth.

Outlying values for the remaining data were rejected on the basis of statistical tests as outlined in American Society for Testing and Materials (1981). After rejection of the outliers, the data remaining for each constituent were used to calculate the means, standard deviations, and percent deviation from the mean for each value. Outliers are not recalculated when determining the means and standard deviations for each determination listed by "method". The total range for each constituent included those values rejected as outliers. Confidence limits about the mean also were calculated; these limits define the range within which the true value may be expected to occur with a confidence level of 95 percent.

The mean, standard deviation, and confidence limits about the mean usually are reported to one more significant figure than the reported value. Statistical information is tabulated for each method used by three or more laboratories to determine a specific constituent. Tables 8, 10, 12, 14, and 16, listing the mean and standard deviation for the constituent determined by each method, and the number of laboratories that used it, follow the analytical-data tables for each SRWS.

LABORATORY PERFORMANCE AND REPORTED VALUES

To facilitate interlaboratory performance comparisons, ratings based on the analyses reported for each SRWS are included as tables 2-4 in this report. Laboratory performance for each constituent is rated on an arbitrary scale of 0 to 4 based on the number of "standard deviations" from the mean as indicated below:

4 (Excellent)-----	0.00 to 0.50 standard deviation
3 (Good)-----	0.51 to 1.00 standard deviation
2 (Satisfactory)-----	1.01 to 1.50 standard deviations
1 (Questionable)-----	1.51 to 2.00 standard deviations
0 (Poor)-----	Greater than 2.00 standard deviations

Averages of the constituent ratings for each Standard Reference Water Sample are given for each laboratory in the tables of overall laboratory performance (tables 2-6).

Laboratories were requested to identify the method used for each determination. The references for these methods are included with the analytical data and are identified in the following listing:

1. American Public Health Association and others, 1980, Standard methods for the examination of water and wastewater [15th ed.]: Washington, D.C., 1134 p.
2. American Society for Testing and Materials, 1982, Annual book of ASTM standards, Part 31: Philadelphia, PA, U.S.A., 1554 p.

3. Kopp, J. F., and McKee, G. F., 1978, Methods for chemical analysis of water and wastes: Cincinnati, Ohio, U.S. Environmental Protection Agency, 460 p.
4. Skougstad, M. W., Fishman, M. J., Friedman, L. C., Erdmann, D. E., and Duncan, S. S., eds., 1979, Methods for determination of inorganic substances in water and fluvial sediments: U.S. Geological Survey Techniques of Water-Resources Investigations, Book 5, Chapter A1, 626 p.
5. Fishman, M. J., and Bradford, W. L., eds., 1982, A supplement to methods for the determinations of inorganic substances in water and fluvial sediments: (Supplement to U.S. Geological Survey Techniques of Water-Resources Investigations, Book 5, Chapter A1), U.S. Geological Survey Open-File Report 82-272, 136 p.
6. Fishman, M. J., and Pyen, Grace, 1979, Determination of selected anions in water by ion chromatography: U.S. Geological Survey Water-Resources Investigations 79-101, 30 p.
7. Other references and instrument manufacturer's operation manuals.

In many instances, virtually the same method is given in several references. In those cases, all references describing that method are listed. If the analytical method used was not included in any of the listed references, analysts were requested to indicate "Other". Reference columns are left blank, if no method or an "other" method was indicated.

Values reported for all constituents determined in each SRWS are listed in tables 7, 9, 11, 13, and 15. Each value has been rounded off, when necessary, to conform to U.S. Geological Survey policy on reporting analytical data as given by Bishop and others (1978).

REFERENCES

- American Society for Testing and Materials, 1981, Annual book of ASTM standards, Part 41, Philadelphia, Pa., 1390 p.
- Bishop, E. E., Eckel, E. B., and others, 1978, Suggestions to Authors of the reports of the, U.S. Geological Survey: Washington, D. C., U.S. Government Printing Office, 6th edition, p. 198.

PARTICIPATING LABORATORIES

U.S. Geological Survey

CALIFORNIA, Sacramento: Yates 067
COLORADO, Denver: Duncan 001
 Schoen 019
 Taylor/Hedley112
 Taylor/Garbarino 118

FLORIDA, Ocala: Kirkland 055
LOUISIANA, Baton Rouge: Garrison 003

Cooperator

ALABAMA, Montgomery: ADEM Environmental Laboratory
ALABAMA, Tuscaloosa: Geological Survey of Alabama

ALASKA, Soldotna: Alaska Dept. of Fish and Game, Limnology Lab

ARKANSAS, Little Rock: Ark. Department of Pollution Control and Ecology

CALIFORNIA, Bryte: California Department of Water Resources Chemical Laboratory
CALIFORNIA, Canoga Park: Global Geochemistry Corp.
CALIFORNIA, Castaic: Department of Water Resources Chemical Laboratory
CALIFORNIA, La Mesa: San Diego Water Utilities Laboratory
CALIFORNIA, La Verne: The Metropolitan Water District of Southern California
CALIFORNIA, Los Gatos: Santa Clara Valley Water District
CALIFORNIA, Mammoth Lakes: Sierra Nevada Aquatic Research Lab
CALIFORNIA, Oakland: East Bay Municipal Utility District
CALIFORNIA, Palm Desert: California Regional Water Quality Control Board
CALIFORNIA, Sausalito: U.S. Corps of Engineers-Pacific Division Laboratory

COLORADO, Alamosa: Bureau of Reclamation
COLORADO, Aurora: Core Laboratories Incorporated
COLORADO, Denver: Denver Water Department - Quality Control Laboratory
COLORADO, Denver: Metropolitan Denver Sewage Disposal District #1
COLORADO, Denver: Colorado Department of Health
COLORADO, Fort Collins: Environmental Services/Water Utilities
COLORADO, Golden: Rockwell International General Laboratories
COLORADO, Parachute: Union Oil Company, Upgrade Laboratory
COLORADO, Pueblo: Board of Water Works
COLORADO, Steamboat Springs: ACZ Inc/Bookcliffs Laboratory Division

FLORIDA, Tampa: Hillsborough County Environmental Protection Commission
FLORIDA, West Palm Beach: South Florida Water Management District

GEORGIA, Athens: Univ. of Ga. Department of Horticulture
GEORGIA, Atlanta: Georgia Department of Natural Resources

ILLINOIS, Champaign: Illinois State Water Survey
ILLINOIS, Champaign: Illinois Environmental Protection Agency
ILLINOIS, Chicago: Illinois Environmental Protection Agency

INDIANA, Indianapolis: Marion County Public Health Laboratory
INDIANA, Indianapolis: Indianapolis Department of Public Works

Cooperator--continued

IOWA, Des Moines: University Hygienic Laboratory - Des Moines Branch

KANSAS, Lawrence: Kansas Geological Survey

KANSAS, Topeka: Kansas Department of Health and Environment

KENTUCKY, Frankfort: Kentucky Natural Resources & Environmental Protection

KENTUCKY, Louisville: Univ. of Louisville, Water Resources Lab

LOUISIANA, Lake Charles: Core Laboratories, Inc.

MAINE, Augusta: Maine Department of Environmental Protection

MARYLAND, Baltimore: Martel Laboratory Services, Inc.

MASSACHUSETTS, Barnstable: Barnstable County Health & Environmental Dept.

MASSACHUSETTS, Wellesley Hills: Massachusetts Department of Public Works

MINNESOTA, Eden Prairie: Braun Eng. and Testing Inc.

MINNESOTA, Minneapolis: Minnesota Public Health Department

MINNESOTA, St. Paul: Metropolitan Waste Control Commission

MONTANA, Butte: Montana Bureau of Mines and Geology

NEVADA, Reno: Water Analysis Laboratory, Desert Research Institute

NEVADA, Reno: Nevada State Health Laboratory

NEVADA, Sparks: City of Sparks, Jt. Treatment Plant

NEVADA, Sutcliffe: Pyramid Lake Fisheries

NEW JERSEY, Tom's River: Ocean County Health Department

NEW JERSEY, Trenton: N.J. Department of Health Environmental and Chemical
Laboratory

NEW MEXICO, Albuquerque: City of Albuquerque Water Resources Laboratory

NEW MEXICO, Gallup: Bureau of Indian Affairs - Natural Resources and Engineering
Laboratory

NEW YORK, Buffalo: Erie County Public Health Laboratory

NEW YORK, Central Islip: Suffolk County Health Services Department

NEW YORK, Hempstead: Nassau County Department of Health

NEW YORK, Millbrook: Institute of Ecosystem Studies

NEW YORK, North Babylon: EcoTest Laboratories, Inc.

NEW YORK, Oakdale: Suffolk County Water Authority

NEW YORK, Rochester: Monroe County Environmental Health Laboratory

NEW YORK, Syracuse: Syracuse Univ. Department of Civil Engineering

NEW YORK, Syracuse: Onondaga County Department of Drainage and Sanitation

NEW YORK, Wantagh: Cedar Creek Advanced Wastewater Treatment Lab

NORTH CAROLINA, Browns Summit: Lake Townsend Water Filtration Plant

NORTH CAROLINA, Charlotte: Mecklenburg County Environmental Health
Department

NORTH CAROLINA, Greensboro: City of Greensboro, Osborne Plant

Cooperator--continued

NORTH DAKOTA, Bismarck: North Dakota State Water Commission
 NORTH DAKOTA, Bismarck: North Dakota State Health Department Laboratory

OHIO, Columbus: Ohio Environmental Protection Agency Water Quality Laboratory
 OHIO, Dayton: The Miami Conservancy District
 OHIO, Medina: Medina County Sanitary Engineering Department
 OHIO, Tiffin: Heidelberg College, Water Quality Laboratory

OKLAHOMA, Norman: Oklahoma Geological Survey

OREGON, Corvallis: U.S. Department of Agriculture, Forestry Sciences Laboratory 37
 OREGON, Sandy: Bureau of Water Works, Water Quality Laboratory 42

PENNSYLVANIA, Harrisburg: Pennsylvania DER, Bureau of Laboratories

SOUTH DAKOTA, Brookings: South Dakota State University, Water Quality Laboratory
 SOUTH DAKOTA, Vermillion: South Dakota Geological Survey

TENNESSEE, Chattanooga: Tennessee Valley Authority, Laboratory Branch

TEXAS, Corpus Christi: Core Laboratories, Inc.

VIRGINIA, Culpeper: Environmental System Service
 VIRGINIA, Manassas: Occoquan Watershed Monitoring Lab
 VIRGINIA, Springfield: VERSAR, Inc.

114 WASHINGTON, Port Orchard: Washington Dept. of Ecology, Manchester Env. Lab 114
 # 11 WASHINGTON, Richland: Battelle, Pacific NW Lab 11
 # 59 WASHINGTON, Richland: Rockwell Hanford Operations 59

WEST VIRGINIA, Morgantown: West Virginia Geologic and Economic Survey

WISCONSIN, Madison: State Laboratory of Hygiene
 WISCONSIN, Milwaukee: Milwaukee Metropolitan Sewerage District

WYOMING, Cheyenne: Department of Environmental Quality, Water Quality Division
 WYOMING, Laramie: Wyoming Department of Agriculture

PUERTO RICO, San Juan: Department of Natural Resources, Laboratory Division

	<u>Battelle</u>	<u>Rockwell</u>	<u>WDOE (Manchester)</u>
# 11 =	50% Excellent (4)	# 59 = 65%	# 114 = 60%
	17% Good (3)	= 20%	= 0%
	0% Satisfactory (2)	= 6%	= 20%
	17% Questionable (1)	= 3%	= 0%
	17% Poor (0)	= 3%	= 20%
<u># Tests</u>	(6)	(29)	(5)

The following laboratories requested test samples, which were shipped to them, but no data were received:

FLORIDA, Palatka: St. John's River Water Management District

FLORIDA, Tallahassee: City of Tallahassee Water Quality Laboratory

GEORGIA, Albany: Water, Gas and Light Commission

GEORGIA, Athens: Soil Testing and Plant Analysis Lab

GEORGIA, Tifton: U.S. Department of Agriculture, SE Watershed Laboratory

KENTUCKY, Bowling Green: Western Kentucky University, Resources Mgmt. Inst. Lab

MISSOURI, Columbia: Environmental Trace Substances Research Center

NEVADA, Boulder City: U.S. Bureau of Reclamation, Lower Colorado Regional
Laboratory

NEW JERSEY, Trenton: N.J. Dept. of Environmental Protection, Office of Science
Research

NEW YORK, Rochester: FEV Wastewater Treatment Facility Laboratory

NEW YORK, Westbury: Nytest Environmental, Inc.

OKLAHOMA, Oklahoma City: Oklahoma Agriculture Department Laboratory

SOUTH CAROLINA, Columbia: SC Water Resources Commission

UTAH, Logan: Ecosystem Research Institute

WYOMING, Casper: Core Laboratories

Table 1.--Explanation of abbreviations and symbols used in computer printout sections

APDC - ammonium pyrrolidine dithiocarbamate
AUTO - automated
AVG - average
BLK - block
CHCL₃ - chloroform
CO'METRIC - colorimetric
DC - direct current
DEV - deviation
DIG - digestion
EDTA - ethylenediaminetetraacetic acid
H₂SO₄ - sulfuric acid
IC - inductively coupled
IGNORED - valued reported as less than detection level and not used in statistical analyses
K & HG SO₄ - potassium & mercuric sulfate
MIBK - methyl isobutyl ketone
NABH₄ - sodium borohydride
ND - not determined
NR - not rated
PCT - percent
PDCA - pyrrolidine dithiocarbamic acid
PERSULF - persulfate
PHOSPHOMOLYBD - phosphomolybdate
REJECT - values identified as an outlier and not used in statistical analyses
SPADNS - sodium 2-(parasulfophenylazo)-1,8-dihydroxy-3,6-naphthalene disulfonate
STD - standard
% - percent
< - less than
> - greater than

Table 2 Standard Reference Water Sample No. 44
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	ALK(CACO3)	B	BR	CA	CL	DSRD 190	F	I	K	MG
1	NR	4	NR	4	4	4	4	ND	3	3
2	0	ND	ND	1	ND	3	ND	ND	4	4
3	4	ND	ND	1	4	4	0	ND	3	3
4	4	NR	ND	1	4	4	2	ND	4	0
5	4	ND	ND	0	3	0	1	ND	4	ND
6	2	0	ND	0	0	4	4	ND	4	3
7	4	2	ND	3	4	4	4	ND	4	0
8	0	NR	ND	3	NR	ND	4	ND	4	4
9	ND	ND	ND	0	0	2	ND	ND	0	0
10	4	ND	ND	4	2	2	4	ND	3	4
12	4	NR	ND	3	4	3	3	ND	3	3
13	0	4	ND	0	3	ND	4	ND	3	0
14	4	3	ND	ND	ND	1	4	ND	ND	4
15	0	ND	ND	3	ND	4	0	ND	3	3
16	4	ND	ND	ND	0	ND	4	ND	4	ND
18	2	ND	ND	1	1	4	4	ND	4	3
19	ND	ND	ND	4	ND	ND	ND	ND	ND	4
20	4	ND	ND	3	NR	4	ND	ND	ND	ND
21	4	ND	ND	2	4	4	ND	ND	2	2
22	4	4	ND	3	4	ND	ND	ND	3	4
23	0	ND	ND	3	ND	ND	3	ND	4	3
24	4	NR	ND	4	4	4	ND	ND	4	4
25	ND	ND	ND	1	0	4	3	ND	0	0
26	4	ND	ND	3	4	4	ND	ND	ND	ND
27	4	ND	ND	4	ND	ND	ND	ND	ND	4
28	4	ND	ND	3	4	ND	ND	ND	4	4
29	4	ND	ND	4	1	4	4	ND	1	4
30	0	1	ND	4	4	4	4	ND	3	4
31	4	ND	ND	ND	ND	ND	ND	ND	ND	ND
32	4	ND	ND	4	4	4	ND	ND	3	3
33	ND	3	ND	4	ND	ND	ND	ND	0	4
34	0	ND	ND	2	2	4	ND	ND	0	2
35	4	ND	ND	4	3	4	3	ND	0	4
37	4	ND	ND	0	ND	2	ND	ND	4	4
39	4	4	ND	1	4	4	4	ND	3	2
40	ND	ND	ND	ND	4	NR	ND	ND	1	4
41	4	NR	NR	4	4	4	2	ND	4	4
43	ND	ND	ND	4	4	ND	ND	ND	ND	4
44	4	NR	ND	4	4	4	4	ND	4	3
45	4	ND	ND	3	3	3	3	ND	3	4
46	4	ND	ND	4	4	3	4	ND	0	2

Table 2 Standard Reference Water Sample No. M4
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.

0 (Poor) > 2.00 Std. Dev.
ND Not determined
NR Not rated

LAB	ALK(CACO3) B	BR	CA	CL	CSRD 130	F	I	K	MG
48	4	ND	3	4	ND	ND	ND	4	3
49	ND	ND	4	ND	ND	ND	ND	4	4
50	3	ND	4	4	4	ND	ND	4	4
51	3	ND	4	0	3	0	ND	0	0
52	4	ND	4	4	4	3	ND	4	4
53	4	ND	4	4	4	4	ND	4	4
55	4	ND	4	3	4	3	ND	4	4
56	3	ND	0	0	4	ND	ND	0	0
57	4	ND	4	4	4	ND	ND	4	4
58	4	ND	4	3	4	3	ND	1	0
59	4	ND	4	4	ND	4	ND	4	4
60	ND	ND	ND	ND	ND	ND	ND	ND	ND
61	2	ND	3	2	3	4	ND	4	4
62	4	ND	0	3	3	2	ND	4	2
64	4	ND	ND	4	ND	4	ND	4	ND
65	ND	ND	2	2	3	ND	ND	3	2
67	ND	NR	ND	4	3	ND	ND	ND	ND
69	4	ND	0	2	1	4	ND	4	4
70	0	ND	ND	0	ND	0	ND	ND	ND
71	1	NR	3	NR	2	3	ND	4	3
72	4	NR	3	4	2	4	ND	3	4
73	4	ND	2	NR	0	4	ND	0	3
74	ND	ND	ND	1	2	ND	ND	ND	ND
75	ND	ND	1	ND	ND	ND	ND	0	0
76	4	ND	1	2	ND	4	ND	4	4
77	4	ND	3	4	1	4	ND	3	3
79	ND	ND	ND	3	ND	ND	ND	ND	ND
84	4	1	0	4	ND	3	ND	4	3
86	4	ND	2	4	3	4	ND	4	4
87	2	ND	ND	ND	ND	ND	ND	ND	ND
88	ND	ND	3	ND	ND	ND	ND	3	4
89	4	0	3	0	4	ND	ND	4	2
90	4	2	4	1	4	3	ND	3	4
91	4	ND	3	4	ND	0	ND	0	4
92	0	ND	0	0	0	ND	ND	0	4
94	4	ND	4	4	4	4	ND	3	4
95	0	ND	4	4	2	3	ND	4	4
96	4	ND	3	0	4	3	ND	4	3
97	3	ND	3	4	2	ND	ND	3	3
99	4	NR	3	4	0	4	ND	0	4
100	4	4	3	0	4	3	ND	4	4

Table 2 Standard Reference Water Sample No. M4
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (poor) > 2.00 Std. Dev.
 3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined
 2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated
 1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	ALK(CACO3) B	BR	CA	CL	DSRD 180	F	I	K	MG
102	ND	ND	0	2	0	3	ND	ND	0
104	ND	ND	ND	ND	ND	ND	ND	ND	ND
105	ND	ND	2	3	ND	4	ND	4	2
107	0	NR	3	2	0	3	ND	4	4
109	4	ND	0	1	0	4	ND	0	2
111	ND	ND	ND	NR	ND	ND	ND	ND	ND
112	ND	ND	3	4	ND	3	ND	ND	4
114	0	ND	ND	ND	4	ND	ND	ND	ND
115	4	ND	0	4	ND	0	ND	2	4
116	ND	ND	ND	ND	ND	ND	ND	ND	ND
117	ND	ND	4	3	ND	4	ND	ND	4
118	ND	ND	ND	ND	ND	ND	ND	ND	ND

Table 2 Standard Reference Water Sample No. M4
Overall Laboratory Performance

LAB	NA	V02-N	V03-N	P, TOTAL	PH	SIO2	SO4	SP. COND.	SR	V	N	AVG.	RATING		
													4 (Excellent)	3 (Good)	2 (Satisfactory)
1	4	ND	ND	NR	4	4	2	3	4	NR	13	3.62	0 (poor) > 2.00 Std. Dev.	ND Not determined	NR Not rated
2	3	ND	ND	0	3	4	ND	0	0	ND	11	2.00	3 (Good) 0.51 to 1.00 Std. Dev.		
3	4	ND	ND	ND	3	ND	4	4	ND	ND	11	3.09	2 (Satisfactory) 1.01 to 1.50 Std. Dev.		
4	4	0	0	3	4	4	3	1	4	4	17	2.71	1 (Questionable) 1.51 to 2.00 Std. Dev.		
5	ND	3	2	ND	4	ND	ND	1	ND	ND	10	2.20			
6	3	ND	2	0	3	3	4	2	0	3	17	2.41			
7	0	ND	ND	ND	1	ND	4	3	ND	ND	12	2.75			
8	4	ND	ND	NR	2	3	3	4	3	NR	11	3.09			
9	0	ND	ND	ND	ND	ND	2	ND	ND	ND	7	0.57			
10	3	3	1	3	3	4	4	ND	ND	ND	14	3.14			
12	3	NR	4	4	3	4	3	4	ND	NR	14	3.43			
13	4	ND	1	ND	4	ND	4	2	ND	ND	12	2.42			
14	3	3	3	3	4	ND	4	4	ND	ND	12	3.33			
15	0	NR	4	3	1	0	3	4	0	ND	14	2.00			
16	4	ND	4	2	3	ND	4	3	ND	ND	10	3.20			
18	4	NR	4	ND	0	ND	4	4	ND	ND	12	2.92			
19	4	ND	ND	ND	ND	4	ND	ND	ND	ND	4	4.00			
20	NR	ND	NR	ND	2	ND	3	4	ND	ND	6	3.33			
21	4	3	0	3	2	ND	4	4	ND	ND	13	2.92			
22	4	ND	4	ND	3	3	4	3	3	ND	13	3.54			
23	3	3	4	ND	3	2	3	0	ND	ND	12	2.58			
24	4	ND	0	ND	ND	4	4	ND	ND	ND	10	3.60			
25	2	NR	4	NR	3	ND	3	1	ND	ND	11	1.91			
26	ND	ND	4	ND	4	ND	ND	4	ND	ND	7	3.86			
27	ND	NR	4	3	4	0	ND	4	ND	ND	8	3.38			
28	4	NR	4	3	3	4	ND	3	4	ND	12	3.67			
29	3	NR	4	ND	3	4	4	3	ND	ND	13	3.31			
30	3	ND	4	3	0	4	4	3	4	4	17	3.12			
31	ND	ND	ND	3	4	ND	ND	ND	ND	ND	3	3.67			
32	1	ND	ND	ND	3	ND	2	4	ND	ND	10	3.20			
33	2	ND	ND	ND	ND	4	ND	ND	4	ND	7	3.00			
34	4	ND	ND	ND	2	ND	0	0	ND	ND	10	1.60			
35	3	ND	ND	ND	2	ND	4	4	ND	ND	10	3.50			
37	4	3	4	4	4	4	ND	4	ND	ND	12	3.42			
39	2	NR	4	NR	4	3	4	4	4	NR	15	3.40			
40	ND	ND	ND	NR	0	NR	ND	ND	ND	ND	4	2.25			
41	4	NR	4	3	3	4	4	4	3	NR	15	3.67			
43	ND	3	4	3	4	4	4	4	ND	ND	10	3.80			
44	4	NR	2	4	3	3	4	4	2	ND	14	3.50			
45	4	ND	4	2	3	3	4	3	4	NR	15	3.33			
46	3	3	4	0	4	3	4	1	ND	ND	15	2.87			

Table 2 Standard Reference Water Sample No. M4
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.
0 (Poor) > 2.00 Std. Dev.
ND Not determined
NR Not rated

LAB	NA	NO2-N	NO3-N	P, TOTAL	PH	SI02	SO4	SP. COND.	SR	V	N	AVG.
45	3	ND	4	3	3	4	4	4	ND	ND	12	3.58
49	3	ND	ND	ND	4	0	1	4	ND	ND	8	3.00
50	4	NR	4	NR	4	ND	3	3	NR	NR	12	3.75
51	NR	NR	4	0	0	ND	0	0	ND	ND	12	1.17
52	3	ND	3	ND	3	4	4	4	ND	ND	14	3.71
53	4	ND	4	4	4	ND	4	3	ND	NG	14	3.93
55	4	3	4	3	3	4	3	4	3	ND	16	3.56
56	0	3	4	0	2	ND	4	3	ND	ND	13	1.77
57	2	NR	4	3	4	ND	4	0	ND	ND	12	3.42
58	0	0	3	3	2	4	4	4	ND	ND	16	2.69
59	4	ND	0	ND	3	4	4	1	4	ND	13	3.38
60	ND	3	ND	4	3	4	ND	ND	ND	ND	4	3.53
61	4	ND	1	3	3	4	0	4	1	NR	16	2.75
62	3	3	0	4	4	2	4	4	ND	2	16	2.75
64	4	ND	4	3	3	ND	4	3	ND	ND	10	3.70
65	2	NR	3	NR	2	ND	3	4	ND	ND	10	2.60
67	ND	NR	ND	ND	ND	4	4	4	ND	ND	5	3.80
69	3	3	4	4	4	2	0	1	ND	ND	15	2.67
70	ND	ND	ND	ND	4	ND	ND	0	ND	ND	5	0.80
71	4	NR	4	3	4	2	3	3	ND	NR	13	3.00
72	4	3	4	3	2	4	3	4	4	0	17	3.24
73	3	NR	4	NR	0	ND	2	ND	ND	ND	10	2.20
74	ND	NR	4	ND	0	3	1	4	ND	ND	8	2.13
75	0	ND	3	3	0	ND	2	ND	ND	ND	9	1.13
76	4	3	4	ND	3	ND	ND	4	ND	ND	10	3.30
77	4	3	3	4	4	4	4	4	ND	ND	15	3.47
79	ND	ND	4	0	3	ND	4	3	ND	ND	6	2.83
84	3	ND	4	NR	3	4	1	2	ND	ND	13	2.77
86	0	3	4	3	4	ND	4	4	ND	ND	14	3.36
87	ND	3	0	3	4	4	ND	1	ND	ND	7	2.43
88	3	ND	ND	ND	ND	ND	ND	ND	ND	ND	4	3.25
89	4	3	0	3	4	ND	3	2	ND	ND	14	2.57
90	3	3	0	1	4	2	4	2	3	ND	17	2.76
91	3	3	4	ND	4	ND	4	1	ND	ND	12	2.83
92	3	3	0	3	3	ND	0	4	ND	ND	13	1.54
94	4	NR	4	ND	4	2	3	4	ND	ND	13	3.69
95	4	0	4	4	4	3	2	3	0	4	18	2.83
96	3	NR	4	NR	4	3	3	3	2	NR	15	3.13
97	4	NR	4	NR	ND	ND	ND	ND	ND	ND	8	3.25
99	3	NR	4	3	3	3	4	4	2	NR	15	3.00
100	4	ND	1	ND	3	3	4	3	ND	ND	14	3.14

Table 2 Standard Reference Water Sample No. M4
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.

0 (Poor) > 2.00 Std. Dev.
ND Not determined
NR Not rated

LAB	NA	N02-N	N03-N	P, TOTAL	PH	SI02	S04	SP. COND.	SR	V	N	AVG.
102	ND	1	3	1	3	3	3	2	ND	ND	13	1.77
104	ND	3	4	0	4	ND	ND	ND	ND	ND	4	2.75
105	3	NR	3	3	ND	ND	2	4	ND	ND	11	3.00
107	0	3	4	0	2	4	0	2	ND	NR	15	2.07
109	4	NR	1	0	3	4	4	0	ND	ND	15	2.07
111	ND	NR	3	3	ND	2	ND	2	ND	ND	4	2.50
112	3	ND	4	ND	ND	ND	4	ND	3	1	10	3.20
114	ND	ND	ND	ND	2	ND	ND	4	ND	ND	4	2.50
115	2	ND	4	NR	4	0	4	2	3	NR	14	2.57
116	ND	ND	ND	ND	4	ND	ND	ND	ND	ND	1	4.00
117	ND	ND	ND	ND	ND	ND	4	ND	ND	ND	5	3.80
118	ND	ND	ND	ND	ND	ND	ND	ND	3	ND	1	3.00

Table 3 Standard Reference Water Sample No. 193
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.
0 (Poor) > 2.00 Std. Dev.
ND Not determined
NR Not rated

LAB	ACIDCACC03	AL	AS	BA	BE	CD	CO	CR 101	CU
1	3	4	4	NR	1	4	2	0	4
2	ND	ND	ND	4	ND	ND	ND	ND	4
3	ND	ND	ND	ND	ND	ND	ND	ND	4
4	0	2	4	4	2	2	3	ND	4
5	ND	2	4	4	ND	4	ND	4	4
6	ND	4	4	0	0	2	2	4	4
7	ND	4	3	NR	1	3	NR	0	3
8	ND	NR	4	4	4	NR	NR	4	4
9	ND	ND	ND	ND	ND	ND	ND	4	4
10	ND	3	4	4	ND	3	ND	4	4
12	3	4	3	4	4	4	ND	4	4
13	ND	ND	ND	ND	ND	ND	ND	ND	ND
14	3	2	4	0	4	4	ND	4	3
15	ND	ND	ND	ND	ND	3	ND	4	ND
16	4	ND	ND	ND	NR	NR	ND	ND	ND
17	ND	NR	ND	NR	ND	4	ND	4	4
19	ND	ND	ND	ND	3	2	3	ND	4
21	0	2	ND	ND	4	1	4	ND	4
22	ND	4	3	3	ND	ND	ND	ND	4
23	ND	4	ND	4	ND	4	4	ND	ND
24	3	4	ND	4	ND	NR	ND	ND	4
25	ND	ND	ND	ND	ND	0	ND	0	4
26	ND	3	4	1	4	4	0	4	4
27	ND	ND	ND	ND	ND	ND	ND	ND	ND
28	ND	ND	2	4	4	4	4	3	ND
29	ND	ND	1	ND	ND	ND	ND	2	4
30	ND	ND	3	4	4	4	2	0	0
32	4	4	ND	NR	ND	2	ND	0	3
33	ND	0	4	0	4	4	4	3	3
34	0	NR	ND	4	ND	4	CN	3	0
35	ND	ND	ND	ND	ND	ND	ND	ND	ND
36	ND	2	4	3	4	4	NR	4	4
40	ND	0	ND	C	4	0	ND	4	4
41	ND	NR	4	4	4	NR	NR	4	4
44	ND	4	ND	ND	3	3	CN	4	4
45	2	4	ND	4	4	4	ND	4	4
46	ND	2	C	2	ND	4	CN	0	0
50	ND	4	4	3	0	2	NR	4	3
51	ND	ND	4	ND	NR	3	NR	NR	NR
52	ND	ND	ND	ND	ND	ND	CN	CN	CN
54	GN	ND	4	4	ND	ND	ND	ND	ND

Table 3 Standard Reference Water Sample No. 193
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.
ND Not determined
NR Not rated

LAB	ACI3ACAC03 AG	AL	AS	BA	BE	CD	CO	CR TOT	CU
55	ND	ND	ND	CN	ND	ND	CN	ND	4
56	ND	ND	ND	ND	ND	0	ND	1	4
57	ND	4	4	3	0	4	ND	4	4
58	ND	3	ND	2	1	2	NR	1	3
59	ND	ND	4	4	4	ND	ND	ND	3
60	ND	ND	ND	ND	ND	2	ND	4	3
61	4	2	0	3	0	3	4	0	4
62	3	NR	2	4	4	2	ND	4	0
64	ND	NR	NR	NR	ND	0	CN	1	0
65	ND	ND	NR	NR	4	4	CN	NR	4
67	ND	ND	4	ND	ND	ND	CN	3	4
69	ND	4	0	4	ND	1	ND	3	2
70	ND	3	ND	4	ND	4	ND	4	4
71	4	NR	0	NR	NR	NR	NR	NR	2
72	ND	4	3	4	4	4	4	3	3
73	ND	NR	3	ND	ND	4	ND	4	4
75	ND	ND	ND	CN	ND	ND	ND	ND	0
76	ND	ND	ND	ND	ND	ND	ND	ND	ND
77	3	4	3	4	ND	4	0	4	4
78	ND	NR	1	0	CN	1	CN	4	4
79	ND	ND	ND	ND	ND	1	ND	3	3
84	ND	NR	0	4	0	4	NR	3	4
86	ND	4	4	2	ND	4	ND	4	4
91	ND	NR	ND	ND	ND	NR	ND	NR	1
92	0	ND	ND	CN	ND	0	CN	NR	NR
94	3	4	4	3	4	4	ND	4	4
95	4	4	3	0	NR	0	NR	NR	4
96	ND	4	4	4	4	3	NR	4	4
97	ND	ND	ND	4	ND	1	ND	0	4
99	4	4	3	4	4	NR	2	0	3
100	CN	ND	CN	ND	ND	ND	ND	ND	ND
101	ND	3	ND	ND	ND	ND	CN	ND	ND
102	CN	4	4	3	4	1	3	4	4
104	CN	ND	ND	CN	ND	2	ND	3	4
105	ND	4	4	1	ND	1	CN	3	4
107	2	NR	NR	4	NR	NR	NR	NR	NR
112	ND	NR	ND	4	4	2	ND	ND	2
116	ND	0	3	0	1	4	3	2	4
117	ND	NR	ND	ND	ND	2	CN	ND	4
118	ND	ND	ND	4	4	4	CN	4	4

Table 3 Standard Reference Water Sample No. 193 Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
 3 (Good) 0.51 to 1.00 Std. Dev.
 2 (Satisfactory) 1.01 to 1.50 Std. Dev.
 1 (Questionable) 1.51 to 2.00 Std. Dev.
 0 (Poor) > 2.00 Std. Dev.
 ND Not determined
 NR Not rated

LAB	FE	HG	LI	MN	MO	NI	PB	SB	SE	SR
1	2	NR	3	3	3	3	4	4	1	3
2	3	ND	2	2	ND	ND	2	ND	NR	0
3	4	ND	3	4	4	0	4	ND	ND	ND
4	4	ND	4	4	ND	ND	2	ND	1	5
5	ND	ND	ND	4	ND	ND	2	ND	ND	5
6	0	4	0	0	3	4	2	3	4	6
7	3	4	4	2	ND	ND	0	ND	0	7
8	4	ND	ND	0	NR	4	4	ND	4	8
9	4	ND	ND	3	ND	4	0	ND	ND	9
10	4	3	ND	4	ND	3	3	ND	3	10
12	4	0	ND	4	4	4	4	NR	3	12
13	0	ND	ND	0	ND	ND	ND	ND	ND	13
14	3	NR	ND	4	4	4	4	ND	3	14
15	0	ND	ND	ND	ND	ND	ND	ND	0	15
16	ND	ND	ND	ND	ND	ND	ND	ND	ND	16
18	4	NR	ND	4	ND	NR	NR	4	4	18
19	4	ND	4	4	3	ND	4	ND	4	19
21	4	ND	4	3	ND	3	4	ND	0	21
22	4	ND	3	3	ND	ND	ND	ND	0	22
23	4	4	4	0	ND	ND	3	ND	ND	23
24	4	ND	ND	4	ND	ND	ND	ND	ND	24
25	3	ND	ND	4	ND	1	0	ND	ND	25
26	4	ND	4	4	ND	1	1	NR	3	26
27	0	ND	ND	ND	ND	ND	ND	ND	ND	27
28	4	ND	4	4	4	4	1	ND	4	28
29	ND	ND	ND	ND	ND	ND	ND	ND	4	29
30	4	4	2	4	4	2	ND	3	4	30
32	0	ND	ND	4	ND	NR	NR	ND	ND	32
33	0	ND	3	0	3	2	3	0	4	33
34	0	NR	ND	0	ND	0	3	ND	0	34
35	4	ND	ND	4	ND	ND	ND	ND	ND	35
39	0	NR	NR	0	NR	4	1	1	4	39
40	0	ND	ND	2	ND	NR	NR	ND	4	40
41	4	NR	3	4	NR	ND	ND	ND	3	41
44	3	NR	NR	3	ND	ND	4	ND	3	44
45	4	ND	2	4	ND	3	3	ND	4	45
46	4	ND	3	3	ND	ND	0	ND	0	46
50	4	NR	ND	4	0	NR	3	NR	2	50
51	1	ND	ND	ND	NR	1	0	NR	ND	51
52	ND	2	ND	ND	ND	ND	ND	ND	ND	52
54	ND	ND	ND	2	ND	ND	ND	ND	4	54

Table 3 Standard Reference Water Sample No. 193
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.
0 (Poor) > 2.00 Std. Dev.
ND Not determined
NR Not rated

LAB	FE	HG	LI	MN	MO	NI	PB	SE	SZ
55	4	ND	VD	4	ND	ND	VD	ND	3
56	4	4	VD	4	ND	ND	ND	ND	ND
57	3	0	VD	2	ND	4	4	3	ND
58	3	4	ND	4	ND	0	4	3	0
59	4	ND	3	4	ND	ND	ND	ND	4
60	ND	NR	ND	ND	ND	4	4	ND	ND
61	4	0	NR	3	3	3	0	4	0
62	1	0	ND	0	ND	NR	4	2	ND
64	ND	NR	ND	ND	ND	2	1	ND	ND
65	2	ND	ND	3	ND	NR	4	4	ND
67	3	ND	ND	3	ND	ND	ND	4	ND
69	0	0	ND	2	ND	1	0	ND	ND
70	2	4	ND	0	ND	ND	1	3	ND
71	4	NR	NR	4	NR	NR	NR	ND	CN
72	3	NR	NR	3	NR	NR	4	NR	4
73	4	NR	ND	2	ND	NR	NR	NR	4
75	0	ND	ND	0	ND	ND	NR	ND	ND
76	0	ND	ND	CN	CN	ND	ND	ND	ND
77	4	3	VD	4	CN	ND	4	3	CN
78	4	NR	CN	CN	CN	NR	ND	4	GN
79	ND	3	VD	ND	ND	ND	4	CN	GN
84	4	NR	VD	4	CN	CN	0	3	GN
86	ND	3	VD	CN	CN	ND	3	4	GN
91	2	ND	VD	0	CN	NR	NR	CN	GN
92	ND	ND	VD	CN	ND	NR	NR	CN	ND
94	2	3	VD	4	CN	4	4	0	GN
95	4	3	0	4	1	NR	4	NR	0
96	4	NR	VD	4	4	4	4	3	4
97	3	NR	ND	2	CN	3	2	0	GN
99	4	4	4	3	3	4	4	4	4
100	4	ND	ND	4	CN	VD	ND	CN	CN
101	4	ND	VD	3	4	4	4	3	0
102	4	4	3	3	4	4	4	3	0
104	ND	3	VD	CN	ND	3	2	CN	CN
105	2	3	VD	2	CN	3	3	4	GN
107	4	NR	NR	2	NR	NR	NR	0	GN
112	4	ND	NR	2	1	1	1	CN	4
116	2	0	ND	4	CN	4	0	NR	CN
117	3	NR	ND	3	ND	4	ND	2	GN
118	ND	ND	ND	ND	ND	4	4	CN	4

Table 3 Standard Reference Water Sample No. T93
Overall Laboratory Performance

RATINGS 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.
 3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined
 2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated
 1 (Questionable) 1.51 to 2.00 Std. Dev.

AB	TL	ZN	N	Avg.
1	3	4	20	2.90
2	ND	ND	8	2.63
3	ND	ND	2	4.00
4	ND	4	20	2.65
5	ND	0	3	2.88
6	2	0	21	2.05
7	ND	3	12	2.50
8	NR	3	12	2.92
9	ND	3	3	2.75
10	ND	4	14	3.50
12	NR	4	17	3.53
13	ND	ND	2	0.00
14	ND	3	14	3.07
15	ND	ND	4	1.75
16	ND	ND	1	4.00
18	ND	4	7	3.86
19	ND	4	11	3.55
21	ND	4	14	2.93
22	ND	4	11	2.91
23	ND	4	9	3.33
24	ND	4	7	3.86
25	ND	4	3	2.00
26	ND	3	15	2.50
27	ND	ND	1	0.00
28	1	4	17	3.47
29	ND	ND	3	2.33
30	NR	4	17	3.53
32	ND	1	10	2.20
33	0	0	19	2.21
34	ND	0	11	1.27
35	ND	ND	2	4.00
39	3	2	17	2.94
40	ND	0	12	1.58
41	NR	4	11	3.64
44	NR	4	12	3.67
45	ND	3	17	3.47
46	ND	1	12	1.75
50	3	3	15	2.73
51	NR	NR	7	1.71
52	ND	ND	2	2.50
54	ND	4	5	3.60

Table 3 Standard Reference Water Sample No. 193
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.
 3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined
 2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated
 1 (Questionable) 1.51 to 2.00 Std. Dev.

.AB	TL	ZN	N	Avg.
55	ND	4	5	3.80
56	ND	4	7	3.14
57	3	4	16	3.13
58	ND	4	16	2.31
59	ND	4	9	3.78
60	ND	4	6	3.50
61	NR	4	20	2.25
62	ND	4	15	2.47
64	ND	ND	6	1.00
65	ND	3	8	3.38
67	ND	2	6	3.33
69	ND	4	13	1.62
70	ND	4	12	3.08
71	NR	4	6	3.00
72	NR	4	17	3.41
73	ND	NR	9	2.67
75	ND	0	4	0.00
76	ND	ND	1	0.00
77	ND	4	14	3.43
78	ND	ND	8	2.63
79	ND	3	6	2.83
84	ND	1	13	2.62
86	ND	1	11	3.36
91	ND	2	4	1.25
92	ND	3	3	1.00
94	ND	4	16	3.44
95	NR	3	14	2.43
96	NR	4	17	3.53
97	ND	4	11	2.36
99	NR	4	20	3.30
100	ND	ND	2	4.00
101	ND	0	4	2.25
102	4	4	21	3.33
104	ND	4	7	2.85
105	ND	4	13	2.65
107	NR	1	7	2.00
112	ND	4	12	2.53
116	ND	4	16	2.19
117	ND	4	7	3.14
118	4	ND	7	4.00

Table 4 Standard Reference Water Sample No. N15
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.

3 (Poor) > 2.00 Std. Dev.
ND Not determined
NR Not rated

-AB	NH3-N	VO2-N	VO3-N	ORG-N	P, TOTAL	PO4-P	N	AVG.
1	4	4	4	3	3	4	5	3.67
2	ND	ND	ND	ND	0	ND	1	0.00
4	ND	4	3	ND	4	4	4	3.75
5	2	0	0	3	ND	ND	4	1.25
6	1	ND	4	ND	0	0	4	1.25
7	ND	0	0	ND	ND	0	3	0.00
8	0	ND	ND	ND	0	ND	2	0.00
9	1	ND	ND	ND	ND	ND	1	1.00
10	4	2	4	3	4	4	5	3.50
12	4	NR	4	3	4	1	5	3.20
13	1	2	0	2	4	4	6	2.17
14	4	2	ND	4	ND	1	4	2.75
15	2	4	ND	4	3	3	5	3.20
16	4	ND	0	ND	4	ND	3	2.67
17	2	4	4	ND	ND	ND	3	3.33
18	4	ND	3	ND	ND	ND	2	3.50
21	3	2	0	3	4	4	6	2.67
23	ND	4	4	ND	ND	ND	2	4.00
24	ND	ND	3	ND	ND	ND	1	3.00
25	4	2	3	2	2	4	5	2.83
26	2	4	3	4	4	1	6	3.00
27	4	0	4	3	4	4	6	3.17
28	2	ND	ND	NR	3	ND	2	2.50
29	4	4	3	ND	ND	ND	3	3.67
30	2	ND	3	ND	2	4	4	2.75
31	0	ND	ND	ND	4	3	3	2.33
40	NR	ND	ND	NR	ND	ND	0	0.00
41	4	4	3	4	4	4	6	3.83
42	3	4	4	2	4	4	6	3.50
44	0	4	ND	2	4	2	5	2.40
46	0	4	2	1	4	0	6	1.83
50	4	4	4	4	3	0	6	3.17
51	1	4	3	2	4	2	6	2.67
55	4	4	4	4	3	0	5	3.17
56	0	ND	2	4	4	3	5	2.60
57	4	2	3	4	4	4	6	3.50
58	4	4	4	4	0	4	6	3.33
60	ND	4	ND	ND	4	4	3	4.00
61	3	ND	0	ND	4	3	4	2.50
62	4	4	1	4	3	3	5	3.17
64	4	ND	3	ND	3	ND	3	3.33

Table 4 Standard Reference Water Sample No. N15
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.
 3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined
 2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated
 1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	NH3-N	NO2-N	NO3-N	ORG-N	P, TOTAL	PO4-P	N	Avg.
65	2	4	1	2	4	ND	5	2.60
67	4	2	4	ND	ND	0	4	2.50
69	2	4	1	3	2	3	6	2.50
71	0	0	3	1	2	3	6	1.50
72	4	2	2	4	4	2	6	3.00
73	3	0	4	0	3	1	6	1.83
74	2	2	3	ND	2	ND	4	2.25
75	4	4	2	ND	3	4	5	3.40
76	ND	0	3	ND	ND	4	3	2.33
77	4	2	4	3	3	4	6	3.33
79	4	ND	4	3	3	4	5	3.60
86	4	4	3	1	2	3	6	2.83
87	2	2	0	ND	1	0	5	1.00
89	1	4	0	ND	4	3	5	2.40
91	3	4	4	ND	ND	ND	3	3.67
92	ND	4	4	ND	0	0	4	2.00
94	4	ND	ND	0	4	4	4	3.00
95	2	2	3	3	4	4	6	3.00
97	4	4	4	3	4	1	6	3.33
98	1	ND	ND	4	3	4	4	3.00
99	ND	NR	3	ND	3	ND	2	3.00
102	4	2	4	4	3	0	6	2.83
104	2	4	4	ND	0	0	5	2.00
105	4	4	4	2	0	4	6	3.00
107	4	0	3	0	2	ND	5	1.80
109	4	NR	0	2	0	0	5	1.20
111	4	NR	2	3	4	3	5	3.20
114	2	4	3	ND	ND	4	4	3.25

Table 5 Standard Reference Water Sample No. P9
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.
 3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined
 2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated
 1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	ACID	CAC03	AG	C4	C5	CL	CO	CR	TOT	CJ	F	FE
1	3	NR	1	4	ND	NR	NR	3	3	NR	NR	4
5	ND	0	ND	0	3	ND	ND	4	ND	4	4	ND
6	ND	4	3	4	3	4	4	2	4	0	0	4
8	ND	NR	1	NR	NR	NR	NR	NR	NR	NR	0	NR
9	ND	2	3	NR	NR	NR	NR	NR	NR	NR	4	NR
10	ND	3	3	4	3	ND	ND	3	1	3	3	3
11	ND	ND	4	ND	1	ND	ND	ND	ND	ND	ND	ND
12	4	4	3	4	3	ND	ND	NR	NR	NR	NR	NR
13	ND	ND	0	ND	ND	ND	ND	ND	ND	3	3	0
14	3	3	0	3	3	ND	ND	3	4	4	4	4
15	ND	ND	0	4	ND	ND	ND	0	ND	ND	1	3
16	2	NR	ND	NR	4	ND	ND	ND	ND	ND	4	ND
18	ND	NR	0	4	0	ND	ND	NR	NR	NR	NR	NR
19	ND	ND	4	4	ND	NR	NR	ND	NR	NR	ND	NR
20	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
23	ND	3	0	4	ND	2	2	ND	ND	2	2	3
30	ND	0	4	4	ND	ND	ND	NR	3	ND	4	4
31	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
32	4	0	3	NR	4	ND	ND	NR	0	ND	3	ND
33	ND	2	0	4	ND	4	4	3	0	ND	ND	0
37	ND	ND	4	ND	ND	ND	ND	ND	ND	ND	ND	ND
41	ND	NR	4	NR	NR	NR	NR	NR	NR	NR	NR	NR
44	ND	3	3	4	NR	ND	ND	3	4	NR	NR	NR
48	ND	NR	2	ND	NR	ND	ND	NR	3	NR	NR	NR
54	ND	ND	3	ND	ND	ND	ND	ND	ND	ND	ND	ND
57	ND	ND	3	4	1	ND	ND	NR	3	ND	ND	4
64	ND	4	ND	4	4	ND	ND	NR	0	4	4	ND
65	ND	ND	4	4	NR	ND	ND	NR	4	ND	ND	NR
71	3	NR	4	NR	NR	NR	NR	NR	NR	NR	NR	NR
72	ND	4	3	4	4	2	2	3	4	NR	NR	4
75	ND	ND	1	ND	ND	ND	ND	ND	ND	ND	ND	ND
77	4	2	3	4	4	0	0	4	1	3	3	3
82	ND	ND	4	ND	ND	ND	ND	ND	ND	ND	ND	ND
86	ND	ND	ND	4	3	ND	ND	4	3	2	2	ND
91	ND	NR	0	NR	4	ND	ND	NR	0	NR	NR	NR
94	4	4	4	3	0	ND	ND	3	4	NR	NR	NR
95	1	NR	2	1	0	NR	NR	NR	NR	NR	NR	NR
99	3	NR	4	1	ND	NR	NR	NR	3	ND	ND	NR
101	ND	ND	1	ND	ND	ND	ND	ND	ND	ND	ND	ND
107	1	NR	3	NR	4	NR	NR	NR	NR	1	NR	NR
111	ND	ND	ND	ND	NR	ND	ND	ND	ND	ND	ND	ND
112	ND	NR	4	2	ND	0	0	ND	2	ND	ND	3
115	ND	NR	4	0	ND	0	0	0	0	4	4	4
118	ND	ND	ND	3	ND	ND	ND	ND	4	ND	ND	ND

Table 5 Standard Reference Water Sample No. P9
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.

0 (Poor) > 2.00 Std. Dev.
ND Not determined
NR Not rated

LAB	K	MG	MN	NA	NH3-N	PB	PH	SO4	SP. COND.	TL
1	4	1	NR	4	ND	NR	3	3	ND	ND
5	1	ND	0	ND	ND	0	4	ND	3	ND
6	4	4	4	3	0	4	4	3	3	ND
8	4	1	1	4	0	NR	4	NR	4	NR
9	NR	NR	2	NR	0	2	ND	NR	ND	ND
10	0	0	3	4	ND	ND	3	3	4	ND
11	4	4	ND	3	ND	ND	ND	0	ND	ND
12	4	4	2	3	3	3	3	NR	4	NR
13	0	ND	ND	4	3	ND	0	ND	ND	NR
14	ND	NR	4	4	ND	3	3	NR	0	ND
15	NR	NR	ND	0	4	ND	3	4	4	ND
16	0	ND	ND	0	3	ND	0	3	3	ND
18	2	0	NR	1	NR	NR	3	NR	ND	ND
19	ND	4	4	NR	ND	NR	ND	ND	ND	ND
20	ND	ND	ND	ND	0	ND	3	ND	ND	ND
23	2	4	4	4	ND	3	2	3	0	ND
30	4	4	4	1	3	ND	4	ND	3	NR
31	ND	ND	ND	ND	3	ND	3	ND	ND	ND
32	2	0	NR	4	ND	NR	1	0	4	ND
33	3	0	0	1	ND	3	ND	ND	ND	ND
37	4	4	ND	4	4	ND	4	ND	3	ND
41	4	4	NR	4	3	NR	ND	ND	ND	NR
44	NR	NR	4	NR	3	3	3	NR	4	ND
48	NR	NR	ND	4	3	ND	3	ND	4	ND
54	4	4	4	2	4	ND	3	3	4	ND
57	2	4	4	4	NR	3	3	NR	0	NR
64	ND	ND	ND	ND	3	3	3	ND	4	ND
65	NR	NR	NR	NR	ND	NR	0	NR	4	ND
71	4	1	4	2	ND	NR	3	4	4	NR
72	4	4	4	3	4	4	3	3	ND	NR
75	4	0	0	3	4	ND	1	ND	ND	ND
77	2	4	4	3	3	4	3	3	4	ND
82	4	4	ND	3	3	ND	3	ND	3	ND
86	ND	ND	ND	ND	ND	1	4	ND	4	ND
91	NR	NR	NR	NR	3	NR	4	1	1	ND
94	4	4	4	3	NR	3	4	4	4	ND
95	1	1	4	4	NR	4	4	3	3	NR
99	0	4	1	2	ND	3	3	ND	4	NR
101	3	4	ND	3	4	ND	3	ND	ND	ND
107	2	NR	NR	0	0	NR	3	4	3	NR
111	ND	ND	ND	ND	3	ND	ND	ND	0	ND
112	ND	4	3	3	ND	3	ND	ND	ND	ND
115	0	4	4	0	ND	2	4	ND	4	NR
118	ND	ND	ND	ND	ND	3	ND	ND	ND	ND

Table 5 Standard Reference Water Sample No. P9
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.

0 (Poor) > 2.00 Std. Dev.
ND Not determined
NR Not rated

LAB	ZN	N	AVG.
1	4	12	3.08
5	3	11	2.00
6	4	19	3.21
8	4	10	2.30
9	4	8	2.88
10	3	16	2.59
11	ND	6	2.57
12	3	14	3.36
13	ND	7	1.43
14	3	15	2.93
15	ND	10	2.30
16	ND	9	2.11
18	NR	7	1.43
19	NR	4	4.00
20	ND	2	1.50
23	3	15	2.60
30	4	13	3.23
31	ND	2	3.00
32	2	13	2.08
33	3	13	1.77
37	ND	7	3.36
41	2	6	3.50
44	0	11	3.09
48	ND	6	3.17
54	0	10	3.10
57	0	13	2.59
64	ND	9	3.22
65	NR	4	3.00
71	3	10	3.20
72	4	17	3.59
75	3	8	2.00
77	4	20	3.10
82	ND	7	3.43
86	3	9	3.11
91	NR	7	1.36
94	3	16	3.44
95	3	13	2.38
99	4	12	2.67
101	ND	6	3.00
107	NR	10	2.10
111	ND	2	1.50
112	3	10	2.70
115	3	15	2.20
113	ND	3	3.33

Table 6 Standard Reference Water Sample No. AMW2
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	ACID@CAC03	AG	AL	AS	BA	BE	CD	CL	CO	CR TOT
1	4	NR	4	0	NR	2	4	2	3	2
5	ND	0	2	4	ND	ND	ND	2	ND	3
8	ND	NR	4	3	NR	3	4	ND	3	NR
9	ND	NR	ND	ND	NR	ND	3	4	ND	NR
10	ND	NR	4	4	3	ND	2	4	ND	3
12	2	NR	0	3	0	0	2	4	ND	3
19	ND	NR	ND	ND	4	3	4	ND	4	ND
20	ND	NR	0	ND	ND	ND	ND	ND	ND	ND
21	0	NR	ND	ND	ND	2	3	4	1	1
22	ND	3	4	0	4	ND	2	ND	3	3
23	3	NR	ND	ND	ND	ND	ND	ND	ND	ND
24	ND	NR	4	ND	NR	ND	4	3	ND	ND
30	ND	NR	4	3	4	4	3	4	ND	4
31	ND	NR	ND	ND	ND	ND	ND	ND	ND	ND
33	ND	NR	4	0	4	0	3	ND	2	0
44	ND	NR	3	ND	NR	ND	ND	4	ND	2
45	4	NR	4	4	NR	4	4	4	ND	2
56	ND	NR	ND	ND	ND	ND	2	1	0	1
57	ND	NR	4	4	1	4	3	ND	ND	4
59	ND	NR	4	ND	4	4	ND	4	ND	ND
62	0	NR	ND	3	NR	0	0	ND	ND	4
65	ND	NR	4	4	NR	ND	4	4	ND	NR
71	4	NR	4	3	NR	NR	4	ND	4	4
72	ND	NR	4	3	4	3	0	4	4	4
73	ND	NR	4	2	NR	ND	4	ND	ND	NR
75	ND	NR	ND	ND	NR	ND	0	ND	ND	3
77	3	NR	4	3	ND	ND	3	2	2	3
91	ND	NR	ND	ND	ND	ND	3	4	ND	2
92	0	NR	ND	ND	ND	ND	ND	0	ND	ND
94	3	NR	4	3	3	4	4	0	ND	3
95	4	NR	3	1	NR	2	3	4	4	NR
99	ND	NR	3	ND	4	4	4	ND	4	4
107	3	NR	0	4	0	NR	2	2	3	2

Table 6 Standard Reference Water Sample No. AMW2
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.
 3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined
 2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated
 1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	CU	F	FE	HG	LI	MN	MO	NI	NO3-N	P3
1	2	4	1	3	2	3	3	0	ND	4
5	ND	3	ND	ND	ND	3	ND	ND	ND	2
3	4	3	3	ND	ND	0	NR	3	ND	3
9	4	ND	2	ND	ND	1	ND	4	ND	0
10	4	3	3	ND	ND	1	VD	2	1	4
12	1	2	3	3	ND	2	1	3	NR	4
19	3	ND	4	ND	3	3	NR	ND	ND	NR
20	ND	ND	2	ND	ND	ND	ND	ND	ND	ND
21	3	ND	3	ND	3	4	ND	2	3	3
22	4	ND	3	ND	3	4	ND	3	ND	ND
23	ND	3	0	3	ND	ND	ND	ND	3	3
24	3	ND	4	ND	ND	3	VD	ND	NR	ND
30	3	4	3	3	4	3	3	4	3	ND
31	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
33	3	ND	4	ND	2	2	4	3	ND	0
44	ND	2	0	NR	ND	0	ND	ND	NR	ND
45	1	ND	2	3	ND	1	ND	4	NR	4
56	1	ND	3	1	ND	3	VD	0	3	ND
57	3	ND	3	NR	ND	4	ND	3	3	4
59	4	2	3	ND	4	3	ND	ND	ND	ND
62	3	ND	0	4	ND	2	VD	3	VD	4
65	3	ND	3	ND	ND	2	ND	3	ND	2
71	3	NR	2	NR	NR	4	NR	3	ND	3
72	0	3	4	VD	4	2	NR	3	3	3
73	0	ND	3	NR	ND	0	NR	0	ND	NR
75	0	ND	0	ND	ND	0	ND	0	ND	1
77	4	4	3	3	ND	3	ND	4	3	2
91	3	ND	2	ND	ND	0	ND	0	ND	NR
92	ND	ND	ND	ND	ND	ND	ND	ND	3	ND
94	4	2	4	3	ND	0	VD	3	ND	3
95	0	2	2	2	0	4	NR	4	ND	4
99	4	ND	3	NR	4	2	4	1	NR	0
107	1	2	0	NR	ND	3	NR	4	1	NR

Table 6 Standard Reference Water Sample No. AMW2
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.
3 (Good) 0.51 to 1.00 Std. Dev.
2 (Satisfactory) 1.01 to 1.50 Std. Dev.
1 (Questionable) 1.51 to 2.00 Std. Dev.

0 (Poor) > 2.00 Std. Dev.
ND Not determined
NR Not rated

LAB	PH	SE	SI02	S04	SP. COND.	SR	ZN	N	Avg.
1	3	NR	1	3	4	2	4	23	2.61
5	1	ND	ND	ND	0	ND	0	11	1.82
8	4	NR	2	4	3	4	4	17	3.18
9	ND	ND	ND	3	ND	ND	4	9	2.78
10	0	ND	4	4	2	ND	3	18	2.83
12	4	NR	3	4	4	ND	3	21	2.43
19	ND	ND	4	ND	ND	ND	ND	9	3.56
20	ND	ND	ND	ND	ND	ND	ND	2	1.00
21	3	ND	ND	3	3	ND	4	18	2.61
22	ND	1	4	ND	ND	2	4	16	2.94
23	4	ND	0	4	0	ND	ND	11	2.36
24	ND	ND	3	3	ND	ND	4	9	3.44
30	3	3	4	4	2	3	4	23	3.43
31	4	ND	ND	ND	ND	ND	ND	1	4.00
33	ND	0	4	ND	ND	4	4	19	2.26
44	3	ND	0	0	4	0	0	12	1.50
45	4	NR	2	3	3	4	3	19	3.16
56	4	ND	ND	ND	ND	ND	4	12	1.92
57	3	NR	ND	3	ND	ND	4	15	3.33
59	4	ND	4	3	ND	2	4	14	3.50
62	2	3	ND	ND	ND	ND	3	14	2.21
65	2	NR	ND	4	1	ND	4	13	3.08
71	4	ND	2	ND	2	ND	3	15	3.27
72	3	4	4	1	ND	1	4	23	2.96
73	ND	ND	ND	ND	ND	ND	3	8	2.00
75	1	ND	ND	0	ND	ND	2	11	0.82
77	3	3	4	4	0	ND	0	22	2.91
91	2	ND	ND	4	1	ND	3	11	2.18
92	2	ND	ND	0	4	ND	ND	6	1.50
94	4	NR	4	4	ND	ND	4	19	3.11
95	4	NR	4	3	3	3	4	21	2.86
99	2	ND	4	ND	ND	3	3	17	3.12
107	3	NR	4	4	4	ND	3	19	2.37

Table 7 Standard Reference Water Sample 44 Report for ALK(CAC03)

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	27.0	-0.6	TITRATION, ELECTROMETRIC, AUTOMATED	4
2	58.0	113.5	REJECT TITRATION, COLORIMETRIC, MANUAL	1,2
3	27.0	-0.6	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
4	26.0	-4.3	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
5	26.1	-3.9	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
6	33.3	22.6	TITRATION, COLORIMETRIC, MANUAL	1,2
7	26.0	-4.3	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
8	64.0	135.6	REJECT TITRATION, ELECTROMETRIC, AUTOMATED	4
10	26.0	-4.3	TITRATION, COLORIMETRIC, MANUAL	1,2
12	26.0	-4.3	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
13	40.0	47.3	TITRATION, COLORIMETRIC, MANUAL	1,2
14	28.9	6.4	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
15	15.3	-40.0	TITRATION, COLORIMETRIC, AUTOMATED	3
16	25.1	-7.6	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
18	32.0	17.8	TITRATION, COLORIMETRIC, MANUAL	1,2
20	28.0	3.1	TITRATION, COLORIMETRIC, MANUAL	1,2
21	26.0	-4.3	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
22	27.1	-0.2	TITRATION, ELECTROMETRIC, AUTOMATED	4
23	4.0	-85.3	REJECT TITRATION, COLORIMETRIC, AUTOMATED	3
24	28.0	3.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
26	26.8	-1.3	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
27	28.0	3.1	TITRATION, COLORIMETRIC, MANUAL	1,2
28	26.0	-4.3	TITRATION, COLORIMETRIC, AUTOMATED	3
29	28.0	3.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
30	51.1	88.1	REJECT TITRATION, COLORIMETRIC, MANUAL	1,2
31	25.0	-8.0	TITRATION, COLORIMETRIC, MANUAL	1,2
32	27.1	-0.2	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
34	13.2	-51.4	TITRATION, COLORIMETRIC, AUTOMATED	3
35	26.5	-2.4	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
37	27.0	-0.6	TITRATION, COLORIMETRIC, AUTOMATED	3
39	26.0	-4.3	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
41	27.1	-0.2	TITRATION, ELECTROMETRIC, AUTOMATED	4
44	26.0	-4.3	TITRATION, COLORIMETRIC, MANUAL	1,2
45	28.0	3.1	OTHER	
46	28.5	4.9	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
48	28.0	3.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
50	24.0	-11.6	TITRATION, COLORIMETRIC, MANUAL	1,2
51	31.0	14.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
52	28.0	3.1	TITRATION, ELECTROMETRIC, AUTOMATED	4
53	27.0	-0.6	OTHER	
55	26.4	-2.8	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
56	23.0	-15.3	TITRATION, COLORIMETRIC, MANUAL	1,2
57	27.0	-0.6	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
58	27.2	0.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
59	25.2	-7.2	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4

Table 7 Standard Reference Water Sample M4 Report for ALK(CAC03)

Code Number	Reported value	Pct. dev. from mean	Methods	References
61	34.0	25.2	TITRATION, COLORIMETRIC, AUTOMATED	3
62	27.0	-0.6	TITRATION, COLORIMETRIC, MANUAL	1,2
64	27.0	-0.6	TITRATION, COLORIMETRIC, MANUAL	1,2
69	27.0	-0.6	TITRATION, ELECTROMETRIC, AUTOMATED	4
70	14.0	-48.5	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
71	36.0	32.5	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
72	28.0	3.1	TITRATION, ELECTROMETRIC, AUTOMATED	4
73	25.0	-8.0	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
76	28.0	3.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
77	28.0	3.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
84	26.0	-4.3	OTHER	
86	28.0	3.1	TITRATION, COLORIMETRIC, AUTOMATED	3
87	34.0	25.2	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
89	26.0	-4.3	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
90	28.0	3.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
91	28.7	5.7	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
92	14.0	-48.5	NOT REPORTED	
94	27.8	2.4	TITRATION, COLORIMETRIC, AUTOMATED	3
95	37.2	37.0	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
96	27.0	-0.6	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
97	24.5	-9.8	NOT REPORTED	
99	28.2	3.8	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
100	27.0	-0.6	TITRATION, ELECTROMETRIC, AUTOMATED	4
102	21.5	-20.8	TITRATION, COLORIMETRIC, MANUAL	1,2
105	31.0	14.1	TITRATION, COLORIMETRIC, MANUAL	1,2
107	40.0	47.3	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
109	27.0	-0.6	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
114	43.3	59.4	REJECT TITRATION, COLORIMETRIC, MANUAL	1,2
115	25.4	-6.5	TITRATION, ELECTROMETRIC, AUTOMATED	4

74 Labs had a total range of 4.0 to 64.0 and a mean of 27.16 with a standard deviation of 4.65 and a 95% confidence interval of the mean +/- 1.12.

Table 7 Standard Reference Water Sample M4 Report for B

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 10		IGNORED EMISSION, DC PLASMA	
4	< 0		IGNORED EMISSION, IC PLASMA	
6	50	220.1	OTHER	
7	0	-100.0	COLORIMETRIC, CARMINE (CARMINIC ACID)	2,4
8	< 50		COLORIMETRIC, CURCUMIN	1,2,3,4
12	< 50		IGNORED EMISSION, IC PLASMA	
13	10	-36.0	COLORIMETRIC, AZOMETHINE	5
14	23	47.3	COLORIMETRIC, CURCUMIN	1,2,3,4
22	10	-36.0	EMISSION, IC PLASMA	
24	< 10		IGNORED EMISSION, IC PLASMA	
30	40	156.1	EMISSION, IC PLASMA	
33	8	-48.8	EMISSION, IC PLASMA	
39	10	-36.0	EMISSION, IC PLASMA	
41	< 50		IGNORED COLORIMETRIC, CURCUMIN	1,2,3,4
44	< 50		IGNORED COLORIMETRIC, CURCUMIN	1,2,3,4
52	15	-4.0	COLORIMETRIC, AZOMETHINE	5
53	20	23.0	COLORIMETRIC, CARMINE (CARMINIC ACID)	2,4
58	15	-4.0	EMISSION, DC PLASMA	
61	30	92.1	COLORIMETRIC, CURCUMIN	1,2,3,4
67	< 100		IGNORED COLORIMETRIC, AZOMETHINE	5
71	< 1000		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	7
72	< 250		IGNORED COLORIMETRIC, CURCUMIN	1,2,3,4
74	0	-100.0	COLORIMETRIC, CURCUMIN	1,2,3,4
84	40	156.1	COLORIMETRIC, CURCUMIN	1,2,3,4
89	130	732.3	REJECT COLORIMETRIC, CURCUMIN	1,2,3,4
90	0	-100.0	COLORIMETRIC, CARMINE (CARMINIC ACID)	2,4
95	0	-100.0	COLORIMETRIC, CURCUMIN	1,2,3,4
96	13	-16.8	EMISSION, IC PLASMA	
99	< 100		IGNORED COLORIMETRIC, DIANTHRIMIDE	4
100	14	-10.4	COLORIMETRIC, CARMINE (CARMINIC ACID)	2,4
107	< 10		IGNORED EMISSION, IC PLASMA	
109	20	23.0	COLORIMETRIC, CURCUMIN	1,2,3,4
112	8	-48.8	EMISSION, IC PLASMA	
115	2	-37.2	EMISSION, IC PLASMA	

34 Labs had a total range of 0 to 130 and a mean of 15.6 with a standard deviation of 14.2 and a 95% confidence interval of the mean +/- 6.5.

Table 7 Standard Reference Water Sample M4 Report for BR

Code Number	Reported value	Pct. dev. from mean	Methods	References
41	< 20		IGNORED OTHER	
99	< 100		IGNORED OTHER	
112	10	0.0	OTHER	

Table 7 Standard Reference Water Sample M4 Report for CA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	11.0	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	10.0	-12.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
3	9.3	-14.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	10.0	-12.5	TITRATION, EDTA	1,3
5	7.3	-36.1	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	11.0	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	12.0	5.0	TITRATION, EDTA	1,3
8	12.0	5.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	5.6	-51.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	11.4	-3.2	EMISSION, IC PLASMA	3,5,7
12	12.0	5.0	EMISSION, IC PLASMA	3,5,7
13	16.0	40.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	12.2	6.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
18	10.0	-12.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	11.0	-3.7	EMISSION, IC PLASMA	3,5,7
20	12.0	5.0	TITRATION, EDTA	1,3
21	12.4	8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	12.0	5.0	EMISSION, IC PLASMA	3,5,7
23	12.2	6.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
24	11.0	-3.7	EMISSION, IC PLASMA	3,5,7
25	12.8	12.0	TITRATION, EDTA	1,3
26	12.2	6.8	TITRATION, EDTA	1,3
27	11.2	-2.0	TITRATION, EDTA	1,3
28	12.0	5.0	EMISSION, IC PLASMA	3,5,7
29	11.0	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	11.3	-1.1	EMISSION, IC PLASMA	3,5,7
32	11.3	-1.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	11.4	-0.2	EMISSION, IC PLASMA	3,5,7
34	10.4	-9.0	TITRATION, EDTA	1,3
35	11.4	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	9.2	-19.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
39	13.0	13.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
41	11.6	1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	11.0	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
45	10.7	-5.3	EMISSION, IC PLASMA	3,5,7
46	11.6	1.5	TITRATION, EDTA	1,3
48	12.0	5.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
49	11.3	-1.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
50	11.0	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	11.0	-3.7	EMISSION, IC PLASMA	3,5,7
52	11.0	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
53	11.2	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
55	11.0	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	5.7	-50.1	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	11.0	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

Table 7 Standard Reference Water Sample M4 Report for CA

Code Number	Reported value	Pct. dev. from mean	Methods	References
58	11.5	0.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	11.0	-3.7	EMISSION, IC PLASMA	3,5,7
61	10.9	-4.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
62	7.8	-31.7	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	12.4	8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	13.4	17.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	11.9	4.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
72	10.3	-5.5	EMISSION, IC PLASMA	3,5,7
73	10.3	-9.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	10.0	-12.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
76	9.9	-13.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	12.0	5.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	13.5	13.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	12.4	3.5	EMISSION, IC PLASMA	3,5,7
88	10.8	-5.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
89	12.0	5.0	TITRATION, EDTA	1,3
90	11.0	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	12.0	5.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
92	15.5	35.7	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
94	11.0	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	11.0	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	11.9	4.2	EMISSION, IC PLASMA	3,5,7
97	10.7	-6.3	NOT REPORTED	
99	12.0	5.0	EMISSION, IC PLASMA	3,5,7
100	12.0	5.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
102	15.5	35.7	REJECT TITRATION, EDTA	1,3
105	10.5	-8.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
107	12.3	7.7	EMISSION, IC PLASMA	3,5,7
109	20.0	75.1	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	12.0	5.0	EMISSION, IC PLASMA	3,5,7
115	14.0	22.5	EMISSION, IC PLASMA	3,5,7
117	11.5	0.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

77 Labs had a total range of 5.6 to 20.0 and a mean of 11.42 with a standard deviation of 0.90 and a 95% confidence interval of the mean +/- 0.22.

Table 7 Standard Reference Water Sample M4 Report for CL

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	3.1	7.3	NOT REPORTED	
3	2.7	-6.5	TITRATION, SILVER NITRATE	1,2,4
4	3.0	3.8	TITRATION, MERCURIC NITRATE	1,2,3,4
5	2.3	-20.4	ION SELECTIVE ELECTRODE	1,2,3,4
6	1.4	-51.5	ION CHROMATOGRAPHY	2,3,6
7	3.1	7.3	TITRATION, MERCURIC NITRATE	1,2,3,4
8	< 5.0		IGNORED	1,2,4
9	9.0	211.5	REJECT	1,2,3,4
10	2.0	-30.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,4
12	2.7	-6.5	TITRATION, SILVER NITRATE	2,3,6
13	3.3	14.2	ION CHROMATOGRAPHY	1,2,3,4
16	4.9	69.6	TITRATION, MERCURIC NITRATE	1,2,3,4
18	1.5	-48.1	ION SELECTIVE ELECTRODE	1,2,4
20	< 5.0		IGNORED	1,2,4
21	2.7	-6.5	TITRATION, SILVER NITRATE	1,2,3,4
22	2.6	-10.0	TITRATION, MERCURIC NITRATE	1,2,3,4
24	3.1	7.3	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
25	8.0	175.9	REJECT	1,2,3,4
26	3.0	3.8	ION SELECTIVE ELECTRODE	1,2,4
28	2.6	-10.0	TITRATION, SILVER NITRATE	1,2,3,4
29	4.0	38.5	COLORIMETRIC, FERRIC THIOCYANATE	1,2,4
30	2.8	-3.1	TITRATION, SILVER NITRATE	2,3,6
32	2.9	0.4	ION CHROMATOGRAPHY	1,2,3,4
34	2.1	-27.3	TITRATION, MERCURIC NITRATE	1,2,3,4
35	3.5	21.1	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
39	3.0	3.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
40	2.7	-6.5	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
41	2.8	-3.1	TITRATION, MERCURIC NITRATE	1,2,3,4
43	2.9	0.4	COLORIMETRIC, FERRIC THIOCYANATE	1,2,4
44	3.0	3.8	TITRATION, SILVER NITRATE	1,2,3,4
45	2.4	-16.9	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
46	2.9	0.4	TITRATION, SILVER NITRATE	1,2,4
48	2.8	-3.1	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
50	3.0	3.8	TITRATION, SILVER NITRATE	1,2,4
51	1.0	-65.4	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
52	3.0	3.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
53	3.0	3.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
55	2.5	-13.5	TITRATION, MERCURIC NITRATE	1,2,3,4
56	10.3	256.5	REJECT	1,2,3,4
57	3.0	3.8	TITRATION, MERCURIC NITRATE	1,2,3,4
58	2.5	-13.5	TITRATION, MERCURIC NITRATE	1,2,3,4
59	2.7	-6.5	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
61	3.7	28.1	ION CHROMATOGRAPHY	2,3,6
62	2.5	-13.5	NOT REPORTED	1,2,3,4
64	3.0	3.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
			TITRATION, MERCURIC NITRATE	

Table 7 Standard Reference Water Sample M4 Report for CL

Code Number	Reported value	Pct. dev. from mean	Methods	References
65	2.0	-30.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
67	3.0	3.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
69	2.0	-30.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
70	17.9	519.6	REJECT TITRATION, MERCURIC NITRATE	1,2,3,4
71	< 2.0		IGNORED TITRATION, SILVER NITRATE	1,2,4
72	3.0	3.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
73	< 1.0		IGNORED TITRATION, SILVER NITRATE	1,2,4
74	4.2	45.4	TITRATION, MERCURIC NITRATE	1,2,3,4
76	2.0	-30.8	TITRATION, MERCURIC NITRATE	1,2,3,4
77	3.1	7.3	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
79	3.5	21.1	TITRATION, MERCURIC NITRATE	1,2,3,4
84	3.1	7.3	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
86	2.6	-10.0	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
89	3.0	175.9	REJECT TITRATION, MERCURIC NITRATE	1,2,3,4
90	4.0	38.5	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
91	3.2	10.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
92	0.4	-86.2	REJECT TITRATION, MERCURIC NITRATE	1,2,3,4
94	2.8	-3.1	ION SELECTIVE ELECTRODE	1,2,3,4
95	3.0	3.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
96	5.0	73.1	TITRATION, MERCURIC NITRATE	1,2,3,4
97	2.7	-6.5	NOT REPORTED	
99	2.7	-6.5	ION CHROMATOGRAPHY	2,3,6
100	6.6	128.4	REJECT TITRATION, SILVER NITRATE	1,2,4
102	2.1	-27.3	ION CHROMATOGRAPHY	2,3,6
105	3.5	21.1	TITRATION, MERCURIC NITRATE	1,2,3,4
107	3.7	23.1	TITRATION, SILVER NITRATE	1,2,4
109	4.0	38.5	TITRATION, MERCURIC NITRATE	1,2,3,4
111	< 10.0		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
112	2.8	-3.1	ION CHROMATOGRAPHY	2,3,6
115	2.7	-6.5	ION CHROMATOGRAPHY	2,3,6
117	2.5	-13.5	TITRATION, SILVER NITRATE	1,2,4

76 Labs had a total range of 0.4 to 17.9 and a mean of 2.89 with a standard deviation of 0.70 and a 95% confidence interval of the mean +/- 0.18.

Table 7 Standard Reference Water Sample M4 Report for DSRD 180

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	64	-6.8	RESIDUE ON EVAPORATION	4
2	75	9.3	RESIDUE, FILTRABLE	1,3
3	68	-0.9	RESIDUE ON EVAPORATION	4
4	64	-6.8	RESIDUE, FILTRABLE	1,3
5	32	-53.4	RESIDUE, FILTRABLE	1,3
6	74	7.8	RESIDUE, FILTRABLE	1,3
7	70	2.0	RESIDUE ON EVAPORATION	2,4
9	54	-21.3	RESIDUE, FILTRABLE	1,3
10	56	-18.4	RESIDUE, FILTRABLE	1,3
12	79	15.1	RESIDUE, FILTRABLE	3
14	88	28.2	RESIDUE, FILTRABLE	1,3
15	68	-0.9	RESIDUE ON EVAPORATION	2,4
18	64	-6.8	RESIDUE ON EVAPORATION	4
20	65	-5.3	RESIDUE, FILTRABLE	1,3
21	72	4.9	RESIDUE, FILTRABLE	1,3
24	66	-3.8	RESIDUE ON EVAPORATION	4
25	72	4.9	RESIDUE, FILTRABLE	3
26	70	2.0	RESIDUE, FILTRABLE	3
29	73	6.4	RESIDUE, FILTRABLE	1,3
30	67	-2.4	RESIDUE ON EVAPORATION	2,4
32	71	3.4	RESIDUE, FILTRABLE	1,3
34	74	7.8	NOT REPORTED	
35	67	-2.4	RESIDUE, FILTRABLE	3
37	86	25.3	RESIDUE ON EVAPORATION	2,4
39	70	2.0	RESIDUE ON EVAPORATION	4
40	< 0		IGNORED NOT REPORTED	
41	66	-3.8	RESIDUE, FILTRABLE	1,3
44	70	2.0	RESIDUE, FILTRABLE	1,3
45	75	9.3	RESIDUE ON EVAPORATION	2,4
46	76	10.7	RESIDUE ON EVAPORATION	2,4
50	74	7.8	RESIDUE, FILTRABLE	1,3
51	57	-17.0	RESIDUE, FILTRABLE	1,3
52	68	-0.9	RESIDUE, FILTRABLE	1,3
53	66	-3.8	RESIDUE, FILTRABLE	3
55	68	-0.9	RESIDUE ON EVAPORATION	4
56	66	-3.8	RESIDUE ON EVAPORATION	2,4
57	68	-0.9	RESIDUE, FILTRABLE	1,3
58	72	4.9	RESIDUE, FILTRABLE	1,3
61	57	-17.0	RESIDUE ON EVAPORATION	2,4
62	80	16.6	RESIDUE, FILTRABLE	1,3
65	80	16.6	RESIDUE ON EVAPORATION	4
67	77	12.2	RESIDUE ON EVAPORATION	4
69	92	34.0	RESIDUE, FILTRABLE	1,3
71	56	-18.4	RESIDUE, FILTRABLE	1,3
72	56	-18.4	RESIDUE, FILTRABLE	3

Table 7 Standard Reference Water Sample M4 Report for DSRD 130

Code Number	Reported value	Pct. dev. from mean		Methods	References
73	124	80.7	REJECT	RESIDUE, FILTRABLE	1,3
74	54	-21.3		RESIDUE ON EVAPORATION	2,4
77	88	23.2		RESIDUE, FILTRABLE	1,3
86	58	-15.5		RESIDUE, FILTRABLE	1,3
89	67	-2.4		RESIDUE ON EVAPORATION	2,4
90	73	6.4		RESIDUE, FILTRABLE	1,3
92	11	-34.0	REJECT	RESIDUE, FILTRABLE	1,3
94	69	0.5		RESIDUE, FILTRABLE	1,3
95	54	-21.3		RESIDUE ON EVAPORATION	2,4
96	68	-0.9		RESIDUE ON EVAPORATION	2,4
97	55	-19.9		NOT REPORTED	
99	35	-49.0		RESIDUE ON EVAPORATION	4
100	65	-5.3		RESIDUE ON EVAPORATION	4
102	96	39.9		RESIDUE ON EVAPORATION	2,4
107	96	39.9		RESIDUE ON EVAPORATION	2,4
109	0	-100.0	REJECT	RESIDUE, FILTRABLE	1,3
114	70	2.0		RESIDUE ON EVAPORATION	2,4

62 Labs had a total range of 0 to 124 and a mean of 68.6 with a standard deviation of 11.9 and a 95% confidence interval of the mean +/- 3.1.

Table 7 Standard Reference Water Sample M4 Report for F

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.46	-1.9	ION SELECTIVE ELECTRODE	1,2,3,4
3	0.60	28.0	COLORIMETRIC, ZIRCONIUM ERIOCHROME	4
4	0.60	-14.7	ION SELECTIVE ELECTRODE	1,2,3,4
5	0.35	-25.3	OTHER	
6	0.47	0.3	ION CHROMATOGRAPHY	2,3,6
7	0.47	0.3	OTHER	
8	0.47	0.3	OTHER	
10	0.45	-4.0	OTHER	2,3,6
12	0.43	-8.3	ION CHROMATOGRAPHY	
13	0.47	0.3	OTHER	
14	0.48	2.4	OTHER	1,2,3
15	0.66	40.8	COLORIMETRIC, SPADNS	
16	0.43	2.4	OTHER	
18	0.45	-4.0	OTHER	
23	0.43	-8.3	COLORIMETRIC, ZIRCONYL ALIZARIN	1
25	0.50	6.7	OTHER	
29	0.44	-6.1	OTHER	
30	0.43	2.4	ION CHROMATOGRAPHY	2,3,6
35	0.42	-10.4	OTHER	3
39	0.44	-5.1	COLORIMETRIC, CEROUS ALIZARIN "COMPLEXONE"	
41	0.40	-14.7	OTHER	
44	0.44	-6.1	OTHER	1,2,3,4
45	0.42	-10.4	ION SELECTIVE ELECTRODE	
46	0.47	0.3	OTHER	
50	0.45	-4.0	OTHER	
51	1.13	141.1	OTHER	
52	0.50	6.7	OTHER	
53	0.46	-1.9	COLORIMETRIC, ZIRCONYL ALIZARIN	1
55	0.50	6.7	COLORIMETRIC, ZIRCONIUM ERIOCHROME	4
58	0.42	-10.4	COLORIMETRIC, CEROUS ALIZARIN "COMPLEXONE"	3
59	0.48	2.4	ION CHROMATOGRAPHY	2,3,6
61	0.45	-4.0	OTHER	
62	0.54	15.2	OTHER	1,2,3,4
64	0.48	2.4	ION SELECTIVE ELECTRODE	
69	0.46	-1.9	OTHER	
70	0.29	-38.1	OTHER	
71	0.50	6.7	OTHER	1,2,3,4
72	0.45	-4.0	ION SELECTIVE ELECTRODE	
73	0.44	-6.1	OTHER	1,2,3,4
77	0.47	0.3	ION SELECTIVE ELECTRODE	
84	0.42	-10.4	OTHER	
86	0.45	-4.0	OTHER	
90	0.50	6.7	COLORIMETRIC, SPADNS	1,2,3
91	0.63	34.4	OTHER	1,2,3,4
94	0.44	-6.1	ION SELECTIVE ELECTRODE	

Table 7 Standard Reference Water Sample M4 Report for F

Code Number	Reported value	Pct. dev. from mean	Methods	References
95	0.50	6.7	COLORIMETRIC, CEROUS ALIZARIN "COMPLEXONE"	3
96	0.50	6.7	COLORIMETRIC, LANTHANUM ALIZARIN "COMPLEXONE"	1
99	0.47	0.3	ION CHROMATOGRAPHY	2,3,6
100	0.50	6.7	OTHER	
102	0.53	13.1	OTHER	
105	0.45	-4.0	OTHER	
107	0.41	-12.5	ION SELECTIVE ELECTRODE	1,2,3,4
109	0.46	-1.9	OTHER	
112	0.42	-10.4	ION CHROMATOGRAPHY	2,3,6
115	0.65	33.7	COLORIMETRIC, SPADNS	1,2,3
117	0.48	2.4	OTHER	

56 Labs had a total range of 0.29 to 1.13 and a mean of 0.469 with a standard deviation of 0.063 and a 95% confidence interval of the mean +/- 0.017.

Table 7 Standard Reference Water Sample M4 Report for I

Code Number	Reported value	Pct. dev. from mean	Methods	References
62	1	0.0	COLORIMETRIC, CERIC ARSENIOUS OXIDATION	2,4

Table 7 Standard Reference Water Sample M4 Report for K

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1.3	13.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	1.2	5.2	EMISSION, FLAME	1,2
3	1.3	13.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	1.1	-3.6	EMISSION, FLAME	1,2
5	1.2	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	1.2	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	1.2	5.2	EMISSION, FLAME	1,2
8	1.2	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	0.5	-56.2	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	1.0	-12.3	OTHER	
12	1.0	-12.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	1.3	13.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	1.0	-12.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
16	1.2	5.2	ION SELECTIVE ELECTRODE	1,2,3,4
18	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
21	1.4	22.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	1.0	-12.3	OTHER	
23	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
24	1.1	-3.6	OTHER	
25	1.6	40.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
28	1.2	5.2	OTHER	
29	0.8	-29.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	1.0	-12.3	OTHER	
32	1.3	13.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	1.8	57.8	REJECT EMISSION, IC PLASMA	3
34	1.9	66.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	1.2	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
39	1.0	-12.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
40	0.8	-29.9	NOT REPORTED	
41	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
45	1.0	-12.3	OTHER	
46	0.1	-91.2	REJECT EMISSION, FLAME	1,2
48	1.2	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
49	1.2	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	1.2	5.2	EMISSION, FLAME	1,2
51	0.7	-39.6	OTHER	
52	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
53	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
55	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	2.4	110.4	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
58	0.8	-29.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	1.2	5.2	OTHER	
61	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

Table 7 Standard Reference Water Sample M4 Report for K

Code Number	Reported value	Pct. dev. from mean	Methods	References
62	1.2	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
64	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	1.3	13.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	1.2	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	1.0	-12.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	1.5	31.5	EMISSION, FLAME	1,2
75	0.7	-38.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
76	1.2	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	1.0	-12.3	EMISSION, FLAME	1,2
84	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	1.1	-3.6	OTHER	1,2,3,4
88	1.0	-12.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
89	1.2	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
90	1.0	-12.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	1.5	31.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
92	1.5	31.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	1.0	-12.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	1.2	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	1.1	-3.6	OTHER	1,2,3,4
97	1.3	13.9	NOT REPORTED	1,2,3,4
99	2.1	84.1	REJECT	1,2,3,4
100	1.1	-3.6	OTHER	1,2,3,4
105	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
107	1.1	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
109	1.5	31.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
115	1.4	22.7	OTHER	1,2,3,4

72 Labs had a total range of 0.1 to 2.4 and a mean of 1.14 with a standard deviation of 0.18 and a 95% confidence interval of the mean +/- 0.04.

Table 7 Standard Reference Water Sample M4 Report for M3

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	3.3	11.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
3	3.2	8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	3.7	25.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	3.2	8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	1.2	-59.3	REJECT TITRATION, EDTA	2
8	3.0	1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	2.0	-32.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	2.9	-1.7	EMISSION, IC PLASMA	3,5
12	3.3	11.9	EMISSION, IC PLASMA	3,5
13	4.0	35.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
14	3.0	1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	2.6	-11.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
18	2.6	-11.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	3.0	1.7	EMISSION, IC PLASMA	3,5
21	3.5	18.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	2.9	-1.7	EMISSION, IC PLASMA	3,5
23	3.2	8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
24	2.9	-1.7	EMISSION, IC PLASMA	3,5
25	3.9	32.2	TITRATION, EDTA	2
27	2.9	-1.7	TITRATION, EDTA	2
28	3.0	1.7	EMISSION, IC PLASMA	3,5
29	3.0	1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	2.9	-1.7	EMISSION, IC PLASMA	3,5
32	2.6	-11.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	3.1	5.1	EMISSION, IC PLASMA	3,5
34	2.4	-18.6	TITRATION, EDTA	2
35	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	3.0	1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
39	2.5	-15.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
40	2.8	-5.1	NOT REPORTED	
41	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	3.0	1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	2.6	-11.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
45	2.9	-1.7	EMISSION, IC PLASMA	3,5
46	3.4	15.3	TITRATION, EDTA	2
48	3.2	8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
49	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
50	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	0.6	-79.7	REJECT EMISSION, IC PLASMA	3,5
52	3.0	1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
53	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
55	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	1.8	-39.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

Table 7 Standard Reference Water Sample M4 Report for MG

Code Number	Reported value	Pct. dev. from mean	Methods	References
58	0.3	-32.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	3.1	5.1	EMISSION, IC PLASMA	3,5
61	3.0	1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
62	2.5	-15.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	3.5	13.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	3.3	11.9	ATOMIC ABSORPTION, DIRECT, AIR	1,7
72	2.8	-5.1	EMISSION, IC PLASMA	3,5
73	2.6	-11.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	2.0	-32.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
76	3.1	5.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	3.3	11.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	3.2	3.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	3.0	1.7	EMISSION, IC PLASMA	3,5
88	2.8	-5.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
89	2.4	-18.6	TITRATION, EDTA	2
90	3.0	1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
92	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,7
94	2.9	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	2.8	-5.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	3.2	3.5	EMISSION, IC PLASMA	3,5
97	2.6	-11.8	NOT REPORTED	
99	3.0	1.7	EMISSION, IC PLASMA	3,5
100	3.0	1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
102	0.2	-93.2	TITRATION, EDTA	2
105	3.5	13.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
107	3.1	5.1	EMISSION, IC PLASMA	3,5
109	2.5	-15.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	2.8	-5.1	EMISSION, IC PLASMA	3,5
115	3.1	5.1	EMISSION, IC PLASMA	3,5
117	3.0	1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

77 Labs had a total range of 0.2 to 4.0 and a mean of 2.95 with a standard deviation of 0.37 and a 95% confidence interval of the mean +/- 0.09.

Table 7 Standard Reference Water Sample M4 Report for NA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	4.5	-2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	4.4	-5.1	EMISSION, FLAME	1,2
3	4.8	3.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	4.6	-0.8	EMISSION, FLAME	1,2
6	4.4	-5.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	2.3	-53.4	REJECT EMISSION, FLAME	1,2
8	4.7	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	3.3	-28.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	4.4	-5.1	EMISSION, IC PLASMA	3,5
12	4.2	-9.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	4.8	3.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
14	4.4	-5.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	5.6	20.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
16	4.5	-2.9	OTHER	
18	4.6	-0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	4.6	-0.8	EMISSION, IC PLASMA	3,5
20	< 5.0		IGNORED EMISSION, FLAME	1,2
21	4.5	-2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	4.5	-2.9	EMISSION, IC PLASMA	3,5
23	4.9	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
24	4.6	-0.8	EMISSION, IC PLASMA	3,5
25	5.2	12.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
28	4.8	3.5	EMISSION, IC PLASMA	3,5
29	5.0	7.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	4.4	-5.1	EMISSION, IC PLASMA	3,5
32	3.8	-18.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	4.0	-13.7	EMISSION, IC PLASMA	3,5
34	4.6	-0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
35	4.4	-5.1	EMISSION, FLAME	1,2
37	4.5	-2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
39	4.0	-13.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
41	4.6	-0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	4.5	-2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
45	4.6	-0.8	EMISSION, IC PLASMA	3,5
46	5.0	7.8	EMISSION, FLAME	1,2
48	4.9	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
49	5.1	10.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	4.7	1.4	EMISSION, FLAME	1,2
51	< 5.0		IGNORED EMISSION, IC PLASMA	3,5
52	5.0	7.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
53	4.7	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
55	4.7	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	8.2	76.9	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	5.3	14.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
58	6.1	31.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

Table 7 Standard Reference Water Sample M4 Report for NA

Code Number	Reported value	Pct. dev. from mean	Methods	References
59	4.8	3.5	EMISSION, IC PLASMA	3,5
61	4.8	3.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
62	4.4	-5.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
64	4.7	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	5.3	14.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	4.9	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	4.8	3.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	4.5	-2.9	EMISSION, IC PLASMA	3,5
73	5.1	10.0	EMISSION, FLAME	1,2
75	3.2	-31.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
76	4.5	-2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	4.6	-0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	4.4	-5.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	3.4	-26.7	EMISSION, IC PLASMA	3,5
88	5.1	10.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
89	4.6	-0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
90	5.0	7.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	4.9	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
92	4.4	-5.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	4.6	-0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	4.6	-0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	5.0	7.8	EMISSION, IC PLASMA	3,5
97	4.6	-0.8	NOT REPORTED	
99	5.0	7.8	EMISSION, IC PLASMA	3,5
100	4.7	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
105	4.3	-7.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
107	5.7	22.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
109	4.6	-0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	4.4	-5.1	EMISSION, IC PLASMA	3,5
115	4.1	-11.6	EMISSION, IC PLASMA	3,5

75 Labs had a total range of 2.3 to 8.2 and a mean of 4.64 with a standard deviation of 0.47 and a 95% confidence interval of the mean +/- 0.11.

Table 7 Standard Reference Water Sample M4 Report for NO2-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
4	0.03	476.9	COLORIMETRIC, DIAZOTIZATION	1,3,4
5	0.01	92.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
10	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
12	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
14	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
15	< 0.00	92.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
18	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
21	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
23	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
25	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
27	< 0.00		COLORIMETRIC, DIAZOTIZATION	1,3,4
28	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
29	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
37	< 0.01	92.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
39	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
41	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
43	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
44	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
46	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
50	< 0.00		COLORIMETRIC, DIAZOTIZATION	1,3,4
51	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
55	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
56	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
57	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
58	0.03	476.9	COLORIMETRIC, DIAZOTIZATION	1,3,4
60	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
62	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
65	< 0.02		COLORIMETRIC, DIAZOTIZATION	1,3,4
67	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
69	< 0.01	92.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
71	< 0.05		COLORIMETRIC, DIAZOTIZATION	1,3,4
72	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
73	< 0.05		COLORIMETRIC, DIAZOTIZATION	1,3,4
74	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
76	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
77	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
86	0.01	92.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
87	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
89	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
90	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
91	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
92	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
94	< 0.01		COLORIMETRIC, DIAZOTIZATION	1,3,4
95	0.05	861.5	REJECT	1,3,4
96	< 0.02		IGNORED	1,3,4

Table 7 Standard Reference Water Sample 44 Report for N02-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
97	< 0.01		IGNORED NOT REPORTED	
99	< 0.05		COLORIMETRIC, DIAZOTIZATION	1,3,4
102	0.02	284.6	COLORIMETRIC, DIAZOTIZATION	1,3,4
104	0.01	92.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
105	< 0.00		COLORIMETRIC, DIAZOTIZATION	1,3,4
107	0.00	-100.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
109	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
111	< 0.02		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4

53 Labs had a total range of 0.00 to 0.05 and a mean of 0.005 with a standard deviation of 0.009 and a 95% confidence interval of the mean +/- 0.004.

Table 7 Standard Reference Water Sample No. Report for ND3-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
4	0.30	52.7	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
5	0.14	-25.5	OTHER	
6	0.14	-25.5	ION CHROMATOGRAPHY	2,3,6,7
10	0.27	43.8	COLORIMETRIC, BRUCINE	1,2,3,4
12	0.19	1.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
13	0.26	33.4	COLORIMETRIC, BRUCINE	1,2,3,4
14	0.16	-14.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
15	0.19	1.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
16	0.21	11.8	COLORIMETRIC, BRUCINE	1,2,3,4
18	0.20	6.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
20	< 0.50		COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
21	0.08	-57.4	IGNORED	
22	0.19	1.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
23	0.19	1.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
24	0.40	113.0	REJECT	
25	0.19	1.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
26	0.20	5.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
27	0.18	-4.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
28	0.19	1.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
29	0.18	-4.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
30	0.17	-9.5	ION CHROMATOGRAPHY	2,3,6,7
37	0.21	11.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
39	0.18	-4.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
41	0.18	-4.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
43	0.18	-4.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
44	0.14	-25.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
45	0.20	5.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
46	0.19	1.2	COLORIMETRIC, BRUCINE	1,2,3,4
48	0.20	6.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
50	0.19	1.2	COLORIMETRIC, BRUCINE	1,2,3,4
51	0.17	-9.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
52	0.15	-20.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
53	0.13	-4.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
55	0.18	-4.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
56	0.19	-4.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
57	0.19	1.2	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
58	0.16	-14.8	COLORIMETRIC, BRUCINE	1,2,3,4
59	0.86	357.9	ION CHROMATOGRAPHY	2,3,6,7
61	0.26	38.4	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
62	0.07	-62.7	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
64	0.21	11.8	ION CHROMATOGRAPHY	2,3,6,7
65	0.15	-20.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
69	0.17	-9.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
71	0.20	6.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
72	0.18	-4.2	ION CHROMATOGRAPHY	2,3,6,7

Table 7 Standard Reference Water Sample #4 Report for M03-V

Code Number	Reported value	Pct. dev. from mean	Methods	References
73	0.20	6.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
74	0.18	-4.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
75	0.15	-20.1	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
76	0.19	1.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
77	0.22	17.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
79	0.20	6.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
84	0.20	6.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
86	0.17	-9.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
87	0.32	70.4	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
89	0.30	59.7	COLORIMETRIC, BRUCINE	1,2,3,4
90	0.23	49.1	COLORIMETRIC, BRUCINE	1,2,3,4
91	0.18	-4.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
92	0.01	-94.7	REJECT	1,2,3,4
94	0.19	1.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
95	0.20	6.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
96	0.18	-4.2	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	1,2,3,4
97	0.20	6.5	NOT REPORTED	3
99	0.17	-9.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
100	0.10	-46.8	OTHER	1,2,3,4
102	0.16	-14.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
104	0.19	1.2	COLORIMETRIC, BRUCINE	1,2,3,4
105	0.15	-20.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
107	0.21	11.8	COLORIMETRIC, BRUCINE	1,2,3,4
109	0.10	-46.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
111	0.22	17.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
112	0.18	-4.2	ION CHROMATOGRAPHY	2,3,6,7
115	0.18	-4.2	ION CHROMATOGRAPHY	2,3,6,7

72 Labs had a total range of 0.01 to 0.86 and a mean of 0.189 with a standard deviation of 0.044 and a 95% confidence interval of the mean +/- 0.011.

Table 7 Standard Reference Water Sample M4 Report for P_i TOTAL

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 0.01		IGNORED	
2	0.06	222.6	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
4	0.01	-46.2	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4,	PHOSPHOMOLYBDATE 4
6	0.20	975.3	REJECT	
8	< 0.01		IGNORED	
10	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
12	0.02	7.5	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4,	PHOSPHOMOLYBDATE 4
14	0.03	61.3	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4,	PHOSPHOMOLYBDATE 4
15	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
16	0.00	-100.0	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
21	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
25	< 0.05		IGNORED	
27	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
28	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
30	0.03	61.3	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
31	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
37	0.02	7.5	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
39	< 0.02		IGNORED	
40	< 0.03		IGNORED	
41	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
43	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
44	0.02	7.5	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4,	PHOSPHOMOLYBDATE 4
45	0.00	-100.0	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
46	0.07	276.3	REJECT	
48	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
50	< 0.05		IGNORED	
51	0.06	222.6	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
53	0.02	7.5	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
55	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
56	0.08	330.1	REJECT	
57	0.03	61.3	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4,	PHOSPHOMOLYBDATE 4
58	0.01	-46.2	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4,	PHOSPHOMOLYBDATE 4
60	0.02	7.5	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
61	0.03	61.3	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
62	0.02	7.5	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
64	0.01	-46.2	EMISSION, IC PLASMA	3,5
65	< 0.05		IGNORED	
69	0.02	7.5	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4,	PHOSPHOMOLYBDATE 4
71	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
72	0.03	61.3	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
73	< 0.10		IGNORED	
75	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
77	0.02	7.5	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
79	0.05	163.8	COLORIMETRIC, H2SO4/PERSULF	DIG. ASCORBIC ACID PHOSPHOMOLYBD 1,2,3,4
84	< 0.01		IGNORED	

Table 7 Standard Reference Water Sample M4 Report for P, TOTAL

Code Number	Reported value	Pct. dev. from mean	Methods	References
86	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
87	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
89	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
90	0.04	115.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
92	0.01	-46.2	COLORIMETRIC, BLK DIG, H2SO4, K8HG2SO4, PHOSPHOMOLYBDATE	4
95	0.02	7.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
96	< 0.02		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
97	< 0.01		IGNORED	
99	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
102	0.04	115.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
104	0.39	1996.8	OTHER	1,2,3,4
105	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	3,5
107	0.09	383.9	EMISSION, IC PLASMA	4
109	0.30	1512.9	COLORIMETRIC, BLK DIG, H2SO4, K8HG2SO4, PHOSPHOMOLYBDATE	4
111	0.01	-46.2	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
115	< 0.10		IGNORED	3,5

61 Labs had a total range of 0.00 to 0.39 and a mean of 0.019 with a standard deviation of 0.014 and a 95% confidence interval of the mean +/- 0.004.

Table 7 Standard Reference Water Sample M4 Report for PH

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	7.61	1.6	ELECTROMETRIC	1,2,3,4
2	7.70	2.8	ELECTROMETRIC	1,2,3,4
3	7.80	4.2	ELECTROMETRIC	1,2,3,4
4	7.60	1.5	ELECTROMETRIC	1,2,3,4
5	7.42	-0.9	ELECTROMETRIC	1,2,3,4
6	7.80	4.2	ELECTROMETRIC	1,2,3,4
7	6.76	-2.7	ELECTROMETRIC	1,2,3,4
8	7.90	5.5	ELECTROMETRIC	1,2,3,4
10	7.80	4.2	ELECTROMETRIC	1,2,3,4
12	7.73	3.2	ELECTROMETRIC	1,2,3,4
13	7.58	1.2	ELECTROMETRIC	1,2,3,4
14	7.60	1.5	ELECTROMETRIC	1,2,3,4
15	6.85	-3.5	ELECTROMETRIC	1,2,3,4
16	7.81	4.3	ELECTROMETRIC	1,2,3,4
18	6.40	-14.5	ELECTROMETRIC	1,2,3,4
20	7.90	5.5	ELECTROMETRIC	1,2,3,4
21	7.90	5.5	ELECTROMETRIC	1,2,3,4
22	7.20	-3.8	ELECTROMETRIC	1,2,3,4
23	7.30	-2.5	ELECTROMETRIC	1,2,3,4
25	7.80	4.2	ELECTROMETRIC	1,2,3,4
26	7.60	1.5	ELECTROMETRIC	1,2,3,4
27	7.31	-2.4	ELECTROMETRIC	1,2,3,4
28	7.30	-2.5	ELECTROMETRIC	1,2,3,4
29	7.80	4.2	ELECTROMETRIC	1,2,3,4
30	6.51	-13.1	ELECTROMETRIC	1,2,3,4
31	7.44	-0.6	ELECTROMETRIC	1,2,3,4
32	7.70	2.8	ELECTROMETRIC	1,2,3,4
34	7.90	5.5	ELECTROMETRIC	1,2,3,4
35	7.11	-5.0	ELECTROMETRIC	1,2,3,4
37	7.60	1.5	ELECTROMETRIC	1,2,3,4
39	7.60	1.5	ELECTROMETRIC	1,2,3,4
40	6.56	-11.1	ELECTROMETRIC	1,2,3,4
41	7.27	-2.9	ELECTROMETRIC	1,2,3,4
43	7.60	1.5	ELECTROMETRIC	1,2,3,4
44	7.74	3.4	ELECTROMETRIC	1,2,3,4
45	7.73	3.2	ELECTROMETRIC	1,2,3,4
46	7.65	2.2	ELECTROMETRIC	1,2,3,4
48	7.20	-3.8	ELECTROMETRIC	1,2,3,4
49	7.43	-0.8	ELECTROMETRIC	1,2,3,4
50	7.50	0.2	ELECTROMETRIC	1,2,3,4
51	6.60	-11.9	ELECTROMETRIC	1,2,3,4
52	7.80	4.2	ELECTROMETRIC	1,2,3,4
53	7.60	1.5	ELECTROMETRIC	1,2,3,4
55	7.70	2.8	ELECTROMETRIC	1,2,3,4
56	7.10	-5.2	ELECTROMETRIC	1,2,3,4

Table 7 Standard Reference Water Sample M4 Report for PH

Code Number	Reported value	Pct. dev. from mean	Methods	References
57	7.53	3.6	ELECTROMETRIC	1,2,3,4
58	7.90	5.5	ELECTROMETRIC	1,2,3,4
59	7.74	3.4	ELECTROMETRIC	1,2,3,4
60	7.80	4.2	ELECTROMETRIC	1,2,3,4
61	7.30	-2.5	ELECTROMETRIC	1,2,3,4
62	7.42	-3.9	ELECTROMETRIC	1,2,3,4
64	7.70	2.8	ELECTROMETRIC	1,2,3,4
65	7.10	-5.2	ELECTROMETRIC	1,2,3,4
69	7.55	0.8	ELECTROMETRIC	1,2,3,4
70	7.35	-1.8	ELECTROMETRIC	1,2,3,4
71	7.40	-1.2	ELECTROMETRIC	1,2,3,4
72	7.05	-5.8	ELECTROMETRIC	1,2,3,4
73	6.40	-14.5	ELECTROMETRIC	1,2,3,4
74	6.10	-18.5	REJECT ELECTROMETRIC	1,2,3,4
75	8.58	14.6	ELECTROMETRIC	1,2,3,4
76	7.72	3.1	ELECTROMETRIC	1,2,3,4
77	7.54	0.7	ELECTROMETRIC	1,2,3,4
79	7.81	4.3	ELECTROMETRIC	1,2,3,4
84	7.70	2.8	ELECTROMETRIC	1,2,3,4
86	7.49	0.0	ELECTROMETRIC	1,2,3,4
87	7.40	-1.2	ELECTROMETRIC	1,2,3,4
89	7.45	-3.5	ELECTROMETRIC	1,2,3,4
90	7.60	1.5	ELECTROMETRIC	1,2,3,4
91	7.50	0.2	ELECTROMETRIC	1,2,3,4
92	7.15	-4.5	ELECTROMETRIC	1,2,3,4
94	7.50	0.2	ELECTROMETRIC	1,2,3,4
95	7.56	1.0	ELECTROMETRIC	1,2,3,4
96	7.60	1.5	ELECTROMETRIC	1,2,3,4
99	7.75	3.5	ELECTROMETRIC	1,2,3,4
100	7.70	2.8	ELECTROMETRIC	1,2,3,4
102	7.80	4.2	ELECTROMETRIC	1,2,3,4
104	7.32	-2.2	ELECTROMETRIC	1,2,3,4
107	7.00	-5.5	ELECTROMETRIC	1,2,3,4
109	7.30	-2.5	ELECTROMETRIC	1,2,3,4
114	7.87	5.1	ELECTROMETRIC	1,2,3,4
115	7.60	1.5	ELECTROMETRIC	1,2,3,4
116	7.44	-0.6	ELECTROMETRIC	1,2,3,4

82 Labs had a total range of 6.10 to 8.58 and a mean of 7.488 with a standard deviation of 0.374 and a 95% confidence interval of the mean +/- 0.083.

Table 7 Standard Reference Water Sample M4 Report for 5102

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	6.9	-3.9	NOT REPORTED	
2	6.7	-6.6	COLORIMETRIC, MOLYBDSILICIC ACID	1,2,3
4	7.4	3.1	COLORIMETRIC, MOLYBDSILICIC ACID	1,2,3
6	7.9	10.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
8	3.3	15.6	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
10	7.0	-2.5	EMISSION, IC PLASMA	5
12	7.6	5.9	EMISSION, IC PLASMA	5
15	4.2	-41.5	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
19	7.7	7.3	EMISSION, IC PLASMA	5
22	3.0	11.5	EMISSION, IC PLASMA	5
23	5.6	-22.0	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
24	6.9	-3.9	EMISSION, IC PLASMA	5
27	3.6	-49.8	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
28	7.5	4.5	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
29	7.6	5.9	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
30	7.1	-1.1	EMISSION, IC PLASMA	5
33	7.5	4.5	EMISSION, IC PLASMA	5
37	7.6	5.9	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
39	3.2	14.3	EMISSION, IC PLASMA	5
40	< 0.0		IGNORED	
41	7.4	3.1	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3
43	7.3	1.7	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
44	3.0	11.5	COLORIMETRIC, MOLYBDSILICIC ACID	1,2,3
45	6.5	-9.4	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
46	6.0	-15.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
48	7.1	-1.1	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
49	4.0	-44.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
52	6.7	-6.6	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
55	7.2	0.3	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
58	6.7	-6.6	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
59	7.5	4.5	EMISSION, IC PLASMA	5
60	7.7	7.3	COLORIMETRIC, MOLYBDSILICIC ACID	1,2,3
61	6.8	-5.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
62	5.4	-24.8	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
67	7.5	4.5	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
69	8.5	13.4	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3
71	9.0	25.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
72	7.0	-2.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
74	6.3	-12.2	COLORIMETRIC, MOLYBDSILICIC ACID	1,2,3
77	7.4	3.1	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3
84	6.7	-6.6	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
87	7.5	4.5	COLORIMETRIC, MOLYBDSILICIC ACID	1,2,3
90	5.6	-22.0	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
94	8.7	21.2	COLORIMETRIC, MOLYBDSILICIC ACID	1,2,3
95	6.0	-16.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4

Table 7 Standard Reference Water Sample M4 Report for SI02

Code Number	Reported value	Pct. dev. from mean	Methods	References
96	8.1	12.9	EMISSION, IC PLASMA	5
99	8.2	14.3	EMISSION, IC PLASMA	5
100	6.3	-12.2	COLORIMETRIC, MOLYBDSILICIC ACID	1,2,3
102	8.2	14.3	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
107	7.4	3.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
109	7.6	5.9	COLORIMETRIC, MOLYBDSILICIC ACID	1,2,3
111	8.9	24.0	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
115	10.7	49.1	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3

53 Labs had a total range of 3.6 to 10.7 and a mean of 7.18 with a standard deviation of 1.24 and a 95% confidence interval of the mean +/- 0.34.

Table 7 Standard Reference Water Sample M4 Report for S04

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	14	-28.3	NOT REPORTED	
3	18	-7.8	THORIN TITRATION	2,4
4	16	-18.1	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
6	20	2.4	ION CHROMATOGRAPHY	2,6
7	20	2.4	GRAVIMETRIC, BARIUM SULFATE	1,2,3
8	23	17.8	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
9	25	28.0	COLORIMETRIC, CHLORANILATE	3
10	13	-7.8	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
12	17	-12.9	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
13	20	2.4	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
14	20	2.4	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
15	16	-18.1	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
16	19	-2.7	GRAVIMETRIC, BARIUM SULFATE	1,2,3
18	19	-2.7	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
20	17	-12.9	GRAVIMETRIC, BARIUM SULFATE	1,2,3
21	19	-2.7	GRAVIMETRIC, BARIUM SULFATE	1,2,3
22	19	-2.7	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
23	23	17.8	COLORIMETRIC, CHLORANILATE	3
24	21	7.5	GRAVIMETRIC, BARIUM SULFATE	1,2,3
25	22	12.7	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
29	21	7.5	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
30	19	-2.7	ION CHROMATOGRAPHY	2,6
32	25	28.0	GRAVIMETRIC, BARIUM SULFATE	1,2,3
34	31	58.7	COLORIMETRIC, CHLORANILATE	3
35	20	2.4	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
39	21	7.5	GRAVIMETRIC, BARIUM SULFATE	1,2,3
41	20	2.4	ION CHROMATOGRAPHY	2,6
43	18	-7.8	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
44	18	-7.8	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
45	20	2.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
46	21	7.5	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
48	21	7.5	ION CHROMATOGRAPHY	2,6
49	27	38.3	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
50	17	-12.9	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
51	5	-74.4	REJECT COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
52	20	2.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
53	20	2.4	NOT REPORTED	
55	17	-12.9	THORIN TITRATION	2,4
56	20	2.4	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
57	20	2.4	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
58	20	2.4	COLORIMETRIC, CHLORANILATE	3
59	19	-2.7	ION CHROMATOGRAPHY	2,6
61	8	-59.0	GRAVIMETRIC, BARIUM SULFATE	1,2,3
62	18	-7.8	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
64	18	-7.8	ION CHROMATOGRAPHY	2,6

Table 7 Standard Reference Water Sample M4 Report for S04

Code Number	Reported value	Pct. dev. from mean	Methods	References
55	17	-12.9	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
67	20	2.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
69	29	48.5	GRAVIMETRIC, BARIUM SULFATE	1,2,3
71	23	17.8	GRAVIMETRIC, BARIUM SULFATE	1,2,3
72	22	12.7	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
73	15	-23.2	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
74	26	33.1	GRAVIMETRIC, BARIUM SULFATE	1,2,3
75	15	-23.2	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
77	13	-7.8	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
79	19	-2.7	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
84	26	33.1	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
86	19	-2.7	OTHER	
89	17	-12.9	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
90	21	7.5	GRAVIMETRIC, BARIUM SULFATE	1,2,3
91	21	7.5	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
92	9	-53.9	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
94	17	-12.9	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
95	25	28.0	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
96	16	-18.1	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
99	20	2.4	ION CHROMATOGRAPHY	2,6
100	20	2.4	GRAVIMETRIC, BARIUM SULFATE	1,2,3
102	16	-18.1	ION CHROMATOGRAPHY	2,6
105	24	22.9	GRAVIMETRIC, BARIUM SULFATE	1,2,3
107	7	-64.2	GRAVIMETRIC, BARIUM SULFATE	1,2,3
109	21	7.5	GRAVIMETRIC, BARIUM SULFATE	1,2,3
112	19	-2.7	ION CHROMATOGRAPHY	2,6
115	19	-2.7	ION CHROMATOGRAPHY	2,6
117	20	2.4	TURBIDIMETRIC, BARIUM SULFATE	1,2,3

73 Labs had a total range of 5 to 31 and a mean of 19.5 with a standard deviation of 4.0 and a 95% confidence interval of the mean +/- 0.9.

Table 7 Standard Reference Water Sample M4 Report for SP. COND.

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	116	4.3	NOT REPORTED	
2	93	-16.4	OTHER	
3	111	-0.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
4	100	-10.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
5	121	8.8	ELECTRODELESS, INDUCTIVE CELL-TYPE	2
6	104	-6.5	DIRECT READING INSTRUMENT	4
7	117	5.2	DIRECT READING INSTRUMENT	4
8	112	0.7	DIRECT READING INSTRUMENT	4
12	111	-0.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
13	118	6.1	DIRECT READING INSTRUMENT	4
14	113	1.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
15	113	1.6	DIRECT READING INSTRUMENT	4
16	115	3.4	OTHER	
18	108	-2.9	DIRECT READING INSTRUMENT	4
20	110	-1.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
21	110	-1.1	DIRECT READING INSTRUMENT	4
22	107	-3.8	DIRECT READING INSTRUMENT	4
23	183	64.6	REJECT DIRECT READING INSTRUMENT	4
25	100	-10.1	DIRECT READING INSTRUMENT	4
26	110	-1.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
27	110	-1.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
28	115	3.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
29	115	3.4	DIRECT READING INSTRUMENT	4
30	115	3.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
32	111	-0.2	DIRECT READING INSTRUMENT	4
34	95	-14.6	DIRECT READING INSTRUMENT	4
35	109	-2.0	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
37	113	1.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
39	109	-2.0	DIRECT READING INSTRUMENT	4
41	111	-0.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
43	108	-2.9	DIRECT READING INSTRUMENT	4
44	114	2.5	DIRECT READING INSTRUMENT	4
45	115	3.4	DIRECT READING INSTRUMENT	4
46	124	11.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
48	110	-1.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
49	111	-0.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
50	107	-3.8	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
51	131	17.8	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
52	112	0.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
53	115	3.4	DIRECT READING INSTRUMENT	4
55	113	1.6	DIRECT READING INSTRUMENT	4
56	105	-5.6	ELECTRODELESS, INDUCTIVE CELL-TYPE	2
57	19261	2E+04	REJECT DIRECT READING INSTRUMENT	4
58	111	-0.2	OTHER	
59	100	-10.1	DIRECT READING INSTRUMENT	4

Table 7 Standard Reference Water Sample M4 Report for SP. COND.

Code Number	Reported value	Pct. dev. from mean	Methods	References
61	113	1.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
62	108	-2.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
64	106	-4.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
65	110	-1.1	DIRECT READING INSTRUMENT	4
67	113	1.6	DIRECT READING INSTRUMENT	4
69	123	10.6	DIRECT READING INSTRUMENT	4
70	60	-46.0	REJECT	4
71	105	-5.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
72	110	-1.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
74	113	1.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
76	111	-0.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
77	112	0.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
79	117	5.2	DIRECT READING INSTRUMENT	4
84	102	-8.3	DIRECT READING INSTRUMENT	4
86	110	-1.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
87	122	9.7	DIRECT READING INSTRUMENT	4
89	104	-6.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
90	118	5.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
91	100	-10.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
92	108	-2.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
94	111	-0.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
95	115	3.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
96	106	-4.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
99	108	-2.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
100	115	3.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
102	120	7.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
105	113	1.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
107	120	7.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
109	954	758.0	DIRECT READING INSTRUMENT	4
111	113	5.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
114	114	2.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
115	104	-6.5	DIRECT READING INSTRUMENT	4

77 Labs had a total range of 60 to 19251 and a mean of 111.2 with a standard deviation of 6.5 and a 95% confidence interval of the mean +/- 1.5.

Table 7 Standard Reference Water Sample M4 Report for SR

Code Number	Reported value	Pct. dev. from mean		Methods	References
1	85	-1.9		EMISSION, IC PLASMA	3,5
2	70	-19.2		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
4	90	3.9		OTHER	
6	130	50.1	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
8	80	-7.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
15	0	-100.0	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
22	82	-5.4		EMISSION, IC PLASMA	3,5
28	90	3.9		EMISSION, IC PLASMA	3,5
30	86	-0.7		EMISSION, IC PLASMA	3,5
33	84	-3.0		EMISSION, IC PLASMA	3,5
39	90	3.9		EMISSION, IC PLASMA	3,5
41	80	-7.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
44	94	8.5		OTHER	
45	87	0.4		NOT REPORTED	
55	80	-7.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
59	87	0.4		EMISSION, IC PLASMA	3,5
61	100	15.4		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
72	89	2.7		EMISSION, IC PLASMA	3,5
90	80	-7.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
95	120	38.5	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
96	95	9.7		EMISSION, IC PLASMA	3,5
99	95	9.7		EMISSION, IC PLASMA	3,5
112	91	5.0		EMISSION, IC PLASMA	3,5
115	80	-7.7		EMISSION, IC PLASMA	3,5
118	91	5.0		MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

25 Labs had a total range of 0 to 130 and a mean of 86.6 with a standard deviation of 6.8 and a 95% confidence interval of the mean +/- 3.0.

Table 7 Standard Reference Water Sample M4 Report for V

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 6		IGNORED OTHER	3
4	5	-6.2	ATOMIC ABSORPTION, FLAMELESS	1,3
6	2	-62.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
8	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
12	< 10		IGNORED EMISSION, IC PLASMA	3,5
30	5	-6.2	EMISSION, IC PLASMA	3,5
39	< 10		IGNORED EMISSION, IC PLASMA	3,5
41	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
45	< 5		IGNORED EMISSION, IC PLASMA	3,5
50	< 50		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
61	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
62	1	-81.2	ATOMIC ABSORPTION, FLAMELESS	3
71	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
72	141	2543.8	REJECT EMISSION, IC PLASMA	3,5
95	7	31.3	ATOMIC ABSORPTION, FLAMELESS	3
96	< 5		IGNORED EMISSION, IC PLASMA	3,5
99	< 3		IGNORED EMISSION, IC PLASMA	3,5
107	< 250		IGNORED EMISSION, IC PLASMA	3,5
112	12	125.0	EMISSION, IC PLASMA	3,5
115	< 30		IGNORED EMISSION, IC PLASMA	3,5

20 Labs had a total range of 1 to 141 and a mean of 5.3 with a standard deviation of 3.9 and a 95% confidence interval of the mean +/- 4.1.

Table 8 . Statistics by method for standard reference sample M4

Determination	Method	Range:		Mean	Standard Deviation	N
		from	to			
ALK(CAC03)	OTHER	26.000	- 28.000	27.000	1.000	3
	TITRATION, COLORIMETRIC, AUTOMATED	4.000	- 34.000	22.038	9.918	8
	TITRATION, COLORIMETRIC, MANUAL	21.500	- 58.000	32.012	10.288	17
	TITRATION, ELECTROMETRIC, AUTOMATED	25.400	- 64.000	27.075	0.805	8
	TITRATION, ELECTROMETRIC, MANUAL	14.000	- 40.000	26.955	1.110	29
	OVER-ALL	4.000	- 64.000	27.161	4.653	69
B	COLORIMETRIC, AZOMETHINE	10.000	- 15.000	12.500	1.581	
	COLORIMETRIC, CARMINE (CARMINIC ACID)	0.000	- 20.000	8.500	10.116	4
	COLORIMETRIC, CURCUMIN	0.000	- 130.000	18.833	16.130	6
	EMISSION, IC PLASMA	2.000	- 40.000	8.500	3.674	6
	OVER-ALL	0.000	- 130.000	15.619	14.235	21
BR	OTHER	10.000	- 10.000	10.000	3.674	1
	OVER-ALL	10.000	- 10.000	10.000	14.235	1
CA	ATOMIC ABSORPTION, DIRECT, AIR	5.600	- 20.000	10.964	1.835	44
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	11.300	- 15.500	12.900	2.272	3
	EMISSION, IC PLASMA	10.700	- 14.000	11.541	0.560	17
	TITRATION, EDTA	10.000	- 15.500	11.578	0.897	9
	OVER-ALL	5.600	- 20.000	11.425	0.905	69
CL	COLORIMETRIC, FERRIC THIOCYANATE	1.000	- 9.000	2.752	0.403	21
	ION CHROMATOGRAPHY	1.400	- 2.800	2.733	0.052	6
	ION SELECTIVE ELECTRODE	2.300	- 4.900	3.275	1.132	4
	NOT REPORTED	2.700	- 3.700	3.167	0.503	3
	TITRATION, MERCURIC NITRATE	0.400	- 17.900	3.227	0.742	15
	TITRATION, SILVER NITRATE	1.500	- 8.000	3.567	1.888	12
	OVER-ALL	0.400	- 17.900	2.889	0.700	64
DSRD 180	NOT REPORTED	55.000	- 74.000	64.500	3.082	
	RESIDUE ON EVAPORATION	54.000	- 96.000	71.333	12.998	15
	RESIDUE ON EVAPORATION	35.000	- 80.000	69.111	5.732	9
	RESIDUE, FILTRABLE	0.000	- 124.000	66.071	23.311	28
	RESIDUE, FILTRABLE	56.000	- 79.000	68.333	7.607	6
	OVER-ALL	0.000	- 124.000	68.638	11.945	58
F	COLORIMETRIC, CEROUS ALIZARIN "COMPLEXONE"	0.420	- 0.500	0.453	0.042	3
	COLORIMETRIC, SPADNS	0.500	- 0.660	0.655	0.071	
	ION CHROMATOGRAPHY	0.420	- 0.480	0.458	0.026	6
	ION SELECTIVE ELECTRODE	0.400	- 0.480	0.441	0.029	8
	OTHER	0.290	- 1.130	0.466	0.032	27
	OVER-ALL	0.290	- 1.130	0.469	0.063	55
	K	ATOMIC ABSORPTION, DIRECT, AIR	0.500	- 2.400	1.152	0.177
EMISSION, FLAME		0.100	- 1.500	1.200	0.167	6
OTHER		0.700	- 2.100	1.073	0.174	11
OVER-ALL		0.100	- 2.400	1.141	0.179	66

Table B . Statistics by method for standard reference sample M4

Determination	Method	Range:		Mean	Standard Deviation	N
		from	to			
MG	ATOMIC ABSORPTION, DIRECT, AIR	0.300	4.000	2.930	0.410	46
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	2.900	3.300	2.900	0.000	2
	EMISSION, IC PLASMA	0.600	3.300	3.000	0.137	17
	TITRATION, EDTA	0.200	3.900	2.343	1.275	7
	OVER-ALL	0.200	4.000	2.949	0.371	73
NA	ATOMIC ABSORPTION, DIRECT, AIR	3.200	8.200	4.683	0.509	48
	EMISSION, FLAME	2.300	5.100	4.700	0.297	6
	EMISSION, IC PLASMA	3.400	5.000	4.550	0.293	14
	OVER-ALL	2.300	8.200	4.637	0.471	71
NO2-N	COLORIMETRIC, DIAZOTIZATION	0.000	0.050	0.005	0.009	27
	OVER-ALL	0.000	0.050	0.005	0.009	27
NO3-N	COLORIMETRIC, BRUCINE	0.160	0.300	0.223	0.046	11
	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	0.010	0.300	0.184	0.019	39
	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	0.150	0.320	0.220	0.069	5
	ION CHROMATOGRAPHY	0.140	0.860	0.177	0.023	6
	OTHER	0.100	0.400	0.213	0.163	3
	OVER-ALL	0.010	0.860	0.188	0.044	68
P, TOTAL	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	0.000	0.200	0.022	0.020	37
	COLORIMETRIC, BLK DIG, H2SO4, K3HG2SO4, PHOSPHOMOLYBDATE	0.010	0.300	0.019	0.009	7
	EMISSION, IC PLASMA	0.010	0.090	0.050	0.200	
	OVER-ALL	0.000	0.390	0.019	0.014	43
PH	ELECTROMETRIC	6.100	8.580	7.488	0.374	81
	OVER-ALL	6.100	8.580	7.488	0.374	81
SI02	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4.000	9.000	6.910	1.403	10
	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	3.600	8.900	6.727	1.660	11
	COLORIMETRIC, MOLYBDO-SILICIC ACID	6.300	8.700	7.356	0.797	9
	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	5.600	7.500	6.533	0.781	6
	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	7.400	10.700	8.500	1.556	4
	EMISSION, IC PLASMA	6.900	8.200	7.618	0.475	11
	OVER-ALL	3.600	10.700	7.177	1.235	52
SO4	COLORIMETRIC, CHLORANILATE	20.000	31.000	24.750	4.646	4
	COLORIMETRIC, METHYL THYMOL BLUE	5.000	26.000	19.588	2.395	17
	GRAVIMETRIC, BARIUM SULFATE	7.000	29.000	20.063	5.744	16
	ION CHROMATOGRAPHY	16.000	21.000	19.100	1.370	10
	TURBIDIMETRIC, BARIUM SULFATE	9.000	27.000	18.700	3.600	20
	OVER-ALL	5.000	31.000	19.528	3.982	72
SP. COND.	DIRECT READING INSTRUMENT	95.000	-19261.000	111.179	6.945	28
	OTHER	93.000	115.000	106.333	11.719	3
	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	60.000	954.000	110.842	4.716	38
	OVER-ALL	60.000	-19261.000	111.192	6.482	73
SR	ATOMIC ABSORPTION, DIRECT, AIR	0.000	130.000	92.500	21.876	8

Table 3. Statistics by method for standard reference sample M4

Determination	Method	Range: from	to	Mean	Standard Deviation	N
SR	EMISSION, IC PLASMA	80.000	95.000	87.833	4.726	12
	OVER-ALL	0.000	130.000	86.636	6.828	22
V	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	2.000	2.000	2.000	4.726	1
	ATOMIC ABSORPTION, FLAMELESS	1.000	7.000	4.333	3.055	3
	EMISSION, IC PLASMA	5.000	141.000	9.500	1.871	1
	OVER-ALL	1.000	141.000	5.333	3.933	6

Table 9 Standard Reference Water Sample T93 Report for ACID2CAC03

Code Number	Reported value	Pct. dev. from mean		Methods	References
1	1000	-2.5		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
4	1100	7.3		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
12	1010	-1.5		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
14	1000	-2.5		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
16	1040	1.4		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
21	1140	11.2	REJECT	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
24	1000	-2.5		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
32	1040	1.4		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
34	10	-99.0	REJECT	TITRATION, COLORIMETRIC, MANUAL	1,2,3
45	1060	3.4		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
61	1020	-0.5		TITRATION, COLORIMETRIC, MANUAL	1,2,3
62	1000	-2.5		TITRATION, COLORIMETRIC, MANUAL	1,2,3
71	1030	0.4		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
77	1000	-2.5		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
92	1440	40.4	REJECT	NOT REPORTED	
94	1050	2.4		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
95	1040	1.4		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
99	1030	0.4		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
107	990	-3.5		TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4

19 Labs had a total range of 10 to 1440 and a mean of 1026 with a standard deviation of 29 and a 95% confidence interval of the mean +/- 16.

Table 9 Standard Reference Water Sample T93 Report for AG

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	5.0	55.0	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,2,4
2	2.7	-16.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
4	1.0	-69.0	ATOMIC ABSORPTION, FLAMELESS	3
5	6.0	86.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
6	3.1	-3.9	ATOMIC ABSORPTION, FLAMELESS	3
7	3.5	8.5	ATOMIC ABSORPTION, FLAMELESS	3
8	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
10	4.3	33.3	ATOMIC ABSORPTION, FLAMELESS	3
12	2.8	-13.2	ATOMIC ABSORPTION, FLAMELESS	3
14	0.3	-90.7	ATOMIC ABSORPTION, FLAMELESS	3
18	< 2.5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
21	1.0	-69.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
22	2.5	-22.5	ATOMIC ABSORPTION, FLAMELESS	3
23	2.6	-19.4	ATOMIC ABSORPTION, FLAMELESS	3
26	5.0	55.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
32	9.0	179.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
33	250.0	7650.3	REJECT NOT REPORTED	
34	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
39	1.0	-69.0	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,2,4
40	90.0	2690.1	REJECT OTHER	
41	< 5.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
44	2.9	-10.1	ATOMIC ABSORPTION, FLAMELESS	3
45	3.3	2.3	ATOMIC ABSORPTION, FLAMELESS	3
46	6.0	86.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
50	4.0	24.0	ATOMIC ABSORPTION, FLAMELESS	3
57	3.0	-7.0	NOT REPORTED	
58	1.7	-47.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
61	1.0	-69.0	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,2,4
62	< 50.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
64	0.6	-81.4	ATOMIC ABSORPTION, FLAMELESS	3
69	3.0	-7.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
70	2.0	-38.0	ATOMIC ABSORPTION, FLAMELESS	3
71	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
72	2.9	-10.1	ATOMIC ABSORPTION, FLAMELESS	3
73	< 5.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
77	3.0	-7.0	ATOMIC ABSORPTION, FLAMELESS	3
78	< 10.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
84	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
86	2.7	-16.3	NOT REPORTED	
91	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
94	2.8	-13.2	ATOMIC ABSORPTION, FLAMELESS	3
95	2.6	-19.4	ATOMIC ABSORPTION, FLAMELESS	3
96	3.2	-0.8	ATOMIC ABSORPTION, FLAMELESS	3
99	2.6	-19.4	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,2,4
102	4.0	24.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
105	2.8	-13.2	ATOMIC ABSORPTION, FLAMELESS	3
107	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
112	< 6.0		IGNORED EMISSION, IC PLASMA	3
116	9.0	179.0	EMISSION, IC PLASMA	3
117	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3

50 Labs had a total range of 0.3 to 250.0 and a mean of 3.23 with a standard deviation of 1.98 and a 95% confidence interval of the mean +/- 0.63.

Table 9 Standard Reference Water Sample T93 Report for AL

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	150	-10.4	EMISSION, DC PLASMA	
4	50	-70.1	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
5	179	7.0	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
6	120	-28.3	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
8	250	49.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3, 4
10	125	-25.3	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
12	140	-16.3	EMISSION, IC PLASMA	3, 5
14	110	-34.3	SPECTROPHOTOMETRIC, ERIOCHROME CYANINE	1
22	141	-15.7	EMISSION, IC PLASMA	3, 5
24	150	7.6	EMISSION, IC PLASMA	3, 5
26	212	26.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3, 4
28	230	37.4	EMISSION, IC PLASMA	3, 5
29	95	-43.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3, 4
30	200	19.5	EMISSION, IC PLASMA	3, 5
32	148	-11.6	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3, 4
33	149	-11.0	NOT REPORTED	1, 2, 3, 4
39	160	-4.4	EMISSION, IC PLASMA	3, 5
41	160	-4.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3, 4
44	180	7.6	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
45	170	1.6	ATOMIC ABSORPTION, FLAMELESS (GRAPHITE FURNACE)	
50	170	1.6	ATOMIC ABSORPTION, DIRECT, FLAMELESS	
51	<		IGNORED	
54	160	-4.4	EMISSION, IC PLASMA	3, 5
57	186	11.1	EMISSION, IC PLASMA	3, 5
59	180	7.6	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
61	350	109.1	EMISSION, IC PLASMA	3, 5
62	110	-34.3	ATOMIC ABSORPTION, CHELATION EXTRACTION, NITROUS OXIDE	2, 4
64	<		REJECT	
65	<		IGNORED	
69	776	363.7	ATOMIC ABSORPTION, DIRECT, FLAMELESS (GRAPHITE FURNACE)	3
71	<		IGNORED	
72	273	63.1	EMISSION, IC PLASMA	3, 5
73	190	13.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3, 4
77	195	16.5	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
78	240	43.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3, 4
84	10	-94.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3, 4
86	147	-12.2	EMISSION, IC PLASMA	3, 5
94	149	-11.0	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
95	200	19.5	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
96	163	-2.6	EMISSION, IC PLASMA	3, 5
99	190	13.5	ATOMIC ABSORPTION, CHELATION EXTRACTION, NITROUS OXIDE	2, 4
101	138	-17.5	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
102	160	-4.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3, 4
107	<		IGNORED	
116	192	14.7	EMISSION, IC PLASMA	3, 5

45 Labs had a total range of 10 to 776 and a mean of 167.4 with a standard deviation of 43.7 and a 95% confidence interval of the mean +/- 14.5.

Table 9 Standard Reference Water Sample T93 Report for AS

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	6.0	10.3	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
4	6.0	10.3	ATOMIC ABSORPTION, FLAMELESS	3
5	5.4	-0.7	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
6	5.2	-4.4	ATOMIC ABSORPTION, FLAMELESS	3
7	4.2	-22.8	ATOMIC ABSORPTION, FLAMELESS	3
8	5.0	-8.1	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
10	6.0	10.3	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
12	6.3	15.8	ATOMIC ABSORPTION, FLAMELESS	3
14	6.0	10.3	ATOMIC ABSORPTION, FLAMELESS	3
18	< 5.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
22	2.1	-61.4	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
26	6.0	10.3	ATOMIC ABSORPTION, FLAMELESS	3
30	5.9	6.6	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
33	15.0	175.7	REJECT NOT REPORTED	
34	5.0	-8.1	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
39	4.0	-26.5	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
40	8.8	61.7	ATOMIC ABSORPTION, FLAMELESS	3
41	5.0	-8.1	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
45	5.2	-4.4	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
46	1.8	-66.9	ATOMIC ABSORPTION, HYDRIDE, (ZINC)	1,2,3,4
50	3.1	-43.0	ATOMIC ABSORPTION, FLAMELESS	3
51	5.0	-8.1	ATOMIC ABSORPTION, FLAMELESS	3
57	4.6	-15.4	ATOMIC ABSORPTION, FLAMELESS	3
58	3.9	-30.2	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
61	4.5	-17.3	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
62	4.5	-17.3	ATOMIC ABSORPTION, FLAMELESS	3
64	< 3.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
65	6.6	21.3	ATOMIC ABSORPTION, FLAMELESS	3
67	5.6	2.9	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
69	5.8	6.6	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
70	5.0	-8.1	ATOMIC ABSORPTION, FLAMELESS	3
71	10.0	83.8	ATOMIC ABSORPTION, FLAMELESS	3
72	4.4	-19.1	ATOMIC ABSORPTION, FLAMELESS	3
73	10.0	83.8	ATOMIC ABSORPTION, FLAMELESS	3
77	5.0	-8.1	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
78	4.0	-26.5	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
84	6.0	10.3	ATOMIC ABSORPTION, FLAMELESS	3
86	5.1	-6.3	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
94	6.1	12.1	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
95	< 10.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
96	8.0	47.0	ATOMIC ABSORPTION, FLAMELESS	3
97	6.0	10.3	NOT REPORTED	
99	5.7	4.8	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
102	6.0	10.3	ATOMIC ABSORPTION, FLAMELESS	3
105	4.9	-9.9	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
107	5.0	-8.1	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1
116	57.0	947.7	REJECT EMISSION, IC PLASMA	3

47 Labs had a total range of 1.8 to 57.0 and a mean of 5.44 with a standard deviation of 1.63 and a 95% confidence interval of the mean +/- 0.51.

Table 9 Standard Reference Water Sample T93 Report for BA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 100		ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
2	70	-0.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
4	70	-0.8	EMISSION, IC PLASMA	3,5
6	8	-88.7	ATOMIC ABSORPTION, FLAMELESS	3
7	65	-7.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
8	< 100		ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
10	68	-3.6	EMISSION, IC PLASMA	3,5
12	70	-0.8	EMISSION, IC PLASMA	3,5
14	40	-43.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
18	50		ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
19	66	-6.5	EMISSION, IC PLASMA	3,5
21	60	-15.0	EMISSION, DC PLASMA	3,5
22	64	-9.3	EMISSION, IC PLASMA	3,5
23	80	13.4	EMISSION, IC PLASMA	3,5
24	70	-0.8	EMISSION, DC PLASMA	3,5
26	94	33.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
28	68	-3.6	EMISSION, IC PLASMA	3,5
30	72	2.0	EMISSION, IC PLASMA	3,5
32	< 50		ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
33	67	-5.1	NOT REPORTED	
39	60	-15.0	EMISSION, IC PLASMA	3,5
41	60	-15.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
44	75	6.3	ATOMIC ABSORPTION, FLAMELESS	3
45	70	-0.8	EMISSION, IC PLASMA	3,5
46	57	-19.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
50	80	13.4	EMISSION, DC PLASMA	2
54	73	3.4	EMISSION, IC PLASMA	3,5
57	102	44.5	ATOMIC ABSORPTION, FLAMELESS	3
58	83	17.6	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
59	69	-2.2	EMISSION, IC PLASMA	3,5
61	200	183.4	EMISSION, IC PLASMA	3
62	75	6.3	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
65	< 200		ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	3,5
69	224	217.4	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
70	76	7.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
71	< 500		ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
72	70	-0.8	EMISSION, IC PLASMA	3,5
73	52	-26.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
78	170	140.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
84	60	-15.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
86	85	20.4	EMISSION, IC PLASMA	3,5
94	60	-15.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
95	100	41.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
96	67	-5.1	EMISSION, IC PLASMA	3,5
97	63	-10.7	NOT REPORTED	

Table 9 Standard Reference Water Sample T93 Report for BA

Code Number	Reported value	Pct. dev. from mean	Methods	References
99	69	-2.2	EMISSION, IC PLASMA	3,5
102	60	-15.0	ATOMIC ABSORPTION, FLAMELESS	3
105	100	41.7	ATOMIC ABSORPTION, FLAMELESS	3
107	60	-15.0	EMISSION, IC PLASMA	3,5
112	75	6.3	EMISSION, IC PLASMA	3,5
116	74	4.9	EMISSION, IC PLASMA	3,5
118	65	-7.9	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

52 Labs had a total range of 8 to 224 and a mean of 70.6 with a standard deviation of 12.5 and a 95% confidence interval of the mean +/- 3.9.

Table 9 Standard Reference Water Sample T93 Report for BE

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	10.0	49.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
4	11.0	64.5	ATOMIC ABSORPTION, FLAMELESS	3
6	15.0	124.3	REJECT ATOMIC ABSORPTION, FLAMELESS	3
8	10.0	49.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
12	6.0	-10.3	EMISSION, IC PLASMA	3,5
19	7.8	15.6	EMISSION, IC PLASMA	3,5
21	7.0	4.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
26	7.0	4.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
28	6.1	-3.3	EMISSION, IC PLASMA	3,5
30	7.0	4.7	EMISSION, IC PLASMA	3,5
33	6.4	-4.3	NOT REPORTED	
39	6.0	-10.3	EMISSION, IC PLASMA	3,5
40	7.0	4.7	EMISSION, IC PLASMA	3,5
41	7.0	4.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
45	6.0	-10.3	EMISSION, IC PLASMA	3,5
50	64.0	856.9	REJECT COLORIMETRIC, ALUMINON	1
51	< 10.0		IGNORED EMISSION, IC PLASMA	3,5
57	6.5	-2.8	ATOMIC ABSORPTION, FLAMELESS	3
58	4.0	-40.2	EMISSION, IC PLASMA	3,5
59	6.4	-4.3	EMISSION, IC PLASMA	3,5
61	3.0	-55.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
62	6.5	-2.8	ATOMIC ABSORPTION, FLAMELESS	3
71	< 50.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
72	6.5	-2.8	EMISSION, IC PLASMA	3,5
84	20.0	199.0	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
94	6.7	0.2	ATOMIC ABSORPTION, FLAMELESS	3
95	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
96	6.2	-7.3	EMISSION, IC PLASMA	3,5
99	7.4	10.6	EMISSION, IC PLASMA	3,5
102	6.0	-10.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
107	< 10.0		IGNORED EMISSION, IC PLASMA	3,5
112	6.4	-4.3	EMISSION, IC PLASMA	3,5
116	4.0	-40.2	EMISSION, IC PLASMA	3,5

33 Labs had a total range of 3.0 to 64.0 and a mean of 6.69 with a standard deviation of 1.72 and a 95% confidence interval of the mean +/- 0.69.

Table 9 Standard Reference Water Sample T93 Report for Cd

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	5.0	4.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	4.0	-16.4	ATOMIC ABSORPTION, FLAMELESS	3
5	5.0	4.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	5.6	17.0	ATOMIC ABSORPTION, FLAMELESS	3
7	4.2	-12.3	ATOMIC ABSORPTION, FLAMELESS	3
8	5.0	4.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	5.3	10.7	ANODIC STRIPPING VOLTAMMETRY, DIFFERENTIAL PULSE	2
10	4.2	-12.3	ATOMIC ABSORPTION, FLAMELESS	3
12	4.7	-1.8	ATOMIC ABSORPTION, FLAMELESS	3
14	5.0	4.5	ATOMIC ABSORPTION, FLAMELESS	3
15	4.4	-8.1	ATOMIC ABSORPTION, FLAMELESS	3
16	< 5.0		IGNORED EMISSION, DC PLASMA	7
18	4.5	-6.0	ATOMIC ABSORPTION, FLAMELESS	3
19	4.0	-16.4	EMISSION, IC PLASMA	3,5
21	6.0	25.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
23	5.0	4.5	ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1,4
24	< 4.0		IGNORED EMISSION, IC PLASMA	3,5
25	8.0	67.1	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
26	5.0	4.5	ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1,4
28	4.5	-5.0	EMISSION, IC PLASMA	3,5
30	5.0	4.5	EMISSION, IC PLASMA	3,5
32	4.0	-15.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	4.5	-6.0	NOT REPORTED	
34	5.0	4.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
39	5.0	4.5	ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1,4
40	8.0	67.1	REJECT EMISSION, IC PLASMA	3,5
41	< 5.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	4.4	-8.1	ATOMIC ABSORPTION, FLAMELESS	3
45	4.2	-12.3	EMISSION, IC PLASMA	3,5
46	5.0	4.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	4.0	-16.4	ATOMIC ABSORPTION, FLAMELESS	3
51	4.3	-10.2	ATOMIC ABSORPTION, FLAMELESS	3
56	9.0	83.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	4.9	2.4	ATOMIC ABSORPTION, FLAMELESS	3
58	5.5	14.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
60	4.1	-14.3	ATOMIC ABSORPTION, FLAMELESS	3
61	5.2	8.6	ATOMIC ABSORPTION, EXTRACTION, (PDCA/CHCL3)	2,3
62	4.0	-16.4	ATOMIC ABSORPTION, FLAMELESS	3
64	1.0	-79.1	REJECT ATOMIC ABSORPTION, FLAMELESS	3
65	4.9	2.4	ATOMIC ABSORPTION, FLAMELESS	3
69	6.0	25.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	5.0	4.5	ATOMIC ABSORPTION, FLAMELESS	3
71	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	4.8	0.3	EMISSION, IC PLASMA	3,5
73	5.0	4.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

Table 9 Standard Reference Water Sample T93 Report for CD

Code Number	Reported value	Pct. dev. from mean	Methods	References
77	5.0	4.5	ATOMIC ABSORPTION, FLAMELESS	3
78	6.0	25.4	ATOMIC ABSORPTION, FLAMELESS	3
79	3.3	-20.6	ATOMIC ABSORPTION, FLAMELESS	3
84	5.0	4.5	ATOMIC ABSORPTION, FLAMELESS	3
86	4.8	0.3	ATOMIC ABSORPTION, FLAMELESS	3
91	< 15.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
92	10.0	103.9	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	4.5	-6.0	ATOMIC ABSORPTION, FLAMELESS	3
95	8.4	75.5	REJECT ATOMIC ABSORPTION, FLAMELESS	3
96	4.4	-9.1	ATOMIC ABSORPTION, FLAMELESS	3
97	6.0	25.4	NOT REPORTED	
99	< 1.0		IGNORED ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1,4
102	6.0	25.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	5.5	14.9	ATOMIC ABSORPTION, FLAMELESS	3
105	3.3	-20.6	ATOMIC ABSORPTION, FLAMELESS	3
107	< 10.0		IGNORED EMISSION, IC PLASMA	3,5
112	4.0	-16.4	EMISSION, IC PLASMA	3,5
116	5.0	4.5	EMISSION, IC PLASMA	3,5
117	4.0	-16.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	4.9	2.4	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

65 Labs had a total range of 1.0 to 10.0 and a mean of 4.79 with a standard deviation of 0.61 and a 95% confidence interval of the mean +/- 0.17.

Table 9 Standard Reference Water Sample T93 Report for CO

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	4.0	-22.3	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
4	4.0	-22.3	ATOMIC ABSORPTION, FLAMELESS	3
6	6.4	24.3	ATOMIC ABSORPTION, FLAMELESS	3
8	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	4.3	-15.5	EMISSION, IC PLASMA	3,5
21	7.0	35.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
23	5.0	-2.9	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
26	42.0	715.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
28	4.7	-3.7	EMISSION, IC PLASMA	3,5
33	5.5	6.8	NOT REPORTED	
39	4.0	-22.3	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
41	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	< 5.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
53	< 20.0		IGNORED EMISSION, IC PLASMA	3,5
61	5.0	-2.9	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
71	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	5.6	8.7	EMISSION, IC PLASMA	3,5
77	13.0	152.4	REJECT ATOMIC ABSORPTION, FLAMELESS	3
84	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	7.0	35.9	EMISSION, IC PLASMA	3,5
99	4.0	-22.3	EMISSION, IC PLASMA	3,5
102	6.0	16.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
107	< 10.0		IGNORED EMISSION, IC PLASMA	3,5
112	3.9	-24.3	EMISSION, IC PLASMA	3,5
116	6.0	16.5	EMISSION, IC PLASMA	3,5

26 Labs had a total range of 3.9 to 42.0 and a mean of 5.15 with a standard deviation of 1.10 and a 95% confidence interval of the mean +/- 0.53.

Table 9 Standard Reference Water Sample T93 Report for CR TDT

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	22.0	105.0	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,3,4
4	10.0	-5.8	ATOMIC ABSORPTION, FLAMELESS	3
5	3.6	-12.9	ATOMIC ABSORPTION, FLAMELESS	3
6	13.0	21.1	ATOMIC ABSORPTION, FLAMELESS	3
7	10.6	-1.2	ATOMIC ABSORPTION, FLAMELESS	3
8	< 10.0		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	4.1	-61.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	9.1	-35.2	ATOMIC ABSORPTION, FLAMELESS	3
12	10.2	-5.0	ATOMIC ABSORPTION, FLAMELESS	3
14	12.0	11.8	ATOMIC ABSORPTION, FLAMELESS	3
15	10.2	-5.0	ATOMIC ABSORPTION, FLAMELESS	3
18	8.6	-17.9	ATOMIC ABSORPTION, FLAMELESS	3
21	12.0	11.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	20.0	56.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
26	11.0	2.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
28	11.0	2.5	EMISSION, IC PLASMA	3
30	15.0	39.7	EMISSION, IC PLASMA	3
32	11.0	2.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	12.0	11.8	NOT REPORTED	3
34	14.0	30.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
39	< 10.0		ATOMIC ABSORPTION, FLAMELESS	3
40	13.0		EMISSION, IC PLASMA	3
41	< 20.0	21.1	EMISSION, IC PLASMA	3
44	9.8	-8.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
45	9.0	-15.2	ATOMIC ABSORPTION, FLAMELESS	3
46	5.0	-44.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	9.0	-15.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	< 10.0		ATOMIC ABSORPTION, FLAMELESS	3
56	6.0	-44.1	EMISSION, IC PLASMA	3
57	10.8	0.6	NOT REPORTED	3
58	3.8	-64.6	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
60	9.6	-10.6	ATOMIC ABSORPTION, DIRECT, AIR	3
61	20.0	85.3	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
62	12.0	11.8	ATOMIC ABSORPTION, FLAMELESS	3
64	3.0	-72.1	ATOMIC ABSORPTION, FLAMELESS	3
65	< 30.0		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	14.0	30.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	10.0	-6.8	ATOMIC ABSORPTION, FLAMELESS	3
71	< 20.0		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	8.4	-21.7	EMISSION, IC PLASMA	3
73	9.0	-15.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	10.0	-5.8	ATOMIC ABSORPTION, FLAMELESS	3
78	10.0	-5.8	ATOMIC ABSORPTION, FLAMELESS	3
79	13.0	21.1	ATOMIC ABSORPTION, FLAMELESS	3
84	7.0	-34.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

Table 9 Standard Reference Water Sample 193 Report for CR 101

Code Number	Reported value	Pct. dev. from mean	Methods	References
86	13.7	-0.3	ATOMIC ABSORPTION, FLAMELESS	3
91	< 20.0		IGNORED	1,2,3,4
92	<100.0		IGNORED	1
94	13.5	-2.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	3
95	< 10.0		IGNORED	1,2,3,4
96	10.4	-3.1	ATOMIC ABSORPTION, DIRECT, AIR	3
97	25.0	142.2	EMISSION, IC PLASMA	
99	2.4	-77.5	NOT REPORTED	
102	15.0	39.7	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,3,4
104	7.4	-31.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
105	14.3	33.2	ATOMIC ABSORPTION, FLAMELESS	3
107	< 20.0		IGNORED	3
116	16.0	49.1	EMISSION, IC PLASMA	3

58 Labs had a total range of 2.4 to 26.0 and a mean of 10.73 with a standard deviation of 4.01 and a 95% confidence interval of the mean +/- 1.13.

Table 9 Standard Reference Water Sample T93 Report for CU

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	30	0.8	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
2	32	7.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	29	-2.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	29	-2.5	ATOMIC ABSORPTION, FLAMELESS	3
7	12	-59.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	25	-16.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	23	-5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	30	0.8	EMISSION, IC PLASMA	3,5
12	33	10.9	EMISSION, IC PLASMA	3,5
14	37	24.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
18	32	7.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	31	4.2	EMISSION, IC PLASMA	3,5
21	32	7.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	33	10.9	EMISSION, IC PLASMA	3,5
24	26	-12.6	EMISSION, IC PLASMA	3,5
25	30	0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
26	35	17.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
28	34	14.3	EMISSION, IC PLASMA	3,5
29	21	-29.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	30	0.8	EMISSION, IC PLASMA	3,5
32	55	84.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	25	-16.0	NOT REPORTED	
34	10	-66.4	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
39	29	-2.5	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
40	30	0.8	EMISSION, IC PLASMA	3,5
41	32	7.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	31	4.2	ATOMIC ABSORPTION, FLAMELESS	3
45	34	14.3	EMISSION, IC PLASMA	3,5
46	4	-86.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	35	17.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	< 10		IGNORED EMISSION, IC PLASMA	3,5
55	30	0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	30	0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	30	0.8	ATOMIC ABSORPTION, FLAMELESS	3
58	35	17.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	36	21.0	EMISSION, IC PLASMA	3,5
60	34	14.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
61	29	-2.5	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
62	50	68.0	ATOMIC ABSORPTION, FLAMELESS	3
64	10	-66.4	ATOMIC ABSORPTION, FLAMELESS	3
65	31	4.2	ATOMIC ABSORPTION, FLAMELESS	3
67	27	-9.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	42	41.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	27	-9.3	ATOMIC ABSORPTION, FLAMELESS	3
71	20	-32.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

Table 9 Standard Reference Water Sample 193 Report for CU

Code Number	Reported value	Pct. dev. from mean	Methods	References
72	24	-19.3	EMISSION, IC PLASMA	3,5
73	33	10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	10	-66.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	26	-12.6	ATOMIC ABSORPTION, FLAMELESS	3
78	32	7.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
79	37	24.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	30	0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	32	7.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	45	51.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
92	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	32	7.5	ATOMIC ABSORPTION, FLAMELESS	3
95	30	0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	30	0.8	EMISSION, IC PLASMA	3,5
97	32	7.5	NOT REPORTED	
99	34	14.3	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
102	33	10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	23	-22.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
105	32	7.5	ATOMIC ABSORPTION, FLAMELESS	3
107	< 10		IGNORED EMISSION, IC PLASMA	3,5
112	21	-29.4	EMISSION, IC PLASMA	3,5
116	31	4.2	EMISSION, IC PLASMA	3,5
117	31	4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	31	4.2	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

68 Labs had a total range of 4 to 55 and a mean of 29.8 with a standard deviation of 3.3 and a 95% confidence interval of the mean +/- 2.1.

Table 9 Standard Reference Water Sample T93 Report for FE

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	128	25.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	90	-11.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
3	100	-2.0	OTHER	
4	100	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	10	-90.2	REJECT ATOMIC ABSORPTION, FLAMELESS	3
7	120	17.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	100	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	104	1.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	106	3.9	EMISSION, IC PLASMA	3,5
12	95	-6.9	EMISSION, IC PLASMA	3,5
13	210	105.8	REJECT OTHER	
14	120	17.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	0	-100.0	REJECT OTHER	
18	93	-8.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	94	-7.9	EMISSION, IC PLASMA	3,5
21	93	-3.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	100	-2.0	EMISSION, IC PLASMA	3,5
23	105	2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
24	100	-2.0	EMISSION, IC PLASMA	3,5
25	120	17.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
26	178	74.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
27	306	200.0	REJECT OTHER	
28	110	7.8	EMISSION, IC PLASMA	3,5
30	100	-2.0	EMISSION, IC PLASMA	3,5
32	39	-61.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	1035	914.5	REJECT NOT REPORTED	
34	150	47.0	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
35	93	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
39	220	115.7	REJECT EMISSION, IC PLASMA	3,5
40	154	51.0	EMISSION, IC PLASMA	3,5
41	100	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	90	-11.8	ATOMIC ABSORPTION, FLAMELESS	3
45	105	2.9	EMISSION, IC PLASMA	3,5
46	96	-5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	100	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	60	-41.2	EMISSION, IC PLASMA	3,5
55	110	7.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	100	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	90	-11.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
58	90	-11.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	103	1.0	EMISSION, IC PLASMA	3,5
61	105	2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
62	140	37.2	ATOMIC ABSORPTION, FLAMELESS	3
65	71	-30.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
67	87	-14.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

Table 9 Standard Reference Water Sample T93 Report for PE

Code Number	Reported value	Pct. dev. from mean	Methods	References
69	151	48.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	125	22.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	100	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	115	12.7	EMISSION, IC PLASMA	3,5
73	100	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	40	-60.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
76	0	-100.0	REJECT	1,2,3,4
77	95	-6.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
78	105	2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	100	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	130	27.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	70	-31.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	100	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	101	-1.0	EMISSION, IC PLASMA	3,5
97	121	18.6	NOT REPORTED	
99	97	-4.9	EMISSION, IC PLASMA	3,5
100	93	-8.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
101	100	-2.0	ATOMIC ABSORPTION, FLAMELESS	3
102	107	4.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
105	79	-22.6	ATOMIC ABSORPTION, FLAMELESS	3
107	80	-21.6	EMISSION, IC PLASMA	3,5
112	102	-0.0	EMISSION, IC PLASMA	3,5
116	127	24.5	EMISSION, IC PLASMA	3,5
117	119	15.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

69. Labs had a total range of 0 to 1035 and a mean of 102.0 with a standard deviation of 21.4 and a 95% confidence interval of the mean +/- 5.5.

Table 9 Standard Reference Water Sample T93 Report for HG

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 0.1		IGNORED EMISSION, IC PLASMA	3
4	0.2	-40.5	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
6	0.3	-10.7	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
7	0.2	-40.5	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
10	0.1	-70.2	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
12	2.0	495.2	REJECT ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
14	< 0.5		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
18	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
23	0.2	-40.5	EMISSION, IC PLASMA	3
26	0.4	19.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
30	0.3	-10.7	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
34	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
39	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
41	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
44	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
45	0.0	-100.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
46	0.5	48.8	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
50	< 0.3		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
52	0.0	-100.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
56	0.4	19.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
57	1.0	197.6	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
58	0.2	-40.5	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
60	< 0.3		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
61	1.1	227.4	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
62	1.0	197.6	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
64	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
69	4.0	1090.5	REJECT ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
70	0.4	19.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
71	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
72	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
73	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
77	0.1	-70.2	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
78	< 0.4		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
79	0.5	48.8	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
84	< 0.5		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
86	0.1	-70.2	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
94	0.1	-70.2	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
95	0.1	-70.2	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
96	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
97	< 0.2		IGNORED NOT REPORTED	
99	0.2	-40.5	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
102	0.4	19.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
104	0.5	48.8	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
105	0.1	-70.2	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
107	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
116	75.0	2E+04	REJECT EMISSION, IC PLASMA	3
117	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4

47 Labs had a total range of 0.0 to 75.0 and a mean of 0.34 with a standard deviation of 0.30 and a 95% confidence interval of the mean +/- 0.13.

Table 9 Standard Reference Water Sample 173 Report for LI

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	20.0	-9.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
2	18.0	-18.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
4	20.0	-9.1	EMISSION, FLAME	1
6	4.0	-81.8	OTHER	
19	21.0	-4.5	EMISSION, IC PLASMA	3,5
21	23.0	4.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
22	20.0	-9.1	EMISSION, IC PLASMA	3,5
28	21.0	-4.5	EMISSION, IC PLASMA	3,5
30	25.0	18.2	EMISSION, IC PLASMA	3,5
33	19.0	-13.6	NOT REPORTED	
39	< 10.0		IGNORED	
41	20.0	-9.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
59	25.0	13.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
61	<100.0		EMISSION, IC PLASMA	3,5
71	< 20.0		EMISSION, FLAME	1
72	20.0	-9.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
95	30.0	36.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
99	22.0	0.0	EMISSION, FLAME	1
102	25.0	13.6	EMISSION, IC PLASMA	3,5
112	< 25.0		IGNORED	

20 Labs had a total range of 4.0 to 30.0 and a mean of 22.00 with a standard deviation of 3.23 and a 95% confidence interval of the mean +/- 1.79.

Table 9 Standard Reference Water Sample T93 Report for MN

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	101	3.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	106	8.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
3	100	2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	100	2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
5	100	2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	10	-89.8	REJECT ATOMIC ABSORPTION, FLAMELESS	3
7	104	5.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	120	22.7	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	94	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	100	2.2	EMISSION, IC PLASMA	3,5
12	97	-0.8	EMISSION, IC PLASMA	3,5
13	60	-33.7	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
14	100	2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
18	100	2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	95	0.2	EMISSION, IC PLASMA	3,5
21	101	3.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	94	-3.9	EMISSION, IC PLASMA	3,5
23	110	12.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
24	100	2.2	EMISSION, IC PLASMA	3,5
25	100	2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
26	98	0.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
28	99	1.2	EMISSION, IC PLASMA	3,5
30	100	2.2	EMISSION, IC PLASMA	3,5
32	95	-1.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	164	57.7	REJECT NOT REPORTED	
34	65	-33.5	REJECT ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
35	93	0.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
39	110	12.5	EMISSION, IC PLASMA	3,5
40	90	-3.0	EMISSION, IC PLASMA	3,5
41	100	2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	94	-3.9	ATOMIC ABSORPTION, FLAMELESS	3
45	100	2.2	EMISSION, IC PLASMA	3,5
46	102	4.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	100	2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	80	-13.2	EMISSION, IC PLASMA	3,5
54	104	6.3	EMISSION, IC PLASMA	3,5
55	100	2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	95	-2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	90	-3.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
58	95	-2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	100	2.2	EMISSION, IC PLASMA	3,5
61	94	-3.9	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
62	80	-13.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	103	5.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
67	93	-4.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

Table 9 Standard Reference Water Sample T93 Report for MV

Code Number	Reported value	Pct. dev. from mean		Methods	References
69	105	7.3		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	125	27.8	REJECT	ATOMIC ABSORPTION, FLAMELESS	3
71	100	2.2		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	101	3.3		EMISSION, IC PLASMA	3,5
73	91	-7.0		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	46	-53.0	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	97	-0.8		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	100	2.2		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	80	-13.2		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	100	2.2		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	100	2.2		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	97	-0.8		EMISSION, IC PLASMA	3,5
97	104	6.3		NOT REPORTED	
99	101	3.3		EMISSION, IC PLASMA	3,5
100	93	0.2		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
101	90	-8.0		ATOMIC ABSORPTION, FLAMELESS	3
102	101	3.3		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
105	104	6.3		ATOMIC ABSORPTION, FLAMELESS	3
107	90	-3.0		EMISSION, IC PLASMA	3,5
112	91	-7.0		EMISSION, IC PLASMA	3,5
116	100	2.2		EMISSION, IC PLASMA	3,5
117	93	-4.9		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

67 Labs had a total range of 10 to 154 and a mean of 97.8 with a standard deviation of 6.1 and a 95% confidence interval of the mean +/- 1.6.

Table 9 Standard Reference Water Sample T93 Report for MO

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	15.0	-23.2	ATOMIC ABS, EXTRACTION, 8 HYDROXYQUINOLINE/MIBK, NITROUS OXIDE	4
4	21.0	7.5	ATOMIC ABSORPTION, FLAMELESS	3
6	17.0	-13.0	ATOMIC ABSORPTION, FLAMELESS	3
8	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
12	20.0	2.4	EMISSION, IC PLASMA	3, 5
19	23.0	17.8	EMISSION, IC PLASMA	3, 5
28	19.0	-2.7	EMISSION, IC PLASMA	3, 5
30	13.0	-7.8	EMISSION, IC PLASMA	3, 5
33	17.0	-13.0	NOT REPORTED	
39	< 50.0		IGNORED EMISSION, IC PLASMA	3, 5
41	< 50.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
50	30.0	53.6	ATOMIC ABSORPTION, FLAMELESS	3
51	< 10.0		IGNORED EMISSION, IC PLASMA	3, 5
58	47.0	140.6	REJECT EMISSION, IC PLASMA	3, 5
61	17.0	-13.0	ATOMIC ABS, EXTRACTION, 8 HYDROXYQUINOLINE/MIBK, NITROUS OXIDE	4
71	< 100.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
72	21.5	10.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
95	27.0	39.2	ATOMIC ABSORPTION, FLAMELESS	3
96	20.0	2.4	EMISSION, IC PLASMA	3, 5
99	15.0	-23.2	ATOMIC ABS, EXTRACTION, 8 HYDROXYQUINOLINE/MIBK, NITROUS OXIDE	4
102	21.0	7.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
107	< 60.0		IGNORED EMISSION, IC PLASMA	3, 5
112	11.0	-43.7	EMISSION, IC PLASMA	3, 5

23 Labs had a total range of 11.0 to 47.0 and a mean of 19.53 with a standard deviation of 4.63 and a 95% confidence interval of the mean +/- 2.47.

Table 9 Standard Reference Water Sample T93 Report for NI

Code Number	Reported value	Pct. dev. from mean		Methods	References
1	9.0	-29.5		ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
4	30.0	135.0	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	15.0	17.5		ATOMIC ABSORPTION, FLAMELESS	3
8	13.0	1.8		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	11.1	-13.0		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	9.0	-29.5		ATOMIC ABSORPTION, FLAMELESS	3
12	11.0	-13.8		ATOMIC ABSORPTION, FLAMELESS	3
14	11.0	-13.8		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
21	10.0	-21.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	20.0	56.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
26	35.0	174.2	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
28	12.0	-6.0		OTHER	
30	6.0	-53.0		OTHER	
32	16.0	25.4		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	7.2	-43.6		NOT REPORTED	
34	3.0	-76.5		ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
39	10.0	-21.7		ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
40	< 1.0		IGNORED	OTHER	
45	8.4	-34.2		OTHER	
50	<100.0		IGNORED	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	20.0	56.7		OTHER	
57	12.5	-2.1		ATOMIC ABSORPTION, FLAMELESS	3
58	23.0	80.2		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
60	15.0	17.5		ATOMIC ABSORPTION, FLAMELESS	3
61	9.0	-29.5		ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
62	< 50.0		IGNORED	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
64	6.6	-48.3		ATOMIC ABSORPTION, FLAMELESS	3
65	< 40.0		IGNORED	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	20.0	56.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	< 20.0		IGNORED	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	14.0	9.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	20.0	56.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
78	< 50.0		IGNORED	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	40.0	213.4	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	<100.0		IGNORED	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
92	< 50.0		IGNORED	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	11.1	-13.0		ATOMIC ABSORPTION, FLAMELESS	3
95	< 10.0		IGNORED	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	12.0	-6.0		OTHER	
97	16.0	25.4		NOT REPORTED	
99	12.0	-6.0		ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
102	11.0	-13.8		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	9.2	-27.9		ATOMIC ABSORPTION, FLAMELESS	3
105	17.0	33.2		ATOMIC ABSORPTION, FLAMELESS	3
107	< 30.0		IGNORED	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	22.0	72.4		EMISSION, IC PLASMA	3,5
116	11.0	-13.8		EMISSION, IC PLASMA	3,5
117	14.0	9.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	12.4	-2.9		MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

49 Labs had a total range of 3.0 to 40.0 and a mean of 12.76 with a standard deviation of 4.69 and a 95% confidence interval of the mean +/- 1.59.

Table 9 Standard Reference Water Sample T93 Report for PB

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	12.0	-3.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	6.2	-50.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	10.0	-19.5	ATOMIC ABSORPTION, FLAMELESS	3
5	20.0	61.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	20.0	61.0	ATOMIC ABSORPTION, FLAMELESS	3
7	26.1	110.1	ATOMIC ABSORPTION, FLAMELESS	3
8	11.0	-11.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	27.4	120.6	ANODIC STRIPPING VOLTAMMETRY	2
10	7.9	-36.4	ATOMIC ABSORPTION, FLAMELESS	3
12	11.0	-11.4	ATOMIC ABSORPTION, FLAMELESS	3
14	12.0	-3.4	ATOMIC ABSORPTION, FLAMELESS	3
18	< 5.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
19	< 10.0		IGNORED EMISSION, IC PLASMA	3,5
21	12.0	-3.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
23	7.0	-43.6	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
25	60.0	383.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
26	21.0	69.1	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
28	14.0	12.7	EMISSION, IC PLASMA	3,5
32	< 50.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	9.0	-27.5	NOT REPORTED	
34	8.0	-35.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
39	13.0	4.7	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
40	3.0	-75.8	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
41	< 50.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	14.0	12.7	ATOMIC ABSORPTION, FLAMELESS	3
45	9.5	-23.5	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
46	55.0	342.8	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	7.5	-39.6	ATOMIC ABSORPTION, FLAMELESS	3
51	60.0	383.0	REJECT EMISSION, IC PLASMA	3,5
57	13.4	7.9	ATOMIC ABSORPTION, FLAMELESS	3
58	14.0	12.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
60	14.0	12.7	ATOMIC ABSORPTION, FLAMELESS	3
61	135.0	986.3	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
62	15.0	20.8	ATOMIC ABSORPTION, FLAMELESS	3
64	3.6	-71.0	ATOMIC ABSORPTION, FLAMELESS	3
65	11.7	-5.8	ATOMIC ABSORPTION, FLAMELESS	3
69	24.0	93.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	3.0	-75.8	ATOMIC ABSORPTION, FLAMELESS	3
71	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	14.6	17.5	EMISSION, IC PLASMA	3,5
73	< 50.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	13.0	4.7	ATOMIC ABSORPTION, FLAMELESS	3
79	14.0	12.7	ATOMIC ABSORPTION, FLAMELESS	3
84	10.0	-19.5	ATOMIC ABSORPTION, FLAMELESS	3
86	15.5	24.8	ATOMIC ABSORPTION, FLAMELESS	3

Table 9 Standard Reference Water Sample T93 Report for Pb

Code Number	Reported value	Pct. dev. from mean	Methods	References
91	<200.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
92	<100.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	10.7	-13.9	ATOMIC ABSORPTION, FLAMELESS	3
95	14.0	12.7	ATOMIC ABSORPTION, FLAMELESS	3
96	11.0	-11.4	ATOMIC ABSORPTION, FLAMELESS	3
97	5.0	-59.7	NOT REPORTED	
99	15.0	20.8	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
102	11.0	-11.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	4.6	-63.0	ATOMIC ABSORPTION, FLAMELESS	3
105	7.9	-36.4	ATOMIC ABSORPTION, FLAMELESS	3
107	< 50.0		IGNORED EMISSION, IC PLASMA	3,5
112	21.0	69.1	EMISSION, IC PLASMA	3,5
116	44.0	254.2	REJECT EMISSION, IC PLASMA	3,5
118	11.4	-8.2	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

59 Labs had a total range of 3.0 to 135.0 and a mean of 12.42 with a standard deviation of 5.68 and a 95% confidence interval of the mean +/- 1.71.

Table 9 Standard Reference Water Sample T33 Report for S3

Code Number	Reported Value	Pct. dev. from mean	Methods	References
1	1.0	-15.9	NOT REPORTED	
6	0.8	-32.7	ATOMIC ABSORPTION, FLAMELESS	3
12	< 1.0		IGNORED	3
26	< 5.0		IGNORED	3
28	2.2	85.0	OTHER	
30	0.7	-41.1	ATOMIC ABSORPTION, HYDRIDE	2,4
33	5.8	387.8	REJECT	
39	1.0	-15.9	ATOMIC ABSORPTION, HYDRIDE	2,4
40	2.0	63.2	ATOMIC ABSORPTION, FLAMELESS	3
50	< 5.0		IGNORED	3
51	< 10.0		IGNORED	
61	1.0	-15.9	ATOMIC ABSORPTION, HYDRIDE	2,4
62	1.0	-15.9	ATOMIC ABSORPTION, FLAMELESS	3
71	< 100.0		IGNORED	1,3
72	< 4.0		IGNORED	3
95	< 10.0		IGNORED	3
96	< 20.0		IGNORED	3
99	3.8	219.6	REJECT	
102	1.0	-15.9	ATOMIC ABSORPTION, HYDRIDE	2,4
107	< 100.0		IGNORED	3
116	144.0	1E+04	REJECT EMISSION, IC PLASMA	1,3

21 Labs had a total range of 0.7 to 144.0 and a mean of 1.19 with a standard deviation of 0.53 and a 95% confidence interval of the mean +/- 0.41.

Table 9 Standard Reference Water Sample T93 Report for SE

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	7.0	30.4	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
2	< 2.0		NOT REPORTED	
4	4.0	-25.5	ATOMIC ABSORPTION, FLAMELESS	3
6	5.4	0.6	ATOMIC ABSORPTION, FLAMELESS	3
7	0.4	-92.5	ATOMIC ABSORPTION, FLAMELESS	3
8	5.0	-6.8	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
10	4.9	-8.7	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
12	6.0	11.3	ATOMIC ABSORPTION, FLAMELESS	3
14	5.0	11.8	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
18	5.3	-1.3	ATOMIC ABSORPTION, FLAMELESS	3
22	8.3	54.6	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
26	4.7	-12.4	ATOMIC ABSORPTION, FLAMELESS	3
29	5.5	2.5	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
30	5.3	-1.3	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
34	15.0	179.5	OTHER	
39	5.0	-5.8	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
40	5.6	4.3	ATOMIC ABSORPTION, FLAMELESS	3
41	6.0	11.3	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
44	4.6	-14.3	ATOMIC ABSORPTION, FLAMELESS	3
45	5.4	0.6	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
46	3.2	-40.4	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
50	6.2	15.5	ATOMIC ABSORPTION, FLAMELESS	3
51	6.0	11.8	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
52	6.0	11.8	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
57	4.8	-10.6	ATOMIC ABSORPTION, FLAMELESS	3
53	5.3	8.1	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
61	5.7	5.2	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
62	4.2	-21.8	ATOMIC ABSORPTION, FLAMELESS	3
65	5.1	-5.0	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
67	5.4	0.6	ATOMIC ABSORPTION, HYDRIDE	3
70	6.0	11.8	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
72	< 6.0		ATOMIC ABSORPTION, FLAMELESS	3
73	< 10.0		ATOMIC ABSORPTION, FLAMELESS	3
77	6.0	11.8	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
78	5.4	0.6	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
84	5.0	11.8	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
86	5.6	4.3	ATOMIC ABSORPTION, FLAMELESS	3
94	38.0	608.0	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
95	< 10.0		ATOMIC ABSORPTION, FLAMELESS	3
96	4.6	-14.3	NOT REPORTED	
97	3.0	-44.1	NOT REPORTED	
99	5.7	6.2	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
102	5.0	11.8	ATOMIC ABSORPTION, FLAMELESS	3
105	5.7	6.2	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
107	2.0	-62.7	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
116	< 10.0		EMISSION, IC PLASMA	3
117	6.5	21.1	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4

47 Labs had a total range of 0.4 to 38.0 and a mean of 5.37 with a standard deviation of 0.83 and a 95% confidence interval of the mean +/- 0.23.

Table 9 Standard Reference Water Sample T93 Report for SR

Code Number	Reported value	Pct. dev. from mean		Methods	References
1	70	9.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
2	50	-21.6		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
4	70	9.7		OTHER	
6	100	56.8	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
8	60	-5.9		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
15	0	-100.0	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
19	62	-2.8		EMISSION, IC PLASMA	3,5
22	61	-4.4		EMISSION, IC PLASMA	3,5
28	63	-1.2		EMISSION, IC PLASMA	3,5
30	63	-1.2		EMISSION, IC PLASMA	3,5
33	65	1.9		NOT REPORTED	
39	65	1.9		EMISSION, IC PLASMA	3,5
41	60	-5.9		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
44	64	0.3		OTHER	
45	62	-2.8		EMISSION, IC PLASMA	3,5
54	66	3.5		EMISSION, IC PLASMA	3,5
55	70	9.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
58	43	-24.7		EMISSION, IC PLASMA	3,5
59	63	-1.2		EMISSION, IC PLASMA	3,5
61	100	56.8	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
72	66	3.5		EMISSION, IC PLASMA	3,5
95	30	25.4		ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
96	63	-1.2		EMISSION, IC PLASMA	3,5
99	66	3.5		EMISSION, IC PLASMA	3,5
102	3	-95.3	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
112	65	1.9		EMISSION, IC PLASMA	3,5
118	65	1.9		MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

27 Labs had a total range of 0 to 100 and a mean of 63.3 with a standard deviation of 6.3 and a 95% confidence interval of the mean +/- 2.7.

Table 9 Standard Reference Water Sample 193 Report for TL

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	2.0	-37.3	ATOMIC ABSORPTION, FLAMELESS	3
6	5.3	92.0	ATOMIC ABSORPTION, FLAMELESS	3
8	< 20.0		IGNORED	1,3
12	< 50.0		IGNORED	1,3
28	7.2	125.9	ATOMIC ABSORPTION, DIRECT, AIR	
30	< 1.0		OTHER	
33	150.0	4605.9	OTHER	
39	2.0	-37.3	NOT REPORTED	
41	< 50.0		ATOMIC ABSORPTION, FLAMELESS	3
44	< 0.5		ATOMIC ABSORPTION, DIRECT, AIR	1,3
50	2.0	-37.3	ATOMIC ABSORPTION, FLAMELESS	3
51	< 10.0		OTHER	3
57	1.2	-62.4	ATOMIC ABSORPTION, FLAMELESS	3
61	< 1.0		ATOMIC ABSORPTION, DIRECT, AIR	1,3
71	< 50.0		ATOMIC ABSORPTION, DIRECT, AIR	1,3
72	< 4.0		ATOMIC ABSORPTION, FLAMELESS	3
95	< 100.0		ATOMIC ABSORPTION, FLAMELESS	1,3
96	< 25.0		ATOMIC ABSORPTION, DIRECT, AIR	
99	< 10.0		OTHER	
102	3.0	-5.9	ATOMIC ABSORPTION, FLAMELESS	3
107	< 50.0		ATOMIC ABSORPTION, FLAMELESS	3
118	2.3	-27.8	ATOMIC ABSORPTION, DIRECT, AIR	1,3
			MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

22 Labs had a total range of 1.2 to 150.0 and a mean of 3.19
with a standard deviation of 2.14 and a 95% confidence interval of the mean +/- 1.79.

Table 9 Standard Reference Water Sample T93 Report for ZN

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	30	5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	25	-11.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
5	51	80.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	86	203.7	REJECT ATOMIC ABSORPTION, FLAMELESS	3
7	21	-25.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	23	-18.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	21	-25.8	ANODIC STRIPPING VOLTAMMETRY	
10	29	2.4	EMISSION, IC PLASMA	3,5
12	27	-4.7	EMISSION, IC PLASMA	3,5
14	22	-22.3	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2
18	25	-11.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	24	-15.2	EMISSION, IC PLASMA	3,5
21	27	-4.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	26	-3.2	EMISSION, IC PLASMA	3,5
23	30	5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
24	30	5.9	EMISSION, IC PLASMA	3,5
25	30	5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
26	34	20.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
28	27	-4.7	EMISSION, IC PLASMA	3,5
30	33	16.5	EMISSION, IC PLASMA	3,5
32	48	69.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	56	97.8	NOT REPORTED	
34	2	-92.9	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2
39	40	41.3	EMISSION, IC PLASMA	3,5
40	106	274.3	REJECT EMISSION, IC PLASMA	3,5
41	26	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	28	-1.1	ATOMIC ABSORPTION, FLAMELESS	3
45	35	23.6	EMISSION, IC PLASMA	3,5
46	11	-51.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	20	-29.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	< 10		IGNORED EMISSION, IC PLASMA	3,5
54	29	2.4	EMISSION, IC PLASMA	3,5
55	30	5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	24	-15.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	30	5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
58	31	9.5	ANODIC STRIPPING VOLTAMMETRY	
59	29	2.4	EMISSION, IC PLASMA	3,5
60	31	9.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
61	26	-8.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
62	32	13.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	23	-18.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
67	40	41.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	33	16.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	28	-1.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	30	5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

Table 9 Standard Reference Water Sample T93 Report for Zn

Code Number	Reported value	Pct. dev. from mean	Methods	References
72	28	-1.1	EMISSION, IC PLASMA	3,5
73	<		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	60	111.9	ANODIC STRIPPING VOLTAMMETRY	
77	29	2.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
79	21	-25.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	10	-64.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	3	-71.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	40	41.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
92	20	-29.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	26	-8.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	20	-29.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	25	-11.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
97	29	-1.1	EMISSION, IC PLASMA	3,5
99	30	5.9	NOT REPORTED	
101	51	80.1	EMISSION, IC PLASMA	3,5
102	30	5.9	ATOMIC ABSORPTION, FLAMELESS	3
104	26	-8.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
105	26	-8.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
107	10	-64.7	ATOMIC ABSORPTION, FLAMELESS	3
112	28	-1.1	EMISSION, IC PLASMA	3,5
116	25	-11.7	EMISSION, IC PLASMA	3,5
117	26	-8.2	EMISSION, IC PLASMA	1,2,3,4
			ATOMIC ABSORPTION, DIRECT, AIR	

67 Labs had a total range of 2 to 106 and a mean of 28.3 with a standard deviation of 10.2 and a 95% confidence interval of the mean +/- 2.6.

Table 10. Statistics by method for standard reference sample 193

Determination	Method	Range: from to	Mean	Standard Deviation	N
ACIDSGAC03	TITRATION, COLORIMETRIC, MANUAL	10.000 - 1020.000	1010.000	3.162	9
	TITRATION, ELECTROMETRIC, MANUAL	990.000 - 1140.000	1035.333	41.208	15
	OVER-ALL	10.000 - 1440.000	1025.625	29.205	16
AG	ATOMIC ABSORPTION, DIRECT, AIR	1.000 - 9.000	4.267	2.508	9
	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1.000 - 5.000	2.400	1.890	4
	ATOMIC ABSORPTION, FLAMELESS	0.300 - 4.300	2.642	1.038	19
	NOT REPORTED	2.700 - 250.000	2.850	0.387	35
	OVER-ALL	0.300 - 250.000	3.225	1.979	35
AL	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	10.000 - 776.000	162.778	74.920	9
	ATOMIC ABSORPTION, DIRECT, FLAMELESS	50.000 - 200.000	150.167	43.933	12
	EMISSION, IC PLASMA	140.000 - 275.000	180.500	39.300	12
	OVER-ALL	10.000 - 776.000	167.351	43.659	37
AS	ATOMIC ABSORPTION, FLAMELESS	3.100 - 10.000	6.089	1.903	19
	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	2.100 - 6.000	4.814	1.367	7
	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	3.800 - 5.100	5.093	0.884	14
	OVER-ALL	1.800 - 57.000	5.441	1.627	42
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	40.000 - 224.000	94.067	56.900	15
BA	ATOMIC ABSORPTION, FLAMELESS	8.000 - 102.000	68.571	31.633	7
	EMISSION, IC PLASMA	60.000 - 85.000	68.611	4.132	18
	OVER-ALL	8.000 - 224.000	70.571	12.486	42
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	3.000 - 20.000	7.143	2.410	7
	ATOMIC ABSORPTION, FLAMELESS	6.500 - 15.000	9.140	3.798	5
BE	EMISSION, IC PLASMA	4.000 - 7.800	6.200	1.083	14
	OVER-ALL	3.000 - 64.000	6.689	1.716	25
	ATOMIC ABSORPTION, DIRECT, AIR	4.000 - 10.000	5.900	1.755	15
	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	5.000 - 5.000	5.000	0.000	3
	ATOMIC ABSORPTION, FLAMELESS	1.000 - 8.600	4.600	0.570	25
CO	EMISSION, IC PLASMA	4.000 - 8.000	4.500	0.443	7
	OVER-ALL	1.000 - 10.000	4.787	0.610	52
	ATOMIC ABSORPTION, DIRECT, AIR	6.000 - 42.000	6.500	0.707	13
	ATOMIC ABSORPTION, FLAMELESS	4.000 - 13.000	7.800	4.660	3
	EMISSION, IC PLASMA	3.900 - 7.000	5.071	1.163	7
CR TOT	_OVER-ALL_	3.900 - 42.000	5.150	1.095	16
	ATOMIC ABSORPTION, DIRECT, AIR	3.800 - 20.000	10.915	5.294	13
	ATOMIC ABSORPTION, FLAMELESS	3.000 - 14.300	10.583	1.748	23
	EMISSION, IC PLASMA	8.400 - 16.000	12.300	2.898	6
	NOT REPORTED	6.000 - 26.000	14.667	10.263	3
CU	_OVER-ALL_	2.400 - 26.000	10.734	4.015	47
	ATOMIC ABSORPTION, DIRECT, AIR	4.000 - 55.000	30.091	9.511	33
	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	29.000 - 34.000	31.000	2.646	3

Table 10. Statistics by method for standards reference sample 133

Determination	Method	Range: from to	Mean	Standard Deviation	N
CU	ATOMIC ABSORPTION, FLAMELESS	10.000 - 50.000	29.800	9.636	10
	EMISSION, IC PLASMA	21.000 - 36.000	30.214	4.117	14
	OVER-ALL	4.000 - 55.000	29.754	9.288	65
FE	ATOMIC ABSORPTION, DIRECT, AIR	0.000 - 178.000	99.833	21.365	37
	ATOMIC ABSORPTION, FLAMELESS	10.000 - 140.000	33.800	47.246	5
	EMISSION, IC PLASMA	60.000 - 220.000	102.333	10.411	15
	OTHER	0.000 - 306.000	154.000	132.755	4
	OVER-ALL	0.000 - 1035.000	102.016	21.363	61
HG	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	0.000 - 4.000	0.375	0.450	4
	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	0.000 - 2.000	0.335	0.289	20
	EMISSION, IC PLASMA	0.200 - 75.000	37.600	6.116	
	OVER-ALL	0.000 - 75.000	0.336	0.304	25
	ATOMIC ABSORPTION, DIRECT, AIR	18.000 - 23.000	20.200	1.789	5
LI	EMISSION, FLAME	20.000 - 30.000	25.000	5.000	3
	EMISSION, IC PLASMA	20.000 - 25.000	22.500	2.429	6
	OVER-ALL	4.000 - 30.000	22.000	3.229	15
	ATOMIC ABSORPTION, DIRECT, AIR	46.000 - 120.000	99.147	4.157	34
	ATOMIC ABSORPTION, FLAMELESS	10.000 - 125.000	84.600	43.850	5
MN	EMISSION, IC PLASMA	30.000 - 110.000	98.444	4.233	18
	OVER-ALL	10.000 - 164.000	97.817	6.066	60
	ATOMIC ABSORPTION, DIRECT, AIR	15.000 - 17.000	15.000	0.000	2
	ATOMIC ABSORPTION, DIRECT, AIR	21.000 - 21.500	21.250	0.500	4
	ATOMIC ABSORPTION, FLAMELESS	17.000 - 30.000	23.750	5.852	4
MO	EMISSION, IC PLASMA	11.000 - 47.000	18.500	4.037	6
	OVER-ALL	11.000 - 47.000	19.531	4.631	16
	ATOMIC ABS, EXTRACTION, 8-HYDROXYQUINOLINE/MIBK, NITROUS OXIDE	10.000 - 40.000	19.207	9.269	15
	ATOMIC ABSORPTION, DIRECT, AIR	3.000 - 12.000	8.500	3.873	4
	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	6.600 - 17.000	11.822	3.369	9
NI	ATOMIC ABSORPTION, FLAMELESS	6.000 - 20.000	11.680	5.302	5
	OTHER	3.000 - 40.000	12.764	4.694	36
	OVER-ALL	10.000 - 40.000	19.207	9.269	15
	ATOMIC ABSORPTION, DIRECT, AIR	6.200 - 135.000	21.200	18.670	11
	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	3.000 - 21.000	11.417	6.344	6
PB	ATOMIC ABSORPTION, DIRECT, AIR	3.000 - 26.100	11.082	4.127	22
	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	14.000 - 60.000	30.720	20.418	5
	EMISSION, IC PLASMA	3.000 - 135.000	12.422	5.683	45
	OVER-ALL	0.800 - 2.000	1.200	0.542	4
	ATOMIC ABSORPTION, FLAMELESS	0.700 - 3.800	1.000	0.000	2
SB	ATOMIC ABSORPTION, HYDRIDE	2.200 - 2.200	2.200	0.000	1
	OTHER	0.700 - 144.000	1.189	0.530	9
	OVER-ALL	0.400 - 38.000	5.281	0.725	16
	ATOMIC ABSORPTION, FLAMELESS	2.000 - 8.300	5.679	0.516	19
	ATOMIC ABSORPTION, HYDRIDE				

Table 10. Statistics by method for standard reference sample 193

Determination	Method	Range:		Mean	Standard Deviation	N
		from	to			
SE	_OVER-ALL_	0.400	38.000	5.369	0.832	37
SR	ATOMIC ABSORPTION, DIRECT, AIR	0.000	100.000	59.300	34.551	10
	EMISSION, IC PLASMA	48.000	66.000	63.750	1.765	12
	OVER-ALL	0.000	100.000	63.793	6.339	23
TL	ATOMIC ABSORPTION, DIRECT, AIR	-	-	63.750	1.118	5
	ATOMIC ABSORPTION, FLAMELESS	1.200	5.800	2.040	0.639	1
	OTHER	7.200	7.200	7.200	0.639	1
	OVER-ALL	1.200	150.000	3.188	2.136	8
ZN	ANODIC STRIPPING VOLTAMMETRY	21.000	60.000	37.333	20.257	3
	ATOMIC ABSORPTION, DIRECT, AIR	8.000	51.000	27.333	8.922	36
	ATOMIC ABSORPTION, FLAMELESS	26.000	86.000	47.750	27.909	4
	EMISSION, IC PLASMA	10.000	105.000	28.333	2.968	15
	OVER-ALL	2.000	106.000	28.318	10.190	63

Table 11 Standard Reference Water Sample N15 Report for NH3-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.58	3.6	OTHER	
5	0.45	-19.6	ION SELECTIVE ELECTRODE	1,2,3,4
6	0.40	-28.5	OTHER	
8	0.88	57.3	REJECT COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
9	0.42	-24.9	ION SELECTIVE ELECTRODE	1,2,3,4
10	0.59	5.4	ION SELECTIVE ELECTRODE	1,2,3,4
12	0.55	-1.7	COLORIMETRIC, PHENATE	1,2,3
13	0.41	-26.7	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
14	0.55	-1.7	COLORIMETRIC, PHENATE	1,2,3
15	0.46	-17.8	COLORIMETRIC, PHENATE	1,2,3
16	0.53	-5.3	ION SELECTIVE ELECTRODE	1,2,3,4
18	0.46	-17.8	COLORIMETRIC, PHENATE	1,2,3
20	0.58	3.6	COLORIMETRIC, PHENATE	1,2,3
21	0.62	10.8	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
25	0.60	7.2	OTHER	
26	0.46	-17.8	COLORIMETRIC, PHENATE	1,2,3
27	0.55	-1.7	COLORIMETRIC, PHENATE	1,2,3
28	0.68	21.5	COLORIMETRIC, PHENATE	1,2,3
29	0.53	-5.3	ION SELECTIVE ELECTRODE	1,2,3,4
30	0.66	17.9	ION SELECTIVE ELECTRODE	1,2,3,4
31	0.80	43.0	ION SELECTIVE ELECTRODE	1,2,3,4
40	< 0.10		IGNORED ION SELECTIVE ELECTRODE	1,2,3,4
41	0.56	0.1	COLORIMETRIC, PHENATE	1,2,3
42	0.49	-12.4	ION SELECTIVE ELECTRODE	1,2,3,4
44	0.15	-73.2	REJECT OTHER	
46	1.00	78.7	REJECT ION SELECTIVE ELECTRODE	1,2,3,4
50	0.60	7.2	ION SELECTIVE ELECTRODE	1,2,3,4
51	0.70	25.1	COLORIMETRIC, PHENATE	1,2,3
55	0.52	-7.1	OTHER	
56	0.75	34.0	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
57	0.52	-7.1	COLORIMETRIC, PHENATE	1,2,3
58	0.53	-5.3	ION SELECTIVE ELECTRODE	1,2,3,4
61	0.43	-14.2	ION SELECTIVE ELECTRODE	1,2,3,4
62	0.57	1.9	COLORIMETRIC, PHENATE	1,2,3
64	0.55	-1.7	COLORIMETRIC, PHENATE	1,2,3
65	0.65	15.2	COLORIMETRIC, INDOPHENOL	4
67	0.54	-3.5	COLORIMETRIC, INDOPHENOL	4
69	0.68	21.5	ION SELECTIVE ELECTRODE	1,2,3,4
71	1.12	100.1	REJECT OTHER	
72	0.53	-5.3	COLORIMETRIC, PHENATE	1,2,3
73	0.50	-10.7	COLORIMETRIC, PHENATE	1,2,3
74	0.44	-21.4	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
75	0.57	1.9	COLORIMETRIC, PHENATE	1,2,3
77	0.55	-1.7	COLORIMETRIC, INDOPHENOL	4
79	0.60	7.2	COLORIMETRIC, PHENATE	1,2,3

Table 11 Standard Reference Water Sample N15 Report for NH3-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
86	0.57	1.9	COLORIMETRIC, PHENATE	1,2,3
87	0.46	-17.8	COLORIMETRIC, PHENATE	1,2,3
89	0.70	25.1	ION SELECTIVE ELECTRODE	1,2,3,4
91	0.61	9.0	COLORIMETRIC, PHENATE	1,2,3
94	0.53	3.6	COLORIMETRIC, PHENATE	1,2,3
95	0.68	21.5	COLORIMETRIC, PHENATE	1,2,3
97	0.52	-7.1	NOT REPORTED	
98	0.39	-30.3	COLORIMETRIC, PHENATE	1,2,3
102	0.55	-1.7	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
104	0.69	23.3	ION SELECTIVE ELECTRODE	1,2,3,4
105	0.55	-1.7	ION SELECTIVE ELECTRODE	1,2,3,4
107	0.53	-5.3	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
109	0.52	-7.1	COLORIMETRIC, PHENATE	1,2,3
111	0.55	-1.7	COLORIMETRIC, PHENATE	1,2,3
114	0.67	19.7	COLORIMETRIC, PHENATE	1,2,3

60 Labs had a total range of 0.15 to 1.12 and a mean of 0.560 with a standard deviation of 0.089 and a 95% confidence interval of the mean +/- 0.024.

Table 11 Standard Reference Water Sample M15 Report for N02-V

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
4	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
5	0.06	40.5	COLORIMETRIC, DIAZOTIZATION	1,3,4
7	0.03	-29.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
10	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
12	<		IGNORED	1,3,4
13	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
14	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
15	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
18	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
21	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
23	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
25	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
26	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
27	0.00	-100.0	REJECT	1,3,4
29	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
41	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
42	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
44	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
46	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
50	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
51	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
55	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
57	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
58	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
60	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
62	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
65	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
67	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
69	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
71	0.03	-29.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
72	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
73	0.16	274.7	REJECT	1,3,4
74	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
75	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
76	0.03	-29.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
77	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
86	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
87	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
89	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
91	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
92	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
95	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
97	0.04	-6.3	NOT REPORTED	1,3,4
99	<		IGNORED	1,3,4

Table 11 Standard Reference Water Sample N15 Report for N02-V

Code Number	Reported value	Pct. dev. from mean	Methods	References
102	0.05	17.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
104	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
105	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4
107	0.01	-76.6	REJECT	1,3,4
109	< 0.01		IGNORED	1,3,4
111	< 0.02		IGNORED	1,3,4
114	0.04	-6.3	COLORIMETRIC, DIAZOTIZATION	1,3,4

52 Labs had a total range of 0.00 to 0.16 and a mean of 0.043 with a standard deviation of 0.006 and a 95% confidence interval of the mean +/- 0.002.

Table 11 Standard Reference Water Sample N15 Report for N03-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.98	2.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
4	0.89	-5.6	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
5	0.75	-21.3	ION SELECTIVE ELECTRODE	1,2,3,4
6	0.99	3.9	ION CHROMATOGRAPHY	2,3,6,7
7	1.20	25.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
10	0.93	-2.4	COLORIMETRIC, BRUCINE	1,2,3,4
12	0.93	-2.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
13	0.07	-92.7	REJECT COLORIMETRIC, BRUCINE	1,2,3,4
16	0.12	-87.4	REJECT COLORIMETRIC, BRUCINE	1,2,3,4
18	0.98	2.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
20	1.00	4.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
21	0.24	-74.8	REJECT COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
23	0.94	-1.4	COLORIMETRIC, BRUCINE	1,2,3,4
24	1.00	4.9	OTHER	
25	1.01	6.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
26	1.00	4.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
27	0.96	0.7	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
29	1.03	3.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
30	0.90	-5.6	ION CHROMATOGRAPHY	2,3,6,7
41	0.91	-4.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
42	0.96	0.7	COLORIMETRIC, BRUCINE	1,2,3,4
46	1.05	10.2	COLORIMETRIC, BRUCINE	1,2,3,4
50	0.96	0.7	COLORIMETRIC, BRUCINE	1,2,3,4
51	1.02	7.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
55	0.95	-0.3	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
56	0.84	-11.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
57	0.90	-5.6	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
58	0.92	-3.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
61	0.34	-64.3	REJECT COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
62	0.82	-13.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
64	0.91	-4.5	ION CHROMATOGRAPHY	2,3,6,7
65	0.82	-13.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
67	0.93	-2.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
69	0.81	-15.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
71	1.00	4.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
72	1.04	9.1	ION CHROMATOGRAPHY	2,3,6,7
73	0.97	1.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
74	0.87	-8.7	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
75	1.05	10.2	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
76	1.01	6.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
77	0.92	-3.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
79	0.98	2.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
86	0.89	-6.6	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
87	1.31	37.5	REJECT COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
89	0.70	-26.5	COLORIMETRIC, BRUCINE	1,2,3,4

Table 11 Standard Reference Water Sample N15 Report for N03-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
91	0.92	-3.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
92	0.94	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
95	1.00	4.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
97	0.97	1.8	NOT REPORTED	
99	1.00	4.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
102	0.99	3.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
104	0.99	3.9	COLORIMETRIC, BRUCINE	1,2,3,4
105	0.97	1.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
107	1.03	8.1	COLORIMETRIC, BRUCINE	1,2,3,4
109	0.61	-36.0	REJECT COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
111	1.05	10.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
114	1.02	7.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4

57 Labs had a total range of 0.07 to 1.31 and a mean of 0.953 with a standard deviation of 0.083 and a 95% confidence interval of the mean +/- 0.024.

Table 11 Standard Reference Water Sample N15 Report for JRG-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.53	-23.2	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
5	0.45	-39.0	DIGESTION, DISTILLATION, ION SELECTIVE ELECTRODE	1,2,3
10	0.51	-30.9	DIGESTION, DISTILLATION, ION SELECTIVE ELECTRODE	1,2,3
12	0.91	23.3	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
13	0.33	-55.3	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	2,3,4
14	0.78	5.7	COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	3,7
15	0.77	4.4	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
21	0.94	27.4	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	2,3,4
25	0.38	-43.5	OTHER	
26	0.76	3.0	COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	3,7
27	0.45	-39.0	OTHER	
28	< 0.05		IGNORED COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
40	< 0.10		IGNORED DIGESTION, DISTILLATION, TITRATION	2,3,4
41	0.60	-13.7	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	2,3,4
42	1.10	49.1	DIGESTION, DISTILLATION, ION SELECTIVE ELECTRODE	1,2,3
44	1.09	47.7	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
46	0.10	-36.4	DIGESTION, DISTILLATION, ION SELECTIVE ELECTRODE	1,2,3
50	0.80	3.4	DIGESTION, DISTILLATION, ION SELECTIVE ELECTRODE	1,2,3
51	0.40	-45.8	COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	3,7
55	0.63	-14.6	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
56	0.70	-5.1	NOT REPORTED	
57	0.63	-14.6	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
58	0.80	3.4	DIGESTION, DISTILLATION, TITRATION	2,3,4
62	0.59	-20.0	OTHER	
65	1.15	55.3	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
69	0.52	-29.5	COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	3,7
71	1.35	83.0	DIGESTION, DISTILLATION, TITRATION	2,3,4
72	0.37	17.9	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
73	2.30	211.7	REJECT COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	3,7
77	0.51	-30.9	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
79	1.00	35.5	COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	3,7
86	1.23	66.7	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
94	1.39	83.4	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
95	0.47	-36.3	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	2,3,4
97	1.01	36.9	NOT REPORTED	
98	0.69	-6.5	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
102	0.82	11.1	DIGESTION, DISTILLATION, TITRATION	2,3,4
105	1.13	53.1	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	2,3,4
107	0.02	-97.3	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	2,3,4
109	1.07	45.0	COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	3,7
111	0.56	-24.1		

41 Labs had a total range of 0.02 to 2.30 and a mean of 0.733 with a standard deviation of 0.322 and a 95% confidence interval of the mean +/- 0.106.

Table 11 Standard Reference Water Sample N15 Report for P, TOTAL

Code Number	Reported value	Pct. dev. from mean		Methods	References
1	0.82	4.8		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
2	0.21	-73.1	REJECT	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
4	0.75	-4.1		COLORIMETRIC, BLK DIG, H2SO4, K8HG2SO4, PHOSPHOMOLYBDATE	4
6	1.45	85.4	REJECT	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
8	0.63	-19.4		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
10	0.80	2.3		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
12	0.75	-4.1		COLORIMETRIC, BLK DIG, H2SO4, K8HG2SO4, PHOSPHOMOLYBDATE	4
13	0.78	-0.3		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
15	0.82	4.8		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
16	0.79	1.0		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
21	0.77	-1.5		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
25	0.83	12.5		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
26	0.80	2.3		COLORIMETRIC, BLK DIG, H2SO4, K8HG2SO4, PHOSPHOMOLYBDATE	4
27	0.77	-1.5		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
28	0.73	-6.7		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
30	0.85	10.0		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
31	0.77	-1.5		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
41	0.80	2.3		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
42	0.78	-0.3		OTHER	
44	0.81	3.6		COLORIMETRIC, BLK DIG, H2SO4, K8HG2SO4, PHOSPHOMOLYBDATE	4
46	0.78	-0.3		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
50	0.83	6.1		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
51	0.81	3.6		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
55	0.72	-7.9		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
56	0.78	-0.3		COLORIMETRIC, BLK DIG, H2SO4, K8HG2SO4, PHOSPHOMOLYBDATE	4
57	0.81	3.6		COLORIMETRIC, BLK DIG, H2SO4, K8HG2SO4, PHOSPHOMOLYBDATE	4
58	0.63	-19.4		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
60	0.79	1.0		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
61	0.79	1.0		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
62	0.83	6.1		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
64	0.74	-5.4		EMISSION, IC PLASMA	3,5
65	0.78	-0.3		COLORIMETRIC, BLK DIG, H2SO4, K8HG2SO4, PHOSPHOMOLYBDATE	4
69	0.70	-10.5		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
71	0.70	-10.5		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
72	0.80	2.3		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
73	0.73	-6.7		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
74	0.70	-10.5		OTHER	
75	0.82	4.8		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
77	0.84	7.4		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
79	0.83	6.1		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
86	0.87	11.2		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
87	0.67	-14.3		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
89	0.77	-1.5		COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBDO	1,2,3,4
92	0.10	-87.2	REJECT	COLORIMETRIC, BLK DIG, H2SO4, K8HG2SO4, PHOSPHOMOLYBDATE	4
94	0.73	-0.3		COLORIMETRIC, BLK DIG, H2SO4, K8HG2SO4, PHOSPHOMOLYBDATE	4

Table 11 Standard Reference Water Sample N15 Report for P, TOTAL

Code Number	Reported value	Pct. dev. from mean	Methods	References
95	0.81	3.6	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
97	0.75	-4.1	NOT REPORTED	
98	0.74	-5.4	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
99	0.82	4.8	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
102	0.84	7.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
104	0.92	17.6	OTHER	
105	0.59	-24.6	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
107	0.85	3.7	EMISSION, IC PLASMA	3,5
109	0.93	13.9	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
111	0.81	3.6	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4

55 Labs had a total range of 0.10 to 1.45 and a mean of 0.782 with a standard deviation of 0.067 and a 95% confidence interval of the mean +/- 0.019.

Table 11 Standard Reference Water Sample N15 Report for P04-P

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
4	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
6	0.45	135.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
7	0.35	83.3	REJECT OTHER	
10	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
12	0.22	15.2	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
13	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
14	0.16	-16.2	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
15	0.13	-5.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
21	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
25	0.20	4.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
26	0.22	15.2	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
27	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
30	0.20	4.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
31	0.21	10.0	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
41	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
42	0.20	4.8	OTHER	
44	0.17	-10.9	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
46	0.24	25.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
50	0.24	25.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
51	0.17	-10.9	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
55	0.14	-26.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
56	0.13	-5.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
57	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
58	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
60	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
61	0.21	10.0	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
62	0.13	-5.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
67	0.42	120.0	REJECT COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
69	0.13	-5.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
71	0.13	-5.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
72	0.17	-10.9	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
73	0.22	15.2	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
75	0.20	4.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
76	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
77	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
79	0.20	4.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
86	0.13	-5.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
87	0.33	72.9	REJECT COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
89	0.13	-5.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
92	0.02	-89.5	REJECT COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
94	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
95	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
97	0.16	-16.2	NOT REPORTED	
98	0.19	-0.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4

Table 11 Standard Reference Water Sample N15 Report for P04-P

Code Number	Reported value	Pct. dev. from mean		Methods	References
102	0.26	36.2	REJECT	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
104	0.28	46.7	REJECT	OTHER	
105	0.19	-0.5		COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
109	0.02	-89.5	REJECT	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
111	0.13	-5.7		COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4
114	0.20	4.8		COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1,2,3,4

51 Labs had a total range of 0.02 to 0.45 and a mean of 0.191 with a standard deviation of 0.019 and a 95% confidence interval of the mean +/- 0.006.

Table 12. Statistics by method for standard reference sample N15

Determination	Method	Range: from to	Mean	Standard Deviation	N	
NH3-N	COLORIMETRIC, DISTILLATION, NESSLERIZATION	0.410 - 0.880	0.597	0.168	7	
	COLORIMETRIC, INCORPENOL	0.540 - 0.650	0.545	0.071		
	COLORIMETRIC, PHENAT	0.460 - 0.570	0.527	0.059	3	
	COLORIMETRIC, PHENATE	0.390 - 0.700	0.558	0.078	23	
	ION SELECTIVE ELECTRODE	0.420 - 1.000	0.580	0.107	15	
	OTHER	0.150 - 1.120	0.562	0.320	6	
	-OVER-ALL-	0.150 - 1.120	0.560	0.089	55	
	COLORIMETRIC, DIAZOTIZATION	0.000 - 0.160	0.043	0.006	44	
	-OVER-ALL-	0.000 - 0.160	0.043	0.006	45	
	NO3-N	COLORIMETRIC, BRUCINE	0.070 - 1.050	0.775	0.371	10
COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION		0.240 - 1.200	0.949	0.065	33	
COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION		0.340 - 1.310	0.900	0.410	4	
ION CHROMATOGRAPHY		0.900 - 1.040	0.960	0.067	4	
-OVER-ALL-		0.070 - 1.310	0.953	0.083	51	
OR3-N		COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	0.510 - 1.390	0.867	0.291	12
		COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	0.020 - 1.130	0.582	0.405	6
		COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	0.400 - 2.300	0.755	0.261	6
		DIGESTION, DISTILLATION, ION SELECTIVE ELECTRODE	0.100 - 1.100	0.592	0.378	5
		DIGESTION, DISTILLATION, TITRATION	0.300 - 1.350	0.810	0.100	
	OTHER	0.380 - 0.590	0.473	0.107	3	
	-OVER-ALL-	0.020 - 2.300	0.738	0.322	38	
	P, TOTAL	COLORIMETRIC, H2SO4/PERSULF DIG, ASCORBIC ACID PHOSPHOMOLYBDATE	0.210 - 1.450	0.778	0.069	36
		COLORIMETRIC/BLK DIG/H2SO4, KH2SO4, PHOSPHOMOLYBDATE	0.100 - 0.930	0.778	0.026	9
		OTHER	0.700 - 0.920	0.800	0.111	3
-OVER-ALL-		0.100 - 1.450	0.782	0.067	52	
PO4-P		COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	0.020 - 0.420	0.183	0.019	22
	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	0.020 - 0.450	0.199	0.023	20	
	OTHER	0.200 - 0.350	0.277	0.075	3	
	-OVER-ALL-	0.020 - 0.450	0.191	0.019	43	

Table 13 Standard Reference Water Sample P9 Report for ACIDACAC03

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	2400	-1.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
12	2440	0.5	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
14	2400	-1.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
16	2470	2.6	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
32	2440	0.5	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
71	2400	-1.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
77	2430	0.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
94	2450	1.0	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
95	2525	4.0	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
99	2382	-1.2	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
107	2338	-3.7	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4

11 Labs had a total range of 2338 to 2525 and a mean of 2433 with a standard deviation of 51.6 and a 95% confidence interval of the mean +/- 34.6.

Table 13 Standard Reference Water Sample P9 Report for AG

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1.00		IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,2,4
5	8.00	2380.6	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
6	0.19	-41.1	ATOMIC ABSORPTION, FLAMELESS	3
8	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
10	0.60	86.0	ATOMIC ABSORPTION, FLAMELESS	3
12	0.22	-31.8	ATOMIC ABSORPTION, FLAMELESS	3
14	0.10	-69.0	ATOMIC ABSORPTION, FLAMELESS	3
18	< 2.50		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
23	0.05	-84.5	ATOMIC ABSORPTION, FLAMELESS	3
30	1.00	210.1	OTHER	
32	30.00	9202.3	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
33	0.63	95.3	NOT REPORTED	
41	< 5.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
44	0.16	-50.4	ATOMIC ABSORPTION, FLAMELESS	3
48	< 0.50		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
64	0.40	24.0	ATOMIC ABSORPTION, FLAMELESS	3
71	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
72	0.32	-0.8	ATOMIC ABSORPTION, FLAMELESS	3
77	0.00	-100.0	ATOMIC ABSORPTION, FLAMELESS	3
91	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
94	0.20	-33.0	ATOMIC ABSORPTION, FLAMELESS	3
95	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
99	< 1.00		IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,2,4
107	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
112	< 6.00		IGNORED OTHER	
115	< 0.03		IGNORED OTHER	

26 Labs had a total range of 0.00 to 30.00 and a mean of 0.323 with a standard deviation of 0.291 and a 95% confidence interval of the mean +/- 0.185.

Table 13 Standard Reference Water Sample P9 Report for CA

Code Number	Reported value	Pct. dev. from mean		Methods	References
1	0.17	-32.6		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	0.30	18.9		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	0.33	30.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	0.20	-20.8		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	0.30	18.9		EMISSION, IC PLASMA	3,5,7
11	0.23	-8.9		EMISSION, IC PLASMA	3,5,7
12	0.21	-16.8		EMISSION, IC PLASMA	3,5,7
13	3.00	1088.6	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
14	1.40	454.7	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	2.17	759.7	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
18	0.60	137.7	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	0.24	-4.9		EMISSION, IC PLASMA	3,5,7
23	0.50	98.1	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	0.26	3.0		EMISSION, IC PLASMA	3,5,7
32	0.30	18.9		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	0.35	38.7		NOT REPORTED	
37	0.23	-3.9		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
41	0.27	7.0		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	0.30	18.9		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	0.20	-20.8		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	0.29	14.9		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	0.28	10.9		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	< 1.00		IGNORED	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	0.25	-1.0		OTHER	
72	0.29	14.9		EMISSION, IC PLASMA	3,5,7
75	0.16	-35.6		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	0.22	-12.8		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
82	0.25	-1.0		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	0.66	161.5	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	0.23	-8.9		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	0.31	22.8		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	0.24	-4.9		EMISSION, IC PLASMA	3,5,7
101	0.18	-28.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
107	0.22	-12.8		EMISSION, IC PLASMA	3,5,7
112	0.24	-4.9		EMISSION, IC PLASMA	3,5,7
115	0.27	7.0		EMISSION, IC PLASMA	3,5,7

36 Labs had a total range of 0.16 to 3.00 and a mean of 0.252 with a standard deviation of 0.048 and a 95% confidence interval of the mean +/- 0.018.

Table 13 Standard Reference Water Sample P9 Report for Cd

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1.00	-24.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
5	3.00	125.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	1.20	-9.8	ATOMIC ABSORPTION, FLAMELESS	3
8	< 5.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	1.80	35.3	ANODIC STRIPPING VOLTAMMETRY, DIFFERENTIAL PULSE	2
10	1.10	-17.3	ATOMIC ABSORPTION, FLAMELESS	3
12	1.08	-18.8	ATOMIC ABSORPTION, FLAMELESS	3
14	0.80	-39.9	ATOMIC ABSORPTION, FLAMELESS	3
15	1.10	-17.3	ATOMIC ABSORPTION, FLAMELESS	3
16	< 5.00		IGNORED OTHER	
18	1.00	-24.9	ATOMIC ABSORPTION, FLAMELESS	3
19	1.30	-2.3	EMISSION, IC PLASMA	3,5
23	1.00	-24.9	ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1,4
30	1.00	-24.9	EMISSION, IC PLASMA	3,5
32	< 3.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	1.00	-24.9	NOT REPORTED	
41	< 5.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	1.10	-17.3	ATOMIC ABSORPTION, FLAMELESS	3
57	1.10	-17.3	ATOMIC ABSORPTION, FLAMELESS	3
64	1.10	-17.3	ATOMIC ABSORPTION, FLAMELESS	3
65	1.03	-22.6	ATOMIC ABSORPTION, FLAMELESS	3
71	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	1.50	12.7	EMISSION, IC PLASMA	3,5
77	1.00	-24.9	ATOMIC ABSORPTION, FLAMELESS	3
86	1.00	-24.9	ATOMIC ABSORPTION, FLAMELESS	3
91	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	0.92	-30.9	ATOMIC ABSORPTION, FLAMELESS	3
95	2.60	95.4	ATOMIC ABSORPTION, FLAMELESS	3
99	2.60	95.4	ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1,4
107	<10.00		IGNORED EMISSION, IC PLASMA	3,5
112	0.40	-69.9	EMISSION, IC PLASMA	3,5
115	3.00	125.4	EMISSION, IC PLASMA	3,5
118	0.87	-34.6	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

33 Labs had a total range of 0.40 to 3.00 and a mean of 1.331 with a standard deviation of 0.686 and a 95% confidence interval of the mean +/- 0.277.

Table 13 Standard Reference Water Sample P9 Report for CL

Code Number	Reported value	Pct. dev. from mean	Methods	References
5	0.02	-99.5	ION SELECTIVE ELECTRODE	1,2,3,4
6	0.49	-83.3	ION CHROMATOGRAPHY	2,3,6
8	< 5.00		IGNORED TITRATION, SILVER NITRATE	1,2,4
9	< 1.00		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
10	0.00	-100.0	TITRATION, SILVER NITRATE	1,2,4
11	14.88	255.2	ION CHROMATOGRAPHY	2,3,6
12	0.40	-90.5	ION CHROMATOGRAPHY	2,3,6
14	0.30	-92.8	ION SELECTIVE ELECTRODE	1,2,3,4
16	2.62	-37.5	ION SELECTIVE ELECTRODE	1,2,3,4
18		3839.1	REJECT TITRATION, SILVER NITRATE	1,2,4
32	2.30	-45.1	TITRATION, MERCURIC NITRATE	1,2,3,4
41	< 1.00		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
44	< 0.50		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
57	13.00	210.4	TITRATION, MERCURIC NITRATE	1,2,3,4
64	2.00	-52.3	TITRATION, MERCURIC NITRATE	1,2,3,4
65	< 2.00		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
71	< 2.00		IGNORED TITRATION, SILVER NITRATE	1,2,4
72	2.00	-52.3	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
77	6.71	60.2	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
86	0.00	-100.0	ION CHROMATOGRAPHY	2,3,6
91	1.80	-57.0	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
94	16.80	301.1	ION SELECTIVE ELECTRODE	1,2,3,4
95	31.00	640.1	REJECT COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
107	3.70	-11.7	TITRATION, SILVER NITRATE	1,2,4
111	<10.00		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4

25 Labs had a total range of 0.00 to and a mean of 4.139 with a standard deviation of 5.625 and a 95% confidence interval of the mean +/- 2.995.

Table 13 Standard Reference Water Sample P9 Report for CO

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1.00		IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
6	0.29	-29.7	ATOMIC ABSORPTION, FLAMELESS	3
8	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	< 3.00		IGNORED EMISSION, IC PLASMA	3,5
23	0.00	-100.0	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
33	0.46	11.5	NOT REPORTED	
41	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	0.90	113.2	EMISSION, IC PLASMA	3,5
77	8.33	1919.4	REJECT ATOMIC ABSORPTION, FLAMELESS	3
95	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	< 3.00		IGNORED EMISSION, IC PLASMA	3,5
107	<10.00		IGNORED EMISSION, IC PLASMA	3,5
112	3.00	627.3	REJECT EMISSION, IC PLASMA	3,5
115	35.00	8384.8	REJECT EMISSION, IC PLASMA	3,5

15 Labs had a total range of 0.00 to 35.00 and a mean of 0.413 with a standard deviation of 0.375 and a 95% confidence interval of the mean +/- 0.599.

Table 15 Standard Reference Water Sample p9 Report for CR TOT

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	3.0	85.6	ATOMIC ABSORPTION, EXTRACTION (APDC/MISK)	1,3,4
5	2.0	23.7	ATOMIC ABSORPTION, FLAMELESS	3
6	3.7	128.9	ATOMIC ABSORPTION, FLAMELESS	3
8	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	< 2.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	< 0.0	-100.0	ATOMIC ABSORPTION, FLAMELESS	3
12	< 1.0		ATOMIC ABSORPTION, FLAMELESS	3
14	< 0.4	-75.3	ATOMIC ABSORPTION, FLAMELESS	3
15	170.4	1E+04	ATOMIC ABSORPTION, FLAMELESS	3
18	< 5.0		REJECT ATOMIC ABSORPTION, FLAMELESS	3
30	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
32	< 8.0		IGNORED EMISSION, IC PLASMA	3
33	< 0.3	-81.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
41	< 20.0		NOT REPORTED	
44	< 0.4	-75.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	< 3.0		ATOMIC ABSORPTION, FLAMELESS	3
57	< 1.0		ATOMIC ABSORPTION, FLAMELESS	3
64	< 0.7		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
65	< 30.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
71	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	< 0.3	-81.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	2.2	36.1	EMISSION, IC PLASMA	3
86	2.0	23.7	ATOMIC ABSORPTION, FLAMELESS	3
91	< 20.0		ATOMIC ABSORPTION, FLAMELESS	3
94	0.1	-93.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	< 10.0		ATOMIC ABSORPTION, FLAMELESS	3
99	< 3.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
107	< 20.0		EMISSION, IC PLASMA	3
115	5.0	209.3	EMISSION, IC PLASMA	3

29 Labs had a total range of 0.0 to 170.4 and a mean of 1.62 with a standard deviation of 1.64 and a 95% confidence interval of the mean +/- 1.04.

Table 13 Standard Reference Water Sample P9 Report for CU

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	5.00	26.2	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
6	4.00	1.0	ATOMIC ABSORPTION, FLAMELESS	3
8	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	4.60	16.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	2.00	-49.5	EMISSION, IC PLASMA	3,5
12	< 5.00		IGNORED EMISSION, IC PLASMA	3,5
14	4.00	1.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
18	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	<10.00		IGNORED EMISSION, IC PLASMA	3,5
30	5.00	26.2	EMISSION, IC PLASMA	3,5
32	28.00	607.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	6.30	59.1	NOT REPORTED	
41	< 5.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	3.80	-4.1	ATOMIC ABSORPTION, FLAMELESS	3
48	3.00	-24.3	NOT REPORTED	
57	4.80	21.2	ATOMIC ABSORPTION, FLAMELESS	3
64	8.00	102.0	REJECT ATOMIC ABSORPTION, FLAMELESS	3
65	3.70	-6.6	ATOMIC ABSORPTION, FLAMELESS	3
71	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	3.90	-1.5	EMISSION, IC PLASMA	3,5
77	2.00	-49.5	ATOMIC ABSORPTION, FLAMELESS	3
86	4.60	16.1	ATOMIC ABSORPTION, FLAMELESS	3
91	25.00	531.2	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	3.50	-11.6	ATOMIC ABSORPTION, FLAMELESS	3
95	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	4.80	21.2	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
107	<10.00		IGNORED EMISSION, IC PLASMA	3,5
112	2.50	-36.9	EMISSION, IC PLASMA	3,5
115	20.00	405.0	REJECT EMISSION, IC PLASMA	3,5
118	3.79	-4.3	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

30 Labs had a total range of 2.00 to 28.00 and a mean of 3.96 with a standard deviation of 1.11 and a 95% confidence interval of the mean +/- 0.552.

Table 13 Standard Reference Water Sample P9 Report for F

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1.00		ION SELECTIVE ELECTRODE	1,2,3,4
5	0.06	41.2	OTHER	
6	0.48	1029.4	ION CHROMATOGRAPHY	2,3,6
8	0.25	488.2	OTHER	
10	0.02	-52.9	OTHER	
12	< 0.02		ION CHROMATOGRAPHY	2,3,6
13	0.02	-52.9	OTHER	
14	0.04	-15.9	OTHER	
15	0.10	135.3	COLORIMETRIC, SPADNS	1,2,3
16	0.06	41.2	OTHER	
18	< 0.10		OTHER	
23	0.00	-100.0	COLORIMETRIC, ZIRCONYL ALIZARIN	1
44	< 0.10		ION SELECTIVE ELECTRODE	1,2,3,4
64	0.03	-29.4	ION SELECTIVE ELECTRODE	1,2,3,4
71	< 0.10		OTHER	
72	< 0.10		ION SELECTIVE ELECTRODE	1,2,3,4
77	0.01	-76.5	ION SELECTIVE ELECTRODE	1,2,3,4
86	0.00	-100.0	ION CHROMATOGRAPHY	2,3,6
91	< 0.50		OTHER	
94	< 0.10		ION SELECTIVE ELECTRODE	1,2,3,4
95	< 0.10		COLORIMETRIC, CEROUS ALIZARIN "COMPLEXONE"	3
107	0.11	156.8	ION SELECTIVE ELECTRODE	1,2,3,4
115	0.06	41.2	COLORIMETRIC, SPADNS	1,2,3

23 Labs had a total range of 0.00 to 0.48 and a mean of 0.043 with a standard deviation of 0.037 and a 95% confidence interval of the mean +/- 0.023.

Table 13 Standard Reference Water Sample P9 Report for FE

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	27.0	-31.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	9.4	-73.2	ATOMIC ABSORPTION, FLAMELESS	3
8	< 10.0		IGNORED	1,2,3,4
9	24.5	-37.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	4.0	-89.9	EMISSION, IC PLASMA	3,5
12	< 10.0		IGNORED	3,5
13	180.0	355.3	OTHER	
14	10.0	-74.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	0.2	-99.5	OTHER	
18	< 30.0		IGNORED	1,2,3,4
19	< 3.0		IGNORED	1,2,3,4
23	5.0	-87.3	EMISSION, IC PLASMA	3,5
30	13.0	-67.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
32	86.0	113.0	EMISSION, IC PLASMA	3,5
33	196.0	396.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
41	< 10.0		NOT REPORTED	
44	< 20.0		IGNORED	1,2,3,4
57	20.0	-49.3	ATOMIC ABSORPTION, DIRECT, AIR	3
65	< 50.0		IGNORED	1,2,3,4
71	< 20.0		IGNORED	1,2,3,4
72	23.3	-40.9	EMISSION, IC PLASMA	3,5
77	8.3	-79.0	ATOMIC ABSORPTION, FLAMELESS	3
91	< 50.0		IGNORED	1,2,3,4
94	< 10.0		IGNORED	1,2,3,4
95	< 50.0		IGNORED	1,2,3,4
99	< 10.0		IGNORED	1,2,3,4
107	< 10.0		IGNORED	3,5
112	4.4	-88.8	EMISSION, IC PLASMA	3,5
115	20.0	-49.3	EMISSION, IC PLASMA	3,5

29 Labs had a total range of 0.2 to 196.0 and a mean of 39.44 with a standard deviation of 61.36 and a 95% confidence interval of the mean +/- 32.69.

Table 13 Standard Reference Water Sample P9 Report for K

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.05	-15.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
5	0.11	83.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	0.06	0.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	0.06	0.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	< 0.10		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	0.00	-100.0	OTHER	
11	0.05	-15.7	EMISSION, FLAME	1,2
12	0.05	-15.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	0.20	233.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	< 0.10		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
16	0.42	600.0	OTHER	
18	0.10	65.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
23	0.10	65.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	0.05	-15.7	OTHER	
32	0.10	65.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	0.04	-33.3	NOT REPORTED	
37	0.07	15.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
41	0.05	-15.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	< 0.50		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	< 0.10		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	0.05	0.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	0.02	-65.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	< 0.50		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	0.05	-15.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	0.06	0.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	0.06	0.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	0.03	-50.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
82	0.05	-15.7	EMISSION, FLAME	1,2
91	< 0.05		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	0.05	-15.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	0.11	83.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	0.72	1100.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
101	0.04	-33.3	OTHER	
107	0.09	50.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
115	1.00	1565.7	OTHER	

35 Labs had a total range of 0.00 to 1.00 and a mean of 0.060 with a standard deviation of 0.028 and a 95% confidence interval of the mean +/- 0.012.

Table 13 Standard Reference Water Sample P9 Report for MS

Code Number	Reported Value	Pct. dev. from mean	Methods	References
1	0.04	27.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	0.03	-4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	0.04	27.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	< 0.10		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	0.00	-100.0	EMISSION, IC PLASMA	3,5
11	0.03	-4.2	EMISSION, IC PLASMA	3,5
12	0.03	-4.2	EMISSION, IC PLASMA	3,5
14	< 0.20		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	< 0.26		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
18	0.07	125.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	0.03	-4.2	EMISSION, IC PLASMA	3,5
23	0.03	-4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	0.03	-4.2	EMISSION, IC PLASMA	3,5
32	0.30	358.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	0.06	91.7	EMISSION, IC PLASMA	3,5
37	0.03	-4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
41	0.03	-4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	< 0.10		NOT REPORTED	1,2,3,4
48	< 0.10		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	0.03	-4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	0.03	-4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	< 0.50		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	0.04	27.8	OTHER	1,2,3,4
72	0.03	-4.2	EMISSION, IC PLASMA	3,5
75	0.02	-36.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	0.03	-4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
82	0.03	-4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	< 0.05		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	0.03	-4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	0.04	27.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	0.03	-4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
101	0.03	-4.2	EMISSION, IC PLASMA	3,5
107	< 0.01		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	0.03	-4.2	EMISSION, IC PLASMA	3,5
115	0.03	-4.2	EMISSION, IC PLASMA	3,5

35 Labs had a total range of 0.00 to 0.30 and a mean of 0.031 with a standard deviation of 0.005 and a 95% confidence interval of the mean +/- 0.002.

Table 13 Standard Reference Water Sample P9 Report for MN

Code Number	Reported Value	Pct. dev. from mean	Methods	References
1	< 10.0		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
5	28.0	210.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	9.4	-6.9	ATOMIC ABSORPTION, DIRECT, AIR	3
8	13.0	44.1	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
9	11.9	31.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	7.0	-22.4	ATOMIC ABSORPTION, DIRECT, AIR	3,5
12	6.0	-33.5	EMISSION, IC PLASMA	3,5
14	10.0	10.5	EMISSION, IC PLASMA	1,2,3,4
18	< 10.0		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	9.9	9.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
23	10.0	10.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	8.0	-11.3	EMISSION, IC PLASMA	3,5
32	< 5.0		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	15.0	66.2	EMISSION, IC PLASMA	1,2,3,4
41	< 10.0		NOT REPORTED	3
44	8.0	-11.3	ATOMIC ABSORPTION, DIRECT, AIR	3,5
54	8.2	-9.1	ATOMIC ABSORPTION, FLAMELESS	3
57	8.5	-5.8	EMISSION, IC PLASMA	3,5
65	< 10.0		ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
71	10.0	10.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	10.0	10.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	4.0	-55.7	ATOMIC ABSORPTION, DIRECT, AIR	3,5
77	10.0	10.8	EMISSION, IC PLASMA	1,2,3,4
91	< 10.0		ATOMIC ABSORPTION, DIRECT, AIR	3
94	10.0	10.8	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
95	10.0	10.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	5.3	-41.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
107	< 10.0		EMISSION, IC PLASMA	3,5
112	7.3	-19.1	EMISSION, IC PLASMA	3,5
115	8.0	-11.3	EMISSION, IC PLASMA	3,5

30 Labs had a total range of 4.0 to 28.0 and a mean of 9.02 with a standard deviation of 2.46 and a 95% confidence interval of the mean +/- 1.09.

Table 13 Standard Reference Water Sample 60 Report for Na

Code Number	Reported Value	Pct. dev. from mean	Methods	References
1	0.11	3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	0.03	-25.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	0.09	-15.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	< 0.10		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	0.10	75.2	EMISSION, IC PLASMA	3,5
11	0.07	-34.3	EMISSION, IC PLASMA	3,5
12	0.14	31.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	0.10	-5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
14	0.10	-5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	52.00	58+04	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
16	0.52	357.3	OTHER	
18	0.20	87.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	< 0.20		EMISSION, IC PLASMA	3,5
23	0.10	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	0.19	78.2	EMISSION, IC PLASMA	3,5
32	0.10	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	0.20	87.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	0.11	3.2	NOT REPORTED	
41	0.12	12.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	< 0.50		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	0.10	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	0.16	50.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	0.10	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	< 5.00		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	0.04	-62.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	0.07	-34.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	0.06	-43.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	0.06	-43.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
82	0.03	-25.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	< 0.10		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	0.07	-34.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	0.12	12.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	0.05	-53.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
101	0.08	-25.0	EMISSION, IC PLASMA	3,5
107	0.88	725.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	0.06	-43.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
115	0.23	115.8	EMISSION, IC PLASMA	3,5

37 Labs had a total range of 0.04 to 52.00 and a mean of 0.107 with a standard deviation of 0.048 and a 95% confidence interval of the mean +/- 0.018.

Table 13 Standard Reference Water Sample P9 Report for NH3-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
6	0.20	174.0	OTHER	1,4
8	0.21	187.7	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,2,3,4
9	0.03	-58.9	ION SELECTIVE ELECTRODE	1,2,3
12	0.03	-58.9	COLORIMETRIC, PHENATE	1,4
13	0.03	-58.9	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,2,3
15	0.10	37.0	COLORIMETRIC, PHENATE	1,2,3,4
16	0.04	-45.2	ION SELECTIVE ELECTRODE	1,2,3
18	< 0.01		IGNORED	1,2,3
20	0.43	489.0	REJECT	1,2,3
30	0.04	-45.2	ION SELECTIVE ELECTRODE	1,2,3,4
31	0.12	64.4	ION SELECTIVE ELECTRODE	1,2,3,4
37	0.10	37.0	COLORIMETRIC, INDOPHENOL	4
41	0.03	-58.9	COLORIMETRIC, PHENATE	1,2,3
44	0.02	-72.6	OTHER	1,2,3
48	0.03	-58.9	COLORIMETRIC, PHENATE	4
54	0.06	-17.8	COLORIMETRIC, INDOPHENOL	1,2,3
57	< 0.05		IGNORED	1,2,3
64	0.03	-58.9	COLORIMETRIC, PHENATE	1,2,3
72	0.05	-31.5	COLORIMETRIC, PHENATE	1,2,3
75	0.07	-4.1	COLORIMETRIC, PHENATE	1,2,3
77	0.02	-72.6	COLORIMETRIC, INDOPHENOL	4
82	0.03	-58.9	NOT REPORTED	
91	0.12	64.4	COLORIMETRIC, PHENATE	1,2,3
94	< 0.05		IGNORED	1,2,3
95	< 0.05		IGNORED	1,2,3
101	0.08	9.6	COLORIMETRIC, PHENATE	1,2,3
107	0.21	187.7	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
111	0.03	-58.9	COLORIMETRIC, PHENATE	1,2,3

28 Labs had a total range of 0.02 to 0.43 and a mean of 0.073 with a standard deviation of 0.061 and a 95% confidence interval of the mean +/- 0.027.

Table 13 Standard Reference Water Sample P9 Report for P3

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
5	30.00	263.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	10.00	21.2	ATOMIC ABSORPTION, FLAMELESS	3
8	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	18.70	126.6	ANODIC STRIPPING VOLTAMMETRY	2
12	1.30	-84.2	ATOMIC ABSORPTION, FLAMELESS	3
14	1.00	-87.9	ANODIC STRIPPING VOLTAMMETRY	2
18	< 5.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
19	<10.00		IGNORED EMISSION, IC PLASMA	3,5
23	0.50	-93.9	ANODIC STRIPPING VOLTAMMETRY	2
32	<50.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	3.40	-58.8	NOT REPORTED	
41	<50.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	2.20	-73.3	ATOMIC ABSORPTION, FLAMELESS	3
57	12.80	55.1	ATOMIC ABSORPTION, FLAMELESS	3
64	2.60	-68.5	ATOMIC ABSORPTION, FLAMELESS	3
65	< 2.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
71	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	5.30	-35.8	EMISSION, IC PLASMA	3,5
77	4.80	-41.8	ATOMIC ABSORPTION, FLAMELESS	3
86	23.00	178.7	ATOMIC ABSORPTION, FLAMELESS	3
91	<		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	1.30	-84.2	ATOMIC ABSORPTION, FLAMELESS	3
95	12.00	45.4	ATOMIC ABSORPTION, FLAMELESS	3
99	3.10	-62.4	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
107	<50.00		IGNORED EMISSION, IC PLASMA	3,5
112	3.40	-53.8	EMISSION, IC PLASMA	3,5
115	20.00	142.3	EMISSION, IC PLASMA	3,5
118	1.42	-82.8	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

29 Labs had a total range of 0.50 to 30.00 and a mean of 8.254 with a standard deviation of 8.801 and a 95% confidence interval of the mean +/- 4.242.

Table 13 Standard Reference Water Sample P9 Report for PH

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1.36	-10.9	ELECTROMETRIC	1,2,3,4
5	1.49	-2.4	ELECTROMETRIC	1,2,3,4
6	1.60	4.8	ELECTROMETRIC	1,2,3,4
8	1.50	-1.7	ELECTROMETRIC	1,2,3,4
10	1.36	-10.9	ELECTROMETRIC	1,2,3,4
12	1.33	-12.9	ELECTROMETRIC	1,2,3,4
13	2.67	74.9	REJECT ELECTROMETRIC	1,2,3,4
14	1.40	-3.3	ELECTROMETRIC	1,2,3,4
15	1.35	-10.9	ELECTROMETRIC	1,2,3,4
16	1.93	29.7	ELECTROMETRIC	1,2,3,4
18	1.70	11.4	ELECTROMETRIC	1,2,3,4
20	1.40	-3.3	ELECTROMETRIC	1,2,3,4
23	1.80	17.9	ELECTROMETRIC	1,2,3,4
30	1.58	3.5	ELECTROMETRIC	1,2,3,4
31	1.37	-10.2	ELECTROMETRIC	1,2,3,4
32	1.90	24.5	ELECTROMETRIC	1,2,3,4
37	1.60	4.8	ELECTROMETRIC	1,2,3,4
44	1.33	-12.9	ELECTROMETRIC	1,2,3,4
48	1.40	-3.3	ELECTROMETRIC	1,2,3,4
54	1.35	-11.6	ELECTROMETRIC	1,2,3,4
57	1.37	-10.2	ELECTROMETRIC	1,2,3,4
64	1.40	-3.3	ELECTROMETRIC	1,2,3,4
65	2.09	36.9	ELECTROMETRIC	1,2,3,4
71	1.70	11.4	ELECTROMETRIC	1,2,3,4
72	1.42	-7.0	ELECTROMETRIC	1,2,3,4
75	1.84	20.6	NOT REPORTED	
77	1.38	-9.6	ELECTROMETRIC	1,2,3,4
82	1.36	-10.9	NOT REPORTED	
86	1.47	-3.7	ELECTROMETRIC	1,2,3,4
91	1.55	1.6	ELECTROMETRIC	1,2,3,4
94	1.57	2.9	ELECTROMETRIC	1,2,3,4
95	1.50	-1.7	ELECTROMETRIC	1,2,3,4
99	1.40	-3.3	ELECTROMETRIC	1,2,3,4
101	1.42	-7.0	ELECTROMETRIC	1,2,3,4
107	1.70	11.4	ELECTROMETRIC	1,2,3,4
115	1.44	-5.7	ELECTROMETRIC	1,2,3,4

36 Labs had a total range of 1.33 to 2.67 and a mean of 1.526 with a standard deviation of 0.199 and a 95% confidence interval of the mean +/- 0.053.

Table 13 Standard Reference Water Sample P9 Report for S04

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1.20	-64.0	NOT REPORTED	
6	1.10	-67.0	ION CHROMATOGRAPHY	2,6
8	< 10.00		IGNORED TURBIDIMETRIC, BARIUM SULFATE	1,2,3
9	< 1.00		IGNORED COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
10	1.00	-70.0	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
11	34.20	2429.1	REJECT ION CHROMATOGRAPHY	2,6
12	< 0.10		IGNORED ION CHROMATOGRAPHY	2,6
14	< 2.50		IGNORED TURBIDIMETRIC, BARIUM SULFATE	1,2,3
15	3.70	11.1	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
16	0.40	-88.0	GRAVIMETRIC, BARIUM SULFATE	1,2,3
18	< 0.50		IGNORED COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
23	0.00	-100.0	COLORIMETRIC, CHLORANILATE	3
32	12.80	284.5	GRAVIMETRIC, BARIUM SULFATE	1,2,3
44	< 5.00		IGNORED TURBIDIMETRIC, BARIUM SULFATE	1,2,3
54	0.40	-88.0	ION CHROMATOGRAPHY	2,6
57	< 1.00		IGNORED TURBIDIMETRIC, BARIUM SULFATE	1,2,3
65	< 10.00		IGNORED COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
71	3.00	-9.9	GRAVIMETRIC, BARIUM SULFATE	1,2,3
72	0.50	-35.0	ION CHROMATOGRAPHY	2,6
77	6.23	87.1	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
91	9.40	182.3	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
94	4.31	29.5	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
95	1.00	-70.0	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
107	4.90	47.2	GRAVIMETRIC, BARIUM SULFATE	1,2,3

24 Labs had a total range of 0.00 to 84.20 and a mean of 3.329 with a standard deviation of 3.734 and a 95% confidence interval of the mean +/- 2.068.

Table 13 Standard Reference Water Sample 09 Report for SP. COND.

Code Number	Reported value	Pct. dev. from mean	Methods	References
5	20400	29.5	ELECTRODELESS, INDUCTIVE CELL-TYPE	2
6	20750	31.9	DIRECT READING INSTRUMENT	4
9	19200	22.0	DIRECT READING INSTRUMENT	4
10	14150	-19.1	DIRECT READING INSTRUMENT	4
12	18500	17.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
14	150	-99.0	NOT REPORTED	
15	19300	22.6	DIRECT READING INSTRUMENT	4
16	19500	23.9	DIRECT READING INSTRUMENT	4
23	50	-99.7	NOT REPORTED	
30	19500	23.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
32	18050	14.7	DIRECT READING INSTRUMENT	4
37	23000	45.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
44	19000	20.7	DIRECT READING INSTRUMENT	4
48	19000	20.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
54	19200	22.0	NOT REPORTED	
57	100	-99.4	DIRECT READING INSTRUMENT	4
64	16500	4.9	NOT REPORTED	
65	15800	0.4	NOT REPORTED	
71	16200	2.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
77	19400	23.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
82	20400	29.5	NOT REPORTED	
86	17350	10.3	NOT REPORTED	
91	1250	-92.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
94	19000	20.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
95	19800	25.8	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
99	19300	22.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
107	22500	43.0	DIRECT READING INSTRUMENT	4
111	100	-99.4	DIRECT READING INSTRUMENT	4
115	18900	20.1	NOT REPORTED	

29 Labs had a total range of 50 to 23000 and a mean of 15736 with a standard deviation of 7376 and a 95% confidence interval of the mean +/- 2305.

Table 13 Standard Reference Water Sample P6 Report for TL

Code Number	Reported value	Pct. dev. from mean	Methods	References
6	3.90	15.47	REJECT	
8	<20.00		ATOMIC ABSORPTION, FLAMELESS	3
12	<50.00		IGNORED	1,3
30	<1.00		ATOMIC ABSORPTION, DIRECT, AIR	1,3
41	<50.00		IGNORED	
44	1.60	4.9	ATOMIC ABSORPTION, DIRECT, AIR	1,3
57	<1.00		REJECT	3
71	<50.00		ATOMIC ABSORPTION, FLAMELESS	3
72	<4.00		IGNORED	1,3
95	<100		ATOMIC ABSORPTION, DIRECT, AIR	3
99	<10.00		IGNORED	1,3
107	<50.00		ATOMIC ABSORPTION, FLAMELESS	3
115	<500		IGNORED	1,3
118	1.45	-4.9	OTHER	
14			MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

Labs had a total range of 1.45 to 3.90 and a mean of 1.525 with a standard deviation of 0.274 and a 95% confidence interval of the mean +/- 0.000.

Table 13 Standard Reference Water Sample P9 Report for ZN

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	15.0	-17.4	ANODIC STRIPPING VOLTAMMETRY	
5	28.0	54.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	17.0	-6.4	ATOMIC ABSORPTION, FLAMELESS	3
8	11.0	-39.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	12.0	-33.9	ANODIC STRIPPING VOLTAMMETRY	
10	9.0	-50.5	EMISSION, IC PLASMA	3,5
12	9.0	-50.5	EMISSION, IC PLASMA	3,5
14	7.0	-61.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
18	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	< 3.0		IGNORED EMISSION, IC PLASMA	3,5
23	10.0	-45.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	21.0	15.6	EMISSION, IC PLASMA	3,5
32	40.0	120.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	28.0	54.1	NOT REPORTED	
41	40.0	120.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
44	76.0	318.3	REJECT ATOMIC ABSORPTION, FLAMELESS	3
54	62.0	241.3	EMISSION, IC PLASMA	3,5
57	60.0	230.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	10.0	-45.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	10.6	-41.7	EMISSION, IC PLASMA	3,5
75	5.0	-72.5	ANODIC STRIPPING VOLTAMMETRY	
77	13.0	-28.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	3.0	-83.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	<100.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	9.0	-50.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	10.0	-45.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	11.0	-39.5	EMISSION, IC PLASMA	3,5
107	< 10.0		IGNORED EMISSION, IC PLASMA	3,5
112	5.6	-69.2	EMISSION, IC PLASMA	3,5
115	8.0	-56.0	EMISSION, IC PLASMA	3,5

31 Labs had a total range of 3.0 to 76.0 and a mean of 18.17 with a standard deviation of 15.21 and a 95% confidence interval of the mean +/- 6.69.

Table 1-4. Statistics by method for standard reference sample 29

Determination	Method	Range:		Mean	Standard Deviation
		from	to		
ACID3AC03	TITRATION, ELECTROMETRIC, MANUAL	2338.000	- 2525.000	2426.818	51.557 11
	OVER-ALL	2338.000	- 2525.000	2426.818	51.557 11
AG	ATOMIC ABSORPTION, DIRECT, AIR	3.000	- 30.000	19.000	3.317
	ATOMIC ABSORPTION, FLAMELESS	0.000	- 0.500	0.224	0.177 10
	OTHER	1.000	- 1.000	1.000	0.177 11
	OVER-ALL	0.000	- 30.000	0.323	0.291 12
CA	ATOMIC ABSORPTION, DIRECT, AIR	0.160	- 3.000	0.299	0.135 20
	EMISSION, IC PLASMA	0.210	- 0.300	0.250	0.029 10
OVER-ALL	0.160	- 3.000	0.252	0.068 29	
CD	ATOMIC ABSORPTION, DIRECT, AIR	1.000	- 3.000	2.000	1.000
	ATOMIC ABSORPTION, FLAMELESS	0.800	- 2.500	1.041	0.101 13
	EMISSION, IC PLASMA	0.400	- 3.000	1.440	0.966 5
	OVER-ALL	0.400	- 3.000	1.331	0.686 26
CL	COLORIMETRIC, FERRIC THIOCYANATE	1.900	- 31.000	10.378	13.934 4
	ION CHROMATOGRAPHY	0.000	- 14.830	0.297	0.261 3
	ION SELECTIVE ELECTRODE	0.020	- 16.800	0.980	1.427 3
	TITRATION, MERCURIC NITRATE	2.000	- 13.000	2.150	0.397
	TITRATION, SILVER NITRATE	0.000	- 165.000	1.850	1.360
	OVER-ALL	0.000	- 165.000	4.189	5.625 16
	ATOMIC ABSORPTION, DIRECT, AIR	0.900	- 35.000	1.850	1.360
EMISSION, IC PLASMA	0.900	- 35.000	1.950	1.025	
OVER-ALL	0.000	- 35.000	0.413	0.376 4	
CR TOT	ATOMIC ABSORPTION, DIRECT, AIR	-	-	1.950	1.025
	ATOMIC ABSORPTION, FLAMELESS	0.000	- 170.400	1.350	1.324 8
	EMISSION, IC PLASMA	0.300	- 5.000	2.650	1.533
	OVER-ALL	0.000	- 170.400	1.617	1.643 12
CU	ATOMIC ABSORPTION, DIRECT, AIR	4.000	- 28.000	15.400	12.878 4
	ATOMIC ABSORPTION, FLAMELESS	2.000	- 3.000	3.771	0.914 7
	EMISSION, IC PLASMA	2.000	- 20.000	3.350	1.363 4
	OVER-ALL	2.000	- 28.000	3.961	1.110 18
F	ION CHROMATOGRAPHY	0.000	- 0.430	0.240	0.490
	ION SELECTIVE ELECTRODE	0.010	- 0.110	0.050	0.053 3
	OTHER	0.020	- 0.250	0.040	0.020 5
	OVER-ALL	0.000	- 0.430	0.043	0.037 12
FE	ATOMIC ABSORPTION, DIRECT, AIR	5.000	- 86.000	17.300	9.458 5
	ATOMIC ABSORPTION, FLAMELESS	8.300	- 9.400	9.850	0.742
	EMISSION, IC PLASMA	4.000	- 23.300	12.940	8.804 5
	OVER-ALL	0.200	- 195.000	39.444	61.361 15
K	ATOMIC ABSORPTION, DIRECT, AIR	0.020	- 0.200	0.067	0.026 20

Table 14. Statistics by method for standard reference sample P9

Determination	Method	Range:		Mean	Standard Deviation	N
		from	to			
K	OTHER	0.000	1.000	0.438	0.430	5
	OVER-ALL	0.000	1.000	0.060	0.028	25
MG	ATOMIC ABSORPTION, DIRECT, AIR	0.020	0.300	0.031	0.005	14
	EMISSION, IC PLASMA	0.000	0.300	0.030	0.010	8
	OVER-ALL	0.000	0.300	0.031	0.005	23
MN	ATOMIC ABSORPTION, DIRECT, AIR	4.000	28.000	10.700	1.237	7
	ATOMIC ABSORPTION, FLAMELESS	8.000	10.000	8.725	0.877	4
	EMISSION, IC PLASMA	5.300	10.000	7.744	1.575	9
	OVER-ALL	4.000	28.000	9.023	2.460	22
NA	ATOMIC ABSORPTION, DIRECT, AIR	0.040	62.000	0.093	0.023	21
	EMISSION, IC PLASMA	0.050	0.230	0.117	0.075	6
	OVER-ALL	0.040	62.000	0.107	0.043	29
NH3-N	COLORIMETRIC, DISTILLATION, NESSLERIZATION	0.030	0.210	0.210	0.000	2
	COLORIMETRIC, INDOPHENOL	0.020	0.100	0.060	0.040	3
	COLORIMETRIC, PHENATE	0.030	0.430	0.060	0.034	9
	ION SELECTIVE ELECTRODE	0.030	0.120	0.040	0.000	2
	OVER-ALL	0.020	0.430	0.073	0.061	23
	ANODIC STRIPPING VOLTAMMETRY	0.500	13.700	0.750	0.500	1
	ATOMIC ABSORPTION, DIRECT, AIR	30.000	30.000	30.000	0.500	1
PB	ATOMIC ABSORPTION, DIRECT, AIR	1.300	23.000	7.778	7.335	9
	EMISSION, IC PLASMA	3.400	20.000	9.567	9.085	3
	OVER-ALL	0.500	30.000	8.254	8.801	19
	ELECTROMETRIC	1.330	2.570	1.522	0.195	33
	OVER-ALL	1.330	2.570	1.526	0.199	35
SO4	COLORIMETRIC, METHYL THYMOL BLUE	1.000	9.400	4.928	3.123	5
	GRAVIMETRIC, BARIUM SULFATE	0.400	12.300	5.275	5.345	4
	ION CHROMATOGRAPHY	0.400	84.200	0.667	0.379	3
	TURBIDIMETRIC, BARIUM SULFATE	1.000	1.000	1.000	0.379	1
	OVER-ALL	0.000	84.200	3.329	3.734	15
	DIRECT READING INSTRUMENT	110.000	22430.000	15264.000	8258.354	10
NOT REPORTED	160.000	20410.000	19026.667	1764.445	6	
SP. COND.	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1270.000	23000.000	19300.000	1748.313	9
	OVER-ALL	50.000	23000.000	15736.207	7376.409	29
	DIRECT READING INSTRUMENT	110.000	22430.000	15264.000	8258.354	10
TL	ATOMIC ABSORPTION, DIRECT, AIR	1.600	3.900	1.600	29.552	1
	ATOMIC ABSORPTION, FLAMELESS	1.450	3.900	1.525	0.274	1
	OVER-ALL	1.600	3.900	1.600	29.552	1
ZN	ANODIC STRIPPING VOLTAMMETRY	5.000	15.000	10.667	5.132	3
	ATOMIC ABSORPTION, DIRECT, AIR	3.000	60.000	20.083	17.769	12
	EMISSION, IC PLASMA	5.600	62.000	8.867	1.950	6
	OVER-ALL	3.000	76.000	18.168	16.207	25

Table 15 Standard Reference Water Sample AMW2 Report for ACIDDCAC03

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	900	0.0	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
12	850	-5.6	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
21	1000	11.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
23	940	4.4	TITRATION, COLORIMETRIC, MANUAL	1,2,3
45	900	0.0	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
62	630	-30.0	REJECT TITRATION, COLORIMETRIC, MANUAL	1,2,3
71	900	0.0	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
77	870	-3.3	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
92	1250	33.9	REJECT NOT REPORTED	
94	860	-4.4	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
95	910	1.1	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4
107	870	-3.3	TITRATION, ELECTROMETRIC, MANUAL	1,2,3,4

12 Labs had a total range of 630 to 1250 and a mean of 900 with a standard deviation of 44 and a 95% confidence interval of the mean +/- 32.

Table 15 Standard Reference Water Sample AMW2 Report for AG

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1.0		IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,2,4
5	63.0	1532.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
8	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
12	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
21	10.0	140.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
22	0.3	-92.8	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,2,4
23	1.0	-76.0	ATOMIC ABSORPTION, FLAMELESS	3
33		3E+04	REJECT NOT REPORTED	
44	< 0.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
45	< 2.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
57	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
71	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
72	0.7	-83.2	ATOMIC ABSORPTION, FLAMELESS	3
73	< 50.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
75	10.0	140.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
77	3.0	-23.0	ATOMIC ABSORPTION, FLAMELESS	3
91	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
94	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
95	< 23.7		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
99	< 4.0		IGNORED OTHER	
107	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3

21 Labs had a total range of 0.3 to and a mean of 4.17 with a standard deviation of 4.61 and a 95% confidence interval of the mean +/- 4.84.

Table 15 Standard Reference Water Sample AMW2 Report for AL

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	20.1	7.3	EMISSION, DC PLASMA	
5	29.1	55.4	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
8	20.0	6.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
10	22.1	18.0	EMISSION, IC PLASMA	3,5
12	2.1	-38.8	EMISSION, IC PLASMA	3,5
20	0.0	-100.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
22	20.7	10.5	EMISSION, IC PLASMA	3,5
24	21.0	12.1	EMISSION, IC PLASMA	3,5
30	20.6	10.0	EMISSION, IC PLASMA	3,5
33	20.3	8.4	NOT REPORTED	
44	23.0	22.8	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
45	22.0	17.5	NOT REPORTED	
57	20.0	6.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
59	22.1	18.0	EMISSION, IC PLASMA	3,5
65	21.2	13.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
71	20.2	7.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
72	18.9	0.9	EMISSION, IC PLASMA	3,5
73	19.2	2.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
77	21.0	12.1	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
94	16.2	-13.5	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
95	23.7	26.6	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
99	23.6	26.0	ATOMIC ABSORPTION, CHELATION EXTRACTION, NITROUS OXIDE	2,4
107	3.6	-30.8	EMISSION, IC PLASMA	3,5

23 Labs had a total range of 0.0 to 29.1 and a mean of 18.73 with a standard deviation of 7.08 and a 95% confidence interval of the mean +/- 3.06.

Table 15 Standard Reference Water Sample 4M42 Report for 45

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	20	-62.9	ATOMIC ABSORPTION, HYDRIDE, (NABH ₄)	1,4
5	56	3.9	ATOMIC ABSORPTION, HYDRIDE, (NABH ₄)	1
8	40	-25.8	ATOMIC ABSORPTION, HYDRIDE, (NABH ₄)	1,4
10	58	7.6	SPECTROPHOTOMETRIC, SILVER DIETHYL DITHIOCARBAMATE	2,3,4
12	64	18.8	ATOMIC ABSORPTION, FLAMELESS	3
22	2	-96.3	REJECT ATOMIC ABSORPTION, HYDRIDE, (NABH ₄)	1,4
30	67	24.3	ATOMIC ABSORPTION, HYDRIDE, (NABH ₄)	1
33	362	571.8	REJECT NOT REPORTED	
45	51	-5.3	ATOMIC ABSORPTION, HYDRIDE, (NABH ₄)	1,4
57	47	-12.8	ATOMIC ABSORPTION, FLAMELESS	3
62	46	-14.6	ATOMIC ABSORPTION, FLAMELESS	3
65	55	2.1	ATOMIC ABSORPTION, FLAMELESS	3
71	67	24.3	ATOMIC ABSORPTION, FLAMELESS	3
72	64	18.8	ATOMIC ABSORPTION, FLAMELESS	3
73	70	29.9	ATOMIC ABSORPTION, FLAMELESS	3
77	66	22.5	ATOMIC ABSORPTION, HYDRIDE, (NABH ₄)	1
94	65	20.6	ATOMIC ABSORPTION, HYDRIDE, (NABH ₄)	1
95	30	-44.3	ATOMIC ABSORPTION, FLAMELESS	3
107	50	-7.2	ATOMIC ABSORPTION, HYDRIDE, (NABH ₄)	1

19 Labs had a total range of 2 to 362 and a mean of 53.9 with a standard deviation of 14.0 and a 95% confidence interval of the mean +/- 7.2.

Table 15 Standard Reference Water Sample AMW2 Report for 8A

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	<100.0		IGNORED	
3	<100.0		IGNORED	
10	4.0	-73.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
12	60.0	297.6	EMISSION, IC PLASMA	1,2,3,4
19	6.0	-60.2	EMISSION, IC PLASMA	3,5
22	5.4	-64.2	EMISSION, IC PLASMA	3,5
24	< 10.0		EMISSION, IC PLASMA	3,5
30	6.0	-60.2	EMISSION, IC PLASMA	3,5
33	6.6	-56.3	NOT REPORTED	
44	< 5.0		IGNORED	
45	< 10.0		ATOMIC ABSORPTION, FLAMELESS	3
57	53.0	251.2	EMISSION, IC PLASMA	3,5
59	13.0	-13.9	ATOMIC ABSORPTION, FLAMELESS	3
62	< 20.0		EMISSION, IC PLASMA	3,5
65	< 200.0		IGNORED	
71	< 500.0		ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
72	4.8	-63.2	EMISSION, IC PLASMA	3,5
73	< 0.5		IGNORED	
94	1.2	-92.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
95	< 100.0		IGNORED	
99	6.0	-60.2	ATOMIC ABSORPTION, FLAMELESS	3
107	140.0	827.7	EMISSION, IC PLASMA	1,2,3,4
			REJECT	

22 Labs had a total range of 1.2 to 140.0 and a mean of 15.09 with a standard deviation of 20.72 and a 95% confidence interval of the mean +/- 13.92.

Table 15 Standard Reference Water Sample AMW2 Report for BE

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	10.0	-32.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
8	13.0	-11.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
12	390.0	2546.1	EMISSION, IC PLASMA	3,5
19	17.0	15.3	EMISSION, IC PLASMA	3,5
21	18.0	22.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
30	14.0	-5.0	EMISSION, IC PLASMA	3,5
33	30.0	103.5	NOT REPORTED	
45	15.0	1.8	EMISSION, IC PLASMA	3,5
57	13.7	-7.0	ATOMIC ABSORPTION, FLAMELESS	3
59	14.0	-5.0	EMISSION, IC PLASMA	3,5
62	22.0	49.3	ATOMIC ABSORPTION, FLAMELESS	3
71	< 50.0		IGNORED	
72	16.9	14.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
94	14.0	-5.0	EMISSION, IC PLASMA	3,5
95	10.0	-32.2	ATOMIC ABSORPTION, FLAMELESS	3
99	14.0	-5.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
107	< 10.0		EMISSION, IC PLASMA	3,5
			IGNORED	

17 Labs had a total range of 10.0 to 390.0 and a mean of 14.74 with a standard deviation of 3.22 and a 95% confidence interval of the mean +/- 1.95.

Table 15 Standard Reference Water Sample AMW2 Report for CD

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	130	4.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	123	-1.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	137	10.2	ANODIC STRIPPING VOLTAMMETRY, DIFFERENTIAL PULSE	2
10	93	-25.2	ATOMIC ABSORPTION, FLAMELESS	3
12	151	21.4	ATOMIC ABSORPTION, FLAMELESS	3
19	121	-2.7	EMISSION, IC PLASMA	3,5
21	137	10.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	150	20.6	EMISSION, IC PLASMA	3,5
24	120	-3.5	EMISSION, IC PLASMA	3,5
30	136	9.4	EMISSION, IC PLASMA	3,5
33	139	11.8	NOT REPORTED	
45	117	-5.9	EMISSION, IC PLASMA	3,5
56	154	23.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	139	11.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
62	330	165.4	REJECT ATOMIC ABSORPTION, FLAMELESS	3
65	128	2.9	ATOMIC ABSORPTION, FLAMELESS	3
71	120	-3.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	68	-45.3	EMISSION, IC PLASMA	3,5
73	130	4.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	70	-43.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	110	-11.5	ATOMIC ABSORPTION, FLAMELESS	3
91	140	12.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	125	0.5	ATOMIC ABSORPTION, FLAMELESS	3
95	140	12.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	131	5.3	EMISSION, IC PLASMA	3,5
107	100	-19.6	EMISSION, IC PLASMA	3,5

26 Labs had a total range of 68 to 330 and a mean of 124.4 with a standard deviation of 22.1 and a 95% confidence interval of the mean +/- 9.1.

Table 15 Standard Reference Water Sample AMW2 Report for CL

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	2.2	-79.0	NOT REPORTED	
5	0.5	-95.2	ION SELECTIVE ELECTRODE	1,2,3,4
9	7.3	-30.3	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
10	10.0	-4.5	TITRATION, SILVER NITRATE	1,2,4
12	7.8	-25.5	ION CHROMATOGRAPHY	2,3,6
21	10.9	4.1	TITRATION, MERCURIC NITRATE	1,2,3,4
24	5.3	-44.6	ION SELECTIVE ELECTRODE	1,2,3,4
30	3.6	-17.9	ION CHROMATOGRAPHY	2,3,6
44	8.8	-16.0	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
45	6.9	-34.1	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
56	23.0	119.6	TITRATION, MERCURIC NITRATE	1,2,3,4
59	7.8	-25.5	ION CHROMATOGRAPHY	2,3,6
65	7.0	-33.2	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
72	7.0	-33.2	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
77	0.1	-99.0	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
91	11.0	5.0	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
92	27.1	153.7	TITRATION, MERCURIC NITRATE	1,2,3,4
94	29.5	131.6	ION SELECTIVE ELECTRODE	1,2,3,4
95	8.0	-23.6	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
107	20.2	92.8	TITRATION, SILVER NITRATE	1,2,4

20 Labs had a total range of 0.1 to 29.5 and a mean of 10.43 with a standard deviation of 3.15 and a 95% confidence interval of the mean +/- 3.21.

Table 15 Standard Reference Water Sample AMW2 Report for CO

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	125	-12.5	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
8	130	-9.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	136	-4.8	EMISSION, IC PLASMA	3,5
21	181	26.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	120	-16.0	EMISSION, IC PLASMA	3,5
33	167	16.9	NOT REPORTED	
56	190	33.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	150	5.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	140	-2.0	EMISSION, IC PLASMA	3,5
77	110	-23.0	ATOMIC ABSORPTION, FLAMELESS	3
95	140	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	138	-3.4	EMISSION, IC PLASMA	3,5
107	130	-9.0	EMISSION, IC PLASMA	3,5

13 Labs had a total range of 110 to 190 and a mean of 142.3 with a standard deviation of 23.5 and a 95% confidence interval of the mean +/- 14.2.

Table 15 Standard Reference Water Sample AM2 Report for CR TOT

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	40	93.7	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,3,4
5	12	-41.9	ATOMIC ABSORPTION, FLAMELESS	3
8	<		IGNORED	1,2,3,4
9	<		IGNORED	1,2,3,4
10	10	-51.6	ATOMIC ABSORPTION, FLAMELESS	3
12	3	-61.3	ATOMIC ABSORPTION, FLAMELESS	3
21	47	127.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	28	35.6	ATOMIC ABSORPTION, FLAMELESS	3
30	20	-3.1	EMISSION, IC PLASMA	3
33	116	461.7	EMISSION, IC PLASMA	3
44	6	-70.9	NOT REPORTED	3
45	3	-35.5	ATOMIC ABSORPTION, FLAMELESS	3
56	46	122.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	16	-22.5	ATOMIC ABSORPTION, FLAMELESS	3
62	15	-27.4	ATOMIC ABSORPTION, FLAMELESS	3
65	30		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	20	-3.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	20	-3.1	EMISSION, IC PLASMA	3
73	50		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	10	-51.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	8	-61.3	ATOMIC ABSORPTION, FLAMELESS	3
91	35	69.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	9	-56.4	ATOMIC ABSORPTION, FLAMELESS	3
95	<		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	20	-3.1	EMISSION, IC PLASMA	3
107	40	93.7	EMISSION, IC PLASMA	3

26 Labs had a total range of 3 to 116 and a mean of 20.7 with a standard deviation of 13.9 and a 95% confidence interval of the mean +/- 6.5.

Table 15 Standard Reference Water Sample AMW2 Report for CU

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	5350	4.6	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
8	5100	-0.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	5130	0.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	5060	-1.0	EMISSION, IC PLASMA	3,5
12	5500	7.6	EMISSION, IC PLASMA	3,5
19	5230	2.3	EMISSION, IC PLASMA	3,5
21	4940	-3.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	5100	-0.2	EMISSION, IC PLASMA	3,5
24	4900	-4.2	EMISSION, IC PLASMA	3,5
30	5220	2.1	EMISSION, IC PLASMA	3,5
33	4950	-3.2	NOT REPORTED	
45	5450	6.6	EMISSION, IC PLASMA	3,5
56	4790	-6.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	5000	-2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	5100	-0.2	EMISSION, IC PLASMA	3,5
62	4950	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	5290	3.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	5220	2.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	4570	-10.6	EMISSION, IC PLASMA	3,5
73	6180	20.9	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	2300	-45.2	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	5100	-0.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	5000	-2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	5130	0.3	ATOMIC ABSORPTION, FLAMELESS	3
95	510	-90.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	5180	1.3	EMISSION, IC PLASMA	3,5
107	5440	6.4	EMISSION, IC PLASMA	3,5

27 Labs had a total range of 510 to 6130 and a mean of 5113 with a standard deviation of 214 and a 95% confidence interval of the mean +/- 90.

Table 15 Standard Reference Water Sample AMW2 Report for F

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	2.7	8.0	ION SELECTIVE ELECTRODE	1,2,3,4
5	1.4	-44.0	OTHER	
8	3.9	56.0	OTHER	
10	3.6	44.0	OTHER	
12	4.5	30.0	ION CHROMATOGRAPHY	2,3,6
23	1.7	-32.0	COLORIMETRIC, ZIRCONYL ALIZARIN	1
30	2.6	4.0	ION CHROMATOGRAPHY	2,3,6
44	0.5	-80.0	ION SELECTIVE ELECTRODE	1,2,3,4
59	4.6	84.0	ION CHROMATOGRAPHY	2,3,6
71	< 0.1		IGNORED OTHER	
72	1.7	-32.0	ION SELECTIVE ELECTRODE	1,2,3,4
77	2.4	-4.0	ION SELECTIVE ELECTRODE	1,2,3,4
94	0.3	-88.0	ION SELECTIVE ELECTRODE	1,2,3,4
95	4.4	76.0	COLORIMETRIC, CEROUS ALIZARIN "COMPLEXONE"	3
107	0.7	-72.0	ION SELECTIVE ELECTRODE	1,2,3,4

15 Labs had a total range of 0.3 to 4.6 and a mean of 2.50 with a standard deviation of 1.51 and a 95% confidence interval of the mean +/- 0.87.

Table 15 Standard Reference Water Sample AMW2 Report for FE

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	119.0	-16.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	135.0	-5.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	125.4	-11.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	148.7	4.7	EMISSION, IC PLASMA	3,5
12	150.0	5.6	EMISSION, IC PLASMA	3,5
19	143.4	1.0	EMISSION, IC PLASMA	3,5
20	130.0	-8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
21	149.4	5.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	149.0	4.9	EMISSION, IC PLASMA	3,5
23	113.0	-20.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
24	140.0	-1.4	EMISSION, IC PLASMA	3,5
30	148.0	4.2	EMISSION, IC PLASMA	3,5
33	138.7	-2.3	NOT REPORTED	
44	15.0	-89.4	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
45	157.0	10.5	EMISSION, IC PLASMA	3,5
56	150.0	5.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	135.0	-5.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	150.0	5.6	EMISSION, IC PLASMA	3,5
62	190.0	33.8	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	153.0	7.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	154.0	8.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	140.0	-1.4	EMISSION, IC PLASMA	3,5
73	135.3	-4.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	69.0	-51.4	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	150.0	5.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	127.0	-10.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	147.0	3.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	154.0	8.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	151.0	6.3	EMISSION, IC PLASMA	3,5
107	1.6	-98.9	REJECT EMISSION, IC PLASMA	3,5

30 Labs had a total range of 1.6 to 190.0 and a mean of 115.59 and a 95% confidence interval of the mean +/- 4.68.

Table 15 Standard Reference Water Sample AMW2 Report for HG

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.10	-91.2	EMISSION, IC PLASMA	3
12	2.00	75.1	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
23	0.03	-93.0	EMISSION, IC PLASMA	3
30	0.12	-89.5	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
44	< 0.20		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
45	1.70	43.9	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
56	2.80	145.2	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
57	< 0.20		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
62	1.60	40.4	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
71	< 2.00		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
73	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
77	0.20	-82.5	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
94	0.30	-73.7	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
95	2.52	120.7	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
99	< 2.00		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
107	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4

16 Labs had a total range of 0.03 to 2.80 and a mean of 1.142 with a standard deviation of 1.093 and a 95% confidence interval of the mean +/- 0.782.

Table 15 Standard Reference Water Sample AMW2 Report for LI

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	35	-30.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
19	61	21.3	EMISSION, IC PLASMA	3,5
21	40	-20.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
22	62	23.3	EMISSION, IC PLASMA	3,5
30	46	-8.5	EMISSION, IC PLASMA	3,5
33	35	-30.4	NOT REPORTED	
59	48	-4.6	EMISSION, IC PLASMA	3,5
71	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
72	50	-0.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
95	80	59.0	EMISSION, FLAME	1
99	46	-8.5	EMISSION, IC PLASMA	3,5

11 Labs had a total range of 35 to 80 and a mean of 50.3 with a standard deviation of 13.9 and a 95% confidence interval of the mean +/- 10.0.

Table 15 Standard Reference Water Sample AMW2 Report for MN

Code Number	Reported value	Pct. dev. from mean		Methods	References
1	86.3	-2.8		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
5	85.1	-4.2		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	110.0	23.9	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	82.0	-7.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	96.4	3.6		EMISSION, IC PLASMA	3,5
12	94.0	5.9		EMISSION, IC PLASMA	3,5
19	85.9	-3.3		EMISSION, IC PLASMA	3,5
21	87.3	-1.7		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	88.1	-0.8		EMISSION, IC PLASMA	3,5
24	85.0	-4.3		EMISSION, IC PLASMA	3,5
30	92.8	4.5		EMISSION, IC PLASMA	3,5
33	83.9	-5.5		NOT REPORTED	
44	7.4	-91.7	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
45	95.5	7.5		EMISSION, IC PLASMA	3,5
56	85.0	-4.3		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	87.0	-2.0		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	92.2	3.8		EMISSION, IC PLASMA	3,5
62	83.0	-6.5		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	93.3	5.1		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	89.0	0.2		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	84.2	-5.2		EMISSION, IC PLASMA	3,5
73	43.2	-51.4	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	49.0	-44.8	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	92.0	3.6		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	115.0	29.5	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	9.0	-89.9	REJECT	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	87.6	-1.4		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	94.1	6.0		EMISSION, IC PLASMA	3,5
107	92.8	4.5		EMISSION, IC PLASMA	3,5

29 Labs had a total range of 7.4 to 115.0 and a mean of 88.80 with a standard deviation of 4.38 and a 95% confidence interval of the mean +/- 1.89.

Table 15 Standard Reference Water Sample Aww2 Report for MO

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1	-98.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
8	<		ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3
12	180	165.5	EMISSION, IC PLASMA	3,5
19	<		EMISSION, IC PLASMA	3,5
30	30	-55.8	EMISSION, IC PLASMA	3,5
33	70	3.2	NOT REPORTED	
71	<		ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3
72	<		ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3
95	<		ATOMIC ABSORPTION, FLAMELESS	3
99	53	-14.5	EMISSION, IC PLASMA	3,5
107	<		EMISSION, IC PLASMA	3,5

11 Labs had a total range of 1 to 180 and a mean of 67.3
with a standard deviation of 65.2 and a 95% confidence interval of the mean +/- 84.9.

Table 15 Standard Reference Water Sample Aww2 Report for NI

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	60	-75.5	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
8	210	-14.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	260	6.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	290	18.4	ATOMIC ABSORPTION, DIRECT, AIR	3
12	220	-10.2	ATOMIC ABSORPTION, FLAMELESS	3
21	200	-18.4	ATOMIC ABSORPTION, FLAMELESS	
22	210	-14.3	NOT REPORTED	
30	260	6.1	OTHER	
33	280	14.3	NOT REPORTED	
45	260	6.1	OTHER	
56	660	169.4	NOT REPORTED	
57	270	10.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
62	270	10.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	280	14.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	210	-14.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	270	10.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	600	144.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	150	-38.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	250	2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	400	63.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	210	-14.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	250	2.0	ATOMIC ABSORPTION, FLAMELESS	3
99	310	26.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
107	240	-2.0	OTHER	
			ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4

24 Labs had a total range of 60 to 660 and a mean of 245
with a standard deviation of 38 and a 95% confidence interval of the mean +/- 18.

Table 15 Standard Reference Water Sample AMW2 Report for N03-V

Code Number	Reported value	Pct. dev. from mean	Methods	References
10	4.20	252.1	COLORIMETRIC, BRUCINE	1,2,3,4
12	< 0.01		IGNORED	1,2,3,4
21	0.15	-87.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
23	0.00	-100.0	COLORIMETRIC, BRUCINE	1,2,3,4
24	< 0.10		IGNORED	
30	2.30	92.8	ION CHROMATOGRAPHY	2,3,6,7
44	< 0.01		IGNORED	1,2,3,4
45	< 0.10		IGNORED	1,2,3,4
56	0.20	-33.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
57	0.25	-77.0	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
72	0.04	-25.6	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
77	0.00	-100.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
92	0.24	-77.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
99	< 0.05		IGNORED	1,2,3,4
107	4.55	281.4	COLORIMETRIC, BRUCINE	1,2,3,4

15 Labs had a total range of 0.00 to 4.55 and a mean of 1.193 with a standard deviation of 1.813 and a 95% confidence interval of the mean +/- 1.297.

Table 15 Standard Reference Water Sample AMW2 Report for PB

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	39	-35.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
5	130	113.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	27	-55.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	317	420.6	REJECT	2
10	36	41.2	ANODIC STRIPPING VOLTAMMETRY	3
12	38	-37.6	ATOMIC ABSORPTION, FLAMELESS	3
19	< 20		IGNORED	3,5
21	102	67.5	EMISSION, IC PLASMA	1,2,3,4
23	12	-30.3	ATOMIC ABSORPTION, DIRECT, AIR	1,4
33	189	210.4	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	
45	43	-29.4	NOT REPORTED	1,4
57	40	-34.3	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,2,3,4
62	35	-42.5	ATOMIC ABSORPTION, DIRECT, AIR	3
65	6	-90.1	ATOMIC ABSORPTION, FLAMELESS	3
71	20	-67.2	ATOMIC ABSORPTION, FLAMELESS	1,2,3,4
72	94	54.4	ATOMIC ABSORPTION, DIRECT, AIR	3,5
73	< 500		IGNORED	1,2,3,4
75	150	145.4	ATOMIC ABSORPTION, DIRECT, AIR	2
77	5	-91.8	ANODIC STRIPPING VOLTAMMETRY	3
91	< 200		IGNORED	1,2,3,4
94	30	-50.7	ATOMIC ABSORPTION, FLAMELESS	3
95	50	-17.9	ATOMIC ABSORPTION, DIRECT, AIR	3
99	415	581.6	REJECT	3,5
107	< 50		IGNORED	3,5

24 Labs had a total range of 5 to 415 and a mean of 60.9 with a standard deviation of 52.3 and a 95% confidence interval of the mean +/- 26.3.

Table 15 Standard Reference Water Sample AMW3 Report for PH

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	2.41	-3.4	ELECTROMETRIC	1,2,3,4
5	2.65	6.3	ELECTROMETRIC	1,2,3,4
6	2.50	0.2	ELECTROMETRIC	1,2,3,4
10	2.27	-9.0	ELECTROMETRIC	1,2,3,4
12	2.50	0.2	ELECTROMETRIC	1,2,3,4
21	2.56	2.7	ELECTROMETRIC	1,2,3,4
23	2.50	0.2	ELECTROMETRIC	1,2,3,4
30	2.44	-2.2	ELECTROMETRIC	1,2,3,4
31	2.45	-1.8	ELECTROMETRIC	1,2,3,4
44	2.40	-3.8	ELECTROMETRIC	1,2,3,4
45	2.50	0.2	ELECTROMETRIC	1,2,3,4
56	2.52	1.1	ELECTROMETRIC	1,2,3,4
57	2.41	-3.4	ELECTROMETRIC	1,2,3,4
59	2.50	0.2	ELECTROMETRIC	1,2,3,4
62	2.64	5.9	ELECTROMETRIC	1,2,3,4
65	2.64	5.9	OTHER	
71	2.50	0.2	ELECTROMETRIC	1,2,3,4
72	2.43	-2.6	ELECTROMETRIC	1,2,3,4
75	2.67	7.1	NOT REPORTED	
77	2.40	-3.8	ELECTROMETRIC	1,2,3,4
91	2.60	4.3	ELECTROMETRIC	1,2,3,4
92	2.63	5.5	ELECTROMETRIC	1,2,3,4
94	2.48	-0.6	ELECTROMETRIC	1,2,3,4
95	2.45	-1.8	ELECTROMETRIC	1,2,3,4
99	2.39	-4.2	ELECTROMETRIC	1,2,3,4
107	2.40	-3.8	ELECTROMETRIC	1,2,3,4

26 Labs had a total range of 2.27 to 2.57 and a mean of 2.494 with a standard deviation of 0.099 and a 95% confidence interval of the mean +/- 0.040.

Table 15 Standard Reference Water Sample AMW2 Report for SE

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1.0		IGNORED ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
8	< 1.0		IGNORED ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
12	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
22	22.1	222.2	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
30	0.1	-98.5	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
33	509.0	7319.8	REJECT NOT REPORTED	
45	< 10.0		IGNORED ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
57	< 2.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
62	1.0	-35.4	ATOMIC ABSORPTION, FLAMELESS	3
65	< 2.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
72	10.3	57.4	ATOMIC ABSORPTION, FLAMELESS	3
77	0.3	-95.6	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
94	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
95	< 10.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
107	< 1.0		IGNORED ATOMIC ABSORPTION, HYDRIDE	1,2,3,4

15 Labs had a total range of 0.1 to 509.0 and a mean of 6.85 with a standard deviation of 9.63 and a 95% confidence interval of the mean +/- 11.95.

Table 15 Standard Reference Water Sample AMW2 Report for SI02

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	37	-20.9	OTHER	
8	53	13.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
10	47	0.5	EMISSION, IC PLASMA	5
12	50	7.0	EMISSION, IC PLASMA	5
19	46	-1.6	EMISSION, IC PLASMA	5
22	49	4.3	EMISSION, IC PLASMA	5
23	32	-31.6	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
24	42	-10.2	EMISSION, IC PLASMA	5
30	46	-1.6	EMISSION, IC PLASMA	5
33	45	-3.7	NOT REPORTED	
44	60	28.3	COLORIMETRIC, MOLYBDOUSILICIC ACID	1,2,3
45	55	17.6	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
59	46	-1.6	EMISSION, IC PLASMA	5
71	53	13.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
72	47	0.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
77	44	-5.9	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3
94	45	-3.7	COLORIMETRIC, MOLYBDOUSILICIC ACID	1,2,3
95	45	-3.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
99	48	2.7	EMISSION, IC PLASMA	5
107	45	-3.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4

20 Labs had a total range of 32 to 60 and a mean of 46.3 with a standard deviation of 6.1 and a 95% confidence interval of the mean +/- 2.8.

Table 15 Standard Reference Water Sample AMW2 Report for 504

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	2100	3.9	NOT REPORTED	
8	1350	-4.0	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
9	1720	-10.3	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
10	2000	3.8	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
12	1960	1.7	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
21	1700	-11.8	GRAVIMETRIC, BARIUM SULFATE	1,2,3
23	2040	5.8	COLORIMETRIC, CHLORANILATE	3
24	2050	6.3	GRAVIMETRIC, BARIUM SULFATE	1,2,3
30	2000	3.8	ION CHROMATOGRAPHY	2,5
44	2500	29.7	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
45	1700	-11.3	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
57	2050	6.3	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
59	2070	7.4	ION CHROMATOGRAPHY	2,5
65	1890	-2.0	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
72	1540	-20.1	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
75	1400	-27.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
77	2000	3.8	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
91	1900	-1.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
92	20	-99.0	REJECT TURBIDIMETRIC, BARIUM SULFATE	1,2,3
94	1330	-5.1	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
95	2160	12.1	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
107	2020	4.8	GRAVIMETRIC, BARIUM SULFATE	1,2,3

22 Labs had a total range of 20 to 2500 and a mean of 1928 with a standard deviation of 234 and a 95% confidence interval of the mean +/- 106.

Table 15 Standard Reference Water Sample AMW2 Report for SP. COND.

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	3910	0.6	NOT REPORTED	
5	9240	137.7	REJECT ELECTRODELESS, INDUCTIVE CELL-TYPE	2
8	4350	4.2	DIRECT READING INSTRUMENT	4
10	4160	7.0	DIRECT READING INSTRUMENT	4
12	3920	0.8	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
21	3670	-5.6	DIRECT READING INSTRUMENT	4
23	6100	56.9	REJECT NOT REPORTED	
30	4180	7.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
44	4000	2.9	DIRECT READING INSTRUMENT	4
45	4050	4.2	DIRECT READING INSTRUMENT	4
65	3520	-9.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
71	3590	-7.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
77	5370	63.8	REJECT WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
91	3480	-10.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
92	3850	-1.0	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
95	4050	4.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
107	4000	2.9	DIRECT READING INSTRUMENT	4

17 Labs had a total range of 3480 to 9240 and a mean of 3828 with a standard deviation of 233 and a 95% confidence interval of the mean +/- 134.

Table 15 Standard Reference Water Sample AMW2 Report for SR

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1500	-4.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
3	1600	1.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
22	1500	-4.6	EMISSION, IC PLASMA	3,5
30	1530	-2.7	EMISSION, IC PLASMA	3,5
33	1570	-0.1	NOT REPORTED	
44	1100	-30.0	REJECT OTHER	
45	1600	1.8	EMISSION, IC PLASMA	3,5
59	1500	-4.6	EMISSION, IC PLASMA	3,5
72	1700	8.1	EMISSION, IC PLASMA	3,5
95	1610	2.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
99	1610	2.4	EMISSION, IC PLASMA	3,5

11 Labs had a total range of 1100 to 1700 and a mean of 1572 with a standard deviation of 65 and a 95% confidence interval of the mean +/- 47.

Table 15 Standard Reference Water Sample AMW2 Report for ZN

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	43.6	12.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
5	0.9	-97.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	44.0	13.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	45.4	16.7	ANODIC STRIPPING VOLTAMMETRY	
10	46.6	19.8	EMISSION, IC PLASMA	3,5
12	47.0	20.8	EMISSION, IC PLASMA	3,5
21	44.1	13.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	44.0	13.1	EMISSION, IC PLASMA	3,5
24	42.0	8.0	EMISSION, IC PLASMA	3,5
30	45.4	16.7	EMISSION, IC PLASMA	3,5
33	41.8	7.4	NOT REPORTED	
44	4.4	-88.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
45	45.9	18.0	EMISSION, IC PLASMA	3,5
56	38.5	-1.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	44.0	13.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	45.3	16.4	EMISSION, IC PLASMA	3,5
62	47.0	20.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
65	44.5	14.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	46.8	20.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	40.7	4.6	EMISSION, IC PLASMA	3,5
73	30.5	-21.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
75	24.0	-33.3	ANODIC STRIPPING VOLTAMMETRY	
77	4.2	-89.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	46.5	19.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	43.9	12.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	43.1	10.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
99	46.8	20.3	EMISSION, IC PLASMA	3,5
107	48.4	24.4	EMISSION, IC PLASMA	3,5

28 Labs had a total range of 0.9 to 48.4 and a mean of 38.90 with a standard deviation of 13.60 and a 95% confidence interval of the mean +/- 5.27.

Table 15. Statistics by method for standard reference sample AMW2

Determination	Method	Range: from to	Mean	Standard Deviation	N
ACID3CAC03	TITRATION, ELECTROMETRIC, MANUAL	850.000 - 1000.000	882.500	22.520	8
	OVER-ALL	630.000 - 1250.000	900.000	44.222	10
AG	ATOMIC ABSORPTION, DIRECT, AIR	10.000 - 68.000	10.000	0.000	2
	ATOMIC ABSORPTION, FLAMELESS	0.700 - 3.000	1.567	1.250	3
AL	_OVER-ALL_	0.300 - 1267.000	4.167	4.613	6
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	0.000 - 21.200	20.120	0.716	5
AS	ATOMIC ABSORPTION, DIRECT, FLAMELESS	16.200 - 29.100	22.600	4.667	5
	EMISSION, IC PLASMA	2.100 - 22.100	16.388	8.425	8
BA	_OVER-ALL_	0.000 - 29.100	19.726	7.084	23
	ATOMIC ABSORPTION, FLAMELESS	30.000 - 70.000	55.375	13.627	8
BE	ATOMIC ABSORPTION, HYDRIDE, (NASH4)	2.000 - 51.000	28.250	21.701	4
	ATOMIC ABSORPTION, HYDRIDE, (NASH4)	50.000 - 67.000	60.800	7.493	5
BO	_OVER-ALL_	2.000 - 362.000	53.382	14.032	17
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	-	60.800	3.286	
BR	ATOMIC ABSORPTION, DIRECT, FLAMELESS	1.200 - 53.000	27.100	5.029	
	EMISSION, IC PLASMA	4.000 - 140.000	5.367	0.924	6
BS	_OVER-ALL_	1.200 - 140.000	15.091	20.722	11
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	10.000 - 18.000	12.750	3.775	4
BT	ATOMIC ABSORPTION, DIRECT, FLAMELESS	13.700 - 22.000	13.850	0.387	
	EMISSION, IC PLASMA	14.000 - 390.000	15.150	1.447	6
BU	_OVER-ALL_	10.000 - 390.000	14.739	3.220	13
	ATOMIC ABSORPTION, DIRECT, AIR	70.000 - 154.000	134.778	10.305	9
BV	ATOMIC ABSORPTION, FLAMELESS	93.000 - 330.000	121.400	21.617	5
	EMISSION, IC PLASMA	68.000 - 150.000	117.875	24.942	8
BW	_OVER-ALL_	68.000 - 330.000	124.360	22.136	25
	COLORIMETRIC, FERRIC THIOCYANATE	0.100 - 11.000	7.240	0.451	5
BX	ION CHROMATOGRAPHY	7.800 - 8.600	7.800	0.000	2
	ION SELECTIVE ELECTRODE	0.500 - 29.500	11.933	15.442	3
BY	TITRATION, MERCURIC NITRATE	10.900 - 27.100	20.333	8.423	3
	OVER-ALL	0.100 - 29.500	13.475	8.149	20
C0	ATOMIC ABSORPTION, DIRECT, AIR	130.000 - 190.000	158.200	26.100	5
	EMISSION, IC PLASMA	120.000 - 140.000	132.800	9.075	5
C1	_OVER-ALL_	110.000 - 190.000	142.846	23.540	13
	ATOMIC ABSORPTION, DIRECT, AIR	10.000 - 47.000	31.600	16.257	5
C2	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3.000 - 15.000	9.667	4.153	9
	EMISSION, IC PLASMA	20.000 - 40.000	25.600	8.764	5
C3	_OVER-ALL_	3.000 - 116.000	20.650	13.926	20
	ATOMIC ABSORPTION, DIRECT, AIR	510.000 - 6180.000	5052.000	146.272	10
C4	EMISSION, IC PLASMA	4570.000 - 5500.000	5159.091	268.308	11

Table 1b. Statistics by method for standard reference sample 1MW2

Determination	Method	Range: from to	Mean	Standard Deviation	N
CU	_OVER-ALL_	510.000 - 5130.000	5112.500	214.278	24
F	ION CHROMATOGRAPHY	2.600 - 4.600	4.550	0.224	
	ION SELECTIVE ELECTRODE	0.300 - 2.700	1.383	1.028	6
	OTHER	1.400 - 3.900	2.967	1.365	3
	OVER-ALL	0.300 - 4.600	2.500	1.515	14
FE	ATOMIC ABSORPTION, DIRECT, AIR EMISSION, IC PLASMA	15.000 - 190.000	138.473	13.595	15
	OVER-ALL	1.600 - 157.000	147.710	5.247	10
		1.600 - 190.000	142.035	11.591	26
HG	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	0.120 - 2.800	1.323	1.251	6
	OVER-ALL	0.080 - 2.800	1.142	1.093	10
LI	ATOMIC ABSORPTION, DIRECT, AIR EMISSION, IC PLASMA	35.000 - 50.000	41.667	7.638	3
	OVER-ALL	45.000 - 62.000	52.600	8.173	5
		35.000 - 80.000	50.300	13.945	10
MN	ATOMIC ABSORPTION, DIRECT, AIR EMISSION, IC PLASMA	7.400 - 115.000	75.953	30.760	17
	OVER-ALL	84.200 - 95.400	91.000	4.389	11
		7.400 - 115.000	83.904	4.373	23
MO	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE EMISSION, IC PLASMA	30.000 - 130.000	91.000	1.342	
	OVER-ALL	1.000 - 130.000	89.333	79.753	3
			67.800	68.156	5
NI	ATOMIC ABSORPTION, DIRECT, AIR EMISSION, FLAMELESS	150.000 - 600.000	251.000	24.698	10
	NOT REPORTED	210.000 - 290.000	240.000	43.589	3
	OTHER	200.000 - 660.000	380.000	245.764	3
	OVER-ALL	210.000 - 310.000	260.000	40.825	4
		60.000 - 660.000	245.000	38.320	20
NO3-N	COLORIMETRIC, BRUCINE	0.000 - 4.550	4.375	0.418	
	OVER-ALL	0.000 - 0.240	0.126	0.103	5
		0.000 - 4.550	1.193	1.813	10
PB	ATOMIC ABSORPTION, DIRECT, AIR EMISSION, FLAMELESS	20.000 - 130.000	59.667	45.152	6
	OVER-ALL	5.000 - 86.000	35.714	27.681	7
		94.000 - 415.000	254.500	12.669	
		5.000 - 415.000	60.889	52.801	18
PH	ELECTROMETRIC	2.270 - 2.650	2.480	0.091	24
	OVER-ALL	2.270 - 2.670	2.494	0.099	26
SE	ATOMIC ABSORPTION, FLAMELESS	1.000 - 10.800	5.900	2.214	
	OVER-ALL	0.100 - 22.100	0.200	0.316	
		0.100 - 509.000	6.860	9.629	5
SiO2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE EMISSION, IC PLASMA	45.000 - 53.000	43.600	4.099	5
		42.000 - 50.000	46.750	2.435	8

Table 15. Statistics by method for standard reference sample 6M2

Determination	Method	from	to	Range:	Mean	Standard Deviation	N
SiO2	_OVER-ALL_	52.000	-	60.000	46.750	6.051	20
	COLORIMETRIC, METHYL THYMOL BLUE	1400.000	-	2160.000	1767.500	234.262	8
	GRAVIMETRIC, BARIUM SULFATE	1700.000	-	2050.000	2035.000	3.873	5
	TURBIDIMETRIC, BARIUM SULFATE	20.000	-	2500.000	1972.000	75.299	5
	OVER-ALL	20.000	-	2500.000	1927.519	233.729	21
SP. COND.	DIRECT READING INSTRUMENT	3670.000	-	4150.000	4052.000	65.345	5
	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	3450.000	-	6370.000	3798.571	273.461	7
	OVER-ALL	3450.000	-	9240.000	3597.857	232.749	14
	ATOMIC ABSORPTION, DIRECT, AIR	1500.000	-	1610.000	1605.000	2.236	6
	EMISSION, IC PLASMA	1500.000	-	1700.000	1573.333	78.401	6
ZN	_OVER-ALL_	1100.000	-	1700.000	1572.000	65.115	10
	ATOMIC ABSORPTION, DIRECT, AIR	0.900	-	47.000	35.057	17.008	15
	EMISSION, IC PLASMA	40.700	-	48.400	45.219	2.368	10
	OVER-ALL	0.900	-	48.400	38.904	13.596	28