

**RESULTS OF THE U.S. GEOLOGICAL SURVEY'S ANALYTICAL
EVALUATION PROGRAM FOR STANDARD REFERENCE SAMPLES
DISTRIBUTED IN OCTOBER 1999**

U.S. GEOLOGICAL SURVEY

Open-File Report 00-227



**RESULTS OF THE U.S. GEOLOGICAL SURVEY'S ANALYTICAL
EVALUATION PROGRAM FOR STANDARD REFERENCE SAMPLES
DISTRIBUTED IN OCTOBER 1999**

By Jerry W. Farrar

U.S. GEOLOGICAL SURVEY

Open-File Report 00-227

**Lakewood, Colorado
1999**

DEPARTMENT OF THE INTERIOR

BRUCE BABBITT, Secretary

U.S. GEOLOGICAL SURVEY

Charles G. Groat, Director

For additional information
write to:
Chief, Branch of Quality Systems
U.S. Geological Survey
Box 25046, Mail Stop 401
Denver Federal Center
Denver, Colorado 80225-0046

Copies of this report can be
purchased from:
U.S. Geological Survey
Branch of Information Services
Box 25286
Denver, Colorado 80225-0286

RESULTS OF THE U.S. GEOLOGICAL SURVEY'S ANALYTICAL EVALUATION PROGRAM FOR STANDARD REFERENCE SAMPLES DISTRIBUTED IN OCTOBER 1999

By Jerry W. Farrar

ABSTRACT

This report presents the results of the U.S. Geological Survey's analytical evaluation program for six standard reference samples -- T-159 (trace constituents), M-152 (major constituents), N-63 (nutrient constituents), N-64 (nutrient constituents), P-33 (low ionic strength constituents), and Hg-29 (mercury) -- that were distributed in October 1999 to 149 laboratories enrolled in the U.S. Geological Survey sponsored interlaboratory testing program. Analytical data that were received from 131 of the laboratories were evaluated with respect to overall laboratory performance and relative laboratory performance for each analyte in the six reference samples. Results of these evaluations are presented in tabular form. Also presented are tables and graphs summarizing the analytical data provided by each laboratory for each analyte in the six standard reference samples. The most probable value for each analyte was determined using nonparametric statistics.

INTRODUCTION

The U.S. Geological Survey (USGS) conducts an interlaboratory analytical evaluation program semiannually. This program provides a variety of standard reference samples (SRSs) to accomplish quality assurance testing of laboratories and to provide an adequate supply of samples that contribute to quality control programs of participating laboratories. Natural-matrix reference materials are preferred for use in this interlaboratory evaluation program. A series of samples are prepared and distributed each spring and fall.

The program began in 1962 with a single sample containing major constituents that was prepared from distilled water and reagent grade chemicals. Twenty-three USGS laboratories participated in the first analytical evaluation program. Since that time, objectives of the program have been to:

- (1) evaluate and improve the performance of USGS and other participating laboratories;
- (2) provide a library of carefully prepared, homogeneous, stable reference materials for use in the quality control programs of laboratories;
- (3) identify analytical problem areas;
- (4) identify quality assurance needs with respect to environmental analyses and develop new reference materials to meet these needs; and
- (5) evaluate the accuracy and precision of analytical methods.

A total of 195 USGS and non-USGS laboratories are enrolled in the program, which can currently provide 9 different types of SRSs:

1. Trace constituents.
2. Major constituents.
3. Nutrient constituents.
4. Low ionic-strength constituents.
5. Mercury.
6. Whole water (water with suspended sediment).
7. Acid mine water constituents.
8. Ground-water trace constituents.
9. Ground-water major constituents.

Though this is not a laboratory certification program, participation in this continuing quality assurance program is mandatory for all laboratories providing water-quality data for USGS sponsored reports or storage in the USGS national data bases. Federal, State, Municipal, and University laboratories can participate even though they do not provide data to the USGS. The results from this study can be used to alert participating laboratories of possible deficiencies in their analytical operations and provide reference materials for laboratory quality-control programs. Participating laboratories are identified only by a confidential laboratory code number.

A library of SRSs, from previous evaluations, is available. USGS offices and participating laboratories can request these SRSs for further testing, continuing quality assurance, and quality-control programs by contacting:

U.S. Geological Survey
Branch of Quality Systems
Denver Federal Center, Bldg. 53
P. O. Box 25046 MS 401
Denver, Colorado 80225-0046
(303) 236-1876

This report summarizes the analytical results submitted by 131 of the 149 laboratories that requested and were shipped SRSs for the October 1999 evaluation (table 1). Not all SRSs are requested or necessarily analyzed by all the laboratories; nor do all laboratories enrolled in the program participate in each evaluation. Analytical results for the following, which were mailed the week of October 5, 1999, are presented in this report.

T-159	Trace constituents	N-64	Nutrient constituents
M-152	Major constituents	P-33	Low ionic strength constituents
N-63	Nutrient constituents	Hg-29	Mercury

The USGS requested that analytical results be returned by November 14, 1999 for evaluation and preparation of this report. Laboratories that are providing analytical services to USGS offices are requested to analyze the appropriate SRSs for the same analytes requested by the USGS offices. All laboratories are requested to include the analytical methods used to determine the concentration of each analyte. When analytical method information was provided, it has been included in tables 11 - 16.

Table 1.-Laboratory participants in the analyses of standard reference samples distributed in October 1999

<u>State</u>	<u>City</u>	<u>Participating Laboratory</u>
Alabama	Tuscaloosa	Geological Survey of Alabama
Alaska	Soldotna	Alaska Department of Fish and Game
Arkansas	Arkadelphia	Ouachita Baptist University, Department of Biology
	Fayetteville	University of Arkansas, Arkansas Water Resources Center
	Little Rock	Arkansas Department of Pollution Control and Ecology
California	Davis	University of California, Davis, Division of Environmental Studies
	Los Angeles	Metropolitan Water District, Water Quality Laboratory
	Martinez	Central Contra Costa Sanitary District
	Menlo Park	U.S. Geological Survey, Branch of Regional Research, Western Region
	Oakland	East Bay Municipal Utility District
	San Diego	U.S. Geological Survey, Water Resources Division
	Santa Fe Springs	West Coast Analytical Service, Inc.
	Tahoe City	Tahoe Research Group
	Three Rivers	U.S. Geological Survey, Sequoia Field Station
	West Sacramento	California Department of Water Resources
Colorado	Alamosa	Bureau of Reclamation
	Arvada	Quanterra Environmental Services
	Aurora	Core Laboratories, Inc.
	Boulder	U.S. Geological Survey, Branch of Regional Research, Central Region
	Boulder	U.S. Geological Survey, Branch of Regional Research, Aqueous Crystal Growth
	Colorado Springs	City of Colorado Springs, Environmental Quality Laboratory
	Denver	Metro Wastewater Reclamation
	Denver	U.S. Geological Survey, Mineral Resources Chemistry Project
	Denver	U.S. Geological Survey, Earth Science Investment Program
	Denver	U.S. Geological Survey, National Water Quality Laboratory
	Fort Collins	City of Fort Collins - Water Quality
	Fort Collins	U.S. Department of Agriculture, Forest Service
	Greeley	Central Colorado Water Conservatory District
	Loveland	Northern Colorado Water Conservation
	Northglenn	Northglenn Waste Water Treatment Plant
Pueblo	City of Pueblo Waste Water Treatment Plant	
Westminster	City of Westminster, Semper Water Treatment Plant	
Florida	Bradenton	Manatee County Environmental Management
	Brooksville	Southwest Florida Water Management District
	Ocala	U.S. Geological Survey, Water Resources Division, Quality Water Service Unit
	Orlando	Post, Buckley, Schuh, and Jernigan, Inc.
	Ormond Beach	Environmental Laboratory
	Tallahassee	City of Tallahassee, Water Quality Division
	Tallahassee	Florida Department of Environmental Protection
	Tallahassee	Savannah Laboratories and Environmental Services
	Tampa	Hillsborough County Environmental Protection Commission
West Palm Beach	South Florida Water Management District	
Georgia	Athens	University of Georgia
	Atlanta	U.S. Geological Survey, Water Resources Division
	Marietta	Cobb County Water System
	Stone Mountain	Dekalb County Public Works Department
Hawaii	Honolulu	University of Hawaii, SOEST Analytical Services
Idaho	Boise	U.S. Bureau of Reclamation, Pacific Northwest Regional Laboratory
	Boise	Boise City Water Quality Laboratory
Illinois	Champaign	Illinois Environmental Protection Agency

Table 1-Laboratory participants in the analyses of standard reference samples distributed in October 1999
--continued

State	City	Participating Laboratory
Iowa	Des Moines	University of Iowa Hygienic Laboratory, Des Moines Branch
Kansas	Lawrence	Kansas Geological Survey
	Wichita	City of Wichita, Water and Sewer Department
Kentucky	Frankfort	Division of Environmental Studies
	Lexington	Kentucky Geological Survey
Maine	Orono	University of Maine, Environmental Chemistry Laboratory
Maryland	Baltimore	Maryland Department of Health and Mental Hygiene
	Cambridge	University of Maryland, Horn Point Laboratory
	Ft. Meade	District of Columbia Department of Health
	Solomons	University of Maryland, Chesapeake Biology Laboratory
Michigan	Detroit	Detroit Water and Sewerage Department, Analytical Laboratory
Minnesota	Minneapolis	University of Minnesota, Department of Geology and Geophysics
	St. Paul	Metropolitan Council Environmental Services
	St. Paul	University of Minnesota, Department of Soil Science
Missouri	Jefferson City	Missouri Department of Health
Montana	Butte	Montana Bureau of Mines & Geology
	Helena	State of Montana, Laboratory Services Bureau
	Jefferson City	Montana Tunnels Laboratory
Nebraska	McCook	Olsen Laboratory
Nevada	Las Vegas	University of Nevada, Las Vegas
	Reno	Desert Research Institute
	Reno	Truckee Meadows Water Reclamation
New Jersey	Trenton	Suffolk County Water Authority
New York	Bolton Landing	Darrin Freshwater Institute
	Brewster	New York City Department of Environmental Protection, Brewster Laboratory
	Buffalo	Erie County Public Health Laboratory
	Grahamsville	New York City Department of Environmental Protection, Grahamsville Laboratory
	Hempstead	Nassau County Department of Health
	Milbrook	Institute of Ecosystem Studies
	North Babylon	EcoTest Laboratories
	Rochester	Monroe County Department of Health
	Rochester	Columbia Analytical Services
	Shokan	New York City Department of Environmental Protection, Ben Nessin Laboratory
	Syracuse	Onondaga County DDS
	Syracuse	State University of New York, College of Environmental Science and Forestry
	Troy	U.S. Geological Survey, Water Resources Division
	Troy	Rensselaer Polytechnic Institute
Valhalla	New York City Department of Environmental Protection, Kensico Laboratory	
Yorktown	New York City Department of Environmental Protection, Croton Gatehouse Lab	
North Carolina	Chapel Hill	City of Durham Water Resources Department
	Charlotte	Mecklenburg County Department of Environmental Protection
	Durham	Duke University
North Dakota	Rocky Mount	Tar River Regional Wastewater Treatment Facility
	Bismarck	North Dakota Department of Health, East Laboratory
Ohio	Bismarck	North Dakota State Water Commission
	Cincinnati	U.S. Environmental Protection Agency
	Cincinnati	U.S. Environmental Protection Agency, NERL
	Columbus	City of Columbus
	Cuyahoga Heights	Northeast Ohio Regional Sewer District
	Tiffin	Heidelberg College
Oklahoma	Wooster	Ohio State University, Ohio Agricultural Research and Developmental Center
	Norman	Oklahoma Geological Survey
Oregon	Corvallis	U.S. Department of Agriculture, Forestry Services Laboratory
	Hillsboro	Unified Sewerage Agency of Washington County

Table 1-Laboratory participants in the analyses of standard reference samples distributed in October 1999
 --continued

<u>State</u>	<u>City</u>	<u>Participating Laboratory</u>
Pennsylvania	Harrisburg	Pennsylvania Department of Environmental Protection
	Somerset	Geochemical Testing, Energy Center, Inc.
South Carolina	Cayce	Shealy Environmental Services
	Charleston	College of Charleston-Department of Geology
South Dakota	Brookings	Water Resources Institute-South Dakota State University
Tennessee	Knoxville	University of Tennessee
Texas	College Station	Albion Environmental
	College Station	Intermountain Labs
	Laredo	City of Laredo
Virginia	Chesapeake	City of Chesapeake
	Lorton	Fairfax Co. System Engineering and Monitoring Division
	Norfolk	Old Dominion University-Applied Marine Research
	Richmond	Division Consolidated Lab Service
Vermont	Waterbury	Vermont Agency of Natural Resources
Washington	Richland	Battelle Pacific Northwest
Wisconsin	Madison	Wisconsin State Lab of Hygiene
	Madison	Madison Department of Public Health
	Middleton	U.S. Geological Survey
Wyoming	Laramie	Wyoming Department of Public Health
Middle East Laboratories		
<u>Location</u>	<u>Participating Laboratory</u>	
Gaza	Islamic University of Gaza, Water & Environment Research Center	
Israel	Ministry of Agriculture Laboratory	
	Israeli Hydrologic Service Laboratory	
	Geological Survey of Israel Laboratory	
	Mekeroth Water Company, Eylat Laboratory	
	Public Health Laboratory Beer Shiva	
	Public Health Lab, Tel-Aviv	
	Water Resources Research Center, Institute for Desert Research	
Jordan	Royal Scientific Society, Environmental Research Centre	
	Water Authority of Jordan	
West Bank	Al-Quds University, College of Science and Technology, Water Research Center	
	Bethlehem University, Water and Soil Environmental Research Unit	
	Birzeit University, Center for Environmental & Occupational Health Sciences	
	Birzeit University, Gaza Branch	
	Najah University, Water & Environmental Studies Center	
	Palestinian Water Authority Laboratory	

PREPARATION OF STANDARD REFERENCE SAMPLES

All of the SRSs used in this evaluation were prepared by USGS personnel located in Lakewood, Colorado, and were analyzed for analyte concentrations and physical property values before mailing. A library of these SRSs is maintained, and these SRSs can be requested by participating laboratories and USGS offices for use in their quality-control programs.

Trace constituents sample T-159 was prepared using water collected from the Big Thompson River near Drake, Colorado. The water was pumped through 0.45-, 0.2-, and 0.1-micrometer (μm) filters, in series, into a 1200-liter (L) polypropylene drum. The water was continuously circulated and passed through a 0.1- μm filter and ultraviolet sterilizer for 24 hours. Following this circulation, the water was acidified to pH 1.5 with nitric acid and chlorinated to 5 parts per million (ppm) free chlorine with sodium hypochlorite. The trace constituent concentrations were adjusted by adding reagent grade chemicals. The sample was circulated an additional 24 hours prior to bottling. During bottling, the sample was pumped through an ultraviolet sterilizer and a 0.1- μm filter. The polypropylene and fluorinated ethylene propylene bottles and caps used were acid leached with 0.2N HNO_3 , deionized-water rinsed, and autoclave sterilized.

Major constituents sample M-152 was prepared using water collected from the Fall River near Idaho Springs, Colorado. The water was pumped through 0.45- 0.2- and 0.1- μm filters, in series, into a 1200-L polypropylene drum. The water was chlorinated to 5-ppm free chlorine with sodium hypochlorite, continuously circulated, and passed through an ultraviolet sterilizer for 24 hours prior to bottling. The major constituent concentrations were adjusted by adding reagent grade chemicals. The sample was circulated an additional 24 hours prior to bottling. During bottling the sample was pumped through an ultraviolet sterilizer and a 0.1- μm filter. The polypropylene bottles and caps used were acid leached with 0.2N HNO_3 , deionized-water rinsed, and autoclave sterilized.

Nutrient constituents sample N-63 was prepared using deionized water. These samples were prepared the week prior to the mailing for this SRS evaluation. The water was pumped through 0.45-, 0.2-, and 0.1- μm filters, in series, into a 25-L polypropylene drum. The desired nutrient concentrations were obtained by adding reagent-grade chemicals. The sample was continuously circulated through an ultraviolet sterilizer for 24 hours prior to being bottled. The 30-milliliter (mL) glass vials used were new, amber, acid leached with 0.2N HCl , deionized-water rinsed, and autoclave sterilized.

Nutrient constituents sample N-64 was prepared using water collected from the Fall River near Idaho Springs, Colorado. These samples were prepared the week prior to the mailing for this SRS evaluation. The water was pumped through 0.45- 0.2- and 0.1- μm filters, in series, into a 200-L polypropylene drum and continuously circulated and passed through an ultraviolet sterilizer for 24 hours. The desired nutrient concentrations were obtained by adding reagent-grade chemicals. The sample was passed through a 0.1- μm filter while bottling. The 250-mL polyethylene bottles used were new, amber, acid leached 0.2N HCl , deionized-water rinsed, and autoclave sterilized.

Low ionic strength constituents sample P-33 was prepared in a 400-L polypropylene drum using snowmelt from Alice, Colorado. The water was pumped into the drum through 0.45- 0.2- and 0.1- μm filters in series. The desired phosphate and fluoride concentrations were obtained by adding reagent-grade chemicals. Prior to bottling, the sample was continuously mixed for 24 hours while being circulated through a 0.1- μm filter and an ultraviolet sterilizer. During bottling the sample was pumped through an ultraviolet sterilizer and a 0.1- μm filter. The 500-mL polypropylene bottles used were acid leached with 0.2N HNO_3 , deionized-water rinsed, and autoclave sterilized.

Mercury sample Hg-29 was prepared using water collected from the Fall River near Idaho Springs, Colorado. The sample was prepared in a 200-L polypropylene drum. The river water was pumped into this drum through 0.45-, 0.2-, and 0.1- μm filters in series. The water was continuously circulated and passed through a 0.1- μm filter and ultraviolet sterilizer for 48 hours. Nitric acid (5-percent, by volume) and dichromate compound (0.05-percent, by weight) were added to stabilize the sample. The desired mercury concentration was obtained by adding a mercury standard solution. Following an additional 24 hours of circulation, the sample was bottled. The 250-mL glass bottles and tetrafluoroethylene fluorocarbon resin caps used were new, acid leached with 0.2N HNO_3 , and deionized-water rinsed, and autoclave sterilized.

LABORATORY ANALYSES

The participating laboratories were asked to determine constituents that are summarized in table 2. The number of analytes varied from 28 in T-159 (trace constituents) to 1 in Hg-29 (mercury).

Table 2. -Analytes determined in standard reference samples distributed in October 1999

[mg/L, milligrams per liter; µg/L, micrograms per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius]

Constituent or property		Units	T-159	M-152	N-63	N-64	P-33	Hg-29
Acidity	Acidity as CaCO ₃	mg/L					X	
Alk	Alkalinity as CaCO ₃	mg/L		X				
Ag	Silver	µg/L	X					
Al	Aluminum	µg/L	X					
As	Arsenic	µg/L	X					
B	Boron	µg/L	X	X				X
Ba	Barium	µg/L	X					
Be	Beryllium	µg/L	X					
Ca	Calcium	mg/L	X	X				X
Cd	Cadmium	µg/L	X					
Cl	Chloride	mg/L		X			X	
Co	Cobalt	µg/L	X					
Cr	Chromium	µg/L	X					
Cu	Copper	µg/L	X					
DSRD	Dissolved solids	mg/L		X				
F	Fluoride	mg/L		X			X	
Fe	Iron	µg/L	X					
Hg	Mercury	µg/L						X
K	Potassium	mg/L	X	X			X	
Li	Lithium	µg/L	X					
Mg	Magnesium	mg/L	X	X			X	
Mn	Manganese	µg/L	X					
Mo	Molybdenum	µg/L	X					
Na	Sodium	mg/L	X	X			X	
NH ₃ as N	Ammonia	mg/L				X	X	
NH ₃ +Org N as N	Ammonia + Organic N	mg/L				X	X	
Ni	Nickel	µg/L	X					
NO ₃ as N	Nitrate	mg/L				X	X	
Pb	Lead	µg/L	X					
pH		unit		X			X	
PO ₄ as P	Orthophosphate	mg/L				X	X	X
total P as P	Phosphorus	mg/L		X		X	X	
Sb	Antimony	µg/L	X					
Se	Selenium	µg/L	X					
SiO ₂	Silica	mg/L	X	X				
SO ₄	Sulfate	mg/L		X			X	
Sp Cond	Specific conductance	µS/cm		X			X	
Sr	Strontium	µg/L	X	X				
Tl	Thallium	µg/L	X					
U	Uranium	µg/L	X					
V	Vanadium	µg/L	X	X				
Zn	Zinc	µg/L	X					

Laboratories were requested to identify the method used for each constituent according to analytical method codes in table 3.

Table 3. Analytical method codes

Code	Method
0	Other
1	Atomic absorption: direct, air
2	Atomic absorption: direct, nitrous oxide
3	Atomic absorption: graphite furnace
4	Inductively coupled plasma
5	Direct current plasma
6	Inductively coupled plasma/mass spectrometry
7	Ion chromatography
8	Atomic absorption: cold vapor
9	Atomic fluorescence
10	Atomic absorption: extraction [<i>specify chelating agents</i>]
11	Atomic absorption: hydride [<i>specify reducing agent</i>]
12	Flame emission
20	Titration: colorimetric [<i>specify color reagent</i>]
21	Titration: electrometric [<i>specify reducing or oxidizing agent/color reagent</i>]
22	Colorimetric: [<i>specify reducing or oxidizing agent/color reagent</i>]
40	Ion selective electrode
41	Electrometric [<i>pH and specific conductance</i>]
50	Gravimetric: [<i>specify filtration, evaporation, and so forth</i>]
51	Turbidimetric

Participating laboratories were also asked to identify the method used, such as those references listed next, to further define the methods.

1. American Public Health Association, American Water Works Association, and Water Environment Federation, 1995, Standard methods for the examination of water and wastewater (19th ed.): Washington, D.C., American Public Health Association, variable pagination.
2. American Society for Testing and Materials, 1995, Annual book of ASTM standards: Philadelphia, v. 11.0, and v. 11.02.
3. Kopp, J.F., and McKee, G.F., 1979, Methods for chemical analysis of water and wastes: Cincinnati, U.S. Environmental Protection Agency, EPA 600/4-79-020, rev. 1983, 460 p.
4. Fishman, M.J., and Friedman, L.C., eds., 1989. Methods for determination of inorganic substances in water and fluvial sediments (3rd ed.): U.S. Geological Survey Techniques of Water-Resources Investigations, Book 5, Chapter A1, 545 p.
5. Miscellaneous manufacturer's instrument manuals or references.

LABORATORY PERFORMANCE RATINGS

To facilitate laboratory intercomparison, laboratory performance ratings that are based on the analyses reported for each SRS are included in tables 4 through 17 in this report. For each SRS, averages of all the analyte ratings and the number of analyte values reported are given for each participating laboratory. In some cases, laboratory reported values in tables 4 - 17 might have been reformatted in terms of significant figures to meet publication criteria. For example, a reported value of 15 may have been changed to 15.0 or a value of 102.86 may have been changed to 102.9 in these tables. However, the actual reported values by all the laboratories were used to calculate the statistical results and performance ratings presented in the report.

Laboratory determination of each analyte is rated on a scale 4 to 0, based on the absolute Z-value, as listed below:

Rating	Absolute Z-value
4 (Excellent)	0.00 to 0.50
3 (Good)	0.51 to 1.00
2 (Satisfactory)	1.01 to 1.50
1 (Marginal)	1.51 to 2.00
0 (Unsatisfactory)	Greater than 2.00

A weighted laboratory rating greater than 2.0 is considered satisfactory, and ratings less than 2.0 are considered unsatisfactory. Ratings are based on the relative performance of laboratories on specific samples and should be reviewed and evaluated on a case-by-case basis for each laboratory considering such factors as methods used and data needs of specific USGS projects using the laboratory data.

STATISTICAL PRESENTATION OF DATA

Data in this report have been evaluated using nonparametric statistics as described by Hoaglin and others (1983). This statistical approach is a resistant statistic because outliers do not influence the median, as they influence the mean in traditional parametric statistics. Analytical data for each analyte are presented in tabular and graphical forms in tables 11 through 16. Tabulated data for each analyte include the laboratory code number; reported values; analytical method; most probable value (MPV); number of reported analyses; excluding less than values (N); data range; Z-value; and the F-pseudostandard deviation. The Z-value is equivalent to the Z-score of traditional statistics, being the number of deviations the reported value is from the MPV. The F-pseudostandard deviation is equivalent to the standard deviation (σ) of traditional statistics when the data has a Gaussian distribution. If an analyte has a sufficient number of analyses by a given method, usually 7, the F-pseudostandard deviation for that analytical method is reported in the block of data listed for each analyte.


The median value is considered the MPV. The median (midpoint) divides the ordered data into halves and is designated the MPV. The hinges include the middle 50-percent of the data and are the mid-values of the upper and lower halves of the data. The hinges are similar to quartiles, but are not mathematically equivalent. The range of data between the upper hinge (Uh) and the lower hinge (Lh) by comparison of the H-spr value to the Gaussian distribution relation; 67.45 percent of the data "hinges" between plus and minus 1σ , resulting in a H-spr of $2 \times 0.6745 = 1.349\sigma$. This relation allows the calculation of the F-pseudostandard deviation = (H-spr)/1.349. Laboratories reporting "less than" values are not performance rated unless their reported "less than" values are greater than two Z-values from the MPV.

The graphical plot of the reported data is shown in figure 1. The upper and lower boundaries of the graphical plots generally are +3 and -3 F-pseudosigma deviations from the median. Computer-program scaling constraints do not permit these boundaries to always be graphed at exactly these values as shown in the graphical plot. Reported values are grouped by analytical method in ascending order of value. Lines designate the MPV, Uh, Lh, and the UWL and LWL at +2 and -2 F-pseudosigma, respectively. "Less than" values are not plotted.

In some cases, if the F-pseudosigma is less than 5 percent of the MPV, the rating criterion is set to 5 percent of the MPV, as shown in Table 11, page 41.

The term "insufficient data" is included in some of the tables and is used when the number of analyses is less than seven or the calculated F-pseudosigma is greater than the MPV.

In some cases the F-pseudosigma is equal to or greater than the MPV. This results in an MPV = insufficient data. An estimated MPV may be calculated from the available data for a single analytical method, this estimated concentration is denoted by MPV = Estimated. Estimated values are not used to rate laboratories.



NOTE: vertical scale is the concentration value of the individual analyte in appropriate units (see table 2). Horizontal scale is the laboratory reported values separated by method (different symbols) and plotted by increasing values. Numbers next to each symbol at the bottom of the figure are analytical method codes that are described in table 3. Laboratory-reported results greater than 3 F-pseudosigma from the median are not shown on the graphs.

Figure 1. -Statistical parameters shown on reported-data graphs in tables 13 - 20

REFERENCE

Hoaglin, D.C., Mosteller, F., and Tukey, J.W., Eds. 1983, Understanding robust and exploratory data analysis: New York, NY, John Wiley, Inc., p. 38-41.

Table 4. Overall laboratory performance ratings for standard reference samples distributed in October 1999

[SRS, standard reference sample; Lab, laboratory number; OWR, overall weighted rating for all sample types; OLR, overall laboratory rating for reported values of a sample type; V/64, number of reported values of 65 total possible values from all sample types; V/28, V/15, V/5, V/5, V/10 and V/1 are number of reported values possible for I-159, M-152, N-63, N-64, P-33 and Hg-29 respectively; NR, not rated; --, not reported.]

Lab	SRS =		T-159		M-152		N-63		N-64		P-33		Hg-29	
	OWR	V/64	OLR	V/28	OLR	V/15	OLR	V/5	OLR	V/5	OLR	V/10	OLR	V/1
1	3.7	63	3.7	28	3.9	14	4.0	5	3.6	5	3.1	10	4	1
2	2.3	8									2.3	8		
3	3.1	55	3.2	26	3.4	14			3.0	5	2.0	9	4	1
10	3.2	31	2.2	9	3.5	11	4.0	5	3.8	5			1	1
11	2.5	62	2.8	26	2.9	15	1.2	5	2.2	5	1.5	10	4	1
12	1.9	25	2.4	7	2.0	6	2.0	3	1.4	5	2.0	3	0	1
13	2.7	45	2.6	24	2.8	13	2.0	4	3.3	4			NR	
18	3.0	10					2.6	5	3.4	5				
21	3.7	6	3.0	1			3.8	5						
23	2.6	37	2.4	21	3.3	6	1.8	5	3.8	5				
24	3.0	29	3.1	17	2.8	12								
25	2.6	21			1.7	7	3.8	4	2.3	4	3.0	6		
26	2.2	32	2.3	18	2.1	12			1.5	2				
32	3.6	35	3.7	26	3.8	8							1	1
33	3.2	34	2.3	10	3.8	11	2.7	3			3.7	10		
36	1.8	10					1.4	5	2.2	5				
38	3.7	27			3.7	9	4.0	5	3.4	5	3.8	8		
39	2.2	35	2.1	22	2.1	12							4	1
42	3.0	50	3.4	25	2.6	14	1.5	2	3.5	2	2.4	7		
43	3.5	4			3.5	4								
45	2.2	26	2.4	8	2.5	10					2.0	7	0	1
46	2.9	54	3.2	22	2.1	12	2.6	5	3.2	5	2.9	9	4	1
48	2.5	50	3.0	24	2.7	11	0.6	5			2.1	9	2	1
50	3.3	40	3.4	22	3.1	12	3.8	5					2	1
51	2.8	19	2.3	4	3.0	10	3.0	5						
53	3.8	4					4.0	2	3.5	2				
57	3.0	40	3.1	27	2.6	13								
59	2.8	44	2.6	25	3.5	2	2.4	5	2.6	5	3.5	6	4	1
64	3.0	32	2.2	5	3.7	10	4.0	4	3.5	4	2.1	9		
69	3.5	2					4.0	1	3.0	1				
70	2.5	51	2.8	26	2.4	14	1.0	5	2.2	5			4	1
76	3.4	28	3.6	15	3.8	8	3.0	2	2.0	3				
81	2.3	55	2.2	24	2.0	14	2.2	5			3.1	11	2	1
83	3.7	6					3.5	2	4.0	2	3.5	2		
85	3.1	20			2.9	12	3.8	4	2.8	4				
86	2.8	49	2.5	24	3.4	13	1.7	3			3.0	8	4	1
87	2.2	32	2.4	17	1.6	9			3.0	5			0	1
89	2.5	57	2.1	22	2.2	13	3.6	5	3.4	5	2.7	11	4	1
90	3.5	6					4.0	3	3.0	3				
93	2.5	47	1.8	18	2.8	11	2.5	4	2.3	4	3.6	10		
96	2.8	32	2.8	15	2.8	6	2.4	5	3.2	5			4	1
97	2.7	27	2.5	12	2.8	11	2.8	4						
102	1.4	53	1.3	22	1.7	12	1.8	5	1.8	5	0.9	9		
105	1.9	9					1.8	4	2.0	5				
107	3.1	15	2.2	6	4.0	3			3.5	4	4.0	2		
109	2.6	16	2.8	6	2.4	10								
110	3.0	1					3.0	1						
113	3.4	51	3.4	19	3.3	12	4.0	4	3.8	5	3.0	10	4	1
114	1.8	24	1.5	11	2.3	7	2.0	3	1.7	3				
121	3.1	27	3.0	20	3.3	7								
126	2.1	13	2.1	12									2	1
127	3.4	63	3.3	26	3.9	15	3.0	5	3.4	5	3.5	11	3	1
129	3.2	10					4.0	5	2.4	5				
131	2.2	31	2.2	20	2.2	11								
134	3.7	64	3.7	28	3.8	15	3.6	5	2.6	5	3.9	10	4	1
138	3.7	60	3.7	25	3.7	15	4.0	5	3.6	5	3.9	9	3	1
140	2.4	47	2.9	15	1.9	12	2.2	5	1.4	5	3.0	10		
141	2.1	50	2.3	26	1.8	13					2.0	10	2	1
142	2.7	45	2.8	26	2.3	13	3.2	5					4	1
143	3.5	19			3.4	5	3.8	5	3.2	5	3.5	4		
144	2.2	15	2.3	14									1	1
145	2.4	54	1.7	20	3.0	14	3.6	5	2.8	5	2.1	9	2	1
146	2.8	45	3.5	20	2.1	9	2.2	5	2.6	5	3.0	5	0	1
147	3.6	8	3.6	7									4	1
151	3.2	42	3.4	23	3.3	12	2.0	3	1.3	3			4	1
154	2.8	50	2.9	26	3.1	13	2.2	5	2.8	5			0	1
155	2.9	15					3.0	5	3.2	5	2.4	5		
158	2.4	10					2.4	5	2.4	5				
180	2.4	46	1.6	16	2.6	10	3.0	5	3.2	5	3.1	9	0	1
183	2.2	17			2.2	5	2.0	4	2.3	3	2.2	5		

Table 4. Overall laboratory performance ratings for standard reference samples distributed in October 1999--Continued

[SRS, standard reference sample; Lab, laboratory number; OWR, overall weighted rating for all sample types; OLR, overall laboratory rating for reported values of a sample type; V/64, number of reported values of 65 total possible values from all sample types; V/28,V/15,V/5,V/5,V/10 and V/1 are number of reported values possible for I-159, M-152, N-63, N-64, P-33 and Hg-29 respectively; NR, not rated; --, not reported.]

Lab	SRS =		T-159		M-152		N-63		N-64		P-33		Hg-29	
	OWR	V/64	OLR	V/28	OLR	V/15	OLR	V/5	OLR	V/5	OLR	V/10	OLR	V/1
185	3.2	29	3.3	4	3.4	9	3.3	4	3.3	4	3.0	8		
190	2.7	40	3.0	14	2.5	12	3.0	2	2.5	2	2.3	10		
191	3.2	41	3.4	23	3.2	10					2.9	8		
193	2.5	33	2.4	14	2.8	6	2.7	3	2.0	4	2.8	5	4	1
196	3.1	7			3.0	3					3.3	4		
198	2.7	21	2.4	12			2.8	4	3.3	4			3	1
200	3.3	4							3.3	4				
203	2.9	36	2.8	18	2.9	10					2.9	7	3	1
204	2.3	28	2.1	20	4.0	1			2.2	5	4.0	2		
205	2.5	4					1.0	2	4.0	2				
208	2.5	6			3.5	2			3.0	2	1.0	2		
209	2.9	25	2.4	5	2.9	7	3.0	3	3.3	3	3.1	7		
212	2.6	52	3.0	26	2.9	15	1.2	5	1.8	5			0	1
213	3.0	10			3.3	3	3.0	2	2.8	4			3	1
215	1.7	50	1.8	20	1.9	12	0.8	4	3.4	5	0.9	8	0	1
219	3.0	36	3.3	26	2.3	9							3	1
224	2.5	10					2.4	5	2.6	5				
227	3.0	34	3.0	10	3.5	8	3.6	5	2.2	5	2.5	6		
228	2.0	9									2.0	9		
234	3.2	51	3.2	27	3.5	15	3.3	4	1.8	4			4	1
236	2.2	34	2.1	22	2.4	12								
237	3.5	8									3.5	8		
238	3.4	5									3.4	5		
241	2.9	49	3.0	23	3.4	11	3.3	4			2.1	10	2	1
243	2.8	11			2.7	3	2.3	3	2.7	3	4.0	2		
244	3.8	4			3.5	2					4.0	2		
247	2.7	61	2.4	27	2.7	15	2.0	5	3.0	5	3.5	8	4	1
253	2.4	13			2.2	5	2.0	4	3.0	4				
254	3.3	23	3.1	15	3.5	8								
255	3.1	33	2.8	15	3.3	6	4.0	3	3.8	4	3.0	4	1	1
259	3.1	34	3.1	21	3.5	12							0	1
265	3.4	48	3.3	28	3.7	12					2.9	7	4	1
268	2.7	20	1.8	4	3.0	8					2.8	8		
273	1.7	44	1.8	20	1.0	13					2.4	11		
274	1.1	27	1.6	5	1.4	11					0.6	11		
275	1.4	11			1.4	11								
277	1.7	24	1.6	13	1.5	10							4	1
279	2.6	20	3.3	4	2.5	8					2.4	8		
282	2.3	61	2.6	26	2.3	14	0.2	5	2.4	5	2.7	10	3	1
285	0.4	8					0.8	4	0.0	4				
287	1.6	39	1.6	15	1.6	8	0.0	3	1.0	3	2.1	10		
291	4.0	1							4.0	1				
292	2.6	34	2.6	16	3.0	10	2.3	4	2.3	4				
298	4.0	1											4	1
301	2.6	11					3.0	4			2.4	7		
302	1.8	13	2.3	4	1.7	9								
304	3.3	22	3.1	14	3.6	7							4	1
305	1.6	13	1.6	9	1.8	4								
306	1.6	8					2.5	4	0.8	4				
307	2.3	27	2.5	14	2.3	8	2.0	4					0	1
308	1.3	9	1.0	6	4.0	1	0.0	1					2	1
309	0.3	14	0.4	8	0.2	6								
312	2.3	10	0.0	1	1.3	3	3.5	2	0.0	1	4.0	3		
313	3.5	10					3.0	5	4.0	5				
314	1.6	5					1.6	5						
316	3.8	10					3.8	5	3.8	5				
317	2.7	3					2.7	3						
318	4.0	5					4.0	5						
319	1.8	10					1.6	5	2.0	5				
320	2.5	10					2.8	5	2.2	5				
321	2.9	11					3.3	4			2.7	7		

Table 5. Laboratory performance ratings for standard reference sample T-159 (trace constituents)

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte = Ag (Silver)				Al (Aluminum)		As (Arsenic)		B (Boron)		Ba (Barium)		Be (Beryllium)		Ca (Calcium)	
	MPV = 9.67		mg/L		31.9	mg/L	28.4	mg/L	26.4	mg/L	38.1	mg/L	10.8	mg/L	25.5	mg/L
	F-pseudostigma = 0.91	V/28	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	3.7	28	9.49	4	31.2	4	29.3	3	28.2	3	38.1	4	11.0	4	25.9	4
3	3.2	26	9.48	4	31.0	4	28.1	4	31.9	1	35.4	2	10.7	4	24.9	4
10	2.2	9	--	--	--	--	31.0	1	--	--	--	--	--	--	--	--
11	2.8	26	10.00	4	120.0	0	28.0	4	26.0	4	37.0	3	11.0	4	26.0	4
12	2.4	7	9.40	4	--	--	--	--	--	--	--	--	--	--	26.0	4
13	2.6	24	9.99	4	31.0	4	27.7	4	--	--	37.3	4	11.0	4	25.0	4
21	3.0	1	--	--	--	--	--	--	--	--	--	--	--	--	--	--
23	2.4	21	11.60	0	32.2	4	24.2	0	--	--	37.8	4	9.7	0	27.4	2
24	3.1	17	--	--	--	--	--	--	25.0	4	39.5	3	--	--	25.4	4
26	2.3	18	--	--	33.1	4	27.5	3	24.2	3	38.5	4	11.2	3	15.5	0
32	3.7	26	9.40	4	30.2	4	28.6	4	27.1	4	38.2	4	10.4	3	25.7	4
33	2.3	10	--	--	70.0	0	--	--	--	--	43.0	0	--	--	27.0	2
39	2.1	22	8.91	3	48.5	0	28.0	4	17.5	0	36.6	3	--	--	24.0	2
42	3.4	25	9.67	4	< 30	NR	29.6	3	< 30	NR	38.0	4	--	--	26.3	3
45	2.4	8	--	--	--	--	--	--	--	--	--	--	--	--	23.9	2
46	3.2	22	10.10	4	34.6	3	26.2	2	25.5	4	39.1	3	10.7	4	25.1	4
48	3.0	24	9.40	4	31.0	4	27.9	4	26.3	4	38.0	4	10.7	4	29.2	0
50	3.4	22	9.90	4	31.9	4	27.8	4	28.0	3	40.5	2	10.6	4	--	--
51	2.3	4	--	--	--	--	--	--	--	--	--	--	--	--	27.2	2
57	3.1	27	10.30	3	32.1	4	31.3	1	27.4	4	37.3	4	11.2	3	25.2	4
59	2.6	25	9.00	3	32.2	4	31.4	1	--	--	38.9	4	11.8	1	27.3	2
64	2.2	5	--	--	--	--	--	--	--	--	--	--	--	--	33.4	0
70	2.8	26	8.00	1	31.4	4	29.6	3	< 50	NR	38.1	4	12.0	0	26.5	3
76	3.6	15	--	--	31.1	4	--	--	--	--	38.7	4	11.3	3	25.1	4
81	2.2	24	7.00	0	42.0	0	30.0	3	--	--	36.0	2	10.0	2	26.3	3
86	2.5	24	8.90	3	--	--	26.6	2	13.0	0	37.6	4	10.6	4	25.8	4
87	2.4	17	10.30	3	--	--	25.6	1	--	--	38.5	4	--	--	24.9	4
89	2.1	22	8.91	3	36.0	2	28.4	4	--	--	< 50	NR	13.5	0	24.7	3
93	1.8	18	10.10	4	24.9	1	26.3	2	--	--	--	--	9.9	1	24.6	3
96	2.8	15	10.70	2	--	--	28.3	4	--	--	< 100	NR	10.4	3	--	--
97	2.5	12	--	--	--	--	29.2	4	--	--	40.8	2	--	--	--	--
102	1.3	22	40.00	0	29.4	3	22.6	0	--	--	43.9	0	10.2	2	27.0	2
107	2.2	6	--	--	--	--	--	--	--	--	--	--	--	--	25.2	4
109	2.8	6	--	--	--	--	--	--	--	--	--	--	--	--	25.0	4
113	3.4	19	10.45	3	--	--	30.0	3	--	--	37.8	4	11.1	3	25.9	4
114	1.5	11	12.00	0	--	--	--	--	--	--	--	--	12.0	0	--	--
121	3.0	20	9.10	3	--	--	28.5	4	--	--	37.0	3	--	--	25.4	4
126	2.1	12	11.00	2	--	--	31.0	1	--	--	< 200	NR	10.6	4	25.0	4
127	3.3	26	9.45	4	103.0	0	26.6	2	25.5	4	36.4	3	10.7	4	25.4	4
131	2.2	20	< 30	NR	24.0	0	< 40	NR	24.0	3	35.0	1	--	--	25.3	4
134	3.7	28	9.56	4	32.0	4	29.7	3	26.5	4	37.3	4	10.5	3	25.3	4
138	3.7	25	9.43	4	30.7	4	28.4	4	27.3	4	38.2	4	10.5	3	25.2	4
140	2.9	15	8.60	2	--	--	--	--	--	--	50.0	0	--	--	25.5	4
141	2.3	26	10.70	2	26.6	2	32.1	0	50.1	0	40.8	2	11.6	2	26.9	2
142	2.8	26	12.70	0	< 30	NR	29.0	4	< 30	NR	40.1	2	10.8	4	27.0	2
144	2.3	14	9.20	3	--	--	24.5	0	--	--	--	--	8.9	0	--	--
145	1.7	20	--	--	< 180	NR	46.0	0	34.0	0	39.0	3	11.0	4	26.6	3
146	3.5	20	11.40	1	< 36.9	NR	29.4	3	--	--	39.6	3	10.9	4	25.2	4
147	3.6	7	--	--	32.3	4	28.0	4	--	--	--	--	--	--	--	--
151	3.4	23	9.60	4	27.4	2	27.8	4	--	--	37.6	4	10.1	2	25.7	4
154	2.9	26	9.90	4	29.5	3	28.8	4	29.0	3	35.7	2	10.9	4	25.2	4
180	1.6	16	10.70	2	< 51.6	NR	--	--	63.2	0	40.8	2	13.0	0	26.5	3
185	3.3	4	--	--	--	--	--	--	--	--	--	--	--	--	25.1	4
190	3.0	14	9.20	3	32.8	4	28.7	4	--	--	--	--	--	--	33.1	0
191	3.4	23	--	--	32.0	4	29.2	4	25.0	4	39.0	3	10.9	4	26.3	3

Table 5. Laboratory performance ratings for standard reference sample T-159 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte = Ag (Silver)				Al (Aluminum)		As (Arsenic)		B (Boron)		Ba (Barium)		Be (Beryllium)		Ca (Calcium)	
	MPV =	9.67	mg/L		31.9	mg/L	28.4	mg/L	26.4	mg/L	38.1	mg/L	10.8	mg/L	25.5	mg/L
	F-pseudosigma =	0.91			3.7		1.6		3.0		1.9		0.4		0.8	
OLR	V/28	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	
193	2.4	14	10.20	3	--	--	23.2	0	--	--	--	--	10.5	3	25.8	4
198	2.4	12	< 50	NR	< 50	NR	29.1	4	--	--	< 50	NR	11.4	2	26.3	3
203	2.8	18	9.00	3	29.0	3	27.5	3	--	--	40.0	2	--	--	24.8	3
204	2.1	20	5.00	0	28.2	2	26.0	1	--	--	35.7	2	8.0	0	25.8	4
209	2.4	5	--	--	32.5	4	--	--	--	--	--	--	--	--	25.2	4
212	3.0	26	9.83	4	68.0	0	27.4	3	28.2	3	38.7	4	11.7	1	24.2	2
215	1.8	20	8.00	1	--	--	25.0	0	--	--	37.0	3	11.0	4	25.5	4
219	3.3	26	11.00	2	28.0	2	30.0	3	--	--	36.0	2	11.0	4	25.7	4
227	3.0	10	--	--	30.5	4	--	--	--	--	--	--	--	--	25.2	4
234	3.2	27	10.00	4	38.4	1	29.4	3	27.0	4	37.6	4	10.8	4	25.2	4
236	2.1	22	10.00	4	55.0	0	< 45	NR	24.0	3	39.0	3	11.0	4	24.4	3
241	3.0	23	9.38	4	28.7	3	31.2	1	--	--	39.5	3	9.9	1	25.8	4
247	2.4	27	9.84	4	31.4	4	31.0	1	64.0	0	40.9	1	13.0	0	24.8	3
254	3.1	15	--	--	30.7	4	--	--	--	--	--	--	--	--	26.1	4
255	2.8	15	10.80	2	33.1	4	30.3	2	22.8	2	--	--	--	--	25.3	4
259	3.1	21	9.60	4	37.1	2	27.5	3	12.0	0	38.0	4	--	--	25.4	4
265	3.3	28	9.00	3	27.0	2	28.5	4	25.0	4	38.0	4	11.0	4	25.5	4
268	1.8	4	--	--	--	--	--	--	--	--	--	--	--	--	25.5	4
273	1.8	20	9.10	3	31.6	4	--	--	19.5	0	46.9	0	--	--	27.4	2
274	1.6	5	--	--	--	--	--	--	--	--	--	--	--	--	25.6	4
277	1.6	13	9.30	4	--	--	27.6	4	--	--	34.1	0	--	--	--	--
279	3.3	4	--	--	--	--	--	--	--	--	--	--	--	--	25.8	4
282	2.6	26	12.20	0	47.0	0	31.6	1	27.4	4	36.8	3	10.7	4	25.8	4
287	1.6	15	--	--	71.6	0	19.0	0	< 1	0	--	--	--	--	26.5	3
292	2.6	16	17.00	0	--	--	--	--	--	--	35.0	1	9.0	0	25.8	4
302	2.3	4	--	--	--	--	--	--	--	--	--	--	--	--	25.2	4
304	3.1	14	7.70	0	--	--	--	--	--	--	37.2	4	10.7	4	--	--
305	1.6	9	9.04	3	--	--	20.2	0	--	--	--	--	--	--	26.2	3
307	2.5	14	9.75	4	--	--	27.7	4	--	--	--	--	--	--	24.0	2
308	1.0	6	--	--	--	--	--	--	--	--	--	--	--	--	--	--
309	0.4	8	11.00	2	--	--	54.0	0	--	--	--	--	--	--	27.6	1
312	0.0	1	--	--	--	--	--	--	--	--	--	--	--	--	--	--

Table 5. Laboratory performance ratings for standard reference sample T-159 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value										
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00										
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00										
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)												
Analyte = Cd (Cadmium)		Co (Cobalt)		Cr (Chromium)		Cu (Copper)		Fe (Iron)		K (Potassium)		Li (Lithium)				
MPV = 24.0 µg/L		13.3 mg/L		26.8 mg/L		33.4 mg/L		48.9 mg/L		1.52 mg/L		8.97 mg/L				
F-pseudosigma = 1.6		0.9		1.8		2.5		6.2		0.13		1.53				
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	23.8	4	13.3	4	28.0	3	32.7	4	50.9	4	1.47	4	7.92	4		
3	22.7	3	12.8	3	27.1	4	32.7	4	73.7	0	1.57	4	< 10	NR		
10	24.0	4	--	--	28.0	3	35.0	3	56.0	2	--	--	--	--		
11	25.0	3	13.0	4	28.0	3	35.0	3	90.0	0	1.61	3	11.00	2		
12	24.5	4	--	--	--	--	31.0	3	--	--	--	--	--	--		
13	20.5	0	16.3	0	26.6	4	19.3	0	44.3	3	1.57	4	--	--		
21	--	--	--	--	--	--	--	--	54.8	3	--	--	--	--		
23	19.0	0	--	--	26.0	4	31.1	3	44.2	3	1.48	4	--	--		
24	24.2	4	14.2	2	31.9	0	37.9	1	50.2	4	1.48	4	--	--		
26	--	--	14.5	2	--	--	< 2	0	38.4	1	1.00	0	8.00	4		
32	24.1	4	13.6	4	26.4	4	35.0	3	--	--	1.50	4	--	--		
33	--	--	--	--	--	--	--	--	50.0	4	1.52	4	--	--		
39	22.0	2	11.0	0	26.0	4	32.7	4	40.7	2	1.42	3	--	--		
42	24.0	4	13.6	4	27.5	4	33.6	4	40.9	2	1.45	3	9.23	4		
45	--	--	--	--	41.9	0	34.0	4	--	--	1.69	2	--	--		
46	22.5	3	--	--	27.1	4	34.4	4	45.9	4	1.50	4	--	--		
48	23.7	4	12.3	2	26.1	4	30.3	2	46.0	4	1.47	4	--	--		
50	25.2	3	13.3	4	28.8	2	35.2	3	47.2	4	--	--	7.60	4		
51	--	--	--	--	--	--	--	--	--	--	2.25	0	--	--		
57	23.2	4	13.0	4	25.0	3	34.4	4	49.0	4	2.70	0	11.10	2		
59	25.4	3	13.3	4	29.5	2	34.9	3	34.5	0	1.60	3	7.15	3		
64	--	--	--	--	--	--	--	--	--	--	1.47	4	--	--		
70	23.4	4	12.4	2	26.0	4	29.4	1	51.2	4	1.48	4	--	--		
76	24.4	4	13.1	4	28.2	3	--	--	--	--	--	--	--	--		
81	24.0	4	14.0	3	30.0	1	37.0	2	42.0	2	1.77	1	--	--		
86	22.6	3	12.4	2	26.6	4	32.6	4	33.8	0	1.50	4	--	--		
87	25.1	3	--	--	26.8	4	34.8	3	45.9	4	1.49	4	--	--		
89	26.3	2	15.6	0	25.1	3	31.3	3	43.9	3	1.54	4	--	--		
93	21.7	2	--	--	26.4	4	--	--	42.9	3	1.23	0	--	--		
96	24.6	4	12.9	4	25.7	3	32.7	4	60.0	1	--	--	--	--		
97	--	--	--	--	--	--	35.5	3	41.0	2	1.45	3	--	--		
102	25.4	3	15.0	1	28.2	3	32.0	3	54.7	3	1.14	0	--	--		
107	--	--	--	--	--	--	--	--	60.0	1	1.84	0	--	--		
109	--	--	--	--	--	--	--	--	52.6	3	1.71	2	--	--		
113	24.8	4	--	--	--	--	34.2	4	46.7	4	1.57	4	--	--		
114	26.0	2	--	--	33.0	0	30.5	2	42.2	2	--	--	--	--		
121	25.1	3	11.6	1	24.5	2	30.4	2	53.0	3	1.47	4	--	--		
126	27.9	0	--	--	25.3	3	41.0	0	--	--	--	--	--	--		
127	26.0	2	13.0	4	27.3	4	34.6	4	49.7	4	1.57	4	< 10	NR		
131	22.0	2	12.0	2	24.0	1	47.0	0	50.0	4	1.60	3	8.90	4		
134	23.7	4	13.4	4	26.9	4	32.4	4	48.8	4	1.50	4	8.66	4		
138	23.8	4	14.0	3	27.1	4	33.2	4	51.5	4	1.43	3	--	--		
140	22.9	3	--	--	30.0	1	34.3	4	50.0	4	1.40	3	--	--		
141	26.4	1	14.1	3	25.0	3	35.4	3	39.4	1	1.67	2	10.50	3		
142	24.9	3	12.8	3	26.8	4	31.5	3	40.0	2	1.49	4	8.47	4		
144	26.6	1	12.2	2	29.9	1	30.5	2	56.0	2	--	--	--	--		
145	27.0	1	18.0	0	32.0	0	41.0	0	80.0	0	1.61	3	< 20	NR		
146	23.9	4	13.5	4	27.8	3	32.5	4	< 50	NR	1.58	4	--	--		
147	23.1	3	--	--	--	--	32.1	3	53.0	3	--	--	--	--		
151	24.1	4	--	--	23.6	1	33.4	4	32.5	0	1.47	4	7.60	4		
154	23.0	3	13.1	4	26.8	4	36.6	2	46.5	4	2.10	0	--	--		
180	26.8	1	17.6	0	26.8	4	37.1	2	56.5	2	< 1.63	NR	--	--		
185	--	--	--	--	--	--	--	--	--	--	1.69	2	--	--		
190	26.0	2	--	--	26.4	4	33.4	4	56.8	2	--	--	--	--		
191	23.5	4	13.7	4	27.7	4	32.3	4	38.0	1	1.37	2	--	--		

Table 5. Laboratory performance ratings for standard reference sample T-159 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value										
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00										
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00										
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)												
Analyte = Cd (Cadmium)		Co (Cobalt)		Cr (Chromium)		Cu (Copper)		Fe (Iron)		K (Potassium)		Li (Lithium)				
MPV = 24.0 µg/L		13.3 mg/L		26.8 mg/L		33.4 mg/L		48.9 mg/L		1.52 mg/L		8.97 mg/L				
F-pseudosigma = 1.6		0.9		1.8		2.5		6.2		0.13		1.53				
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
193	20.5	0	--	--	22.9	0	34.4	4	71.8	0	1.72	2	--	--		
198	27.5	0	--	--	< 50	NR	< 50	NR	51.0	4	2.01	0	--	--		
203	21.8	2	--	--	4.8	0	33.0	4	48.0	4	1.62	3	--	--		
204	21.6	1	--	--	25.6	3	27.8	0	46.5	4	1.51	4	--	--		
209	--	--	--	--	--	--	--	--	--	--	2.06	0	--	--		
212	23.7	4	12.6	3	27.6	4	30.0	2	51.3	4	1.42	3	--	--		
215	24.0	4	12.0	2	22.0	0	13.0	0	27.0	0	--	--	--	--		
219	24.0	4	14.0	3	28.0	3	33.0	4	--	--	1.40	3	7.90	4		
227	24.8	4	--	--	--	--	37.5	1	52.1	3	--	--	--	--		
234	25.0	3	13.0	4	26.6	4	34.5	4	49.6	4	1.50	4	13.20	0		
236	20.0	0	16.0	0	27.0	4	30.0	2	47.0	4	1.42	3	< 10	NR		
241	25.4	3	--	--	26.4	4	32.5	4	49.0	4	1.78	1	--	--		
247	27.0	1	13.7	4	28.5	3	35.6	3	37.0	1	1.47	4	9.15	4		
254	--	--	14.0	3	--	--	33.2	4	50.4	4	1.58	4	7.30	3		
255	24.5	4	--	--	29.8	1	32.9	4	50.9	4	--	--	--	--		
259	24.0	4	13.0	4	35.2	0	27.0	0	47.8	4	1.47	4	--	--		
265	23.0	3	13.0	4	27.0	4	35.0	3	50.0	4	1.55	4	8.00	4		
268	--	--	--	--	--	--	--	--	--	--	1.96	0	--	--		
273	23.3	4	14.1	3	18.0	0	35.3	3	42.4	2	1.49	4	17.80	0		
274	--	--	--	--	--	--	--	--	--	--	1.70	2	--	--		
277	21.2	1	11.8	1	25.2	3	30.8	2	43.3	3	--	--	--	--		
279	--	--	--	--	--	--	--	--	--	--	1.50	4	--	--		
282	25.0	3	14.9	1	28.9	2	36.1	2	47.7	4	1.63	3	--	--		
287	29.0	0	--	--	28.3	3	35.0	3	40.0	2	1.73	1	--	--		
292	--	--	--	--	25.0	3	34.0	4	46.0	4	1.60	3	--	--		
302	--	--	--	--	--	--	--	--	--	--	1.54	4	--	--		
304	23.0	3	--	--	26.0	4	35.0	3	49.0	4	1.50	4	--	--		
305	11.9	0	--	--	24.4	2	31.6	3	--	--	--	--	--	--		
307	20.8	1	--	--	24.5	2	31.5	3	50.0	4	--	--	--	--		
308	20.7	0	--	--	--	--	37.0	2	--	--	--	--	--	--		
309	--	--	--	--	--	--	--	--	--	--	2.32	0	--	--		
312	--	--	--	--	--	--	--	--	--	--	--	--	--	--		

Table 5. Laboratory performance ratings for standard reference sample T-159 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value								
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00								
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00								
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)										
Analyte = Mg (Magnesium)		Mn (Manganese)		Mo (Molybdenum)		Na (Sodium)		Ni (Nickel)		Pb (Lead)		Sb (Antimony)		
MPV =	5.60	mg/L	22.0	mg/L	41.4	mg/L	100	mg/L	22.2	mg/L	16.6	mg/L	13.9	mg/L
F-pseudosigma =	0.21		1.6		2.6		5		1.6		1.2		1.1	
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	5.62	4	23.5	3	42.0	4	99	4	21.9	4	16.2	4	14.0	4
3	5.58	4	21.4	4	41.6	4	99	4	21.7	4	16.5	4	13.4	4
10	--	--	25.0	1	--	--	--	--	--	--	16.0	3	--	--
11	5.52	4	23.0	3	41.0	4	87	0	22.0	4	15.0	2	15.0	2
12	6.60	0	--	--	--	--	94	2	--	--	--	--	--	--
13	5.55	4	24.2	2	42.9	3	110	1	27.9	0	18.2	2	13.1	3
21	--	--	--	--	--	--	--	--	--	--	--	--	--	--
23	5.40	3	21.0	3	41.1	4	93	2	18.6	0	16.0	3	14.2	4
24	5.49	4	23.1	3	42.5	4	103	3	23.1	3	--	--	--	--
26	3.20	0	23.2	3	41.6	4	85	0	--	--	--	--	--	--
32	5.50	4	21.8	4	39.5	3	97	3	23.9	2	17.9	2	13.9	4
33	5.90	2	20.0	2	--	--	100	4	--	--	--	--	--	--
39	4.97	0	18.5	0	--	--	97	3	22.2	4	16.0	3	14.0	4
42	5.41	3	22.2	4	42.4	4	93	2	22.9	4	17.7	3	14.2	4
45	5.27	2	--	--	--	--	103	3	--	--	--	--	14.6	3
46	5.50	4	21.9	4	39.2	3	104	3	16.9	0	15.7	3	11.9	1
48	6.33	0	21.2	4	44.0	3	113	0	20.6	2	17.6	3	14.7	3
50	--	--	21.7	4	45.6	1	--	--	22.5	4	16.1	4	15.0	2
51	5.76	3	--	--	--	--	101	4	--	--	--	--	--	--
57	5.56	4	21.7	4	35.0	0	97	3	23.7	3	16.1	4	14.0	4
59	6.17	0	24.4	2	41.2	4	106	2	23.9	2	16.7	4	15.9	1
64	4.92	0	--	--	--	--	96	3	--	--	--	--	--	--
70	5.50	4	20.6	3	44.2	2	100	4	21.2	3	18.9	1	13.8	4
76	5.64	4	23.7	2	41.2	4	--	--	22.8	4	17.3	3	--	--
81	5.43	3	22.0	4	38.0	2	99	4	20.0	2	17.0	4	11.0	0
86	5.80	3	21.0	3	34.9	0	100	4	21.3	3	17.4	3	22.0	0
87	5.40	3	< 2.0	0	34.7	0	99	4	28.8	0	17.6	3	--	--
89	5.59	4	18.9	1	--	--	42	0	23.3	3	15.6	3	11.9	1
93	5.17	1	17.4	0	--	--	109	1	19.5	1	15.5	3	11.1	0
96	--	--	20.0	2	--	--	--	--	23.3	3	17.6	3	13.9	4
97	--	--	--	--	45.9	1	98	4	--	--	17.6	3	15.0	2
102	6.95	0	23.9	2	--	--	85	0	26.4	0	18.6	1	10.0	0
107	5.57	4	30.0	0	--	--	102	4	--	--	--	--	--	--
109	5.50	4	21.5	4	--	--	20	0	--	--	--	--	--	--
113	5.78	3	22.9	3	--	--	100	4	20.8	3	18.9	1	12.5	2
114	--	--	21.5	4	--	--	--	--	20.6	2	26.5	0	--	--
121	5.60	4	23.0	3	--	--	100	4	19.9	2	18.7	1	--	--
126	--	--	24.0	2	--	--	102	4	23.3	3	18.0	2	--	--
127	5.55	4	21.8	4	53.5	0	99	4	21.4	3	15.5	3	13.8	4
131	5.60	4	21.0	3	42.0	4	100	4	30.0	0	14.0	0	--	--
134	5.40	3	22.0	4	39.8	3	105	3	22.8	4	16.9	4	13.0	3
138	5.57	4	21.4	4	39.5	3	100	4	21.5	4	16.3	4	15.4	2
140	5.60	4	20.0	2	--	--	101	4	21.4	3	17.1	4	--	--
141	5.75	3	23.5	3	41.0	4	101	4	23.4	3	15.7	3	11.4	0
142	5.95	2	13.0	0	42.3	4	107	2	22.1	4	16.4	4	16.6	0
144	--	--	--	--	--	--	--	--	21.9	4	17.0	4	12.2	2
145	5.85	3	23.0	3	46.0	1	101	4	29.0	0	< 84	NR	--	--
146	5.45	3	22.2	4	40.4	4	95	3	22.6	4	17.0	4	< 15.4	NR
147	--	--	--	--	--	--	--	--	--	--	16.7	4	--	--
151	5.71	4	21.2	4	42.7	4	104	3	21.2	3	16.6	4	13.6	4
154	5.50	4	21.6	4	38.7	2	92	1	22.0	4	16.8	4	12.4	2
180	5.75	3	23.2	3	45.5	1	103	3	28.8	0	< 40.2	NR	< 40.7	NR
185	5.69	4	--	--	--	--	105	3	--	--	--	--	--	--
190	--	--	23.0	3	--	--	--	--	22.8	4	15.9	3	--	--
191	5.78	3	24.0	2	--	--	102	4	22.6	4	16.2	4	--	--

Table 5. Laboratory performance ratings for standard reference sample T-159 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value								
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00								
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00								
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)										
Analyte =	Mg (Magnesium)	Mn (Manganese)	Mo (Molybdenum)	Na (Sodium)	Ni (Nickel)	Pb (Lead)	Sb (Antimony)							
MPV =	5.60 mg/L	22.0 mg/L	41.4 mg/L	100 mg/L	22.2 mg/L	16.6 mg/L	13.9 mg/L							
F-pseudosigma =	0.21	1.6	2.6	5	1.6	1.2	1.1							
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating						
193	5.55	4	--	--	--	--	102	4	< 25	NR	17.6	3	13.4	4
198	5.87	3	30.3	0	--	--	106	2	< 50	NR	16.2	4	14.9	3
203	5.82	3	22.0	4	--	--	97	3	30.0	0	15.8	3	--	--
204	5.73	4	19.1	1	--	--	102	4	17.6	0	14.8	1	13.2	3
209	5.62	4	--	--	--	--	112	0	--	--	--	--	--	--
212	5.50	4	21.7	4	43.7	3	103	3	20.5	2	17.3	3	14.0	4
215	5.60	4	22.0	4	38.0	2	107	2	19.0	0	16.0	3	10.0	0
219	5.00	0	21.0	3	39.0	3	99	4	22.0	4	16.0	3	14.0	4
227	5.46	4	23.2	3	--	--	--	--	--	--	17.0	4	--	--
234	5.73	4	19.5	1	39.7	3	99	4	22.4	4	15.5	3	13.3	3
236	5.43	3	22.0	4	35.0	0	97	3	20.0	2	< 20	NR	64.0	0
241	5.25	2	21.3	4	42.3	4	103	3	21.5	4	16.1	4	13.9	4
247	5.67	4	22.4	4	45.2	2	99	4	24.6	1	19.2	0	14.2	4
254	5.78	3	23.1	3	--	--	105	2	22.3	4	--	--	--	--
255	5.60	4	23.3	3	--	--	--	--	24.9	1	17.2	3	--	--
259	5.55	4	20.5	3	38.4	2	101	4	22.7	4	16.7	4	--	--
265	5.75	3	23.5	3	42.0	4	100	4	22.0	4	16.5	4	13.0	3
268	5.80	3	--	--	--	--	72	0	--	--	--	--	--	--
273	6.10	1	23.6	3	--	--	110	1	35.7	0	14.4	1	--	--
274	5.03	0	--	--	--	--	117	0	--	--	--	--	--	--
277	--	--	18.3	0	--	--	--	--	20.7	3	11.2	0	--	--
279	5.80	3	--	--	--	--	93	2	--	--	--	--	--	--
282	5.78	3	23.3	3	42.5	4	101	4	22.0	4	16.0	3	13.1	3
287	4.99	0	23.0	3	--	--	94	2	--	--	19.8	0	--	--
292	5.50	4	21.0	3	37.0	1	97	3	25.0	1	16.0	3	--	--
302	6.40	0	--	--	--	--	109	1	--	--	--	--	--	--
304	5.60	4	--	--	--	--	--	--	23.0	3	28.0	0	--	--
305	--	--	--	--	--	--	--	--	20.5	2	14.2	1	--	--
307	6.25	0	12.0	0	--	--	100	4	22.5	4	16.5	4	--	--
308	5.36	3	24.7	1	--	--	--	--	--	--	19.2	0	--	--
309	--	--	5.5	0	--	--	40	0	--	--	29.0	0	--	--
312	--	--	--	--	--	--	--	--	--	--	--	--	--	--

Table 5. Laboratory performance ratings for standard reference sample T-159 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value									
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00									
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00									
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)											
Analyte =	Se (Selenium)	SiO ₂ (Silica)		Sr (Strontium)		Tl (Thallium)		U (Uranium)		V (Vanadium)		Zn (Zinc)			
MPV =	5.49 mg/L	11.5 mg/L		190 mg/L		13.8 mg/L		5.04 mg/L		14.4 mg/L		19.2 mg/L			
F-pseudostigma =	0.83	0.7		7		1.6		0.32		1.7		1.9			
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	
1	5.40	4	11.8	4	186	4	13.3	4	4.81	3	17.0	1	18.8	4	
3	7.22	0	10.9	3	179	2	12.7	3	--	--	14.1	4	17.6	3	
10	4.60	2	--	--	--	--	--	--	--	--	--	--	22.0	1	
11	6.00	3	12.5	2	200	2	12.0	2	--	--	14.0	4	< 25	NR	
12	--	--	--	--	--	--	--	--	--	--	--	--	35.0	0	
13	8.00	0	11.4	4	--	--	13.9	4	--	--	13.2	3	17.3	2	
21	--	--	--	--	--	--	--	--	--	--	--	--	--	--	
23	4.90	3	--	--	--	--	9.7	0	--	--	--	--	18.8	4	
24	--	--	12.1	3	196	3	--	--	--	--	--	--	18.8	4	
26	--	--	12.0	3	--	--	--	--	--	--	15.6	3	18.6	4	
32	5.40	4	11.8	4	189	4	14.0	4	5.35	3	13.6	4	19.0	4	
33	--	--	11.2	4	206	1	--	--	--	--	--	--	--	--	
39	5.10	4	--	--	--	--	18.2	0	--	--	10.8	0	21.9	2	
42	6.19	3	10.7	2	187	4	14.9	3	--	--	15.1	4	20.2	3	
45	--	--	--	--	--	--	--	--	--	--	--	--	21.0	3	
46	5.61	4	--	--	--	--	--	--	--	--	15.2	3	17.3	2	
48	4.90	3	--	--	--	--	13.7	4	--	--	13.9	4	16.8	2	
50	5.40	4	--	--	195	3	13.9	4	--	--	14.7	4	18.0	3	
51	--	--	--	--	--	--	--	--	--	--	--	--	--	--	
57	4.30	2	11.2	4	190	4	14.6	3	--	--	14.0	4	21.5	2	
59	5.90	4	--	--	187	4	14.9	3	--	--	15.4	3	20.2	3	
64	--	--	11.4	4	--	--	--	--	--	--	--	--	--	--	
70	5.55	4	12.1	3	201	2	15.7	2	5.60	1	13.6	4	17.3	2	
76	--	--	12.1	3	--	--	13.5	4	5.12	4	--	--	--	--	
81	4.00	1	--	--	181	3	15.0	3	--	--	13.0	3	22.0	1	
86	4.67	3	--	--	192	4	9.5	0	--	--	16.9	1	17.1	2	
87	< 2	0	--	--	--	--	--	--	--	--	--	--	23.7	0	
89	2.00	0	11.0	3	--	--	25.2	0	--	--	16.3	2	18.0	3	
93	4.10	1	--	--	--	--	--	--	--	--	13.0	3	17.2	2	
96	5.50	4	--	--	--	--	--	--	--	--	16.9	1	< 10	0	
97	--	--	11.4	4	--	--	15.7	2	--	--	18.0	0	--	--	
102	3.50	0	--	--	197	3	--	--	--	--	16.6	2	11.0	0	
107	--	--	--	--	--	--	--	--	--	--	--	--	--	--	
109	--	--	--	--	--	--	--	--	--	--	--	--	--	--	
113	5.19	4	--	--	189	4	13.8	4	--	--	--	--	20.4	3	
114	--	--	--	--	--	--	34.3	0	--	--	--	--	19.5	4	
121	--	--	--	--	190	4	--	--	4.80	3	12.2	2	19.0	4	
126	--	--	--	--	--	--	9.8	0	--	--	--	--	--	--	
127	4.99	3	11.4	4	181	3	12.3	3	< 200	NR	14.2	4	18.7	4	
131	< 60	NR	10.6	2	163	0	--	--	--	--	--	--	20.5	3	
134	5.80	4	11.3	4	183	3	13.2	4	5.00	4	13.9	4	20.2	3	
138	5.47	4	--	--	183	3	13.9	4	--	--	14.0	4	19.0	4	
140	--	--	12.4	2	--	--	--	--	--	--	--	--	20.0	4	
141	6.60	2	--	--	201	2	13.6	4	--	--	14.3	4	20.9	3	
142	6.04	3	11.4	4	198	3	13.6	4	5.08	4	13.8	4	14.8	0	
144	5.60	4	--	--	--	--	15.0	3	--	--	--	--	20.0	4	
145	--	--	11.8	4	192	4	--	--	--	--	20.0	0	23.0	0	
146	5.28	4	--	--	--	--	11.7	2	--	--	15.2	3	20.0	4	
147	--	--	--	--	--	--	--	--	--	--	--	--	18.5	4	
151	5.50	4	--	--	188	4	13.6	4	--	--	--	--	19.2	4	
154	6.50	2	10.9	3	175	1	12.9	3	--	--	16.0	3	16.5	2	
180	< 67.9	NR	--	--	--	--	< 54.9	NR	--	--	19.3	0	< 24.4	NR	
185	--	--	--	--	--	--	--	--	--	--	--	--	--	--	
190	4.81	3	12.3	2	--	--	--	--	--	--	--	--	19.9	4	
191	6.80	1	11.7	4	191	4	12.7	3	5.10	4	--	--	18.3	4	

Table 5. Laboratory performance ratings for standard reference sample T-159 (trace constituents)--Continued

[MPV, most probable value; µg/L, micrograms per liter; mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/28, number of reported values of 28 possible values; RV, reported value; <, less than; NR, not rated --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value										
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00										
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00										
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)												
Analyte =	Se (Selenium)	SiO ₂ (Silica)	Sr (Strontium)	Tl (Thallium)	U (Uranium)	V (Vanadium)	Zn (Zinc)									
MPV =	5.49 mg/L	11.5 mg/L	190 mg/L	13.8 mg/L	5.04 mg/L	14.4 mg/L	19.2 mg/L									
F-pseudostigma =	0.83	0.7	7	1.6	0.32	1.7	1.9									
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
193	4.57	2	--	--	--	--	--	--	--	--	--	--	< 25	NR		
198	<10	NR	--	--	--	--	14.3	4	--	--	--	--	< 50	NR		
203	5.30	4	11.7	4	--	--	--	--	--	--	--	--	18.0	3		
204	5.30	4	--	--	--	--	12.1	2	--	--	--	--	16.5	2		
209	--	--	--	--	--	--	--	--	--	--	--	--	--	--		
212	5.21	4	12.2	2	189	4	13.9	4	--	--	11.7	1	19.4	4		
215	6.00	3	--	--	--	--	--	--	--	--	24.0	0	38.0	0		
219	5.50	4	11.0	3	193	4	14.0	4	4.80	3	14.0	4	20.0	4		
227	--	--	5.8	0	--	--	--	--	--	--	--	--	20.5	3		
234	6.60	2	11.6	4	182	3	15.1	3	--	--	14.4	4	22.4	1		
236	<10	NR	6.0	0	192	4	--	--	--	--	< 10	0	16.0	1		
241	6.55	2	11.1	4	--	--	12.8	3	--	--	16.2	2	22.4	1		
247	6.19	3	13.0	0	192	4	15.7	2	--	--	15.0	4	23.3	0		
254	--	--	11.9	3	199	3	--	--	4.16	0	--	--	20.2	3		
255	7.15	1	--	--	--	--	--	--	--	--	--	--	20.2	3		
259	--	--	11.7	4	190	4	--	--	--	--	--	--	18.2	3		
265	9.00	0	5.8	0	190	4	13.0	4	4.70	2	14.0	4	18.0	3		
268	--	--	--	--	--	--	--	--	--	--	--	--	--	--		
273	--	--	12.8	1	211	0	--	--	--	--	--	--	18.7	4		
274	--	--	10.6	2	--	--	--	--	--	--	--	--	--	--		
277	1.07	0	--	--	--	--	--	--	--	--	--	--	9.5	0		
279	--	--	--	--	--	--	--	--	--	--	--	--	--	--		
282	5.60	4	5.5	0	--	--	8.5	0	6.80	0	14.4	4	19.8	4		
287	5.20	4	--	--	--	--	--	--	--	--	--	--	18.0	3		
292	6.00	3	--	--	--	--	--	--	--	--	--	--	19.0	4		
302	--	--	--	--	--	--	--	--	--	--	--	--	--	--		
304	--	--	11.5	4	--	--	--	--	--	--	12.0	2	19.9	4		
305	--	--	--	--	--	--	--	--	--	--	--	--	15.4	0		
307	4.82	3	--	--	--	--	--	--	--	--	--	--	25.0	0		
308	--	--	--	--	--	--	--	--	--	--	--	--	26.5	0		
309	--	--	5.9	0	--	--	--	--	--	--	--	--	--	--		
312	--	--	15.1	0	--	--	--	--	--	--	--	--	--	--		

Table 6. Laboratory performance ratings for standard reference sample M-152 (major constituents)

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/15, number of reported values of 15 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte = Alkalinity ¹				B (Boron)		Ca (Calcium)		Cl (Chloride)		DSRD (Dissolved Solids)	
	MPV = 6.00 mg/L		RV		92.2 µg/L		3.63 mg/L		17.3 mg/L		66.5 mg/L	
	F_pseudosigma = 1.49	OLR	V/15	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV
1	3.9	14	5.86	NR	93.0	4	3.69	4	16.9	4	66.0	4
3	3.4	14	6.00	NR	91.0	4	3.69	4	18.4	2	71.0	3
10	3.5	11	4.10	NR	--	--	3.60	4	17.5	4	66.0	4
11	2.9	15	5.00	NR	93.0	4	3.77	3	18.0	3	68.0	4
12	2.0	6	--	--	--	--	--	--	14.0	0	72.0	3
13	2.8	13	7.55	NR	--	--	3.59	4	16.7	3	61.3	3
23	3.3	6	6.10	NR	--	--	--	--	17.5	4	--	--
24	2.8	12	6.74	NR	97.0	3	3.88	2	17.0	4	--	--
25	1.7	7	11.00	NR	--	--	--	--	17.8	3	36.0	0
26	2.1	12	8.40	NR	85.6	3	2.60	0	19.1	1	66.0	4
32	3.8	8	--	--	91.5	4	3.70	4	--	--	--	--
33	3.8	11	3.74	NR	--	--	3.70	4	16.8	3	--	--
38	3.7	9	6.34	NR	--	--	3.58	4	--	--	63.0	4
39	2.1	12	6.50	NR	74.2	0	3.31	1	20.0	0	--	--
42	2.6	14	7.56	NR	91.4	4	3.55	4	17.5	4	--	--
43	3.5	4	< 20	NR	--	--	--	--	17.0	4	74.0	3
45	2.5	10	< 20	NR	--	--	3.41	2	18.5	2	70.0	4
46	2.1	12	0.22	NR	92.2	4	3.58	4	--	--	106.0	0
48	2.7	11	< 5	NR	94.0	4	3.98	1	18.0	3	71.0	3
50	3.1	12	4.99	NR	96.7	3	3.79	3	16.6	3	63.0	4
51	3.0	10	4.30	NR	--	--	3.31	1	17.9	3	74.0	3
57	2.6	13	6.00	NR	96.9	3	3.63	4	22.0	0	--	--
59	3.5	2	4.60	NR	--	--	--	--	--	--	--	--
64	3.7	10	--	--	--	--	3.67	4	17.5	4	--	--
70	2.4	14	7.00	NR	129.0	0	3.98	1	17.3	4	70.0	4
76	3.8	8	--	--	--	--	3.58	4	17.4	4	72.0	3
81	2.0	14	5.15	NR	--	--	3.36	1	17.2	4	40.0	0
85	2.9	12	4.75	NR	--	--	4.19	0	18.4	2	55.0	2
86	3.4	13	--	--	81.2	2	3.60	4	17.3	4	--	--
87	1.6	9	--	--	--	--	2.40	0	16.0	2	88.0	0
89	2.2	13	6.10	NR	--	--	2.39	0	16.8	3	72.0	3
93	2.8	11	4.81	NR	--	--	3.58	4	19.0	1	--	--
96	2.8	6	4.50	NR	--	--	--	--	16.0	2	71.0	3
97	2.8	11	6.30	NR	--	--	--	--	16.4	2	74.0	3
102	1.7	12	--	--	--	--	3.97	1	17.7	4	--	--
107	4.0	3	10.80	NR	--	--	--	--	17.3	4	--	--
109	2.4	10	7.91	NR	--	--	3.80	3	18.0	3	55.0	2
113	3.3	12	6.00	NR	--	--	3.75	3	17.9	3	55.0	2
114	2.3	7	5.00	NR	--	--	--	--	16.6	3	46.0	0
121	3.3	7	--	--	--	--	3.60	4	--	--	--	--
127	3.9	15	6.45	NR	93.6	4	3.61	4	17.4	4	64.5	4
131	2.2	11	--	--	90.0	4	3.60	4	18.3	2	--	--
134	3.8	15	6.83	NR	92.1	4	3.72	3	17.1	4	68.4	4
138	3.7	15	6.50	NR	93.8	4	3.67	4	17.3	4	70.0	4
140	1.9	12	--	--	--	--	3.60	4	15.2	0	100.0	0
141	1.8	13	5.95	NR	81.5	2	3.81	2	18.2	3	44.5	0
142	2.3	13	7.00	NR	97.9	3	2.98	0	18.2	3	64.0	4
143	3.4	5	4.00	NR	--	--	--	--	16.8	3	72.0	3
145	3.0	14	4.80	NR	104.0	1	3.85	2	17.7	4	--	--
146	2.1	9	4.66	NR	--	--	3.35	1	17.2	4	44.0	0

¹ Laboratories were not rated for alkalinity because the long term stability needs to be verified. The MPV for alkalinity is provided as an estimated value.

Table 6. Laboratory performance ratings for standard reference sample M-152 (major constituents)--Continued

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/15, number of reported values of 15 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Lab	Analyte = Alkalinity ¹				B (Boron)		Ca (Calcium)		Cl (Chloride)		DSRD (Dissolved Solids)	
	OLR	V/15	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
			MPV = 6.00 mg/L		92.2 µg/L		3.63 mg/L		17.3 mg/L		66.5 mg/L	
			F_pseudosigma = 1.49		7.8		0.18		0.9		8.2	
151	3.3	12	6.00	NR	--	--	3.88	2	17.1	4	61.0	3
154	3.1	13	10.00	NR	--	--	3.60	4	17.2	4	--	--
180	2.6	10	6.90	NR	< 12.6	0	3.80	3	18.5	2	--	--
183	2.2	5	--	--	--	--	--	--	21.2	0	--	--
185	3.4	9	5.05	NR	--	--	3.61	4	17.5	4	--	--
190	2.5	12	5.60	NR	--	--	5.24	0	16.2	2	73.5	3
191	3.2	10	--	--	--	--	3.79	3	17.4	4	--	--
193	2.8	6	4.00	NR	--	--	3.27	1	--	--	--	--
196	3.0	3	--	--	--	--	--	--	16.7	3	--	--
203	2.9	10	3.80	NR	--	--	3.60	4	17.6	4	--	--
204	4.0	1	3.65	NR	--	--	--	--	17.0	4	--	--
208	3.5	2	--	--	--	--	--	--	17.3	4	--	--
209	2.9	7	--	--	--	--	3.06	0	18.7	1	--	--
212	2.9	15	6.02	NR	91.9	4	3.55	4	16.7	3	78.0	2
213	3.3	3	5.00	NR	--	--	--	--	16.3	2	--	--
215	1.9	12	6.00	NR	80.0	1	3.70	4	18.7	1	66.0	4
219	2.3	9	--	--	--	--	3.63	4	16.0	2	--	--
227	3.5	8	6.00	NR	--	--	3.48	3	16.5	3	61.0	3
234	3.5	15	7.80	NR	84.6	3	3.65	4	17.6	4	67.0	4
236	2.4	12	6.40	NR	83.0	2	3.47	3	--	--	72.0	3
241	3.4	11	5.00	NR	--	--	3.80	3	17.7	4	--	--
243	2.7	3	--	--	--	--	--	--	--	--	--	--
244	3.5	2	13.10	NR	--	--	--	--	--	--	--	--
247	2.7	15	5.98	NR	143.0	0	3.58	4	17.4	4	63.0	4
253	2.2	5	3.60	NR	--	--	--	--	17.0	4	--	--
254	3.5	8	--	--	--	--	3.78	3	17.4	4	--	--
255	3.3	6	--	--	93.8	4	3.64	4	--	--	--	--
259	3.5	12	4.00	NR	85.0	3	3.70	4	17.6	4	--	--
265	3.7	12	--	--	94.0	4	3.65	4	17.8	3	--	--
268	3.0	8	24.40	NR	--	--	2.95	0	16.6	3	--	--
273	1.0	13	9.10	NR	61.4	0	4.34	0	15.7	1	66.0	4
274	1.4	11	9.86	NR	--	--	2.37	0	34.4	0	--	--
275	1.4	11	10.00	NR	--	--	3.00	0	20.0	0	60.0	3
277	1.5	10	5.35	NR	--	--	3.73	3	14.9	0	--	--
279	2.5	8	--	--	--	--	3.00	0	17.0	4	--	--
282	2.3	14	5.08	NR	97.5	3	3.58	4	15.8	1	102.0	0
287	1.6	8	7.00	NR	--	--	2.58	0	16.2	2	--	--
292	3.0	10	6.50	NR	--	--	5.40	0	17.3	4	--	--
302	1.7	9	6.00	NR	--	--	3.91	1	15.3	0	--	--
304	3.6	7	--	--	--	--	3.70	4	--	--	--	--
305	1.8	4	--	--	--	--	--	--	--	--	31.0	0
307	2.3	8	6.00	NR	--	--	4.95	0	19.0	1	--	--
308	4.0	1	--	--	--	--	--	--	--	--	--	--
309	0.2	6	11.60	NR	--	--	3.91	1	--	--	--	--
312	1.3	3	11.50	NR	--	--	--	--	--	--	--	--

¹ Laboratories were not rated for alkalinity because the long term stability needs to be verified. The MPV for alkalinity is provided as an estimated value.

Table 6. Laboratory performance ratings for standard reference sample M-152 (major constituents)--Continued

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/15, number of reported values of 15 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = F (Fluoride)		K (Potassium)		Mg (Magnesium)		Na (Sodium)		(total Phosphorus) as P		
MPV =	1.98 mg/L		0.92 mg/L		0.722 mg/L		16.3 mg/L		0.386	
F-pseudostigma =	0.12		0.09		0.047		0.9		0.024	
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	1.98	4	0.89	4	0.730	4	16.9	3	--	--
3	2.04	4	1.39	0	0.719	4	16.1	4	0.378	4
10	2.05	3	0.92	4	0.750	3	16.1	4	--	--
11	1.60	0	0.64	0	0.700	4	15.5	3	0.390	4
12	--	--	--	--	--	--	14.0	0	0.420	2
13	1.83	2	0.69	0	0.740	4	17.2	2	0.366	3
23	2.08	3	--	--	--	--	--	--	0.348	1
24	--	--	0.96	4	0.740	4	16.4	4	--	--
25	1.64	0	--	--	--	--	--	--	0.390	4
26	--	--	0.57	0	0.450	0	13.2	0	--	--
32	--	--	0.96	4	0.730	4	16.7	4	--	--
33	1.88	3	0.88	4	0.710	4	16.1	4	--	--
38	--	--	0.92	4	0.746	3	15.3	2	0.381	4
39	1.98	4	0.89	4	0.591	0	15.6	3	0.370	3
42	1.96	4	0.89	4	0.616	0	15.5	3	0.316	0
43	--	--	--	--	--	--	--	--	--	--
45	2.00	4	1.02	2	0.630	1	18.2	0	--	--
46	2.03	4	23.00	0	0.692	3	16.6	4	0.416	2
48	--	--	0.95	4	0.770	2	17.0	3	--	--
50	1.93	4	1.04	2	0.740	4	17.0	3	--	--
51	--	--	1.19	0	0.730	4	16.0	4	0.385	4
57	1.70	0	1.15	0	0.736	4	15.4	3	0.410	2
59	2.00	4	--	--	--	--	--	--	--	--
64	--	--	0.89	4	0.700	4	15.8	4	0.400	3
70	1.96	4	1.05	2	0.746	3	17.0	3	0.300	0
76	2.04	4	--	--	0.743	4	15.6	3	--	--
81	1.98	4	0.80	2	0.623	0	15.3	2	0.391	4
85	1.99	4	0.98	3	0.734	4	16.3	4	0.409	3
86	2.06	3	0.90	4	0.759	3	16.4	4	0.388	4
87	--	--	0.98	3	0.600	0	16.6	4	0.393	4
89	1.85	2	0.88	4	0.613	0	13.7	0	0.387	4
93	0.92	0	0.79	2	0.720	4	15.4	3	0.430	1
96	2.07	3	--	--	--	--	--	--	--	--
97	2.09	3	0.88	4	--	--	16.0	4	0.380	4
102	3.17	0	0.52	0	0.940	0	50.7	0	0.371	3
107	2.00	4	--	--	--	--	--	--	--	--
109	1.95	4	0.77	1	0.650	1	16.0	4	--	--
113	2.00	4	1.02	2	0.784	2	16.9	3	0.376	4
114	1.93	4	--	--	--	--	--	--	0.370	3
121	--	--	0.94	4	0.700	4	15.6	3	--	--
127	2.05	3	0.84	3	0.745	4	16.1	4	0.394	4
131	2.21	1	1.00	3	0.640	1	15.0	2	0.530	0
134	1.96	4	0.90	4	0.709	4	16.8	3	0.390	4
138	1.95	4	0.87	3	0.706	4	16.3	4	0.401	3
140	2.03	4	0.83	2	0.700	4	16.4	4	0.260	0
141	2.10	3	1.02	2	0.806	1	17.9	1	--	--
142	2.09	3	< 1	NR	< 1	NR	16.9	3	0.173	0
143	--	--	--	--	--	--	--	--	0.375	4
145	2.00	4	1.01	3	0.790	2	16.5	4	0.390	4
146	2.06	3	< 0.860	NR	0.675	2	15.2	2	--	--

Table 6. Laboratory performance ratings for standard reference sample M-152 (major constituents)--Continued

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/15, number of reported values of 15 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = F (Fluoride)	K (Potassium)		Mg (Magnesium)		Na (Sodium)		(total Phosphorus) as P			
MPV =	1.98 mg/L	0.92 mg/L	0.722 mg/L	16.3 mg/L	0.386					
F-pseudostigma =	0.12	0.09	0.047	0.9	0.024					
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
151	2.04	4	0.92	4	0.830	0	16.4	4	--	--
154	1.88	3	1.00	3	0.730	4	15.5	3	0.450	0
180	2.08	3	< 1.63	NR	0.735	4	16.8	3	--	--
183	2.00	4	--	--	--	--	--	--	0.371	3
185	--	--	0.90	4	0.715	4	16.9	3	0.382	4
190	1.96	4	0.91	4	0.910	0	16.2	4	0.367	3
191	1.99	4	0.83	3	0.770	2	17.0	3	0.360	2
193	--	--	0.95	4	0.677	3	16.5	4	--	--
196	1.80	2	--	--	--	--	--	--	--	--
203	--	--	0.92	4	0.620	0	15.0	2	0.374	4
204	--	--	--	--	--	--	--	--	--	--
208	--	--	--	--	--	--	--	--	--	--
209	--	--	0.95	4	0.683	3	16.2	4	--	--
212	1.71	0	0.85	3	0.731	4	15.5	3	0.302	0
213	--	--	--	--	--	--	--	--	0.390	4
215	1.87	3	--	--	0.660	2	17.2	2	0.420	2
219	1.63	0	0.91	4	0.700	4	16.0	4	--	--
227	--	--	--	--	0.698	3	--	--	0.376	4
234	2.18	1	0.81	2	0.722	4	16.2	4	0.380	4
236	--	--	0.84	3	0.670	2	15.6	3	0.410	2
241	1.90	3	0.70	0	0.720	4	16.5	4	--	--
243	--	--	--	--	--	--	--	--	0.320	0
244	--	--	--	--	--	--	--	--	--	--
247	1.73	1	0.94	4	0.737	4	16.3	4	0.327	0
253	--	--	--	--	--	--	--	--	0.453	0
254	--	--	0.95	4	0.740	4	17.0	3	--	--
255	2.04	4	--	--	0.716	4	--	--	0.416	2
259	2.00	4	0.66	0	0.700	4	16.5	4	--	--
265	2.10	3	0.88	4	0.740	4	16.2	4	--	--
268	--	--	1.10	1	0.740	4	16.2	4	--	--
273	1.48	0	1.03	2	0.810	1	18.8	0	0.260	0
274	1.79	1	0.98	3	0.950	0	12.3	0	0.390	4
275	2.00	4	1.00	3	0.800	1	15.0	2	0.050	0
277	1.00	0	0.98	3	1.100	0	14.0	0	0.280	0
279	--	--	0.94	4	0.600	0	16.8	3	--	--
282	2.14	2	1.25	0	0.747	3	16.6	4	0.405	3
287	1.79	1	0.82	2	0.641	1	14.3	0	--	--
292	1.92	4	1.00	3	0.600	0	16.5	4	0.400	3
302	1.94	4	1.05	2	0.899	0	18.2	0	--	--
304	--	--	0.90	4	0.690	3	16.7	4	--	--
305	1.72	0	--	--	--	--	--	--	0.386	4
307	--	--	--	--	0.750	3	16.5	4	0.410	2
308	--	--	--	--	--	--	--	--	--	--
309	--	--	1.44	0	0.520	0	5.1	0	--	--
312	--	--	--	--	--	--	--	--	--	--

Table 6. Laboratory performance ratings for standard reference sample M-152 (major constituents)--Continued

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/15, number of reported values of 15 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = pH		SiO ₂ (Silica)		SO ₄ (Sulfate)		Sp Cond		Sr (Strontium)		V (Vanadium)		
MPV =	6.60	4.50 mg/L		3.76 mg/L		113 µS/cm		18.3 µg/L		30.0 µg/L		
F-pseudostigma =	0.22	0.39		0.33		3		0.9		1.9		
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	6.71	4	4.58	4	3.71	4	114	4	17.8	3	30.5	4
3	6.69	4	4.12	3	< 10	NR	113	4	17.9	4	28.6	3
10	6.67	4	4.87	3	4.10	2	113	4	--	--	--	--
11	6.92	3	4.53	4	4.00	3	110	3	19.0	3	29.0	3
12	6.35	3	--	--	--	--	114	4	--	--	--	--
13	6.83	3	22.50	0	3.85	4	115	4	--	--	30.6	4
23	6.64	4	--	--	3.72	4	111	4	--	--	--	--
24	6.60	4	4.68	4	4.58	0	106	2	22.0	0	31.7	3
25	6.80	3	--	--	5.00	0	120	2	--	--	--	--
26	7.03	2	4.40	4	3.90	4	112	4	--	--	31.6	3
32	--	--	4.80	3	--	--	--	--	18.9	3	30.2	4
33	6.72	4	4.56	4	3.78	4	115	4	18.6	4	--	--
38	6.70	4	4.57	4	--	--	112	4	--	--	--	--
39	6.59	4	4.09	2	3.60	4	--	--	--	--	19.9	0
42	6.37	3	4.09	2	4.20	2	119	2	17.1	2	31.7	3
43	6.34	3	--	--	< 10	NR	113	4	--	--	--	--
45	6.57	4	--	--	4.21	2	114	4	--	--	--	--
46	7.43	0	--	--	2.06	0	207	0	--	--	30.0	4
48	6.40	3	--	--	3.00	0	108	3	--	--	30.0	4
50	6.30	3	4.41	4	3.02	0	113	4	--	--	--	--
51	6.58	4	--	--	3.93	3	113	4	--	--	--	--
57	6.70	4	4.30	3	< 5	NR	110	3	18.7	4	29.1	4
59	--	--	4.80	3	--	--	--	--	--	--	--	--
64	6.80	3	4.15	3	3.82	4	112	4	--	--	--	--
70	5.98	1	4.67	4	3.82	4	109	3	21.6	0	< 50	NR
76	--	--	--	--	3.79	4	113	4	--	--	--	--
81	6.58	4	4.79	3	3.06	0	112	4	16.0	0	26.0	0
85	6.80	3	4.80	3	3.60	4	110	3	--	--	--	--
86	6.62	4	--	--	3.74	4	107	2	17.8	3	31.5	3
87	6.01	1	--	--	< 5	NR	70	0	--	--	--	--
89	5.58	0	4.20	3	3.73	4	118	3	--	--	32.1	2
93	6.68	4	--	--	3.80	4	111	4	--	--	29.1	4
96	6.75	4	--	--	4.20	2	116	3	--	--	--	--
97	6.74	4	4.45	4	2.87	0	117	3	--	--	38.7	0
102	--	--	4.84	3	3.44	3	117	3	20.5	0	31.7	3
107	--	--	4.41	4	--	--	--	--	--	--	--	--
109	6.00	1	--	--	3.29	2	116	3	--	--	--	--
113	6.54	4	4.38	4	--	--	114	4	18.1	4	--	--
114	6.41	3	--	--	5.69	0	117	3	--	--	--	--
121	--	--	4.49	4	--	--	--	--	18.0	4	36.0	0
127	6.60	4	4.43	4	3.66	4	114	4	18.2	4	29.9	4
131	--	--	4.10	2	4.12	2	--	--	17.5	3	--	--
134	6.64	4	4.37	4	3.48	3	113	4	18.3	4	29.7	4
138	6.80	3	4.66	4	3.78	4	113	4	17.4	3	29.0	3
140	5.31	0	4.93	2	1.00	0	110	3	--	--	--	--
141	5.96	1	--	--	3.70	4	100	0	20.6	0	30.6	4
142	6.18	2	3.86	1	3.93	3	115	4	21.9	0	30.0	4
143	6.58	4	--	--	--	--	108	3	--	--	--	--
145	6.75	4	4.26	3	3.69	4	111	4	19.0	3	35.0	0
146	6.72	4	--	--	< 5	NR	127	0	--	--	28.3	3

Table 6. Laboratory performance ratings for standard reference sample M-152 (major constituents)--Continued

[MPV, most probable value; mg/L, micrograms per liter; mg/L, milligrams per liter; mS/cm, microsiemens per centimeter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/15, number of reported values of 15 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = pH	SiO ₂ (Silica)		SO ₄ (Sulfate)		Sp Cond		Sr (Strontium)		V (Vanadium)			
MPV =	6.60	4.50 mg/L		3.76 mg/L		113 μS/cm		18.3 μg/L		30.0 μg/L		
F-pseudostigma =	0.22	0.39		0.33		3		0.9		1.9		
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating		
151	6.65	4	4.49	4	3.76	4	112	4	17.4	3	--	--
154	6.62	4	4.20	3	4.59	0	114	4	18.0	4	30.0	4
180	6.60	4	--	--	3.91	4	108	3	--	--	63.8	0
183	6.92	3	--	--	--	--	103	1	--	--	--	--
185	6.52	4	--	--	3.78	4	132	0	--	--	--	--
190	6.43	3	4.86	3	3.20	1	116	3	--	--	--	--
191	--	--	4.22	3	3.74	4	--	--	18.0	4	--	--
193	--	--	4.91	2	--	--	117	3	--	--	--	--
196	--	--	--	--	3.74	4	--	--	--	--	--	--
203	6.55	4	5.13	1	3.31	2	114	4	--	--	--	--
204	--	--	--	--	--	--	--	--	--	--	--	--
208	--	--	--	--	4.00	3	--	--	--	--	--	--
209	6.70	4	--	--	3.69	4	--	--	--	--	--	--
212	6.71	4	4.66	4	3.96	3	118	3	18.8	3	28.9	3
213	6.51	4	--	--	--	--	--	--	--	--	--	--
215	6.60	4	--	--	6.00	0	98	0	--	--	24.0	0
219	--	--	4.00	2	4.77	0	--	--	19.7	1	--	--
227	6.56	4	--	--	3.84	4	111	4	--	--	--	--
234	6.44	4	4.55	4	4.05	3	114	4	18.2	4	28.1	3
236	6.23	2	2.30	0	--	--	113	4	18.0	4	27.0	1
241	6.50	4	4.51	4	3.64	4	110	3	--	--	30.2	4
243	6.74	4	--	--	--	--	114	4	--	--	--	--
244	6.77	3	--	--	--	--	112	4	--	--	--	--
247	6.76	4	4.60	4	3.01	0	113	4	17.0	2	33.0	1
253	6.81	3	--	--	< 1	0	113	4	--	--	--	--
254	--	--	4.59	4	3.80	4	--	--	19.5	2	--	--
255	--	--	--	--	3.40	2	--	--	--	--	--	--
259	6.50	4	4.48	4	3.80	4	114	4	19.0	3	--	--
265	7.00	2	4.50	4	3.70	4	--	--	18.5	4	30.0	4
268	6.46	4	--	--	3.77	4	112	4	--	--	--	--
273	7.10	1	6.83	0	8.54	0	115	4	--	--	--	--
274	5.70	0	4.11	3	0.32	0	114	4	--	--	--	--
275	6.00	1	--	--	3.10	1	125	0	--	--	--	--
277	6.31	3	--	--	3.81	4	120	2	--	--	--	--
279	6.24	2	--	--	3.68	4	108	3	--	--	--	--
282	6.47	4	2.17	0	3.79	4	127	0	--	--	29.2	4
287	6.76	4	--	--	3.46	3	--	--	--	--	--	--
292	6.75	4	--	--	3.70	4	113	4	--	--	--	--
302	6.76	4	--	--	1.64	0	115	4	--	--	--	--
304	--	--	4.30	3	--	--	--	--	17.5	3	30.0	4
305	6.82	3	--	--	--	--	--	--	--	--	--	--
307	6.56	4	8.00	0	--	--	113	4	--	--	--	--
308	6.51	4	--	--	--	--	--	--	--	--	--	--
309	7.74	0	--	--	1.83	0	--	--	--	--	--	--
312	6.51	4	5.54	0	--	--	100	0	--	--	--	--

Table 7. Laboratory performance ratings for standard reference sample N-63 (nutrient constituents)

[MPV, most probable value mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/5, number of reported values of 5 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value						
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00						
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00						
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)								
Analyte = NH ₃ as N (Ammonia)			NH ₃ + Org N as N (Ammonia + Organic N)			NO ₃ as N (Nitrate)		total P as P (total Phosphorus)		PO ₄ as P (Orthophosphate as P)		
MPV = 0.150 mg/L			0.200 mg/L			0.084 mg/L		0.158 mg/L		0.142 mg/L		
F-pseudosigma = 0.021			0.066			0.008		0.010		0.010		
Lab	OLR	V/5	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	4.0	5	0.149	4	0.188	4	0.084	4	0.158	4	0.142	4
10	4.0	5	0.140	4	0.190	4	0.080	4	0.158	4	0.145	4
11	1.2	5	0.430	0	0.300	1	0.080	4	0.140	1	0.190	0
12	2.0	3	--	--	--	--	0.090	3	0.130	0	0.134	3
13	2.0	4	0.123	2	--	--	0.072	1	0.162	4	0.158	1
18	2.6	5	0.262	0	0.267	2	0.077	3	0.157	4	0.143	4
21	3.8	5	0.151	4	0.164	3	0.085	4	0.157	4	0.140	4
23	1.8	5	0.152	4	0.700	0	0.085	4	0.271	0	0.127	1
25	3.8	4	0.130	3	--	--	0.080	4	0.160	4	0.138	4
33	2.7	3	0.150	4	--	--	0.070	1	--	--	0.150	3
36	1.4	5	1.227	0	0.260	3	0.086	4	0.180	0	0.120	0
38	4.0	5	0.150	4	0.170	4	0.084	4	0.163	4	0.139	4
42	1.5	2	--	--	--	--	0.107	0	--	--	0.134	3
46	2.6	5	0.128	2	0.260	3	0.100	0	0.158	4	0.146	4
48	0.6	5	0.400	0	0.530	0	0.160	0	0.150	3	0.074	0
50	3.8	5	0.143	4	0.230	4	0.082	4	0.158	4	0.135	3
51	3.0	5	0.160	4	0.160	3	0.100	0	0.153	4	0.144	4
53	4.0	2	--	--	--	--	0.080	4	--	--	0.138	4
59	2.4	5	0.110	1	0.200	4	0.090	3	0.100	0	0.140	4
64	4.0	4	0.150	4	--	--	0.080	4	0.158	4	0.141	4
69	4.0	1	--	--	--	--	0.080	4	--	--	--	--
70	1.0	5	0.105	0	0.450	0	0.090	3	0.205	0	0.152	2
76	3.0	2	0.149	4	--	--	0.075	2	--	--	--	--
81	2.2	5	0.173	2	0.131	2	0.092	2	0.138	1	0.138	4
83	3.5	2	0.130	3	--	--	--	--	--	--	0.140	4
85	3.8	4	0.152	4	--	--	0.087	4	0.168	3	0.143	4
86	1.7	3	0.187	1	--	--	0.017	0	0.159	4	--	--
89	3.6	5	0.144	4	0.240	3	0.085	4	0.161	4	0.147	3
90	4.0	3	0.156	4	0.181	4	0.080	4	--	--	--	--
93	2.5	4	0.130	3	--	--	0.080	4	0.150	3	0.167	0
96	2.4	5	0.132	3	0.166	3	0.089	3	0.142	1	0.129	2
97	2.8	4	0.150	4	0.120	2	0.080	4	< 0.18	NR	0.160	1
102	1.8	5	0.170	3	0.200	4	0.000	0	0.147	2	0.116	0
105	1.8	4	0.140	4	< 1	NR	0.070	1	0.220	0	0.152	2
110	3.0	1	0.165	3	--	--	--	--	--	--	--	--
113	4.0	4	0.145	4	< 0.5	NR	0.082	4	0.160	4	0.146	4
114	2.0	3	0.120	2	--	--	0.100	0	0.160	4	--	--
127	3.0	5	0.133	3	0.178	4	0.093	2	0.155	4	0.131	2
129	4.0	5	0.144	4	0.187	4	0.081	4	0.162	4	0.145	4
134	3.6	5	0.148	4	0.135	3	0.090	3	0.158	4	0.139	4
138	4.0	5	0.152	4	0.226	4	0.086	4	0.158	4	0.139	4
140	2.2	5	0.160	4	0.190	4	0.078	3	0.110	0	0.110	0
142	3.2	5	0.156	4	0.156	3	0.086	4	0.148	3	0.132	2
143	3.8	5	0.170	3	0.200	4	0.086	4	0.154	4	0.143	4
145	3.6	5	0.160	4	0.200	4	0.090	3	0.160	4	0.150	3
146	2.2	5	0.100	0	0.230	4	0.082	4	0.150	3	1.800	0
151	2.0	3	0.140	4	--	--	0.093	2	--	--	0.420	0
154	2.2	5	0.149	4	0.308	1	0.050	0	0.164	3	0.137	3
155	3.0	5	0.129	2	0.210	4	0.097	1	0.161	4	0.144	4
158	2.4	5	0.157	4	0.156	3	0.146	0	0.142	1	0.142	4
180	3.0	5	0.152	4	0.226	4	0.088	3	0.179	1	0.151	3
183	2.0	4	0.340	0	--	--	0.680	0	0.158	4	0.144	4
185	3.3	4	0.129	2	--	--	0.083	4	0.157	4	0.135	3
190	3.0	2	0.171	2	--	--	--	--	--	--	0.139	4
193	2.7	3	0.101	0	< 0.2	NR	0.080	4	0.159	4	--	--

Table 7. Laboratory performance ratings for standard reference sample N-63 (nutrient constituents)--Continued

[MPV, most probable value mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/5, number of reported values of 5 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value						
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00						
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00						
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)								
Analyte = NH ₃ as N (Ammonia)												
MPV = 0.150 mg/L												
F-pseudosigma = 0.021												
Lab	OLR	V/5	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
198	2.8	4	0.145	4	--	--	0.084	4	0.139	1	0.132	2
205	1.0	2	0.124	2	--	--	0.128	0	--	--	--	--
209	3.0	3	0.163	3	0.157	3	0.088	3	--	--	--	--
212	1.2	5	0.150	4	0.444	0	0.240	0	0.135	0	0.154	2
213	3.0	2	< 1	NR	< 1	NR	--	--	0.160	4	0.130	2
215	0.8	4	0.120	2	--	--	0.070	1	0.220	0	0.180	0
224	2.4	5	0.181	2	0.459	0	0.088	3	0.150	3	0.144	4
227	3.6	5	0.141	4	0.168	4	0.081	4	0.156	4	0.132	2
234	3.3	4	0.152	4	--	--	0.080	4	0.150	3	0.152	2
241	3.3	4	0.133	3	--	--	0.077	3	0.160	4	0.137	3
243	2.3	3	0.198	0	--	--	0.080	4	0.150	3	--	--
247	2.0	5	0.073	0	0.187	4	0.076	3	0.127	0	0.134	3
253	2.0	4	0.151	4	--	--	0.102	0	0.249	0	0.144	4
255	4.0	3	0.145	4	--	--	< 0.1	NR	0.158	4	0.140	4
282	0.2	5	0.380	0	0.370	0	0.104	0	0.178	1	0.176	0
285	0.8	4	0.648	0	0.698	0	0.069	1	0.144	2	--	--
287	0.0	3	0.248	0	--	--	0.250	0	--	--	0.190	0
292	2.3	4	0.100	0	--	--	0.080	4	0.170	2	0.135	3
301	3.0	4	0.144	4	--	--	0.068	1	0.165	3	0.140	4
306	2.5	4	0.165	3	0.192	4	0.063	0	0.166	3	--	--
307	2.0	4	0.127	2	--	--	0.093	2	0.178	1	0.148	3
308	0.0	1	< 0.15	NR	0.520	0	< 0.15	NR	< 0.15	NR	< 0.15	NR
312	3.5	2	0.134	3	--	--	--	--	--	--	0.143	4
313	3.0	5	0.163	3	0.278	2	0.081	4	0.164	3	0.136	3
314	1.6	5	0.165	3	0.487	0	0.091	3	0.170	2	0.166	0
316	3.8	5	0.154	4	0.199	4	0.084	4	0.166	3	0.140	4
317	2.7	3	0.157	4	--	--	0.115	0	--	--	0.142	4
318	4.0	5	0.156	4	0.200	4	0.081	4	0.155	4	0.145	4
319	1.6	5	0.060	0	0.230	4	0.100	0	0.180	0	0.146	4
320	2.8	5	0.120	2	0.214	4	0.081	4	0.148	3	0.127	1
321	3.3	4	0.150	4	--	--	0.078	3	0.158	4	0.156	2

Table 8. Laboratory performance ratings for standard reference sample N-64 (nutrient constituents)

[MPV, most probable value mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/5, number of reported values of 5 possible values; RV, reported value; <, less than; --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value							
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00							
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00							
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)									
Analyte = NH ₃ as N (Ammonia)			NH ₃ + Org N as N (Ammonia + Organic N)			NO ₃ as N (Nitrate)			total P as P (total Phosphorus)			PO ₄ as P (Orthophosphate as P)	
MPV = 1.38 mg/L			1.54 mg/L			1.26 mg/L			0.883 mg/L			0.860 mg/L	
F-pseudosigma = 0.08			0.13			0.04			0.041			0.034	
Lab	OLR	V/5	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating	
1	3.6	5	1.43	3	1.51	4	1.25	4	0.862	4	0.891	3	
3	3.0	5	1.28	2	1.52	4	1.31	2	0.884	4	0.826	3	
10	3.8	5	1.38	4	1.49	4	1.24	3	0.877	4	0.867	4	
11	2.2	5	1.63	0	1.80	1	1.29	3	0.900	4	0.900	3	
12	1.4	5	1.10	0	2.40	0	1.24	3	0.787	0	0.840	4	
13	3.3	4	1.33	3	--	--	1.26	4	0.889	4	0.816	2	
18	3.4	5	1.32	3	1.65	3	1.23	3	0.887	4	0.869	4	
23	3.8	5	1.35	4	1.57	4	1.26	4	0.860	3	0.850	4	
25	2.3	4	1.31	3	--	--	1.24	3	0.810	1	0.909	2	
26	1.5	2	--	--	--	--	1.65	0	--	--	0.830	3	
36	2.2	5	0.02	0	1.60	4	1.24	3	1.610	0	0.840	4	
38	3.4	5	1.41	4	1.41	3	1.26	4	0.920	3	0.836	3	
42	3.5	2	--	--	--	--	1.29	3	--	--	0.871	4	
46	3.2	5	1.34	3	1.66	3	1.26	4	0.894	4	0.904	2	
53	3.5	2	--	--	--	--	1.24	3	--	--	0.856	4	
59	2.6	5	1.39	4	1.50	4	1.34	0	0.800	1	0.840	4	
64	3.5	4	1.36	4	--	--	1.21	2	0.875	4	0.846	4	
69	3.0	1	--	--	--	--	1.24	3	--	--	--	--	
70	2.2	5	0.76	0	1.90	0	1.26	4	0.885	4	0.901	3	
76	2.0	3	1.40	4	--	--	1.16	0	--	--	0.799	2	
83	4.0	2	1.38	4	--	--	--	--	--	--	0.860	4	
85	2.8	4	1.41	4	--	--	1.32	1	0.947	2	0.880	4	
87	3.0	5	1.42	4	1.64	3	1.26	4	0.988	0	0.881	4	
89	3.4	5	1.34	3	1.50	4	1.29	3	0.889	4	0.884	3	
90	3.0	3	1.50	2	1.46	3	1.25	4	--	--	--	--	
93	2.3	4	1.30	2	--	--	1.29	3	0.880	4	1.040	0	
96	3.2	5	1.30	3	1.52	4	1.31	2	0.851	3	0.869	4	
102	1.8	5	1.22	1	1.48	4	1.21	2	0.831	2	0.717	0	
105	2.0	5	1.34	3	1.56	4	1.24	3	1.120	0	1.230	0	
107	3.5	4	1.40	4	--	--	1.29	3	0.856	3	0.848	4	
113	3.8	5	1.38	4	1.44	3	1.27	4	0.871	4	0.855	4	
114	1.7	3	1.28	2	--	--	1.36	0	0.840	3	--	--	
127	3.4	5	1.39	4	1.64	3	1.24	3	0.920	3	0.863	4	
129	2.4	5	1.28	2	2.09	0	1.31	2	0.877	4	0.855	4	
134	2.6	5	1.51	1	1.32	1	1.27	4	0.911	3	0.859	4	
138	3.6	5	1.35	4	1.60	4	1.25	4	0.886	4	0.815	2	
140	1.4	5	1.65	0	1.68	2	1.20	1	0.770	0	0.850	4	
143	3.2	5	1.35	4	1.49	4	1.16	0	0.877	4	0.861	4	
145	2.8	5	1.45	3	1.65	3	1.25	4	0.910	3	0.925	1	
146	2.6	5	1.40	4	1.60	4	1.30	2	0.860	3	0.700	0	
151	1.3	3	1.50	2	--	--	1.31	2	--	--	2.720	0	
154	2.8	5	1.37	4	1.54	4	1.24	3	0.981	0	0.891	3	
155	3.2	5	1.40	4	1.57	4	1.09	0	0.904	4	0.860	4	
158	2.4	5	1.46	3	1.30	1	1.32	1	0.878	4	0.893	3	
180	3.2	5	1.43	3	1.56	4	1.30	2	0.883	4	0.888	3	
183	2.3	3	1.50	2	--	--	--	--	0.830	2	0.830	3	
185	3.3	4	1.39	4	--	--	1.27	4	0.907	3	0.912	2	
190	2.5	2	1.51	1	--	--	--	--	--	--	0.878	4	
193	2.0	4	1.08	0	1.54	4	1.28	3	0.803	1	--	--	
198	3.3	4	1.37	4	--	--	1.26	4	0.882	4	0.782	1	
200	3.3	4	1.40	4	--	--	1.22	2	0.910	3	0.850	4	
204	2.2	5	1.43	3	1.87	0	1.31	2	0.871	4	0.904	2	
205	4.0	2	1.42	4	--	--	1.26	4	--	--	--	--	
208	3.0	2	--	--	--	--	1.31	2	--	--	0.850	4	
209	3.3	3	1.43	3	1.43	3	1.25	4	--	--	--	--	

Table 8. Laboratory performance ratings for standard reference sample N-64 (nutrient constituents)--Continuec

[MPV, most probable value mg/L, milligrams per liter; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/5, number of reported values of 5 possible values; RV, reported value; <, less than; --, not reported.]

Rating		Absolute Z-value		Rating		Absolute Z-value						
4 (Excellent)		0.00 - 0.50		1 (Marginal)		1.51 - 2.00						
3 (Good)		0.51 - 1.00		0 (Unsatisfactory)		greater than 2.00						
2 (Satisfactory)		1.01 - 1.50		NR (Not Rated)								
Analyte = NH ₃ as N (Ammonia)		NH ₃ + Org N as N (Ammonia + Organic N)		NO ₃ as N (Nitrate)		total P as P (total Phosphorus)		PO ₄ as P (Orthophosphate as P)				
MPV = 1.38 mg/L		1.54 mg/L		1.26 mg/L		0.883 mg/L		0.860 mg/L				
F-pseudosigma = 0.08		0.13		0.04		0.041		0.034				
Lab	OLR	V/5	RV	Rating	RV	Rating	RV	Rating	RV	Rating		
212	1.8	5	1.44	3	1.44	3	1.16	0	0.800	1	0.904	2
213	2.8	4	1.40	4	2.30	0	--	--	0.850	3	0.850	4
215	3.4	5	1.37	4	1.50	4	1.25	4	0.920	3	0.920	2
224	2.6	5	1.54	1	1.51	4	1.32	1	0.875	4	0.896	3
227	2.2	5	1.49	2	2.35	0	1.22	2	0.858	3	0.866	4
234	1.8	4	1.42	4	--	--	1.32	1	0.820	2	0.670	0
243	2.7	3	1.47	2	--	--	1.27	4	0.830	2	--	--
247	3.0	5	1.31	3	1.45	3	1.24	3	0.822	2	0.846	4
253	3.0	4	1.41	4	--	--	1.38	0	0.883	4	0.855	4
255	3.8	4	1.42	4	--	--	1.27	4	0.920	3	0.856	4
282	2.4	5	1.30	2	1.21	0	1.27	4	0.842	3	0.831	3
285	0.0	4	0.70	0	0.72	0	1.18	0	0.987	0	--	--
287	1.0	3	2.28	0	--	--	1.17	0	--	--	0.830	3
291	4.0	1	--	--	--	--	1.27	4	--	--	--	--
292	2.3	4	0.98	0	--	--	1.30	2	0.920	3	0.845	4
306	0.8	4	0.78	0	0.89	0	0.62	0	0.911	3	--	--
312	0.0	1	1.03	0	--	--	--	--	--	--	--	--
313	4.0	5	1.38	4	1.56	4	1.27	4	0.899	4	0.863	4
316	3.8	5	1.42	4	1.49	4	1.26	4	0.913	3	0.865	4
319	2.0	5	1.38	4	1.46	3	1.23	3	1.070	0	1.000	0
320	2.2	5	1.37	4	2.66	0	1.28	3	2.000	0	0.876	4

Table 9. Laboratory performance ratings for standard reference sample P-33 (low ionic strength constituents)

[MPV, most probable value; mg/L, milligrams per liter; μ S/cm, microseimens per centimeter at 25 °C; Lab, laboratory number; OLR, overall laboratory rating for all reported values; V/10, number of reported values of 10 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Acidity as CaCO ₃		Ca (Calcium)		Cl (Chloride)		F (Fluoride)		K (Potassium)		Mg (Magnesium)				
MPV = Insufficient data		0.322 mg/L		2.38 mg/L		0.105 mg/L		0.110 mg/L		0.100 mg/L				
F-pseudostigma =		0.043		0.18		0.039		0.015		0.015				
Lab	OLR	V/10	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating		
1	3.1	10	2.50	4	0.319	4	2.51	3	< 0.1	NR	0.117	4	0.099	4
2	2.3	8	--	--	0.242	1	1.79	0	0.112	4	0.112	4	0.119	2
3	2.0	9	< 10	NR	0.354	3	1.69	0	0.137	3	0.532	0	0.134	0
11	1.5	10	--	--	0.400	1	2.00	0	0.080	3	< 0.05	0	0.100	4
12	2.0	3	--	--	--	--	--	--	--	--	--	--	--	--
25	3.0	6	3.00	4	--	--	2.40	4	0.090	4	--	--	--	--
33	3.7	10	--	--	0.310	4	2.39	4	0.100	4	0.100	3	0.090	3
38	3.8	8	0.31	4	0.320	4	--	--	--	--	0.120	3	0.099	4
42	2.4	7	--	--	0.302	4	2.31	4	0.110	4	0.107	4	0.068	0
45	2.0	7	--	--	--	--	2.46	4	0.085	3	0.280	0	--	--
46	2.9	9	--	--	0.322	4	2.35	4	0.075	3	0.157	0	0.102	4
48	2.1	9	--	--	0.346	3	2.00	0	--	--	0.132	2	0.116	2
59	3.5	6	--	--	--	--	2.38	4	0.150	2	--	--	--	--
64	2.1	9	--	--	0.390	1	2.43	4	--	--	0.060	0	0.110	3
81	3.1	11	1.40	4	0.321	4	2.46	4	0.105	4	0.082	1	0.094	4
83	3.5	2	--	--	--	--	--	--	0.140	3	--	--	--	--
86	3.0	8	--	--	0.281	3	2.29	3	0.100	4	--	--	0.111	3
89	2.7	11	3.53	4	0.170	0	2.36	4	0.100	4	0.100	3	0.100	4
93	3.6	10	--	--	0.310	4	2.38	4	0.070	3	0.120	3	0.080	2
102	0.9	9	--	--	0.230	0	4.77	0	0.820	0	0.060	0	0.130	1
107	4.0	2	--	--	--	--	--	--	--	--	--	--	--	--
113	3.0	10	--	--	0.362	3	2.46	4	0.114	4	0.145	0	0.111	3
127	3.5	11	4.10	4	0.337	4	2.18	2	0.094	4	0.106	4	0.115	2
134	3.9	10	--	--	0.320	4	2.30	4	0.115	4	0.111	4	0.100	4
138	3.9	9	--	--	0.340	4	2.36	4	< 0.10	NR	0.110	4	0.095	4
140	3.0	10	--	--	0.337	4	2.28	3	0.138	3	0.091	2	0.100	4
141	2.0	10	61.00	0	0.357	3	2.30	4	0.110	4	0.177	0	0.140	0
143	3.5	4	--	--	--	--	2.59	2	--	--	--	--	--	--
145	2.1	9	--	--	0.350	3	2.38	4	0.070	3	0.040	0	< 0.19	NR
146	3.0	5	< 5.79	NR	< 0.313	NR	2.13	2	0.130	3	< 1	NR	< 0.5	NR
155	2.4	5	--	--	0.468	0	--	--	--	--	--	--	0.211	0
180	3.1	9	--	--	0.358	3	2.43	4	0.171	1	< 1.63	NR	0.115	2
183	2.2	5	--	--	--	--	5.04	0	0.140	3	--	--	--	--
185	3.0	8	--	--	0.326	4	2.42	4	--	--	0.109	4	0.098	4
190	2.3	10	--	--	0.800	0	1.81	0	0.103	4	0.120	3	0.190	0
191	2.9	8	--	--	0.320	4	2.34	4	0.080	3	0.080	1	0.100	4
193	2.8	5	--	--	0.290	3	--	--	--	--	0.100	3	0.095	4
196	3.3	4	--	--	--	--	2.21	3	0.074	3	--	--	--	--
203	2.9	7	--	--	0.260	2	2.60	2	--	--	0.120	3	< 0.10	NR
204	4.0	2	--	--	--	--	--	--	--	--	--	--	--	--
208	1.0	2	--	--	--	--	2.60	2	--	--	--	--	--	--
209	3.1	7	--	--	0.247	1	2.30	4	--	--	0.098	3	0.094	4
215	0.9	8	2.00	4	0.230	0	2.70	1	--	--	--	--	--	--
227	2.5	6	--	--	0.322	4	2.77	0	--	--	--	--	0.096	4
228	2.0	9	20.70	0	0.240	1	2.44	4	--	--	0.040	0	0.050	0
237	3.5	8	1.01	4	0.320	4	2.58	2	--	--	0.110	4	0.090	3
238	3.4	5	--	--	0.320	4	2.58	2	--	--	0.110	4	0.100	4
241	2.1	10	--	--	0.292	3	2.08	1	0.100	4	0.128	2	0.126	1
243	4.0	2	--	--	--	--	--	--	--	--	--	--	--	--
244	4.0	2	--	--	--	--	--	--	--	--	--	--	--	--
247	3.5	8	--	--	0.335	4	2.22	3	0.080	3	< 0.20	NR	< 0.2	NR
255	3.0	4	--	--	0.334	4	--	--	< 0.5	NR	--	--	0.098	4
265	2.9	7	--	--	0.340	4	2.10	1	< 0.1	NR	0.110	4	0.100	4
268	2.8	8	--	--	0.400	1	2.39	4	--	--	0.150	0	0.110	3
273	2.4	11	3.05	4	0.290	3	2.21	3	0.119	4	0.110	4	0.100	4
274	0.6	11	15.99	0	1.980	0	19.68	0	0.002	0	0.100	3	0.950	0
279	2.4	8	--	--	0.950	0	2.30	4	--	--	0.130	2	0.180	0
282	2.7	10	9.85	2	0.310	4	1.63	0	0.160	2	--	--	0.095	4
287	2.1	10	1.00	4	0.403	1	3.80	0	0.240	0	0.098	3	0.110	3
301	2.4	7	--	--	0.291	3	--	--	--	--	0.110	4	0.089	3
312	4.0	3	--	--	--	--	--	--	--	--	--	--	--	--
321	2.7	7	--	--	0.370	2	2.40	4	--	--	--	--	--	--

Table 9. Laboratory performance ratings for standard reference sample P-33 (low ionic strength constituents)--Continued

[MPV, most probable value; mg/L, milligrams per liter; µs/cm, microseimens per centimeter at 25 °C; Lab, laboratory number; OLR, overall laboratory rating

for all reported values; V/10, number of reported values of 10 possible values; RV, reported value; <, less than; NR, not rated; --, not reported.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Na (Sodium)		pH		PO ₄ as P		SO ₄ (Sulfate)		Specific Conductance		
MPV =	1.42 mg/L		4.67		0.045		1.41 mg/L		20.8	
F-pseudostigma =	0.10		0.13		0.007		0.17		2.4	
Lab	RV	Rating	RV	Rating	RV	Rating	RV	Rating	RV	Rating
1	1.37	4	5.01	2	0.025	0	1.43	4	18.1	2
2	1.93	0	4.50	3	--	--	1.42	4	--	--
3	1.53	2	4.39	2	0.046	4	< 10	NR	20.8	4
11	0.41	0	5.58	0	0.040	3	2.00	0	19.7	4
12	--	--	4.04	0	0.039	3	--	--	19.1	3
25	--	--	4.84	3	0.038	3	--	--	28.0	0
33	1.37	4	4.64	4	0.050	3	1.39	4	20.7	4
38	1.31	3	4.70	4	0.042	4	--	--	21.6	4
42	1.21	1	--	--	< 0.05	NR	2.32	0	--	--
45	4.09	0	4.44	3	--	--	1.84	0	20.9	4
46	1.40	4	4.90	3	0.047	4	--	--	61.0	0
48	1.46	4	6.00	0	0.045	4	1.00	0	21.0	4
59	--	--	4.70	4	0.050	3	1.39	4	21.0	4
64	1.31	3	5.01	2	0.032	1	1.37	4	17.1	1
81	1.50	3	4.60	4	0.037	2	1.20	2	17.2	2
83	--	--	--	--	0.042	4	--	--	--	--
86	1.35	3	4.67	4	--	--	1.61	2	18.0	2
89	1.16	0	4.39	2	0.042	4	1.44	4	25.0	1
93	1.39	4	4.78	4	0.045	4	1.37	4	20.6	4
102	50.06	0	--	--	0.039	3	1.99	0	22.0	4
107	--	--	4.67	4	--	--	--	--	20.8	4
113	1.60	1	4.70	4	0.047	4	1.39	4	19.0	3
127	1.42	4	4.66	4	0.053	2	1.43	4	22.0	4
134	1.45	4	4.67	4	0.040	3	1.45	4	21.0	4
138	1.40	4	4.67	4	0.043	4	1.32	3	20.6	4
140	1.48	3	4.69	4	0.050	3	1.00	0	21.0	4
141	1.57	2	4.50	3	--	--	1.30	3	16.0	1
143	--	--	4.68	4	0.048	4	--	--	20.0	4
145	1.42	4	4.50	3	< 0.01	0	1.60	2	< 1	0
146	1.45	4	4.61	4	< 0.021	NR	< 5	NR	24.0	2
155	--	--	4.56	4	0.044	4	--	--	20.8	4
180	1.46	4	4.70	4	0.044	4	1.41	4	18.0	2
183	--	--	4.80	3	0.039	3	--	--	18.0	2
185	1.47	3	4.26	1	--	--	1.37	4	64.5	0
190	1.47	3	4.60	4	0.051	3	1.21	2	21.0	4
191	1.46	4	--	--	0.110	0	1.55	3	--	--
193	1.28	2	--	--	--	--	--	--	17.3	2
196	--	--	--	--	0.048	4	1.52	3	--	--
203	1.22	1	4.58	4	0.045	4	< 2.5	NR	19.7	4
204	--	--	4.67	4	--	--	--	--	21.4	4
208	--	--	--	--	--	--	2.00	0	--	--
209	1.29	2	4.69	4	--	--	1.45	4	--	--
215	0.86	0	5.20	0	0.060	1	2.00	0	16.9	1
227	--	--	4.70	4	--	--	1.82	0	19.5	3
228	1.55	2	4.59	4	--	--	1.28	3	21.5	4
237	1.39	4	4.55	3	--	--	1.39	4	--	--
238	1.50	3	--	--	--	--	--	--	--	--
241	1.36	3	4.42	2	0.036	2	1.29	3	11.0	0
243	--	--	4.63	4	--	--	--	--	20.8	4
244	--	--	4.67	4	--	--	--	--	20.8	4
247	1.43	4	4.70	4	0.040	3	1.30	3	21.0	4
255	--	--	--	--	0.045	4	2.30	0	--	--
265	1.40	4	7.00	0	--	--	1.30	3	--	--
268	1.45	4	4.68	4	--	--	1.41	4	18.0	2
273	0.14	0	6.31	0	0.289	0	1.11	1	22.7	3
274	4.94	0	3.58	0	0.065	0	0.32	0	19.9	4
279	1.54	2	4.75	4	--	--	1.48	4	18.9	3
282	1.37	4	4.35	2	0.041	3	1.30	3	22.4	3
287	1.32	3	4.75	4	0.580	0	1.27	3	--	--
301	1.33	3	4.75	4	0.256	0	--	--	0.1	0
312	--	--	4.67	4	0.045	4	--	--	21.7	4
321	1.41	4	5.15	0	0.048	4	1.46	4	17.0	1

Table 10. Laboratory performance ratings for standard reference sample Hg-29 (Mercury)

[MPV, most probable value; ug/L, micrograms per liter; Lab, laboratory number; V/1, number of reported values of 1 value; RV, reported value; <, less than; NR, not rated.]

Rating	Absolute Z-value	Rating	Absolute Z-value
4 (Excellent)	0.00 - 0.50	1 (Marginal)	1.51 - 2.00
3 (Good)	0.51 - 1.00	0 (Unsatisfactory)	greater than 2.00
2 (Satisfactory)	1.01 - 1.50	NR (Not Rated)	

Analyte = Hg (Mercury)
 MPV = 0.463 µg/L
 F-pseudostigma = 0.052

Lab	V/1	RV	Rating
1	1	0.438	4
3	1	0.463	4
10	1	0.560	1
11	1	0.440	4
12	1	0.600	0
13		< 4	NR
32	1	0.550	1
39	1	0.457	4
45	1	0.110	0
46	1	0.448	4
48	1	0.390	2
50	1	0.540	2
59	1	0.480	4
70	1	0.474	4
81	1	0.388	2
86	1	0.453	4
87	1	0.600	0
89	1	0.470	4
96	1	0.460	4
113	1	0.450	4
126	1	0.540	2
127	1	0.419	3
134	1	0.470	4
138	1	0.420	3
141	1	0.530	2
142	1	0.460	4
144	1	0.550	1
145	1	0.390	2
146	1	0.584	0
147	1	0.437	4
151	1	0.450	4
154	1	0.200	0
180	1	0.300	0
193	1	0.466	4
198	1	0.508	3
203	1	0.500	3
212	1	0.569	0
213	1	0.500	3
215	1	0.600	0
219	1	0.490	3
234	1	0.440	4
241	1	0.410	2
247	1	0.480	4
255	1	0.359	1
259	1	0.680	0
265	1	0.470	4
277	1	0.475	4
282	1	0.432	3
298	1	0.480	4
304	1	0.440	4
307	1	0.340	0
308	1	0.390	2

Table 11. *Statistical summary of reported data for standard reference sample T-159 (trace constituents)*

Definition of analytical methods, abbreviations, and symbols

Analytical methods and codes

0. Other/Not reported	=	
1. AA: direct, air	=	atomic absorption: direct air
2. AA: direct, N ₂ O	=	atomic absorption: direct, nitrous oxide
3. AA: graphite furnace	=	atomic absorption: graphite furnace
4. ICP	=	inductively coupled plasma
5. DCP	=	direct current plasma
6. ICP/MS	=	inductively coupled plasma / mass spectrometry
7. IC	=	ion chromatography
8. AA: cold vapor	=	atomic absorption: cold vapor
11. AA: hydride	=	atomic absorption: hydride (reducing agent specified)
12. AA: flame emission	=	atomic absorption: flame emission
20. Titration color	=	Titration colorimetric (specify color reagent)
22. Color	=	colorimetric (color reagent specified)


Abbreviations and figure symbols

N	=	number of analyses--(excluding less than values)
MPV	=	most probable value
F-pseudosigma	=	nonparametric statistic deviation
Uh	=	upper hinge value
Lh	=	lower hinge value
Uwl	=	upper warning limit - - - -
Lwl	=	lower warning limit - - - -
Ucl	=	upper warning limit
Lcl	=	lower warning limit - - - -
µg/L	=	micrograms per liter
mg/L	=	milligrams per liter
Lab	=	laboratory code number
NR	=	not rated, less than value reported or insufficient data
<	=	less than
--	=	not reported


<u>Constituent</u>	<u>page</u>	<u>Constituent</u>	<u>page</u>
Ag Silver	36	Mg Magnesium	50
Al Aluminum	37	Mn Manganese	51
As Arsenic	38	Mo Molybdenum	52
B Boron	39	Na Sodium	53
Ba Barium	40	Ni Nickel	54
Be Beryllium	41	Pb Lead	55
Ca Calcium	42	Sb Antimony	56
Cd Cadmium	43	Se Selenium	57
Co Cobalt	44	SiO ₂ Silica	58
Cr Chromium	45	Sr Strontium	59
Cu Copper	46	Tl Thallium	60
Fe Iron	47	U Uranium	61
K Potassium	48	V Vanadium	62
Li Lithium	49	Zn Zinc	63

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Ag (Silver) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY						
N =	3	16	24	16	1. AA: direct, air	MPV = 9.67
Minimum =	8.60	7.00	7.70	5.00	3. AA: graphite furnace	F-pseudosigma = 0.91
Maximum =	12.00	12.20	40.00	12.70	4. ICP	N = 59
Median =		9.30	10.00	9.46	6. ICP/MS	Uh = 10.38
F-pseudosigma =		1.02	0.87	0.44		Lh = 9.15

Lab	Rating	Z-value	1	3	4	6	Lab	Rating	Z-value	1	3	4	6
1	4	-0.20	--	--	--	9.49	255	2	1.24	--	10.80	--	--
3	4	-0.21	--	--	9.48	--	259	4	-0.08	--	--	9.60	--
11	4	0.36	--	--	10.00	--	265	3	-0.74	--	--	--	9.00
12	4	-0.30	--	9.40	--	--	273	3	-0.63	--	--	9.10	--
13	4	0.35	--	--	9.99	--	277	4	-0.41	--	--	9.30	--
23	0	2.13	--	--	11.60	--	282	0	2.79	--	12.20	--	--
32	4	-0.30	--	--	--	9.40	292	0	8.07	--	--	17.00	--
39	3	-0.84	--	--	8.91	--	304	0	-2.17	--	--	7.70	--
42	4	0.00	--	--	--	9.67	305	3	-0.69	--	9.04	--	--
46	4	0.47	--	--	10.10	--	307	4	0.09	--	9.75	--	--
48	4	-0.30	--	--	--	9.40	309	2	1.46	--	--	11.00	--
50	4	0.25	--	--	--	9.90							
57	3	0.69	--	--	10.30	--							
59	3	-0.74	--	--	--	9.00							
70	1	-1.84	--	8.00	--	--							
81	0	-2.94	--	7.00	--	--							
86	3	-0.85	--	--	8.90	--							
87	3	0.69	--	--	10.30	--							
89	3	-0.84	--	8.91	--	--							
93	4	0.47	--	--	10.10	--							
96	2	1.13	--	10.70	--	--							
102	0	33.40	--	--	40.00	--							
113	3	0.86	--	10.45	--	--							
114	0	2.57	12.00	--	--	--							
121	3	-0.63	--	--	--	9.10							
126	2	1.46	11.00	--	--	--							
127	4	-0.24	--	9.45	--	--							
131	NR	--	--	--	< 30	--							
134	4	-0.12	--	--	9.56	--							
138	4	-0.26	--	--	--	9.43							
140	2	-1.18	8.60	--	--	--							
141	2	1.13	--	--	10.70	--							
142	0	3.34	--	--	--	12.70							
144	3	-0.52	--	9.20	--	--							
146	1	1.91	--	--	11.40	--							
151	4	-0.08	--	--	--	9.60							
154	4	0.25	--	--	9.90	--							
180	2	1.13	--	--	10.70	--							
190	3	-0.52	--	9.20	--	--							
193	3	0.58	--	10.20	--	--							
198	NR	--	--	--	< 50	--							
203	3	-0.74	--	9.00	--	--							
204	0	-5.14	--	--	--	5.00							
212	4	0.18	--	--	--	9.83							
215	1	-1.84	--	8.00	--	--							
219	2	1.46	--	--	--	11.00							
234	4	0.36	--	--	10.00	--							
236	4	0.36	--	--	10.00	--							
241	4	-0.32	--	--	--	9.38							
247	4	0.19	--	--	--	9.84							

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Al (Aluminum) Concentration Unit : mg/L



SUMMARY

N =	4	24	1	18
Minimum =	31.0	24.0	70.0	27.0
Maximum =	71.6	120.0		47.0
Median =		32.2		31.2
F-pseudostigma =		7.6		2.4


- 3. AA: graphite furnace
- 4. ICP
- 5. DCP
- 6. ICP/MS

MPV =	31.9
F-pseudostigma =	3.7
N =	47
Uh =	35.3
Lh =	30.4


Lab	Rating	Z-value	3	4	5	6
1	4	-0.19	--	--	--	31.2
3	4	-0.25	--	31.0	--	--
11	0	24.01	--	120.0	--	--
13	4	-0.25	31.0	--	--	--
23	4	0.08	--	32.2	--	--
26	4	0.33	--	33.1	--	--
32	4	-0.46	--	--	--	30.2
33	0	10.38	--	--	70.0	--
39	0	4.52	--	48.5	--	--
42	NR	--	--	--	--	< 30
46	3	0.74	--	34.6	--	--
48	4	-0.25	--	--	--	31.0
50	4	0.00	--	--	--	31.9
57	4	0.05	--	32.1	--	--
59	4	0.08	--	--	--	32.2
70	4	-0.14	--	--	--	31.4
76	4	-0.22	--	--	--	31.1
81	0	2.75	--	42.0	--	--
89	2	1.12	36.0	--	--	--
93	1	-1.91	--	24.9	--	--
102	3	-0.68	--	29.4	--	--
127	0	19.38	--	103.0	--	--
131	0	-2.15	--	24.0	--	--
134	4	0.03	--	32.0	--	--
138	4	-0.33	--	--	--	30.7
141	2	-1.44	--	26.6	--	--
142	NR	--	--	< 30	--	--
145	NR	--	--	< 180	--	--
146	NR	--	--	< 36.9	--	--
147	4	0.11	--	--	--	32.3
151	2	-1.23	--	--	--	27.4
154	3	-0.65	--	29.5	--	--
180	NR	--	--	< 51.6	--	--
190	4	0.25	32.8	--	--	--
191	4	0.03	--	--	--	32.0
198	NR	--	--	< 50	--	--
203	3	-0.79	--	29.0	--	--
204	2	-1.01	--	--	--	28.2
209	4	0.16	--	32.5	--	--
212	0	9.84	--	68.0	--	--
219	2	-1.06	--	--	--	28.0
227	4	-0.38	--	30.5	--	--
234	1	1.77	--	38.4	--	--
236	0	6.30	--	55.0	--	--
241	3	-0.87	--	--	--	28.7
247	4	-0.14	--	--	--	31.4
254	4	-0.33	--	30.7	--	--
255	4	0.33	--	--	--	33.1
259	2	1.42	--	37.1	--	--
265	2	-1.34	--	--	--	27.0

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : As (Arsenic) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	22	13	19	5	3. AA: graphite furnace	MPV =	28.4		
Minimum =	19.0	22.6	26.0	25.6	4. ICP	F-pseudosigma =	1.6		
Maximum =	31.0	54.0	31.4	31.6	6. ICP/MS	N =	59		
Median =	28.2	28.1	28.6		11na. AA: hydride NaBH4	Uh =	29.7		
F-pseudosigma =	2.2	2.8	1.4			Lh =	27.5		


Lab	Rating	Z-value	3	4	6	11na	Lab	Rating	Z-value	3	4	6	11na
1	3	0.56	29.3	--	--	--	241	1	1.76	--	--	31.2	--
3	4	-0.19	--	28.1	--	--	247	1	1.63	--	--	31.0	--
10	1	1.63	--	--	--	31.0	255	2	1.19	--	--	30.3	--
11	4	-0.25	--	28.0	--	--	259	3	-0.56	--	27.5	--	--
13	4	-0.44	27.7	--	--	--	265	4	0.06	--	--	28.5	--
23	0	-2.64	--	24.2	--	--	277	4	-0.50	--	27.6	--	--
26	3	-0.56	--	--	--	27.5	282	1	2.01	--	--	--	31.6
32	4	0.13	--	--	28.6	--	292	0	-5.90	19.0	--	--	--
39	4	-0.25	28.0	--	--	--	305	0	-5.12	20.2	--	--	--
42	3	0.75	--	--	29.6	--	307	4	-0.44	27.7	--	--	--
46	2	-1.38	26.2	--	--	--	309	0	16.06	--	54.0	--	--
48	4	-0.31	--	--	27.9	--							
50	4	-0.38	--	--	27.8	--							
57	1	1.82	--	31.3	--	--							
59	1	1.88	--	--	31.4	--							
70	3	0.75	--	--	29.6	--							
81	3	1.00	30.0	--	--	--							
86	2	-1.13	--	--	--	26.6							
87	1	-1.76	--	--	--	25.6							
89	4	0.00	28.4	--	--	--							
93	2	-1.32	--	26.3	--	--							
96	4	-0.06	28.3	--	--	--							
97	4	0.50	29.2	--	--	--							
102	0	-3.64	--	22.6	--	--							
113	3	1.00	30.0	--	--	--							
121	4	0.06	--	--	28.5	--							
126	1	1.63	31.0	--	--	--							
127	2	-1.13	26.6	--	--	--							
131	NR	--	--	< 40	--	--							
134	3	0.82	--	29.7	--	--							
138	4	0.00	--	--	28.4	--							
141	0	2.32	--	32.1	--	--							
142	4	0.38	--	--	29.0	--							
144	0	-2.45	24.5	--	--	--							
145	0	11.04	--	46.0	--	--							
146	3	0.63	--	29.4	--	--							
147	4	-0.25	--	--	28.0	--							
151	4	-0.38	--	--	27.8	--							
154	4	0.25	28.8	--	--	--							
190	4	0.19	28.7	--	--	--							
191	4	0.50	--	--	29.2	--							
193	0	-3.26	23.2	--	--	--							
198	4	0.44	29.1	--	--	--							
203	3	-0.56	27.5	--	--	--							
204	1	-1.51	--	--	26.0	--							
212	3	-0.63	--	--	27.4	--							
215	0	-2.13	25.0	--	--	--							
219	3	1.00	--	--	30.0	--							
234	3	0.63	29.4	--	--	--							
236	NR	--	--	< 45	--	--							

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : B (Boron) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY					
N =	0	24	6	3. AA: graphite furnace	MPV = 26.4
Minimum =		12.0	25.0	4. ICP	F-pseudosigma = 3.0
Maximum =		64.0	28.0	6. ICP/MS	N = 30
Median =		26.3			Uh = 28.2
F-pseudosigma =		3.4			Lh = 24.2


Lab	Rating	Z-value	3	4	6
1	3	0.61	--	28.2	--
3	1	1.85	--	31.9	--
11	4	-0.13	--	26.0	--
24	4	-0.47	--	25.0	--
26	3	-0.74	--	24.2	--
32	4	0.24	--	--	27.1
39	0	-3.00	--	17.5	--
42	NR	--	--	--	< 30
46	4	-0.30	--	25.5	--
48	4	-0.03	--	--	26.3
50	3	0.54	--	--	28.0
57	4	0.34	--	27.4	--
70	NR	--	--	< 50	--
86	0	-4.52	--	13.0	--
127	4	-0.30	--	25.5	--
131	3	-0.81	--	24.0	--
134	4	0.03	--	26.5	--
138	4	0.30	--	27.3	--
141	0	7.99	--	50.1	--
142	NR	--	--	< 30	--
145	0	2.56	--	34.0	--
154	3	0.88	--	29.0	--
180	0	12.41	--	63.2	--
191	4	-0.47	--	--	25.0
212	3	0.61	--	28.2	--
234	4	0.20	--	27.0	--
236	3	-0.81	--	24.0	--
247	0	12.68	--	64.0	--
255	2	-1.21	--	22.8	--
259	0	-4.86	--	12.0	--
265	4	-0.47	--	--	25.0
273	0	-2.33	--	19.5	--
282	4	0.34	--	--	27.4
287	0	-8.47	< 1	--	--

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte: Ba (Barium) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY									
1. AA: direct, air	4. ICP								
2. AA: direct, nitrous oxide	5. DCP								
3. AA: graphite furnace	6. ICP/MS								
N =	1	0	2	34	1	16			
Minimum =	50.0		38.5	34.1	43.0	35.7			
Maximum =			40.8	46.9		40.9			
Median =				37.8		38.2			
F-pseudostigma =				1.5		1.4			
1. AA: direct, air	MPV =	38.1							
2. AA: direct, nitrous oxide	F-pseudostigma =	1.9							
3. AA: graphite furnace	N =	54							
4. ICP	Uh =	39.5							
5. DCP	Lh =	37.0							
6. ICP/MS									
Lab	Rating	Z-value	1	2	3	4	5	6	
1	4	0.03	--	--	--	38.1	--	--	
3	2	-1.43	--	--	--	35.4	--	--	
11	3	-0.57	--	--	--	37.0	--	--	
13	4	-0.40	--	--	--	37.3	--	--	
23	4	-0.13	--	--	--	37.8	--	--	
24	3	0.78	--	--	--	39.5	--	--	
26	4	0.24	--	--	--	38.5	--	--	
32	4	0.08	--	--	--	--	--	38.2	
33	0	2.67	--	--	--	--	43.0	--	
39	3	-0.78	--	--	--	36.6	--	--	
42	4	-0.03	--	--	--	--	--	38.0	
46	3	0.57	--	--	--	39.1	--	--	
48	4	-0.03	--	--	--	--	--	38.0	
50	2	1.32	--	--	--	--	--	40.5	
57	4	-0.40	--	--	--	37.3	--	--	
59	4	0.46	--	--	--	--	--	38.9	
70	4	0.03	--	--	--	--	--	38.1	
76	4	0.33	--	--	--	--	--	38.7	
81	2	-1.11	--	--	--	36.0	--	--	
86	4	-0.24	--	--	--	37.6	--	--	
87	4	0.24	--	--	38.5	--	--	--	
89	NR		--	--	< 50	--	--	--	
96	NR		--	--	< 100	--	--	--	
97	2	1.48	--	--	40.8	--	--	--	
102	0	3.16	--	--	--	43.9	--	--	
113	4	-0.13	--	--	--	37.8	--	--	
121	3	-0.57	--	--	--	--	--	37.0	
126	NR		--	--	< 200	--	--	--	
127	3	-0.89	--	--	--	36.4	--	--	
131	1	-1.65	--	--	--	35.0	--	--	
134	4	-0.40	--	--	--	37.3	--	--	
138	4	0.08	--	--	--	38.2	--	--	
140	0	6.45	50.0	--	--	--	--	--	
141	2	1.48	--	--	--	40.8	--	--	
142	2	1.11	--	--	--	--	--	40.1	
145	3	0.51	--	--	--	39.0	--	--	
146	3	0.84	--	--	--	39.6	--	--	
151	4	-0.24	--	--	--	--	--	37.6	
154	2	-1.27	--	--	--	35.7	--	--	
180	2	1.48	--	--	--	40.8	--	--	
191	3	0.51	--	--	--	--	--	39.0	
198	NR		--	--	--	< 50	--	--	
203	2	1.05	--	--	--	40.0	--	--	
204	2	-1.27	--	--	--	--	--	35.7	
212	4	0.35	--	--	--	38.7	--	--	
215	3	-0.57	--	--	--	37.0	--	--	
219	2	-1.11	--	--	--	--	--	36.0	
234	4	-0.24	--	--	--	37.6	--	--	
236	3	0.51	--	--	--	39.0	--	--	
241	3	0.78	--	--	--	--	--	39.5	

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Be (Beryllium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY					
N =	1	7	26	16	
Minimum =	12.0	8.9	9.0	8.0	2. AA: direct, nitrous oxide
Maximum =		13.5	13.0	13.0	3. AA: graphite furnace
Median =		10.5	10.9	10.8	4. ICP
F-pseudostigma =		0.6	0.3	0.5	6. ICP/MS
					MPV = 10.8
					F-pseudostigma = 0.4
					Rating Criterion = 0.5
					N = 50
					Uh = 11.1
					Lh = 10.5


Lab	Rating	Z-value	2	3	4	6
1	4	0.31	--	--	--	11.0
3	4	-0.19	--	--	10.7	--
11	4	0.37	--	--	11.0	--
13	4	0.37	--	--	11.0	--
23	0	-2.02	--	--	9.7	--
26	3	0.74	--	--	11.2	--
32	3	-0.74	--	--	--	10.4
42	4	0.00	--	--	--	10.8
46	4	-0.19	--	--	10.7	--
48	4	-0.19	--	--	--	10.7
50	4	-0.37	--	--	--	10.6
57	3	0.74	--	--	11.2	--
59	1	1.85	--	--	--	11.8
70	0	2.22	--	--	--	12.0
76	3	0.96	--	--	--	11.3
81	2	-1.48	--	10.0	--	--
86	4	-0.37	--	--	10.6	--
89	0	5.00	--	13.5	--	--
93	1	-1.59	--	--	9.9	--
96	3	-0.74	--	10.4	--	--
102	2	-1.11	--	--	10.2	--
113	3	0.56	--	--	11.1	--
114	0	2.22	12.0	--	--	--
126	4	-0.37	--	10.6	--	--
127	4	-0.19	--	--	10.7	--
134	3	-0.56	--	--	10.5	--
138	3	-0.56	--	--	10.5	--
141	2	1.48	--	--	11.6	--
142	4	0.00	--	--	--	10.8
144	0	-3.52	--	8.9	--	--
145	4	0.37	--	--	11.0	--
146	4	0.19	--	--	10.9	--
151	2	-1.30	--	--	--	10.1
154	4	0.19	--	--	10.9	--
180	0	4.07	--	--	13.0	--
191	4	0.19	--	--	--	10.9
193	3	-0.56	--	10.5	--	--
198	2	1.11	--	11.4	--	--
204	0	-5.19	--	--	--	8.0
212	1	1.67	--	--	11.7	--
215	4	0.37	--	--	11.0	--
219	4	0.37	--	--	--	11.0
234	4	0.00	--	--	10.8	--
236	4	0.37	--	--	11.0	--
241	1	-1.76	--	--	--	9.9
247	0	4.07	--	--	--	13.0
265	4	0.37	--	--	11.0	--
282	4	-0.19	--	--	--	10.7
292	0	-3.33	--	--	9.0	--
304	4	-0.19	--	--	10.7	--

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Ca (Calcium) Concentration Unit : mg/L




ANALYTICAL METHOD CODE



SUMMARY																	
Lab	Rating	Z-value	0	1	4	6	12	20	Lab	Rating	Z-value	0	1	4	6	12	20
			N =	5	14	46	6	2	2	0. Other		MPV = 25.5					
			Minimum =	15.5	23.9	24.0	25.7	24.7	25.6	1. AA: direct, air		F-pseudosigma = 0.8					
			Maximum =	33.1	27.4	33.4	29.2	25.8	25.8	4. ICP		Rating Criterion = 1.3					
			Median =	25.4		25.5							N = 75				
			F-pseudosigma =	0.6		0.8							Uh = 26.3				
													Lh = 25.2				
													20. Titrate: colorimetric				
1	4	0.31	--	--	25.9	--	--	--	204	4	0.24	--	--	25.8	--	--	--
3	4	-0.47	--	--	24.9	--	--	--	209	4	-0.24	--	--	25.2	--	--	--
11	4	0.39	--	--	26.0	--	--	--	212	2	-1.02	--	--	24.2	--	--	--
12	4	0.39	--	--	26.0	--	--	--	215	4	0.00	--	--	25.5	--	--	--
13	4	-0.39	--	--	25.0	--	--	--	219	4	0.16	--	--	--	25.7	--	--
23	2	1.49	--	27.4	--	--	--	--	227	4	-0.24	--	--	25.2	--	--	--
24	4	-0.08	--	--	25.4	--	--	--	234	4	-0.24	--	--	25.2	--	--	--
26	0	-7.84	15.5	--	--	--	--	--	236	3	-0.88	--	--	24.4	--	--	--
32	4	0.16	--	--	--	25.7	--	--	241	4	0.24	--	--	--	--	25.8	--
33	2	1.18	27.0	--	--	--	--	--	247	3	-0.55	--	--	24.8	--	--	--
39	2	-1.18	--	--	24.0	--	--	--	254	4	0.43	--	--	26.1	--	--	--
42	3	0.63	--	--	26.3	--	--	--	255	4	-0.16	--	--	25.3	--	--	--
45	2	-1.25	--	23.9	--	--	--	--	259	4	-0.08	--	--	25.4	--	--	--
46	4	-0.31	--	--	25.1	--	--	--	265	4	0.00	--	--	25.5	--	--	--
48	0	2.90	--	--	--	29.2	--	--	268	4	0.00	--	25.5	--	--	--	--
51	2	1.33	--	27.2	--	--	--	--	273	2	1.49	--	--	27.4	--	--	--
57	4	-0.24	--	--	25.2	--	--	--	274	4	0.09	--	--	--	--	--	25.6
59	2	1.41	--	--	--	27.3	--	--	279	4	0.24	--	--	--	--	--	25.8
64	0	6.20	--	--	33.4	--	--	--	282	4	0.24	--	--	--	25.8	--	--
70	3	0.78	--	--	26.5	--	--	--	287	3	0.75	--	26.5	--	--	--	--
76	4	-0.31	--	25.1	--	--	--	--	292	4	0.24	--	25.8	--	--	--	--
81	3	0.63	--	--	26.3	--	--	--	302	4	-0.26	25.2	--	--	--	--	--
86	4	0.24	--	--	25.8	--	--	--	304	3	0.55	--	--	26.2	--	--	--
87	4	-0.47	24.9	--	--	--	--	--	307	2	-1.18	--	24.0	--	--	--	--
89	3	-0.63	--	--	--	--	24.7	--	309	1	1.65	--	--	27.6	--	--	--
93	3	-0.71	--	--	24.6	--	--	--									
102	2	1.18	--	--	27.0	--	--	--									
107	4	-0.24	--	25.2	--	--	--	--									
109	4	-0.39	--	25.0	--	--	--	--									
113	4	0.31	--	--	25.9	--	--	--									
121	4	-0.08	--	--	25.4	--	--	--									
126	4	-0.39	--	25.0	--	--	--	--									
127	4	-0.08	--	--	25.4	--	--	--									
131	4	-0.16	--	--	25.3	--	--	--									
134	4	-0.16	--	--	25.3	--	--	--									
138	4	-0.24	--	--	25.2	--	--	--									
140	4	0.00	--	25.5	--	--	--	--									
141	2	1.10	--	--	26.9	--	--	--									
142	2	1.18	--	--	27.0	--	--	--									
145	3	0.86	--	--	26.6	--	--	--									
146	4	-0.24	--	--	25.2	--	--	--									
151	4	0.16	--	25.7	--	--	--	--									
154	4	-0.24	--	--	25.2	--	--	--									
180	3	0.78	--	--	26.5	--	--	--									
185	4	-0.32	--	25.1	--	--	--	--									
190	0	5.96	33.1	--	--	--	--	--									
191	3	0.63	--	--	--	26.3	--	--									
193	4	0.24	--	--	25.8	--	--	--									
198	3	0.63	--	--	26.3	--	--	--									
203	3	-0.57	--	--	24.8	--	--	--									

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Cd (Cadmium) Concentration Unit : mg/L



SUMMARY

N =	2	18	29	19
Minimum =	22.9	11.9	19.0	21.6
Maximum =	26.0	29.0	27.0	25.4
Median =		24.3	23.9	24.0
F-pseudosigma =		3.3	1.7	0.8


- 1. AA: direct, air
- 3. AA: graphite furnace
- 4. ICP
- 6. ICP/MS

MPV =	24.0
F-pseudosigma =	1.6
N =	68
Uh =	25.1
Lh =	23.0

Lab	Rating	Z-value	1	3	4	6	Lab	Rating	Z-value	1	3	4	6
1	4	-0.13	--	--	--	23.8	215	4	0.00	--	--	24.0	--
3	3	-0.82	--	--	22.7	--	219	4	0.00	--	--	--	24.0
10	4	0.00	--	24.0	--	--	227	4	0.50	--	--	24.8	--
11	3	0.63	--	--	25.0	--	234	3	0.63	--	--	25.0	--
12	4	0.31	--	24.5	--	--	236	0	-2.51	--	--	20.0	--
13	0	-2.20	--	--	20.5	--	241	3	0.88	--	--	--	25.4
23	0	-3.14	--	--	19.0	--	247	1	1.88	--	--	27.0	--
24	4	0.13	--	--	24.2	--	255	4	0.31	--	--	--	24.5
32	4	0.06	--	--	--	24.1	259	4	0.00	--	--	24.0	--
39	2	-1.25	--	22.0	--	--	265	3	-0.63	--	--	--	23.0
42	4	0.00	--	--	--	24.0	273	4	-0.44	--	--	23.3	--
46	3	-0.94	--	22.5	--	--	277	1	-1.76	--	--	21.2	--
48	4	-0.19	--	--	--	23.7	282	3	0.63	--	25.0	--	--
50	3	0.75	--	--	--	25.2	287	0	3.14	--	29.0	--	--
57	4	-0.50	--	--	23.2	--	304	3	-0.63	--	--	23.0	--
59	3	0.88	--	--	--	25.4	305	0	-7.57	--	11.9	--	--
70	4	-0.38	--	--	--	23.4	307	1	-2.01	--	20.8	--	--
76	4	0.22	--	--	--	24.4	308	0	-2.07	--	20.7	--	--
81	4	0.00	--	24.0	--	--							
86	3	-0.88	--	--	22.6	--							
87	3	0.69	--	--	25.1	--							
89	2	1.44	--	26.3	--	--							
93	2	-1.44	--	--	21.7	--							
96	4	0.38	--	24.6	--	--							
102	3	0.88	--	--	25.4	--							
113	4	0.50	--	--	24.8	--							
114	2	1.25	26.0	--	--	--							
121	3	0.69	--	--	--	25.1							
126	0	2.45	--	27.9	--	--							
127	2	1.25	--	--	26.0	--							
131	2	-1.25	--	--	22.0	--							
134	4	-0.19	--	--	23.7	--							
138	4	-0.13	--	--	--	23.8							
140	3	-0.69	22.9	--	--	--							
141	1	1.51	--	--	26.4	--							
142	3	0.56	--	--	--	24.9							
144	1	1.63	--	26.6	--	--							
145	1	1.88	--	--	27.0	--							
146	4	-0.06	--	--	23.9	--							
147	3	-0.56	--	--	--	23.1							
151	4	0.06	--	--	--	24.1							
154	3	-0.63	--	--	23.0	--							
180	1	1.76	--	--	26.8	--							
190	2	1.25	--	26.0	--	--							
191	4	-0.31	--	--	--	23.5							
193	0	-2.20	--	20.5	--	--							
198	0	2.20	--	--	27.5	--							
203	2	-1.38	--	21.8	--	--							
204	1	-1.51	--	--	--	21.6							
212	4	-0.19	--	--	23.7	--							

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Co (Cobalt) Concentration Unit : mg/L



SUMMARY

N =	4	25	16
Minimum =	12.0	11.0	11.6
Maximum =	15.6	18.0	14.9
Median =		13.5	13.3
F-pseudosigma =		0.9	0.7


- 3. AA: graphite furnace
- 4. ICP
- 6. ICP/MS

MPV =	13.3
F-pseudosigma =	0.9
N =	45
Uh =	14.0
Lh =	12.8


Lab	Rating	Z-value	3	4	6
1	4	0.00	--	--	13.3
3	3	-0.56	--	12.8	--
11	4	-0.34	--	13.0	--
13	0	3.37	--	16.3	--
24	2	1.01	--	14.2	--
26	2	1.35	--	14.5	--
32	4	0.34	--	--	13.6
39	0	-2.59	--	11.0	--
42	4	0.34	--	--	13.6
48	2	-1.12	--	--	12.3
50	4	0.00	--	--	13.3
57	4	-0.34	--	13.0	--
59	4	0.00	--	--	13.3
70	2	-1.01	--	--	12.4
76	4	-0.25	--	--	13.1
81	3	0.79	--	14.0	--
86	2	-1.01	--	12.4	--
89	0	2.59	15.6	--	--
96	4	-0.45	12.9	--	--
102	1	1.91	--	15.0	--
121	1	-1.91	--	--	11.6
127	4	-0.34	--	13.0	--
131	2	-1.46	--	12.0	--
134	4	0.11	--	13.4	--
138	3	0.79	--	14.0	--
141	3	0.90	--	14.1	--
142	3	-0.56	--	--	12.8
144	2	-1.24	12.2	--	--
145	0	5.28	--	18.0	--
146	4	0.22	--	13.5	--
154	4	-0.22	--	13.1	--
180	0	4.83	--	17.6	--
191	4	0.45	--	--	13.7
212	3	-0.79	--	--	12.6
215	2	-1.46	12.0	--	--
219	3	0.79	--	--	14.0
234	4	-0.34	--	13.0	--
236	0	3.04	--	16.0	--
247	4	0.45	--	--	13.7
254	3	0.79	--	14.0	--
259	4	-0.34	--	13.0	--
265	4	-0.34	--	--	13.0
273	3	0.90	--	14.1	--
277	1	-1.69	--	11.8	--
282	1	1.80	--	--	14.9

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Cr (Chromium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	2	14	28	20	2. AA: direct, nitrous oxide	MPV =	26.8		
Minimum =	30.0	4.8	18.0	23.6	3. AA: graphite furnace	F-pseudosigma =	1.8		
Maximum =	33.0	41.9	35.2	29.8	4. ICP	N =	64		
Median =		25.5	26.8	27.6	6. ICP/MS	Uh =	28.1		
F-pseudosigma =		2.9	0.9	1.5		Lh =	25.7		


Lab	Rating	Z-value	2	3	4	6	Lab	Rating	Z-value	2	3	4	6
1	3	0.66	--	--	--	28.0	234	4	-0.11	--	--	26.6	--
3	4	0.17	--	--	27.1	--	236	4	0.11	--	--	27.0	--
10	3	0.66	--	28.0	--	--	241	4	-0.22	--	--	--	26.4
11	3	0.66	--	--	28.0	--	247	3	0.91	--	--	--	28.5
13	4	-0.11	--	--	26.6	--	255	1	1.65	--	--	--	29.8
23	4	-0.44	--	--	26.0	--	259	0	4.63	--	--	35.2	--
24	0	2.81	--	--	31.9	--	265	4	0.11	--	--	--	27.0
32	4	-0.22	--	--	--	26.4	273	0	-4.85	--	--	18.0	--
39	4	-0.44	--	--	26.0	--	277	3	-0.88	--	--	25.2	--
42	4	0.39	--	--	--	27.5	282	2	1.16	--	--	--	28.9
45	0	8.31	--	41.9	--	--	287	3	0.83	--	28.3	--	--
46	4	0.17	--	--	27.1	--	292	3	-0.99	--	--	25.0	--
48	4	-0.39	--	--	--	26.1	304	4	-0.44	--	--	26.0	--
50	2	1.10	--	--	--	28.8	305	2	-1.34	--	24.4	--	--
57	3	-0.99	--	--	25.0	--	307	2	-1.27	--	24.5	--	--
59	2	1.49	--	--	--	29.5							
70	4	-0.44	--	--	--	26.0							
76	3	0.78	--	--	--	28.2							
81	1	1.76	--	30.0	--	--							
86	4	-0.11	--	--	26.6	--							
87	4	0.00	--	--	26.8	--							
89	3	-0.94	--	25.1	--	--							
93	4	-0.22	--	--	26.4	--							
96	3	-0.61	--	25.7	--	--							
102	3	0.77	--	--	28.2	--							
114	0	3.41	33.0	--	--	--							
121	2	-1.27	--	--	--	24.5							
126	3	-0.83	--	25.3	--	--							
127	4	0.28	--	--	27.3	--							
131	1	-1.54	--	--	24.0	--							
134	4	0.06	--	--	26.9	--							
138	4	0.17	--	--	27.1	--							
140	1	1.76	30.0	--	--	--							
141	3	-0.99	--	--	25.0	--							
142	4	0.00	--	--	--	26.8							
144	1	1.71	--	29.9	--	--							
145	0	2.86	--	--	32.0	--							
146	3	0.55	--	--	27.8	--							
151	1	-1.76	--	--	--	23.6							
154	4	0.00	--	--	26.8	--							
180	4	0.00	--	--	26.8	--							
190	4	-0.22	--	26.4	--	--							
191	4	0.50	--	--	--	27.7							
193	0	-2.15	--	22.9	--	--							
198	NR	--	--	--	< 50	--							
203	0	-12.11	--	4.8	--	--							
204	3	-0.66	--	--	--	25.6							
212	4	0.44	--	--	--	27.6							
215	0	-2.64	--	22.0	--	--							
219	3	0.66	--	--	--	28.0							

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Cu (Copper) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	1	5	12	32	19	0. Other	MPV =	33.4	
Minimum =	31.0	30.5	30.5	13.0	27.8	1. AA: direct, air	F-pseudostigma =	2.5	
Maximum =		41.0	37.0	47.0	36.1	3. AA: graphite furnace	N =	69	
Median =			33.1	34.1	32.9	4. ICP	Uh =	35.0	
F-pseudostigma =			2.7	2.0	3.0	6. ICP/MS	Lh =	31.6	


Lab	Rating	Z-value	0	1	3	4	6	Lab	Rating	Z-value	0	1	3	4	6
1	4	-0.28	--	--	32.7	--	--	212	2	-1.36	--	--	--	--	30.0
3	4	-0.28	--	--	--	32.7	--	215	0	-8.17	--	--	--	13.0	--
10	3	0.64	--	--	35.0	--	--	219	4	-0.16	--	--	--	--	33.0
11	3	0.64	--	--	--	35.0	--	227	1	1.64	--	--	--	37.5	--
12	3	-0.96	31.0	--	--	--	--	234	4	0.44	--	--	--	34.5	--
13	0	-5.64	--	--	--	19.3	--	236	2	-1.36	--	--	--	30.0	--
23	3	-0.92	--	--	--	31.1	--	241	4	-0.36	--	--	--	32.5	--
24	1	1.80	--	--	--	37.9	--	247	3	0.86	--	--	--	35.6	--
26	0	-12.56	--	--	--	< 2	--	254	4	-0.08	--	--	--	33.2	--
32	3	0.64	--	--	--	--	35.0	255	4	-0.20	--	--	--	32.9	--
39	4	-0.28	--	--	--	32.7	--	259	0	-2.56	--	--	--	27.0	--
42	4	0.08	--	--	--	--	33.6	265	3	0.64	--	--	--	35.0	--
45	4	0.24	--	34.0	--	--	--	273	3	0.76	--	--	--	35.3	--
46	4	0.40	--	--	--	34.4	--	277	2	-1.06	--	--	--	30.8	--
48	2	-1.24	--	--	--	--	30.3	282	2	1.08	--	--	--	36.1	--
50	3	0.72	--	--	--	--	35.2	287	3	0.64	--	35.0	--	--	--
57	4	0.40	--	--	--	34.4	--	292	4	0.24	--	--	--	34.0	--
59	3	0.60	--	--	--	--	34.9	304	3	0.64	--	--	--	35.0	--
70	1	-1.60	--	--	--	--	29.4	305	3	-0.71	--	--	31.6	--	--
81	2	1.44	--	--	37.0	--	--	307	3	-0.76	--	--	31.5	--	--
86	4	-0.32	--	--	--	32.6	--	308	2	1.44	--	--	37.0	--	--
87	3	0.56	--	--	--	34.8	--								
89	3	-0.84	--	--	31.3	--	--								
96	4	-0.28	--	--	32.7	--	--								
97	3	0.84	--	--	35.5	--	--								
102	3	-0.56	--	--	--	32.0	--								
113	4	0.32	--	--	--	34.2	--								
114	2	-1.16	--	30.5	--	--	--								
121	2	-1.20	--	--	--	--	30.4								
126	0	3.04	--	41.0	--	--	--								
127	4	0.48	--	--	--	34.6	--								
131	0	5.44	--	--	--	47.0	--								
134	4	-0.40	--	--	--	32.4	--								
138	4	-0.08	--	--	--	33.2	--								
140	4	0.36	--	34.3	--	--	--								
141	3	0.80	--	--	--	35.4	--								
142	3	-0.76	--	--	--	--	31.5								
144	2	-1.16	--	--	30.5	--	--								
145	0	3.04	--	--	--	41.0	--								
146	4	-0.36	--	--	--	32.5	--								
147	3	-0.52	--	--	--	--	32.1								
151	4	0.00	--	--	--	--	33.4								
154	2	1.28	--	--	--	36.6	--								
180	2	1.48	--	--	--	37.1	--								
190	4	0.00	--	--	33.4	--	--								
191	4	-0.44	--	--	--	--	32.3								
193	4	0.40	--	--	34.4	--	--								
198	NR	--	--	--	--	< 50	--								
203	4	-0.16	--	--	--	33.0	--								
204	0	-2.24	--	--	--	--	27.8								

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Fe (Iron) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
N =		2	10	3	47	1	5	0. Other	MPV = 48.9								
Minimum =		49.0	32.5	41.0	27.0	50.0	34.5	1. AA: direct, air	F-pseudosigma = 6.2								
Maximum =		54.8	60.0	56.0	90.0		47.7	3. AA: graphite furnace	N = 68								
Median =			51.3		48.8			4. ICP	Uh = 51.4								
F-pseudosigma =			10.8		5.4			5. DCP	Lh = 43.1								
								6. ICP/MS									
Lab	Rating	Z-value	0	1	3	4	5	6	Lab	Rating	Z-value	0	1	3	4	5	6
1	4	0.33	--	--	--	50.9	--	--	204	4	-0.39	--	--	--	46.5	--	--
3	0	4.03	--	--	--	73.7	--	--	212	4	0.39	--	--	--	51.3	--	--
10	2	1.15	--	56.0	--	--	--	--	215	0	-3.56	--	--	--	27.0	--	--
11	0	6.68	--	--	--	90.0	--	--	227	3	0.52	--	--	--	52.1	--	--
13	3	-0.75	--	--	--	44.3	--	--	234	4	0.11	--	--	--	49.6	--	--
21	3	0.96	54.8	--	--	--	--	--	236	4	-0.31	--	--	--	47.0	--	--
23	3	-0.76	--	--	--	44.2	--	--	241	4	0.02	49.0	--	--	--	--	--
24	4	0.21	--	--	--	50.2	--	--	247	1	-1.93	--	--	--	37.0	--	--
26	1	-1.71	--	--	--	38.4	--	--	254	4	0.24	--	--	--	50.4	--	--
33	4	0.18	--	--	--	--	50.0	--	255	4	0.33	--	--	--	50.9	--	--
39	2	-1.33	--	--	--	40.7	--	--	259	4	-0.18	--	--	--	47.8	--	--
42	2	-1.30	--	--	--	40.9	--	--	265	4	0.18	--	--	--	50.0	--	--
46	4	-0.49	--	--	--	45.9	--	--	273	2	-1.06	--	--	--	42.4	--	--
48	4	-0.47	--	--	--	--	--	46.0	277	3	-0.91	--	--	--	43.3	--	--
50	4	-0.28	--	--	--	--	--	47.2	282	4	-0.20	--	--	--	--	--	47.7
57	4	0.02	--	--	--	49.0	--	--	287	2	-1.45	--	40.0	--	--	--	--
59	0	-2.34	--	--	--	--	--	34.5	292	4	-0.47	--	--	--	46.0	--	--
70	4	0.37	--	--	--	51.2	--	--	304	4	0.02	--	--	--	49.0	--	--
81	2	-1.12	--	--	--	42.0	--	--	307	4	0.18	--	50.0	--	--	--	--
86	0	-2.45	--	--	--	33.8	--	--									
87	4	-0.49	--	--	--	45.9	--	--									
89	3	-0.81	--	--	43.9	--	--	--									
93	3	-0.98	--	--	--	42.9	--	--									
96	1	1.80	--	60.0	--	--	--	--									
97	2	-1.28	--	--	41.0	--	--	--									
102	3	0.94	--	--	--	54.7	--	--									
107	1	1.80	--	60.0	--	--	--	--									
109	3	0.60	--	52.6	--	--	--	--									
113	4	-0.36	--	--	--	46.7	--	--									
114	2	-1.09	--	42.2	--	--	--	--									
121	3	0.67	--	--	--	53.0	--	--									
127	4	0.13	--	--	--	49.7	--	--									
131	4	0.18	--	--	--	50.0	--	--									
134	4	-0.02	--	--	--	48.8	--	--									
138	4	0.42	--	--	--	51.5	--	--									
140	4	0.18	--	50.0	--	--	--	--									
141	1	-1.54	--	--	--	39.4	--	--									
142	2	-1.45	--	--	--	40.0	--	--									
144	2	1.15	--	--	56.0	--	--	--									
145	0	5.05	--	--	--	80.0	--	--									
146	NR	--	--	--	--	< 50	--	--									
147	3	0.67	--	--	--	53.0	--	--									
151	0	-2.67	--	32.5	--	--	--	--									
154	4	-0.39	--	--	--	46.5	--	--									
180	2	1.24	--	--	--	56.5	--	--									
190	2	1.28	--	56.8	--	--	--	--									
191	1	-1.77	--	--	--	--	--	38.0									
193	0	3.72	--	--	--	71.8	--	--									
198	4	0.34	--	--	--	51.0	--	--									
203	4	-0.15	--	--	--	48.0	--	--									

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : K (Potassium) Concentration Unit : mg/L




ANALYTICAL METHOD CODE



SUMMARY																	
		N =	17	35	1	6	2	6	1. AA: direct, air	MPV =	1.52						
		Minimum =	1.40	1.14	1.52	1.37	1.00	1.50	4. ICP	F-pseudosigma =	0.13						
		Maximum =	1.96	2.70		1.63	1.54	2.25	5. DCP	N =	67						
		Median =	1.58	1.50					6. ICP/MS	Uh =	1.65						
		F-pseudosigma =	0.18	0.10					7. Ion chromatography	Lh =	1.47						
									12. Flame emission								
Lab	Rating	Z-value	1	4	5	6	7	12	Lab	Rating	Z-value	1	4	5	6	7	12
1	4	-0.37	1.47	--	--	--	--	--	219	3	-0.90	--	--	--	1.40	--	--
3	4	0.37	--	1.57	--	--	--	--	234	4	-0.15	--	1.50	--	--	--	--
11	3	0.67	--	1.61	--	--	--	--	236	3	-0.75	--	1.42	--	--	--	--
13	4	0.37	--	1.57	--	--	--	--	241	1	1.95	--	--	--	--	--	1.78
23	4	-0.30	--	1.48	--	--	--	--	247	4	-0.37	--	1.47	--	--	--	--
24	4	-0.30	--	1.48	--	--	--	--	254	4	0.45	1.58	--	--	--	--	--
26	0	-3.90	--	--	--	--	1.00	--	259	4	-0.37	--	1.47	--	--	--	--
32	4	-0.15	--	--	--	1.50	--	--	265	4	0.22	--	1.55	--	--	--	--
33	4	0.00	--	--	1.52	--	--	--	268	0	3.30	1.96	--	--	--	--	--
39	3	-0.75	--	1.42	--	--	--	--	273	4	-0.22	--	1.49	--	--	--	--
42	3	-0.52	--	1.45	--	--	--	--	274	2	1.35	--	--	--	--	--	1.70
45	2	1.27	1.69	--	--	--	--	--	279	4	-0.15	--	--	--	--	--	1.50
46	4	-0.15	--	1.50	--	--	--	--	282	3	0.82	--	--	--	1.63	--	--
48	4	-0.37	--	--	--	1.47	--	--	287	1	1.57	1.73	--	--	--	--	--
51	0	5.47	--	--	--	--	--	2.25	292	3	0.60	1.60	--	--	--	--	--
57	0	8.84	--	2.70	--	--	--	--	302	4	0.11	--	--	--	--	1.54	--
59	3	0.60	--	--	--	1.60	--	--	304	4	-0.15	--	1.50	--	--	--	--
64	4	-0.37	1.47	--	--	--	--	--	309	0	6.00	--	2.32	--	--	--	--
70	4	-0.30	--	1.48	--	--	--	--									
81	1	1.87	--	1.77	--	--	--	--									
86	4	-0.15	--	1.50	--	--	--	--									
87	4	-0.22	1.49	--	--	--	--	--									
89	4	0.15	--	--	--	--	--	1.54									
93	0	-2.17	--	1.23	--	--	--	--									
97	3	-0.52	1.45	--	--	--	--	--									
102	0	-2.85	--	1.14	--	--	--	--									
107	0	2.40	1.84	--	--	--	--	--									
109	2	1.42	1.71	--	--	--	--	--									
113	4	0.37	--	1.57	--	--	--	--									
121	4	-0.37	1.47	--	--	--	--	--									
127	4	0.37	--	1.57	--	--	--	--									
131	3	0.60	--	1.60	--	--	--	--									
134	4	-0.19	1.50	--	--	--	--	--									
138	3	-0.67	--	1.43	--	--	--	--									
140	3	-0.90	1.40	--	--	--	--	--									
141	2	1.12	--	1.67	--	--	--	--									
142	4	-0.22	--	1.49	--	--	--	--									
145	3	0.67	--	1.61	--	--	--	--									
146	4	0.45	--	1.58	--	--	--	--									
151	4	-0.37	1.47	--	--	--	--	--									
154	0	4.35	--	2.10	--	--	--	--									
180	NR	--	--	< 1.63	--	--	--	--									
185	2	1.27	1.69	--	--	--	--	--									
191	2	-1.12	--	--	--	1.37	--	--									
193	2	1.50	1.72	--	--	--	--	--									
198	0	3.67	--	2.01	--	--	--	--									
203	3	0.75	--	--	--	--	--	1.62									
204	4	-0.07	--	1.51	--	--	--	--									
209	0	4.04	--	2.06	--	--	--	--									
212	3	-0.75	--	1.42	--	--	--	--									

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Li (Lithium) Concentration Unit : mg/L



SUMMARY

N = 12 6
 Minimum = 7.30 7.15
 Maximum = 17.80 8.00
 Median = 9.19
 F-pseudosigma = 1.84


4. ICP
 6. ICP/MS

MPV = 8.57
 F-pseudosigma = 1.93
 N = 18
 Uh = 10.50
 Lh = 7.90

Lab	Rating	Z-value	4	6
1	4	-0.33	--	7.92
3	NR	--	< 10	--
11	2	1.26	11.00	--
26	4	-0.29	8.00	--
42	4	0.35	9.23	--
50	4	-0.50	--	7.60
57	2	1.32	11.10	--
59	3	-0.73	--	7.15
127	NR	--	< 10	--
131	4	0.17	8.90	--
134	4	0.05	8.66	--
141	3	1.00	10.50	--
142	4	-0.05	8.47	--
145	NR	--	< 20	--
151	4	-0.50	--	7.60
219	4	-0.35	--	7.90
234	0	2.40	13.20	--
236	NR	--	< 10	--
247	4	0.30	9.15	--
254	3	-0.66	7.30	--
265	4	-0.29	--	8.00
273	0	4.79	17.80	--

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued


Analyte : Mg (Magnesium) Concentration Unit : mg/L




SUMMARY																	
N =	4	15	45	1	6	2	0. Other	MPV =	5.60								
Minimum =	3.20	4.99	4.92	5.90	5.00	5.25	1. AA: direct, air	F-pseudosigma =	0.21								
Maximum =	6.40	6.25	6.95		6.33	5.59	4. ICP	Rating Criterion =	0.28								
Median =		5.57	5.60				5. DCP	N =	73								
F-pseudosigma =		0.22	0.19				6. ICP/MS	Uh =	5.78								
							12. Flame emission	Lh =	5.49								
Lab	Rating	Z-value	0	1	4	5	6	12	Lab	Rating	Z-value	0	1	4	5	6	12
1	4	0.07	--	--	5.62	--	--	--	212	4	-0.36	--	--	5.50	--	--	--
3	4	-0.07	--	--	5.58	--	--	--	215	4	0.00	--	--	5.60	--	--	--
11	4	-0.29	--	--	5.52	--	--	--	219	0	-2.14	--	--	--	--	5.00	--
12	0	3.57	--	--	6.60	--	--	--	227	4	-0.50	--	--	5.46	--	--	--
13	4	-0.18	--	--	5.55	--	--	--	234	4	0.46	--	--	5.73	--	--	--
23	3	-0.71	--	5.40	--	--	--	--	236	3	-0.61	--	--	5.43	--	--	--
24	4	-0.39	--	--	5.49	--	--	--	241	2	-1.25	--	--	--	--	--	5.25
26	0	-8.57	3.20	--	--	--	--	--	247	4	0.25	--	--	5.67	--	--	--
32	4	-0.36	--	--	--	--	5.50	--	254	3	0.64	--	--	5.78	--	--	--
33	2	1.07	--	--	--	5.90	--	--	255	4	0.00	--	--	5.60	--	--	--
39	0	-2.25	--	--	4.97	--	--	--	259	4	-0.18	--	--	5.55	--	--	--
42	3	-0.68	--	--	5.41	--	--	--	265	3	0.54	--	--	5.75	--	--	--
45	2	-1.18	--	5.27	--	--	--	--	268	3	0.71	--	5.80	--	--	--	--
46	4	-0.36	--	--	5.50	--	--	--	273	1	1.79	--	--	6.10	--	--	--
48	0	2.61	--	--	--	--	6.33	--	274	0	-2.04	5.03	--	--	--	--	--
51	3	0.57	--	5.76	--	--	--	--	279	3	0.71	5.80	--	--	--	--	--
57	4	-0.14	--	--	5.56	--	--	--	282	3	0.64	--	--	--	--	5.78	--
59	0	2.04	--	--	--	--	6.17	--	287	0	-2.18	--	4.99	--	--	--	--
64	0	-2.43	--	--	4.92	--	--	--	292	4	-0.36	--	5.50	--	--	--	--
70	4	-0.36	--	--	5.50	--	--	--	302	0	2.85	6.40	--	--	--	--	--
76	4	0.15	--	5.64	--	--	--	--	304	4	0.00	--	--	5.60	--	--	--
81	3	-0.61	--	--	5.43	--	--	--	307	0	2.32	--	6.25	--	--	--	--
86	3	0.71	--	--	5.80	--	--	--	308	3	-0.86	--	5.36	--	--	--	--
87	3	-0.71	--	5.40	--	--	--	--									
89	4	-0.04	--	--	--	--	--	5.59									
93	1	-1.54	--	--	5.17	--	--	--									
102	0	4.82	--	--	6.95	--	--	--									
107	4	-0.11	--	5.57	--	--	--	--									
109	4	-0.36	--	5.50	--	--	--	--									
113	3	0.64	--	--	5.78	--	--	--									
121	4	0.00	--	--	5.60	--	--	--									
127	4	-0.18	--	--	5.55	--	--	--									
131	4	0.00	--	--	5.60	--	--	--									
134	3	-0.71	--	--	5.40	--	--	--									
138	4	-0.11	--	--	5.57	--	--	--									
140	4	0.00	--	5.60	--	--	--	--									
141	3	0.54	--	--	5.75	--	--	--									
142	2	1.25	--	--	5.95	--	--	--									
145	3	0.89	--	--	5.85	--	--	--									
146	3	-0.54	--	--	5.45	--	--	--									
151	4	0.39	--	5.71	--	--	--	--									
154	4	-0.36	--	--	5.50	--	--	--									
180	3	0.54	--	--	5.75	--	--	--									
185	4	0.32	--	5.69	--	--	--	--									
191	3	0.64	--	--	--	--	5.78	--									
193	4	-0.18	--	--	5.55	--	--	--									
198	3	0.96	--	--	5.87	--	--	--									
203	3	0.79	--	--	5.82	--	--	--									
204	4	0.46	--	--	5.73	--	--	--									
209	4	0.07	--	--	5.62	--	--	--									

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Mn (Manganese) Concentration Unit : mg/L




ANALYTICAL METHOD CODE



SUMMARY																	
N =	1	10	2	39	1	15	0. Other	MPV =	22.0								
Minimum =	17.4	12.0	18.9	5.5	20.0	19.1	1. AA: direct, air	F-pseudosigma =	1.6								
Maximum =		30.0	23.0	30.3		24.4	3. AA: graphite furnace	N =	68								
Median =		22.3		22.0		21.8	4. ICP	Uh =	23.2								
F-pseudosigma =		3.5		1.4		1.6	5. DCP	Lh =	21.0								
							6. ICP/MS										
Lab	Rating	Z-value	0	1	3	4	5	6	Lab	Rating	Z-value	0	1	3	4	5	6
1	3	0.92	--	--	--	--	--	23.5	215	4	0.00	--	--	--	22.0	--	--
3	4	-0.36	--	--	--	21.4	--	--	219	3	-0.61	--	--	--	--	--	21.0
10	1	1.84	--	25.0	--	--	--	--	227	3	0.74	--	--	--	23.2	--	--
11	3	0.62	--	--	--	23.0	--	--	234	1	-1.53	--	--	--	19.5	--	--
13	2	1.35	--	--	--	24.2	--	--	236	4	0.00	--	--	--	22.0	--	--
23	3	-0.61	--	--	--	21.0	--	--	241	4	-0.43	--	--	--	--	--	21.3
24	3	0.68	--	--	--	23.1	--	--	247	4	0.27	--	--	--	--	--	22.4
26	3	0.74	--	--	--	23.2	--	--	254	3	0.68	--	--	--	23.1	--	--
32	4	-0.12	--	--	--	--	--	21.8	255	3	0.80	--	--	--	--	--	23.3
33	2	-1.22	--	--	--	--	20.0	--	259	3	-0.92	--	--	--	20.5	--	--
39	0	-2.14	--	--	--	18.5	--	--	265	3	0.92	--	--	--	23.5	--	--
42	4	0.13	--	--	--	22.2	--	--	273	3	0.98	--	--	--	23.6	--	--
46	4	-0.06	--	--	--	21.9	--	--	277	0	-2.27	--	--	--	18.3	--	--
48	4	-0.49	--	--	--	--	--	21.2	282	3	0.80	--	--	--	--	--	23.3
50	4	-0.18	--	--	--	--	--	21.7	287	3	0.62	--	23.0	--	--	--	--
57	4	-0.18	--	--	--	21.7	--	--	292	3	-0.61	--	--	--	21.0	--	--
59	2	1.47	--	--	--	--	--	24.4	307	0	-6.13	--	12.0	--	--	--	--
70	3	-0.86	--	--	--	--	--	20.6	308	1	1.66	--	24.7	--	--	--	--
76	2	1.01	--	--	--	--	--	23.7	309	0	-10.11	--	--	--	5.5	--	--
81	4	0.00	--	--	--	22.0	--	--									
86	3	-0.61	--	--	--	21.0	--	--									
87	0	-12.50	--	--	--	< 2.0	--	--									
89	1	-1.90	--	--	18.9	--	--	--									
93	0	-2.82	17.4	--	--	--	--	--									
96	2	-1.22	--	20.0	--	--	--	--									
102	2	1.17	--	--	--	23.9	--	--									
107	0	4.91	--	30.0	--	--	--	--									
109	4	-0.33	--	21.5	--	--	--	--									
113	3	0.55	--	--	--	22.9	--	--									
114	4	-0.30	--	21.5	--	--	--	--									
121	3	0.62	--	--	--	23.0	--	--									
126	2	1.23	--	24.0	--	--	--	--									
127	4	-0.12	--	--	--	21.8	--	--									
131	3	-0.61	--	--	--	21.0	--	--									
134	4	0.00	--	--	--	22.0	--	--									
138	4	-0.36	--	--	--	21.4	--	--									
140	2	-1.22	--	20.0	--	--	--	--									
141	3	0.92	--	--	--	23.5	--	--									
142	0	-5.52	--	--	--	13.0	--	--									
145	3	0.62	--	--	--	23.0	--	--									
146	4	0.13	--	--	--	22.2	--	--									
151	4	-0.49	--	--	--	--	--	21.2									
154	4	-0.24	--	--	--	21.6	--	--									
180	3	0.74	--	--	--	23.2	--	--									
190	3	0.62	--	--	23.0	--	--	--									
191	2	1.23	--	--	--	--	--	24.0									
198	0	5.09	--	--	--	30.3	--	--									
203	4	0.00	--	--	--	22.0	--	--									
204	1	-1.78	--	--	--	--	--	19.1									
212	4	-0.18	--	--	--	21.7	--	--									

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Mo (Molybdenum) Concentration Unit : mg/L



SUMMARY

N =	1	24	17
Minimum =	45.9	34.7	39.0
Maximum =		53.5	45.6
Median =		40.1	42.3
F-pseudosigma =		2.8	1.8


- 3. AA: graphite furnace
- 4. ICP
- 6. ICP/MS

MPV =	41.4
F-pseudosigma =	2.6
N =	42
Uh =	42.7
Lh =	39.2


Lab	Rating	Z-value	3	4	6
1	4	0.21	--	--	42.0
3	4	0.07	--	41.6	--
11	4	-0.16	--	41.0	--
13	3	0.57	--	42.9	--
23	4	-0.12	--	41.1	--
24	4	0.42	--	42.5	--
26	4	0.07	--	41.6	--
32	3	-0.74	--	--	39.5
42	4	0.38	--	--	42.4
46	3	-0.86	--	39.2	--
48	3	0.99	--	--	44.0
50	1	1.61	--	--	45.6
57	0	-2.47	--	35.0	--
59	4	-0.08	--	--	41.2
70	2	1.07	--	--	44.2
76	4	-0.07	--	--	41.2
81	2	-1.32	--	38.0	--
86	0	-2.51	--	34.9	--
87	0	-2.59	--	34.7	--
97	1	1.73	45.9	--	--
127	0	4.66	--	53.5	--
131	4	0.22	--	42.0	--
134	3	-0.62	--	39.8	--
138	3	-0.74	--	--	39.5
141	4	-0.16	--	41.0	--
142	4	0.34	--	--	42.3
145	1	1.77	--	46.0	--
146	4	-0.39	--	40.4	--
151	4	0.49	--	--	42.7
154	2	-1.05	--	38.7	--
180	1	1.57	--	45.5	--
212	3	0.88	--	--	43.7
215	2	-1.32	--	38.0	--
219	3	-0.93	--	--	39.0
234	3	-0.66	--	39.7	--
236	0	-2.47	--	35.0	--
241	4	0.34	--	--	42.3
247	2	1.47	--	--	45.2
259	2	-1.16	--	38.4	--
265	4	0.22	--	--	42.0
282	4	0.42	--	--	42.5
292	1	-1.70	--	37.0	--

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Na (Sodium) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
		N =	14	41	1	7	2	6	1. AA: direct, air	MPV =	100						
		Minimum =	20	40	100	97	85	42	4. ICP	F-pseudosigma =	4						
		Maximum =	105	112		113	109	117	5. DCP	Rating Criterion =	5						
		Median =	99	100		102			6. ICP/MS	N =	71						
		F-pseudosigma =	6	4		4			7. Ion chromatography	Uh =	103						
									12. Flame emission	Lh =	97						
Lab	Rating	Z-value	1	4	5	6	7	12	Lab	Rating	Z-value	1	4	5	6	7	12
1	4	-0.24	--	99	--	--	--	--	209	0	2.40	--	112	--	--	--	--
3	4	-0.12	--	99	--	--	--	--	212	3	0.60	--	103	--	--	--	--
11	0	-2.58	--	87	--	--	--	--	215	2	1.40	--	107	--	--	--	--
12	2	-1.20	--	94	--	--	--	--	219	4	-0.20	--	--	--	99	--	--
13	1	2.00	--	110	--	--	--	--	234	4	-0.20	--	99	--	--	--	--
23	2	-1.44	93	--	--	--	--	--	236	3	-0.66	--	97	--	--	--	--
24	3	0.60	--	103	--	--	--	--	241	3	0.60	--	--	--	--	--	103
26	0	-2.92	--	--	--	--	85	--	247	4	-0.25	--	99	--	--	--	--
32	3	-0.70	--	--	--	97	--	--	254	2	1.02	--	105	--	--	--	--
33	4	0.00	--	--	100	--	--	--	259	4	0.10	--	101	--	--	--	--
39	3	-0.62	--	97	--	--	--	--	265	4	-0.06	--	100	--	--	--	--
42	2	-1.32	--	93	--	--	--	--	268	0	-5.70	72	--	--	--	--	--
45	3	0.60	103	--	--	--	--	--	273	1	2.00	--	110	--	--	--	--
46	3	0.80	--	104	--	--	--	--	274	0	3.46	--	--	--	--	--	117
48	0	2.60	--	--	--	113	--	--	279	2	-1.50	--	--	--	--	--	93
51	4	0.24	--	--	--	--	--	101	282	4	0.20	--	--	--	101	--	--
57	3	-0.60	--	97	--	--	--	--	287	2	-1.24	94	--	--	--	--	--
59	2	1.20	--	--	--	106	--	--	292	3	-0.52	97	--	--	--	--	--
64	3	-0.74	96	--	--	--	--	--	302	1	1.76	--	--	--	--	109	--
70	4	0.00	--	100	--	--	--	--	307	4	-0.08	100	--	--	--	--	--
81	4	-0.24	--	99	--	--	--	--	309	0	-11.96	--	40	--	--	--	--
86	4	0.00	--	100	--	--	--	--									
87	4	-0.24	99	--	--	--	--	--									
89	0	-11.68	--	--	--	--	--	42									
93	1	1.80	--	109	--	--	--	--									
97	4	-0.32	98	--	--	--	--	--									
102	0	-3.00	--	85	--	--	--	--									
107	4	0.40	102	--	--	--	--	--									
109	0	-16.07	20	--	--	--	--	--									
113	4	0.02	--	100	--	--	--	--									
121	4	0.00	--	100	--	--	--	--									
126	4	0.40	102	--	--	--	--	--									
127	4	-0.26	--	99	--	--	--	--									
131	4	0.00	--	100	--	--	--	--									
134	3	1.00	105	--	--	--	--	--									
138	4	0.00	--	100	--	--	--	--									
140	4	0.20	101	--	--	--	--	--									
141	4	0.16	--	101	--	--	--	--									
142	2	1.40	--	107	--	--	--	--									
145	4	0.24	--	101	--	--	--	--									
146	3	-1.00	--	95	--	--	--	--									
151	3	0.86	--	--	--	104	--	--									
154	1	-1.60	--	92	--	--	--	--									
180	3	0.60	--	103	--	--	--	--									
185	3	0.90	--	--	--	--	--	105									
191	4	0.40	--	--	--	102	--	--									
193	4	0.40	--	102	--	--	--	--									
198	2	1.20	--	106	--	--	--	--									
203	3	-0.56	--	97	--	--	--	--									
204	4	0.40	--	102	--	--	--	--									

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Ni (Nickel) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	2	11	29	20	1. AA: direct, air	MPV =	22.2		
Minimum =	20.6	19.0	16.9	17.6	3. AA: graphite furnace	F-pseudosigma =	1.6		
Maximum =	21.4	23.3	35.7	24.9	4. ICP	N =	61		
Median =		22.0	22.7	22.0	6. ICP/MS	Uh =	23.3		
F-pseudosigma =		1.5	3.5	1.2		Lh =	21.2		


Lab	Rating	Z-value	1	3	4	6	Lab	Rating	Z-value	1	3	4	6
1	4	-0.19	--	--	--	21.9	236	2	-1.41	--	--	20.0	--
3	4	-0.32	--	--	21.7	--	241	4	-0.45	--	--	--	21.5
11	4	-0.13	--	--	22.0	--	247	1	1.54	--	--	--	24.6
13	0	3.66	--	--	27.9	--	254	4	0.06	--	--	22.3	--
23	0	-2.31	--	--	18.6	--	255	1	1.73	--	--	--	24.9
24	3	0.58	--	--	23.1	--	259	4	0.32	--	--	22.7	--
32	2	1.09	--	--	--	23.9	265	4	-0.13	--	--	--	22.0
39	4	0.00	--	--	22.2	--	273	0	8.67	--	--	35.7	--
42	4	0.45	--	--	--	22.9	277	3	-0.96	--	--	20.7	--
46	0	-3.40	--	--	16.9	--	282	4	-0.13	--	22.0	--	--
48	2	-1.03	--	--	--	20.6	292	1	1.80	--	--	25.0	--
50	4	0.19	--	--	--	22.5	304	3	0.51	--	--	23.0	--
57	3	0.96	--	--	23.7	--	305	2	-1.08	--	20.5	--	--
59	2	1.09	--	--	--	23.9	307	4	0.19	--	22.5	--	--
70	3	-0.64	--	--	--	21.2							
76	4	0.39	--	--	--	22.8							
81	2	-1.41	--	20.0	--	--							
86	3	-0.58	--	--	21.3	--							
87	0	4.24	--	--	28.8	--							
89	3	0.71	--	23.3	--	--							
93	1	-1.73	--	--	19.5	--							
96	3	0.71	--	23.3	--	--							
102	0	2.70	--	--	26.4	--							
113	3	-0.90	--	--	20.8	--							
114	2	-1.03	20.6	--	--	--							
121	2	-1.48	--	--	--	19.9							
126	3	0.71	--	23.3	--	--							
127	3	-0.51	--	21.4	--	--							
131	0	5.01	--	--	30.0	--							
134	4	0.39	--	--	22.8	--							
138	4	-0.45	--	--	--	21.5							
140	3	-0.51	21.4	--	--	--							
141	3	0.77	--	--	23.4	--							
142	4	-0.06	--	--	--	22.1							
144	4	-0.19	--	21.9	--	--							
145	0	4.37	--	--	29.0	--							
146	4	0.26	--	--	22.6	--							
151	3	-0.64	--	--	--	21.2							
154	4	-0.13	--	--	22.0	--							
180	0	4.24	--	--	28.8	--							
190	4	0.39	--	22.8	--	--							
191	4	0.26	--	--	--	22.6							
193	NR		--	< 25	--	--							
198	NR		--	--	< 50	--							
203	0	5.01	--	--	30.0	--							
204	0	-2.95	--	--	--	17.6							
212	2	-1.09	--	--	--	20.5							
215	0	-2.06	--	19.0	--	--							
219	4	-0.13	--	--	--	22.0							
234	4	0.13	--	--	22.4	--							

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Pb (Lead) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	1	2	25	18	20	0. Other	MPV =	16.6	
Minimum =	17.0	17.1	14.2	11.2	14.8	1. AA: direct, air	F-pseudosigma =	1.2	
Maximum =		26.5	19.8	29.0	19.2	3. AA: graphite furnace	N =	66	
Median =			16.5	16.3	16.7	4. ICP	Uh =	17.6	
F-pseudosigma =			1.2	1.3	1.1	6. ICP/MS	Lh =	16.0	


Lab	Rating	Z-value	0	1	3	4	6	Lab	Rating	Z-value	0	1	3	4	6
1	4	-0.34	--	--	--	--	16.2	219	3	-0.51	--	--	--	--	16.0
3	4	-0.08	--	--	--	16.5	--	227	4	0.34	17.0	--	--	--	--
10	3	-0.51	--	--	16.0	--	--	234	3	-0.93	--	--	--	15.5	--
11	2	-1.35	--	--	--	15.0	--	236	NR		--	--	--	< 20	--
13	2	1.35	--	--	18.2	--	--	241	4	-0.42	--	--	--	--	16.1
23	3	-0.51	--	--	--	16.0	--	247	0	2.15	--	--	--	--	19.2
32	2	1.10	--	--	--	--	17.9	255	3	0.51	--	--	--	--	17.2
39	3	-0.51	--	--	16.0	--	--	259	4	0.08	--	--	--	16.7	--
42	3	0.93	--	--	--	--	17.7	265	4	-0.08	--	--	--	--	16.5
46	3	-0.76	--	--	15.7	--	--	273	1	-1.85	--	--	--	14.4	--
48	3	0.84	--	--	--	--	17.6	277	0	-4.55	--	--	--	11.2	--
50	4	-0.42	--	--	--	--	16.1	282	3	-0.51	--	--	16.0	--	--
57	4	-0.42	--	--	--	16.1	--	287	0	2.70	--	--	19.8	--	--
59	4	0.08	--	--	--	--	16.7	292	3	-0.51	--	--	16.0	--	--
70	1	1.94	--	--	--	--	18.9	304	0	9.61	--	--	--	28.0	--
76	3	0.59	--	--	--	--	17.3	305	1	-2.03	--	--	14.2	--	--
81	4	0.34	--	--	17.0	--	--	307	4	-0.08	--	--	16.5	--	--
86	3	0.67	--	--	17.4	--	--	308	0	2.19	--	--	19.2	--	--
87	3	0.84	--	--	--	17.6	--	309	0	10.45	--	--	--	29.0	--
89	3	-0.84	--	--	15.6	--	--								
93	3	-0.93	--	--	--	15.5	--								
96	3	0.84	--	--	17.6	--	--								
97	3	0.84	--	--	17.6	--	--								
102	1	1.69	--	--	--	18.6	--								
113	1	1.94	--	--	18.9	--	--								
114	0	8.35	--	26.5	--	--	--								
121	1	1.77	--	--	--	--	18.7								
126	2	1.18	--	--	18.0	--	--								
127	3	-0.93	--	--	15.5	--	--								
131	0	-2.19	--	--	--	14.0	--								
134	4	0.25	--	--	--	16.9	--								
138	4	-0.25	--	--	--	--	16.3								
140	4	0.42	--	17.1	--	--	--								
141	3	-0.76	--	--	--	15.7	--								
142	4	-0.17	--	--	--	--	16.4								
144	4	0.34	--	--	17.0	--	--								
145	NR		--	--	--	< 84	--								
146	4	0.34	--	--	--	17.0	--								
147	4	0.08	--	--	--	--	16.7								
151	4	0.00	--	--	--	--	16.6								
154	4	0.17	--	--	16.8	--	--								
180	NR		--	--	--	< 40.2	--								
190	3	-0.59	--	--	15.9	--	--								
191	4	-0.34	--	--	--	--	16.2								
193	3	0.84	--	--	17.6	--	--								
198	4	-0.34	--	--	16.2	--	--								
203	3	-0.67	--	--	15.8	--	--								
204	1	-1.52	--	--	--	--	14.8								
212	3	0.59	--	--	--	17.3	--								
215	3	-0.51	--	--	16.0	--	--								

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Sb (Antimony) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY

N =	16	11	15	1	3. AA: graphite furnace	MPV =	13.9
Minimum =	10.0	10.0	13.0	13.1	4. ICP	F-pseudosigma =	1.1
Maximum =	22.0	64.0	16.6		6. ICP/MS	N =	43
Median =	13.3	13.4	14.0		11na. AA: hydride NaBH4	Uh =	14.4
F-pseudosigma =	1.7	1.6	0.7			Lh =	13.0

Lab	Rating	Z-value	3	4	6	11na
1	4	0.09	14.0	--	--	--
3	4	-0.47	--	13.4	--	--
11	2	1.04	--	15.0	--	--
13	3	-0.76	13.1	--	--	--
23	4	0.28	--	14.2	--	--
32	4	0.00	--	--	13.9	--
39	4	0.09	--	14.0	--	--
42	4	0.28	--	--	14.2	--
45	3	0.66	14.6	--	--	--
46	1	-1.90	11.9	--	--	--
48	3	0.76	--	--	14.7	--
50	2	1.04	--	--	15.0	--
57	4	0.09	--	14.0	--	--
59	1	1.90	--	--	15.9	--
70	4	-0.09	--	--	13.8	--
81	0	-2.75	11.0	--	--	--
86	0	7.69	22.0	--	--	--
89	1	-1.90	11.9	--	--	--
93	0	-2.66	--	11.1	--	--
96	4	0.00	13.9	--	--	--
97	2	1.04	15.0	--	--	--
102	0	-3.70	--	10.0	--	--
113	2	-1.33	--	12.5	--	--
127	4	-0.09	13.8	--	--	--
134	3	-0.85	13.0	--	--	--
138	2	1.42	--	--	15.4	--
141	0	-2.37	--	11.4	--	--
142	0	2.56	--	--	16.6	--
144	2	-1.61	12.2	--	--	--
146	NR		--	< 15.4	--	--
151	4	-0.28	--	--	13.6	--
154	2	-1.42	12.4	--	--	--
180	NR		--	< 40.7	--	--
193	4	-0.47	13.4	--	--	--
198	3	0.95	14.9	--	--	--
204	3	-0.66	--	--	13.2	--
212	4	0.09	--	--	14.0	--
215	0	-3.70	10.0	--	--	--
219	4	0.09	--	--	14.0	--
234	3	-0.57	--	13.3	--	--
236	0	47.58	--	64.0	--	--
241	4	0.00	--	--	13.9	--
247	4	0.32	--	--	14.2	--
265	3	-0.85	--	--	13.0	--
282	3	-0.76	--	--	--	13.1

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Se (Selenium) Concentration Unit : mg/L



SUMMARY

N =	18	10	16	5
Minimum =	4.00	1.07	4.90	2.00
Maximum =	8.00	7.22	7.15	9.00
Median =	5.35	5.19	5.53	
F-pseudosigma =	0.73	1.85	0.59	


3. AA: graphite furnace
 4. ICP
 6. ICP/MS
 11na. AA: hydride NaBH4

MPV =	5.49
F-pseudosigma =	0.83
N =	49
Uh =	6.02
Lh =	4.90


Lab	Rating	Z-value	3	4	6	11na
1	4	-0.10	5.40	--	--	--
3	0	2.09	--	7.22	--	--
10	2	-1.07	--	--	--	4.60
11	3	0.62	--	6.00	--	--
13	0	3.03	8.00	--	--	--
23	3	-0.70	--	4.90	--	--
32	4	-0.10	--	--	5.40	--
39	4	-0.46	--	5.10	--	--
42	3	0.85	--	--	6.19	--
46	4	0.15	5.61	--	--	--
48	3	-0.70	--	--	4.90	--
50	4	-0.10	--	--	5.40	--
57	2	-1.43	4.30	--	--	--
59	4	0.50	--	--	5.90	--
70	4	0.08	--	--	5.55	--
81	1	-1.79	4.00	--	--	--
86	3	-0.98	--	--	--	4.67
87	0	-4.20	--	--	--	< 2
89	0	-4.20	--	--	--	2.00
93	1	-1.67	--	4.10	--	--
96	4	0.02	5.50	--	--	--
102	0	-2.39	--	3.50	--	--
113	4	-0.36	5.19	--	--	--
127	3	-0.60	4.99	--	--	--
131	NR	--	--	< 60	--	--
134	4	0.38	5.80	--	--	--
138	4	-0.02	--	--	5.47	--
141	2	1.34	--	6.60	--	--
142	3	0.67	--	--	6.04	--
144	4	0.14	5.60	--	--	--
146	4	-0.25	--	5.28	--	--
151	4	0.02	--	--	5.50	--
154	2	1.22	6.50	--	--	--
180	NR	--	--	< 67.9	--	--
190	3	-0.81	4.81	--	--	--
191	1	1.58	--	--	6.80	--
193	2	-1.10	4.57	--	--	--
198	NR	--	<10	--	--	--
203	4	-0.22	5.30	--	--	--
204	4	-0.22	--	--	5.30	--
212	4	-0.33	--	--	5.21	--
215	3	0.62	6.00	--	--	--
219	4	0.02	--	--	5.50	--
234	2	1.34	--	6.60	--	--
236	NR	--	--	<10	--	--
241	2	1.28	--	--	6.55	--
247	3	0.85	--	--	6.19	--
255	1	2.01	--	--	7.15	--
265	0	4.23	--	--	--	9.00
277	0	-5.32	--	1.07	--	--

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : SiO₂ (Silica) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	1	25	1	6	9	0. Other	MPV =	11.5	
Minimum =	12.8	5.8	11.2	5.5	10.6	4. ICP	F-pseudostigma =	0.7	
Maximum =		12.5		12.1	15.1	5. DCP	N =	42	
Median =		11.4			12.1	6. ICP/MS	Uh =	12.0	
F-pseudostigma =		0.7			0.7	22mb. Color: molybdate blue	Lh =	11.0	

Lab	Rating	Z-value	0	4	5	6	22mb
1	4	0.47	--	11.8	--	--	--
3	3	-0.74	--	10.9	--	--	--
11	2	1.42	--	12.5	--	--	--
13	4	-0.07	--	11.4	--	--	--
24	3	0.88	--	12.1	--	--	--
26	3	0.74	--	12.0	--	--	--
32	4	0.47	--	--	--	11.8	--
33	4	-0.34	--	--	11.2	--	--
42	2	-1.01	--	10.7	--	--	--
57	4	-0.34	--	11.2	--	--	--
64	4	-0.07	--	11.4	--	--	--
70	3	0.88	--	--	--	--	12.1
76	3	0.84	--	--	--	12.1	--
89	3	-0.61	--	--	--	--	11.0
97	4	-0.07	--	--	--	--	11.4
121	4	0.47	--	11.8	--	--	--
127	4	-0.07	--	11.4	--	--	--
131	2	-1.15	--	10.6	--	--	--
134	4	-0.15	--	11.3	--	--	--
140	2	1.24	--	--	--	--	12.4
142	4	-0.07	--	11.4	--	--	--
145	4	0.42	--	11.8	--	--	--
154	3	-0.74	--	10.9	--	--	--
190	2	1.15	--	--	--	--	12.3
191	4	0.34	--	--	--	11.7	--
203	4	0.31	--	--	--	--	11.7
212	2	1.01	--	12.2	--	--	--
219	3	-0.61	--	--	--	11.0	--
227	0	-7.62	--	5.8	--	--	--
234	4	0.20	--	11.6	--	--	--
236	0	-7.42	--	6.0	--	--	--
241	4	-0.47	--	--	--	11.1	--
247	0	2.09	--	--	--	--	13.0
254	3	0.54	--	11.9	--	--	--
259	4	0.28	--	11.7	--	--	--
265	0	-7.62	--	5.8	--	--	--
273	1	1.75	12.8	--	--	--	--
274	2	-1.17	--	--	--	--	10.6
282	0	-7.99	--	--	--	5.5	--
304	4	0.07	--	11.5	--	--	--
309	0	-7.49	--	5.9	--	--	--
312	0	4.87	--	--	--	--	15.1

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Sr (Strontium) Concentration Unit : mg/L



SUMMARY

N =	27	1	7
Minimum =	163	206	186
Maximum =	211		195
Median =	190		189
F-pseudostigma =	10		3


4. ICP
5. DCP
6. ICP/MS

MPV =	190
F-pseudostigma =	7
Rating Criterion =	10
N =	35
Uh =	196
Lh =	187

Lab	Rating	Z-value	4	5	6
1	4	-0.42	--	--	186
3	2	-1.16	179	--	--
11	2	1.05	200	--	--
24	3	0.63	196	--	--
32	4	-0.11	--	--	189
33	1	1.68	--	206	--
42	4	-0.32	187	--	--
50	3	0.53	--	--	195
57	4	0.00	190	--	--
59	4	-0.32	--	--	187
70	2	1.16	201	--	--
81	3	-0.95	181	--	--
86	4	0.21	192	--	--
102	3	0.69	197	--	--
113	4	-0.11	189	--	--
121	4	0.00	190	--	--
127	3	-0.95	181	--	--
131	0	-2.84	163	--	--
134	3	-0.74	183	--	--
138	3	-0.74	183	--	--
141	2	1.16	201	--	--
142	3	0.84	198	--	--
145	4	0.21	192	--	--
151	4	-0.21	--	--	188
154	1	-1.58	175	--	--
191	4	0.11	--	--	191
212	4	-0.11	189	--	--
219	4	0.32	--	--	193
234	3	-0.84	182	--	--
236	4	0.21	192	--	--
247	4	0.21	192	--	--
254	3	0.89	199	--	--
259	4	0.00	190	--	--
265	4	0.00	190	--	--
273	0	2.21	211	--	--

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Tl (Thallium) Concentration Unit : mg/L





SUMMARY

N =	1	12	9	18	0. Other	MPV =	13.8
Minimum =	34.3	8.5	9.7	12.1	3. AA: graphite furnace	F-pseudosigma =	1.6
Maximum =		25.2	18.2	15.7	4. ICP	N =	40
Median =		13.9	12.7	13.8	6. ICP/MS	Uh =	14.9
F-pseudosigma =		2.7	1.9	0.5		Lh =	12.8

Lab	Rating	Z-value	0	3	4	6
1	4	-0.29	--	--	--	13.3
3	3	-0.67	--	--	12.7	--
11	2	-1.10	--	--	12.0	--
13	4	0.09	--	13.9	--	--
23	0	-2.54	--	--	9.7	--
32	4	0.15	--	--	--	14.0
39	0	2.79	--	--	18.2	--
42	3	0.72	--	--	--	14.9
48	4	-0.04	--	--	--	13.7
50	4	0.09	--	--	--	13.9
57	3	0.53	--	--	14.6	--
59	3	0.72	--	--	--	14.9
70	2	1.22	--	--	--	15.7
76	4	-0.14	--	--	--	13.5
81	3	0.78	--	15.0	--	--
86	0	-2.70	--	9.5	--	--
89	0	7.18	--	25.2	--	--
97	2	1.22	--	15.7	--	--
113	4	0.04	--	13.8	--	--
114	0	12.89	34.3	--	--	--
126	0	-2.48	--	9.8	--	--
127	3	-0.92	--	--	12.3	--
134	4	-0.35	--	13.2	--	--
138	4	0.09	--	--	--	13.9
141	4	-0.10	--	--	13.6	--
142	4	-0.10	--	--	--	13.6
144	3	0.78	--	15.0	--	--
146	2	-1.29	--	--	11.7	--
151	4	-0.10	--	--	--	13.6
154	3	-0.54	--	12.9	--	--
180	NR	--	--	--	< 54.9	--
191	3	-0.67	--	--	--	12.7
198	4	0.34	--	14.3	--	--
204	2	-1.04	--	--	--	12.1
212	4	0.09	--	--	--	13.9
219	4	0.15	--	--	--	14.0
234	3	0.84	--	--	15.1	--
241	3	-0.60	--	--	--	12.8
247	2	1.22	--	--	--	15.7
265	4	-0.48	--	--	--	13.0
282	0	-3.30	--	8.5	--	--

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : U (Uranium) Concentration Unit : mg/L




SUMMARY						
N =	1	0	10	1	0. Other	MPV = 5.04
Minimum =	4.16		4.70	6.80	4. ICP	F-pseudostigma = 0.32
Maximum =			5.60		6. ICP/MS	N = 12
Median =			5.04		9. Atomic fluorescence	Uh = 5.24
F-pseudostigma =			0.24			Lh = 4.80

Lab	Rating	Z-value	0	4	6	9
1	3	-0.71	--	--	4.81	--
32	3	0.96	--	--	5.35	--
70	1	1.74	--	--	5.60	--
76	4	0.25	--	--	5.12	--
121	3	-0.74	--	--	4.80	--
127	NR		--	< 200	--	--
134	4	-0.12	--	--	5.00	--
142	4	0.12	--	--	5.08	--
191	4	0.19	--	--	5.10	--
219	3	-0.74	--	--	4.80	--
254	0	-2.73	4.16	--	--	--
265	2	-1.05	--	--	4.70	--
282	0	5.46	--	--	--	6.80

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : V (Vanadium) Concentration Unit : mg/L




SUMMARY						
N =	4	23	13	3. AA: graphite furnace	MPV =	14.4
Minimum =	14.2	10.8	11.7	4. ICP	F-pseudosigma =	1.7
Maximum =	18.0	24.0	17.0	6. ICP/MS	N =	40
Median =		14.1	14.4		Uh =	16.1
F-pseudosigma =		1.7	1.0		Lh =	13.9


Lab	Rating	Z-value	3	4	6
1	1	1.57	--	--	17.0
3	4	-0.15	--	14.1	--
11	4	-0.21	--	14.0	--
13	3	-0.69	--	13.2	--
26	3	0.75	--	15.6	--
32	4	-0.45	--	--	13.6
39	0	-2.13	--	10.8	--
42	4	0.45	--	--	15.1
46	3	0.51	--	15.2	--
48	4	-0.27	--	--	13.9
50	4	0.21	--	--	14.7
57	4	-0.21	--	14.0	--
59	3	0.63	--	--	15.4
70	4	-0.45	--	--	13.6
81	3	-0.81	--	13.0	--
86	1	1.53	--	16.9	--
89	2	1.17	16.3	--	--
93	3	-0.81	--	13.0	--
96	1	1.53	16.9	--	--
97	0	2.19	18.0	--	--
102	2	1.35	--	16.6	--
121	2	-1.29	--	12.2	--
127	4	-0.09	14.2	--	--
134	4	-0.27	--	13.9	--
138	4	-0.21	--	14.0	--
141	4	-0.03	--	14.3	--
142	4	-0.33	--	--	13.8
145	0	3.39	--	20.0	--
146	3	0.51	--	15.2	--
154	3	0.99	--	16.0	--
180	0	2.97	--	19.3	--
212	1	-1.59	--	--	11.7
215	0	5.79	--	24.0	--
219	4	-0.21	--	--	14.0
234	4	0.03	--	14.4	--
236	0	-2.59	--	< 10	--
241	2	1.11	--	--	16.2
247	4	0.39	--	--	15.0
265	4	-0.21	--	14.0	--
282	4	0.03	--	--	14.4
304	2	-1.41	--	12.0	--

Table 11. Statistical summary of reported data for standard reference water sample T-159 (trace constituents)--Continued

Analyte : Zn (Zinc) Concentration Unit : mg/L



ANALYTICAL METHOD CODE



SUMMARY																		
Lab	Rating	Z-value	1	3	4	6	20	51	Lab	Rating	Z-value	1	3	4	6	20	51	
N = 9 3 31 20 1 1 Minimum = 15.4 18.0 9.5 14.8 18.8 20.9 Maximum = 26.5 35.0 38.0 23.3 Median = 20.0 18.8 19.0 F-pseudostigma = 1.1 2.3 1.4									1. AA: direct, air 3. AA: graphite furnace 4. ICP 6. ICP/MS 20. Titrate: colorimetric 51. Turbidimetric								MPV = 19.2 F-pseudostigma = 1.9 N = 65 Uh = 20.5 Lh = 18.0	
1	4	-0.22	--	--	--	18.8	--	--	215	0	10.14	--	--	38.0	--	--	--	
3	3	-0.86	--	--	17.6	--	--	--	219	4	0.43	--	--	--	20.0	--	--	
10	1	1.51	--	22.0	--	--	--	--	227	3	0.70	--	--	20.5	--	--	--	
11	NR	--	--	--	< 25	--	--	--	234	1	1.73	--	--	22.4	--	--	--	
12	0	8.53	--	35.0	--	--	--	--	236	1	-1.73	--	--	16.0	--	--	--	
13	2	-1.03	--	--	17.3	--	--	--	241	1	1.73	--	--	--	22.4	--	--	
23	4	-0.22	--	--	--	--	18.8	--	247	0	2.21	--	--	--	23.3	--	--	
24	4	-0.22	--	--	18.8	--	--	--	254	3	0.54	--	--	20.2	--	--	--	
26	4	-0.32	--	--	18.6	--	--	--	255	3	0.54	--	--	--	20.2	--	--	
32	4	-0.11	--	--	--	19.0	--	--	259	3	-0.54	--	--	18.2	--	--	--	
39	2	1.46	--	--	21.9	--	--	--	265	3	-0.65	--	--	18.0	--	--	--	
42	3	0.54	--	--	--	20.2	--	--	273	4	-0.27	--	--	18.7	--	--	--	
45	3	0.97	21.0	--	--	--	--	--	277	0	-5.23	--	--	9.5	--	--	--	
46	2	-1.03	--	--	17.3	--	--	--	282	4	0.32	--	--	--	19.8	--	--	
48	2	-1.30	--	--	--	16.8	--	--	287	3	-0.65	18.0	--	--	--	--	--	
50	3	-0.65	--	--	--	18.0	--	--	292	4	-0.11	--	--	19.0	--	--	--	
57	2	1.24	--	--	21.5	--	--	--	304	4	0.38	--	--	19.9	--	--	--	
59	3	0.54	--	--	--	20.2	--	--	305	0	-2.05	15.4	--	--	--	--	--	
70	2	-1.03	--	--	--	17.3	--	--	307	0	3.13	25.0	--	--	--	--	--	
81	1	1.51	--	--	22.0	--	--	--	308	0	3.94	26.5	--	--	--	--	--	
86	2	-1.13	--	--	17.1	--	--	--										
87	0	2.43	--	--	23.7	--	--	--										
89	3	-0.65	--	18.0	--	--	--	--										
93	2	-1.08	--	--	17.2	--	--	--										
96	0	-5.21	< 10	--	--	--	--	--										
102	0	-4.42	--	--	11.0	--	--	--										
113	3	0.65	--	--	20.4	--	--	--										
114	4	0.16	19.5	--	--	--	--	--										
121	4	-0.11	--	--	--	19.0	--	--										
127	4	-0.27	--	--	18.7	--	--	--										
131	3	0.70	--	--	20.5	--	--	--										
134	3	0.54	--	--	20.2	--	--	--										
138	4	-0.11	--	--	--	19.0	--	--										
140	4	0.43	20.0	--	--	--	--	--										
141	3	0.92	--	--	--	--	--	20.9										
142	0	-2.37	--	--	--	14.8	--	--										
144	4	0.43	20.0	--	--	--	--	--										
145	0	2.05	--	--	23.0	--	--	--										
146	4	0.43	--	--	20.0	--	--	--										
147	4	-0.38	--	--	--	18.5	--	--										
151	4	0.00	--	--	--	19.2	--	--										
154	2	-1.46	--	--	16.5	--	--	--										
180	NR	--	--	--	< 24.4	--	--	--										
190	4	0.38	19.9	--	--	--	--	--										
191	4	-0.49	--	--	--	18.3	--	--										
193	NR	--	< 25	--	--	--	--	--										
198	NR	--	--	--	< 50	--	--	--										
203	3	-0.65	--	--	18.0	--	--	--										
204	2	-1.46	--	--	--	16.5	--	--										
212	4	0.11	--	--	--	19.4	--	--										


Table 12. *Statistical summary of reported data for standard reference sample M-152 (major constituents)*

Definition of analytical methods, abbreviations, and symbols		
<u>Analytical methods</u>		
0. Other/Not reported		
1. AA: direct, air	=	atomic absorption: direct, air
2. AA: direct, N ₂ O	=	atomic absorption: direct, nitrous oxide
3. AA: graphite furnace	=	atomic absorption: graphite furnace
4. ICP	=	inductively coupled plasma
5. DCP	=	direct current plasma
6. ICP/MS	=	inductively coupled plasma / mass spectrometry
7. IC	=	ion chromatography
12. Flame emission	=	flame emission
20. Titrate: color	=	titration: colorimetric (color reagent specified)
21. Titrate: electro	=	titration: electrometric
22. Color:	=	colorimetric (color reagent specified)
40. Ion electrode	=	ion selective electrode
41. Electro	=	electrometric: (type meter specified)
50. Gravimetric	=	gravimetric: (precipitate specified)
51. Turbidimetric	=	turbidimetric: (precipitate specified)
<hr/>		
<u>Abbreviations and figure symbols</u>		
N	=	number of analyses--(excluding less than values)
MPV	=	most probable value ---
F-pseudostigma	=	nonparametric statistic deviation
Uh	=	upper hinge value
Lh	=	lower hinge value
Uwl	=	upper warning limit ----
Lwl	=	lower warning limit ----
Ucl	=	upper warning limit
Lcl	=	lower warning limit ----
µg/L	=	micrograms per liter
mg/L	=	milligrams per liter
µS/cm	=	microsiemens per centimeter at 25 ⁰ C
Lab	=	laboratory code number
NR	=	not rated, less than value reported or insufficient data
<	=	less than
--	=	not reported
<hr/>		
<u>Constituent</u>		<u>page</u>
Alk Alkalinity as CaCO ₃		65
B Boron		66
Ca Calcium		67
Cl Chloride		68
DSRD Dissolved solids		69
F Fluoride		70
K Potassium		71
Mg Magnesium		72
Na Sodium		73
total P Phosphorus		74
pH		75
SiO ₂ Silica		76
SO ₄ Sulfate		77
Sp Con Specific Conductance		78
Sr Strontium		79
V Vanadium		80


Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte : Alkalinity (as CaCO₃)

Concentration Unit : mg/L



ANALYTICAL METHOD CODE



SUMMARY									
N =	1	8	53	2	4	0. Other	MPV =	6.00	
Minimum =	11.50	3.60	3.65	4.80	0.22	20. Titrate: colorimetric	F-pseudostigma =	1.49	
Maximum =		11.60	24.40	7.80	6.90	21. Titrate: electrometric	N =	68	
Median =		6.05	6.00			22. Colorimetric	Uh =	7.00	
F-pseudostigma =		2.91	1.30			41. Direct reading	Lh =	5.00	


Lab	Rating	Z-value	0	20	21	22	41	Lab	Rating	Z-value	0	20	21	22	41
1	NR	-0.09	--	--	5.86	--	--	215	NR	0.00	--	--	6.00	--	--
3	NR	0.00	--	--	6.00	--	--	227	NR	0.00	--	--	6.00	--	--
10	NR	-1.28	--	--	4.10	--	--	234	NR	1.21	--	--	--	7.80	--
11	NR	-0.67	--	--	5.00	--	--	236	NR	0.27	--	--	6.40	--	--
13	NR	1.04	--	--	7.55	--	--	241	NR	-0.67	--	--	5.00	--	--
23	NR	0.07	--	6.10	--	--	--	244	NR	4.78	--	--	13.10	--	--
24	NR	0.50	--	--	6.74	--	--	247	NR	-0.01	--	--	5.98	--	--
25	NR	3.36	--	--	11.00	--	--	253	NR	-1.61	--	3.60	--	--	--
26	NR	1.61	--	--	8.40	--	--	259	NR	-1.35	--	--	4.00	--	--
33	NR	-1.52	--	--	3.74	--	--	268	NR	12.38	--	--	24.40	--	--
38	NR	0.23	--	--	6.34	--	--	273	NR	2.09	--	--	9.10	--	--
39	NR	0.34	--	--	6.50	--	--	274	NR	2.60	--	9.86	--	--	--
42	NR	1.05	--	--	7.56	--	--	275	NR	2.69	--	10.00	--	--	--
43	NR	--	--	--	< 20	--	--	277	NR	-0.44	--	--	5.35	--	--
45	NR	--	--	--	< 20	--	--	282	NR	-0.62	--	--	5.08	--	--
46	NR	-3.89	--	--	--	--	0.22	287	NR	0.67	--	--	7.00	--	--
48	NR	--	< 5	--	--	--	--	292	NR	0.34	--	--	6.50	--	--
50	NR	-0.68	--	--	4.99	--	--	302	NR	0.00	--	6.00	--	--	--
51	NR	-1.14	--	--	4.30	--	--	307	NR	0.00	--	6.00	--	--	--
57	NR	0.00	--	6.00	--	--	--	309	NR	3.77	--	11.60	--	--	--
59	NR	-0.94	--	--	4.60	--	--	312	NR	3.70	11.50	--	--	--	--
70	NR	0.67	--	--	7.00	--	--								
81	NR	-0.57	--	--	5.15	--	--								
85	NR	-0.84	--	--	4.75	--	--								
89	NR	0.07	--	--	6.10	--	--								
93	NR	-0.80	--	--	4.81	--	--								
96	NR	-1.01	--	--	4.50	--	--								
97	NR	0.20	--	--	6.30	--	--								
107	NR	3.23	--	--	10.80	--	--								
109	NR	1.29	--	--	7.91	--	--								
113	NR	0.00	--	--	6.00	--	--								
114	NR	-0.67	--	--	5.00	--	--								
127	NR	0.30	--	--	6.45	--	--								
134	NR	0.56	--	--	--	--	6.83								
138	NR	0.34	--	--	6.50	--	--								
141	NR	-0.03	--	--	5.95	--	--								
142	NR	0.67	--	--	7.00	--	--								
143	NR	-1.35	--	--	4.00	--	--								
145	NR	-0.81	--	--	--	4.80	--								
146	NR	-0.90	--	--	4.66	--	--								
151	NR	0.00	--	--	6.00	--	--								
154	NR	2.69	--	--	10.00	--	--								
180	NR	0.61	--	--	--	--	6.90								
185	NR	-0.64	--	--	5.05	--	--								
190	NR	-0.27	--	--	--	--	5.60								
193	NR	-1.35	--	--	4.00	--	--								
203	NR	-1.48	--	--	3.80	--	--								
204	NR	-1.58	--	--	3.65	--	--								
212	NR	0.01	--	--	6.02	--	--								
213	NR	-0.67	--	--	5.00	--	--								

Laboratories were not rated for alkalinity because the long term stability of alkalinity needs to be verified. The MPV for alkalinity is provided as an estimated value.


Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte : B (Boron)

Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY					
N =	1	28	3	0. Other	MPV = 92.2
Minimum =	74.2	12.6	93.0	4. ICP	F-pseudosigma = 7.8
Maximum =		143.0	94.0	6. ICP/MS	N = 32
Median =		92.0			Uh = 95.4
F-pseudosigma =		8.9			Lh = 84.8


Lab	Rating	Z-value	0	4	6
1	4	0.10	--	--	93.0
3	4	-0.15	--	--	91.0
11	4	0.11	--	--	93.0
24	3	0.62	--	--	97.0
26	3	-0.84	--	--	85.6
32	4	-0.09	--	--	91.5
39	0	-2.30	74.2	--	--
42	4	-0.10	--	--	91.4
46	4	0.00	--	--	92.2
48	4	0.23	--	--	94.0
50	3	0.58	--	--	96.7
57	3	0.61	--	--	96.9
70	0	4.71	--	--	129.0
86	2	-1.40	--	--	81.2
127	4	0.18	--	--	93.6
131	4	-0.28	--	--	90.0
134	4	0.00	--	--	92.1
138	4	0.21	--	--	93.8
141	2	-1.36	--	--	81.5
142	3	0.73	--	--	97.9
145	1	1.51	--	--	104.0
180	0	-10.17	--	--	< 12.6
212	4	-0.03	--	--	91.9
215	1	-1.56	--	--	80.0
234	3	-0.97	--	--	84.6
236	2	-1.17	--	--	83.0
247	0	6.50	--	--	143.0
255	4	0.21	--	--	93.8
259	3	-0.92	--	--	85.0
265	4	0.23	--	--	94.0
273	0	-3.93	--	--	61.4
282	3	0.68	--	--	97.5

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte : Ca (Calcium) Concentration Unit : mg/L




ANALYTICAL METHOD CODE



SUMMARY																	
N =	6	14	2	45	4	3	0. Other	MPV =	3.63								
Minimum =	2.37	2.58	2.40	2.98	3.63	2.60	1. AA: direct, air	F-pseudostigma =	0.18								
Maximum =	3.70	5.40	3.58	4.34	3.98	5.24	2. AA: direct, nitrous oxide	N =	74								
Median =		3.61		3.65			4. ICP	Uh =	3.79								
F-pseudostigma =		0.29		0.13			6. ICP/MS	Lh =	3.55								
							7. Ion chromatography										
Lab	Rating	Z-value	0	1	2	4	6	7	Lab	Rating	Z-value	0	1	2	4	6	7
1	4	0.34	--	--	--	3.69	--	--	215	4	0.39	--	--	--	3.70	--	--
3	4	0.34	--	--	--	3.69	--	--	219	4	0.00	--	--	--	--	3.63	--
10	4	-0.17	--	3.60	--	--	--	--	227	3	-0.84	--	--	--	3.48	--	--
11	3	0.79	--	--	--	3.77	--	--	234	4	0.11	--	--	--	3.65	--	--
13	4	-0.22	--	--	--	3.59	--	--	236	3	-0.90	--	--	--	3.47	--	--
24	2	1.41	--	--	--	3.88	--	--	241	3	0.96	--	3.80	--	--	--	--
26	0	-5.79	--	--	--	--	2.60	--	247	4	-0.28	--	--	--	3.58	--	--
32	4	0.39	--	--	--	--	3.70	--	254	3	0.84	--	--	--	3.78	--	--
33	4	0.39	3.70	--	--	--	--	--	255	4	0.06	--	--	--	3.64	--	--
38	4	-0.28	--	--	3.58	--	--	--	259	4	0.39	--	--	--	3.70	--	--
39	1	-1.80	3.31	--	--	--	--	--	265	4	0.11	--	--	--	3.65	--	--
42	4	-0.45	--	--	--	3.55	--	--	268	0	-3.82	--	2.95	--	--	--	--
45	2	-1.24	--	3.41	--	--	--	--	273	0	3.99	--	--	--	4.34	--	--
46	4	-0.28	--	--	--	3.58	--	--	274	0	-7.08	2.37	--	--	--	--	--
48	1	1.97	--	--	--	--	3.98	--	275	0	-3.54	3.00	--	--	--	--	--
50	3	0.90	--	--	--	3.79	--	--	277	3	0.56	--	--	--	3.73	--	--
51	1	-1.80	--	3.31	--	--	--	--	279	0	-3.54	3.00	--	--	--	--	--
57	4	0.00	--	--	--	3.63	--	--	282	4	-0.28	--	--	--	3.58	--	--
64	4	0.22	--	--	--	3.67	--	--	287	0	-5.90	--	2.58	--	--	--	--
70	1	1.97	--	--	--	3.98	--	--	292	0	9.95	--	5.40	--	--	--	--
76	4	-0.31	--	3.58	--	--	--	--	302	1	1.60	--	--	--	--	--	3.91
81	1	-1.52	--	--	--	3.36	--	--	304	4	0.39	--	--	--	3.70	--	--
85	0	3.15	--	4.19	--	--	--	--	307	0	7.42	--	4.95	--	--	--	--
86	4	-0.17	--	--	--	3.60	--	--	309	1	1.57	--	--	--	3.91	--	--
87	0	-6.91	--	--	2.40	--	--	--									
89	0	-6.97	2.39	--	--	--	--	--									
93	4	-0.28	--	--	--	3.58	--	--									
102	1	1.91	--	--	--	3.97	--	--									
109	3	0.96	--	3.80	--	--	--	--									
113	3	0.67	--	--	--	3.75	--	--									
121	4	-0.17	--	--	--	3.60	--	--									
127	4	-0.11	--	--	--	3.61	--	--									
131	4	-0.17	--	--	--	3.60	--	--									
134	3	0.51	--	--	--	3.72	--	--									
138	4	0.22	--	--	--	3.67	--	--									
140	4	-0.17	--	3.60	--	--	--	--									
141	2	1.01	--	--	--	3.81	--	--									
142	0	-3.65	--	--	--	2.98	--	--									
145	2	1.24	--	--	--	3.85	--	--									
146	1	-1.57	--	--	--	3.35	--	--									
151	2	1.41	--	3.88	--	--	--	--									
154	4	-0.17	--	--	--	3.60	--	--									
180	3	0.96	--	--	--	3.80	--	--									
185	4	-0.11	--	3.61	--	--	--	--									
190	0	9.05	--	--	--	--	--	5.24									
191	3	0.90	--	--	--	--	3.79	--									
193	1	-2.02	--	--	--	3.27	--	--									
203	4	-0.17	--	--	--	3.60	--	--									
209	0	-3.19	--	--	--	3.06	--	--									
212	4	-0.45	--	--	--	3.55	--	--									

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued


Analyte : Cl (Chloride) Concentration Unit : mg/L




SUMMARY																	
N =	1	49	10	5	13	2	0. Other		MPV =	17.3							
Minimum =	18.0	14.9	16.3	17.0	14.0	16.6	7. Ion chromatography		F-pseudostigma =	0.9							
Maximum =		19.1	34.4	18.0	21.2	17.0	20. Titrate: colorimetric		N =	80							
Median =		17.4	18.9		16.8		21. Titrate: electrometric		Uh =	18.0							
F-pseudostigma =		0.7	2.2		0.9		22. Colorimetric		Lh =	16.8							
							40. Ion selective electrode										
Lab	Rating	Z-value	0	7	20	21	22	40	Lab	Rating	Z-value	0	7	20	21	22	40
1	4	-0.48	--	16.9	--	--	--	--	190	2	-1.26	--	16.2	--	--	--	--
3	2	1.21	--	--	--	--	18.4	--	191	4	0.03	--	17.4	--	--	--	--
10	4	0.20	--	--	--	--	17.5	--	196	3	-0.70	--	16.7	--	--	--	--
11	3	0.76	--	--	18.0	--	--	--	203	4	0.31	--	--	--	17.6	--	--
12	0	-3.74	--	--	--	--	14.0	--	204	4	-0.37	--	--	--	17.0	--	--
13	3	-0.70	--	16.7	--	--	--	--	208	4	-0.03	--	17.3	--	--	--	--
23	4	0.20	--	17.5	--	--	--	--	209	1	1.59	--	18.7	--	--	--	--
24	4	-0.37	--	--	--	--	17.0	--	212	3	-0.70	--	16.7	--	--	--	--
25	3	0.53	--	17.8	--	--	--	--	213	2	-1.15	--	--	16.3	--	--	--
26	1	2.00	--	19.1	--	--	--	--	215	1	1.55	--	--	18.7	--	--	--
33	3	-0.59	--	16.8	--	--	--	--	219	2	-1.49	--	16.0	--	--	--	--
39	0	3.01	--	--	20.0	--	--	--	227	3	-0.88	--	16.5	--	--	--	--
42	4	0.20	--	17.5	--	--	--	--	234	4	0.31	--	17.6	--	--	--	--
43	4	-0.37	--	--	--	--	--	17.0	241	4	0.42	--	17.7	--	--	--	--
45	2	1.32	--	18.5	--	--	--	--	247	4	0.08	--	17.4	--	--	--	--
48	3	0.76	18.0	--	--	--	--	--	253	4	-0.40	--	--	17.0	--	--	--
50	3	-0.82	--	--	--	--	16.6	--	254	4	0.08	--	17.4	--	--	--	--
51	3	0.60	--	17.9	--	--	--	--	259	4	0.31	--	17.6	--	--	--	--
57	0	5.26	--	--	22.0	--	--	--	265	3	0.53	--	17.8	--	--	--	--
64	4	0.20	--	17.5	--	--	--	--	268	3	-0.82	--	16.6	--	--	--	--
70	4	-0.03	--	17.3	--	--	--	--	273	1	-1.86	--	15.7	--	--	--	--
76	4	0.06	--	17.4	--	--	--	--	274	0	19.23	--	--	34.4	--	--	--
81	4	-0.14	--	--	--	17.2	--	--	275	0	3.01	--	--	20.0	--	--	--
85	2	1.21	--	18.4	--	--	--	--	277	0	-2.73	--	14.9	--	--	--	--
86	4	-0.03	--	17.3	--	--	--	--	279	4	-0.42	--	--	17.0	--	--	--
87	2	-1.49	--	--	--	--	16.0	--	282	1	-1.71	--	15.8	--	--	--	--
89	3	-0.59	--	16.8	--	--	--	--	287	2	-1.23	--	16.2	--	--	--	--
93	1	1.88	--	19.0	--	--	--	--	292	4	-0.03	--	17.3	--	--	--	--
96	2	-1.49	--	--	--	--	16.0	--	302	0	-2.31	--	15.3	--	--	--	--
97	2	-1.04	--	--	--	--	16.4	--	307	1	1.88	--	--	19.0	--	--	--
102	4	0.46	--	17.7	--	--	--	--									
107	4	-0.03	--	--	--	17.3	--	--									
109	3	0.76	--	--	--	18.0	--	--									
113	3	0.65	--	17.9	--	--	--	--									
114	3	-0.82	--	--	--	--	--	16.6									
127	4	0.08	--	17.4	--	--	--	--									
131	2	1.10	--	18.3	--	--	--	--									
134	4	-0.24	--	17.1	--	--	--	--									
138	4	-0.03	--	17.3	--	--	--	--									
140	0	-2.37	--	--	--	--	15.2	--									
141	3	0.98	--	18.2	--	--	--	--									
142	3	0.98	--	18.2	--	--	--	--									
143	3	-0.59	--	--	--	--	16.8	--									
145	4	0.40	--	17.7	--	--	--	--									
146	4	-0.14	--	--	--	--	17.2	--									
151	4	-0.25	--	17.1	--	--	--	--									
154	4	-0.14	--	--	--	--	17.2	--									
180	2	1.32	--	18.5	--	--	--	--									
183	0	4.31	--	--	--	--	21.2	--									
185	4	0.17	--	17.5	--	--	--	--									

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte : DSRD (Dissolved solids) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY

N =	1	45	4. ICP	MPV =	66.5
Minimum =	68.0	31.0	50. Gravimetric	F-pseudosigma =	8.2
Maximum =		106.0		N =	46
Median =		66.0		Uh =	72.0
F-pseudosigma =		8.2		Lh =	61.0


Lab	Rating	Z-value	4	50
1	4	-0.06	--	66.0
3	3	0.55	--	71.0
10	4	-0.06	--	66.0
11	4	0.18	68.0	--
12	3	0.67	--	72.0
13	3	-0.64	--	61.3
25	0	-3.74	--	36.0
26	4	-0.06	--	66.0
38	4	-0.43	--	63.0
43	3	0.92	--	74.0
45	4	0.43	--	70.0
46	0	4.84	--	106.0
48	3	0.55	--	71.0
50	4	-0.43	--	63.0
57	3	0.92	--	74.0
70	4	0.43	--	70.0
76	3	0.67	--	72.0
81	0	-3.25	--	40.0
85	2	-1.41	--	55.0
87	0	2.64	--	88.0
89	3	0.67	--	72.0
96	3	0.55	--	71.0
97	3	0.92	--	74.0
109	2	-1.41	--	55.0
113	2	-1.41	--	55.0
114	0	-2.51	--	46.0
127	4	-0.25	--	64.5
134	4	0.23	--	68.4
138	4	0.43	--	70.0
140	0	4.11	--	100.0
141	0	-2.70	--	44.5
142	4	-0.31	--	64.0
143	3	0.67	--	72.0
146	0	-2.76	--	44.0
151	3	-0.67	--	61.0
190	3	0.86	--	73.5
212	2	1.41	--	78.0
215	4	-0.06	--	66.0
227	3	-0.67	--	61.0
234	4	0.06	--	67.0
236	3	0.67	--	72.0
247	4	-0.43	--	63.0
273	4	-0.06	--	66.0
275	3	-0.80	--	60.0
282	0	4.35	--	102.0
305	0	-4.35	--	31.0

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte : F (Fluoride) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY					
N =	28	1	34	7. Ion chromatography	MPV = 1.98
Minimum =	0.92	1.79	1.60	20. Titrate: colorimetric	F-pseudosigma = 0.12
Maximum =	3.17	--	2.09	40. Ion selective electrode	N = 63
Median =	1.99	--	1.99		Uh = 2.04
F-pseudosigma =	0.21	--	0.07		Lh = 1.88


Lab	Rating	Z-value	7	20	40	Lab	Rating	Z-value	7	20	40
1	4	0.00	--	--	1.98	247	1	-2.04	1.73	--	--
3	4	0.49	--	--	2.04	255	4	0.49	--	--	2.04
10	3	0.57	--	--	2.05	259	4	0.16	--	--	2.00
11	0	-3.10	--	--	1.60	265	3	0.98	2.10	--	--
13	2	-1.22	1.83	--	--	273	0	-4.08	1.48	--	--
23	3	0.82	2.08	--	--	274	1	-1.55	--	1.79	--
25	0	-2.77	1.64	--	--	275	4	0.16	--	--	2.00
33	3	-0.82	1.88	--	--	277	0	-7.99	1.00	--	--
39	4	0.00	--	--	1.98	282	2	1.30	2.14	--	--
42	4	-0.16	1.96	--	--	287	1	-1.55	1.79	--	--
45	4	0.16	2.00	--	--	292	4	-0.49	1.92	--	--
46	4	0.41	--	--	2.03	302	4	-0.36	1.94	--	--
50	4	-0.41	--	--	1.93	305	0	-2.12	--	--	1.72
57	0	-2.28	--	--	1.70						
59	4	0.16	--	--	2.00						
70	4	-0.16	--	--	1.96						
76	4	0.50	2.04	--	--						
81	4	0.00	1.98	--	--						
85	4	0.08	--	--	1.99						
86	3	0.65	2.06	--	--						
89	2	-1.06	--	--	1.85						
93	0	-8.64	0.92	--	--						
96	3	0.73	--	--	2.07						
97	3	0.90	--	--	2.09						
102	0	9.70	3.17	--	--						
107	4	0.16	--	--	2.00						
109	4	-0.24	--	--	1.95						
113	4	0.16	--	--	2.00						
114	4	-0.41	--	--	1.93						
127	3	0.57	2.05	--	--						
131	1	1.87	2.21	--	--						
134	4	-0.16	--	--	1.96						
138	4	-0.24	--	--	1.95						
140	4	0.41	--	--	2.03						
141	3	0.98	2.10	--	--						
142	3	0.90	--	--	2.09						
145	4	0.16	2.00	--	--						
146	3	0.65	--	--	2.06						
151	4	0.49	--	--	2.04						
154	3	-0.82	--	--	1.88						
180	3	0.82	2.08	--	--						
183	4	0.16	--	--	2.00						
190	4	-0.16	--	--	1.96						
191	4	0.08	1.99	--	--						
196	2	-1.47	1.80	--	--						
212	0	-2.20	--	--	1.71						
215	3	-0.90	--	--	1.87						
219	0	-2.85	1.63	--	--						
234	1	1.63	2.18	--	--						
241	3	-0.65	--	--	1.90						

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte : K (Potassium) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
		N =	2	21	32	4	3	5	0. Other	MPV =	0.92						
		Minimum =	0.88	0.70	0.52	0.83	0.57	0.88	1. AA: direct, air	F-pseudosigma =	0.09						
		Maximum =	1.00	4.00	23.00	0.96	1.05	1.19	4. ICP	N =	67						
		Median =	0.92		0.94				6. ICP/MS	Uh =	1.00						
		F-pseudosigma =	0.05		0.13				7. Ion chromatography	Lh =	0.88						
									12. Flame emission								
Lab	Rating	Z-value	0	1	4	6	7	12	Lab	Rating	Z-value	0	1	4	6	7	12
1	4	-0.33	--	0.89	--	--	--	--	219	4	-0.11	--	--	--	0.91	--	--
3	0	5.13	--	--	1.39	--	--	--	234	2	-1.16	--	--	0.81	--	--	--
10	4	0.00	--	0.92	--	--	--	--	236	3	-0.87	--	--	0.84	--	--	--
11	0	-3.06	--	--	0.64	--	--	--	241	0	-2.40	--	0.70	--	--	--	--
13	0	-2.51	--	--	0.69	--	--	--	247	4	0.17	--	--	0.94	--	--	--
24	4	0.44	--	--	0.96	--	--	--	254	4	0.33	--	0.95	--	--	--	--
26	0	-3.82	--	--	--	--	0.57	--	259	0	-2.84	--	--	0.66	--	--	--
32	4	0.44	--	--	--	0.96	--	--	265	4	-0.44	--	--	0.88	--	--	--
33	4	-0.44	0.88	--	--	--	--	--	268	1	1.97	--	1.10	--	--	--	--
38	4	0.00	--	0.92	--	--	--	--	273	2	1.20	--	--	1.03	--	--	--
39	4	-0.33	--	--	0.89	--	--	--	274	3	0.66	--	--	--	--	--	0.98
42	4	-0.34	--	--	0.89	--	--	--	275	3	0.87	1.00	--	--	--	--	--
45	2	1.09	--	1.02	--	--	--	--	277	3	0.66	--	--	0.98	--	--	--
46	0	241.18	--	--	23.00	--	--	--	279	4	0.22	--	--	--	--	--	0.94
48	4	0.31	--	--	--	0.95	--	--	282	0	3.60	--	--	1.25	--	--	--
50	2	1.31	--	--	1.04	--	--	--	287	2	-1.13	--	0.82	--	--	--	--
51	0	2.95	--	--	--	--	--	1.19	292	3	0.87	--	1.00	--	--	--	--
57	0	2.51	--	--	1.15	--	--	--	302	2	1.46	--	--	--	--	1.05	--
64	4	-0.33	--	0.89	--	--	--	--	304	4	-0.22	--	--	0.90	--	--	--
70	2	1.42	--	--	1.05	--	--	--	309	0	5.68	--	--	1.44	--	--	--
81	2	-1.35	--	--	0.80	--	--	--									
85	3	0.60	--	0.98	--	--	--	--									
86	4	-0.23	--	--	0.90	--	--	--									
87	3	0.66	--	0.98	--	--	--	--									
89	4	-0.44	--	--	--	--	--	0.88									
93	2	-1.42	--	--	0.79	--	--	--									
97	4	-0.44	--	0.88	--	--	--	--									
102	0	-4.37	--	--	0.52	--	--	--									
109	1	-1.64	--	0.77	--	--	--	--									
113	2	1.05	--	--	1.02	--	--	--									
121	4	0.22	--	0.94	--	--	--	--									
127	3	-0.92	--	0.84	--	--	--	--									
131	3	0.87	--	--	1.00	--	--	--									
134	4	-0.27	--	0.90	--	--	--	--									
138	3	-0.51	--	--	0.87	--	--	--									
140	2	-1.04	--	0.83	--	--	--	--									
141	2	1.09	--	--	1.02	--	--	--									
142	NR	--	--	--	< 1	--	--	--									
145	3	0.98	--	--	1.01	--	--	--									
146	NR	--	--	--	< 0.86	--	--	--									
151	4	0.00	--	0.92	--	--	--	--									
154	3	0.87	--	--	1.00	--	--	--									
180	NR	--	--	--	< 1.63	--	--	--									
185	4	-0.22	--	0.90	--	--	--	--									
190	4	-0.11	--	--	--	--	0.91	--									
191	3	-0.98	--	--	--	0.83	--	--									
193	4	0.33	--	0.95	--	--	--	--									
203	4	0.00	--	--	--	--	--	0.92									
209	4	0.36	--	--	0.95	--	--	--									
212	3	-0.73	--	--	0.85	--	--	--									

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte : Mg (Magnesium) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
N =	3	16	46	3	3	2	0. Other	MPV = 0.722									
Minimum =	0.613	0.600	0.520	0.730	0.450	0.600	1. AA: direct, air	F-pseudosigma = 0.047									
Maximum =	0.800	4.000	1.100	0.770	0.910	0.950	4. ICP	N = 73									
Median =		0.725	0.720				6. ICP/MS	Uh = 0.746									
F-pseudosigma =		0.073	0.037				7. Ion chromatography	Lh = 0.683									
							20. Titrate: colorimetric										
Lab	Rating	Z-value	0	1	4	6	7	20	Lab	Rating	Z-value	0	1	4	6	7	20
1	4	0.17	--	--	0.730	--	--	--	215	2	-1.33	--	--	0.660	--	--	--
3	4	-0.06	--	--	0.719	--	--	--	219	4	-0.47	--	--	0.700	--	--	--
10	3	0.60	--	0.750	--	--	--	--	227	3	-0.51	--	--	0.698	--	--	--
11	4	-0.47	--	--	0.700	--	--	--	234	4	0.00	--	--	0.722	--	--	--
13	4	0.39	--	--	0.740	--	--	--	236	2	-1.11	--	--	0.670	--	--	--
24	4	0.39	--	--	0.740	--	--	--	241	4	-0.04	--	0.720	--	--	--	--
26	0	-5.82	--	--	--	--	0.450	--	247	4	0.32	--	--	0.737	--	--	--
32	4	0.17	--	--	--	0.730	--	--	254	4	0.39	--	--	0.740	--	--	--
33	4	-0.26	0.710	--	--	--	--	--	255	4	-0.13	--	--	0.716	--	--	--
38	3	0.51	--	0.746	--	--	--	--	259	4	-0.47	--	--	0.700	--	--	--
39	0	-2.81	--	--	0.591	--	--	--	265	4	0.39	--	--	0.740	--	--	--
42	0	-2.27	--	--	0.616	--	--	--	268	4	0.39	--	0.740	--	--	--	--
45	1	-1.97	--	0.630	--	--	--	--	273	1	1.88	--	--	0.810	--	--	--
46	3	-0.64	--	--	0.692	--	--	--	274	0	4.88	--	--	--	--	--	0.950
48	2	1.03	--	--	--	0.770	--	--	275	1	1.67	0.800	--	--	--	--	--
50	4	0.39	--	--	0.740	--	--	--	277	0	8.09	--	--	1.100	--	--	--
51	4	0.17	--	0.730	--	--	--	--	279	0	-2.61	--	--	--	--	--	0.600
57	4	0.30	--	--	0.736	--	--	--	282	3	0.54	--	--	0.747	--	--	--
64	4	-0.47	--	--	0.700	--	--	--	287	1	-1.73	--	0.641	--	--	--	--
70	3	0.51	--	--	0.746	--	--	--	292	0	-2.61	--	0.600	--	--	--	--
76	4	0.45	--	0.743	--	--	--	--	302	0	3.79	--	--	--	--	0.899	--
81	0	-2.12	--	--	0.623	--	--	--	304	3	-0.69	--	--	0.690	--	--	--
85	4	0.26	--	0.734	--	--	--	--	307	3	0.60	--	0.750	--	--	--	--
86	3	0.79	--	--	0.759	--	--	--	309	0	-4.33	--	--	0.520	--	--	--
87	0	-2.61	--	0.600	--	--	--	--									
89	0	-2.33	0.613	--	--	--	--	--									
93	4	-0.04	--	--	0.720	--	--	--									
102	0	4.67	--	--	0.940	--	--	--									
109	1	-1.54	--	0.650	--	--	--	--									
113	2	1.33	--	--	0.784	--	--	--									
121	4	-0.47	--	--	0.700	--	--	--									
127	4	0.49	--	--	0.745	--	--	--									
131	1	-1.76	--	--	0.640	--	--	--									
134	4	-0.27	--	--	0.709	--	--	--									
138	4	-0.34	--	--	0.706	--	--	--									
140	4	-0.47	--	0.700	--	--	--	--									
141	1	1.80	--	--	0.806	--	--	--									
142	NR	--	--	--	< 1	--	--	--									
145	2	1.46	--	--	0.790	--	--	--									
146	2	-1.01	--	--	0.675	--	--	--									
151	0	2.31	--	0.830	--	--	--	--									
154	4	0.17	--	--	0.730	--	--	--									
180	4	0.28	--	--	0.735	--	--	--									
185	4	-0.15	--	0.715	--	--	--	--									
190	0	4.03	--	--	--	--	0.910	--									
191	2	1.03	--	--	--	0.770	--	--									
193	3	-0.96	--	--	0.677	--	--	--									
203	0	-2.18	--	--	0.620	--	--	--									
209	3	-0.84	--	--	0.683	--	--	--									
212	4	0.19	--	--	0.731	--	--	--									

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte : Na (Sodium) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
N =	17	44	1	3	3	6	1. AA: direct, air	MPV =	16.3								
Minimum =	14.3	5.1	16.1	16.7	13.2	12.3	4. ICP	F-pseudosigma =	0.9								
Maximum =	18.2	50.7		17.0	18.2	16.9	5. DCP	N =	74								
Median =	16.3	16.3					6. ICP/MS	Uh =	16.8								
F-pseudosigma =	0.4	1.0					7. Ion chromatography	Lh =	15.5								
							12. Flame emission										
Lab	Rating	Z-value	1	4	5	6	7	12	Lab	Rating	Z-value	1	4	5	6	7	12
1	3	0.70	--	16.9	--	--	--	--	209	4	-0.05	--	16.2	--	--	--	--
3	4	-0.16	--	16.1	--	--	--	--	212	3	-0.81	--	15.5	--	--	--	--
10	4	-0.16	16.1	--	--	--	--	--	215	2	1.03	--	17.2	--	--	--	--
11	3	-0.81	--	15.5	--	--	--	--	219	4	-0.27	--	16.0	--	--	--	--
12	0	-2.43	--	14.0	--	--	--	--	234	4	-0.05	--	16.2	--	--	--	--
13	2	1.03	--	17.2	--	--	--	--	236	3	-0.73	--	15.6	--	--	--	--
24	4	0.16	--	16.4	--	--	--	--	241	4	0.27	16.5	--	--	--	--	--
26	0	-3.29	--	--	--	--	13.2	--	247	4	0.10	--	16.3	--	--	--	--
32	4	0.49	--	--	--	16.7	--	--	254	3	0.80	--	17.0	--	--	--	--
33	4	-0.16	--	--	16.1	--	--	--	259	4	0.27	--	16.5	--	--	--	--
38	2	-1.03	15.3	--	--	--	--	--	265	4	-0.05	--	16.2	--	--	--	--
39	3	-0.70	--	15.6	--	--	--	--	268	4	-0.05	16.2	--	--	--	--	--
42	3	-0.81	--	15.5	--	--	--	--	273	0	2.75	--	18.8	--	--	--	--
45	0	2.10	18.2	--	--	--	--	--	274	0	-4.24	--	--	--	--	--	12.3
46	4	0.38	--	16.6	--	--	--	--	275	2	-1.35	--	--	--	--	--	15.0
48	3	0.81	--	--	--	17.0	--	--	277	0	-2.48	--	14.0	--	--	--	--
50	3	0.81	--	17.0	--	--	--	--	279	3	0.59	--	--	--	--	--	16.8
51	4	-0.27	--	--	--	--	--	16.0	282	4	0.38	--	16.6	--	--	--	--
57	3	-0.92	--	15.4	--	--	--	--	287	0	-2.16	14.3	--	--	--	--	--
64	4	-0.49	15.8	--	--	--	--	--	292	4	0.27	16.5	--	--	--	--	--
70	3	0.81	--	17.0	--	--	--	--	302	0	2.14	--	--	--	--	18.2	--
76	3	-0.67	15.6	--	--	--	--	--	304	4	0.49	--	16.7	--	--	--	--
81	2	-1.03	--	15.3	--	--	--	--	307	4	0.27	16.5	--	--	--	--	--
85	4	0.05	16.3	--	--	--	--	--	309	0	-11.99	--	5.1	--	--	--	--
86	4	0.16	--	16.4	--	--	--	--									
87	4	0.38	16.6	--	--	--	--	--									
89	0	-2.75	--	--	--	--	--	13.7									
93	3	-0.92	--	15.4	--	--	--	--									
97	4	-0.27	16.0	--	--	--	--	--									
102	0	37.15	--	50.7	--	--	--	--									
109	4	-0.31	16.0	--	--	--	--	--									
113	3	0.67	--	16.9	--	--	--	--									
121	3	-0.70	--	15.6	--	--	--	--									
127	4	-0.16	--	16.1	--	--	--	--									
131	2	-1.35	--	15.0	--	--	--	--									
134	3	0.54	16.8	--	--	--	--	--									
138	4	0.05	--	16.3	--	--	--	--									
140	4	0.16	16.4	--	--	--	--	--									
141	1	1.78	--	17.9	--	--	--	--									
142	3	0.70	--	16.9	--	--	--	--									
145	4	0.27	--	16.5	--	--	--	--									
146	2	-1.13	--	15.2	--	--	--	--									
151	4	0.16	16.4	--	--	--	--	--									
154	3	-0.81	--	15.5	--	--	--	--									
180	3	0.59	--	16.8	--	--	--	--									
185	3	0.71	--	--	--	--	--	16.9									
190	4	-0.05	--	--	--	--	16.2	--									
191	3	0.81	--	--	--	17.0	--	--									
193	4	0.27	--	16.5	--	--	--	--									
203	2	-1.40	--	15.0	--	--	--	--									

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte : total P as P (total Phosphorus as Phosphorus) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	1	4	3	1	47	0. Other	MPV = 0.386		
Minimum =	0.360	0.316	0.260	0.374	0.050	4. ICP	F-pseudosigma = 0.024		
Maximum =		0.410	0.530		0.453	7. Ion chromatography	N = 56		
Median =					0.387	20. Titrate: colorimetric	Uh = 0.401		
F-pseudosigma =					0.022	22m. Color:phosphomolybdate	Lh = 0.369		

Lab	Rating	Z-value	0	4	7	20	22m	Lab	Rating	Z-value	0	4	7	20	22m
3	4	-0.32	--	--	--	--	0.378	275	0	-14.14	--	--	--	--	0.050
11	4	0.19	--	--	--	--	0.390	277	0	-4.45	--	--	0.280	--	--
12	2	1.45	--	--	--	--	0.420	282	3	0.82	--	--	--	--	0.405
13	3	-0.82	--	--	--	--	0.366	292	3	0.61	--	--	--	--	0.400
23	1	-1.58	--	--	--	--	0.348	305	4	0.02	--	--	--	--	0.386
25	4	0.19	--	--	--	--	0.390	307	2	1.03	--	--	--	--	0.410
38	4	-0.19	--	--	--	--	0.381								
39	3	-0.65	--	--	--	--	0.370								
42	0	-2.93	--	0.316	--	--	--								
46	2	1.29	--	--	--	--	0.416								
51	4	-0.02	--	--	--	--	0.385								
57	2	1.03	--	--	--	--	0.410								
64	3	0.61	--	--	--	--	0.400								
70	0	-3.60	--	--	--	--	0.300								
81	4	0.23	--	--	--	--	0.391								
85	3	0.99	--	--	--	--	0.409								
86	4	0.11	--	0.388	--	--	--								
87	4	0.32	--	--	--	--	0.393								
89	4	0.06	--	--	--	--	0.387								
93	1	1.88	--	--	--	--	0.430								
97	4	-0.23	--	--	--	--	0.380								
102	3	-0.61	--	--	--	--	0.371								
113	4	-0.40	--	--	--	--	0.376								
114	3	-0.65	--	--	--	--	0.370								
127	4	0.36	--	--	--	--	0.394								
131	0	6.09	--	--	0.530	--	--								
134	4	0.19	--	--	--	--	0.390								
138	3	0.65	--	--	--	--	0.401								
140	0	-5.29	--	--	--	--	0.260								
142	0	-8.96	--	--	--	--	0.173								
143	4	-0.44	--	--	--	--	0.375								
145	4	0.19	--	--	--	--	0.390								
154	0	2.72	--	--	--	--	0.450								
183	3	-0.61	--	--	--	--	0.371								
185	4	-0.15	--	--	--	--	0.382								
190	3	-0.78	--	--	--	--	0.367								
191	2	-1.07	0.360	--	--	--	--								
203	4	-0.48	--	--	--	0.374	--								
212	0	-3.52	--	--	--	--	0.302								
213	4	0.19	--	--	--	--	0.390								
215	2	1.45	--	--	--	--	0.420								
227	4	-0.40	--	0.376	--	--	--								
234	4	-0.23	--	--	--	--	0.380								
236	2	1.03	--	0.410	--	--	--								
243	0	-2.76	--	--	--	--	0.320								
247	0	-2.47	--	--	--	--	0.327								
253	0	2.85	--	--	--	--	0.453								
255	2	1.29	--	--	--	--	0.416								
273	0	-5.29	--	--	0.260	--	--								
274	4	0.19	--	--	--	--	0.390								

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte : pH



ANALYTICAL METHOD CODE


—○— 41

SUMMARY			
N =	79	41. Direct reading	MPV = 6.60
Minimum =	5.31		F-pseudostigma = 0.22
Maximum =	7.74		Rating Criterion = 0.33
Median =	6.60		N = 79
F-pseudostigma =	0.22		Uh = 6.75
			Lh = 6.45


Lab	Rating	Z-value	41	Lab	Rating	Z-value	41
1	4	0.33	6.71	209	4	0.30	6.70
3	4	0.27	6.69	212	4	0.33	6.71
10	4	0.21	6.67	213	4	-0.27	6.51
11	3	0.97	6.92	215	4	0.00	6.60
12	3	-0.76	6.35	227	4	-0.12	6.56
13	3	0.70	6.83	234	4	-0.48	6.44
23	4	0.12	6.64	236	2	-1.12	6.23
24	4	0.00	6.60	241	4	-0.30	6.50
25	3	0.61	6.80	243	4	0.42	6.74
26	2	1.30	7.03	244	3	0.52	6.77
33	4	0.36	6.72	247	4	0.48	6.76
38	4	0.30	6.70	253	3	0.64	6.81
39	4	-0.03	6.59	259	4	-0.30	6.50
42	3	-0.70	6.37	265	2	1.21	7.00
43	3	-0.79	6.34	268	4	-0.42	6.46
45	4	-0.09	6.57	273	1	1.52	7.10
46	0	2.52	7.43	274	0	-2.73	5.70
48	3	-0.61	6.40	275	1	-1.82	6.00
50	3	-0.91	6.30	277	3	-0.88	6.31
51	4	-0.06	6.58	279	2	-1.09	6.24
57	4	0.30	6.70	282	4	-0.39	6.47
64	3	0.61	6.80	287	4	0.48	6.76
70	1	-1.88	5.98	292	4	0.45	6.75
81	4	-0.06	6.58	302	4	0.48	6.76
85	3	0.61	6.80	305	3	0.67	6.82
86	4	0.06	6.62	307	4	-0.12	6.56
87	1	-1.79	6.01	308	4	-0.27	6.51
89	0	-3.09	5.58	309	0	3.45	7.74
93	4	0.24	6.68	312	4	-0.27	6.51
96	4	0.45	6.75				
97	4	0.42	6.74				
109	1	-1.82	6.00				
113	4	-0.18	6.54				
114	3	-0.58	6.41				
127	4	0.00	6.60				
134	4	0.12	6.64				
138	3	0.61	6.80				
140	0	-3.91	5.31				
141	1	-1.94	5.96				
142	2	-1.27	6.18				
143	4	-0.06	6.58				
145	4	0.45	6.75				
146	4	0.36	6.72				
151	4	0.15	6.65				
154	4	0.06	6.62				
180	4	0.00	6.60				
183	3	0.97	6.92				
185	4	-0.24	6.52				
190	3	-0.52	6.43				
203	4	-0.15	6.55				

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte: SiO₂ (Silica) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
		N =	1	27	1	3	14	7	0. Other	MPV =	4.50						
		Minimum =	8.00	2.17	4.56	4.22	4.11	4.20	4. ICP	F-pseudosigma =	0.39						
		Maximum =		22.50		4.80	5.13	5.54	5. DCP	N =	53						
		Median =		4.37			4.74	4.66	6. ICP/MS	Uh =	4.79						
		F-pseudosigma =		0.30			0.28	0.25	22. Colorimetric	Lh =	4.26						
									22mb. Color: molybdate blue								
Lab	Rating	Z-value	0	4	5	6	22	22mb	Lab	Rating	Z-value	0	4	5	6	22	22mb
1	4	0.20	--	--	--	--	4.58	--	304	3	-0.51	--	4.30	--	--	--	--
3	3	-0.97	--	4.12	--	--	--	--	307	0	8.91	8.00	--	--	--	--	--
10	3	0.94	--	--	--	--	4.87	--	312	0	2.65	--	--	--	--	--	5.54
11	4	0.08	--	4.53	--	--	--	--									
13	0	45.82	--	22.50	--	--	--	--									
24	4	0.46	--	4.68	--	--	--	--									
26	4	-0.25	--	4.40	--	--	--	--									
32	3	0.76	--	--	--	4.80	--	--									
33	4	0.15	--	--	4.56	--	--	--									
38	4	0.18	--	--	--	--	4.57	--									
39	2	-1.04	--	4.09	--	--	--	--									
42	2	-1.04	--	4.09	--	--	--	--									
50	4	-0.23	--	4.41	--	--	--	--									
57	3	-0.51	--	4.30	--	--	--	--									
59	3	0.76	--	--	--	--	4.80	--									
64	3	-0.89	--	4.15	--	--	--	--									
70	4	0.43	--	--	--	--	4.67	--									
81	3	0.74	--	--	--	--	--	4.79									
85	3	0.76	--	--	--	--	4.80	--									
89	3	-0.76	--	--	--	--	--	4.20									
97	4	-0.13	--	--	--	--	4.45	--									
102	3	0.87	--	--	--	--	4.84	--									
107	4	-0.23	--	--	--	--	4.41	--									
113	4	-0.31	--	--	--	--	--	4.38									
121	4	-0.03	--	4.49	--	--	--	--									
127	4	-0.18	--	4.43	--	--	--	--									
131	2	-1.02	--	4.10	--	--	--	--									
134	4	-0.33	--	4.37	--	--	--	--									
138	4	0.41	--	--	--	--	--	4.66									
140	2	1.09	--	--	--	--	4.93	--									
142	1	-1.63	--	3.86	--	--	--	--									
145	3	-0.61	--	4.26	--	--	--	--									
151	4	-0.03	--	--	--	--	4.49	--									
154	3	-0.76	--	4.20	--	--	--	--									
190	3	0.92	--	--	--	--	--	4.86									
191	3	-0.71	--	--	--	4.22	--	--									
193	2	1.04	--	--	--	--	4.91	--									
203	1	1.60	--	--	--	--	5.13	--									
212	4	0.41	--	4.66	--	--	--	--									
219	2	-1.27	--	4.00	--	--	--	--									
234	4	0.13	--	4.55	--	--	--	--									
236	0	-5.60	--	2.30	--	--	--	--									
241	4	0.03	--	--	--	4.51	--	--									
247	4	0.25	--	--	--	--	--	4.60									
254	4	0.23	--	4.59	--	--	--	--									
259	4	-0.05	--	4.48	--	--	--	--									
265	4	0.00	--	4.50	--	--	--	--									
273	0	5.93	--	6.83	--	--	--	--									
274	3	-0.99	--	--	--	--	4.11	--									
282	0	-5.93	--	2.17	--	--	--	--									

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

SO₄ (Sulfate) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
		N =	3	1	47	8	1	10	4. ICP	MPV =	3.76						
		Minimum =	1.83	4.77	1.64	2.06	3.29	0.32	6. ICP/MS	F-pseudosigma =	0.33						
		Maximum =	4.00		8.54	4.58	3.29	6.00	7. Ion chromatography	N =	69						
		Median =			3.78	3.19		4.15	22. Colorimetric	Uh =	3.93						
		F-pseudosigma =			0.13	0.00		1.18	50. Gravimetric	Lh =	3.48						
									51. Turbidimetric								
Lab	Rating	Z-value	4	6	7	22	50	51	Lab	Rating	Z-value	4	6	7	22	50	51
1	4	-0.15	--	--	3.71	--	--	--	208	3	0.72	--	--	4.00	--	--	--
3	NR	--	--	--	--	< 10	--	--	209	4	-0.21	--	--	3.69	--	--	--
10	2	1.02	--	--	--	--	--	4.10	212	3	0.60	--	--	3.96	--	--	--
11	3	0.72	4.00	--	--	--	--	--	215	0	6.72	--	--	--	--	--	6.00
13	4	0.27	--	--	3.85	--	--	--	219	0	3.03	--	4.77	--	--	--	--
23	4	-0.12	--	--	3.72	--	--	--	227	4	0.24	--	--	3.84	--	--	--
24	0	2.46	--	--	--	4.58	--	--	234	3	0.87	--	--	4.05	--	--	--
25	0	3.72	--	--	5.00	--	--	--	241	4	-0.36	--	--	3.64	--	--	--
26	4	0.42	--	--	3.90	--	--	--	247	0	-2.25	--	--	3.01	--	--	--
33	4	0.06	--	--	3.78	--	--	--	253	0	-8.26	--	--	--	--	--	< 1
39	4	-0.48	--	--	3.60	--	--	--	254	4	0.12	--	--	3.80	--	--	--
42	2	1.32	--	--	4.20	--	--	--	255	2	-1.08	--	--	--	3.40	--	--
43	NR	--	--	--	--	--	--	--	259	4	0.12	--	--	3.80	--	--	--
45	2	1.35	--	--	4.21	--	--	--	265	4	-0.18	--	--	3.70	--	--	--
46	0	-5.10	--	--	--	2.06	--	--	268	4	0.03	--	--	3.77	--	--	--
48	0	-2.28	--	--	--	--	--	3.00	273	0	14.33	--	--	8.54	--	--	--
50	0	-2.22	--	--	--	3.02	--	--	274	0	-10.31	--	--	--	--	--	0.32
51	3	0.51	--	--	3.93	--	--	--	275	1	-1.98	--	--	--	--	--	3.10
57	NR	--	--	--	--	< 5	--	--	277	4	0.15	--	--	3.81	--	--	--
64	4	0.18	--	--	3.82	--	--	--	279	4	-0.24	--	--	--	3.68	--	--
70	4	0.18	--	--	3.82	--	--	--	282	4	0.09	--	--	3.79	--	--	--
76	4	0.08	--	--	3.79	--	--	--	287	3	-0.90	--	--	3.46	--	--	--
81	0	-2.10	--	--	--	3.06	--	--	292	4	-0.18	--	--	3.70	--	--	--
85	4	-0.48	--	--	3.60	--	--	--	302	0	-6.36	--	--	1.64	--	--	--
86	4	-0.06	--	--	3.74	--	--	--	309	0	-5.79	1.83	--	--	--	--	--
87	NR	--	--	--	--	--	--	< 5									
89	4	-0.09	--	--	3.73	--	--	--									
93	4	0.12	--	--	3.80	--	--	--									
96	2	1.32	--	--	--	--	--	4.20									
97	0	-2.67	--	--	--	2.87	--	--									
102	3	-0.96	--	--	3.44	--	--	--									
109	2	-1.41	--	--	--	--	3.29	--									
114	0	5.79	--	--	--	--	--	5.69									
127	4	-0.30	3.66	--	--	--	--	--									
131	2	1.08	--	--	4.12	--	--	--									
134	3	-0.84	--	--	3.48	--	--	--									
138	4	0.06	--	--	3.78	--	--	--									
140	0	-8.27	--	--	--	--	--	1.00									
141	4	-0.18	--	--	3.70	--	--	--									
142	3	0.51	--	--	3.93	--	--	--									
145	4	-0.21	--	--	3.69	--	--	--									
146	NR	--	--	--	--	--	--	< 5									
151	4	0.00	--	--	3.76	--	--	--									
154	0	2.49	--	--	--	--	--	4.59									
180	4	0.45	--	--	3.91	--	--	--									
185	4	0.06	--	--	3.78	--	--	--									
190	1	-1.68	--	--	3.20	--	--	--									
191	4	-0.06	--	--	3.74	--	--	--									
196	4	-0.06	--	--	3.74	--	--	--									
203	2	-1.35	--	--	--	3.31	--	--									

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte: Sp Cond (Specific Conductance) Concentration Unit : mS/cm



ANALYTICAL METHOD CODE




SUMMARY					
N =	1	73	4. ICP	MPV =	113
Minimum =	110	70	41. Direct reading	F-pseudostigma =	3
Maximum =		207		Rating Criterion =	6
Median =		113		N =	74
F-pseudostigma =		3		Uh =	115
				Lh =	111

Lab	Rating	Z-value	4	41	Lab	Rating	Z-value	4	41
1	4	0.18	--	114	193	3	0.71	--	117
3	4	0.00	--	113	203	4	0.14	--	114
10	4	0.00	--	113	212	3	0.88	--	118
11	3	-0.57	110	--	215	0	-2.69	--	98
12	4	0.18	--	114	227	4	-0.35	--	111
13	4	0.35	--	115	234	4	0.18	--	114
23	4	-0.28	--	111	236	4	0.00	--	113
24	2	-1.24	--	106	241	3	-0.53	--	110
25	2	1.24	--	120	243	4	0.18	--	114
26	4	-0.18	--	112	244	4	-0.18	--	112
33	4	0.35	--	115	247	4	0.00	--	113
38	4	-0.16	--	112	253	4	0.05	--	113
42	2	1.13	--	119	259	4	0.18	--	114
43	4	0.00	--	113	268	4	-0.18	--	112
45	4	0.18	--	114	273	4	0.35	--	115
46	0	16.64	--	207	274	4	0.16	--	114
48	3	-0.88	--	108	275	0	2.12	--	125
50	4	0.00	--	113	277	2	1.24	--	120
51	4	0.00	--	113	279	3	-0.88	--	108
57	3	-0.53	--	110	282	0	2.48	--	127
64	4	-0.18	--	112	292	4	0.00	--	113
70	3	-0.71	--	109	302	4	0.35	--	115
76	4	-0.07	--	113	307	4	0.00	--	113
81	4	-0.18	--	112	312	0	-2.34	--	100
85	3	-0.53	--	110					
86	2	-1.06	--	107					
87	0	-7.65	--	70					
89	3	0.88	--	118					
93	4	-0.44	--	111					
96	3	0.53	--	116					
97	3	0.71	--	117					
102	3	0.71	--	117					
109	3	0.53	--	116					
113	4	0.11	--	114					
114	3	0.71	--	117					
127	4	0.18	--	114					
134	4	0.02	--	113					
138	4	0.00	--	113					
140	3	-0.53	--	110					
141	0	-2.30	--	100					
142	4	0.35	--	115					
143	3	-0.88	--	108					
145	4	-0.35	--	111					
146	0	2.48	--	127					
151	4	-0.18	--	112					
154	4	0.18	--	114					
180	3	-0.88	--	108					
183	1	-1.77	--	103					
185	0	3.29	--	132					
190	3	0.53	--	116					

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte: Sr (Strontium) Concentration Unit : mg/L




SUMMARY						
N =	26	1	6	4. ICP	MPV =	18.3
Minimum =	16.0	18.6	17.4	5. DCP	F-pseudostandard deviation =	0.9
Maximum =	22.0		19.7	6. ICP/MS	N =	33
Median =	18.2				Uh =	19.0
F-pseudostandard deviation =	0.9				Lh =	17.8

Lab	Rating	Z-value	4	5	6
1	3	-0.52	--	--	17.8
3	4	-0.40	17.9	--	--
11	3	0.83	19.0	--	--
24	0	4.20	22.0	--	--
32	3	0.72	--	--	18.9
33	4	0.38	--	18.6	--
42	2	-1.30	17.1	--	--
57	4	0.49	18.7	--	--
70	0	3.75	21.6	--	--
81	0	-2.54	16.0	--	--
86	3	-0.52	17.8	--	--
102	0	2.52	20.5	--	--
113	4	-0.18	18.1	--	--
121	4	-0.29	18.0	--	--
127	4	-0.07	18.2	--	--
131	3	-0.85	17.5	--	--
134	4	0.07	18.3	--	--
138	3	-0.97	17.4	--	--
141	0	2.63	20.6	--	--
142	0	4.09	21.9	--	--
145	3	0.83	19.0	--	--
151	3	-0.97	--	--	17.4
154	4	-0.29	18.0	--	--
191	4	-0.29	--	--	18.0
212	3	0.61	18.8	--	--
219	1	1.62	--	--	19.7
234	4	-0.07	18.2	--	--
236	4	-0.29	18.0	--	--
247	2	-1.42	17.0	--	--
254	2	1.34	19.5	--	--
259	3	0.83	19.0	--	--
265	4	0.27	--	--	18.5
304	3	-0.85	17.5	--	--

Table 12. Statistical summary of reported data for standard reference water sample M-152 (major constituents)--Continued

Analyte: V (Vanadium) Concentration Unit : mg/L



SUMMARY

N =	3	28	6
Minimum =	29.9	19.9	30.0
Maximum =	38.7	63.8	31.7
Median =		29.8	
F-pseudosigma =		2.1	

- 3. AA: graphite furnace
- 4. ICP
- 6. ICP/MS

MPV =	30.0
F-pseudosigma =	1.9
N =	37
Uh =	31.6
Lh =	29.0

Lab	Rating	Z-value	3	4	6
1	4	0.26	--	--	30.5
3	3	-0.73	--	28.6	--
11	3	-0.52	--	29.0	--
13	4	0.31	--	30.6	--
24	3	0.88	--	31.7	--
26	3	0.83	--	31.6	--
32	4	0.10	--	--	30.2
39	0	-5.24	--	19.9	--
42	3	0.88	--	--	31.7
46	4	0.00	--	30.0	--
48	4	0.00	--	--	30.0
57	4	-0.47	--	29.1	--
70	NR	--	--	< 50	--
81	0	-2.08	--	26.0	--
86	3	0.78	--	31.5	--
89	2	1.09	32.1	--	--
93	4	-0.47	--	29.1	--
97	0	4.51	38.7	--	--
102	3	0.88	--	31.7	--
121	0	3.11	--	36.0	--
127	4	-0.05	29.9	--	--
134	4	-0.18	--	29.7	--
138	3	-0.52	--	29.0	--
141	4	0.31	--	30.6	--
142	4	0.00	--	--	30.0
145	0	2.59	--	35.0	--
146	3	-0.88	--	28.3	--
154	4	0.00	--	30.0	--
180	0	17.54	--	63.8	--
212	3	-0.57	--	28.9	--
215	0	-3.11	--	24.0	--
234	3	-0.99	--	28.1	--
236	1	-1.56	--	27.0	--
241	4	0.10	--	--	30.2
247	1	1.56	--	33.0	--
265	4	0.00	--	30.0	--
282	4	-0.42	--	29.2	--
304	4	0.00	--	30.0	--

Table 13. *Statistical summary of reported data for standard reference sample N-63 (nutrient constituents)*


Definition of analytical methods, abbreviations, and symbols		
<u>Analytical methods</u>		
0. Other/Not reported		
4. ICP	=	inductively coupled plasma
5. DCP	=	direct coupled plasma
7. IC	=	ion chromatography
20. Titrate: color	=	titration: colorimetric (color reagent specified)
21. Titrate: electro	=	titration: electrometric
22. Color	=	colorimetric (color reagent specified)
40. Ion electrode	=	ion selective electrode

<u>Abbreviations and figure symbols</u>		
N	=	number of analyses--(excluding less than values)
MPV	=	most probable value ---
F-pseudostigma	=	nonparametric statistic deviation
Uh	=	upper hinge value
Lh	=	lower hinge value
Uwl	=	upper warning limit - - - -
Lwl	=	lower warning limit - - - -
Ucl	=	upper warning limit
Lcl	=	lower warning limit - - - -
mg/L	=	milligrams per liter
Lab	=	laboratory code number
NR	=	not rated, less than value reported or insufficient data
<	=	less than
--	=	not reported

<u>Constituent</u>		<u>page</u>
NH ₃ as N	Ammonia as nitrogen	82
NH ₃ +Org N as N	Ammonia plus organic nitrogen as nitrogen	83
NO ₃ as N	Nitrate as nitrogen	84
Total P as P	Total Phosphorus as phosphorus	85
PO ₄ as P	Orthophosphate as phosphorus	86

Table 13. Statistical summary of reported data for standard reference water sample N-63 (nutrient constituents)--Continued

Analyte : NH₃ as N (Ammonia as nitrogen) Concentration Unit : mg/L




ANALYTICAL METHOD CODE

—x— 22 —□— 22n —*— 22p —●— 40


SUMMARY																		
N =	3	1	12	2	51	11	0. Other	MPV =	0.150									
Minimum =	0.060	0.248	0.130	0.141	0.073	0.100	7. Ion chromatography	F-pseudosigma =	0.021									
Maximum =	1.227		0.430	0.144	0.262	0.340	22. Colorimetric	N =	80									
Median =			0.155		0.149	0.140	22n. Color: Nesslerization	Uh =	0.160									
F-pseudosigma =			0.091		0.020	0.024	22p. Color: phenate	Lh =	0.132									
							40. Ion selective electrode											
Lab	Rating	Z-value	0	7	22	22n	22p	40	Lab	Rating	Z-value	0	7	22	22n	22p	40	
1	4	-0.05	--	--	--	--	0.149	--	193	0	-2.36	--	--	--	--	--	0.101	--
10	4	-0.48	--	--	--	--	--	0.140	198	4	-0.24	--	--	--	--	--	0.145	--
11	0	13.49	--	--	0.430	--	--	--	205	2	-1.25	--	--	--	--	--	0.124	--
13	2	-1.30	--	--	--	--	0.123	--	209	3	0.63	--	--	--	--	--	--	0.163
18	0	5.40	--	--	--	--	0.262	--	212	4	0.00	--	--	0.150	--	--	--	--
21	4	0.05	--	--	--	--	0.151	--	213	NR	< 1	--	--	--	--	--	--	--
23	4	0.10	--	--	--	--	0.152	--	215	2	-1.45	--	--	--	--	--	0.120	--
25	3	-0.96	--	--	0.130	--	--	--	224	2	1.49	--	--	--	--	--	0.181	--
33	4	0.00	--	--	--	--	0.150	--	227	4	-0.43	--	--	--	0.141	--	--	--
36	0	51.89	1.227	--	--	--	--	--	234	4	0.10	--	--	--	--	--	--	0.152
38	4	0.00	--	--	--	--	0.150	--	241	3	-0.84	--	--	--	--	--	--	0.133
46	2	-1.06	--	--	--	--	0.128	--	243	0	2.31	--	--	--	--	--	0.198	--
48	0	12.04	--	--	0.400	--	--	--	247	0	-3.71	--	--	--	--	--	0.073	--
50	4	-0.34	--	--	0.143	--	--	--	253	4	0.05	--	--	0.151	--	--	--	--
51	4	0.48	--	--	--	--	--	0.160	255	4	-0.24	--	--	--	--	--	0.145	--
59	1	-1.93	--	--	--	--	0.110	--	282	0	11.08	--	--	0.380	--	--	--	--
64	4	0.00	--	--	--	--	0.150	--	285	0	23.99	0.648	--	--	--	--	--	--
70	0	-2.17	--	--	--	--	0.105	--	287	0	4.72	--	0.248	--	--	--	--	--
76	4	-0.07	--	--	--	--	0.149	--	292	0	-2.41	--	--	--	--	--	0.100	--
81	2	1.11	--	--	--	--	0.173	--	301	4	-0.29	--	--	--	--	--	0.144	--
83	3	-0.96	--	--	--	--	0.130	--	306	3	0.72	--	--	--	--	--	--	0.165
85	4	0.10	--	--	--	--	0.152	--	307	2	-1.11	--	--	--	--	--	--	0.127
86	1	1.78	--	--	--	--	0.187	--	308	NR	--	--	< 0.15	--	--	--	--	--
89	4	-0.29	--	--	--	--	0.144	--	312	3	-0.77	--	--	--	--	--	0.134	--
90	4	0.29	--	--	--	--	0.156	--	313	3	0.63	--	--	--	--	--	0.163	--
93	3	-0.96	--	--	--	--	0.130	--	314	3	0.72	--	--	--	--	--	0.165	--
96	3	-0.87	--	--	--	--	0.132	--	316	4	0.18	--	--	0.154	--	--	--	--
97	4	0.00	--	--	0.150	--	--	--	317	4	0.34	--	--	--	--	--	0.157	--
102	3	0.96	--	--	--	--	0.170	--	318	4	0.29	--	--	--	--	--	0.156	--
105	4	-0.48	--	--	--	--	0.140	--	319	0	-4.34	0.060	--	--	--	--	--	--
110	3	0.72	--	--	0.165	--	--	--	320	2	-1.45	--	--	--	--	--	0.120	--
113	4	-0.24	--	--	--	--	0.145	--	321	4	0.00	--	--	--	--	--	0.150	--
114	2	-1.45	--	--	--	--	--	0.120										
127	3	-0.82	--	--	--	--	0.133	--										
129	4	-0.29	--	--	--	0.144	--	--										
134	4	-0.10	--	--	--	--	0.148	--										
138	4	0.10	--	--	--	--	0.152	--										
140	4	0.48	--	--	0.160	--	--	--										
142	4	0.29	--	--	0.156	--	--	--										
143	3	0.96	--	--	--	--	0.170	--										
145	4	0.48	--	--	--	--	0.160	--										
146	0	-2.41	--	--	--	--	0.100	--										
151	4	-0.48	--	--	--	--	--	0.140										
154	4	-0.05	--	--	--	--	0.149	--										
155	2	-1.01	--	--	--	--	0.129	--										
158	4	0.34	--	--	--	--	0.157	--										
180	4	0.10	--	--	--	--	0.152	--										
183	0	9.15	--	--	--	--	--	0.340										
185	2	-1.01	--	--	--	--	0.129	--										
190	2	1.01	--	--	--	--	0.171	--										

Table 13. Statistical summary of reported data for standard reference water sample N-63 (nutrient constituents)--Continued

Analyte : NH₃ + Organic N as N (Ammonia + organic nitrogen as nitrogen) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
N =	3	2	15	2	24	2	0. Other	MPV = 0.200									
Minimum =	0.157	0.192	0.120	0.170	0.131	0.160	20. Titrate: colorimetric	F-pseudostigma = 0.066									
Maximum =	0.698	0.230	0.530	0.187	0.700	0.190	22. Colorimetric	N = 48									
Median =			0.230		0.200		22n. Color: Nesslerization	Uh = 0.267									
F-pseudostigma =			0.164		0.053		22p. Color: phenate	Lh = 0.178									
							40. Ion selective electrode										
Lab	Rating	Z-value	0	20	22	22n	22p	40	Lab	Rating	Z-value	0	20	22	22n	22p	40
1	4	-0.18	--	--	--	--	0.188	--	319	4	0.45	--	0.230	--	--	--	--
10	4	-0.15	--	--	--	--	--	0.190	320	4	0.21	--	--	0.214	--	--	--
11	1	1.52	--	--	0.300	--	--	--									
18	2	1.02	--	--	--	--	0.267	--									
21	3	-0.55	--	--	--	--	0.164	--									
23	0	7.58	--	--	--	--	0.700	--									
36	3	0.91	0.260	--	--	--	--	--									
38	4	-0.45	--	--	--	0.170	--	--									
46	3	0.91	--	--	0.260	--	--	--									
48	0	5.00	--	--	0.530	--	--	--									
50	4	0.45	--	--	0.230	--	--	--									
51	3	-0.61	--	--	--	--	--	0.160									
59	4	0.00	--	--	--	--	0.200	--									
70	0	3.79	--	--	--	--	0.450	--									
81	2	-1.05	--	--	--	--	0.131	--									
89	3	0.61	--	--	--	--	0.240	--									
90	4	-0.29	--	--	0.181	--	--	--									
96	3	-0.52	--	--	--	--	0.166	--									
97	2	-1.21	--	--	0.120	--	--	--									
102	4	0.00	--	--	--	--	0.200	--									
105	NR		--	--	<1	--	--	--									
113	NR		--	--	--	--	<0.5	--									
127	4	-0.33	--	--	--	--	0.178	--									
129	4	-0.20	--	--	--	0.187	--	--									
134	3	-0.99	--	--	--	--	0.135	--									
138	4	0.39	--	--	--	--	0.226	--									
140	4	-0.15	--	--	0.190	--	--	--									
142	3	-0.67	--	--	0.156	--	--	--									
143	4	0.00	--	--	--	--	0.200	--									
145	4	0.00	--	--	--	--	0.200	--									
146	4	0.45	--	--	--	--	0.230	--									
154	1	1.64	--	--	--	--	0.308	--									
155	4	0.15	--	--	--	--	0.210	--									
158	3	-0.67	--	--	--	--	0.156	--									
180	4	0.39	--	--	--	--	0.226	--									
193	NR		--	--	< 0.2	--	--	--									
209	3	-0.65	0.157	--	--	--	--	--									
212	0	3.70	--	--	0.444	--	--	--									
213	NR		--	< 1	--	--	--	--									
224	0	3.93	--	--	--	--	0.459	--									
227	4	-0.49	--	--	0.168	--	--	--									
247	4	-0.20	--	--	--	--	0.187	--									
282	0	2.58	--	--	0.370	--	--	--									
285	0	7.55	0.698	--	--	--	--	--									
306	4	-0.13	--	0.192	--	--	--	--									
308	0	4.85	--	--	0.520	--	--	--									
313	2	1.18	--	--	--	--	0.278	--									
314	0	4.35	--	--	0.487	--	--	--									
316	4	-0.01	--	--	0.199	--	--	--									
318	4	0.00	--	--	--	--	0.200	--									

Table 13. Statistical summary of reported data for standard reference water sample N-63 (nutrient constituents)--Continued

Analyte : NO₃ as N (Nitrate as nitrogen) Concentration Unit : mg/L




ANALYTICAL METHOD CODE




SUMMARY																	
		N =	4	27	9	32	4	3	0. Other	MPV =	0.084						
		Minimum =	0.07	0.00	0.07	0.02	0.08	0.05	7. Ion chromatography	F-pseudosigma =	0.008						
		Maximum =	0.68	0.25	0.15	0.13	0.16	0.08	22. Colorimetric	N =	79						
		Median =		0.081	0.082	0.086			22cd. Cd diazotization	Uh =	0.091						
		F-pseudosigma =		0.008	0.016	0.007			22h. Color: hydrazine diazotization	Lh =	0.080						
									22sulf. Color: Sulfanilamide								
Lab	Rating	Z-value	0	7	22	22cd	22h	22sulf	Lab	Rating	Z-value	0	7	22	22cd	22h	22sulf
1	4	0.03	--	--	--	--	--	0.084	185	4	-0.10	--	0.083	--	--	--	--
10	4	-0.49	--	--	--	0.080	--	--	193	4	-0.49	--	0.080	--	--	--	--
11	4	-0.49	--	--	--	--	--	0.080	198	4	0.03	--	--	--	0.084	--	--
12	3	0.80	--	--	--	0.090	--	--	205	0	5.68	--	--	--	0.128	--	--
13	1	-1.52	--	0.072	--	--	--	--	209	3	0.54	--	0.088	--	--	--	--
18	3	-0.87	--	--	--	0.077	--	--	212	0	20.07	--	0.240	--	--	--	--
21	4	0.15	--	--	--	--	0.085	--	215	1	-1.77	--	--	--	0.070	--	--
23	4	0.15	--	0.085	--	--	--	--	224	3	0.54	--	0.088	--	--	--	--
25	4	-0.49	--	0.080	--	--	--	--	227	4	-0.36	--	0.081	--	--	--	--
33	1	-1.77	--	0.070	--	--	--	--	234	4	-0.49	--	0.080	--	--	--	--
36	4	0.28	0.086	--	--	--	--	--	241	3	-0.87	--	0.077	--	--	--	--
38	4	0.03	--	--	--	0.084	--	--	243	4	-0.49	--	--	--	0.080	--	--
42	0	2.98	--	0.107	--	--	--	--	247	3	-1.00	--	0.076	--	--	--	--
46	0	2.08	--	--	--	0.100	--	--	253	0	2.34	--	--	0.102	--	--	--
48	0	9.79	--	--	--	--	0.160	--	255	NR	--	--	--	--	< 0.1	--	--
50	4	-0.23	--	--	0.082	--	--	--	282	0	2.60	--	--	0.104	--	--	--
51	0	2.08	--	0.100	--	--	--	--	285	1	-1.90	0.069	--	--	--	--	--
53	4	-0.49	0.080	--	--	--	--	--	287	0	21.35	--	0.250	--	--	--	--
59	3	0.80	--	--	--	0.090	--	--	292	4	-0.49	--	0.080	--	--	--	--
64	4	-0.49	--	--	0.080	--	--	--	301	1	-2.03	--	0.068	--	--	--	--
69	4	-0.49	--	--	--	0.080	--	--	306	0	-2.65	--	--	--	0.063	--	--
70	3	0.80	--	--	--	0.090	--	--	307	2	1.18	--	--	--	0.093	--	--
76	2	-1.13	--	0.075	--	--	--	--	308	NR	--	--	--	< 0.15	--	--	--
81	2	1.05	--	--	--	--	0.092	--	313	4	-0.33	--	--	--	0.081	--	--
85	4	0.41	--	--	--	0.087	--	--	314	3	0.93	--	--	--	0.091	--	--
86	0	-8.58	--	--	--	0.017	--	--	316	4	-0.03	--	--	--	0.084	--	--
89	4	0.15	--	--	--	0.085	--	--	317	0	4.01	--	--	--	0.115	--	--
90	4	-0.49	--	--	--	--	0.080	--	318	4	-0.36	--	--	--	0.081	--	--
93	4	-0.49	--	0.080	--	--	--	--	319	0	2.08	--	--	--	0.100	--	--
96	3	0.67	--	--	--	0.089	--	--	320	4	-0.36	--	--	--	0.081	--	--
97	4	-0.49	--	--	0.080	--	--	--	321	3	-0.75	--	0.078	--	--	--	--
102	0	-10.77	--	0.000	--	--	--	--									
105	1	-1.77	--	--	0.070	--	--	--									
113	4	-0.23	--	--	--	0.082	--	--									
114	0	2.08	--	--	--	0.100	--	--									
127	2	1.14	--	0.093	--	--	--	--									
129	4	-0.36	--	0.081	--	--	--	--									
134	3	0.80	--	--	--	0.090	--	--									
138	4	0.28	--	0.086	--	--	--	--									
140	3	-0.75	--	--	0.078	--	--	--									
142	4	0.28	--	--	0.086	--	--	--									
143	4	0.28	--	--	--	0.086	--	--									
145	3	0.80	--	--	--	0.090	--	--									
146	4	-0.23	--	--	--	0.082	--	--									
151	2	1.18	--	0.093	--	--	--	--									
154	0	-4.34	--	--	--	--	--	0.050									
155	1	1.68	--	--	--	0.097	--	--									
158	0	7.99	--	--	0.146	--	--	--									
180	3	0.54	--	0.088	--	--	--	--									
183	0	76.60	0.680	--	--	--	--	--									

Table 13. Statistical summary of reported data for standard reference water sample N-63 (nutrient constituents)--Continued

Analyte : total P as P (total Phosphorus as phosphorus) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	2	2	1	64	0. Other	MPV = 0.158			
Minimum =	0.14	0.16	0.08	0.10	4. ICP	F-pseudosigma = 0.010			
Maximum =	0.18	0.16		0.27	7. Ion chromatography	N = 69			
Median =				0.16	22m. Color:phosphomolybdate	Uh = 0.164			
F-pseudosigma =				0.010		Lh = 0.150			

Lab	Rating	Z-value	0	4	7	22m	Lab	Rating	Z-value	0	4	7	22m
1	4	0.00	--	--	--	0.158	227	4	-0.19	--	0.156	--	--
10	4	0.00	--	--	--	0.158	234	3	-0.77	--	--	--	0.150
11	1	-1.73	--	--	--	0.140	241	4	0.19	--	--	--	0.160
12	0	-2.70	--	--	--	0.130	243	3	-0.77	--	--	--	0.150
13	4	0.39	--	--	--	0.162	247	0	-2.99	--	--	--	0.127
18	4	-0.10	--	--	--	0.157	253	0	8.77	--	--	--	0.249
21	4	-0.10	--	--	--	0.157	255	4	0.00	--	--	--	0.158
23	0	10.89	--	--	--	0.271	282	1	1.93	--	--	--	0.178
25	4	0.19	--	--	--	0.160	285	2	-1.35	0.144	--	--	--
36	0	2.12	0.180	--	--	--	292	2	1.16	--	--	--	0.170
38	4	0.48	--	--	--	0.163	301	3	0.67	--	--	--	0.165
46	4	0.00	--	--	--	0.158	306	3	0.77	--	--	--	0.166
48	3	-0.77	--	--	--	0.150	307	1	1.93	--	--	--	0.178
50	4	0.00	--	--	--	0.158	308	NR	--	--	--	--	< 0.15
51	4	-0.48	--	--	--	0.153	313	3	0.58	--	--	--	0.164
59	0	-5.59	--	--	--	0.100	314	2	1.16	--	--	--	0.170
64	4	0.00	--	--	--	0.158	316	3	0.74	--	--	--	0.166
70	0	4.53	--	--	--	0.205	318	4	-0.29	--	--	--	0.155
81	1	-1.93	--	--	--	0.138	319	0	2.12	--	--	--	0.180
85	3	0.96	--	--	--	0.168	320	3	-0.96	--	--	--	0.148
86	4	0.10	--	0.159	--	--	321	4	0.00	--	--	--	0.158
89	4	0.29	--	--	--	0.161							
93	3	-0.77	--	--	--	0.150							
96	1	-1.54	--	--	--	0.142							
97	NR	--	--	--	--	< 0.18							
102	2	-1.06	--	--	--	0.147							
105	0	5.97	--	--	--	0.220							
113	4	0.19	--	--	--	0.160							
114	4	0.19	--	--	--	0.160							
127	4	-0.29	--	--	--	0.155							
129	4	0.39	--	--	--	0.162							
134	4	0.00	--	--	--	0.158							
138	4	0.00	--	--	--	0.158							
140	0	-4.63	--	--	--	0.110							
142	3	-0.96	--	--	--	0.148							
143	4	-0.39	--	--	--	0.154							
145	4	0.19	--	--	--	0.160							
146	3	-0.77	--	--	--	0.150							
154	3	0.58	--	--	--	0.164							
155	4	0.30	--	--	--	0.161							
158	1	-1.54	--	--	--	0.142							
180	1	2.02	--	--	--	0.179							
183	4	0.00	--	--	--	0.158							
185	4	-0.10	--	--	--	0.157							
193	4	0.10	--	--	--	0.159							
198	1	-1.83	--	--	--	0.139							
212	0	-2.22	--	--	--	0.135							
213	4	0.19	--	--	--	0.160							
215	0	5.97	--	--	--	0.220							
224	3	-0.77	--	--	--	0.150							


Table 13. Statistical summary of reported data for standard reference water sample N-63 (nutrient constituents)--Continued

Analyte : PO₄ as P (Orthophosphate as phosphorus)

Concentration Unit : mg/L



ANALYTICAL METHOD CODE



SUMMARY					
N =	4	10	59	0. Other	MPV = 0.142
Minimum =	0.120	0.078	0.074	7. Ion chromatography	F-pseudosigma = 0.010
Maximum =	0.145	0.420	1.800	22m. Color:phosphomolybdate	N = 73
Median =		0.137	0.142		Uh = 0.148
F-pseudosigma =		0.017	0.006		Lh = 0.135


Lab	Rating	Z-value	0	7	22m	Lab	Rating	Z-value	0	7	22m
1	4	0.00	--	--	0.142	213	2	-1.25	--	--	0.130
10	4	0.31	--	--	0.145	215	0	3.94	--	--	0.180
11	0	4.98	--	--	0.190	224	4	0.21	--	--	0.144
12	3	-0.83	--	--	0.134	227	2	-1.04	0.132	--	--
13	1	1.66	--	0.158	--	234	2	1.04	--	0.152	--
18	4	0.10	--	--	0.143	241	3	-0.52	--	0.137	--
21	4	-0.21	--	--	0.140	247	3	-0.83	--	0.134	--
23	1	-1.56	--	--	0.127	253	4	0.21	--	--	0.144
25	4	-0.42	--	--	0.138	255	4	-0.21	--	--	0.140
33	3	0.83	--	0.150	--	282	0	3.53	--	--	0.176
36	0	-2.28	0.120	--	--	287	0	4.98	--	0.190	--
38	4	-0.31	--	--	0.139	292	3	-0.73	--	--	0.135
42	3	-0.83	--	0.134	--	301	4	-0.21	--	--	0.140
46	4	0.42	--	--	0.146	307	3	0.62	--	--	0.148
48	0	-7.06	--	--	0.074	308	NR		--	--	< 0.15
50	3	-0.73	--	--	0.135	312	4	0.10	--	--	0.143
51	4	0.21	--	--	0.144	313	3	-0.62	--	--	0.136
53	4	-0.42	--	--	0.138	314	0	2.49	--	--	0.166
59	4	-0.21	--	--	0.140	316	4	-0.17	--	--	0.140
64	4	-0.10	--	--	0.141	317	4	0.00	--	--	0.142
70	2	1.04	--	--	0.152	318	4	0.31	0.145	--	--
81	4	-0.42	--	--	0.138	319	4	0.42	--	--	0.146
83	4	-0.21	--	--	0.140	320	1	-1.56	--	--	0.127
85	4	0.10	--	--	0.143	321	2	1.45	--	--	0.156
89	3	0.52	--	--	0.147						
93	0	2.59	--	--	0.167						
96	2	-1.35	--	--	0.129						
97	1	1.87	--	--	0.160						
102	0	-2.70	--	0.116	--						
105	2	1.04	--	--	0.152						
113	4	0.42	--	--	0.146						
127	2	-1.14	--	0.131	--						
129	4	0.31	--	--	0.145						
134	4	-0.31	--	--	0.139						
138	4	-0.31	--	--	0.139						
140	0	-3.32	--	--	0.110						
142	2	-1.04	--	--	0.132						
143	4	0.10	--	--	0.143						
145	3	0.83	--	--	0.150						
146	0	172.05	--	--	1.800						
151	0	28.85	--	0.420	--						
154	3	-0.52	--	--	0.137						
155	4	0.22	--	--	0.144						
158	4	0.00	--	--	0.142						
180	3	0.93	--	--	0.151						
183	4	0.21	0.144	--	--						
185	3	-0.73	--	--	0.135						
190	4	-0.31	--	--	0.139						
198	2	-1.04	--	--	0.132						
212	2	1.25	--	--	0.154						

Table 14. *Statistical summary of reported data for standard reference sample N-64 (nutrient constituents)*


Definition of analytical methods, abbreviations, and symbols		
<u>Analytical methods</u>		
0. Other/Not reported		
4. ICP	=	inductively coupled plasma
5. DCP	=	direct coupled plasma
7. IC	=	ion chromatography
20. Titrate: color	=	titration: colorimetric (color reagent specified)
21. Titrate: electro	=	titration: electrometric
22. Color	=	colorimetric (color reagent specified)
40. Ion electrode	=	ion selective electrode
<hr/>		
<u>Abbreviations and figure symbols</u>		
	N =	number of analyses--(excluding less than values)
	MPV =	most probable value ---
	F-pseudostigma =	nonparametric statistic deviation
	Uh =	upper hinge value
	Lh =	lower hinge value
	Uwl =	upper warning limit - - - -
	Lwl =	lower warning limit - - - -
	Ucl =	upper warning limit
	Lcl =	lower warning limit - - - -
	mg/L =	milligrams per liter
	Lab =	laboratory code number
	NR =	not rated, less than value reported or insufficient data
	< =	less than
	-- =	not reported
<hr/>		
<u>Constituent</u>		<u>page</u>
NH ₃ as N	Ammonia as nitrogen	88
NH ₃ +Org N as N	Ammonia plus organic nitrogen as nitrogen	89
NO ₃ as N	Nitrate as nitrogen	90
Total P as P	Total Phosphorus as phosphorus	91
PO ₄ as P	Orthophosphate as phosphorus	92

Table 14. Statistical summary of reported data for standard reference water sample N-64 (nutrient constituents)--Continued

Analyte : NH₃ as N (Ammonia as Nitrogen) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	2	1	30	29	9	0. Other	MPV =	1.38	
Minimum =	0.02	1.40	1.08	0.76	0.70	20. Titrate: colorimetric	F-pseudostigma =	0.08	
Maximum =	2.28		1.65	1.63	1.50	22. Colorimetric	N =	70	
Median =			1.38	1.39	1.38	22p. Color: phenate	Uh =	1.43	
F-pseudostigma =			0.05	0.07	0.33	40. Ion selective electrode	Lh =	1.32	


Lab	Rating	Z-value	0	20	22	22p	40	Lab	Rating	Z-value	0	20	22	22p	40
1	3	0.57	--	--	--	1.43	--	212	3	0.69	--	--	1.44	--	--
3	2	-1.25	--	--	--	1.28	--	213	4	0.21	--	1.40	--	--	--
10	4	-0.04	--	--	--	--	1.38	215	4	-0.16	--	--	1.37	--	--
11	0	3.00	--	--	--	1.63	--	224	1	1.86	--	--	--	1.54	--
12	0	-3.43	--	--	--	1.10	--	227	2	1.30	--	--	1.49	--	--
13	3	-0.64	--	--	--	1.33	--	234	4	0.45	--	--	--	--	1.42
18	3	-0.76	--	--	--	1.32	--	243	2	1.01	--	--	--	1.47	--
23	4	-0.40	--	--	--	1.35	--	247	3	-0.89	--	--	--	1.31	--
25	3	-0.89	--	--	--	1.31	--	253	4	0.36	--	--	1.41	--	--
36	0	-16.60	0.02	--	--	--	--	255	4	0.45	--	--	1.42	--	--
38	4	0.27	--	--	--	1.41	--	282	2	-1.01	--	--	1.30	--	--
46	3	-0.52	--	--	--	1.34	--	285	0	-8.33	--	--	--	--	0.70
59	4	0.08	--	--	--	1.39	--	287	0	10.88	2.28	--	--	--	--
64	4	-0.28	--	--	--	1.36	--	292	0	-4.89	--	--	--	--	0.98
70	0	-7.57	--	--	--	0.76	--	306	0	-7.27	--	--	--	--	0.78
76	4	0.24	--	--	1.40	--	--	312	0	-4.28	--	--	--	1.03	--
83	4	-0.04	--	--	1.38	--	--	313	4	-0.04	--	--	1.38	--	--
85	4	0.33	--	--	1.41	--	--	316	4	0.50	--	--	1.42	--	--
87	4	0.45	--	--	1.42	--	--	319	4	-0.04	--	--	1.38	--	--
89	3	-0.52	--	--	--	1.34	--	320	4	-0.16	--	--	1.37	--	--
90	2	1.42	--	--	1.50	--	--								
93	2	-1.01	--	--	--	1.30	--								
96	3	-0.97	--	--	--	1.30	--								
102	1	-1.98	--	--	1.22	--	--								
105	3	-0.52	--	--	1.34	--	--								
107	4	0.25	--	--	--	1.40	--								
113	4	0.00	--	--	1.38	--	--								
114	2	-1.25	--	--	--	--	1.28								
127	4	0.08	--	--	--	1.39	--								
129	2	-1.24	--	--	1.28	--	--								
134	1	1.54	--	--	1.51	--	--								
138	4	-0.40	--	--	1.35	--	--								
140	0	3.24	--	--	1.65	--	--								
143	4	-0.40	--	--	1.35	--	--								
145	3	0.81	--	--	--	1.45	--								
146	4	0.21	--	--	--	1.40	--								
151	2	1.42	--	--	--	--	1.50								
154	4	-0.16	--	--	1.37	--	--								
155	4	0.23	--	--	--	1.40	--								
158	3	0.93	--	--	--	1.46	--								
180	3	0.57	--	--	1.43	--	--								
183	2	1.42	--	--	--	--	1.50								
185	4	0.08	--	--	--	1.39	--								
190	1	1.54	--	--	--	1.51	--								
193	0	-3.68	--	--	1.08	--	--								
198	4	-0.21	--	--	1.37	--	--								
200	4	0.21	--	--	--	1.40	--								
204	3	0.57	--	--	1.43	--	--								
205	4	0.45	--	--	1.42	--	--								
209	3	0.55	--	--	--	--	1.43								

Table 14. Statistical summary of reported data for standard reference water sample N-64 (nutrient constituents)--Continued

Analyte : NH₃ + Organic N as N (Ammonia + organic nitrogen as nitrogen) Concentration Unit : mg/L




ANALYTICAL METHOD CODE



SUMMARY									
N =	2	3	31	10	1	0. Other	MPV =	1.54	
Minimum =	1.43	0.89	1.21	1.30	0.72	20. Titrate: colorimetric	F-pseudostigma =	0.13	
Maximum =	1.60	2.30	2.66	1.65		22. Colorimetric	N =	47	
Median =			1.54	1.54		22p. Color: phenate	Uh =	1.65	
F-pseudostigma =			0.14	0.10		40. Ion selective electrode	Lh =	1.47	

Lab	Rating	Z-value	0	20	22	22p	40
1	4	-0.23	--	--	1.51	--	--
3	4	-0.15	--	1.52	--	--	--
10	4	-0.39	--	--	1.49	--	--
11	1	2.00	--	--	1.80	--	--
12	0	6.63	--	--	2.40	--	--
18	3	0.85	--	--	--	1.65	--
23	4	0.23	--	--	1.57	--	--
36	4	0.46	1.60	--	--	--	--
38	3	-1.00	--	--	1.41	--	--
46	3	0.93	--	--	1.66	--	--
59	4	-0.31	--	--	1.50	--	--
70	0	2.78	--	--	1.90	--	--
87	3	0.77	--	--	1.64	--	--
89	4	-0.31	--	--	--	1.50	--
90	3	-0.62	--	--	1.46	--	--
96	4	-0.19	--	--	--	1.52	--
102	4	-0.46	--	--	1.48	--	--
105	4	0.15	--	--	1.56	--	--
113	3	-0.81	--	--	1.44	--	--
127	3	0.77	--	--	--	1.64	--
129	0	4.23	--	--	2.09	--	--
134	1	-1.70	--	--	1.32	--	--
138	4	0.46	--	--	1.60	--	--
140	2	1.08	--	--	1.68	--	--
143	4	-0.39	--	--	1.49	--	--
145	3	0.85	--	--	--	1.65	--
146	4	0.46	--	--	--	1.60	--
154	4	0.00	--	--	1.54	--	--
155	4	0.20	--	--	--	1.57	--
158	1	-1.83	--	--	--	1.30	--
180	4	0.15	--	--	1.56	--	--
193	4	0.00	--	--	1.54	--	--
204	0	2.54	--	--	1.87	--	--
209	3	-0.89	1.43	--	--	--	--
212	3	-0.77	--	--	1.44	--	--
213	0	5.86	--	2.30	--	--	--
215	4	-0.31	--	--	1.50	--	--
224	4	-0.27	--	--	--	1.51	--
227	0	6.27	--	--	2.35	--	--
247	3	-0.69	--	--	--	1.45	--
282	0	-2.54	--	--	1.21	--	--
285	0	-6.36	--	--	--	--	0.72
306	0	-5.00	--	0.89	--	--	--
313	4	0.15	--	--	1.56	--	--
316	4	-0.36	--	--	1.49	--	--
319	3	-0.62	--	--	1.46	--	--
320	0	8.63	--	--	2.66	--	--


Table 14. Statistical summary of reported data for standard reference water sample N-64 (low ionic strength constituents)--Continu
 Analyte : NO₃ as N (Nitrate as nitrogen) Concentration Unit : mg/L




SUMMARY													
N =	6	22	30	13	0. Other	MPV =	1.26						
Minimum =	1.17	1.16	1.16	0.62	7. Ion chromatography	F-pseudosigma =	0.04						
Maximum =	1.27	1.65	1.38	1.36	22. Colorimetric	Rating Criterion =	0.06						
Median =		1.27	1.26	1.29	22cd. Cd diazotization	N =	71						
F-pseudosigma =		0.05	0.02	0.06		Uh =	1.29						
						Lh =	1.24						
Lab	Rating	Z-value	0	7	22	22cd	Lab	Rating	Z-value	0	7	22	22cd
1	4	-0.16	--	--	1.25	--	208	2	0.79	--	1.31	--	--
3	2	0.79	--	--	--	1.31	209	4	-0.24	--	1.25	--	--
10	3	-0.32	--	--	1.24	--	212	0	-1.59	--	1.16	--	--
11	3	0.48	--	--	1.29	--	215	4	-0.16	--	--	1.25	--
12	3	-0.32	--	--	1.24	--	224	1	0.92	--	1.32	--	--
13	4	0.00	--	1.26	--	--	227	2	-0.60	--	1.22	--	--
18	3	-0.48	--	--	--	1.23	234	1	0.95	--	1.32	--	--
23	4	0.00	--	--	1.26	--	243	4	0.16	--	--	--	1.27
25	3	-0.32	--	1.24	--	--	247	3	-0.32	--	1.24	--	--
26	0	6.19	--	1.65	--	--	253	0	1.94	--	--	1.38	--
36	3	-0.32	1.24	--	--	--	255	4	0.16	--	--	1.27	--
38	4	0.02	--	--	1.26	--	282	4	0.16	--	--	1.27	--
42	3	0.41	--	1.29	--	--	285	0	-1.27	1.18	--	--	--
46	4	0.00	--	--	1.26	--	287	0	-1.43	1.17	--	--	--
53	3	-0.32	1.24	--	--	--	291	4	0.16	1.27	--	--	--
59	0	1.27	--	--	--	1.34	292	2	0.63	--	1.30	--	--
64	2	-0.79	--	--	1.21	--	306	0	-10.10	--	--	--	0.62
69	3	-0.32	--	--	1.24	--	313	4	0.16	--	--	1.27	--
70	4	0.00	--	--	1.26	--	316	4	0.07	--	--	1.26	--
76	0	-1.60	--	1.16	--	--	319	3	-0.48	1.23	--	--	--
85	1	0.95	--	--	1.32	--	320	3	0.32	--	--	1.28	--
87	4	0.00	--	--	1.26	--							
89	3	0.48	--	--	--	1.29							
90	4	-0.16	--	--	1.25	--							
93	3	0.48	--	1.29	--	--							
96	2	0.73	--	--	--	1.31							
102	2	-0.79	--	1.21	--	--							
105	3	-0.32	--	--	1.24	--							
107	3	0.41	--	--	--	1.29							
113	4	0.10	--	--	1.27	--							
114	0	1.59	--	--	--	1.36							
127	3	-0.32	--	1.24	--	--							
129	2	0.76	--	1.31	--	--							
134	4	0.16	--	--	1.27	--							
138	4	-0.16	--	1.25	--	--							
140	1	-0.89	--	--	1.20	--							
143	0	-1.59	--	--	1.16	--							
145	4	-0.16	--	--	--	1.25							
146	2	0.63	--	--	--	1.30							
151	2	0.84	--	1.31	--	--							
154	3	-0.32	--	--	1.24	--							
155	0	-2.73	--	--	--	1.09							
158	1	0.95	--	--	1.32	--							
180	2	0.63	--	1.30	--	--							
185	4	0.11	--	1.27	--	--							
193	3	0.32	--	1.28	--	--							
198	4	0.00	--	--	1.26	--							
200	2	-0.63	--	--	--	1.22							
204	2	0.79	--	--	1.31	--							
205	4	0.03	--	--	1.26	--							

Table 14. Statistical summary of reported data for standard reference water sample N-64 (nutrient constituents)--Continued

Analyte : total P as P (total Phosphorus as phosphorus) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY						
N =	1	1	59	0. Other	MPV =	0.883
Minimum =	1.610	0.858	0.770	4. ICP	F-pseudostandard deviation =	0.041
Maximum =			2.000	22m. Color:phosphomolybdate	Rating Criterion =	0.044
Median =			0.883		N =	61
F-pseudostandard deviation =			0.042		Uh =	0.911
					Lh =	0.856

Lab	Rating	Z-value	0	4	22m	Lab	Rating	Z-value	0	4	22m
1	4	-0.48	--	--	0.862	247	2	-1.38	--	--	0.822
3	4	0.02	--	--	0.884	253	4	0.00	--	--	0.883
10	4	-0.14	--	--	0.877	255	3	0.84	--	--	0.920
11	4	0.39	--	--	0.900	282	3	-0.93	--	--	0.842
12	0	-2.17	--	--	0.787	285	0	2.36	--	--	0.987
13	4	0.14	--	--	0.889	292	3	0.84	--	--	0.920
18	4	0.09	--	--	0.887	306	3	0.63	--	--	0.911
23	3	-0.52	--	--	0.860	313	4	0.36	--	--	0.899
25	1	-1.65	--	--	0.810	316	3	0.67	--	--	0.913
36	0	16.47	1.610	--	--	319	0	4.24	--	--	1.070
38	3	0.84	--	--	0.920	320	0	25.30	--	--	2.000
46	4	0.25	--	--	0.894						
59	1	-1.88	--	--	0.800						
64	4	-0.18	--	--	0.875						
70	4	0.05	--	--	0.885						
85	2	1.45	--	--	0.947						
87	0	2.38	--	--	0.988						
89	4	0.14	--	--	0.889						
93	4	-0.07	--	--	0.880						
96	3	-0.72	--	--	0.851						
102	2	-1.18	--	--	0.831						
105	0	5.37	--	--	1.120						
107	3	-0.61	--	--	0.856						
113	4	-0.27	--	--	0.871						
114	3	-0.97	--	--	0.840						
127	3	0.84	--	--	0.920						
129	4	-0.14	--	--	0.877						
134	3	0.63	--	--	0.911						
138	4	0.07	--	--	0.886						
140	0	-2.56	--	--	0.770						
143	4	-0.14	--	--	0.877						
145	3	0.61	--	--	0.910						
146	3	-0.52	--	--	0.860						
154	0	2.22	--	--	0.981						
155	4	0.48	--	--	0.904						
158	4	-0.11	--	--	0.878						
180	4	0.00	--	--	0.883						
183	2	-1.20	--	--	0.830						
185	3	0.54	--	--	0.907						
193	1	-1.81	--	--	0.803						
198	4	-0.02	--	--	0.882						
200	3	0.61	--	--	0.910						
204	4	-0.27	--	--	0.871						
212	1	-1.88	--	--	0.800						
213	3	-0.75	--	--	0.850						
215	3	0.84	--	--	0.920						
224	4	-0.18	--	--	0.875						
227	3	-0.57	--	0.858	--						
234	2	-1.43	--	--	0.820						
243	2	-1.20	--	--	0.830						


Table 14. Statistical summary of reported data for standard reference water sample N-64 (nutrient constituents)--Continued

Analyte : PO₄ as P (Orthophosphate as phosphorus)

Concentration Unit : mg/L



ANALYTICAL METHOD CODE



SUMMARY					
N =	2	9	54	0. Other	MPV = 0.860
Minimum =	0.830	0.670	0.700	7. Ion chromatography	F-pseudosigma = 0.034
Maximum =	0.840	2.720	1.230	22m. Color:phosphomolybdate	Rating Criterion = 0.043
Median =		0.846	0.864		N = 65
F-pseudosigma =		0.035	0.032		Uh = 0.891
					Lh = 0.845

Lab	Rating	Z-value	0	7	22m	Lab	Rating	Z-value	0	7	22m
1	3	0.72	--	--	0.891	213	4	-0.23	--	--	0.850
3	3	-0.79	--	--	0.826	215	2	1.40	--	--	0.920
10	4	0.16	--	--	0.867	224	3	0.84	--	--	0.896
11	3	0.93	--	--	0.900	227	4	0.14	--	--	0.866
12	4	-0.47	--	--	0.840	234	0	-4.42	--	0.670	--
13	2	-1.02	--	0.816	--	247	4	-0.33	--	0.846	--
18	4	0.21	--	--	0.869	253	4	-0.12	--	--	0.855
23	4	-0.23	--	--	0.850	255	4	-0.09	--	--	0.856
25	2	1.14	--	--	0.909	282	3	-0.67	--	--	0.831
26	3	-0.70	--	0.830	--	287	3	-0.70	0.830	--	--
36	4	-0.47	0.840	--	--	292	4	-0.35	--	--	0.845
38	3	-0.56	--	--	0.836	313	4	0.07	--	--	0.863
42	4	0.26	--	0.871	--	316	4	0.10	--	--	0.865
46	2	1.02	--	--	0.904	319	0	3.26	--	--	1.000
53	4	-0.09	--	--	0.856	320	4	0.37	--	--	0.876
59	4	-0.47	--	--	0.840						
64	4	-0.33	--	--	0.846						
70	3	0.95	--	--	0.901						
76	2	-1.42	--	0.799	--						
83	4	0.00	--	--	0.860						
85	4	0.47	--	--	0.880						
87	4	0.49	--	--	0.881						
89	3	0.56	--	--	0.884						
93	0	4.19	--	--	1.040						
96	4	0.21	--	--	0.869						
102	0	-3.33	--	--	0.717						
105	0	8.60	--	--	1.230						
107	4	-0.28	--	--	0.848						
113	4	-0.12	--	--	0.855						
127	4	0.07	--	0.863	--						
129	4	-0.12	--	--	0.855						
134	4	-0.02	--	--	0.859						
138	2	-1.05	--	--	0.815						
140	4	-0.23	--	--	0.850						
143	4	0.02	--	--	0.861						
145	1	1.51	--	--	0.925						
146	0	-3.72	--	--	0.700						
151	0	43.26	--	2.720	--						
154	3	0.72	--	--	0.891						
155	4	0.00	--	--	0.860						
158	3	0.77	--	--	0.893						
180	3	0.65	--	--	0.888						
183	3	-0.70	--	--	0.830						
185	2	1.21	--	--	0.912						
190	4	0.42	--	--	0.878						
198	1	-1.81	--	--	0.782						
200	4	-0.23	--	--	0.850						
204	2	1.02	--	--	0.904						
208	4	-0.23	--	0.850	--						
212	2	1.02	--	--	0.904						

Table 15. *Statistical summary of reported data for standard reference sample P-33 (low ionic strength constituents)*

Definition of analytical methods, abbreviations, and symbols

Analytical methods

0. Other/Not reported	=	
1. AA: direct, air	=	atomic absorption: direct, air
2. AA: direct, N ₂ O	=	atomic absorption: direct, nitrous oxide
3. AA: graphite furnace	=	atomic absorption: graphite furnace
4. ICP	=	inductively coupled plasma
5. DCP	=	direct current plasma
6. ICP/MS	=	inductively coupled plasma / mass spectrometry
7. IC	=	ion chromatography
12. Flame emission	=	flame emission
20. Titrate: color	=	titration: colorimetric (color reagent specified)
21. Titrate: electro	=	titration: electrometric
22. Color:	=	colorimetric (color reagent specified)
40. Ion electrode	=	ion selective electrode
41. Electro	=	electrometric: (type meter specified)
50. Gravimetric	=	gravimetric: (precipitate specified)
51. Turbidimetric	=	turbidimetric: (precipitate specified)

Abbreviations and figure symbols

N	=	number of analyses--(excluding less than values)
MPV	=	most probable value ---
F-pseudostigma	=	nonparametric statistic deviation
Uh	=	upper hinge value
Lh	=	lower hinge value
Uwl	=	upper warning limit - - - -
Lwl	=	lower warning limit - - - -
Ucl	=	upper warning limit
Lcl	=	lower warning limit - - - -
mg/L	=	milligrams per liter
μS/cm	=	microsiemens per centimeter at 25 ^o C
Lab	=	laboratory code number
NR	=	not rated, less than value reported or insufficient data
<	=	less than
--	=	not reported

<u>Constituent</u>	<u>page</u>
Acid Acidity as CaCO ₃	94
Ca Calcium	95
Cl Chloride	96
F Fluoride	97
K Potassium	98
Mg Magnesium	99
Na Sodium	100
pH	101
PO ₄ as P Orthophosphate as Phosphorus	102
SO ₄ Sulfate	103
Sp Cond Specific Conductance	104

Table 15. Statistical summary of reported data for standard reference water sample P-33 (low ionic strength constituents)--Continu

Analyte: Acidity (as CaCO₃) Concentration Unit : mg/L


INSUFFICIENT DATA

SUMMARY				
N =	2	12	20. Titrate: colorimetric	MPV = Insufficient data
Minimum =	2.00	0.31	21. Titrate: electrometric	N = 14
Maximum =	9.85	61.00		
Median =		3.03		
F-pseudosigma =		6.00		


Lab	Rating	Z-value	20	21
1	NR	--	--	2.50
3	NR	--	--	< 10
25	NR	--	--	3.00
38	NR	--	--	0.31
81	NR	--	--	1.40
89	NR	--	--	3.53
127	NR	--	--	4.10
141	NR	--	--	61.00
146	NR	--	--	< 5.79
215	NR	--	2.00	--
228	NR	--	--	20.70
237	NR	--	--	1.01
273	NR	--	--	3.05
274	NR	--	--	15.99
282	NR	--	9.85	--
287	NR	--	--	1.00

Table 15. Statistical summary of reported data for standard reference water sample P-33 (low ionic strength constituents)--Continu

Analyte: Ca (Calcium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	3	6	27	5	3	3	0. Other	MPV =	0.322
Minimum =	0.320	0.326	0.230	0.240	0.170	0.468	1. AA: direct, air	F-pseudosigma =	0.043
Maximum =	0.346	0.403	0.400	0.800	0.292	1.980	4. ICP	N =	47
Median =			0.321				7. Ion chromatography	Uh =	0.356
F-pseudosigma =			0.033				12. Flame emission	Lh =	0.297
							20. Titrate: colorimetric		


Lab	Rating	Z-value	0	1	4	7	12	20
1	4	-0.07	--	--	0.319	--	--	--
2	1	-1.84	--	--	--	0.242	--	--
3	3	0.74	--	--	0.354	--	--	--
11	1	1.80	--	--	0.400	--	--	--
33	4	-0.28	--	--	--	0.310	--	--
38	4	-0.05	0.320	--	--	--	--	--
42	4	-0.46	--	--	0.302	--	--	--
46	4	0.00	--	--	0.322	--	--	--
48	3	0.55	0.346	--	--	--	--	--
64	1	1.57	--	0.390	--	--	--	--
81	4	-0.02	--	--	0.321	--	--	--
86	3	-0.95	--	--	0.281	--	--	--
89	0	-3.51	--	--	--	0.170	--	--
93	4	-0.28	--	--	0.310	--	--	--
102	0	-2.12	--	--	0.230	--	--	--
113	3	0.92	--	--	0.362	--	--	--
127	4	0.35	--	--	0.337	--	--	--
134	4	-0.05	--	--	0.320	--	--	--
138	4	0.42	--	--	0.340	--	--	--
140	4	0.35	--	0.337	--	--	--	--
141	3	0.81	--	--	0.357	--	--	--
145	3	0.65	--	--	0.350	--	--	--
146	NR	--	--	<	0.313	--	--	--
155	0	3.36	--	--	--	--	0.468	--
180	3	0.83	--	--	0.358	--	--	--
185	4	0.09	--	0.326	--	--	--	--
190	0	11.02	--	--	--	0.800	--	--
191	4	-0.05	0.320	--	--	--	--	--
193	3	-0.74	--	--	0.290	--	--	--
203	2	-1.43	--	--	0.260	--	--	--
209	1	-1.73	--	--	0.247	--	--	--
215	0	-2.12	--	--	0.230	--	--	--
227	4	0.00	--	--	0.322	--	--	--
228	1	-1.89	--	--	--	0.240	--	--
237	4	-0.05	--	--	0.320	--	--	--
238	4	-0.05	--	--	--	0.320	--	--
241	3	-0.69	--	--	--	--	0.292	--
247	4	0.30	--	--	0.335	--	--	--
255	4	0.28	--	--	0.334	--	--	--
265	4	0.42	--	--	0.340	--	--	--
268	1	1.80	--	0.400	--	--	--	--
273	3	-0.74	--	--	0.290	--	--	--
274	0	38.23	--	--	--	--	1.980	--
279	0	14.48	--	--	--	--	0.950	--
282	4	-0.28	--	--	0.310	--	--	--
287	1	1.87	--	0.403	--	--	--	--
301	3	-0.71	--	--	--	--	0.291	--
321	2	1.11	--	0.370	--	--	--	--

Table 15. Statistical summary of reported data for standard reference water sample P-33 (low ionic strength constituents)--Continu

Analyte: Cl (Chloride) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY

N =	1	37	4	2	6	0. Other	MPV = 2.38
Minimum =	2.00	1.63	2.00	2.46	1.69	7. Ion chromatography	F-pseudosigma = 0.18
Maximum =		4.77	19.68	2.60	5.04	20. Titrate: colorimetric	N = 50
Median =		2.38				21. Titrate: electrometric	Uh = 2.46
F-pseudosigma =		0.13				22. Colorimetric	Lh = 2.22

Lab	Rating	Z-value	0	7	20	21	22
1	3	0.73	--	2.51	--	--	--
2	0	-3.31	--	1.79	--	--	--
3	0	-3.88	--	--	--	--	1.69
11	0	-2.14	--	--	2.00	--	--
25	4	0.11	--	2.40	--	--	--
33	4	0.06	--	2.39	--	--	--
42	4	-0.39	--	2.31	--	--	--
45	4	0.45	--	2.46	--	--	--
46	4	-0.17	--	--	--	--	2.35
48	0	-2.14	2.00	--	--	--	--
59	4	0.00	--	2.38	--	--	--
64	4	0.28	--	2.43	--	--	--
81	4	0.45	--	--	--	2.46	--
86	3	-0.51	--	2.29	--	--	--
89	4	-0.11	--	2.36	--	--	--
93	4	0.00	--	2.38	--	--	--
102	0	13.43	--	4.77	--	--	--
113	4	0.45	--	2.46	--	--	--
127	2	-1.12	--	2.18	--	--	--
134	4	-0.45	--	2.30	--	--	--
138	4	-0.11	--	2.36	--	--	--
140	3	-0.56	--	--	--	--	2.28
141	4	-0.45	--	2.30	--	--	--
143	2	1.18	--	--	--	--	2.59
145	4	0.00	--	2.38	--	--	--
146	2	-1.41	--	--	--	--	2.13
180	4	0.28	--	2.43	--	--	--
183	0	14.95	--	--	--	--	5.04
185	4	0.20	--	2.42	--	--	--
190	0	-3.20	--	1.81	--	--	--
191	4	-0.22	--	2.34	--	--	--
196	3	-0.96	--	2.21	--	--	--
203	2	1.24	--	--	--	2.60	--
208	2	1.24	--	2.60	--	--	--
209	4	-0.45	--	2.30	--	--	--
215	1	1.80	--	--	2.70	--	--
227	0	2.19	--	2.77	--	--	--
228	4	0.34	--	2.44	--	--	--
237	2	1.12	--	2.58	--	--	--
238	2	1.12	--	2.58	--	--	--
241	1	-1.69	--	2.08	--	--	--
247	3	-0.90	--	2.22	--	--	--
265	1	-1.57	--	2.10	--	--	--
268	4	0.06	--	2.39	--	--	--
273	3	-0.98	--	2.21	--	--	--
274	0	97.24	--	--	19.68	--	--
279	4	-0.45	--	--	2.30	--	--
282	0	-4.22	--	1.63	--	--	--
287	0	7.98	--	3.80	--	--	--
321	4	0.11	--	2.40	--	--	--

Table 15. Statistical summary of reported data for standard reference water sample P-33 (low ionic strength constituents)--Continu

Analyte: F (Fluoride) Concentration Unit : mg/L



SUMMARY

N =	18	1	14
Minimum =	0.070	0.002	0.075
Maximum =	0.820		0.150
Median =	0.100		0.115
F-pseudostandard deviation =	0.027		0.028


7. Ion chromatography
 22. Colorimetric
 40. Ion selective electrode

MPV =	0.105
F-pseudostandard deviation =	0.039
N =	33
Uh =	0.137
Lh =	0.085


Lab	Rating	Z-value	7	22	40
1	NR	--	--	--	< 0.1
2	4	0.18	0.112	--	--
3	3	0.83	--	--	0.137
11	3	-0.65	--	--	0.080
25	4	-0.39	0.090	--	--
33	4	-0.13	0.100	--	--
42	4	0.13	0.110	--	--
45	3	-0.52	0.085	--	--
46	3	-0.78	--	--	0.075
59	2	1.17	--	--	0.150
81	4	0.00	--	--	0.105
83	3	0.91	--	--	0.140
86	4	-0.13	0.100	--	--
89	4	-0.13	--	--	0.100
93	3	-0.91	0.070	--	--
102	0	18.55	0.820	--	--
113	4	0.23	--	--	0.114
127	4	-0.28	0.094	--	--
134	4	0.26	--	--	0.115
138	NR	--	--	--	< 0.10
140	3	0.86	--	--	0.138
141	4	0.13	--	--	0.110
145	3	-0.91	0.070	--	--
146	3	0.65	--	--	0.130
180	1	1.71	0.171	--	--
183	3	0.91	--	--	0.140
190	4	-0.05	0.103	--	--
191	3	-0.65	0.080	--	--
196	3	-0.80	0.074	--	--
241	4	-0.13	--	--	0.100
247	3	-0.65	0.080	--	--
255	NR	--	--	--	< 0.5
265	NR	--	< 0.1	--	--
273	4	0.36	0.119	--	--
274	0	-2.67	--	0.002	--
282	2	1.43	0.160	--	--
287	0	3.50	0.240	--	--

Table 15. Statistical summary of reported data for standard reference water sample P-33 (low ionic strength constituents)--Continu

Analyte: K (Potassium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




N =	11	13	1	2	5	6	1. AA: direct, air	MPV = 0.110
Minimum =	0.060	0.040	0.100	0.080	0.040	0.100	4. ICP	F-pseudostigma = 0.015
Maximum =	0.280	0.532	0.100	0.132	0.120	0.130	5. DCP	N = 38
Median =	0.109	0.110					6. ICP/MS	Uh = 0.120
F-pseudostigma =	0.014	0.035					7. Ion chromatography	Lh = 0.100
							12. Flame emission	


Lab	Rating	Z-value	1	4	5	6	7	12
1	4	0.47	0.117	--	--	--	--	--
2	4	0.13	--	--	--	0.112	--	--
3	0	28.46	--	0.532	--	--	--	--
11	0	-4.00	--	< 0.05	--	--	--	--
33	3	-0.67	--	--	0.100	--	--	--
38	3	0.67	0.120	--	--	--	--	--
42	4	-0.20	--	0.107	--	--	--	--
45	0	11.47	0.280	--	--	--	--	--
46	0	3.17	--	0.157	--	--	--	--
48	2	1.48	--	--	--	0.132	--	--
64	0	-3.37	0.060	--	--	--	--	--
81	1	-1.89	--	0.082	--	--	--	--
89	3	-0.67	--	--	--	--	0.100	--
93	3	0.67	--	0.120	--	--	--	--
102	0	-3.37	--	0.060	--	--	--	--
113	0	2.36	--	0.145	--	--	--	--
127	4	-0.27	0.106	--	--	--	--	--
134	4	0.07	0.111	--	--	--	--	--
138	4	0.00	--	0.110	--	--	--	--
140	2	-1.28	0.091	--	--	--	--	--
141	0	4.52	--	0.177	--	--	--	--
145	0	-4.72	--	0.040	--	--	--	--
146	NR	--	--	< 1	--	--	--	--
180	NR	--	--	< 1.63	--	--	--	--
185	4	-0.07	0.109	--	--	--	--	--
190	3	0.67	--	--	--	0.120	--	--
191	1	-2.02	--	--	0.080	--	--	--
193	3	-0.67	0.100	--	--	--	--	--
203	3	0.67	--	--	--	--	0.120	--
209	3	-0.81	--	0.098	--	--	--	--
228	0	-4.72	--	--	--	0.040	--	--
237	4	0.00	--	--	--	0.110	--	--
238	4	0.00	--	--	--	0.110	--	--
241	2	1.21	--	--	--	--	0.128	--
247	NR	--	--	< 0.20	--	--	--	--
265	4	0.00	--	0.110	--	--	--	--
268	0	2.70	0.150	--	--	--	--	--
273	4	0.00	--	0.110	--	--	--	--
274	3	-0.67	--	--	--	--	0.100	--
279	2	1.35	--	--	--	--	0.130	--
287	3	-0.81	0.098	--	--	--	--	--
301	4	0.00	--	--	--	--	0.110	--

Table 15. Statistical summary of reported data for standard reference water sample P-33 (low ionic strength constituents)--Continu

Analyte: Mg (Magnesium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	3	7	23	2	4	3	0. Other	MPV =	0.100
Minimum =	0.089	0.098	0.068	0.100	0.050	0.180	1. AA: direct, air	F-pseudosigma =	0.015
Maximum =	0.126	0.110	0.140	0.116	0.190	0.950	4. ICP	N =	42
Median =		0.100	0.100				6. ICP/MS	Uh =	0.115
F-pseudosigma =		0.008	0.012				7. Ion chromatography	Lh =	0.095
							20. Titrate: colorimetric		


Lab	Rating	Z-value	0	1	4	6	7	20
1	4	-0.07	--	--	0.099	--	--	--
2	2	1.28	--	--	--	--	0.119	--
3	0	2.29	--	--	0.134	--	--	--
11	4	0.00	--	--	0.100	--	--	--
33	3	-0.67	0.090	--	--	--	--	--
38	4	-0.07	--	0.099	--	--	--	--
42	0	-2.14	--	--	0.068	--	--	--
46	4	0.13	--	--	0.102	--	--	--
48	2	1.08	--	--	--	0.116	--	--
64	3	0.67	--	0.110	--	--	--	--
81	4	-0.40	--	--	0.094	--	--	--
86	3	0.74	--	--	0.111	--	--	--
89	4	0.00	--	0.100	--	--	--	--
93	2	-1.35	--	--	0.080	--	--	--
102	1	2.02	--	--	0.130	--	--	--
113	3	0.74	--	--	0.111	--	--	--
127	2	1.01	--	--	0.115	--	--	--
134	4	-0.03	--	--	0.100	--	--	--
138	4	-0.34	--	--	0.095	--	--	--
140	4	0.00	--	0.100	--	--	--	--
141	0	2.70	--	--	0.140	--	--	--
145	NR	--	--	--	< 0.19	--	--	--
146	NR	--	--	--	< 0.5	--	--	--
155	0	7.49	--	--	--	--	0.211	--
180	2	1.01	--	--	0.115	--	--	--
185	4	-0.13	--	0.098	--	--	--	--
190	0	6.07	--	--	--	--	0.190	--
191	4	0.00	--	--	--	0.100	--	--
193	4	-0.34	--	--	0.095	--	--	--
203	NR	--	--	--	< 0.10	--	--	--
209	4	-0.40	--	--	0.094	--	--	--
227	4	-0.27	--	--	0.096	--	--	--
228	0	-3.37	--	--	--	--	0.050	--
237	3	-0.67	--	--	0.090	--	--	--
238	4	0.00	--	--	--	--	0.100	--
241	1	1.75	0.126	--	--	--	--	--
247	NR	--	--	--	< 0.2	--	--	--
255	4	-0.13	--	--	0.098	--	--	--
265	4	0.00	--	--	0.100	--	--	--
268	3	0.67	--	0.110	--	--	--	--
273	4	0.00	--	--	0.100	--	--	--
274	0	57.33	--	--	--	--	0.950	--
279	0	5.40	--	--	--	--	0.180	--
282	4	-0.34	--	--	0.095	--	--	--
287	3	0.67	--	0.110	--	--	--	--
301	3	-0.74	0.089	--	--	--	--	--

Table 15. Statistical summary of reported data for standard reference water sample P-33 (low ionic strength constituents)--Continu

Analyte: Na (Sodium) Concentration Unit : mg/L



ANALYTICAL METHOD CODE




SUMMARY									
N =	8	24	1	2	5	6	1. AA: direct, air	MPV =	1.42
Minimum =	1.31	0.14	1.37	1.46	1.39	1.16	4. ICP	F-pseudosigma =	0.10
Maximum =	4.09	50.06		1.46	1.93	4.94	5. DCP	N =	46
Median =	1.43	1.40					6. ICP/MS	Uh =	1.48
F-pseudosigma =	0.11	0.12					7. Ion chromatography	Lh =	1.33
							12. Flame emission		

Lab	Rating	Z-value	1	4	5	6	7	12
1	4	-0.43	--	1.37	--	--	--	--
2	0	4.96	--	--	--	--	1.93	--
3	2	1.10	--	1.53	--	--	--	--
11	0	-9.62	--	0.41	--	--	--	--
33	4	-0.43	--	--	1.37	--	--	--
38	3	-1.00	1.31	--	--	--	--	--
42	1	-1.96	--	1.21	--	--	--	--
45	0	25.59	4.09	--	--	--	--	--
46	4	-0.14	--	1.40	--	--	--	--
48	4	0.43	--	--	--	1.46	--	--
64	3	-1.00	1.31	--	--	--	--	--
81	3	0.81	--	1.50	--	--	--	--
86	3	-0.62	--	1.35	--	--	--	--
89	0	-2.44	--	--	--	--	--	1.16
93	4	-0.24	--	1.39	--	--	--	--
102	0	465.41	--	50.06	--	--	--	--
113	1	1.77	--	1.60	--	--	--	--
127	4	0.05	--	1.42	--	--	--	--
134	4	0.33	1.45	--	--	--	--	--
138	4	-0.14	--	1.40	--	--	--	--
140	3	0.57	1.48	--	--	--	--	--
141	2	1.48	--	1.57	--	--	--	--
145	4	0.05	--	1.42	--	--	--	--
146	4	0.33	--	1.45	--	--	--	--
180	4	0.43	--	1.46	--	--	--	--
185	3	0.54	--	--	--	--	--	1.47
190	3	0.53	--	--	--	--	1.47	--
191	4	0.43	--	--	--	1.46	--	--
193	2	-1.29	--	1.28	--	--	--	--
203	1	-1.87	--	1.22	--	--	--	--
209	2	-1.16	--	1.29	--	--	--	--
215	0	-5.31	--	0.86	--	--	--	--
228	2	1.29	--	--	--	--	1.55	--
237	4	-0.24	--	--	--	--	1.39	--
238	3	0.81	--	--	--	--	1.50	--
241	3	-0.53	--	--	--	--	--	1.36
247	4	0.14	--	1.43	--	--	--	--
265	4	-0.14	--	1.40	--	--	--	--
268	4	0.33	1.45	--	--	--	--	--
273	0	-12.20	--	0.14	--	--	--	--
274	0	33.73	--	--	--	--	--	4.94
279	2	1.20	--	--	--	--	--	1.54
282	4	-0.43	--	1.37	--	--	--	--
287	3	-0.91	1.32	--	--	--	--	--
301	3	-0.77	--	--	--	--	--	1.33
321	4	-0.05	1.41	--	--	--	--	--

Table 15. Statistical summary of reported data for standard reference water sample P-33 (low ionic strength constituents)--Continu

Analyte: pH



ANALYTICAL METHOD CODE

—○— 41

SUMMARY

N = 53
 Minimum = 3.58
 Maximum = 7.00
 Median = 4.67
 F-pseudostigma = 0.13


41. Direct reading

MPV = 4.67
 F-pseudostigma = 0.13
 Rating Criterion = 0.23
 N = 53
 Uh = 4.75
 Lh = 4.58

Lab	Rating	Z-value	41	Lab	Rating	Z-value	41
1	2	1.46	5.01	301	4	0.34	4.75
2	3	-0.72	4.50	312	4	0.00	4.67
3	2	-1.20	4.39	321	0	2.06	5.15
11	0	3.90	5.58				
12	0	-2.70	4.04				
25	3	0.73	4.84				
33	4	-0.13	4.64				
38	4	0.13	4.70				
45	3	-0.99	4.44				
46	3	0.99	4.90				
48	0	5.70	6.00				
59	4	0.13	4.70				
64	2	1.46	5.01				
81	4	-0.30	4.60				
86	4	0.00	4.67				
89	2	-1.20	4.39				
93	4	0.47	4.78				
107	4	0.00	4.67				
113	4	0.13	4.70				
127	4	-0.04	4.66				
134	4	0.00	4.67				
138	4	0.00	4.67				
140	4	0.09	4.69				
141	3	-0.73	4.50				
143	4	0.04	4.68				
145	3	-0.73	4.50				
146	4	-0.26	4.61				
155	4	-0.47	4.56				
180	4	0.13	4.70				
183	3	0.56	4.80				
185	1	-1.76	4.26				
190	4	-0.30	4.60				
203	4	-0.39	4.58				
204	4	0.00	4.67				
209	4	0.09	4.69				
215	0	2.27	5.20				
227	4	0.13	4.70				
228	4	-0.34	4.59				
237	3	-0.51	4.55				
241	2	-1.07	4.42				
243	4	-0.17	4.63				
244	4	0.00	4.67				
247	4	0.13	4.70				
265	0	9.98	7.00				
268	4	0.04	4.68				
273	0	7.00	6.31				
274	0	-4.67	3.58				
279	4	0.34	4.75				
282	2	-1.37	4.35				
287	4	0.34	4.75				

Table 15. Statistical summary of reported data for standard reference water sample P-33 (low ionic strength constituents)--Continu

Analyte: PO₄ as P (Orthophosphate as Phosphorus) Concentration Unit : mg/L



SUMMARY


N =	1	9	30
Minimum =	0.256	0.036	0.025
Maximum =		0.580	0.065
Median =		0.050	0.045
F-pseudosigma =		0.052	0.005

0. Other	MPV = 0.045
7. Ion chromatography	F-pseudosigma = 0.007
22m. Color:phosphomolybdate	N = 40
	Uh = 0.050
	Lh = 0.040

Lab	Rating	Z-value	0	7	22m
1	0	-2.70	--	--	0.025
3	4	0.13	--	--	0.046
11	3	-0.67	--	--	0.040
12	3	-0.81	--	--	0.039
25	3	-0.94	--	--	0.038
33	3	0.67	--	0.050	--
38	4	-0.40	--	--	0.042
42	NR	--	--	< 0.05	--
46	4	0.30	--	--	0.047
48	4	0.00	--	--	0.045
59	3	0.67	--	0.050	--
64	1	-1.75	--	--	0.032
81	2	-1.08	--	0.037	--
83	4	-0.40	--	--	0.042
89	4	-0.40	--	--	0.042
93	4	0.00	--	--	0.045
102	3	-0.86	--	--	0.039
113	4	0.27	--	--	0.047
127	2	1.04	--	--	0.053
134	3	-0.67	--	--	0.040
138	4	-0.28	--	--	0.043
140	3	0.67	--	--	0.050
143	4	0.40	--	--	0.048
145	0	-4.72	--	--	< 0.01
146	NR	--	--	< 0.021	--
155	4	-0.16	--	--	0.044
180	4	-0.13	--	--	0.044
183	3	-0.77	--	--	0.039
190	3	0.81	--	--	0.051
191	0	8.77	--	0.110	--
196	4	0.40	--	0.048	--
203	4	0.00	--	--	0.045
215	1	2.02	--	--	0.060
241	2	-1.21	--	0.036	--
247	3	-0.67	--	0.040	--
255	4	0.00	--	--	0.045
273	0	32.92	--	0.289	--
274	0	2.70	--	--	0.065
282	3	-0.54	--	--	0.041
287	0	72.17	--	0.580	--
301	0	28.46	0.256	--	--
312	4	0.00	--	--	0.045
321	4	0.40	--	--	0.048

Table 15. Statistical summary of reported data for standard reference water sample P-33 (low ionic strength constituents)--Continu

Analyte: SO₄ (Sulfate) Concentration Unit : mg/L



SUMMARY

N = 34 4 5
 Minimum = 1.11 1.20 0.32
 Maximum = 2.32 2.30 2.00
 Median = 1.41
 F-pseudostandard deviation = 0.13


7. Ion chromatography
 22mtb. Color: methyl thymol blue
 51. Turbidimetric

MPV = 1.41
 F-pseudostandard deviation = 0.17
 N = 43
 Uh = 1.54
 Lh = 1.30

Lab	Rating	Z-value	7	22mtb	51
1	4	0.11	1.43	--	--
2	4	0.03	1.42	--	--
3	NR	--	--	< 10	--
11	0	3.39	2.00	--	--
33	4	-0.11	1.39	--	--
42	0	5.22	2.32	--	--
45	0	2.47	1.84	--	--
48	0	-2.35	--	--	1.00
59	4	-0.11	1.39	--	--
64	4	-0.23	1.37	--	--
81	2	-1.21	--	1.20	--
86	2	1.15	1.61	--	--
89	4	0.17	1.44	--	--
93	4	-0.23	1.37	--	--
102	0	3.33	1.99	--	--
113	4	-0.11	1.39	--	--
127	4	0.11	1.43	--	--
134	4	0.23	1.45	--	--
138	3	-0.52	1.32	--	--
140	0	-2.35	--	--	1.00
141	3	-0.63	--	--	1.30
145	2	1.09	1.60	--	--
146	NR	--	--	--	< 5
180	4	0.00	1.41	--	--
185	4	-0.24	1.37	--	--
190	2	-1.15	--	1.21	--
191	3	0.80	1.55	--	--
196	3	0.63	1.52	--	--
203	NR	--	--	< 2.5	--
208	0	3.39	2.00	--	--
209	4	0.23	1.45	--	--
215	0	3.39	--	--	2.00
227	0	2.35	1.82	--	--
228	3	-0.75	1.28	--	--
238	4	-0.11	1.39	--	--
241	3	-0.69	1.29	--	--
247	3	-0.63	1.30	--	--
255	0	5.11	--	2.30	--
265	3	-0.63	1.30	--	--
268	4	0.00	1.41	--	--
273	1	-1.72	1.11	--	--
274	0	-6.26	--	--	0.32
279	4	0.40	--	1.48	--
282	3	-0.63	1.30	--	--
287	3	-0.80	1.27	--	--
321	4	0.29	1.46	--	--

Table 15. Statistical summary of reported data for standard reference water sample P-33 (low ionic strength constituents)--Continu

Analyte: Sp Cond (Specific Conductance) Concentration Unit : mS/cm



ANALYTICAL METHOD CODE

○— 41

SUMMARY

Minimum = 0.1
 Maximum = 64.5
 Median = 20.8
 F-pseudosigma = 2.4

41. Direct reading

MPV = 20.8
 F-pseudosigma = 2.4
 N = 49
 Uh = 21.4
 Lh = 18.1

Lab	Rating	Z-value	41
1	2	-1.10	18.1
3	4	0.00	20.8
11	4	-0.45	19.7
12	3	-0.69	19.1
25	0	2.94	28.0
33	4	-0.04	20.7
38	4	0.33	21.6
45	4	0.04	20.9
46	0	16.43	61.0
48	4	0.08	21.0
59	4	0.08	21.0
64	1	-1.51	17.1
81	2	-1.47	17.2
86	2	-1.14	18.0
89	1	1.72	25.0
93	4	-0.08	20.6
102	4	0.49	22.0
107	4	0.00	20.8
113	3	-0.74	19.0
127	4	0.49	22.0
134	4	0.09	21.0
138	4	-0.08	20.6
140	4	0.08	21.0
141	1	-1.96	16.0
143	4	-0.33	20.0
145	0	-8.25	< 1
146	2	1.31	24.0
155	4	0.00	20.8
180	2	-1.14	18.0
183	2	-1.14	18.0
185	0	17.86	64.5
190	4	0.08	21.0
193	2	-1.43	17.3
203	4	-0.45	19.7
204	4	0.25	21.4
215	1	-1.59	16.9
227	3	-0.53	19.5
228	4	0.29	21.5
241	0	-4.01	11.0
243	4	0.00	20.8
244	4	0.00	20.8
247	4	0.08	21.0
268	2	-1.14	18.0
273	3	0.78	22.7
274	4	-0.37	19.9
279	3	-0.78	18.9
282	3	0.65	22.4
301	0	-8.48	0.1
312	4	0.37	21.7
321	1	-1.55	17.0

Table 16. *Statistical summary of reported data for standard reference sample Hg-29 (mercury)*

Definition of analytical methods, abbreviations, and symbols

Analytical methods

- 0. Other/Not reported
 - 4. ICP = inductively coupled plasma
 - 6. ICP/MS = inductively coupled plasma / mass spectrometry
 - 8. AA: cold vapor = atomic absorption: cold vapor
 - 9. Atomic fluorescence
-


Abbreviations and figure symbols

- N = number of analyses--(excluding less than values)
 - MPV = most probable value ---
 - F-pseudostigma = nonparametric statistic deviation
 - Uh = upper hinge value
 - Lh = lower hinge value
 - Uwl = upper warning limit ---
 - Lwl = lower warning limit ---
 - Ucl = upper warning limit
 - Lcl = lower warning limit ---
 - µg/L = micrograms per liter
 - Lab = laboratory code number
 - NR = not rated, less than value reported or insufficient data
 - < = less than
 - = not reported
-


<u>Constituent</u>	<u>page</u>
Hg Mercury	106

Table 16. Statistical summary of reported data for standard reference water sample Hg-29 (mercury)--Continued

Analyte: Hg (Mercury) Concentration Unit : mg/L



ANALYTICAL METHOD CODE



SUMMARY						
N =	2	3	44	2	0. Other	MPV = 0.463
Minimum =	0.480	0.410	0.110	0.440	6. ICP/MS	F-pseudosigma = 0.052
Maximum =	0.490	0.550	0.680	0.453	8. AA: cold vapor	N = 51
Median =			0.465		9. Atomic fluorescence	Uh = 0.504
F-pseudosigma =			0.069			Lh = 0.435

Lab	Rating	Z-value	0	6	8	9
1	4	-0.49	--	--	0.438	--
3	4	0.00	--	--	0.463	--
10	1	1.88	--	--	0.560	--
11	4	-0.45	--	--	0.440	--
12	0	2.66	--	--	0.600	--
13	NR	--	--	--	< 4	--
32	1	1.69	--	0.550	--	--
39	4	-0.12	--	--	0.457	--
45	0	-6.85	--	--	0.110	--
46	4	-0.29	--	--	0.448	--
48	2	-1.42	--	--	0.390	--
50	2	1.49	--	--	0.540	--
59	4	0.33	--	--	0.480	--
70	4	0.21	--	--	0.474	--
81	2	-1.46	--	--	0.388	--
86	4	-0.19	--	--	--	0.453
87	0	2.66	--	--	0.600	--
89	4	0.14	--	--	0.470	--
96	4	-0.06	--	--	0.460	--
113	4	-0.25	--	--	0.450	--
126	2	1.49	--	--	0.540	--
127	3	-0.85	--	--	0.419	--
134	4	0.14	--	--	0.470	--
138	3	-0.83	--	--	0.420	--
141	2	1.30	--	--	0.530	--
142	4	-0.06	--	--	0.460	--
144	1	1.69	--	--	0.550	--
145	2	-1.42	--	--	0.390	--
146	0	2.35	--	--	0.584	--
147	4	-0.50	--	--	0.437	--
151	4	-0.25	--	--	0.450	--
154	0	-5.10	--	--	0.200	--
180	0	-3.16	--	--	0.300	--
193	4	0.06	--	--	0.466	--
198	3	0.87	--	--	0.508	--
203	3	0.72	--	--	0.500	--
212	0	2.06	--	--	0.569	--
213	3	0.72	--	--	0.500	--
215	0	2.66	--	--	0.600	--
219	3	0.52	0.490	--	--	--
234	4	-0.45	--	--	0.440	--
241	2	-1.03	--	0.410	--	--
247	4	0.33	--	--	0.480	--
255	1	-2.02	--	--	0.359	--
259	0	4.21	--	--	0.680	--
265	4	0.14	--	0.470	--	--
277	4	0.23	--	--	0.475	--
282	3	-0.60	--	--	0.432	--
298	4	0.33	0.480	--	--	--
304	4	-0.45	--	--	--	0.440

Table 17. Most probable values for constituents and properties in standard reference samples distributed in October 1999
 [MPV, most probable value; N, number of samples; µg/L, microgram per liter; mg/L, milligram per liter; µS/cm, microsiemens per centimeter at 25 degrees Celsius.]

T-159 (trace constituents)

Analyte	MPV	F-pseudosigma	N	Analyte	MPV	F-pseudosigma	N
Ag	9.67 µg/L	0.91	59	Mg	5.60 mg/L	0.21	73
Al	31.9 µg/L	3.7	47	Mn	22.0 µg/L	1.6	68
As	28.4 µg/L	1.6	59	Mo	41.4 µg/L	2.6	42
B	26.4 µg/L	3.0	30	Na	100 mg/L	4	71
Ba	38.1 µg/L	1.9	54	Ni	22.2 µg/L	1.6	61
Be	10.8 µg/L	0.4	50	Pb	16.6 µg/L	1.2	66
Ca	25.5 mg/L	0.8	75	Sb	13.9 µg/L	1.1	43
Cd	24.0 µg/L	1.6	68	Se	5.49 µg/L	0.83	49
Co	13.3 µg/L	0.9	45	SiO ₂	11.5 mg/L	0.7	42
Cr	26.8 µg/L	1.8	64	Sr	190 µg/L	7	35
Cu	33.4 µg/L	2.5	69	Tl	13.8 µg/L	1.6	40
Fe	48.9 µg/L	6.2	68	U	5.04 µg/L	0.32	12
K	1.52 mg/L	0.13	67	V	14.4 µg/L	1.7	40
Li	8.57 µg/L	1.93	18	Zn	19.2 µg/L	1.9	65

M-152 (major constituents)

Analyte	MPV	F-pseudosigma	N	Analyte	MPV	F-pseudosigma	N
Alkalinity as CaCO ₃	6.00 mg/L	1.49	68	Na	16.3 mg/L	0.9	74
B	92.2 µg/L	7.8	32	total P as P	0.386 mg/L	0.024	56
Ca	3.63 mg/L	0.18	74	pH	6.60 units	0.22	79
Cl	17.3 mg/L	0.9	80	SiO ₂	4.50 mg/L	0.39	53
DSRD	66.5 mg/L	8.2	46	SO ₄	3.76 mg/L	0.33	69
F	1.98 mg/L	0.12	63	Sp Cond	113 µS/cm	3	74
K	0.92 mg/L	0.09	67	Sr	18.3 µg/L	0.9	33
Mg	0.722 mg/L	0.047	73	V	30.0 µg/L	1.9	37

N-63 (nutrient constituents)

Analyte	MPV	F-pseudosigma	N
NH ₃ as N	0.150 mg/L	0.021	80
NH ₃ +OrgN as N	0.200 mg/L	0.066	48
NO ₃ as N	0.084 mg/L	0.008	79
total P as P	0.158 mg/L	0.010	69
PO ₄ as P	0.142 mg/L	0.010	73

N-64 (nutrient constituents)

Analyte	MPV	F-pseudosigma	N
NH ₃ as N	1.38 mg/L	0.08	70
NH ₃ +OrgN as N	1.54 mg/L	0.13	47
NO ₃ as N	1.26 mg/L	0.04	71
total P as P	0.883 mg/L	0.041	61
PO ₄ as P	0.860 mg/L	0.034	65

P-33 (low ionic strength constituents)

Analyte	MPV	F-pseudosigma	N	Analyte	MPV	F-pseudosigma	N
Acidity	Insufficient data			Na	1.42 mg/L	0.10	46
Ca	0.322 mg/L	0.043	47	pH	4.67 units	0.13	53
Cl	2.38 mg/L	0.18	50	PO ₄ as P	0.045 mg/L	0.007	40
F	0.105 mg/L	0.039	33	SO ₄	1.41 mg/L	0.17	43
K	0.110 mg/L	0.015	38	Sp Cond	20.8 µS/cm	2.4	49
Mg	0.100 mg/L	0.015	42				

Hg-29 (mercury)

Analyte	MPV	F-pseudosigma	N
Hg	0.463 µg/L	0.052	51