

UNITED STATES  
DEPARTMENT OF THE INTERIOR  
GEOLOGICAL SURVEY

REPORT OF THE U.S. GEOLOGICAL SURVEY'S ANALYTICAL EVALUATION  
PROGRAM—STANDARD REFERENCE WATER SAMPLES M6, M94 (MAJOR  
CONSTITUENTS), T95 (TRACE CONSTITUENTS), N16 (NUTRIENTS), P8  
(PRECIPITATION SNOWMELT), AND SED3 (SEDIMENT).

---

Denver, Colorado

July 1986

## CONTENTS

	<u>Page</u>
Abstract . . . . .	1
Introduction . . . . .	1
Purpose and plan . . . . .	1
Preparation of samples . . . . .	2
Determinations . . . . .	3
Statistical evaluation . . . . .	6
Laboratory performance and reported values . . . . .	6
References . . . . .	7
Participating laboratories . . . . .	8

## TABLES

Table 1.	Explanation of abbreviations and symbols used in computer-printout tables . . . . .	12
2.	Overall laboratory performance; Standard Reference Water Sample M6 (major constituents) . . . . .	13
3.	Overall laboratory performance; Standard Reference Water Sample M94 (major constituents). . . . .	15
4.	Overall laboratory performance; Standard Reference Water Sample T95 (trace constituents) . . . . .	17
5.	Overall laboratory performance; Standard Reference Water Sample N16 (nutrients) . . . . .	20
6.	Overall laboratory performance; Standard Reference Water Sample P8 (precipitation snowmelt) . . . . .	21
7.	Overall laboratory performance; Standard Reference Water Sample SED3 (sediment) . . . . .	24
8.	Analytical data; Standard Reference Water Sample M6 (major constituents) . . . . .	25
9.	Statistics by method for Standard Reference Water Sample M6 (major constituents) . . . . .	42
10.	Analytical data; Standard Reference Water Sample M94 (major constituents). . . . .	44
11.	Statistics by method for Standard Reference Water Sample M94 (major constituents). . . . .	61
12.	Analytical data; Standard Reference Water Sample T95 (trace constituents) . . . . .	63
13.	Statistics by method for Standard Reference Water Sample T95 (trace constituents) . . . . .	83
14.	Analytical data; Standard Reference Water Sample N16 (nutrients) . . . . .	85
15.	Statistics by method for Standard Reference Water Sample N16 (nutrients) . . . . .	91
16.	Analytical data; Standard Reference Water Sample P8 (precipitation snowmelt) . . . . .	92
17.	Statistics by method for Standard Reference Water Sample P8 (precipitation snowmelt) . . . . .	101
18.	Analytical data; Standard Reference Water Sample SED3 (sediment) . . . . .	102
19.	Statistics by method for Standard Reference Water Sample SED3 (sediment) . . . . .	109

Comments, suggestions, or questions regarding these samples or this program may be made by calling (303) 236-3612 (FTS 776-3612), or by writing to Victor J. Janzer, SRWS Program, U.S. Geological Survey, 5293 Ward Road, Arvada, CO 80002.

## ABSTRACT

The U.S. Geological Survey began an interlaboratory testing program of reference water samples in 1962. Program objectives then were, and now are, to provide a means for participating water laboratories to: (1) Identify analytical problem areas; (2) ascertain the accuracy and precision of common water analyses and analytical methods; and (3) provide reference samples for quality-assurance testing. Participation in this continuing quality-assurance program is mandatory for all domestic laboratories providing water-analysis data for U.S. Geological Survey use.

This report presents analytical data submitted by the laboratories that analyzed the reference samples distributed in **May 1986**. Relative performance ratings achieved by the laboratories for each determination, statistical evaluation of the data, and data summaries are given in 19 tables.

## INTRODUCTION

The U.S. Geological Survey began an interlaboratory testing program of reference water samples in 1962 with a single major-constituent reference sample prepared from distilled water and reagent grade chemicals. Principal objectives of this continuing program are to provide a means for participating water laboratories to: (1) Identify analytical problem areas; (2) ascertain the accuracy and precision of analytical methods for determining various constituents and physical properties of water; and (3) provide reference samples for continuing quality-assurance testing of U.S. Geological Survey and various cooperator and contract laboratories. Twenty-three U.S. Geological Survey laboratories participated in the 1962 effort to determine 6 constituents in a single major-constituent Standard Reference Water Sample (SRWS). Today, more than 120 domestic laboratories, both Survey and non-Survey, participate in the program, which currently uses up to ten SRS types: (1) major constituents; (2) trace constituents; (3) nutrients; (4) herbicides; (5) insecticides; (6) water and suspended-sediment mixture for trace metals; (7) precipitation snowmelt; (8) priority pollutants; (9) acid mine drainage; and (10) a sediment (bed material) for trace metals.

Participation in this continuing quality-assurance program is mandatory for all laboratories providing water-analysis data for U.S. Geological Survey use. Major constituent, trace-constituent, and nutrient SRWS are prepared and distributed to participating laboratories twice each year. One or more of the other SRS types also may be included. **This report presents analytical data submitted by the laboratories that analyzed the reference samples distributed to them in May 1986. Samples were analyzed during May and June, and data were requested to be submitted by June 13. Data received through July 23, 1986, have been included in this report. Relative performance ratings achieved by the laboratories for each determination, and statistical evaluations of the data are given in 19 tables.**

## PURPOSE AND PLAN

Periodic participation in reference sample analysis provides the means to alert participating laboratories to deficiencies in their analytical operations, and provides reference solutions for continuing quality assurance testing. Standard Reference Water Samples are prepared and distributed approximately every 6 months for analysis by U.S. Geological Survey and other cooperating laboratories. These analyses provide independent and objective evaluations of water-quality data provided by some of these laboratories for Survey use and publication. Non-U.S. Geological Survey laboratories participating in these studies are identified only by a confidential code number whereas participating U.S. Geological Survey laboratories are identified by location, name and code number.

This report summarizes the analytical results submitted by 110 of the 131 laboratories that requested and were shipped samples for this round of testing. The original date of June 13, 1986, that was specified as a deadline for data return was extended several times.

Samples which were distributed during May 1986 included SRWS M6 and M94 (major constituents), SRWS T95 (trace constituents), SRWS N16 (nutrients), P8 (precipitation snowmelt), and SED3 (sediment). Not all samples are requested nor necessarily analyzed by all laboratories, nor do all laboratories enrolled in the program participate in each round of analyses. Each participating laboratory was asked to perform at least those determinations that it makes routinely on the respective sample type, and to indicate the analytical methods used for each constituent. When method information was provided, it has been included in the respective data tables.

## PREPARATION OF SAMPLES

SRWS M94 (major constituents), N16 (nutrients), and T95 (trace constituents) were prepared from natural surface waters. SRWS P8 (precipitation) was prepared from melted snow.

Samples M94, T95, N16, and P8 were prepared by allowing the suspended sediment to settle for several days in the collection drums. The partly clarified sample was then filtered sequentially through a 10  $\mu\text{m}$  (micrometer) nominal size prefilter, then a 5  $\mu\text{m}$  nominal size intermediate filter and finally a 0.45  $\mu\text{m}$  membrane filter, into large clean polyethylene drums.

The cold tap water used to prepare M6 was filtered into a clean polyethylene drum using only the 0.45  $\mu\text{m}$  filters. Thymol [1.25 mg/L (milligrams per liter)] was added to M6 to inhibit fungal and bacterial growth. No other additions were made to M6. Natural trace metal abundances in T95, however, were supplemented by the addition of selected constituents as solutions of reagent grade chemicals. These included aluminum, iron, manganese, antimony, arsenic, barium, boron, beryllium, cadmium, chromium, cobalt, copper, lead, lithium, molybdenum, nickel, selenium, silver, thallium and zinc. Solutions for M6, M94, and T95 were mixed overnight with a motor driven Teflon<sup>1</sup>-coated stirrer, after which they were again filtered through 0.45  $\mu\text{m}$  membrane filters, followed by a 0.2  $\mu\text{m}$  "final membrane filter", then passed through a flow-through ultraviolet [254-nm (nanometer)] sterilizer and packaged under ultraviolet radiation, in dry-heat sterilized 1-L Teflon or autoclaved polypropylene bottles. Due to persistent problems of bacterial and fungal growths when using thymol as a preservative, free chlorine was used as a preservative in samples M94 and T95. Prior to bottling these samples, sodium hypochlorite was added to achieve initial concentrations of several parts per million free chlorine. No additions were made to P8.

Natural nutrient concentration levels in SRWS N16 were supplemented by the addition of ammonium, nitrate, nitrite, and orthophosphate "ions", and organic nitrogen and phosphorous as dissolved reagent grade chemicals. It was preserved by the addition of mercuric chloride (50 mg/L). Sodium chloride (450 mg/L) was also added. This is equivalent to the U.S. Geological Survey technique for field preservation of nutrient samples, using mercuric chloride and sodium chloride. The sample was then mixed overnight with a motor-driven, Teflon-coated stirrer, packaged in polyethylene bottles, without sterilization, and stored in the dark at 4 °C (Celsius), until needed. With the exception of the nutrient samples which were shipped in iced coolers, the samples for this round-robin testing were shipped at ambient temperatures.

SRS SED3 (bed material) represents a continuation of efforts to prepare a natural sediment reference material for selected major and trace constituents, to be determined by "soft digestion" (dilute acid) techniques. SED 3 was prepared by air drying, gently

<sup>1</sup>/ The use of the trade name in this report is for identification purposes only and does not constitute endorsement by the U.S. Geological Survey.

disaggregating any lumps formed on drying, and then sieving a natural fine-grained sandy sediment. All material passing a 125 micrometer-size sieve was retained. The resulting material was well mixed and then packaged in 20 ml plastic vials.

Analytical results for SED3 have been tabulated and evaluated in the same manner as the water reference samples. Although individual laboratory ratings for SED3 have been determined and tabulated (table 7), results should only be considered as indicative of a laboratory's capabilities on what may or may not be a suitable reference material. Numerous questions remain regarding suitable digestion and analysis procedures, as well as homogeneity and suitability of SED3 as an acceptable reference material.

### DETERMINATIONS

Abbreviations or symbols are listed below for each determination made on the various SRS. These abbreviations and symbols are used in tables 2-19. Additional abbreviations and symbols used in tables are explained in table 1.

Standard Reference Water Samples in this listing include: M6, M94 (major constituents), T95 (trace constituents), N16 (nutrients), P8 (precipitation snowmelt), and SED3 (sediment).

Abbreviation/Symbol	M6	M94	T95	N16	P8	SED3
	(mg/L) <u>1/</u>	(mg/L) <u>1/</u>	(µg/L) <u>2/</u>	(mg/L)	(µg/L) <u>3/</u>	(µg/g) <u>4/</u>
ALK(CACO3) = Alkalinity (as CaCO <sub>3</sub> )	x	x				
ACID@CACO3 = Acidity (as CaCO <sub>3</sub> )			x		x	
AG = Silver			x		x	
AL = Aluminum			x			x
AS = Arsenic			x			
B = Boron	x	x	x			x
BA = Barium			x			x
BE = Beryllium			x			x
BR = Bromide	x	x				
C, inorganic = total inorganic carbon						x
C, total = total carbon						x
CA = Calcium	x	x	x		x	x
CD = Cadmium			x		x	x
CL = Chloride	x	x			x	
CO = Cobalt			x		x	x

Standard Reference Water Samples in this listing include: M6, M94 (major constituents), T95 (trace constituents), N16 (nutrients), P8 (precipitation snowmelt), and SED3 (sediment), (continued).

Abbreviation/Symbol		M6 (mg/L) <u>1/</u>	M94 (mg/L) <u>1/</u>	T95 (µg/L) <u>2/</u>	N16 (mg/L)	P8 (µg/L) <u>3/</u>	SED3 (µg/g) <u>4/</u>
CR TOT	= Chromium, total			x		x	x
CU	= Copper			x		x	x
DSRD 180	= Dissolved solids, 180°C	x	x				
F	= Fluoride	x	x			x	
FE	= Iron			x		x	x
HG	= Mercury			x			x
I	= Iodide	x	x				
K	= Potassium	x	x	x		x	x
LI	= Lithium			x			x
MG	= Magnesium	x	x	x		x	x
MN	= Manganese			x		x	x
MO	= Molybdenum			x			x
NA	= Sodium	x	x	x		x	x
NH3-N	= Ammonia as nitrogen				x	x	
NI	= Nickel			x			x
NO2-N	= Nitrite as nitrogen	x	x		x		
NO3-N	= Nitrate as nitrogen	x	x		x		
ORG-N	= Organic nitrogen as nitrogen				x		
PB	= Lead			x		x	x
PH	= pH	x	x			x	
PO4-P	= Orthophosphate as phosphorus				x		
P, TOTAL	= Phosphorus, Total as phosphorus	x	x		x		

Standard Reference Water Samples in this listing include: M6, M94 (major constituents), T95 (trace constituents), N16 (nutrients), P8 (precipitation snowmelt), and SED3 (sediment), (continued).

Abbreviation/Symbol		M6	M94	T95	N16	P8	SED3
		(mg/L) <u>1/</u>	(mg/L) <u>1/</u>	(µg/L) <u>2/</u>	(mg/L)	(µg/L) <u>3/</u>	(µg/g) <u>4/</u>
SB	= Antimony			x			
SE	= Selenium			x			x
SIO2	= Silica	x	x	x			
SO4	= Sulfate	x	x			x	
SP.COND.	= Specific conductance	x	x			x	
SR	= Strontium	x	x	x			x
TL	= Thallium			x		x	
V	= Vanadium	x	x	x			x
ZN	= Zinc			x		x	x

1/ Results in mg/L except specific conductance (microsiemens or micromhos per centimeter at 25 °C); pH (units); boron, bromide, iodide, strontium, and vanadium (micrograms per liter).

2/ Results in µg/L except acidity, calcium, magnesium, potassium, silica, and sodium (milligrams per liter).

3/ Results in µg/L except specific conductance (microsiemens per centimeter at 25 °C); pH (units); acidity, ammonia, calcium, chloride, fluoride, magnesium, potassium, sodium and sulfate (milligrams per liter).

4/ Results in µg/g except aluminum, calcium, iron, magnesium potassium and sodium, (milligrams per gram).



## STATISTICAL EVALUATION

Statistical evaluations of the data were made to estimate the most probable value (MPV) for each of the constituents determined. Reported values of "less than" were considered as "not determined" and were not used (ignored) in the computation of the means, standard deviations, and so forth.

Outlying values for the remaining data were rejected on the basis of statistical tests as outlined in American Society for Testing and Materials (1981). After rejection of the outliers, the data remaining for each constituent were used to calculate the means, standard deviations, and percent deviation from the mean for each value. Outliers are not recalculated when determining the means and standard deviations for each determination listed by "method". The total range for each constituent included those values rejected as outliers. Confidence limits about the mean also were calculated; these limits define the range within which the true value may be expected to occur with a confidence level of 95 percent.

The mean, standard deviation, and confidence limits about the mean usually are reported to one more significant figure than the reported value. Statistical information is tabulated for each method used by three or more laboratories to determine a specific constituent. Tables 9, 11, 13, 15, 17, and 19, listing the mean and standard deviation for the constituent determined by each method, and the number of laboratories that used it, follow the analytical-data tables for each SRWS.

## LABORATORY PERFORMANCE AND REPORTED VALUES

To facilitate interlaboratory performance comparisons, ratings based on the analyses reported for each SRWS are included as tables 2-4 in this report. Laboratory performance for each constituent is rated on an arbitrary scale of 0 to 4 based on the number of "standard deviations" from the mean as indicated below:

4 (Excellent)	-----0.00 to 0.50 standard deviation
3 (Good)	-----0.51 to 1.00 standard deviation
2 (Satisfactory)	-----1.01 to 1.50 standard deviations
1 (Questionable)	-----1.51 to 2.00 standard deviations
0 (Poor)	-----Greater than 2.00 standard deviations

Averages of the constituent ratings for each Standard Reference Water Sample are given for each laboratory in the tables of overall laboratory performance (tables 2-6).

Laboratories were requested to identify the method used for each determination. The references for these methods are included with the analytical data and are identified in the following listing:

1. American Public Health Association and others, 1980, Standard methods for the examination of water and wastewater [15th ed.]: Washington, D.C., 1134 p.
2. American Society for Testing and Materials, 1982, Annual book of ASTM standards, Part 31: Philadelphia, PA, U.S.A., 1554 p.

3. Kopp, J. F., and McKee, G. F., 1978, Methods for chemical analysis of water and wastes: Cincinnati, Ohio, U.S. Environmental Protection Agency, 460 p.
4. Skougstad, M. W., Fishman, M. J., Friedman, L. C., Erdmann, D. E., and Duncan, S. S., eds., 1979, Methods for determination of inorganic substances in water and fluvial sediments: U.S. Geological Survey Techniques of Water-Resources Investigations, Book 5, Chapter A1, 626 p.
5. Fishman, M. J., and Bradford, W. L., eds., 1982, A supplement to methods for the determinations of inorganic substances in water and fluvial sediments: (Supplement to U.S. Geological Survey Techniques of Water-Resources Investigations, Book 5, Chapter A1), U.S. Geological Survey Open-File Report 82-272, 136 p.
6. Fishman, M. J., and Pyen, Grace, 1979, Determination of selected anions in water by ion chromatography: U.S. Geological Survey Water-Resources Investigations 79-101, 30 p.
7. Other references and instrument manufacturer's operation manuals.

In many instances, virtually the same method is given in several references. In those cases, all references describing that method are listed. If the analytical method used was not included in any of the listed references, analysts were requested to indicate "Other". Reference columns are left blank, if no method or an "other" method was indicated.

Values reported for all constituents determined in each SRS are listed in tables 8, 10, 12, 14, 16, and 18. Each value has been rounded off, when necessary, to conform to U.S. Geological Survey policy on reporting analytical data as given by Bishop and others (1978).

#### REFERENCES

- American Society for Testing and Materials, 1981, Annual book of ASTM standards, Part 41, Philadelphia, Pa., 1390 p.
- Bishop, E. E., Eckel, E. B., and others, 1978, Suggestions to Authors of the reports of the, U.S. Geological Survey: Washington, D. C., U.S. Government Printing Office, 6th edition, p. 198.

## PARTICIPATING LABORATORIES

### U.S. Geological Survey

CALIFORNIA, Sacramento: Yates 108

COLORADO, Denver: Arozarena 057

Reddy 119

Taylor/Garbarino 144

Taylor/Hedley 051

FLORIDA, Ocala: Kirkland 085

LOUISIANA, Baton Rouge: Garrison 077

### Cooperator

ALABAMA, Montgomery: ADEM Environmental Laboratory

ALASKA, Anchorage: Anchorage Water and Wastewater Utility

ARKANSAS, Little Rock: Ark. Department of Pollution Control and Ecology

CALIFORNIA, Bryte: California Department of Water Resources Chemical Laboratory

CALIFORNIA, Castaic: Department of Water Resources Chemical Laboratory

CALIFORNIA, La Mesa: San Diego Water Utilities Laboratory

CALIFORNIA, La Verne: The Metropolitan Water District of Southern California

CALIFORNIA, Lakeside: Helix Water District

CALIFORNIA, Mammoth Lakes: Sierra Nevada Aquatic Research Lab

CALIFORNIA, Oakland: East Bay Municipal Utility District

CALIFORNIA, Palm Desert: California Regional Water Quality Control Board

CALIFORNIA, Santa Barbara: University of California

CALIFORNIA, Santa Fe Springs: West Coast Analytical Service, Inc.

CALIFORNIA, Sausalito: U.S. Corps of Engineers-Pacific Division Laboratory

COLORADO, Alamosa: Bureau of Reclamation

COLORADO, Arvada: Rocky Mountain Analytical Laboratory

COLORADO, Aurora: Core Laboratories Incorporated

COLORADO, Denver: Metropolitan Denver Sewage Disposal District #1

COLORADO, Denver: Colorado Department of Health

COLORADO, Fort Collins: Environmental Services/Water Utilities

COLORADO, Golden: Rockwell International General Laboratories

COLORADO, Parachute: Union Oil Company, Upgrade Laboratory

COLORADO, Steamboat Springs: ACZ Inc/Bookcliffs Laboratory Division

FLORIDA, Tallahassee: City of Tallahassee Water Quality Laboratory

FLORIDA, Tampa: Hillsborough County Environmental Protection Commission

GEORGIA, Albany: Water, Gas and Light Commission

GEORGIA, Athens: Soil Testing and Plant Analysis Laboratory

GEORGIA, Athens: Univ. of Ga. Department of Horticulture

GEORGIA, Atlanta: Georgia Department of Natural Resources

ILLINOIS, Champaign: Illinois State Water Survey

ILLINOIS, Champaign: Illinois Environmental Protection Agency

ILLINOIS, Chicago: Illinois Environmental Protection Agency

Cooperator--continued

INDIANA, Indianapolis: Marion County Public Health Laboratory  
INDIANA, Indianapolis: Indianapolis Department of Public Works

IOWA, Des Moines: University Hygienic Laboratory - Des Moines Branch

KANSAS, Lawrence: Kansas Geological Survey  
KANSAS, Topeka: Kansas Department of Health and Environment

KENTUCKY, Bowling Green: Western Kentucky University  
KENTUCKY, Frankfort: Kentucky Natural Resources & Environmental Protection  
KENTUCKY, Louisville: Univ. of Louisville, Water Resources Lab

LOUISIANA, Lake Charles: Core Laboratories, Inc.

MAINE, Augusta: Maine Department of Environmental Protection

MARYLAND, Baltimore: Martel Laboratory Services, Inc.

MASSACHUSETTS, Wellesley Hills: Massachusetts Department of Public Works

MINNESOTA, Minneapolis: Braun Eng. and Testing Inc.  
MINNESOTA, Minneapolis: Minnesota Public Health Department  
MINNESOTA, St. Paul: Metropolitan Waste Control Commission

MISSOURI, Jefferson City: Missouri Dept. of Natural Resources

MONTANA, Butte: Montana Bureau of Mines and Geology

NEVADA, Boulder City: BOR, Lower Colorado Regional Lab  
NEVADA, Las Vegas: Clark County Sanitation District  
NEVADA, Reno: Water Analysis Laboratory, Desert Research Institute  
NEVADA, Reno: Nevada State Health Laboratory  
NEVADA, Sparks: City of Sparks, Jt. Treatment Plant  
NEVADA, Sutcliffe: Pyramid Lake Fisheries

NEW JERSEY, Tom's River: Ocean County Health Department  
NEW JERSEY, Trenton: N.J. Department of Health Environmental and Chemical  
Laboratory

NEW MEXICO, Albuquerque: City of Albuquerque Water Resources Laboratory  
NEW MEXICO, Gallup: Bureau of Indian Affairs - Natural Resources and Engineering  
Laboratory

NEW YORK, Buffalo: Erie County Public Health Laboratory  
NEW YORK, Central Islip: Suffolk County Health Services Department  
NEW YORK, Hempstead: Nassau County Department of Health  
NEW YORK, Millbrook: Institute of Ecosystem Studies  
NEW YORK, New York City: New York City Health Department  
NEW YORK, North Babylon: EcoTest Laboratories, Inc.

Cooperator--continued

NEW YORK, Oakdale: Suffolk County Water Authority  
NEW YORK, Rochester: FEV Wastewater Treatment Facility Laboratory  
NEW YORK, Rochester: Monroe County Environmental Health Laboratory  
NEW YORK, Syracuse: Syracuse Univ. Department of Civil Engineering  
NEW YORK, Westbury: Nytest Environmental, Inc.

NORTH CAROLINA, Browns Summit: Lake Townsend Water Filtration Plant  
NORTH CAROLINA, Charlotte: Mecklenburg County Environmental Health  
Department  
NORTH CAROLINA, Greensboro: City of Greensboro, Osborne Plant

NORTH DAKOTA, Bismarck: North Dakota State Water Commission

OHIO, Columbus: Ohio Environmental Protection Agency Water Quality Laboratory  
OHIO, Dayton: The Miami Conservancy District  
OHIO, Medina: Medina County Sanitary Engineering Department  
OHIO, Tiffin: Heidelberg College, Water Quality Laboratory

OKLAHOMA, Norman: Oklahoma Geological Survey  
OKLAHOMA, Oklahoma City: Oklahoma Dept. of Agriculture Laboratory

OREGON, Corvallis: U.S. Department of Agriculture, Forestry Sciences Laboratory  
OREGON, Sandy: Bureau of Water Works, Water Quality Laboratory

PENNSYLVANIA, Harrisburg: Pennsylvania DER, Bureau of Laboratories

SOUTH DAKOTA, Brookings: South Dakota State University, Water Quality Laboratory  
SOUTH DAKOTA, Vermillion: South Dakota Geological Survey

TENNESSEE, Chattanooga: Tennessee Valley Authority, Laboratory Branch

TEXAS, Corpus Christi: Core Laboratories, Inc.  
TEXAS, Tyler: Core Laboratories, Inc.

UTAH, Logan: Ecosystem Research Institute

VIRGINIA, Culpeper: Environmental System Service  
VIRGINIA, Richmond: Commonwealth of VA DGS  
VIRGINIA, Springfield: VERSAR, Inc.

WASHINGTON, Richland: Battelle, Pacific NW Lab  
WASHINGTON, Richland: Rockwell Hanford Operations

WEST VIRGINIA, Morgantown: West Virginia Geologic and Economic Survey

WISCONSIN, Madison: State Laboratory of Hygiene  
WISCONSIN, Milwaukee: Milwaukee Metropolitan Sewerage District

WYOMING, Casper: Core Laboratories, Inc.  
WYOMING, Cheyenne: Department of Environmental Quality, Water Quality Division  
WYOMING, Laramie: Wyoming Department of Agriculture

PUERTO RICO, San Juan: Department of Natural Resources, Laboratory Division

Cooperator--continued

The following laboratories requested test samples, which were shipped to them, but no data were received:

ALABAMA, Tuscaloosa: Geological Survey of Alabama  
ALASKA, Fairbanks: Alaska Geology and Geophysics Survey  
ALASKA, Soldotna: Alaska Dept. of Fish and Game, Limnology Lab  
ARIZONA, Phoenix: Arizona State Laboratory  
CALIFORNIA, Berkeley: Calif. Department of Health, Radiation Laboratory  
CALIFORNIA, Menlo Park: U.S. Geological Survey  
COLORADO, Denver: Center for Environmental Sciences, Univ. of Colo./Denver  
COLORADO, Denver: Denver Water Department - Quality Control Laboratory  
COLORADO, Pueblo: Board of Water Works  
FLORIDA, Palatka: St. John's River Water Management District  
FLORIDA, West Palm Beach: South Florida Water Management District  
GEORGIA, Tifton: U.S. Department of Agriculture, SE Watershed Laboratory  
MAINE, Orono: Univeristy of Maine  
MASSACHUSETTS, Barnstable: Barnstable County Health & Environmental Dept.  
MISSOURI, Columbia: Environmental Trace Substances Research Center  
NEW YORK, Syracuse: Onondaga County Department of Drainage and Sanitation  
NEW YORK, Syracuse: Syracuse Univ. Heroy Laboratory  
NEW YORK, Wantagh: Cedar Creek Advanced Wastewater Treatment Lab  
SOUTH CAROLINA, Columbia: SC Water Resources Commission

**Table 1.--Explanation of abbreviations and symbols used in computer printout sections**

---

APDC - ammonium pyrrolidine dithiocarbamate  
AUTO - automated  
AVG - average  
BLK - block  
CHCL<sub>3</sub> - chloroform  
CO'METRIC - colorimetric  
DC - direct current  
DEV - deviation  
DIG - digestion  
EDTA - ethylenediaminetetraacetic acid  
H<sub>2</sub>SO<sub>4</sub> - sulfuric acid  
IC - inductively coupled  
IGNORED - valued reported as less than detection level and not used in statistical analyses  
K & HG SO<sub>4</sub> - potassium & mercuric sulfate  
MIBK - methyl isobutyl ketone  
NABH<sub>4</sub> - sodium borohydride  
ND - not determined  
NR - not rated  
PCT - percent  
PDCA - pyrrolidine dithiocarbamic acid  
PERSULF - persulfate  
PHOSPHOMOLYBD - phosphomolybdate  
REJECT - values identified as an outlier and not used in statistical analyses  
SPADNS - sodium 2-(parasulfophenylazo)-1,8-dihydroxy-3,6-naphthalene disulfonate  
STD - standard  
% - percent  
< - less than  
> - greater than  
--- - dashes in values columns of Statistics by Methods indicate methods used but without sufficient data to calculate meaningful means and standard deviations

---

Table 2 Standard Reference Water Sample No. M6 (Major Constituents)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	ALK(CACO3)	B	BR	CA	CL	DSRD 180	F	I	K	MG
1	2	NR	NR	4	4	4	3	ND	2	3
2	0	3	ND	ND	ND	3	3	ND	3	3
3	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
4	4	ND	ND	4	ND	4	3	ND	ND	ND
6	4	2	NR	4	4	2	3	ND	4	3
7	2	ND	ND	4	4	4	3	ND	4	3
8	3	3	ND	4	0	4	3	ND	4	3
9	4	4	ND	1	4	ND	0	ND	3	3
10	4	ND	ND	ND	ND	ND	3	ND	ND	ND
13	4	NR	ND	3	4	0	3	ND	3	3
14	ND	3	ND	ND	4	4	ND	ND	ND	ND
15	4	NR	ND	4	2	4	3	ND	4	3
16	4	ND	ND	ND	4	ND	3	ND	ND	ND
17	ND	ND	ND	0	ND	ND	ND	ND	3	3
19	0	ND	ND	0	0	4	3	ND	3	3
20	0	ND	ND	2	0	3	0	ND	3	3
22	0	NR	NR	2	0	4	3	NR	2	0
24	4	ND	ND	3	4	3	3	ND	4	0
25	2	3	ND	3	3	3	3	ND	2	3
27	2	ND	ND	3	4	3	3	ND	3	3
29	ND	ND	ND	2	1	4	3	ND	4	0
30	3	4	ND	4	4	4	3	ND	4	3
33	3	ND	ND	3	3	2	3	ND	4	3
34	4	NR	ND	4	2	4	3	ND	4	3
36	4	ND	ND	4	4	3	3	ND	3	3
37	3	ND	ND	0	4	3	3	ND	1	3
38	3	4	NR	3	4	4	3	ND	4	3
40	4	3	ND	4	2	3	3	ND	4	3
41	2	NR	ND	4	4	0	3	ND	4	3
42	ND	ND	ND	4	4	ND	ND	ND	3	3
43	4	ND	ND	2	3	2	ND	ND	4	3
44	3	ND	ND	ND	ND	ND	ND	ND	ND	ND
45	4	ND	ND	ND	ND	4	ND	ND	ND	ND
46	2	NR	ND	3	4	3	3	ND	0	3
47	ND	0	ND	3	1	3	3	ND	4	3
48	2	NR	ND	2	4	3	3	ND	0	3
49	0	3	ND	4	4	ND	3	ND	3	3
50	2	ND	ND	1	ND	0	3	ND	2	3
51	ND	3	ND	2	4	ND	3	ND	ND	3
52	4	3	ND	4	4	ND	ND	ND	4	3
53	4	ND	ND	2	3	4	ND	ND	2	0
56	4	ND	ND	0	3	4	3	ND	0	0
57	4	3	NR	4	3	3	3	ND	4	3
58	3	ND	ND	ND	2	3	3	ND	ND	ND
63	3	NR	ND	4	4	3	3	ND	0	0
64	4	ND	ND	1	ND	4	ND	ND	4	3
65	ND	ND	ND	ND	4	ND	ND	ND	ND	ND
67	2	2	ND	0	4	ND	ND	ND	3	3
69	0	ND	ND	ND	ND	ND	ND	ND	ND	ND
70	3	ND	ND	4	2	4	0	ND	ND	3
72	4	ND	ND	4	3	0	ND	ND	3	3
73	3	4	ND	4	2	ND	ND	ND	4	3
74	ND	ND	ND	ND	4	ND	ND	ND	ND	ND
76	3	ND	ND	4	3	3	0	ND	2	3
77	4	ND	ND	3	4	4	3	ND	4	3
79	2	ND	ND	ND	3	4	ND	ND	ND	ND
80	4	1	ND	4	4	2	3	ND	2	3
81	3	ND	ND	4	2	3	3	ND	0	3
82	ND	ND	ND	ND	0	ND	ND	ND	ND	ND
83	3	NR	ND	4	4	2	3	ND	0	3
84	3	ND	ND	0	4	0	ND	ND	4	3
85	3	NR	ND	3	2	4	3	ND	4	3
86	1	ND	ND	4	4	ND	3	ND	4	3
91	1	ND	ND	2	0	3	3	ND	2	3
93	3	0	ND	4	4	4	3	ND	4	3
94	0	ND	ND	ND	ND	2	ND	ND	ND	ND
95	ND	ND	ND	ND	0	ND	ND	ND	ND	ND
96	2	ND	ND	4	2	ND	3	ND	4	0
98	4	1	ND	3	4	3	3	ND	4	3
99	1	0	ND	0	0	1	3	ND	2	0
102	4	ND	ND	2	2	0	ND	ND	2	0
103	2	NR	ND	3	4	ND	3	ND	0	3
104	0	ND	ND	0	3	2	3	ND	4	3
107	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
108	ND	2	ND	4	4	4	ND	ND	4	3
109	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
110	ND	4	ND	4	ND	ND	ND	ND	0	3
111	ND	3	ND	4	ND	ND	ND	ND	ND	3
112	4	NR	ND	4	3	3	3	ND	2	3
113	0	ND	ND	4	ND	ND	0	ND	3	0
118	4	ND	ND	4	4	4	ND	ND	NR	3
121	ND	ND	ND	4	4	3	ND	ND	4	3
122	ND	ND	ND	1	2	3	ND	ND	0	0
123	1	2	ND	4	2	4	3	ND	4	3
124	0	ND	ND	ND	3	ND	ND	ND	ND	ND
127	ND	NR	NR	1	ND	ND	ND	NR	ND	3
142	3	4	NR	4	3	4	3	ND	3	3
144	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND



Table 2 Standard Reference Water Sample No. M6 (Major Constituents)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev.  
2 (Satisfactory) 1.01 to 1.50 Std. Dev.  
1 (Questionable) 1.51 to 2.00 Std. Dev.  
0 (Poor) > 2.00 Std. Dev.  
ND Not determined  
NR Not rated

LAB	NA	NO2-N	NO3-N	P, TOTAL	PH	SI02	SO4	SP. COND.	SR	V	N	Avg.
1	4	NR	1	NR	3	2	3	3	0	ND	14	2.71
2	ND	NR	4	NR	4	ND	2	0	ND	ND	10	2.50
3	ND	ND	ND	ND	3	ND	ND	3	ND	ND	2	3.00
4	3	ND	2	NR	0	ND	3	4	ND	ND	9	3.00
6	4	NR	4	3	4	4	4	2	4	NR	16	3.44
7	3	NR	4	0	0	ND	3	4	ND	ND	13	2.92
8	4	NR	2	NR	0	3	4	3	4	NR	15	2.93
9	4	ND	4	3	4	3	3	1	ND	ND	14	2.93
10	ND	ND	ND	ND	4	ND	3	1	ND	ND	5	3.00
13	2	NR	3	2	4	4	2	3	4	2	16	2.88
14	ND	NR	3	3	3	3	3	4	ND	ND	9	3.33
15	4	NR	1	NR	0	3	4	4	3	3	15	3.07
16	ND	ND	ND	ND	4	ND	ND	3	ND	ND	5	3.60
17	2	ND	ND	ND	ND	0	0	4	ND	ND	7	1.71
19	0	ND	0	ND	0	ND	0	0	ND	ND	12	0.83
20	2	NR	1	0	4	4	0	0	ND	ND	14	1.57
22	0	NR	0	NR	4	0	3	4	4	0	15	1.73
24	4	NR	4	ND	2	3	0	4	4	1	15	2.67
25	0	ND	4	ND	3	3	1	2	ND	ND	14	2.57
27	4	NR	1	ND	4	ND	4	3	ND	ND	12	3.00
29	3	4	4	3	3	ND	1	0	ND	ND	13	2.46
30	3	NR	4	4	4	4	4	4	ND	ND	14	3.71
33	0	ND	2	NR	3	ND	0	3	ND	ND	12	2.17
34	3	NR	4	3	3	2	4	4	ND	ND	14	3.36
36	0	NR	4	ND	3	4	4	4	3	ND	14	3.77
37	1	NR	4	3	3	ND	4	4	4	ND	13	2.77
38	4	NR	2	NR	3	3	3	4	4	NR	14	2.43
40	4	4	2	3	4	4	4	4	4	ND	16	3.38
41	3	4	3	3	0	2	4	4	4	NR	16	2.94
42	4	NR	4	3	4	4	3	ND	ND	ND	10	3.60
43	3	NR	2	NR	ND	ND	0	4	ND	ND	10	2.70
44	ND	0	ND	NR	3	ND	ND	ND	ND	ND	3	2.00
45	ND	ND	3	ND	ND	ND	4	4	ND	ND	4	3.75
46	1	NR	0	ND	0	ND	4	4	ND	ND	12	2.25
47	3	ND	0	4	3	1	2	4	4	3	16	2.56
48	3	NR	1	2	2	3	3	4	ND	ND	14	2.50
49	4	ND	4	4	4	4	4	3	4	3	16	3.38
50	4	4	4	ND	3	ND	0	4	ND	ND	12	2.50
51	4	ND	2	NR	ND	ND	4	ND	4	4	9	3.22
52	4	ND	4	ND	4	4	4	4	4	4	14	3.86
53	0	NR	1	NR	3	ND	3	3	3	ND	12	2.33
56	3	4	2	2	2	ND	4	2	ND	ND	14	2.50
57	3	NR	4	NR	2	4	4	4	4	3	16	3.44
58	ND	4	2	NR	4	ND	3	4	ND	ND	9	3.11
63	0	ND	3	NR	4	1	0	1	0	NR	14	1.86
64	4	4	4	3	2	4	4	4	ND	ND	12	3.42
65	ND	NR	3	2	ND	ND	ND	ND	ND	ND	3	3.00
67	2	ND	4	ND	4	ND	3	1	4	ND	12	2.67
69	ND	ND	ND	NR	3	ND	ND	0	ND	ND	3	1.00
70	0	ND	1	ND	4	ND	4	3	ND	ND	11	2.55
72	3	ND	3	ND	ND	ND	0	0	ND	ND	10	2.30
73	4	NR	3	3	4	4	ND	3	4	4	14	3.50
74	ND	NR	4	NR	4	ND	ND	4	ND	ND	4	4.00
76	3	NR	4	ND	2	ND	0	1	ND	ND	12	2.33
77	4	ND	ND	ND	4	ND	4	3	ND	ND	11	3.64
79	ND	NR	4	4	3	ND	3	0	ND	ND	8	2.88
80	4	ND	4	NR	4	3	4	3	4	3	16	3.13
81	3	NR	1	NR	0	ND	0	1	ND	ND	12	1.92
82	ND	ND	1	ND	ND	2	ND	3	ND	ND	4	1.50
83	2	4	0	3	3	3	4	3	ND	NR	15	2.73
84	4	ND	4	3	4	ND	1	0	ND	ND	12	2.50
85	4	4	3	3	3	3	1	4	4	4	16	3.25
86	4	NR	4	4	4	ND	3	2	ND	ND	12	3.33
91	3	NR	4	4	3	0	4	3	ND	ND	14	2.50
93	4	NR	1	0	3	1	3	3	4	NR	16	2.75
94	ND	ND	ND	ND	3	ND	ND	3	ND	ND	4	2.00
95	ND	NR	4	NR	4	4	ND	ND	ND	ND	4	3.00
96	3	ND	4	2	1	ND	4	4	4	ND	12	2.75
98	3	0	NR	NR	4	4	0	4	4	3	16	2.94
99	0	NR	3	ND	4	0	4	0	ND	ND	14	1.29
102	3	ND	ND	ND	4	ND	3	4	ND	ND	10	2.40
103	4	ND	4	ND	4	3	4	3	4	NR	13	3.15
104	4	ND	3	3	4	2	0	4	4	NR	15	2.60
107	ND	ND	4	3	ND	ND	2	ND	ND	ND	3	3.00
108	4	NR	4	0	2	4	2	4	ND	ND	13	3.15
109	ND	NR	0	NR	2	ND	ND	4	ND	ND	3	2.00
110	3	ND	ND	0	ND	4	ND	ND	4	1	9	2.56
111	4	ND	ND	ND	4	ND	ND	3	ND	ND	6	3.50
112	4	4	4	0	2	3	0	ND	0	ND	15	2.60
113	ND	NR	1	NR	1	ND	ND	ND	ND	ND	7	1.29
118	3	ND	4	NR	3	4	4	3	ND	ND	11	3.64
121	4	NR	4	NR	ND	ND	4	ND	ND	ND	8	3.75
122	0	ND	NR	ND	4	ND	3	4	ND	ND	9	1.89
123	4	ND	2	ND	3	4	4	3	ND	ND	14	3.07
124	ND	ND	ND	ND	1	ND	ND	ND	ND	ND	3	1.33
127	0	ND	ND	NR	ND	ND	ND	ND	4	ND	4	2.00
142	4	NR	1	NR	0	4	4	2	4	3	16	3.06
144	ND	ND	ND	ND	ND	ND	ND	ND	4	ND	1	4.00

Table 3 Standard Reference Water Sample No. M94 (Major Constituents)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	ALK(CACO3)	B	BR	CA	CL	DSRD 180	F	I	X	MG
1	3	4	3	4	2	4	3	ND	4	3
2	3	ND	ND	4	0	ND	4	ND	ND	3
3	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
4	3	ND	ND	3	4	4	3	ND	ND	ND
6	4	4	4	3	3	3	3	ND	4	3
7	2	ND	ND	3	1	4	3	ND	4	3
8	2	3	ND	4	0	3	2	ND	4	3
9	3	3	ND	3	4	ND	3	ND	4	3
12	4	ND	ND	ND	4	3	2	ND	ND	ND
13	4	NR	ND	4	3	2	3	ND	0	4
14	ND	0	ND	ND	4	4	ND	ND	ND	ND
15	3	3	ND	4	3	4	4	ND	3	2
16	4	ND	ND	ND	3	ND	4	ND	ND	ND
19	3	ND	4	0	0	2	0	ND	2	4
20	0	ND	ND	3	0	0	0	ND	1	3
22	0	0	3	4	0	4	0	NR	1	0
23	2	ND	ND	ND	ND	ND	ND	ND	ND	ND
24	3	ND	ND	4	4	4	3	ND	4	2
25	1	3	ND	2	0	4	1	ND	2	4
27	1	ND	ND	4	4	4	3	ND	3	4
29	ND	ND	ND	3	1	0	3	ND	4	0
30	0	ND	ND	4	4	4	3	ND	1	3
33	3	ND	ND	0	0	4	4	ND	2	4
34	4	NR	ND	4	1	4	1	ND	4	4
36	4	ND	4	4	4	1	2	ND	4	4
37	4	ND	ND	0	4	4	4	ND	1	3
38	0	4	4	3	2	4	4	ND	3	4
40	3	3	ND	4	4	4	4	ND	3	4
41	2	1	ND	1	4	2	2	ND	2	4
42	ND	ND	ND	3	4	ND	ND	ND	2	4
43	4	ND	ND	2	4	4	ND	ND	3	1
44	3	ND	ND	ND	ND	ND	ND	ND	ND	ND
45	3	ND	ND	ND	ND	4	4	ND	ND	ND
46	0	NR	ND	3	4	3	4	ND	4	2
47	ND	0	ND	3	3	3	0	ND	3	3
48	2	2	ND	3	3	3	4	ND	2	4
49	0	4	ND	4	4	ND	2	ND	4	4
50	1	ND	ND	0	ND	0	4	ND	3	0
52	1	2	ND	4	3	ND	ND	ND	4	4
53	3	ND	ND	2	4	4	ND	ND	4	0
54	0	4	1	1	2	ND	4	ND	2	3
56	4	ND	ND	0	3	0	0	ND	0	0
57	4	4	1	3	3	4	3	ND	4	4
58	4	ND	ND	ND	4	1	4	ND	ND	ND
59	2	ND	ND	3	1	3	ND	ND	0	0
61	ND	ND	ND	3	ND	ND	ND	ND	3	2
62	4	ND	ND	4	4	4	4	ND	2	3
63	4	NR	ND	2	4	3	4	ND	0	1
64	2	ND	ND	1	ND	2	ND	ND	3	4
65	ND	ND	ND	ND	2	ND	ND	ND	ND	ND
68	4	4	ND	3	4	4	3	ND	4	2
69	2	ND	ND	ND	ND	ND	ND	ND	ND	ND
70	3	ND	ND	4	4	4	0	ND	ND	2
72	4	ND	ND	2	4	0	ND	ND	4	4
73	4	3	ND	4	2	ND	ND	ND	3	4
74	ND	ND	ND	ND	4	ND	ND	ND	ND	ND
77	3	ND	ND	4	4	3	0	ND	2	1
79	3	ND	ND	ND	4	4	ND	ND	ND	ND
80	4	4	ND	4	3	0	3	ND	0	4
81	3	ND	ND	3	4	4	3	ND	0	3
82	ND	ND	ND	ND	0	ND	ND	ND	ND	ND
83	3	NR	ND	4	4	0	3	ND	2	1
84	3	ND	ND	0	0	0	ND	ND	3	0
85	3	4	ND	4	3	4	4	ND	4	3
86	0	ND	ND	4	3	ND	4	ND	3	4
87	3	ND	NR	2	2	ND	3	ND	4	4
90	3	ND	ND	3	4	ND	1	ND	4	4
91	2	ND	ND	3	3	2	3	ND	2	4
93	4	0	ND	4	4	4	2	ND	4	4
94	3	ND	ND	ND	ND	4	ND	ND	ND	ND
95	ND	ND	ND	ND	0	4	ND	ND	ND	ND
97	4	0	ND	4	4	1	1	ND	2	4
98	3	4	ND	4	4	2	3	ND	4	0
99	1	3	ND	2	4	0	4	ND	4	2
102	4	ND	ND	1	3	0	ND	ND	4	1
103	3	4	ND	3	4	ND	0	ND	4	3
104	4	ND	ND	4	3	3	3	ND	4	3
105	4	3	ND	4	1	0	4	ND	2	3
107	3	ND	ND	ND	3	3	ND	ND	ND	ND
108	ND	4	ND	4	3	3	ND	ND	3	4
110	ND	4	ND	4	ND	ND	ND	ND	2	3
111	ND	3	ND	4	ND	ND	ND	ND	1	3
112	4	NR	ND	0	4	3	4	ND	4	3
113	4	ND	ND	0	ND	ND	ND	ND	3	2
118	3	ND	ND	2	3	4	ND	ND	2	2
121	ND	ND	ND	4	3	2	ND	ND	4	4
122	ND	ND	ND	3	0	2	ND	ND	0	2
123	2	0	ND	4	4	4	2	ND	3	3
124	0	ND	ND	ND	3	ND	ND	ND	ND	ND
127	ND	NR	0	2	ND	ND	ND	NR	ND	0
142	3	4	4	4	3	4	1	ND	3	4

Table 3 Standard Reference Water Sample No. M94 (Major Constituents)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	NA	NO2-N	NO3-N	P <sub>i</sub>	TOTAL	PH	SI02	SO4	SP. COND.	SR	V	N	Avg.
1	3	NR	0	2	3	2	2	4	4	4	ND	17	3.06
2	3	ND	2	4	3	4	4	2	3	ND	ND	12	2.92
3	ND	ND	ND	4	4	4	ND	ND	1	ND	ND	3	3.00
4	3	ND	4	4	4	4	ND	2	0	ND	ND	11	3.09
6	4	NR	4	4	4	3	4	4	4	4	NR	17	3.65
7	0	NR	4	4	4	4	ND	0	3	ND	ND	13	2.69
8	2	NR	2	4	4	2	4	3	4	4	NR	16	2.88
9	3	ND	4	4	4	4	3	3	3	ND	ND	14	3.14
12	ND	NR	3	2	3	3	4	4	2	ND	ND	10	3.10
13	2	3	4	3	2	4	4	2	4	4	0	17	2.82
14	ND	NR	3	2	1	4	4	0	3	ND	ND	9	2.33
15	4	NR	3	4	3	4	4	0	4	0	NR	16	3.00
16	ND	ND	ND	ND	2	2	ND	ND	3	ND	ND	5	3.20
19	0	4	0	0	0	0	ND	0	0	ND	ND	14	1.36
20	4	NR	0	0	4	4	3	0	4	ND	ND	14	1.57
22	1	NR	4	2	3	3	0	0	4	2	0	18	1.72
23	ND	3	4	4	2	2	3	ND	3	ND	0	7	3.00
24	3	NR	4	0	3	3	4	0	4	1	0	16	2.69
25	0	ND	3	ND	4	4	4	3	4	ND	ND	14	2.50
27	4	NR	2	ND	4	4	ND	4	3	ND	ND	12	3.33
29	1	3	4	2	4	4	ND	1	0	ND	ND	13	2.00
30	4	NR	4	4	4	4	4	3	4	ND	ND	14	3.29
33	3	ND	4	2	0	0	ND	4	3	ND	ND	13	2.54
34	1	NR	3	3	ND	3	4	0	3	ND	ND	14	2.71
36	3	NR	3	ND	2	2	4	4	4	0	ND	15	3.13
37	0	NR	4	3	4	4	ND	4	4	ND	ND	13	3.08
38	2	NR	4	3	4	4	4	2	3	ND	NR	16	2.81
40	3	NR	1	4	2	2	4	3	0	ND	ND	15	3.07
41	2	NR	0	4	4	4	1	3	3	0	NR	17	2.00
42	4	3	4	0	0	4	4	2	ND	ND	ND	11	3.09
43	4	NR	0	4	2	2	ND	0	4	ND	ND	12	2.67
44	ND	0	ND	2	4	4	ND	ND	ND	ND	ND	4	2.25
45	ND	ND	4	ND	ND	ND	ND	2	ND	ND	ND	4	3.25
46	2	NR	1	ND	3	3	ND	3	4	ND	ND	12	2.75
47	2	ND	2	0	3	3	4	0	3	4	3	16	2.31
48	4	NR	4	4	4	4	4	2	3	ND	ND	15	3.20
49	3	ND	ND	3	3	3	4	3	3	4	3	15	3.20
50	4	0	4	ND	4	4	ND	4	3	ND	ND	12	2.50
52	4	ND	4	ND	3	3	4	4	3	3	2	14	3.21
53	4	NR	3	4	2	2	ND	3	3	2	ND	13	2.92
54	3	NR	4	4	3	3	2	3	4	1	4	17	2.65
56	3	3	0	0	0	3	ND	4	3	ND	ND	14	1.64
57	4	NR	4	4	4	4	4	4	4	4	3	18	3.61
58	ND	1	2	2	2	4	ND	4	4	ND	ND	10	3.00
59	0	ND	2	3	3	4	ND	3	2	ND	ND	12	1.92
61	4	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND	4	3.00
62	3	ND	3	ND	0	0	ND	4	3	ND	ND	12	3.17
63	0	ND	1	0	0	0	0	0	2	0	NR	15	1.67
64	3	3	4	4	4	0	4	ND	4	ND	ND	12	2.83
65	ND	NR	4	4	4	ND	ND	ND	ND	ND	ND	3	3.33
68	4	ND	4	ND	3	3	3	4	3	3	ND	15	3.47
69	ND	ND	ND	3	2	2	ND	ND	0	ND	ND	4	1.75
70	3	ND	1	ND	3	3	ND	4	0	ND	ND	11	2.55
72	1	ND	4	4	4	4	ND	0	0	4	ND	12	2.58
73	2	3	4	3	4	4	4	0	3	4	3	15	3.27
74	ND	NR	4	4	0	3	ND	ND	4	4	ND	5	3.20
77	4	ND	ND	ND	3	3	ND	4	4	ND	ND	11	2.91
79	ND	3	0	0	1	1	ND	2	4	4	ND	9	2.33
80	4	ND	4	4	4	2	4	4	3	4	3	17	3.18
81	0	NR	0	0	0	0	ND	0	0	ND	ND	13	1.54
82	ND	NR	3	4	4	ND	4	ND	3	ND	ND	5	2.80
83	4	NR	4	4	4	4	3	4	3	ND	NR	14	2.93
84	0	ND	4	4	3	3	ND	1	1	ND	ND	12	1.50
85	4	3	4	4	4	4	4	3	3	2	ND	17	3.53
86	4	3	4	4	4	4	4	0	0	ND	ND	13	3.00
87	4	ND	4	4	3	3	ND	2	3	ND	ND	12	3.17
90	4	ND	0	0	1	1	ND	0	0	ND	ND	12	1.75
91	4	3	2	2	2	2	0	3	4	ND	ND	15	2.47
93	4	NR	2	2	2	2	0	4	3	4	NR	16	2.94
94	ND	ND	ND	ND	2	2	ND	ND	4	ND	ND	4	3.25
95	ND	3	4	4	3	3	4	ND	3	ND	ND	8	3.13
97	4	ND	4	3	4	4	ND	3	4	ND	ND	14	3.14
98	0	NR	2	4	4	4	4	4	4	4	NR	16	2.94
99	4	4	4	4	3	3	2	0	0	ND	ND	16	2.69
102	4	ND	ND	ND	0	0	ND	0	3	ND	ND	10	2.00
103	3	ND	4	ND	3	3	4	4	2	4	NR	14	3.21
104	3	ND	0	3	0	0	0	2	4	4	0	16	2.31
105	4	ND	2	0	3	3	4	4	4	ND	ND	15	3.00
107	ND	ND	4	4	2	2	ND	3	2	ND	ND	8	3.00
108	4	NR	4	0	4	4	4	2	4	ND	ND	13	3.31
110	2	ND	ND	0	ND	0	0	ND	ND	4	1	9	2.22
111	3	ND	ND	3	4	4	ND	ND	3	ND	ND	8	3.00
112	3	3	3	4	4	4	4	4	ND	0	ND	15	3.13
113	3	3	ND	0	2	2	ND	ND	ND	ND	ND	8	2.13
118	4	ND	4	4	4	4	4	3	3	ND	ND	13	3.23
121	3	NR	4	3	ND	ND	ND	4	ND	ND	ND	9	3.44
122	2	ND	ND	ND	3	3	ND	3	3	ND	ND	8	1.88
123	4	ND	1	ND	3	3	4	4	3	ND	ND	14	2.93
124	ND	ND	ND	ND	3	3	ND	ND	ND	ND	NR	3	2.00
127	0	ND	ND	4	ND	ND	ND	ND	ND	4	NR	6	1.67
142	2	NR	3	4	0	0	4	4	3	4	4	18	3.22

Table 4 Standard Reference Water Sample No. T95 (Trace Constituents)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	ACID@CACO3	AG	AL	AS	B	BA	BE	CA	CD	CO
1	ND	ND	3	NR	4	4	NR	4	NR	NR
2	ND	3	ND	3	2	ND	ND	ND	NR	ND
4	ND	NR	NR	NR	ND	NR	ND	ND	4	NR
6	ND	NR	NR	NR	ND	1	NR	0	NR	NR
7	ND	NR	NR	NR	0	ND	NR	2	4	ND
8	4	NR	NR	NR	0	NR	NR	4	NR	NR
9	ND	0	3	NR	ND	2	NR	ND	NR	1
10	ND	NR	ND	NR	ND	ND	ND	ND	4	ND
13	ND	NR	3	NR	3	4	4	4	4	4
15	ND	2	3	NR	0	4	ND	4	3	NR
16	ND	3	ND	3	ND	3	ND	ND	4	ND
18	ND	ND	NR	ND	ND	NR	ND	ND	4	ND
19	4	ND	ND	ND	ND	ND	ND	ND	ND	ND
20	ND	NR	NR	NR	ND	NR	NR	3	NR	ND
22	ND	NR	NR	NR	4	3	NR	1	NR	NR
24	ND	4	4	NR	ND	0	4	4	4	4
25	ND	4	2	NR	4	NR	ND	3	NR	4
27	ND	NR	ND	NR	ND	0	ND	4	NR	ND
29	ND	ND	ND	ND	ND	ND	ND	4	NR	ND
30	4	3	4	3	ND	0	ND	3	3	3
32	ND	4	ND	ND	ND	ND	ND	4	0	ND
33	ND	ND	NR	NR	NR	ND	ND	1	NR	ND
34	ND	NR	4	NR	ND	1	NR	ND	NR	NR
36	ND	NR	NR	3	ND	ND	ND	3	4	ND
37	2	NR	3	NR	ND	0	NR	0	NR	ND
38	1	NR	0	NR	4	4	NR	3	3	ND
40	ND	3	4	NR	4	3	ND	4	3	ND
41	4	NR	NR	0	1	NR	NR	0	NR	NR
43	ND	NR	ND	2	ND	0	ND	2	3	ND
46	ND	2	4	2	ND	0	ND	ND	3	ND
47	ND	4	4	4	4	2	NR	3	3	NR
48	2	NR	NR	NR	4	NR	NR	4	NR	NR
49	ND	3	3	3	4	ND	ND	3	0	ND
50	ND	NR	ND	ND	ND	ND	NR	0	NR	ND
51	ND	ND	3	ND	4	3	ND	3	4	ND
52	ND	ND	ND	ND	4	3	ND	4	ND	ND
53	ND	ND	NR	NR	3	NR	ND	1	NR	ND
54	4	NR	3	NR	3	NR	NR	3	NR	NR
56	ND	3	ND	0	ND	ND	NR	0	4	ND
57	ND	NR	3	NR	ND	3	NR	3	NR	NR
58	ND	NR	NR	NR	ND	4	NR	4	NR	0
59	ND	2	ND	ND	ND	ND	ND	1	0	4
61	ND	ND	ND	ND	ND	ND	ND	2	ND	ND
63	ND	4	0	4	2	1	NR	ND	0	4
68	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
69	ND	ND	3	0	ND	3	0	ND	0	2
70	ND	0	ND	0	ND	ND	ND	4	ND	ND
72	ND	ND	ND	ND	ND	ND	ND	3	ND	ND
73	ND	3	0	0	4	3	NR	3	2	NR
74	ND	1	ND	4	ND	ND	ND	ND	4	ND
76	ND	NR	NR	NR	ND	NR	NR	ND	NR	NR
77	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
79	ND	ND	ND	ND	ND	ND	ND	ND	3	ND
80	4	NR	ND	4	4	3	NR	4	NR	NR
81	ND	NR	ND	NR	ND	ND	ND	3	NR	NR
83	4	NR	NR	NR	NR	NR	NR	4	NR	NR
84	ND	NR	4	NR	ND	3	ND	0	NR	ND
85	ND	ND	ND	ND	3	ND	ND	ND	NR	2
86	ND	ND	ND	ND	ND	ND	ND	4	ND	ND
90	ND	NR	ND	NR	ND	ND	ND	ND	NR	ND
93	ND	NR	NR	NR	0	NR	NR	3	NR	NR
94	ND	0	ND	ND	ND	0	ND	3	0	ND
95	ND	ND	ND	ND	ND	ND	ND	ND	3	ND
96	1	4	ND	NR	ND	3	ND	0	2	ND
98	ND	NR	3	NR	4	3	NR	2	NR	NR
102	3	ND	ND	ND	ND	ND	ND	2	0	ND
103	ND	NR	NR	ND	4	3	NR	4	NR	NR
104	4	NR	4	3	ND	NR	NR	4	4	NR
107	ND	ND	ND	2	ND	ND	ND	ND	3	ND
108	ND	ND	ND	NR	4	ND	ND	4	ND	ND
110	ND	0	3	0	4	4	4	3	0	4
111	ND	ND	ND	ND	4	ND	ND	4	ND	ND
112	2	ND	4	NR	4	4	NR	0	NR	ND
113	ND	ND	ND	ND	ND	ND	ND	3	ND	ND
118	ND	NR	ND	NR	ND	NR	ND	ND	3	ND
121	ND	ND	ND	3	ND	ND	ND	3	4	ND
122	ND	2	ND	3	ND	ND	ND	ND	4	ND
123	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
124	ND	NR	ND	ND	2	NR	NR	0	NR	NR
127	ND	NR	3	NR	4	3	NR	3	NR	NR
142	ND	NR	3	NR	4	3	NR	3	NR	NR
144	ND	ND	ND	ND	ND	3	ND	ND	3	ND

Table 4 Standard Reference Water Sample No. T95 (Trace Constituents)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	CR TOT	CU	FE	HG	K	LI	MG	MN	MO	PA
1	NR	3	NR	2	4	4	4	NR	ND	4
2	NR	2	0	3	3	ND	4	4	ND	ND
4	NR	3	NR	ND	ND	ND	ND	NR	NR	4
6	NR	4	NR	3	4	3	4	NR	NR	4
7	NR	NR	NR	3	4	ND	4	NR	NR	2
8	NR	NR	NR	NR	4	1	0	NR	NR	2
9	4	4	NR	2	ND	ND	ND	NR	ND	ND
10	3	ND	ND	0	ND	ND	ND	NR	ND	ND
13	NR	3	NR	ND	0	1	4	NR	4	0
15	2	3	0	4	3	2	2	ND	3	3
16	4	4	2	3	ND	ND	ND	3	ND	ND
18	2	0	2	ND	ND	ND	ND	3	ND	4
19	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
20	NR	NR	NR	NR	2	ND	0	NR	NR	4
22	0	4	NR	2	4	NR	2	NR	NR	4
24	3	2	2	4	4	ND	2	4	3	4
25	NR	2	NR	ND	4	ND	4	NR	0	3
27	NR	ND	NR	4	4	ND	4	NR	ND	ND
29	NR	3	NR	ND	0	ND	0	NR	ND	4
30	4	0	4	4	0	ND	4	3	ND	4
32	0	2	0	ND	4	ND	0	4	ND	3
33	NR	4	NR	4	3	4	4	NR	ND	0
34	NR	2	4	3	ND	ND	ND	NR	NR	ND
36	ND	4	2	ND	4	3	4	NR	ND	4
37	4	3	4	0	4	4	0	NR	ND	4
38	2	NR	NR	4	1	ND	2	NR	NR	3
40	2	1	3	4	3	ND	3	3	NR	4
41	NR	NR	NR	4	4	4	4	NR	NR	3
43	3	3	NR	2	4	ND	4	3	ND	4
46	NR	ND	4	0	ND	ND	ND	4	NR	ND
47	3	4	2	ND	4	4	0	3	3	0
48	3	NR	NR	0	2	ND	3	NR	ND	4
49	3	4	1	4	4	3	3	3	ND	3
50	ND	ND	NR	NR	3	0	3	NR	ND	3
51	ND	4	3	ND	ND	4	2	3	ND	4
52	ND	3	ND	ND	2	3	4	ND	ND	4
53	NR	3	NR	ND	4	ND	0	NR	ND	4
54	2	3	4	4	3	4	4	4	3	3
56	2	4	1	0	0	ND	0	NR	ND	4
57	NR	NR	3	0	4	3	4	3	NR	4
58	3	NR	0	2	4	ND	4	NR	ND	3
59	1	4	4	ND	4	4	0	3	ND	2
61	ND	ND	ND	ND	3	ND	1	ND	ND	0
63	4	3	4	0	0	4	0	4	NR	0
68	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
69	2	2	3	0	ND	ND	3	0	ND	0
70	ND	1	ND	ND	ND	ND	4	ND	ND	4
72	ND	ND	ND	ND	3	ND	4	ND	ND	4
73	1	3	4	ND	3	4	4	3	4	3
74	ND	4	ND	0	ND	ND	ND	ND	ND	ND
76	NR	1	NR	4	ND	ND	ND	ND	NR	ND
77	ND	ND	NR	ND	ND	ND	ND	3	ND	ND
79	0	4	ND	ND	ND	ND	ND	3	ND	ND
80	NR	3	NR	4	2	ND	4	NR	ND	4
81	NR	3	4	NR	0	ND	3	NR	ND	2
83	NR	NR	NR	4	3	NR	2	NR	NR	4
84	NR	4	4	4	3	ND	3	NR	ND	0
85	NR	3	NR	ND	ND	ND	ND	NR	ND	ND
86	ND	ND	NR	ND	2	ND	4	NR	ND	0
90	ND	4	NR	2	ND	ND	ND	NR	ND	ND
93	NR	1	2	ND	4	ND	3	NR	NR	4
94	NR	NR	NR	ND	ND	ND	ND	0	ND	0
95	2	3	ND	3	ND	ND	ND	ND	ND	ND
96	3	ND	4	NR	1	ND	0	ND	ND	4
98	NR	3	4	4	4	ND	2	NR	NR	2
102	ND	2	ND	ND	0	ND	2	ND	ND	3
103	NR	NR	NR	ND	4	1	4	NR	NR	4
104	4	2	NR	0	4	ND	4	NR	NR	4
107	3	4	ND	ND	ND	ND	ND	ND	ND	ND
108	NR	3	NR	ND	3	ND	4	NR	2	4
110	0	0	2	ND	4	4	4	4	2	2
111	ND	ND	ND	ND	0	ND	4	ND	ND	4
112	2	4	2	4	4	4	4	3	ND	4
113	2	3	0	ND	3	ND	4	3	ND	ND
118	NR	3	NR	3	ND	ND	4	NR	ND	4
121	3	3	2	4	4	ND	4	3	ND	4
122	4	3	0	3	ND	ND	ND	1	ND	ND
123	ND	ND	4	ND	ND	ND	ND	4	ND	ND
124	NR	NR	NR	ND	ND	ND	0	NR	ND	ND
127	4	4	1	NR	NR	NR	0	NR	4	0
142	NR	4	1	2	3	NR	3	NR	3	4
144	ND	4	ND	ND	ND	ND	ND	ND	ND	ND

Table 4 Standard Reference Water Sample No. T95 (Trace Constituents)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	NI	PB	SB	SE	SI02	SR	TL	V	ZN	N	Avg.
1	NR	NR	NR	3	ND	4	ND	ND	4	13	3.62
2	NR	4	ND	ND	ND	ND	ND	ND	1	11	2.64
4	ND	NR	ND	ND	ND	ND	ND	ND	4	3	2.64
6	NR	NR	ND	NR	ND	4	NR	NR	4	10	2.64
7	NR	NR	NR	4	ND	ND	NR	ND	NR	7	2.64
8	NR	4	4	0	3	4	NR	NR	4	3	2.64
9	0	NR	ND	4	ND	ND	ND	ND	3	3	2.64
10	3	3	ND	ND	ND	ND	ND	ND	ND	5	2.64
13	NR	NR	NR	3	4	4	ND	4	4	6	2.64
15	3	NR	ND	4	4	0	NR	NR	ND	19	2.58
16	ND	3	ND	4	ND	ND	ND	ND	1	12	3.08
18	ND	4	ND	NR	ND	ND	ND	ND	NR	7	2.71
19	ND	ND	ND	ND	ND	ND	ND	ND	ND	1	4.00
20	NR	NR	ND	ND	4	ND	ND	ND	NR	5	2.60
22	NR	NR	NR	NR	1	1	NR	2	NR	12	2.33
24	4	4	ND	NR	ND	2	ND	3	1	21	3.14
25	2	NR	ND	4	4	ND	ND	ND	NR	13	3.08
27	ND	NR	ND	4	ND	ND	ND	ND	ND	6	3.33
29	NR	NR	ND	ND	ND	ND	ND	ND	4	6	2.50
30	4	3	ND	1	4	ND	ND	ND	3	21	2.90
32	3	0	ND	ND	0	ND	ND	ND	0	14	1.71
33	NR	NR	ND	ND	ND	ND	ND	ND	NR	6	2.67
34	4	3	ND	4	ND	3	ND	NR	3	10	3.10
36	ND	ND	ND	NR	0	1	ND	ND	4	13	3.08
37	3	NR	ND	2	ND	ND	ND	ND	3	14	2.07
38	4	3	3	4	3	ND	NR	NR	2	17	2.71
40	4	3	ND	3	3	ND	ND	ND	4	19	3.21
41	NR	4	NR	4	2	0	NR	NR	4	14	2.71
43	3	3	NR	4	ND	ND	NR	ND	2	15	2.80
46	ND	NR	ND	2	ND	ND	ND	ND	ND	9	2.33
47	2	3	0	4	4	4	NR	3	0	24	2.79
48	NR	NR	NR	2	ND	ND	NR	NR	4	10	3.20
49	4	ND	ND	4	4	4	ND	3	4	21	3.00
50	ND	NR	ND	3	ND	ND	ND	ND	ND	6	2.00
51	4	ND	ND	ND	ND	4	ND	ND	3	14	3.43
52	ND	ND	ND	0	4	4	ND	ND	3	12	3.17
53	NR	ND	ND	3	ND	4	ND	ND	4	8	3.25
54	4	4	NR	4	3	2	NR	3	4	22	3.23
56	ND	0	NR	2	ND	ND	ND	ND	4	14	1.79
57	3	NR	NR	3	4	4	NR	NR	0	15	2.87
58	1	NR	NR	3	ND	ND	ND	ND	3	12	2.58
59	4	3	ND	ND	ND	ND	ND	ND	4	15	2.67
61	ND	ND	ND	ND	ND	ND	ND	ND	ND	4	1.50
63	4	NR	2	4	ND	0	NR	NR	4	21	2.29
68	ND	ND	ND	3	ND	ND	ND	ND	ND	1	3.00
69	4	1	3	3	ND	ND	ND	3	2	19	1.84
70	0	ND	ND	0	ND	ND	ND	ND	2	9	1.67
72	ND	ND	ND	ND	ND	ND	ND	ND	4	5	1.60
73	4	0	3	4	0	4	ND	3	3	24	2.71
74	ND	1	ND	2	ND	ND	ND	ND	4	8	2.50
76	NR	4	NR	3	ND	ND	NR	ND	NR	4	3.00
77	ND	ND	ND	ND	ND	ND	ND	ND	ND	1	3.00
79	3	4	ND	ND	ND	ND	ND	ND	ND	6	2.83
80	NR	NR	ND	3	ND	4	ND	NR	4	13	3.62
81	0	NR	ND	4	ND	ND	ND	ND	0	9	2.11
83	NR	NR	NR	NR	4	ND	NR	NR	4	8	3.63
84	3	NR	ND	4	ND	ND	ND	ND	4	12	3.00
85	3	3	ND	ND	ND	ND	ND	ND	4	6	3.00
86	ND	ND	ND	ND	ND	ND	ND	ND	ND	4	2.50
90	4	ND	ND	NR	ND	ND	ND	ND	2	4	3.00
93	NR	NR	ND	1	1	4	NR	NR	4	11	2.45
94	0	NR	ND	ND	ND	ND	ND	ND	2	8	0.63
95	ND	ND	ND	ND	4	ND	ND	ND	4	6	3.17
96	3	1	ND	0	ND	ND	ND	ND	ND	13	2.00
98	NR	NR	NR	2	4	4	NR	NR	4	14	3.21
102	ND	ND	ND	ND	ND	ND	ND	ND	0	8	1.50
103	NR	NR	ND	ND	3	3	ND	NR	4	10	3.40
104	3	4	NR	4	0	2	NR	0	2	18	2.89
107	ND	NR	ND	ND	ND	ND	ND	ND	0	5	2.40
108	NR	ND	ND	3	ND	ND	ND	ND	NR	8	3.38
110	2	0	0	0	1	2	ND	4	0	26	2.04
111	ND	ND	NR	ND	ND	ND	ND	ND	ND	5	3.20
112	3	NR	NR	1	4	4	NR	ND	2	18	3.06
113	2	ND	ND	ND	ND	ND	ND	ND	2	9	2.44
118	NR	NR	ND	ND	ND	ND	ND	ND	4	6	3.50
121	4	4	ND	4	ND	ND	ND	ND	4	15	3.53
122	3	ND	ND	3	ND	ND	ND	ND	0	11	2.36
123	ND	ND	ND	ND	ND	ND	ND	ND	ND	2	4.00
124	NR	NR	ND	ND	ND	ND	ND	ND	4	4	1.50
127	4	3	NR	0	ND	4	NR	NR	4	15	2.73
142	NR	NR	NR	4	4	4	NR	NR	4	15	3.27
144	4	3	ND	ND	ND	4	NR	ND	ND	6	3.50

Table 5 Standard Reference Water Sample No. N16 (Nutrients)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	NH3-N	NO2-N	NO3-N	ORG-N	P, TOTAL	PO4-P	N	Avg.
2	4	4	2	4	4	4	6	3.67
3	ND	ND	ND	ND	4	NR	1	4.00
6	4	4	3	4	3	4	6	3.67
7	1	3	4	2	4	4	6	3.00
8	4	4	3	4	4	3	6	3.67
10	0	ND	4	ND	ND	4	3	2.67
12	1	4	4	ND	4	4	5	3.40
13	3	1	4	4	3	3	6	3.00
14	2	1	0	ND	ND	0	4	0.75
15	ND	2	ND	ND	4	4	3	3.33
20	4	4	0	3	ND	4	5	3.00
22	3	0	3	ND	2	0	5	1.60
23	3	4	3	ND	2	4	5	3.20
24	ND	0	4	ND	ND	4	3	2.67
25	4	ND	1	ND	ND	1	3	2.00
27	2	2	4	ND	ND	1	4	2.25
29	4	0	1	3	1	0	6	1.50
30	4	3	3	2	3	4	6	3.17
32	3	ND	2	ND	ND	ND	2	2.50
33	0	ND	1	ND	0	ND	3	0.33
34	3	1	3	3	3	2	6	2.50
35	4	4	2	4	4	4	6	3.67
36	4	0	3	ND	ND	4	4	2.75
37	3	3	3	3	3	4	6	3.17
38	4	4	4	3	4	2	6	3.50
40	4	1	4	3	4	4	6	3.33
41	3	0	0	ND	3	4	6	2.00
43	4	0	4	3	0	4	6	3.67
44	4	0	ND	ND	0	ND	3	1.33
45	ND	ND	4	ND	ND	ND	1	4.00
46	ND	2	2	ND	ND	0	3	1.33
47	ND	ND	4	ND	2	1	3	2.33
48	4	4	4	2	4	4	6	3.17
49	ND	ND	3	ND	4	3	3	3.33
53	4	4	2	0	3	ND	5	2.60
55	1	0	0	3	0	0	6	0.67
56	1	0	4	ND	0	0	5	1.00
57	4	4	3	4	3	3	6	3.50
58	4	4	2	3	4	3	6	3.33
59	ND	4	3	ND	4	3	4	3.50
65	3	4	3	0	4	4	6	3.00
69	4	ND	ND	ND	0	0	3	1.33
72	4	2	4	ND	4	ND	4	3.50
73	3	ND	ND	3	4	ND	3	3.33
74	0	4	2	2	3	3	6	2.33
76	3	3	4	2	4	4	6	3.33
79	0	0	0	1	4	4	6	1.50
81	4	0	4	ND	3	0	5	2.20
82	4	ND	2	4	4	4	5	3.60
83	4	4	4	0	4	3	6	3.17
84	ND	ND	1	ND	3	4	3	2.67
85	4	4	4	4	4	4	6	4.00
86	3	4	0	4	3	2	6	2.67
91	4	4	2	4	4	3	6	3.50
93	2	4	0	ND	3	4	5	2.60
94	4	ND	ND	ND	ND	ND	1	4.00
95	4	1	2	4	4	4	6	3.17
96	4	ND	2	ND	4	ND	3	3.33
99	1	2	4	4	2	3	6	2.67
104	0	ND	3	ND	3	4	4	2.50
108	0	2	4	ND	0	0	5	1.20
109	3	4	ND	ND	0	0	4	1.75
113	3	3	ND	ND	0	2	4	2.00
121	4	4	3	4	4	ND	5	3.80
124	0	ND	ND	ND	ND	ND	1	0.00
126	ND	ND	4	ND	1	2	2	2.50
128	3	4	3	3	3	2	6	3.00
142	4	4	3	3	3	2	6	3.17

Table 6 Standard Reference Water Sample No. P8 (Precipitation Snowmelt)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	ACID@CACO3	AG	CA	CD	CL	CO	CR TOT	CU	F	FE
1	NR	NR	0	NR	NR	ND	ND	NR	2	NR
2	3	NR	4	3	0	ND	NR	NR	3	4
6	ND	NR	2	NR	4	NR	NR	NR	NR	NR
8	NR	NR	3	NR	ND	NR	NR	NR	4	NR
10	ND	NR	ND	NR	ND	ND	NR	ND	ND	ND
13	4	NR	4	NR	NR	ND	NR	NR	NR	NR
15	ND	NR	NR	3	NR	NR	NR	NR	NR	NR
19	1	ND	3	ND	3	ND	NR	ND	NR	NR
22	ND	NR	0	NR	4	NR	NR	NR	NR	NR
26	ND	ND	4	ND	3	ND	NR	NR	ND	NR
30	3	3	3	3	NR	NR	4	4	3	NR
33	ND	ND	1	ND	ND	ND	ND	ND	ND	ND
37	3	NR	2	NR	0	ND	4	NR	NR	NR
38	3	ND	2	3	NR	ND	NR	NR	3	NR
40	ND	NR	4	NR	NR	ND	3	NR	3	NR
41	0	NR	NR	NR	0	NR	NR	NR	3	NR
43	ND	NR	4	NR	4	ND	2	4	ND	NR
44	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
47	ND	4	4	3	3	NR	NR	3	3	3
49	ND	2	4	4	2	ND	2	1	4	3
50	ND	NR	NR	0	ND	ND	ND	ND	NR	NR
51	ND	ND	4	3	4	ND	ND	3	3	NR
53	4	NR	NR	NR	NR	ND	NR	3	ND	NR
57	NR	NR	3	NR	4	NR	2	3	1	NR
64	3	ND	4	ND	ND	ND	ND	ND	ND	ND
67	ND	ND	1	ND	4	ND	ND	ND	ND	3
71	ND	ND	2	ND	4	ND	ND	ND	ND	ND
72	ND	ND	4	ND	ND	ND	ND	ND	ND	ND
82	ND	ND	ND	ND	1	ND	ND	ND	ND	ND
83	3	NR	NR	NR	NR	NR	NR	NR	NR	NR
86	ND	ND	1	ND	NR	ND	ND	ND	NR	ND
102	2	ND	4	4	2	ND	ND	0	ND	ND
104	3	NR	4	NR	NR	NR	NR	NR	2	NR
109	ND	ND	ND	ND	ND	ND	ND	ND	ND	ND
110	ND	0	0	0	ND	ND	0	0	ND	4
111	ND	ND	3	ND	ND	ND	ND	ND	ND	ND
112	2	ND	1	NR	NR	ND	NR	NR	NR	NR
113	ND	ND	0	ND	3	ND	ND	1	ND	1
118	ND	ND	ND	ND	3	ND	ND	NR	ND	NR
119	ND	ND	ND	ND	4	ND	ND	ND	ND	ND
124	ND	NR	NR	NR	0	NR	NR	NR	ND	NR
125	ND	ND	2	ND	3	ND	ND	ND	ND	ND
142	ND	NR	1	3	3	NR	NR	1	3	2
144	ND	ND	ND	3	ND	ND	ND	4	ND	ND



Table 6 Standard Reference Water Sample No. P8 (Precipitation Snowmelt)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	K	MG	MN	NA	NH3-N	NO3-N	NO3-N	PB	PH	SC4
1	NR	NR	NR	4	NR	NR	NR	4	NR	1
2	ND	NR	4	4	ND	ND	2	4	NR	3
6	3	2	NR	1	3	3	NR	4	3	4
8	4	3	NR	4	4	4	0	4	NR	4
10	1	ND	NR	3	NR	ND	NR	3	ND	2
13	2	4	NR	0	2	ND	NR	4	NR	0
15	NR	NR	ND	NR	NR	ND	NR	3	NR	4
19	2	4	ND	3	ND	ND	ND	ND	4	ND
22	NR	NR	NR	0	NR	4	NR	1	4	4
26	4	4	ND	3	3	4	ND	4	3	4
30	3	3	3	3	3	ND	2	4	NR	3
33	ND	ND	ND	ND	ND	ND	ND	0	0	0
37	2	0	1	4	NR	NR	NR	4	0	4
38	NR	2	NR	NR	3	ND	NR	4	3	2
40	2	4	3	3	NR	ND	0	4	NR	4
41	1	2	NR	3	3	ND	NR	2	0	1
43	3	4	2	2	NR	ND	NR	4	NR	NR
44	ND	ND	ND	ND	0	ND	ND	0	ND	ND
47	4	3	4	4	ND	ND	3	1	C	0
49	ND	4	3	ND	ND	0	ND	ND	4	ND
50	NR	NR	NR	NR	4	ND	NR	2	NR	3
51	ND	3	4	3	ND	0	3	ND	3	ND
53	NR	NR	NR	NR	ND	ND	NR	3	0	0
57	2	3	3	4	1	0	NR	1	3	3
64	NR	4	ND	2	3	3	ND	4	ND	3
67	4	3	1	4	4	3	ND	4	3	4
71	4	2	ND	4	2	4	ND	4	4	2
72	4	3	ND	3	3	ND	ND	ND	ND	3
82	ND	ND	ND	ND	4	ND	ND	ND	ND	4
83	NR	NR	NR	NR	NR	ND	NR	4	0	3
86	ND	ND	NR	ND	0	ND	ND	4	4	4
102	0	3	ND	0	ND	ND	ND	1	ND	4
104	2	4	NR	4	3	ND	NR	3	1	4
109	ND	ND	ND	ND	NR	ND	ND	ND	ND	2
110	0	0	2	0	ND	ND	0	ND	ND	ND
111	ND	0	ND	ND	ND	ND	ND	0	ND	2
112	3	1	NR	4	NR	ND	NR	4	NR	ND
113	2	1	0	ND	ND	ND	ND	4	1	ND
118	ND	ND	NR	NR	ND	3	ND	4	4	3
119	ND	ND	ND	ND	ND	1	ND	ND	3	ND
124	ND	NR	NR	ND	ND	ND	NR	ND	ND	ND
125	4	4	ND	4	ND	ND	ND	ND	3	3
142	NR	3	4	0	NR	ND	NR	1	3	3
144	ND	ND	ND	ND	ND	ND	3	ND	ND	ND

Table 6 Standard Reference Water Sample No. P8 (Precipitation Snowmelt)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev.  
2 (Satisfactory) 1.01 to 1.50 Std. Dev.  
1 (Questionable) 1.51 to 2.00 Std. Dev.

0 (Poor) > 2.00 Std. Dev.  
ND Not determined  
NR Not rated

LAB	SP. COND.	TL	N	Avg.
1	ND	NR	5	2.00
2	ND	3	12	3.08
6	NR	3	11	2.91
8	NR	NR	10	3.40
10	ND	ND	4	2.25
13	NR	3	9	2.78
15	NR	ND	4	2.25
19	ND	ND	7	2.86
22	NR	NR	7	2.43
26	ND	ND	10	3.60
30	ND	3	16	3.13
33	ND	ND	4	0.25
37	ND	3	12	2.25
38	NR	NR	9	2.78
40	ND	4	11	3.09
41	NR	NR	10	1.60
43	NR	3	10	3.30
44	ND	ND	2	0.00
47	ND	0	16	2.63
49	ND	3	13	2.77
50	ND	ND	4	2.25
51	ND	4	12	3.08
53	ND	NR	5	2.00
57	NR	2	15	2.33
64	ND	ND	8	3.25
67	ND	3	13	3.15
71	ND	ND	10	3.30
72	ND	ND	6	3.17
82	ND	ND	3	3.00
83	NR	0	5	2.00
86	ND	ND	5	2.60
102	ND	3	11	2.09
104	NR	3	11	3.00
109	ND	ND	1	2.00
110	ND	4	12	0.83
111	ND	ND	4	1.25
112	NR	NR	6	2.50
113	ND	2	9	1.33
118	ND	NR	5	3.40
119	ND	ND	3	2.67
124	ND	NR	1	0.00
125	ND	ND	6	3.33
142	NR	4	13	2.38
144	ND	ND	3	3.33

Table 7 Standard Reference Water Sample No. SED3 (Bed Material)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	AL	B	BA	BE	C, INORG	C, TOTAL	CA	CD	CO	CR TOT
8	3	ND	4	NR	ND	ND	4	NR	4	4
13	4	ND	3	4	ND	ND	4	NR	NR	2
20	1	ND	4	3	ND	ND	NR	ND	ND	3
22	4	NR	3	NR	ND	ND	4	NR	3	0
30	1	ND	1	ND	ND	ND	4	3	1	2
33	ND	ND	ND	ND	ND	ND	ND	2	ND	2
37	1	ND	2	3	ND	ND	3	NR	ND	3
38	4	2	0	NR	ND	ND	1	NR	ND	4
40	4	ND	3	ND	ND	ND	4	NR	ND	1
41	0	ND	0	0	ND	ND	0	4	0	1
43	3	ND	0	3	ND	ND	4	4	ND	3
50	ND	ND	ND	0	ND	ND	0	NR	ND	ND
59	ND	ND	ND	ND	ND	ND	1	3	3	3
80	4	ND	4	4	ND	ND	4	3	ND	4
84	2	ND	2	NR	ND	ND	4	NR	ND	4
104	4	2	0	NR	ND	ND	2	NR	1	4
112	4	2	3	2	ND	ND	2	NR	3	4
127	ND	NR	3	4	ND	ND	4	NR	3	4
135	4	4	3	2	ND	ND	4	1	4	4
142	2	ND	4	4	ND	ND	4	2	4	3
144	4	4	4	4	ND	ND	4	4	0	4

Table 7 Standard Reference Water Sample No. SED3 (Bed Material)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	CU	FE	HG	K	LI	MG	MN	MO	NA	NI
8	NR	4	NR	3	3	4	4	NR	4	3
13	2	3	NR	3	3	4	3	NR	4	3
20	4	1	NR	4	ND	4	ND	ND	3	4
22	4	4	NR	4	3	4	3	NR	2	3
30	4	3	3	2	ND	3	2	ND	4	2
33	3	3	0	2	ND	0	0	ND	4	4
37	2	3	3	3	ND	3	0	ND	1	2
38	1	2	NR	ND	ND	0	2	NR	ND	3
40	2	3	3	4	ND	4	4	ND	1	4
41	1	0	ND	0	1	0	4	NR	2	0
43	3	2	ND	2	ND	3	2	ND	4	4
50	ND	0	0	0	0	0	4	ND	4	ND
59	3	4	ND	4	3	3	2	ND	1	3
80	4	2	3	4	ND	4	4	ND	2	3
84	2	2	0	0	ND	2	4	ND	NR	1
104	3	ND	NR	0	ND	1	4	3	1	3
112	4	4	NR	4	4	4	3	4	4	4
127	4	0	NR	ND	NR	4	1	2	3	4
135	3	4	NR	4	4	4	2	NR	4	3
142	2	4	NR	4	1	4	4	ND	3	3
144	4	4	ND	ND	4	4	4	ND	4	4

Table 7 Standard Reference Water Sample No. SED3 (Bed Material)  
Overall Laboratory Performance

RATING 4 (Excellent) 0.00 to 0.50 Std. Dev. 0 (Poor) > 2.00 Std. Dev.  
3 (Good) 0.51 to 1.00 Std. Dev. ND Not determined  
2 (Satisfactory) 1.01 to 1.50 Std. Dev. NR Not rated  
1 (Questionable) 1.51 to 2.00 Std. Dev.

LAB	PB	SE	SR	V	ZN	N	Avg.
8	0	NR	4	3	3	16	3.38
13	NR	NR	3	3	4	16	3.25
20	3	ND	ND	ND	4	12	3.17
22	3	0	4	4	3	18	3.06
30	4	ND	ND	ND	1	16	2.50
33	3	ND	ND	ND	1	12	2.00
37	3	3	ND	ND	1	16	2.25
38	4	NR	ND	1	3	13	2.08
40	2	3	ND	ND	4	15	3.07
41	0	3	1	1	1	20	0.95
43	3	NR	ND	ND	3	15	2.87
50	0	0	ND	ND	ND	11	0.73
59	1	ND	ND	ND	3	14	2.64
80	4	3	3	3	2	19	3.42
84	NR	NR	ND	ND	1	12	2.25
104	3	3	0	4	4	17	2.24
112	2	1	4	2	4	21	3.19
127	4	NR	2	4	3	16	3.06
135	1	4	1	NR	3	20	3.15
142	3	NR	4	3	3	19	3.21
144	2	ND	3	4	4	19	3.63

Table 8 Standard Reference Water Sample M6 Report for ALK(CAC03)

Code Number	Reported value	Pct. dev. from mean	Methods	Reference
1	28	10.0	'TITRATION, ELECTROMETRIC'	4
2	20	-21.4	'TITRATION, ELECTROMETRIC'	4
4	26	2.2	'TITRATION, COLORIMETRIC'	3
6	25	-1.7	'TITRATION, ELECTROMETRIC'	4
7	28	10.0	'TITRATION, ELECTROMETRIC'	4
8	24	-5.7	'TITRATION, ELECTROMETRIC'	4
9	25	-1.7	OTHER	
10	26	2.2	'TITRATION, COLORIMETRIC'	3
13	26	2.2	'TITRATION, ELECTROMETRIC'	4
15	26	2.2	'TITRATION, COLORIMETRIC'	3
16	25	-1.7	'TITRATION, ELECTROMETRIC'	4
19	30	17.9	'TITRATION, ELECTROMETRIC'	4
20	52	104.4	'TITRATION, ELECTROMETRIC'	4
22	13	-48.9	REJECT	3
24	25	-1.7	REJECT	4
25	28	10.0	'TITRATION, ELECTROMETRIC'	4
27	28	10.0	'TITRATION, COLORIMETRIC'	3
30	27	6.1	'TITRATION, COLORIMETRIC'	3
33	24	-5.7	'TITRATION, ELECTROMETRIC'	4
34	25	-1.7	NOT REPORTED	
36	25	-1.7	'TITRATION, ELECTROMETRIC'	4
37	24	-5.7	'TITRATION, ELECTROMETRIC'	4
38	24	-5.7	'TITRATION, COLORIMETRIC'	3
40	26	2.2	'TITRATION, ELECTROMETRIC'	4
41	28	10.0	'TITRATION, COLORIMETRIC'	3
43	25	-1.7	'TITRATION, ELECTROMETRIC'	4
44	24	-5.7	'TITRATION, COLORIMETRIC'	3
45	26	2.2	'TITRATION, ELECTROMETRIC'	4
46	28	10.0	'TITRATION, ELECTROMETRIC'	4
48	28	10.0	'TITRATION, ELECTROMETRIC'	4
49	50	96.5	REJECT	4
50	23	-9.6	'TITRATION, ELECTROMETRIC'	4
52	25	-1.7	'TITRATION, ELECTROMETRIC'	4
53	25	-1.7	'TITRATION, ELECTROMETRIC'	4
56	26	2.2	'TITRATION, ELECTROMETRIC'	4
57	25	-1.7	'TITRATION, ELECTROMETRIC'	4
58	24	-5.7	'TITRATION, ELECTROMETRIC'	4
63	24	-5.7	'TITRATION, ELECTROMETRIC'	4
64	26	2.2	'TITRATION, COLORIMETRIC'	3
67	23	-9.6	'TITRATION, COLORIMETRIC'	3
69	31	21.8	NOT REPORTED	
70	27	6.1	'TITRATION, COLORIMETRIC'	3
72	25	-1.7	'TITRATION, ELECTROMETRIC'	4
73	24	-5.7	OTHER	
75	24	-5.7	'TITRATION, COLORIMETRIC'	3
77	26	2.2	'TITRATION, ELECTROMETRIC'	4
79	23	-9.6	'TITRATION, ELECTROMETRIC'	4
80	26	2.2	OTHER	
81	24	-5.7	'TITRATION, ELECTROMETRIC'	4
83	24	-5.7	'TITRATION, COLORIMETRIC'	3
84	24	-5.7	'TITRATION, ELECTROMETRIC'	4
85	24	-5.7	'TITRATION, ELECTROMETRIC'	4
86	29	14.0	'TITRATION, COLORIMETRIC'	3
91	22	-13.5	'TITRATION, ELECTROMETRIC'	4
93	24	-5.7	'TITRATION, COLORIMETRIC'	3
94	38	49.3	REJECT	4
96	23	-9.6	'TITRATION, ELECTROMETRIC'	4
98	26	2.2	'TITRATION, ELECTROMETRIC'	4
99	29	14.0	'TITRATION, COLORIMETRIC'	3
102	25	-1.7	'TITRATION, ELECTROMETRIC'	4
103	23	-9.6	'TITRATION, ELECTROMETRIC'	4
104	30	17.9	'TITRATION, COLORIMETRIC'	3
112	26	2.2	'TITRATION, ELECTROMETRIC'	4
113	16	-37.1	REJECT	4
118	26	2.2	'TITRATION, ELECTROMETRIC'	4
123	29	14.0	'TITRATION, ELECTROMETRIC'	4
124	20	-21.4	NOT REPORTED	
142	24	-5.7	'TITRATION, ELECTROMETRIC'	4

68 Labs had a total range of 13 to 52 and a mean of 25.4 with a standard deviation of 2.2 and a 95% confidence interval of the mean +/- 0.6.

Table 8 Standard Reference Water Sample M6 Report for B

Code Number	Reported value	Pct. dev. from mean	Methods	References	
1	< 10		IGNORED	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
2	< 20	-42.4	IGNORED	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
6	< 60	72.8	IGNORED	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
8	< 48	38.2	IGNORED	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
9	< 45	29.6	IGNORED	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
13	< 250		IGNORED	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
14	< 20	-42.4	IGNORED	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
15	< 50		IGNORED	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
22	< 100		IGNORED	EMISSION, IC PLASMA	3
25	< 51	46.9	IGNORED	EMISSION, DC PLASMA	7
34	< 0		IGNORED	COLORIMETRIC, AZOMETHINE	5
38	< 29	-16.5	IGNORED	EMISSION, IC PLASMA	3
40	< 20	-42.4	IGNORED	EMISSION, IC PLASMA	3
41	< 20		IGNORED	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
46	< 0		IGNORED	COLORIMETRIC, CARMINE (CARMINIC ACID)	2, 4
47	< 230	562.3	REJECT	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	7
48	< 50		IGNORED	EMISSION, IC PLASMA	3
49	< 23	-33.8	IGNORED	EMISSION, IC PLASMA	3
51	< 18	-48.2	IGNORED	EMISSION, IC PLASMA	3
52	< 23	-33.8	IGNORED	EMISSION, IC PLASMA	3
57	< 21	-39.5	IGNORED	EMISSION, DC PLASMA	7
63	< 40		IGNORED	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	7
67	< 13	-62.6	IGNORED	EMISSION, IC PLASMA	3
73	< 41	18.1	IGNORED	EMISSION, IC PLASMA	3
80	< 75	116.0	IGNORED	EMISSION, IC PLASMA	3
83	< 2000		IGNORED	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	7
85	< 20		IGNORED	COLORIMETRIC, DIANTHRIMIDE	4
93	< 137	294.5	REJECT	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
98	< 67	92.9	IGNORED	EMISSION, IC PLASMA	3
99	< 80	130.4	IGNORED	COLORIMETRIC, AZOMETHINE	5
103	< 3		IGNORED	EMISSION, IC PLASMA	3
108	< 10	-71.2	IGNORED	COLORIMETRIC, AZOMETHINE	5
110	< 29	-16.5	IGNORED	EMISSION, IC PLASMA	3
111	< 21	-39.5	IGNORED	EMISSION, IC PLASMA	3
112	< 0		IGNORED	EMISSION, DC PLASMA	7
123	< 10	-71.2	IGNORED	COLORIMETRIC, CARMINE (CARMINIC ACID)	2, 4
127	< 20		IGNORED	OTHER	3
142	< 40	15.2	IGNORED	EMISSION, IC PLASMA	3

38 Labs had a total range of 10 to 230 and a mean of 34.7  
with a standard deviation of 21.1 and a 95% confidence interval of the mean +/- 9.3.

Table 8 Standard Reference Water Sample M6 Report for BR

Code Number	Reported value	Pct. dev. from mean	Methods	References	
1	< 100		IGNORED	COLORIMETRIC, CATALYTIC OXIDATION	2, 4
6	< 50		IGNORED	ION CHROMATOGRAPHY	2, 3, 6
19	< 8	-38.5	REJECT	ION CHROMATOGRAPHY	2, 3, 6
22	< 200		IGNORED	ION CHROMATOGRAPHY	2, 3, 6
38	< 20		IGNORED	ION CHROMATOGRAPHY	2, 3, 6
51	< 18	38.5	REJECT	ION CHROMATOGRAPHY	2, 3, 6
57	< 5		IGNORED	COLORIMETRIC, CHLORAMINE-T	1
127	< 100		IGNORED	OTHER	
142	< 200		IGNORED	ION CHROMATOGRAPHY	2, 3, 6

9 Labs had a total range of 8 to 18.  
INSUFFICIENT DATA TO DEFINE MEAN AND STANDARD OF DEVIATION.

Table 8 Standard Reference Water Sample M6 Report for CA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	26	0.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	26	0.7	TITRATION, EDTA	1,3
6	25	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	26	0.7	EMISSION, IC PLASMA	3,5,7
8	26	0.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
9	29	12.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	24	-7.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	26	0.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
17	18	-30.3	REJECT	1,2,3,4
19	32	23.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
20	28	8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	28	8.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
24	24	-7.0	EMISSION, IC PLASMA	3,5,7
25	27	4.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
27	24	-7.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	23	-10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	25	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	24	-7.0	NOT REPORTED	1,2,3,4
34	26	0.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
36	25	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	21	-18.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	27	4.6	EMISSION, IC PLASMA	3,5,7
40	26	0.7	EMISSION, IC PLASMA	3,5,7
41	26	0.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
42	25	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	28	8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
46	27	4.6	TITRATION, EDTA	1,3
47	24	-7.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	28	8.5	EMISSION, IC PLASMA	3,5,7
49	25	-3.2	EMISSION, IC PLASMA	3,5,7
50	22	-14.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	28	8.5	EMISSION, IC PLASMA	3,5,7
52	25	-3.2	EMISSION, IC PLASMA	3,5,7
53	28	8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	44	70.4	REJECT	1,3
57	25	-3.2	TITRATION, EDTA	3,5,7
63	25	-3.2	EMISSION, IC PLASMA	3,5,7
64	22	-14.8	TITRATION, EDTA	1,3
67	30	16.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	26	0.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	25	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	25	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
76	25	-3.2	EMISSION, IC PLASMA	3,5,7
77	27	4.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	25	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
81	26	0.7	EMISSION, IC PLASMA	3,5,7
83	25	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	30	16.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
85	27	4.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	26	0.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	23	-10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	25	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	25	-3.2	OTHER	1,7
98	24	-7.0	EMISSION, IC PLASMA	3,5,7
99	30	16.2	TITRATION, EDTA	1,3
102	23	-10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
103	27	4.6	EMISSION, IC PLASMA	3,5,7
104	257	895.4	REJECT	1,7
108	26	0.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
110	26	0.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
111	26	0.7	EMISSION, IC PLASMA	3,5,7
112	25	-3.2	EMISSION, IC PLASMA	3,5,7
113	26	0.7	OTHER	1,2,3,4
118	26	0.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	26	0.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
122	29	12.3	OTHER	1,7
123	26	0.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	22	-14.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
142	26	0.7	OTHER	1,7
			EMISSION, IC PLASMA	3,5,7

69 Labs had a total range of 18 to 257 and a mean of 25.8 with a standard deviation of 2.0 and a 95% confidence interval of the mean +/- 0.5.

Table 8 Standard Reference Water Sample M6 Report for CL

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	13.0	-2.2	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
6	13.3	0.0	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
7	13.0	-2.2	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
8	10.8	-18.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
9	13.5	1.5	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
13	13.0	-2.2	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
14	13.8	3.8	TITRATION, MERCURIC NITRATE	1,2,3,4
15	12.0	-9.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
16	13.3	0.0	TITRATION, MERCURIC NITRATE	1,2,3,4
19	6.4	-51.9	ION CHROMATOGRAPHY	2,3,6,7
20	10.0	-24.8	TITRATION, SILVER NITRATE	1,2,4,7
22	16.4	23.3	ION CHROMATOGRAPHY	2,3,6,7
24	13.6	2.3	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
25	14.0	5.3	TITRATION, SILVER NITRATE	1,2,4
27	13.0	-2.2	TITRATION, SILVER NITRATE	1,2,4
29	15.0	12.8	TITRATION, SILVER NITRATE	1,2,4
30	13.2	-0.7	TITRATION, MERCURIC NITRATE	1,2,3,4
33	14.0	5.3	NOT REPORTED	
34	14.6	9.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
36	13.7	3.0	ION CHROMATOGRAPHY	2,3,6,7
37	13.7	3.0	ION SELECTIVE ELECTRODE	1,2,3,4
38	13.0	-8.2	ION CHROMATOGRAPHY	2,3,6,7
40	12.0	-9.8	TITRATION, SILVER NITRATE	1,2,4
41	12.0	-2.2	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
42	13.0	-2.2	TITRATION, SILVER NITRATE	1,2,4
43	14.0	5.3	TITRATION, MERCURIC NITRATE	1,2,3,4
46	13.1	-1.5	TITRATION, MERCURIC NITRATE	1,2,3,4
47	15.3	15.1	ION CHROMATOGRAPHY	2,3,6,7
48	13.0	-2.2	TITRATION, MERCURIC NITRATE	1,2,3,4
49	13.2	-0.7	ION CHROMATOGRAPHY	2,3,6,7
51	12.9	-3.0	ION CHROMATOGRAPHY	2,3,6,7
52	13.0	-2.2	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
53	14.0	5.3	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
56	14.1	6.0	TITRATION, SILVER NITRATE	1,2,4
57	14.0	5.3	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
58	12.0	-9.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
63	13.0	-2.2	ION CHROMATOGRAPHY	2,3,6,7
65	13.0	-2.2	TITRATION, MERCURIC NITRATE	1,2,3,4
67	13.5	1.5	ION CHROMATOGRAPHY	2,3,6,7
70	14.8	11.3	TITRATION, MERCURIC NITRATE	1,2,3,4
72	14.0	5.3	TITRATION, SILVER NITRATE	1,2,4
73	12.0	-9.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
74	12.9	-3.0	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
76	14.0	5.3	TITRATION, SILVER NITRATE	1,2,4
77	13.0	-2.2	TITRATION, SILVER NITRATE	1,2,4
79	14.0	5.3	TITRATION, MERCURIC NITRATE	1,2,3,4
80	12.0	-9.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
81	12.0	-9.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
82	17.5	31.6	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
83	13.0	-2.2	TITRATION, SILVER NITRATE	1,2,4
84	13.5	1.5	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
85	12.0	-9.8	TITRATION, MERCURIC NITRATE	1,2,3,4
86	13.0	-2.2	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
91	11.0	-17.3	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
93	13.4	0.8	TITRATION, SILVER NITRATE	1,2,4
95	17.9	34.6	TITRATION, SILVER NITRATE	1,2,4
96	11.9	-10.5	ION SELECTIVE ELECTRODE	1,2,3,4
98	13.0	-2.2	TITRATION, MERCURIC NITRATE	1,2,3,4
99	16.0	20.3	TITRATION, MERCURIC NITRATE	1,2,3,4
102	14.4	8.3	TITRATION, MERCURIC NITRATE	1,2,3,4
103	13.0	-2.2	ION CHROMATOGRAPHY	2,3,6,7
104	14.0	5.3	TITRATION, MERCURIC NITRATE	1,2,3,4
108	13.0	-2.2	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
112	14.0	5.3	TITRATION, MERCURIC NITRATE	1,2,3,4
118	13.0	-2.2	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
121	13.4	0.8	TITRATION, SILVER NITRATE	1,2,4
122	14.6	9.8	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
123	12.0	-9.8	TITRATION, SILVER NITRATE	1,2,4
124	14.0	5.3	NOT REPORTED	
142	14.0	5.3	ION CHROMATOGRAPHY	2,3,6,7

70 Labs had a total range of 6.4 to 17.9 and a mean of 13.30 with a standard deviation of 1.09 and a 95% confidence interval of the mean +/- 0.27.

Table 8 Standard Reference Water Sample M6 Report for DSRD 180

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	178	-0.3	OTHER	
2	164	-8.1	RESIDUE, FILTRABLE	1, 3
4	170	-4.7	RESIDUE, FILTRABLE	1, 3
6	155	-13.2	RESIDUE, FILTRABLE	1, 3
7	169	-5.3	RESIDUE, FILTRABLE	1, 3
8	182	2.0	RESIDUE, FILTRABLE	1, 3
13	238	33.4	RESIDUE, FILTRABLE	1, 3
14	174	-2.5	RESIDUE ON EVAPORATION	2, 4
15	170	-4.7	RESIDUE, FILTRABLE	1, 3
19	185	3.7	RESIDUE ON EVAPORATION	2, 4
20	160	-10.4	RESIDUE, FILTRABLE	1, 3
22	184	3.1	RESIDUE, FILTRABLE	1, 3
24	167	-6.4	NOT REPORTED	
25	190	6.5	RESIDUE ON EVAPORATION	2, 4
27	159	-10.9	RESIDUE ON EVAPORATION	2, 4
29	169	-5.3	RESIDUE, FILTRABLE	1, 3
30	174	-2.5	RESIDUE, FILTRABLE	1, 3
33	156	-12.6	NOT REPORTED	
34	180	0.9	RESIDUE ON EVAPORATION	2, 4
36	167	-6.4	RESIDUE, FILTRABLE	1, 3
37	160	-10.4	RESIDUE, FILTRABLE	1, 3
38	170	-4.7	RESIDUE, FILTRABLE	1, 3
40	166	-7.0	RESIDUE ON EVAPORATION	2, 4
41	222	24.4	RESIDUE ON EVAPORATION	2, 4
43	148	-17.1	RESIDUE ON EVAPORATION	2, 4
45	184	3.1	RESIDUE ON EVAPORATION	2, 4
46	163	-8.7	RESIDUE, FILTRABLE	1, 3
47	190	6.5	RESIDUE, FILTRABLE	1, 3
48	159	-10.9	RESIDUE, FILTRABLE	1, 3
50	125	-30.0	NOT REPORTED	
53	183	2.5	RESIDUE ON EVAPORATION	2, 4
56	188	5.3	RESIDUE ON EVAPORATION	2, 4
57	166	-7.0	RESIDUE ON EVAPORATION	2, 4
58	196	9.8	RESIDUE, FILTRABLE	1, 3
63	168	-5.9	RESIDUE ON EVAPORATION	2, 4
64	180	0.9	RESIDUE ON EVAPORATION	2, 4
70	170	-4.7	RESIDUE, FILTRABLE	1, 3
72	17	-90.5	REJECT	
76	160	-10.4	RESIDUE, FILTRABLE	2, 4
77	180	0.9	RESIDUE, FILTRABLE	1, 3
79	180	0.9	RESIDUE ON EVAPORATION	2, 4
80	203	0.9	RESIDUE ON EVAPORATION	2, 4
81	163	-8.7	RESIDUE ON EVAPORATION	2, 4
83	207	16.0	RESIDUE, FILTRABLE	1, 3
84	231	29.4	RESIDUE, FILTRABLE	1, 3
85	175	-1.9	RESIDUE ON EVAPORATION	2, 4
91	196	9.8	RESIDUE ON EVAPORATION	2, 4
93	182	2.0	RESIDUE, FILTRABLE	1, 3
94	151	-15.4	RESIDUE, FILTRABLE	1, 3
98	160	-10.4	RESIDUE, FILTRABLE	1, 3
99	210	17.7	RESIDUE ON EVAPORATION	2, 4
102	227	27.2	RESIDUE ON EVAPORATION	2, 4
104	208	16.5	RESIDUE ON EVAPORATION	2, 4
108	178	-0.3	RESIDUE ON EVAPORATION	2, 4
112	189	5.9	RESIDUE, FILTRABLE	1, 3
118	188	5.3	RESIDUE, FILTRABLE	1, 3
121	196	9.8	RESIDUE, FILTRABLE	1, 3
122	162	-9.2	RESIDUE, FILTRABLE	1, 3
123	174	-2.5	RESIDUE ON EVAPORATION	2, 4
142	181	1.4	RESIDUE, FILTRABLE	1, 3

60 Labs had a total range of 17 to 238 and a mean of 178.5 with a standard deviation of 20.9 and a 95% confidence interval of the mean +/- 5.5.



Table 8 Standard Reference Water Sample M6 Report for F

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
2	0.8	-5.7	OTHER	1,2,3,4
4	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
6	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
7	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
8	0.9	6.1	COLORIMETRIC, CEROUS ALIZARIN "COMPLEXONE"	3
9	0.7	-17.5	ION SELECTIVE ELECTRODE	1,2,3,4
10	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
13	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
15	0.9	6.1	OTHER	1,2,3,4
16	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
19	0.4	-52.8	REJECT ION CHROMATOGRAPHY	2,3,6
20	1.2	41.5	REJECT ION SELECTIVE ELECTRODE	1,2,3,4
22	0.8	-5.7	ION CHROMATOGRAPHY	2,3,6
24	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
25	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
27	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
29	0.9	6.1	OTHER	1,2,3,4
30	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
33	0.8	-5.7	NOT REPORTED	1,2,3,4
34	0.8	-5.7	ION SELECTIVE ELECTRODE	2,3,6
36	0.9	6.1	ION CHROMATOGRAPHY	1,2,3,4
37	0.8	-5.7	ION SELECTIVE ELECTRODE	2,3,6
38	0.9	6.1	ION CHROMATOGRAPHY	1,2,3,4
40	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
41	0.9	6.1	COLORIMETRIC, LANTHANUM ALIZARIN "COMPLEXONE"	1
46	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
47	0.9	6.1	ION CHROMATOGRAPHY	2,3,6
48	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
49	0.8	-5.7	ION CHROMATOGRAPHY	2,3,6
50	0.9	6.1	NOT REPORTED	1,2,3,4
51	0.8	-5.7	ION CHROMATOGRAPHY	2,3,6
56	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
57	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
58	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
63	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
70	1.0	17.9	COLORIMETRIC, SPADNS	1,2,3
76	1.1	29.7	REJECT ION SELECTIVE ELECTRODE	1,2,3,4
77	0.9	6.1	COLORIMETRIC, ZIRCONIUM ERIOCHROME	4
80	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
81	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
83	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
85	0.8	-5.7	COLORIMETRIC, ZIRCONIUM ERIOCHROME	4
86	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
91	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
93	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
96	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
98	0.9	6.1	COLORIMETRIC, LANTHANUM ALIZARIN "COMPLEXONE"	1
99	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
103	0.9	6.1	ION CHROMATOGRAPHY	2,3,6
104	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
112	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4
113	0.7	-17.5	ION SELECTIVE ELECTRODE	1,2,3,4
123	0.9	6.1	ION SELECTIVE ELECTRODE	1,2,3,4
142	0.8	-5.7	ION SELECTIVE ELECTRODE	1,2,3,4

55 Labs had a total range of 0.4 to 1.2 and a mean of 0.85 with a standard deviation of 0.06 and a 95% confidence interval of the mean +/- 0.02.

Table 8 Standard Reference Water Sample M6 Report for I

Code Number	Reported value	Pct. dev. from mean	Methods	References
22	< 1000		IGNORED ION CHROMATOGRAPHY	
57	0		COLORIMETRIC, CERIC ARSENIOS OXIDATION	2,4
127	< 100		IGNORED OTHER	

3 Labs had a total range of 0 to 0. INSUFFICIENT DATA FOR DETERMINATION OF MEAN AND STANDARD OF DEVIATION.

Table 8 Standard Reference Water Sample M6 Report for K

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.70	-17.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	0.77	-9.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	0.87	-2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	0.80	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	0.86	0.9	EMISSION, IC PLASMA	3
9	0.75	-12.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	0.78	-8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	0.80	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
17	0.96	12.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	0.79	-7.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	0.76	-10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	1.00	17.3	EMISSION, IC PLASMA	3
24	0.90	5.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	1.00	17.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
27	0.79	-7.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	0.90	5.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	0.86	0.9	EMISSION, FLAME, PHOTOMETRIC	1,2
33	0.90	5.5	NOT REPORTED	
34	0.83	-2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
36	0.75	-12.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	0.68	-20.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	0.90	5.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
40	0.90	5.5	EMISSION, IC PLASMA	3
41	0.90	5.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
42	0.76	-10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	0.80	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
46	1.17	37.2	EMISSION, FLAME, PHOTOMETRIC	1,2
47	0.83	-2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	0.56	-34.3	EMISSION, FLAME, PHOTOMETRIC	1,2
49	0.75	-12.0	EMISSION, IC PLASMA	3
50	0.73	-14.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
52	0.88	3.2	EMISSION, IC PLASMA	3
53	1.00	17.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	1.49	74.7	REJECT EMISSION, FLAME, PHOTOMETRIC	1,2
57	0.85	-0.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
63	3.34	291.7	REJECT EMISSION, FLAME, PHOTOMETRIC	1,2
64	0.88	3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
67	0.92	7.9	EMISSION, FLAME, PHOTOMETRIC	1,2
72	0.75	-12.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	0.86	0.9	EMISSION, IC PLASMA	3
76	1.00	17.3	EMISSION, FLAME, PHOTOMETRIC	1,2
77	0.90	5.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	0.70	-17.9	OTHER	
81	1.60	87.6	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	1.10	29.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	0.89	4.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
85	0.80	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	0.80	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	0.70	-17.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	0.87	2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	0.90	5.5	ION SELECTIVE ELECTRODE	1,2,3,4
98	0.86	0.9	EMISSION, IC PLASMA	3
99	1.00	17.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
102	1.00	17.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
103	1.16	36.0	EMISSION, IC PLASMA	3
104	0.84	-1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
108	0.80	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	2.00	134.5	REJECT EMISSION, IC PLASMA	3
112	0.98	14.9	EMISSION, IC PLASMA	3
113	0.74	-13.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	< 1.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	0.82	-3.8	OTHER	
122	2.35	175.6	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
123	0.80	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
142	0.76	-10.9	EMISSION, IC PLASMA	3

65 Labs had a total range of 0.56 to 3.34 and a mean of 0.853 with a standard deviation 0.115 and a 95% confidence interval of the mean +/- 0.030.

Table 8 Standard Reference Water Sample M6 Report for MG

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	11	5.7	EMISSION, IC PLASMA	3,5
8	10	-3.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
9	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
17	11	5.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
19	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	7	-32.7	REJECT EMISSION, IC PLASMA	3,5
24	9	-13.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
27	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	24	130.6	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	10	-3.9	NOT REPORTED	
34	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
36	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	11	5.7	EMISSION, IC PLASMA	3,5
40	10	-3.9	EMISSION, IC PLASMA	3,5
41	10	-3.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
42	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
46	10	-3.9	TITRATION, EDTA	2
47	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	11	5.7	EMISSION, IC PLASMA	3,5
49	10	-3.9	EMISSION, IC PLASMA	3,5
50	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	11	5.7	EMISSION, IC PLASMA	3,5
52	10	-3.9	EMISSION, IC PLASMA	3,5
53	12	15.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	16	53.8	REJECT TITRATION, EDTA	2
57	11	5.7	EMISSION, IC PLASMA	3,5
63	12	15.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
64	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
67	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	10	-3.9	EMISSION, IC PLASMA	3,5
76	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	10	-3.9	EMISSION, IC PLASMA	3,5
81	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	11	5.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
84	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
85	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	9	-13.5	OTHER	
98	11	5.7	EMISSION, IC PLASMA	3,5
99	13	24.9	REJECT TITRATION, EDTA	2
102	9	-13.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
103	11	5.7	EMISSION, IC PLASMA	3,5
104	11	5.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
108	10	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	11	5.7	EMISSION, IC PLASMA	3,5
111	11	5.7	EMISSION, IC PLASMA	3,5
112	10	-3.9	OTHER	
113	9	-13.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	10	-3.9	TITRATION, EDTA	2
122	8	-23.1	REJECT EMISSION, IC PLASMA	3,5
123	11	5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	10	-3.9	OTHER	
142	11	5.7	EMISSION, IC PLASMA	3,5

69 Labs had a total range of 7 to 24 and a mean of 10.4 with a standard deviation of 0.7 and a 95% confidence interval of the mean +/- 0.2.

Table 8 Standard Reference Water Sample M6 Report for NA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	8.3	-0.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
4	8.0	-3.8	EMISSION, FLAME	1, 2, 3, 4
6	8.4	1.0	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
7	8.0	-3.8	EMISSION, IC PLASMA	3, 5
8	8.3	-0.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
9	8.1	-2.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
13	7.6	-8.6	EMISSION, IC PLASMA	3, 5
15	8.1	-2.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
17	9.1	9.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
19	6.3	-24.2	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
20	9.1	9.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
22	1.0	-88.0	REJECT EMISSION, IC PLASMA	3, 5
24	8.2	-1.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
25	7.0	-15.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
27	8.2	-1.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
29	8.8	5.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
30	7.8	-6.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
33	7.0	-15.8	NOT REPORTED	1, 2, 3, 4
34	8.8	5.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
36	9.5	14.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
37	7.3	-12.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
38	8.6	3.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
40	8.1	-2.6	EMISSION, IC PLASMA	3, 5
41	7.9	-5.0	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
42	8.3	-0.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
43	8.7	4.6	EMISSION, FLAME	1, 2, 3, 4
46	9.2	10.6	EMISSION, FLAME	1, 2, 3, 4
47	7.9	-5.0	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
48	8.8	5.8	EMISSION, IC PLASMA	3, 5
49	8.2	-1.4	EMISSION, IC PLASMA	3, 5
50	8.1	-2.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
51	8.4	1.0	EMISSION, IC PLASMA	3, 5
52	8.2	-1.4	EMISSION, IC PLASMA	3, 5
53	11.6	39.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
56	8.8	5.8	EMISSION, FLAME	1, 2, 3, 4
57	8.7	4.6	EMISSION, IC PLASMA	3, 5
63	16.8	102.0	REJECT EMISSION, FLAME	1, 2, 3, 4
64	8.2	-1.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
67	7.6	-8.6	EMISSION, FLAME	1, 2, 3, 4
70	11.6	39.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
72	8.0	-3.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
73	8.6	3.4	EMISSION, IC PLASMA	3, 5
76	8.0	-3.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
77	8.5	2.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
80	8.2	-1.4	EMISSION, IC PLASMA	3, 5
81	7.9	-5.0	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
83	8.9	7.0	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
84	8.5	2.2	EMISSION, FLAME	1, 2, 3, 4
85	8.2	-1.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
86	8.5	2.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
91	8.0	-3.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
93	8.2	-1.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
96	7.8	-6.2	OTHER	1, 2, 3, 4
98	8.0	-3.8	EMISSION, IC PLASMA	3, 5
99	10.0	20.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
102	8.7	4.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
103	8.2	-1.4	EMISSION, IC PLASMA	3, 5
104	8.4	1.0	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
108	8.3	-0.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
110	8.8	5.8	EMISSION, IC PLASMA	3, 5
111	8.3	-0.2	EMISSION, IC PLASMA	3, 5
112	8.5	2.2	OTHER	3, 5
118	8.0	-3.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
121	8.5	2.2	EMISSION, FLAME	1, 2, 3, 4
122	7.1	-14.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
123	8.3	-0.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
127	10.0	20.3	OTHER	1, 2, 3, 4
142	8.2	-1.4	EMISSION, IC PLASMA	3, 5

68 Labs had a total range of 1.0 to 16.8 and a mean of 8.32 with a standard deviation of 0.57 and a 95% confidence interval of the mean +/- 0.14.

Table 8 Standard Reference Water Sample M6 Report for NO2-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 0.02		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
2	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
6	< 0.00		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
7	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
8	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
13	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
14	< 0.00		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
15	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
20	< 0.01		IGNORED ION CHROMATOGRAPHY	2,3,6
22	< 0.20		IGNORED ION CHROMATOGRAPHY	2,3,6
24	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
27	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
29	0.01	-95.9	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
30	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
34	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
36	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
37	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
38	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
40	0.02	-91.8	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
41	0.14	-42.3	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
42	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
43	< 0.02		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
44	1.38	469.1	ION CHROMATOGRAPHY	2,3,6
46	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
48	< 0.00		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
50	0.02	-91.8	NOT REPORTED	
53	< 0.02		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
56	0.01	-95.9	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
57	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
58	0.02	-91.8	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
64	0.01	-95.9	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
65	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
73	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
74	< 0.02		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
76	< 0.00		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
79	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
81	< 0.05		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
83	0.01	-95.9	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
85	0.01	-95.9	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
86	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
91	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
93	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
95	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
98	1.27	423.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
99	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
108	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
109	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
112	0.01	-95.9	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
113	< 0.01		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
121	< 0.01		IGNORED OTHER	
142	< 0.10		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4

51 Labs had a total range of 0.01 to 1.38 and a mean of 0.243 with a standard deviation of 0.508 and a 95% confidence interval of the mean +/- 0.322.

Table 8 Standard Reference Water Sample M6 Report for NO3-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1.0	-15.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
2	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
4	1.3	9.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
6	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
7	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
8	1.3	9.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
9	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
13	1.1	-7.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
14	1.1	-7.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
15	1.0	-15.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
19	4.0	236.7	REJECT ION CHROMATOGRAPHY	2, 3, 6, 7
20	1.4	17.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
22	1.7	43.1	REJECT ION CHROMATOGRAPHY	2, 3, 6, 7
24	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
25	1.2	1.0	OTHER	
27	1.0	-15.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
29	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
30	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
33	1.3	9.4	NOT REPORTED	
34	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
36	1.2	1.0	ION CHROMATOGRAPHY	2, 3, 6, 7
37	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
38	1.3	9.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
40	1.1	-7.4	COLORIMETRIC, BRUCINE	1, 2, 3, 4
41	1.1	-7.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
42	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
43	1.3	9.4	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
45	1.1	-7.4	OTHER	
46	0.8	-32.7	REJECT COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
47	1.5	26.3	COLORIMETRIC, BRUCINE	1, 2, 3, 4
48	1.4	17.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
49	1.2	1.0	ION CHROMATOGRAPHY	2, 3, 6, 7
50	1.2	1.0	NOT REPORTED	
51	1.3	9.4	ION CHROMATOGRAPHY	2, 3, 6, 7
52	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
53	1.0	-15.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
56	1.3	9.4	COLORIMETRIC, BRUCINE	1, 2, 3, 4
57	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
58	1.3	9.4	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
63	1.1	-7.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
64	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
65	1.1	-7.4	OTHER	
67	1.2	1.0	NOT REPORTED	
70	1.0	-15.8	COLORIMETRIC, BRUCINE	1, 2, 3, 4
72	1.1	-7.4	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
73	1.1	-7.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
74	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
76	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
79	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
80	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
81	1.4	17.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
82	1.0	-15.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
83	1.6	34.7	REJECT COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
84	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
85	1.1	-7.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
86	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
91	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
92	1.4	17.8	OTHER	
95	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
96	1.2	1.0	COLORIMETRIC, BRUCINE	1, 2, 3, 4
98	< 0.0		IGNORED COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
99	1.1	-7.4	COLORIMETRIC, BRUCINE	1, 2, 3, 4
103	1.2	1.0	ION CHROMATOGRAPHY	2, 3, 6, 7
104	1.1	-7.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
107	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
108	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
109	0.1	-91.6	REJECT COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
112	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
113	1.0	-15.8	ION SELECTIVE ELECTRODE	1, 2, 3, 4
118	1.2	1.0	OTHER	
121	1.2	1.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
122	< 0.1		IGNORED NOT REPORTED	
123	1.3	9.4	OTHER	
142	1.0	-15.8	ION CHROMATOGRAPHY	2, 3, 6, 7

74 Labs had a total range of 0.1 to 4.0 and a mean of 1.19 with a standard deviation of 0.11 and a 95% confidence interval of the mean +/- 0.03.

Table 8 Standard Reference Water Sample M6 Report for P, TOTAL

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 0.02		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
2	< 0.01		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
4	< 0.05		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
6	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
7	0.05	197.6	REJECT COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
8	< 0.01		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
9	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
13	0.03	78.6	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
14	0.01	-40.5	NOT REPORTED	
15	< 0.01		IGNORED COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
20	0.13	673.8	REJECT COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
22	< 1.00		IGNORED EMISSION, IC PLASMA	3, 5
29	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
30	0.02	19.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
33	< 0.10		IGNORED NOT REPORTED	
34	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
37	0.01	-40.5	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
38	< 0.01		IGNORED COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
40	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
41	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
42	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
43	< 0.01		IGNORED COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
44	< 0.01		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
47	0.02	19.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
48	0.03	78.6	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
49	0.02	19.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
51	< 0.06		IGNORED EMISSION, IC PLASMA	3, 5
53	< 0.05		IGNORED COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
56	0.03	78.6	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
57	< 0.01		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
58	< 0.02		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
63	< 0.01		IGNORED COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
64	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
65	0.03	78.6	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
69	< 0.01		IGNORED OTHER	
73	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
74	< 0.01		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
79	0.02	19.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
80	< 0.01		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
81	< 0.05		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
83	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
84	0.01	-40.5	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
85	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
86	0.02	19.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
91	0.02	19.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
93	7.74	5E+04	REJECT OTHER	
95	< 0.01		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
96	0.03	78.6	OTHER	
98	< 0.02		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
104	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
107	0.01	-40.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
108	0.04	138.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
109	< 0.00		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
110	0.20	1090.5	REJECT EMISSION, IC PLASMA	3, 5
112	0.10	495.2	REJECT COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
113	< 0.01		IGNORED COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
118	< 0.02		IGNORED COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
121	< 0.01		IGNORED OTHER	
127	< 0.05		IGNORED OTHER	
142	< 0.10		IGNORED EMISSION, IC PLASMA	3, 5

60 Labs had a total range of 0.01 to 7.74 and a mean of 0.017 with a standard deviation of 0.009 and a 95% confidence interval of the mean +/- 0.004.

Table 8 Standard Reference Water Sample M6 Report for PH

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	7.70	2.5	ELECTROMETRIC	1,2,3,4
2	7.60	1.2	ELECTROMETRIC	1,2,3,4
3	7.67	2.1	ELECTROMETRIC	1,2,3,4
4	8.35	11.2	ELECTROMETRIC	1,2,3,4
6	7.41	-1.3	ELECTROMETRIC	1,2,3,4
7	8.10	7.8	ELECTROMETRIC	1,2,3,4
8	6.94	-7.6	ELECTROMETRIC	1,2,3,4
9	7.60	1.2	ELECTROMETRIC	1,2,3,4
10	7.50	-0.1	ELECTROMETRIC	1,2,3,4
13	7.45	-0.8	ELECTROMETRIC	1,2,3,4
14	7.30	-2.8	ELECTROMETRIC	1,2,3,4
15	8.10	7.8	ELECTROMETRIC	1,2,3,4
16	7.49	-0.3	ELECTROMETRIC	1,2,3,4
19	6.88	-8.4	ELECTROMETRIC	1,2,3,4
20	7.40	-1.5	ELECTROMETRIC	1,2,3,4
22	7.40	-1.5	ELECTROMETRIC	1,2,3,4
24	7.79	3.7	ELECTROMETRIC	1,2,3,4
25	7.30	-2.8	ELECTROMETRIC	1,2,3,4
27	7.60	1.2	ELECTROMETRIC	1,2,3,4
29	7.70	2.5	ELECTROMETRIC	1,2,3,4
30	7.39	-1.6	ELECTROMETRIC	1,2,3,4
33	6.17	-17.9	REJECT NOT REPORTED	
34	7.66	2.0	ELECTROMETRIC	1,2,3,4
36	7.59	1.1	ELECTROMETRIC	1,2,3,4
37	7.70	2.5	ELECTROMETRIC	1,2,3,4
38	7.34	-2.3	ELECTROMETRIC	1,2,3,4
40	7.88	4.9	ELECTROMETRIC	1,2,3,4
41	6.40	-14.8	REJECT	
42	7.54	0.4	ELECTROMETRIC	1,2,3,4
44	7.68	2.3	ELECTROMETRIC	1,2,3,4
46	6.27	-16.5	REJECT	
47	7.30	-2.8	ELECTROMETRIC	1,2,3,4
48	7.85	4.5	ELECTROMETRIC	1,2,3,4
49	7.50	-0.1	ELECTROMETRIC	1,2,3,4
50	7.35	-2.1	ELECTROMETRIC	1,2,3,4
52	7.40	-1.5	ELECTROMETRIC	1,2,3,4
53	7.27	-3.2	ELECTROMETRIC	1,2,3,4
56	7.55	0.5	ELECTROMETRIC	1,2,3,4
57	7.80	3.9	ELECTROMETRIC	1,2,3,4
58	7.61	1.3	ELECTROMETRIC	1,2,3,4
63	7.60	1.2	ELECTROMETRIC	1,2,3,4
64	7.80	3.9	ELECTROMETRIC	1,2,3,4
67	7.50	-0.1	ELECTROMETRIC	1,2,3,4
69	7.77	3.5	NOT REPORTED	
70	7.47	-0.5	ELECTROMETRIC	1,2,3,4
73	7.40	-1.5	ELECTROMETRIC	1,2,3,4
74	7.40	-1.5	ELECTROMETRIC	1,2,3,4
76	7.20	-4.1	ELECTROMETRIC	1,2,3,4
77	7.40	-1.5	ELECTROMETRIC	1,2,3,4
79	7.30	-2.8	ELECTROMETRIC	1,2,3,4
80	7.55	0.5	ELECTROMETRIC	1,2,3,4
81	6.90	-8.1	ELECTROMETRIC	1,2,3,4
83	7.70	2.5	ELECTROMETRIC	1,2,3,4
84	7.42	-1.2	ELECTROMETRIC	1,2,3,4
85	7.60	1.2	ELECTROMETRIC	1,2,3,4
86	7.40	-1.5	ELECTROMETRIC	1,2,3,4
91	7.26	-3.3	ELECTROMETRIC	1,2,3,4
93	7.75	3.2	ELECTROMETRIC	1,2,3,4
94	7.70	2.5	ELECTROMETRIC	1,2,3,4
95	7.63	1.6	ELECTROMETRIC	1,2,3,4
96	6.96	-7.3	ELECTROMETRIC	1,2,3,4
98	7.50	-0.1	ELECTROMETRIC	1,2,3,4
99	7.41	-1.3	ELECTROMETRIC	1,2,3,4
102	7.60	1.2	ELECTROMETRIC	1,2,3,4
103	7.54	0.4	ELECTROMETRIC	1,2,3,4
104	7.42	-1.2	ELECTROMETRIC	1,2,3,4
108	7.80	3.9	ELECTROMETRIC	1,2,3,4
109	7.18	-4.4	ELECTROMETRIC	1,2,3,4
111	7.60	1.2	ELECTROMETRIC	1,2,3,4
112	7.86	4.6	ELECTROMETRIC	1,2,3,4
113	7.08	-5.7	ELECTROMETRIC	1,2,3,4
118	7.70	2.5	ELECTROMETRIC	1,2,3,4
122	7.60	1.2	ELECTROMETRIC	1,2,3,4
123	7.70	2.5	ELECTROMETRIC	1,2,3,4
124	7.00	-6.8	NOT REPORTED	
142	6.90	-8.1	ELECTROMETRIC	1,2,3,4

76 Labs had a total range of 6.17 to 8.35 and a mean of 7.511 with a standard deviation of 0.278 and a 95% confidence interval of the mean +/- 0.065.



Table 8 Standard Reference Water Sample M6 Report for SIO2

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	12.7	16.9	EMISSION, IC PLASMA	5
6	11.0	1.3	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3
8	10.0	-7.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
9	10.0	-7.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
13	11.2	3.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
14	9.6	-11.6	NOT REPORTED	
15	12.0	10.5	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
17	6.0	-44.8	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
20	11.0	1.3	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
22	8.2	-24.5	EMISSION, IC PLASMA	5
24	10.2	-6.1	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
25	10.0	-7.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
30	11.0	1.3	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3
34	9.2	-15.3	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
36	10.4	-4.2	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
38	10.6	-2.4	EMISSION, IC PLASMA	5
40	11.0	1.3	EMISSION, IC PLASMA	5
41	9.0	-17.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
42	10.8	-0.6	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
47	13.0	19.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
48	12.0	10.5	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
49	10.5	-3.3	EMISSION, IC PLASMA	5
52	11.0	1.3	EMISSION, IC PLASMA	5
57	11.0	1.3	EMISSION, IC PLASMA	5
63	13.0	19.7	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
64	11.2	3.1	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
73	11.3	4.0	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
80	10.0	-7.9	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
82	12.5	15.1	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
83	12.0	10.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
85	10.0	-7.9	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
91	7.9	-27.3	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
93	8.4	-22.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
95	11.0	1.3	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
98	10.8	-0.6	EMISSION, IC PLASMA	5
99	13.8	27.1	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3
103	10.1	-7.0	EMISSION, IC PLASMA	5
104	12.6	16.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
108	11.0	1.3	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
110	11.0	1.3	EMISSION, IC PLASMA	5
112	12.0	10.5	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
118	11.0	1.3	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
123	11.0	1.3	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
142	11.0	1.3	EMISSION, IC PLASMA	5

44 Labs had a total range of 6.0 to 13.8 and a mean of 10.86 with a standard deviation of 1.27 and a 95% confidence interval of the mean +/- 0.39.

Table 8 Standard Reference Water Sample M6 Report for SO4

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	76	3.8	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
2	79	7.9	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
4	76	3.8	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
6	75	2.4	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
7	76	3.8	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
8	73	-0.3	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
9	70	-4.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
10	78	6.5	ION CHROMATOGRAPHY	2,6,7
13	80	9.3	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
14	70	-4.4	GRAVIMETRIC, BARIUM SULFATE	1,2,3
15	73	-0.3	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
17	121	65.3	REJECT TURBIDIMETRIC, BARIUM SULFATE	1,2,3
19	34	-53.6	REJECT ION CHROMATOGRAPHY	2,6,7
20	8	-89.1	REJECT TURBIDIMETRIC, BARIUM SULFATE	1,2,3
22	76	3.8	ION CHROMATOGRAPHY	2,6,7
24	26	-64.5	REJECT COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
25	65	-11.2	THORIN TITRATION	2,4
27	75	2.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
29	63	-13.9	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
30	72	-1.7	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
33	90	22.9	NOT REPORTED	
34	74	1.1	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
36	72	-1.7	ION CHROMATOGRAPHY	2,6,7
37	71	-3.0	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
38	77	5.2	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
40	74	1.1	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
41	75	2.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
42	69	-5.8	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
43	60	-18.0	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
45	74	1.1	GRAVIMETRIC, BARIUM SULFATE	1,2,3
46	73	-0.3	GRAVIMETRIC, BARIUM SULFATE	1,2,3
47	79	7.9	ION CHROMATOGRAPHY	2,6,7
48	78	6.5	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
49	74	1.1	ION CHROMATOGRAPHY	2,6,7
50	50	-31.7	REJECT TURBIDIMETRIC, BARIUM SULFATE	1,2,3
51	74	1.1	ION CHROMATOGRAPHY	2,6,7
52	74	1.1	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
53	70	-4.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
56	75	2.4	ION CHROMATOGRAPHY	2,6,7
57	75	2.4	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
58	76	3.8	GRAVIMETRIC, BARIUM SULFATE	1,2,3
63	46	-37.2	REJECT ION CHROMATOGRAPHY	2,6,7
67	76	3.8	ION CHROMATOGRAPHY	2,6,7
70	73	-0.3	OTHER	
72	60	-18.0	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
76	60	-18.0	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
77	74	1.1	THORIN TITRATION	2,4
79	77	5.2	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
80	73	-0.3	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
81	95	29.8	REJECT COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
83	71	-3.0	GRAVIMETRIC, BARIUM SULFATE	1,2,3
84	63	-13.9	COLORIMETRIC, CHLORANILATE	3
85	82	12.0	THORIN TITRATION	2,4
86	70	-4.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
91	73	-0.3	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
93	76	3.8	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
96	73	-0.3	GRAVIMETRIC, BARIUM SULFATE	1,2,3
98	62	-15.3	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
99	75	2.4	GRAVIMETRIC, BARIUM SULFATE	1,2,3
102	68	-7.1	TURBIDIMETRIC, BARIUM SULFATE	1,2,3
103	74	1.1	ION CHROMATOGRAPHY	2,6,7
104	45	-38.5	REJECT GRAVIMETRIC, BARIUM SULFATE	1,2,3
107	79	7.9	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
108	80	9.3	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
112	54	-26.2	REJECT TURBIDIMETRIC, BARIUM SULFATE	1,2,3
118	75	2.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
121	73	-0.3	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
122	70	-4.4	COLORIMETRIC, CHLORANILATE	3
123	74	1.1	GRAVIMETRIC, BARIUM SULFATE	1,2,3
142	74	1.1	ION CHROMATOGRAPHY	2,6,7

70 Labs had a total range of 8 to 121 and a mean of 73.2 with a standard deviation of 5.4 and a 95% confidence interval of the mean +/- 1.4.

Table 8 Standard Reference Water Sample M6 Report for SP. COND.

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	261	-3.7	DIRECT READING INSTRUMENT	4
2	181	-33.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
3	280	3.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
4	270	-0.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
6	259	-4.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
7	272	0.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
8	278	2.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
9	253	-6.7	DIRECT READING INSTRUMENT	4
10	253	-6.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
13	278	2.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
14	275	1.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
15	270	-0.4	DIRECT READING INSTRUMENT	4
16	281	3.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
17	273	0.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
19	179	-34.0	DIRECT READING INSTRUMENT	4
20	1300	379.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
22	268	-1.2	DIRECT READING INSTRUMENT	4
24	272	0.3	DIRECT READING INSTRUMENT	4
25	277	2.2	DIRECT READING INSTRUMENT	4
27	258	-4.8	DIRECT READING INSTRUMENT	4
29	240	-11.5	DIRECT READING INSTRUMENT	4
30	273	0.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
33	260	-4.1	NOT REPORTED	
34	276	1.8	DIRECT READING INSTRUMENT	4
36	276	1.8	DIRECT READING INSTRUMENT	4
37	276	1.8	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
38	270	-0.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
40	276	1.8	DIRECT READING INSTRUMENT	4
41	271	-0.1	DIRECT READING INSTRUMENT	4
43	274	1.1	DIRECT READING INSTRUMENT	4
46	276	1.8	DIRECT READING INSTRUMENT	4
47	270	-0.4	DIRECT READING INSTRUMENT	4
48	276	1.8	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
49	282	4.0	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
50	269	-0.8	DIRECT READING INSTRUMENT	4
52	271	-0.1	DIRECT READING INSTRUMENT	4
53	282	4.0	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
56	256	-5.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
57	275	1.4	DIRECT READING INSTRUMENT	4
58	271	-0.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
63	291	7.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
64	272	0.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
67	253	-6.7	DIRECT READING INSTRUMENT	4
69	210	-22.6	REJECT NOT REPORTED	
70	260	-4.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
72	230	-15.2	REJECT DIRECT READING INSTRUMENT	4
73	278	2.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
74	268	-1.2	DIRECT READING INSTRUMENT	4
76	293	8.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
77	277	2.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
79	297	9.5	ELECTRODELESS, INDUCTIVE CELL-TYPE	2
80	277	2.2	DIRECT READING INSTRUMENT	4
81	250	-7.8	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
82	280	3.3	DIRECT READING INSTRUMENT	4
83	263	-3.0	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
84	240	-11.5	DIRECT READING INSTRUMENT	4
85	275	1.4	DIRECT READING INSTRUMENT	4
86	256	-5.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
91	260	-4.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
93	279	2.9	DIRECT READING INSTRUMENT	4
94	260	-4.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
96	271	-0.1	DIRECT READING INSTRUMENT	4
98	272	0.3	DIRECT READING INSTRUMENT	4
99	298	9.9	DIRECT READING INSTRUMENT	4
102	274	1.1	DIRECT READING INSTRUMENT	4
103	263	-3.0	DIRECT READING INSTRUMENT	4
104	266	-1.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
108	270	-0.4	DIRECT READING INSTRUMENT	4
109	272	0.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
111	278	2.5	DIRECT READING INSTRUMENT	4
118	280	3.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
122	276	1.8	DIRECT READING INSTRUMENT	4
123	278	2.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
142	284	4.7	DIRECT READING INSTRUMENT	4

74 Labs had a total range of 179 to 1300 and a mean of 271.1 with a standard deviation of 11.2 and a 95% confidence interval of the mean +/- 2.7.

Table 8 Standard Reference Water Sample M6 Report for SR

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1	-99.8	EMISSION, IC PLASMA	3,5
6	500	14.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
8	500	14.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
13	480	10.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
15	550	26.4	OTHER	
22	470	8.0	EMISSION, IC PLASMA	3,5
24	510	17.2	ATOMIC ABSORPTION, FLAMELESS	7
36	334	-23.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
41	360	-17.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
47	440	1.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
49	495	13.7	EMISSION, IC PLASMA	3,5
51	501	15.1	EMISSION, IC PLASMA	3,5
52	470	8.0	EMISSION, IC PLASMA	3,5
53	536	23.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
57	492	13.1	EMISSION, IC PLASMA	3,5
63	7	-98.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
67	490	12.6	EMISSION, IC PLASMA	3,5
73	490	12.6	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7
80	510	17.2	ATOMIC ABSORPTION, FLAMELESS	7
85	500	14.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
93	504	15.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
98	500	14.9	EMISSION, IC PLASMA	3,5
103	496	14.0	EMISSION, IC PLASMA	3,5
104	500	14.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
110	500	14.9	EMISSION, IC PLASMA	3,5
112	1	-99.8	OTHER	
127	500	14.9	OTHER	
142	480	10.3	EMISSION, IC PLASMA	3,5
144	503	15.6	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

29 Labs had a total range of 1 to 550 and a mean of 435.2 with a standard deviation of 155.5 and a 95% confidence interval of the mean +/- 59.1.

Table 8 Standard Reference Water Sample M6 Report for V

Code Number	Reported value	Pct. dev. from mean	Methods	References
6	<100.0		IGNORED	1,3
8	<100.0		IGNORED	1,3
13	12.5	97.1	EMISSION, IC PLASMA	3,5
15	3.8	-40.1	ATOMIC ABSORPTION, FLAMELESS	3,5
22	29.0	357.3	REJECT	3,5
24	13.6	114.5	ATOMIC ABSORPTION, FLAMELESS	3,5
38	< 10.0		IGNORED	3,5
41	<200.0		IGNORED	1,3
47	3.6	-43.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	3
49	3.0	-52.7	ATOMIC ABSORPTION, FLAMELESS	3,5
52	5.0	-21.2	EMISSION, IC PLASMA	3,5
57	2.0	-68.5	COLORIMETRIC, CATALYTIC OXIDATION	4
63	< 40.0		IGNORED	1,3
73	6.0	-5.4	EMISSION, IC PLASMA	3,5
80	4.0	-36.9	EMISSION, IC PLASMA	3,5
83	<100.0		IGNORED	1,3
93	<100.0		IGNORED	1,3
98	4.0	-36.9	EMISSION, IC PLASMA	3,5
103	< 2.0		IGNORED	3,5
104	<100.0		IGNORED	1,3
110	15.0	136.5	EMISSION, IC PLASMA	3,5
142	3.6	-43.2	EMISSION, IC PLASMA	3,5

22 Labs had a total range of 2.0 to 29.0 and a mean of 6.34 with a standard deviation of 4.57 and a 95% confidence interval of the mean +/- 2.91.

Table 9 . Statistics by method for standard reference sample M6

Determin- ation	Method	Range: from to	Mean	Standard Deviation	N
ALK(CACO3)	'TITRATION, COLORIMETRIC'	13.000 - 30.000	26.000	2.114	18
	'TITRATION, ELECTROMETRIC'	16.000 - 52.000	25.256	1.996	39
	NOT REPORTED	20.000 - 31.000	25.000	5.568	3
	OTHER	24.000 - 26.000	25.000	1.000	3
	_OVER-ALL_	13.000 - 52.000	25.444	2.198	63
B	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	230.000 - 230.000	-----	-----	-
	COLORIMETRIC, AZOMETHINE	10.000 - 80.000	-----	-----	-
	COLORIMETRIC, CURCUMIN	20.000 - 137.000	38.600	17.883	5
	EMISSION, DC PLASMA	21.000 - 51.000	-----	-----	-
	EMISSION, IC PLASMA	13.000 - 75.000	33.250	19.545	12
_OVER-ALL_	10.000 - 230.000	34.727	21.054	22	
BR	ION CHROMATOGRAPHY	8.000 - 18.000	-----	-----	-
	_OVER-ALL_	8.000 - 18.000	-----	-----	-
CA	ATOMIC ABSORPTION, DIRECT, AIR	21.000 - 32.000	25.750	2.395	36
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	18.000 - 257.000	26.000	1.225	5
	EMISSION, IC PLASMA	24.000 - 28.000	26.063	1.237	16
	OTHER	22.000 - 26.000	24.500	1.732	4
	TITRATION, EDTA	25.000 - 44.000	27.000	2.160	4
_OVER-ALL_	18.000 - 257.000	25.818	2.037	66	
CL	COLORIMETRIC, FERRIC THIOCYANATE	10.800 - 17.500	12.908	0.972	24
	ION CHROMATOGRAPHY	6.400 - 16.400	13.800	1.166	10
	TITRATION, MERCURIC NITRATE	12.000 - 16.000	13.707	0.945	15
	TITRATION, SILVER NITRATE	10.000 - 17.900	13.377	0.851	13
	_OVER-ALL_	6.400 - 17.900	13.297	1.086	67
DSRD 180	NOT REPORTED	125.000 - 167.000	149.333	21.779	3
	RESIDUE ON EVAPORATION	17.000 - 231.000	185.542	21.437	24
	RESIDUE, FILTRABLE	151.000 - 238.000	174.034	14.451	29
	_OVER-ALL_	17.000 - 238.000	178.475	20.920	59
F	ION CHROMATOGRAPHY	0.400 - 0.900	0.857	0.053	7
	ION SELECTIVE ELECTRODE	0.700 - 1.200	0.835	0.060	34
	OTHER	0.800 - 0.900	-----	-----	-
	_OVER-ALL_	0.400 - 1.200	0.848	0.061	52
K	ATOMIC ABSORPTION, DIRECT, AIR	0.680 - 2.350	0.838	0.096	40
	EMISSION, FLAME, PHOTOMETRIC	0.560 - 3.340	1.000	0.313	6
	EMISSION, IC PLASMA	0.750 - 2.000	0.895	0.124	10
	_OVER-ALL_	0.560 - 3.340	0.853	0.115	59
	MG	ATOMIC ABSORPTION, DIRECT, AIR	9.000 - 24.000	10.324	0.669
ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE		10.000 - 12.000	10.833	0.753	6
EMISSION, IC PLASMA		7.000 - 11.000	10.667	0.488	15
OTHER		9.000 - 10.000	-----	-----	-
TITRATION, EDTA		10.000 - 16.000	12.250	2.872	4
_OVER-ALL_	7.000 - 24.000	10.406	0.660	64	
NA	ATOMIC ABSORPTION, DIRECT, AIR	6.300 - 11.600	8.245	0.660	38
	EMISSION, FLAME	7.600 - 16.800	8.471	0.528	7
	EMISSION, IC PLASMA	1.000 - 8.800	8.287	0.329	15
	OTHER	7.800 - 10.000	8.767	1.124	3
	_OVER-ALL_	1.000 - 16.800	8.316	0.575	63
NO2-N	COLORIMETRIC, DIAZOTIZATION	0.010 - 1.270	0.013	0.005	8
	ION CHROMATOGRAPHY	1.380 - 1.380	-----	-----	-
	_OVER-ALL_	0.010 - 1.380	0.243	0.508	12
NO3-N	COLORIMETRIC, BRUCINE	1.000 - 1.500	1.200	0.179	6
	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	0.100 - 1.600	1.181	0.098	43
	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	1.100 - 1.300	-----	-----	-
	ION CHROMATOGRAPHY	1.000 - 4.000	1.267	0.234	6
	NOT REPORTED	1.200 - 1.300	-----	-----	-
	OTHER	1.100 - 1.400	1.217	0.117	6
_OVER-ALL_	0.100 - 4.000	1.188	0.108	67	
P, TOTAL	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	0.010 - 0.100	0.015	0.007	21
	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	0.010 - 0.130	0.020	0.012	4
	EMISSION, IC PLASMA	0.200 - 0.200	-----	-----	-
	OTHER	0.030 - 7.740	-----	-----	-
	_OVER-ALL_	0.010 - 7.740	0.017	0.009	28
PH	ELECTROMETRIC	6.270 - 8.350	7.514	0.274	71
	NOT REPORTED	6.170 - 7.770	6.980	0.800	3
	_OVER-ALL_	6.170 - 8.350	7.511	0.278	73
SIO2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	6.000 - 13.000	10.220	2.115	10
	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	10.000 - 13.000	11.078	0.967	9
	COLORIMETRIC, MOLYBDSILICIC ACID	7.900 - 12.500	10.943	1.553	7
	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	9.200 - 11.000	-----	-----	-
	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	11.000 - 13.800	-----	-----	-
	EMISSION, IC PLASMA	8.200 - 12.700	10.778	0.319	9
	_OVER-ALL_	6.000 - 13.800	10.861	1.269	43
SO4	COLORIMETRIC, METHYL THYMOL BLUE	26.000 - 95.000	74.000	3.378	18
	GRAVIMETRIC, BARIUM SULFATE	45.000 - 76.000	73.250	1.982	8
	ION CHROMATOGRAPHY	34.000 - 79.000	75.200	2.098	10
	THORIN TITRATION	65.000 - 82.000	73.667	8.505	3
	TURBIDIMETRIC, BARIUM SULFATE	8.000 - 121.000	69.632	8.757	19
	_OVER-ALL_	8.000 - 121.000	73.213	5.392	61
SP. COND.	DIRECT READING INSTRUMENT	179.000 - 298.000	270.324	11.353	34
	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	181.000 - 1300.000	271.545	10.459	33
	_OVER-ALL_	179.000 - 1300.000	271.145	11.229	69

Table 9 . Statistics by method for standard reference sample M6

Determination	Method	Range:		Mean	Standard Deviation	N
		from	to			
SR	ATOMIC ABSORPTION, DIRECT, AIR	7.000	536.000	465.400	67.079	10
	EMISSION, IC PLASMA	1.000	501.000	489.400	11.937	10
	OTHER	1.000	550.000	-----	-----	-
	_OVER-ALL_	1.000	550.000	435.172	155.478	29
V	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	-----	-----	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	3.600	13.600	-----	-----	-
	EMISSION, IC PLASMA	3.000	29.000	6.638	4.532	8
	_OVER-ALL_	2.000	29.000	6.342	4.572	12

Table 10 Standard Reference Water Sample M94 Report for ALK(CACO3)

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	250	2.4	'TITRATION, ELECTROMETRIC'	4
2	240	-1.7	'TITRATION, ELECTROMETRIC'	4
4	240	-1.7	'TITRATION, COLORIMETRIC'	3
6	246	0.8	'TITRATION, ELECTROMETRIC'	4
7	253	3.7	'TITRATION, ELECTROMETRIC'	4
8	236	-3.3	'TITRATION, ELECTROMETRIC'	4
9	240	-1.7	OTHER	
12	247	1.2	'TITRATION, ELECTROMETRIC'	4
13	245	0.4	'TITRATION, ELECTROMETRIC'	4
15	248	1.6	'TITRATION, COLORIMETRIC'	3
16	246	0.8	'TITRATION, ELECTROMETRIC'	4
19	240	-1.7	'TITRATION, ELECTROMETRIC'	4
20	500	104.9	REJECT	
22	215	-11.9	REJECT	
23	236	-3.3	'TITRATION, ELECTROMETRIC'	4
24	251	2.8	'TITRATION, ELECTROMETRIC'	4
25	258	5.7	'TITRATION, COLORIMETRIC'	3
27	259	6.1	'TITRATION, COLORIMETRIC'	3
30	264	8.2	'TITRATION, ELECTROMETRIC'	4
33	239	-2.1	NOT REPORTED	
34	241	-1.3	'TITRATION, ELECTROMETRIC'	4
36	247	1.2	'TITRATION, ELECTROMETRIC'	4
37	244	-0.0	'TITRATION, COLORIMETRIC'	3
38	224	-8.2	'TITRATION, ELECTROMETRIC'	4
40	250	2.4	'TITRATION, COLORIMETRIC'	3
41	252	3.2	'TITRATION, ELECTROMETRIC'	4
43	246	0.8	'TITRATION, ELECTROMETRIC'	4
44	240	-1.7	'TITRATION, COLORIMETRIC'	3
45	248	1.6	'TITRATION, ELECTROMETRIC'	4
46	222	-9.0	'TITRATION, ELECTROMETRIC'	4
48	234	-4.1	'TITRATION, ELECTROMETRIC'	4
49	488	99.9	REJECT	
50	244	-0.0	'TITRATION, ELECTROMETRIC'	4
52	230	-5.8	'TITRATION, ELECTROMETRIC'	4
53	250	2.4	'TITRATION, ELECTROMETRIC'	4
54	261	6.9	'TITRATION, COLORIMETRIC'	3
56	247	1.2	'TITRATION, ELECTROMETRIC'	4
57	247	1.2	'TITRATION, ELECTROMETRIC'	4
58	244	-0.0	'TITRATION, ELECTROMETRIC'	4
59	233	-4.5	'TITRATION, ELECTROMETRIC'	4
62	241	-1.3	'TITRATION, ELECTROMETRIC'	4
63	245	0.4	'TITRATION, COLORIMETRIC'	3
64	252	3.2	'TITRATION, COLORIMETRIC'	3
68	244	-0.0	'TITRATION, COLORIMETRIC'	3
69	235	-3.7	NOT REPORTED	
70	240	-1.7	'TITRATION, COLORIMETRIC'	3
72	244	-0.0	'TITRATION, ELECTROMETRIC'	4
73	245	0.4	OTHER	
77	248	1.6	'TITRATION, ELECTROMETRIC'	4
79	240	-1.7	'TITRATION, ELECTROMETRIC'	4
80	242	-0.8	OTHER	
81	240	-1.7	'TITRATION, ELECTROMETRIC'	4
83	248	1.6	'TITRATION, COLORIMETRIC'	3
84	249	2.0	'TITRATION, ELECTROMETRIC'	4
85	250	2.4	'TITRATION, ELECTROMETRIC'	4
86	280	14.7	REJECT	
87	250	2.4	'TITRATION, COLORIMETRIC'	3
90	240	-1.7	'TITRATION, COLORIMETRIC'	3
91	235	-3.7	'TITRATION, ELECTROMETRIC'	4
93	242	-0.8	'TITRATION, COLORIMETRIC'	3
94	248	1.6	'TITRATION, ELECTROMETRIC'	4
97	246	0.8	OTHER	
98	239	-2.1	'TITRATION, ELECTROMETRIC'	4
99	256	4.9	'TITRATION, COLORIMETRIC'	3
102	247	1.2	'TITRATION, ELECTROMETRIC'	4
103	238	-2.5	'TITRATION, ELECTROMETRIC'	4
104	243	-0.4	'TITRATION, COLORIMETRIC'	3
105	244	-0.0	'TITRATION, ELECTROMETRIC'	4
107	240	-1.7	'TITRATION, ELECTROMETRIC'	4
112	244	-0.0	'TITRATION, ELECTROMETRIC'	4
113	244	-0.0	'TITRATION, ELECTROMETRIC'	4
118	248	1.6	'TITRATION, ELECTROMETRIC'	4
123	255	4.5	'TITRATION, ELECTROMETRIC'	4
124	225	-7.8	NOT REPORTED	
142	240	-1.7	'TITRATION, ELECTROMETRIC'	4

75 Labs had a total range of 215 to 500 and a mean of 244.1 with a standard deviation of 7.8 and a 95% confidence interval of the mean +/- 1.8.

Table 10 Standard Reference Water Sample M94 Report for B

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	240	7.2	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
6	230	2.7	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
8	200	-10.7	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
9	306	36.7	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
13	< 250		IGNORED	1, 2, 3, 4
14	60	-73.2	REJECT	1, 2, 3, 4
15	243	8.6	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
22	141	-37.0	COLORIMETRIC, CURCUMIN	1, 2, 3, 4
25	257	14.8	EMISSION, IC PLASMA	1, 2, 3, 4
34	< 1		IGNORED	3
38	214	-4.4	COLORIMETRIC, AZOMETHINE	7
40	200	-10.7	EMISSION, IC PLASMA	5
41	170	-24.1	EMISSION, IC PLASMA	3
46	< 1		IGNORED	3
47	2000	793.4	COLORIMETRIC, CARMINE (CARMINIC ACID)	1, 2, 3, 4
48	180	-19.6	REJECT	2, 4
49	230	2.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	7
52	262	17.0	EMISSION, IC PLASMA	3
54	230	2.7	EMISSION, IC PLASMA	3
57	216	-3.5	COLORIMETRIC, AZOMETHINE	3
63	< 40		IGNORED	5
68	215	-4.0	EMISSION, DC PLASMA	7
73	200	-10.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	7
80	225	0.5	COLORIMETRIC, AZOMETHINE	5
83	< 2000		IGNORED	3
85	220	-1.7	EMISSION, IC PLASMA	7
93	313	39.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	3
97	480	114.4	COLORIMETRIC, DIANTHRIMIDE	4
98	240	7.2	REJECT	1, 2, 3, 4
99	250	11.7	COLORIMETRIC, CURCUMIN	2, 4
103	210	-6.2	EMISSION, IC PLASMA	2, 4
105	200	-10.7	COLORIMETRIC, CARMINE (CARMINIC ACID)	3
108	230	2.7	EMISSION, IC PLASMA	5
110	220	-1.7	COLORIMETRIC, AZOMETHINE	3
111	206	-8.0	EMISSION, IC PLASMA	5
112	< 1		IGNORED	3
123	420	87.6	EMISSION, DC PLASMA	7
127	< 20		REJECT	2, 4
142	220	-1.7	IGNORED	3
			OTHER	
			EMISSION, IC PLASMA	3

39 Labs had a total range of 60 to 2000 and a mean of 223.9 with a standard deviation of 35.3 and a 95% confidence interval of the mean +/- 13.7.

Table 10 Standard Reference Water Sample M94 Report for BR

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	600	52.7	COLORIMETRIC, CATALYTIC OXIDATION	2, 4
6	320	-18.6	ION CHROMATOGRAPHY	2, 3, 6
19	412	4.8	ION CHROMATOGRAPHY	2, 3, 6
22	230	-41.5	ION CHROMATOGRAPHY	2, 3, 6
36	420	6.9	ION CHROMATOGRAPHY	2, 3, 6
38	410	4.3	ION CHROMATOGRAPHY	2, 3, 6
54	745	89.6	ION CHROMATOGRAPHY	2, 3, 6
57	< 1	-100.0	IGNORED	2, 4
87	< 1		IGNORED	1
127	5900	1401.3	ION CHROMATOGRAPHY	2, 3, 6
142	400	1.8	REJECT	2, 3, 6
			OTHER	
			ION CHROMATOGRAPHY	2, 3, 6

11 Labs had a total range of 0 to 5900 and a mean of 442.0 with a standard deviation of 161.8.



Table 10 Standard Reference Water Sample M94 Report for CA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	99	-1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	100	-0.5	OTHER	1,2,3,4
4	104	3.5	TITRATION, EDTA	1,3
6	97	-3.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	104	3.5	EMISSION, IC PLASMA	3,5,7
8	101	0.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
9	105	4.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	102	1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	100	-0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	121	20.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	96	-4.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
22	103	2.5	EMISSION, IC PLASMA	3,5,7
24	100	-0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	108	7.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
27	99	-1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	95	-5.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	100	-0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	86	-14.4	NOT REPORTED	
34	98	-2.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
36	100	-0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	84	-16.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	104	3.5	EMISSION, IC PLASMA	3,5,7
40	99	-1.5	EMISSION, IC PLASMA	3,5,7
41	112	11.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
42	96	-4.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	110	7.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
46	106	3.5	TITRATION, EDTA	1,3
47	97	-3.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	106	3.5	EMISSION, IC PLASMA	3,5,7
49	100	-0.5	EMISSION, IC PLASMA	3,5,7
50	79	-21.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
52	100	-0.5	EMISSION, IC PLASMA	3,5,7
53	108	7.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	113	12.5	EMISSION, IC PLASMA	3,5,7
56	180	79.2	REJECT TITRATION, EDTA	1,3
57	96	-4.4	EMISSION, IC PLASMA	3,5,7
59	106	5.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
61	107	6.5	EMISSION, IC PLASMA	3,5,7
62	102	1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
63	108	7.5	TITRATION, EDTA	1,3
64	90	-10.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
68	106	5.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	100	-0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	93	-7.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	98	-2.5	EMISSION, IC PLASMA	3,5,7
77	100	-0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	101	0.5	EMISSION, IC PLASMA	3,5,7
81	104	3.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	100	-0.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
84	143	42.3	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
85	100	-0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	102	1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
87	93	-7.4	OTHER	
90	95	-5.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	97	-3.4	ATOMIC ABSORPTION, DIRECT, AIR	1,7
93	100	-0.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
97	100	-0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
98	90	-10.4	EMISSION, IC PLASMA	3,5,7
99	108	7.5	TITRATION, EDTA	1,3
102	90	-10.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
103	105	4.5	EMISSION, IC PLASMA	3,5,7
104	102	1.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
105	100	-0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
108	100	-0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	100	-0.5	EMISSION, IC PLASMA	3,5,7
111	101	0.5	EMISSION, IC PLASMA	3,5,7
112	10	-90.0	REJECT OTHER	
113	84	-16.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	110	9.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	99	-1.5	OTHER	
122	104	3.5	NOT REPORTED	
123	100	-0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	110	9.5	OTHER	
142	100	-0.5	EMISSION, IC PLASMA	3,5,7

74 Labs had a total range of 10 to 180 and a mean of 100.5 with a standard deviation of 6.9 and a 95% confidence interval of the mean +/- Table

Table 10 Standard Reference Water Sample M94 Report for CL

Code Number	Reported value	Pct. dev. from mean		Methods	References
1	69	6.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
2	6	-90.8	REJECT	ION SELECTIVE ELECTRODE	1, 2, 3, 4
4	66	-1.5		TITRATION, SILVER NITRATE	1, 2, 4
6	63	-3.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
7	70	7.6		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
8	92	41.4	REJECT	COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
9	65	-0.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
12	65	-0.1		TITRATION, SILVER NITRATE	1, 2, 4
13	63	-3.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
14	65	-0.1		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
15	63	-3.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
16	63	-3.1		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
19	59	-9.3		ION CHROMATOGRAPHY	2, 3, 6, 7
20	58	-10.8		TITRATION, SILVER NITRATE	1, 2, 4
22	74	13.8		ION CHROMATOGRAPHY	2, 3, 6, 7
24	65	-0.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
25	50	-23.1	REJECT	TITRATION, SILVER NITRATE	1, 2, 4
27	65	-0.1		TITRATION, SILVER NITRATE	1, 2, 4
29	70	7.6		TITRATION, SILVER NITRATE	1, 2, 4
30	65	-0.1		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
33	59	-9.3		NOT REPORTED	
34	70	7.6		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
36	66	1.5		ION CHROMATOGRAPHY	2, 3, 6, 7
37	65	-0.1		ION SELECTIVE ELECTRODE	1, 2, 3, 4
38	69	6.1		ION CHROMATOGRAPHY	2, 3, 6, 7
40	65	-0.1		TITRATION, SILVER NITRATE	1, 2, 4
41	66	1.5		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
42	64	-1.6		TITRATION, SILVER NITRATE	1, 2, 4
43	66	1.5		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
46	65	-0.1		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
47	63	-3.1		ION CHROMATOGRAPHY	2, 3, 6, 7
48	63	-3.1		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
49	65	-0.1		ION CHROMATOGRAPHY	2, 3, 6, 7
52	63	-3.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
53	65	-0.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
56	69	6.1		ION CHROMATOGRAPHY	2, 3, 6, 7
57	67	3.0		TITRATION, SILVER NITRATE	1, 2, 4
58	63	-3.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
59	64	-1.6		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
62	70	7.6		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
63	65	-0.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
65	69	6.1		ION CHROMATOGRAPHY	2, 3, 6, 7
68	66	1.5		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
70	64	-1.6		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
72	66	1.5		TITRATION, SILVER NITRATE	1, 2, 4
73	61	-6.2		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
74	65	-0.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
77	64	-1.6		TITRATION, SILVER NITRATE	1, 2, 4
79	64	-1.6		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
80	63	-3.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
81	66	1.5		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
82	72	10.7		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
83	65	-0.1		TITRATION, SILVER NITRATE	1, 2, 4
84	46	-29.3	REJECT	COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
85	68	4.5		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
86	68	4.5		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
87	62	-4.7		ION CHROMATOGRAPHY	2, 3, 6, 7
90	66	1.5		TITRATION, SILVER NITRATE	1, 2, 4
91	63	-3.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
93	64	-1.6		TITRATION, SILVER NITRATE	1, 2, 4
95	76	16.8	REJECT	TITRATION, SILVER NITRATE	1, 2, 4
97	64	-1.6		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
98	65	-0.1		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
99	66	1.5		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
102	63	-3.1		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
103	64	-1.6		ION CHROMATOGRAPHY	2, 3, 6, 7
104	68	4.5		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
105	70	7.6		TITRATION, SILVER NITRATE	1, 2, 4
107	63	-3.1		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
108	63	-3.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
112	66	1.5		TITRATION, MERCURIC NITRATE	1, 2, 3, 4
118	63	-3.1		COLORIMETRIC, FERRIC THIOCYANATE	1, 2, 3, 4
121	63	-3.1		TITRATION, SILVER NITRATE	1, 2, 4
122	57	-12.4		NOT REPORTED	
123	66	1.5		TITRATION, SILVER NITRATE	1, 2, 4
124	63	-3.1		NOT REPORTED	
142	63	-3.1		ION CHROMATOGRAPHY	2, 3, 6, 7

78 Labs had a total range of 6 to 92 and a mean of 65.0 with a standard deviation of 3.0 and a 95% confidence interval of the mean +/- 0.7.

Table 10 Standard Reference Water Sample M94 Report for DSRD 180

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	950	0.2	RESIDUE, FILTRABLE	1,3
4	940	-0.9	RESIDUE, FILTRABLE	1,3
6	930	-2.0	RESIDUE, FILTRABLE	1,3
7	953	0.5	RESIDUE, FILTRABLE	1,3
8	964	1.6	RESIDUE, FILTRABLE	1,3
12	967	1.9	RESIDUE, FILTRABLE	1,3
13	976	2.9	RESIDUE, FILTRABLE	1,3
14	960	1.2	RESIDUE ON EVAPORATION	2,4
15	940	-0.9	RESIDUE, FILTRABLE	1,3
19	917	-3.3	RESIDUE ON EVAPORATION	2,4
20	880	-7.2	RESIDUE, FILTRABLE	1,3
22	950	0.2	RESIDUE, FILTRABLE	1,3
24	954	0.6	RESIDUE, FILTRABLE	1,3
25	956	0.8	RESIDUE ON EVAPORATION	2,4
27	945	-0.4	RESIDUE ON EVAPORATION	2,4
29	864	-8.9	RESIDUE, FILTRABLE	1,3
30	957	0.9	RESIDUE, FILTRABLE	1,3
33	940	-0.9	NOT REPORTED	
34	952	0.4	RESIDUE ON EVAPORATION	2,4
36	901	-5.0	RESIDUE, FILTRABLE	1,3
37	935	-1.4	RESIDUE, FILTRABLE	1,3
38	940	-0.9	RESIDUE, FILTRABLE	1,3
40	937	-1.2	RESIDUE ON EVAPORATION	2,4
41	984	3.7	RESIDUE ON EVAPORATION	2,4
43	960	1.2	RESIDUE ON EVAPORATION	2,4
45	956	0.8	RESIDUE ON EVAPORATION	2,4
46	933	-1.6	RESIDUE, FILTRABLE	1,3
47	936	-1.3	RESIDUE, FILTRABLE	1,3
48	930	-2.0	RESIDUE, FILTRABLE	1,3
50	760	-19.9	REJECT NOT REPORTED	
53	957	0.9	RESIDUE ON EVAPORATION	2,4
56	1074	13.2	REJECT RESIDUE ON EVAPORATION	2,4
57	940	-0.9	RESIDUE, FILTRABLE	1,3
58	994	4.8	RESIDUE, FILTRABLE	1,3
59	964	1.6	RESIDUE ON EVAPORATION	2,4
62	946	-0.3	RESIDUE, FILTRABLE	1,3
63	972	2.5	RESIDUE ON EVAPORATION	2,4
64	982	3.5	RESIDUE ON EVAPORATION	2,4
68	947	-0.2	RESIDUE, FILTRABLE	1,3
70	950	0.2	RESIDUE, FILTRABLE	1,3
72	92	-90.3	REJECT RESIDUE ON EVAPORATION	2,4
77	963	1.5	RESIDUE ON EVAPORATION	2,4
79	950	0.2	RESIDUE ON EVAPORATION	2,4
80	1066	12.4	REJECT RESIDUE ON EVAPORATION	2,4
81	941	-0.8	RESIDUE, FILTRABLE	1,3
83	312	-67.1	REJECT RESIDUE, FILTRABLE	1,3
84	1068	12.6	REJECT RESIDUE ON EVAPORATION	2,4
85	938	-1.1	RESIDUE ON EVAPORATION	2,4
91	980	3.3	RESIDUE, FILTRABLE	1,3
93	960	1.2	RESIDUE, FILTRABLE	1,3
94	955	0.7	RESIDUE, FILTRABLE	1,3
95	941	-0.8	RESIDUE ON EVAPORATION	2,4
97	974	2.7	RESIDUE, FILTRABLE	1,3
98	912	-3.8	RESIDUE ON EVAPORATION	2,4
99	884	-6.8	RESIDUE ON EVAPORATION	2,4
102	1017	7.2	RESIDUE ON EVAPORATION	2,4
104	1010	6.5	RESIDUE ON EVAPORATION	2,4
105	929	-2.1	RESIDUE, FILTRABLE	1,3
107	934	-1.5	RESIDUE, FILTRABLE	1,3
108	969	2.2	RESIDUE ON EVAPORATION	2,4
112	966	1.8	RESIDUE, FILTRABLE	1,3
118	952	0.4	RESIDUE, FILTRABLE	1,3
121	981	3.4	RESIDUE, FILTRABLE	1,3
122	910	-4.1	NOT REPORTED	
123	946	-0.3	RESIDUE ON EVAPORATION	2,4
142	940	-0.9	RESIDUE, FILTRABLE	1,3

66 Labs had a total range of 92 to 1074 and a mean of 948.5 with a standard deviation of 27.4 and a 95% confidence interval of the mean +/- 7.1.

Table 10 Standard Reference Water Sample M94 Report for F

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	2.2	-4.5	ION SELECTIVE ELECTRODE	1,2,3,4
2	2.3	-0.2	OTHER	1,2,3,4
4	2.2	-4.5	ION SELECTIVE ELECTRODE	1,2,3,4
6	2.2	-4.5	ION SELECTIVE ELECTRODE	1,2,3,4
7	2.4	4.2	ION SELECTIVE ELECTRODE	1,2,3,4
8	2.5	8.5	COLORIMETRIC, GEROUS ALIZARIN "COMPLEXONE"	3
9	2.2	-4.5	ION SELECTIVE ELECTRODE	1,2,3,4
12	2.5	8.5	COLORIMETRIC, SPADNS	1,2,3
13	2.2	-4.5	ION SELECTIVE ELECTRODE	1,2,3,4
15	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
16	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
19	1.6	-30.5	REJECT ION CHROMATOGRAPHY	2,3,6
20	3.0	30.2	REJECT ION SELECTIVE ELECTRODE	1,2,3,4
22	2.9	25.9	REJECT ION CHROMATOGRAPHY	2,3,6
24	2.2	-4.5	ION SELECTIVE ELECTRODE	1,2,3,4
25	2.1	-8.8	ION SELECTIVE ELECTRODE	1,2,3,4
27	2.2	-4.5	ION SELECTIVE ELECTRODE	1,2,3,4
29	2.4	4.2	OTHER	1,2,3,4
30	2.4	4.2	ION SELECTIVE ELECTRODE	1,2,3,4
33	2.3	-0.2	NOT REPORTED	
34	2.1	-8.8	ION SELECTIVE ELECTRODE	1,2,3,4
36	2.5	8.5	ION CHROMATOGRAPHY	2,3,6
37	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
38	2.3	-0.2	ION CHROMATOGRAPHY	2,3,6
40	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
41	2.5	8.5	COLORIMETRIC, LANTHANUM ALIZARIN "COMPLEXONE"	1
46	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
47	2.0	-13.2	ION CHROMATOGRAPHY	2,3,6
48	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
49	2.5	8.5	ION CHROMATOGRAPHY	2,3,6
50	2.3	-0.2	NOT REPORTED	
54	2.3	-0.2	ION CHROMATOGRAPHY	2,3,6
56	2.6	12.9	ION SELECTIVE ELECTRODE	1,2,3,4
57	2.4	4.2	ION SELECTIVE ELECTRODE	1,2,3,4
58	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
62	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
63	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
68	2.2	-4.5	ION SELECTIVE ELECTRODE	1,2,3,4
70	2.6	12.9	COLORIMETRIC, SPADNS	1,2,3
77	3.4	47.6	REJECT COLORIMETRIC, ZIRCONIUM ERIOCHROME	4
80	2.4	4.2	ION SELECTIVE ELECTRODE	1,2,3,4
81	2.4	4.2	ION SELECTIVE ELECTRODE	1,2,3,4
83	2.2	-4.5	ION SELECTIVE ELECTRODE	1,2,3,4
85	2.3	-0.2	COLORIMETRIC, ZIRCONIUM ERIOCHROME	4
86	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
87	2.2	-4.5	ION CHROMATOGRAPHY	2,3,6
90	2.1	-8.8	ION SELECTIVE ELECTRODE	1,2,3,4
91	2.2	-4.5	ION SELECTIVE ELECTRODE	1,2,3,4
93	2.5	8.5	ION SELECTIVE ELECTRODE	1,2,3,4
97	2.1	-8.8	COLORIMETRIC, LANTHANUM ALIZARIN "COMPLEXONE"	1
98	2.4	4.2	COLORIMETRIC, LANTHANUM ALIZARIN "COMPLEXONE"	1
99	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
103	2.8	21.5	REJECT ION CHROMATOGRAPHY	2,3,6
104	2.2	-4.5	ION SELECTIVE ELECTRODE	1,2,3,4
105	2.3	-0.2	COLORIMETRIC, SPADNS	1,2,3
112	2.3	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
123	2.5	8.5	ION SELECTIVE ELECTRODE	1,2,3,4
142	2.1	-8.8	ION SELECTIVE ELECTRODE	1,2,3,4

58 Labs had a total range of 1.6 to 3.4 and a mean of 2.30 with a standard deviation of 0.13 and a 95% confidence interval of the mean +/- 0.04.

Table 10 Standard Reference Water Sample M94 Report for I

Code Number	Reported value	Pct. dev. from mean	Methods	References
22	< 1000		IGNORED ION CHROMATOGRAPHY	2,3,6
57	10	0.0	COLORIMETRIC, CERIC ARSENIOS OXIDATION	2,4
127	< 100		IGNORED OTHER	

3 Labs had a total range of 10 to 10.  
INSUFFICIENT DATA FOR DETERMINATION OF MEAN AND STANDARD OF DEVIATION.

Table 10 Standard Reference Water Sample M94 Report for K

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	6.0	-2.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	6.3	2.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	6.0	-2.8	EMISSION, IC PLASMA	3
8	6.1	-1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	6.0	-2.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	8.4	36.1	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	6.8	10.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	7.0	13.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	5.0	-19.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	5.0	-19.0	EMISSION, IC PLASMA	3
24	5.9	-4.4	NOT REPORTED	
25	7.0	13.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
27	6.6	6.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	6.2	0.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	7.4	19.9	EMISSION, FLAME, PHOTOMETRIC	1,2
33	5.4	-12.5	NOT REPORTED	
34	5.9	-4.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
36	6.3	2.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	6.2	0.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	7.4	19.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
40	5.8	-6.0	EMISSION, IC PLASMA	3
41	6.9	11.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
42	5.3	-14.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	6.5	5.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
46	6.3	2.1	EMISSION, FLAME, PHOTOMETRIC	1,2
47	6.6	6.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	7.0	13.4	EMISSION, FLAME, PHOTOMETRIC	1,2
49	5.9	-4.4	EMISSION, IC PLASMA	3
50	6.7	8.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
52	6.2	0.4	EMISSION, IC PLASMA	3
53	6.4	3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	5.3	-14.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
56	9.0	45.8	REJECT EMISSION, FLAME, PHOTOMETRIC	1,2
57	5.9	-4.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	4.7	-23.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
61	6.6	6.9	EMISSION, IC PLASMA	3
62	5.3	-14.1	EMISSION, FLAME, PHOTOMETRIC	1,2
63	2.1	-66.0	REJECT EMISSION, FLAME, PHOTOMETRIC	1,2
64	6.8	10.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
68	6.3	2.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	5.9	-4.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	5.7	-7.7	EMISSION, IC PLASMA	3
77	7.1	15.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	7.5	21.5	OTHER	
81	10.5	70.1	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	7.0	13.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	6.8	10.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
85	6.2	0.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	5.8	-6.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
87	6.2	0.4	OTHER	
90	6.1	-1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	5.5	-10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	6.1	-1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
97	5.3	-14.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
98	5.9	-4.4	EMISSION, IC PLASMA	3
99	6.0	-2.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
102	6.0	-2.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
103	5.9	-4.4	EMISSION, IC PLASMA	3
104	6.1	-1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
105	7.0	13.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
108	5.7	-7.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	7.0	13.4	EMISSION, IC PLASMA	3
111	5.2	-15.8	NOT REPORTED	
112	6.2	0.4	EMISSION, IC PLASMA	3
113	5.6	-9.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	7.0	13.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	6.0	-2.8	OTHER	
122	15.6	152.7	REJECT NOT REPORTED	
123	5.7	-7.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
142	5.7	-7.7	EMISSION, IC PLASMA	3

70 Labs had a total range of 2.1 to 15.6 and a mean of 6.17 with a standard deviation of 0.63 and a 95% confidence interval of the mean +/- 0.16.

Table 10 Standard Reference Water Sample M94 Report for MG

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	48	2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	48	2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	48	2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	48	2.7	EMISSION, IC PLASMA	3,5
8	48	2.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
9	48	2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	46	-1.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	44	-5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	47	0.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	45	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	30	-35.8	REJECT EMISSION, IC PLASMA	3,5
24	44	-5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	46	-1.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
27	47	0.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	97	107.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	48	2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	47	0.6	NOT REPORTED	
34	47	0.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
36	47	0.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	45	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	51	9.1	EMISSION, IC PLASMA	3,5
40	46	-1.6	EMISSION, IC PLASMA	3,5
41	46	-1.6	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
42	47	0.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	51	9.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
46	44	-5.9	TITRATION, EDTA	2
47	49	4.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	47	0.6	EMISSION, IC PLASMA	3,5
49	46	-1.6	EMISSION, IC PLASMA	3,5
50	34	-27.3	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
52	47	0.6	EMISSION, IC PLASMA	3,5
53	54	15.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	45	-3.7	EMISSION, IC PLASMA	3,5
56	61	30.5	REJECT TITRATION, EDTA	2
57	47	0.6	EMISSION, IC PLASMA	3,5
59	58	24.1	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
61	50	7.0	EMISSION, IC PLASMA	3,5
62	48	2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
63	42	-10.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
68	46	-1.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
68	50	7.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	44	-5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	46	-1.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	46	-1.6	EMISSION, IC PLASMA	3,5
77	51	9.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	47	0.6	EMISSION, IC PLASMA	3,5
81	45	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	51	9.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
84	61	30.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
85	48	2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	47	0.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
87	46	-1.6	OTHER	
90	42	-10.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	44	-5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	47	0.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
97	46	-1.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
98	41	-12.3	EMISSION, IC PLASMA	3,5
99	44	-5.9	TITRATION, EDTA	2
102	42	-10.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
103	48	2.7	EMISSION, IC PLASMA	3,5
104	49	4.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
105	48	2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
108	46	-1.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	45	-3.7	EMISSION, IC PLASMA	3,5
111	48	2.7	EMISSION, IC PLASMA	3,5
112	48	2.7	OTHER	
113	44	-5.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	50	7.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	47	0.6	TITRATION, EDTA	2
122	44	-5.9	NOT REPORTED	
123	48	2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	55	17.7	REJECT OTHER	
142	46	-1.6	EMISSION, IC PLASMA	3,5

73 Labs had a total range of 30 to 97 and a mean of 46.7 with a standard deviation of 2.4 and a 95% confidence interval of the mean +/- 0.6.

Table 10 Standard Reference Water Sample M94 Report for NA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	142	2.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	135	-2.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	143	3.3	EMISSION, FLAME	1,2,3,4
6	138	-0.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	152	9.8	EMISSION, IC PLASMA	3,5
8	145	4.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	135	-2.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	132	-4.6	EMISSION, IC PLASMA	3,5
15	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	105	-24.1	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	128	-7.5	EMISSION, IC PLASMA	3,5
24	142	2.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	122	-11.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
27	139	0.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	148	6.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	138	-0.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	135	-2.5	NOT REPORTED	
34	149	7.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
36	144	4.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	121	-12.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	145	4.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
40	135	-2.5	EMISSION, IC PLASMA	3,5
41	131	-5.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
42	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
46	145	4.8	EMISSION, FLAME	1,2
47	130	-6.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	140	1.2	EMISSION, IC PLASMA	3,5
49	135	-2.5	EMISSION, IC PLASMA	3,5
50	139	0.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
52	138	-0.3	EMISSION, IC PLASMA	3,5
53	139	0.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	134	-3.2	EMISSION, IC PLASMA	3,5
56	133	-3.9	EMISSION, FLAME	1,2
57	137	-1.0	EMISSION, IC PLASMA	3,5
59	163	17.8	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
61	140	1.2	EMISSION, IC PLASMA	3,5
62	144	4.0	EMISSION, FLAME	1,2
63	201	45.2	REJECT EMISSION, FLAME	1,2
68	133	-3.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	139	0.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
70	142	2.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	150	8.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	147	6.2	EMISSION, IC PLASMA	3,5
77	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	141	1.9	EMISSION, IC PLASMA	3,5
81	123	-11.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	137	-1.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	174	25.7	REJECT EMISSION, FLAME	1,2
85	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
87	138	-0.3	OTHER	
90	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
91	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	138	-0.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
97	137	-1.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
98	126	-9.0	EMISSION, IC PLASMA	3,5
99	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
102	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
103	135	-2.5	EMISSION, IC PLASMA	3,5
104	142	2.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
105	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
108	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	130	-6.1	EMISSION, IC PLASMA	3,5
111	142	2.6	EMISSION, IC PLASMA	3,5
112	144	4.0	OTHER	
113	135	-2.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	143	3.3	EMISSION, FLAME	1,2
122	145	4.8	NOT REPORTED	
123	140	1.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	160	15.6	REJECT OTHER	
142	130	-6.1	EMISSION, IC PLASMA	3,5

74 Labs had a total range of 105 to 201 and a mean of 138.4 with a standard deviation of 6.1 and a 95% confidence interval of the mean +/- 1.5.

Table 10 Standard Reference Water Sample M94 Report for NO2-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	<0.020		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
6	<0.002		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
7	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
8	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
12	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
13	0.002	-67.2	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
14	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
15	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
19	0.006	-1.6	IGNORED ION CHROMATOGRAPHY	2,3,6
20	<0.010		IGNORED ION CHROMATOGRAPHY	2,3,6
21	<0.200		IGNORED ION CHROMATOGRAPHY	2,3,6
22	<0.002	-67.2	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
23	0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
24	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
27	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
29	0.010	63.9	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
30	<0.005		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
34	<0.000		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
36	<0.005		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
37	<0.005		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
38	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
40	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
41	0.480	7768.9	REJECT COLORIMETRIC, DIAZOTIZATION	1,3,4
42	0.002	-67.2	REJECT COLORIMETRIC, DIAZOTIZATION	1,3,4
43	<0.020		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
44	3.400	6E+04	REJECT ION CHROMATOGRAPHY	2,3,6
46	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
48	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
50	0.020	227.9	NOT REPORTED	
53	<0.020		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
54	<0.050		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
56	0.003	-50.8	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
57	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
58	0.017	178.7	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
64	0.001	-83.6	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
65	<0.001		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
73	0.003	-50.8	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
74	<0.020		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
79	0.003	-50.8	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
81	<0.050		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
83	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
85	0.010	63.9	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
86	0.002	-67.2	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
91	0.003	-50.8	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
92	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
93	0.003		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
95	<0.020	-50.8	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
98	<0.020		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
99	0.004	-34.4	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
108	<0.010		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
112	0.010	63.9	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
113	0.009	47.5	IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4
121	<0.010		IGNORED OTHER	
142	<0.100		IGNORED COLORIMETRIC, DIAZOTIZATION	1,3,4

52 Labs had a total range of 0.001 to 3.400 and a mean of .0061 with a standard deviation of .0055 and a 95% confidence interval of the mean +/- .0027.



Table 10 Standard Reference Water Sample M94 Report for NO3-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	3.3	-24.3	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
2	4.8	10.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
4	4.4	0.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
6	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
7	4.4	0.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
8	4.9	12.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
9	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
12	4.7	7.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
13	4.5	3.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
14	4.6	5.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
15	4.1	-5.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
19	22.8	423.1	REJECT ION CHROMATOGRAPHY	2, 3, 6, 7
20	0.6	-86.2	REJECT COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
22	4.3	-1.4	ION CHROMATOGRAPHY	2, 3, 6, 7
23	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
24	4.4	0.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
25	4.7	7.8	OTHER	
27	3.8	-12.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
29	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
30	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
33	4.5	3.2	NOT REPORTED	
34	4.1	-5.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
36	4.6	5.5	ION CHROMATOGRAPHY	2, 3, 6, 7
37	4.4	0.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
38	4.6	5.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
40	5.0	14.7	COLORIMETRIC, BRUCINE	1, 2, 3, 4
41	4.1	-5.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
42	4.4	0.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
43	5.2	19.3	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
45	4.2	-3.6	OTHER	
46	3.6	-17.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
47	4.9	12.4	COLORIMETRIC, BRUCINE	1, 2, 3, 4
48	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
50	4.4	0.9	NOT REPORTED	
52	4.2	-3.6	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
53	4.1	-5.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
56	4.2	-3.6	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
57	5.2	19.3	COLORIMETRIC, BRUCINE	1, 2, 3, 4
58	4.4	0.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
59	3.9	-10.5	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
62	4.0	-8.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
63	5.0	14.7	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
64	4.5	3.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
65	4.4	0.9	OTHER	
68	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
70	3.6	-17.4	COLORIMETRIC, BRUCINE	1, 2, 3, 4
72	4.3	-1.4	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
73	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
74	4.2	-3.6	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
79	3.5	-19.7	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
80	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
81	5.4	23.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
82	4.0	-8.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
83	4.8	10.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
84	4.4	0.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
85	4.4	0.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
86	4.4	0.9	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
87	4.5	3.2	ION CHROMATOGRAPHY	2, 3, 6, 7
90	6.4	46.8	REJECT ION SELECTIVE ELECTRODE	1, 2, 3, 4
91	3.9	-10.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
93	4.8	10.1	OTHER	
95	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
97	4.5	3.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
98	4.8	10.1	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
99	4.3	-1.4	COLORIMETRIC, BRUCINE	1, 2, 3, 4
103	4.4	0.9	ION CHROMATOGRAPHY	2, 3, 6, 7
104	3.5	-19.7	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
105	3.8	-12.8	COLORIMETRIC, BRUCINE	1, 2, 3, 4
107	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
108	4.5	3.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
112	4.0	-8.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
118	4.3	-1.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
121	4.4	0.9	OTHER	
123	5.0	14.7	OTHER	
142	4.6	5.5	ION CHROMATOGRAPHY	2, 3, 6, 7

76 Labs had a total range of 0.6 to 22.8 and a mean of 4.36 with a standard deviation of 0.40 and a 95% confidence interval of the mean +/- 0.09.

Table 10 Standard Reference Water Sample M94 Report for P, TOTAL

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1.01	-12.2	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
2	1.13	-1.8	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
3	1.18	2.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
4	1.14	-0.9	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
6	1.17	1.7	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
7	1.15	-0.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
8	1.12	-2.7	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
9	1.12	-2.7	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
12	1.00	-13.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
13	1.26	9.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
14	1.03	-10.5	NOT REPORTED	
15	1.17	1.7	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
20	0.06	-94.8	REJECT COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
22	1.00	-13.1	EMISSION, IC PLASMA	3,5
23	1.18	2.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
24	1.42	23.4	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
29	1.27	10.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
30	1.18	2.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
33	1.00	-13.1	NOT REPORTED	
34	1.08	-6.2	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
37	1.09	-5.3	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
38	1.10	-4.4	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
40	1.15	-0.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
41	3.16	0.8	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
42	3.49	203.3	REJECT COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
43	1.15	-0.1	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
44	1.02	-11.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
47	1.40	21.7	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
48	1.18	2.5	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
49	1.21	5.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
53	1.16	0.8	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
54	1.14	-0.9	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
56	0.04	-96.5	REJECT COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
57	1.10	-4.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
58	1.03	-10.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
59	1.22	6.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
63	7.00	508.3	REJECT COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
64	1.15	-0.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
65	1.11	-3.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
69	1.07	-7.0	OTHER	
72	1.10	-4.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
73	1.22	6.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
74	1.19	3.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
79	0.60	-47.9	REJECT COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
80	1.12	-2.7	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
81	0.82	-28.7	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
82	1.16	0.8	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
83	1.10	-4.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
84	1.16	0.8	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
85	1.20	4.3	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
86	1.10	-4.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
87	1.20	4.3	OTHER	
90	1.70	47.7	REJECT OTHER	
91	1.00	-13.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
93	1.00	-13.1	OTHER	
95	1.17	1.7	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
97	1.21	5.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
98	1.15	-0.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
99	1.15	-0.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
104	1.24	7.8	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
105	0.57	-50.5	REJECT COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
107	1.18	2.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
108	1.47	27.7	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
110	1.50	30.3	EMISSION, IC PLASMA	3,5
111	1.24	7.8	EMISSION, IC PLASMA	3,5
112	1.10	-4.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
113	0.34	-70.5	REJECT COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1,2,3,4
118	1.19	3.4	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
121	1.08	-6.2	OTHER	
127	1.20	4.3	OTHER	
142	1.20	4.3	EMISSION, IC PLASMA	3,5

71 Labs had a total range of 0.04 to 7.00 and a mean of 1.151 with a standard deviation of 0.111 and a 95% confidence interval of the mean +/- 0.028.

Table 10 Standard Reference Water Sample M94 Report for PH

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	8.50	-0.8	ELECTROMETRIC	1, 2, 3, 4
2	8.62	0.6	ELECTROMETRIC	1, 2, 3, 4
3	8.54	-0.4	ELECTROMETRIC	1, 2, 3, 4
4	8.56	-0.1	ELECTROMETRIC	1, 2, 3, 4
6	8.63	0.7	ELECTROMETRIC	1, 2, 3, 4
7	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
8	8.44	-1.5	ELECTROMETRIC	1, 2, 3, 4
9	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
12	8.50	-0.8	ELECTROMETRIC	1, 2, 3, 4
13	8.68	1.3	ELECTROMETRIC	1, 2, 3, 4
14	8.40	-2.0	ELECTROMETRIC	1, 2, 3, 4
15	8.63	0.7	ELECTROMETRIC	1, 2, 3, 4
16	8.67	1.2	ELECTROMETRIC	1, 2, 3, 4
19	8.21	-4.2	REJECT	1, 2, 3, 4
20	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
22	8.65	0.9	ELECTROMETRIC	1, 2, 3, 4
23	8.46	-1.3	ELECTROMETRIC	1, 2, 3, 4
24	8.64	0.8	ELECTROMETRIC	1, 2, 3, 4
25	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
27	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
29	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
30	8.58	0.1	ELECTROMETRIC	1, 2, 3, 4
33	7.70	-10.2	REJECT	1, 2, 3, 4
34	8.65	0.9	ELECTROMETRIC	1, 2, 3, 4
36	8.69	1.4	ELECTROMETRIC	1, 2, 3, 4
37	8.70	1.5	ELECTROMETRIC	1, 2, 3, 4
38	8.55	-0.2	ELECTROMETRIC	1, 2, 3, 4
40	8.68	1.3	ELECTROMETRIC	1, 2, 3, 4
41	8.00	-6.7	REJECT	1, 2, 3, 4
42	8.57	-0.0	ELECTROMETRIC	1, 2, 3, 4
43	8.70	1.5	ELECTROMETRIC	1, 2, 3, 4
44	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
46	8.50	-0.8	ELECTROMETRIC	1, 2, 3, 4
47	8.50	-0.8	ELECTROMETRIC	1, 2, 3, 4
48	8.55	-0.2	ELECTROMETRIC	1, 2, 3, 4
49	8.64	0.8	ELECTROMETRIC	1, 2, 3, 4
50	8.57	-0.0	NOT REPORTED	1, 2, 3, 4
52	8.50	-0.8	ELECTROMETRIC	1, 2, 3, 4
53	8.44	-1.5	ELECTROMETRIC	1, 2, 3, 4
54	8.63	0.7	ELECTROMETRIC	1, 2, 3, 4
56	8.65	0.9	ELECTROMETRIC	1, 2, 3, 4
57	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
58	8.59	0.2	ELECTROMETRIC	1, 2, 3, 4
59	8.59	0.2	ELECTROMETRIC	1, 2, 3, 4
62	8.24	-3.9	REJECT	1, 2, 3, 4
63	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
64	8.80	2.7	ELECTROMETRIC	1, 2, 3, 4
68	8.50	-0.8	ELECTROMETRIC	1, 2, 3, 4
69	8.71	1.6	NOT REPORTED	1, 2, 3, 4
70	8.63	0.7	ELECTROMETRIC	1, 2, 3, 4
72	8.59	0.2	ELECTROMETRIC	1, 2, 3, 4
73	8.50	-0.8	ELECTROMETRIC	1, 2, 3, 4
74	8.32	-2.9	ELECTROMETRIC	1, 2, 3, 4
77	8.50	-0.8	ELECTROMETRIC	1, 2, 3, 4
79	8.40	-2.0	ELECTROMETRIC	1, 2, 3, 4
80	8.47	-1.2	ELECTROMETRIC	1, 2, 3, 4
81	8.30	-3.2	ELECTROMETRIC	1, 2, 3, 4
83	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
84	8.51	-0.7	ELECTROMETRIC	1, 2, 3, 4
85	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
86	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
87	8.63	0.7	ELECTROMETRIC	1, 2, 3, 4
90	8.72	1.7	ELECTROMETRIC	1, 2, 3, 4
91	8.67	1.2	ELECTROMETRIC	1, 2, 3, 4
93	8.67	1.2	ELECTROMETRIC	1, 2, 3, 4
94	8.70	1.5	ELECTROMETRIC	1, 2, 3, 4
95	8.63	0.7	ELECTROMETRIC	1, 2, 3, 4
97	8.53	-0.5	ELECTROMETRIC	1, 2, 3, 4
98	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
99	8.51	-0.7	ELECTROMETRIC	1, 2, 3, 4
102	8.35	-2.6	ELECTROMETRIC	1, 2, 3, 4
103	8.63	0.7	ELECTROMETRIC	1, 2, 3, 4
104	8.12	-5.3	REJECT	1, 2, 3, 4
105	8.50	-0.8	ELECTROMETRIC	1, 2, 3, 4
107	8.47	-1.2	ELECTROMETRIC	1, 2, 3, 4
108	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
111	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
112	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
113	8.45	-1.4	ELECTROMETRIC	1, 2, 3, 4
118	8.60	0.3	ELECTROMETRIC	1, 2, 3, 4
122	8.50	-0.8	NOT REPORTED	1, 2, 3, 4
123	8.64	0.8	ELECTROMETRIC	1, 2, 3, 4
124	8.50	-0.8	NOT REPORTED	1, 2, 3, 4
142	8.36	-2.5	ELECTROMETRIC	1, 2, 3, 4

84 Labs had a total range of 7.70 to 8.80 and a mean of 8.571 with a standard deviation of 0.096 and a 95% confidence interval of the mean +/- 0.022.

Table 10 Standard Reference Water Sample M94 Report for SIO2

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	21	15.9	EMISSION, IC PLASMA	5
2	17	-6.2	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
6	18	-0.7	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3
8	17	-6.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
9	16	-11.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
12	18	-0.7	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3
13	19	4.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
14	19	4.8	NOT REPORTED	
15	19	4.8	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
20	20	10.4	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
22	13	-28.3	EMISSION, IC PLASMA	5
23	16	-11.7	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
24	17	-6.2	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
25	17	-6.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
30	19	4.8	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3
34	16	-11.7	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
36	18	-0.7	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
38	17	-6.2	EMISSION, IC PLASMA	5
40	17	-6.2	EMISSION, IC PLASMA	5
41	14	-22.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
42	19	4.8	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
47	19	4.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
48	19	4.8	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
49	17	-6.2	EMISSION, IC PLASMA	5
52	18	-0.7	EMISSION, IC PLASMA	5
54	21	15.9	EMISSION, IC PLASMA	5
57	18	-0.7	EMISSION, IC PLASMA	5
63	25	38.0	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
64	18	-0.7	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
68	16	-11.7	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
73	18	-0.7	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
80	18	-0.7	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
82	18	-0.7	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
83	20	10.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
85	19	4.8	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
91	13	-28.3	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
93	12	-33.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
95	19	4.8	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
98	18	-0.7	EMISSION, IC PLASMA	5
99	21	15.9	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	3
103	17	-6.2	EMISSION, IC PLASMA	5
104	23	26.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
105	17	-6.2	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
108	19	4.8	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
110	24	32.4	EMISSION, IC PLASMA	5
112	19	4.8	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
118	18	-0.7	COLORIMETRIC, MOLYBDOSILICIC ACID	1,2,3
123	18	-0.7	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	4
142	19	4.8	EMISSION, IC PLASMA	5

49 Labs had a total range of 12 to 25 and a mean of 18.1 with a standard deviation of 2.4 and a 95% confidence interval of the mean +/- 0.7.

Table 10 Standard Reference Water Sample M94 Report for SO4

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	380	0.6	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
2	400	5.9	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
4	400	5.9	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
6	370	-2.0	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
7	455	20.5	REJECT COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
8	390	3.3	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
9	365	-3.3	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
12	385	2.0	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
13	405	7.2	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
14	200	-47.0	REJECT GRAVIMETRIC, BARIUM SULFATE	1, 2, 3
15	340	-10.0	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
19	295	-21.9	REJECT ION CHROMATOGRAPHY	2, 6, 7
20	35	-90.7	REJECT TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
22	395	4.6	ION CHROMATOGRAPHY	2, 6, 7
24	20	-94.7	REJECT COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
25	360	-4.7	THORIN TITRATION	2, 4
27	380	0.6	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
29	410	8.6	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
30	390	3.3	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
33	385	2.0	NOT REPORTED	
34	335	-11.3	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
36	385	2.0	ION CHROMATOGRAPHY	2, 6, 7
37	370	-2.0	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
38	350	-7.3	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
40	360	-4.7	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
41	395	4.6	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
42	355	-6.0	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
43	290	-23.2	REJECT TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
45	350	-7.3	GRAVIMETRIC, BARIUM SULFATE	1, 2, 3
46	390	3.3	GRAVIMETRIC, BARIUM SULFATE	1, 2, 3
47	270	-28.5	REJECT ION CHROMATOGRAPHY	2, 6, 7
48	405	7.2	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
49	390	3.3	ION CHROMATOGRAPHY	2, 6, 7
50	385	2.0	NOT REPORTED	
52	380	0.6	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
53	360	-4.7	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
54	365	-3.3	ION CHROMATOGRAPHY	2, 6, 7
56	375	-0.7	ION CHROMATOGRAPHY	2, 6, 7
57	375	-0.7	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
58	385	2.0	GRAVIMETRIC, BARIUM SULFATE	1, 2, 3
59	365	-3.3	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
62	385	2.0	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
63	240	-36.4	REJECT ION CHROMATOGRAPHY	2, 6, 7
68	380	0.6	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
70	380	0.6	OTHER	
72	320	-15.3	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
77	380	0.6	THORIN TITRATION	2, 4
79	405	7.2	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
80	375	-0.7	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
81	310	-17.9	REJECT COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
83	380	0.6	GRAVIMETRIC, BARIUM SULFATE	1, 2, 3
84	260	-31.1	REJECT COLORIMETRIC, CHLORANILATE	3
85	390	3.3	THORIN TITRATION	2, 4
86	350	-7.3	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
87	400	5.9	ION CHROMATOGRAPHY	2, 6, 7
90	170	-55.0	REJECT TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
91	390	3.3	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
93	370	-2.0	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
97	365	-3.3	COLORIMETRIC, CHLORANILATE	3
98	380	0.6	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
99	355	-6.0	GRAVIMETRIC, BARIUM SULFATE	1, 2, 3
102	135	-64.3	REJECT TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
103	385	2.0	ION CHROMATOGRAPHY	2, 6, 7
104	405	7.2	GRAVIMETRIC, BARIUM SULFATE	1, 2, 3
105	375	-0.7	GRAVIMETRIC, BARIUM SULFATE	1, 2, 3
107	390	3.3	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
108	400	5.9	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
112	370	-2.0	TURBIDIMETRIC, BARIUM SULFATE	1, 2, 3
118	390	3.3	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
121	375	-0.7	COLORIMETRIC, METHYL THYMOL BLUE	1, 3, 4
122	360	-4.7	NOT REPORTED	
123	380	0.6	GRAVIMETRIC, BARIUM SULFATE	1, 2, 3
142	370	-2.0	ION CHROMATOGRAPHY	2, 6, 7

73 Labs had a total range of 20 to 455 and a mean of 377.6 with a standard deviation of 18.4 and a 95% confidence interval of the mean +/- 4.7.

Table 10 Standard Reference Water Sample M94 Report for SP. COND.

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1300	-2.3	DIRECT READING INSTRUMENT	4
2	1370	2.9	DIRECT READING INSTRUMENT	4
3	1450	9.0	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
4	1180	-11.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
6	1350	1.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
7	1280	-3.8	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
8	1360	2.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
9	1260	-5.3	DIRECT READING INSTRUMENT	4
12	1410	6.0	DIRECT READING INSTRUMENT	4
13	1360	2.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
14	1370	2.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
15	1300	-2.3	DIRECT READING INSTRUMENT	4
16	1380	3.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
19	940	-29.4	REJECT DIRECT READING INSTRUMENT	4
20	1300	-2.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
22	1330	-0.1	DIRECT READING INSTRUMENT	4
23	1390	4.4	DIRECT READING INSTRUMENT	4
24	1320	-0.8	DIRECT READING INSTRUMENT	4
25	1350	1.4	DIRECT READING INSTRUMENT	4
27	1260	-5.3	DIRECT READING INSTRUMENT	4
29	1170	-12.1	DIRECT READING INSTRUMENT	4
30	1360	2.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
33	1400	5.2	NOT REPORTED	4
34	1370	2.9	DIRECT READING INSTRUMENT	4
36	1350	1.4	DIRECT READING INSTRUMENT	4
37	1340	0.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
38	1390	4.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
40	720	-45.9	REJECT DIRECT READING INSTRUMENT	4
41	1380	3.7	DIRECT READING INSTRUMENT	4
43	1310	-1.6	DIRECT READING INSTRUMENT	4
46	1330	-0.1	DIRECT READING INSTRUMENT	4
47	1270	-4.6	DIRECT READING INSTRUMENT	4
48	1370	2.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
49	1390	4.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
50	1400	5.2	NOT REPORTED	4
52	1370	2.9	DIRECT READING INSTRUMENT	4
53	1370	2.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
54	1350	1.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
56	1260	-5.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
57	1350	1.4	DIRECT READING INSTRUMENT	4
58	1330	-0.1	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
59	1250	-6.1	DIRECT READING INSTRUMENT	4
62	1370	2.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
63	1410	6.0	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
64	1360	2.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
68	1370	2.9	DIRECT READING INSTRUMENT	4
69	950	-28.6	REJECT NOT REPORTED	4
70	1600	20.2	REJECT WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
72	1100	-17.3	DIRECT READING INSTRUMENT	4
73	1370	2.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
74	1330	-0.1	DIRECT READING INSTRUMENT	4
77	1360	2.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
79	1360	2.2	ELECTRODELESS, INDUCTIVE CELL-TYPE	2
80	1370	2.9	DIRECT READING INSTRUMENT	4
81	1100	-17.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
82	1380	3.7	DIRECT READING INSTRUMENT	4
83	1260	-5.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
84	1200	-9.8	DIRECT READING INSTRUMENT	4
85	1370	2.9	DIRECT READING INSTRUMENT	4
86	1180	-11.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
87	1280	-3.8	DIRECT READING INSTRUMENT	4
90	1100	-17.3	DIRECT READING INSTRUMENT	4
91	1300	-2.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
93	1370	2.9	DIRECT READING INSTRUMENT	4
94	1300	-2.3	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
95	1370	2.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
97	1360	2.2	DIRECT READING INSTRUMENT	4
98	1330	-0.1	DIRECT READING INSTRUMENT	4
99	1330	-0.1	DIRECT READING INSTRUMENT	4
99	1330	-0.1	DIRECT READING INSTRUMENT	4
102	1370	2.9	DIRECT READING INSTRUMENT	4
103	1440	8.2	DIRECT READING INSTRUMENT	4
104	1320	-0.8	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
105	1340	0.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
107	1410	6.0	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
108	1330	-0.1	DIRECT READING INSTRUMENT	4
111	1370	2.9	DIRECT READING INSTRUMENT	4
118	1400	5.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
123	1370	2.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
142	1400	5.2	DIRECT READING INSTRUMENT	4

80 Labs had a total range of 720 to 1600 and a mean of 1331 with a standard deviation of 73 and a 95% confidence interval of the mean +/- 17.

Table 10 Standard Reference Water Sample M94 Report for SR

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1120	4.1	EMISSION, IC PLASMA	3,5
6	1100	2.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
8	1060	-1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
13	1120	4.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
15	1300	20.8	OTHER	
22	965	-10.3	EMISSION, IC PLASMA	3,5
24	920	-14.5	ATOMIC ABSORPTION, FLAMELESS	7
36	860	-20.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
41	690	-35.9	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
47	1030	-4.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
49	1100	2.3	EMISSION, IC PLASMA	3,5
52	1000	-7.0	EMISSION, IC PLASMA	3,5
53	1200	11.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
54	1220	13.4	EMISSION, IC PLASMA	3,5
57	1100	2.3	EMISSION, IC PLASMA	3,5
63	32	-97.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
68	1000	-7.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
73	1080	0.4	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7
80	1100	2.3	ATOMIC ABSORPTION, FLAMELESS	7
85	980	-8.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
93	1106	2.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
98	1080	0.4	EMISSION, IC PLASMA	3,5
103	1082	0.6	EMISSION, IC PLASMA	3,5
104	1100	2.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
110	1070	-0.5	EMISSION, IC PLASMA	3,5
112	1	-99.9	REJECT OTHER	
127	1100	2.3	OTHER	
142	1100	2.3	EMISSION, IC PLASMA	3,5

28 Labs had a total range of 1 to 1300 and a mean of 1076 with a standard deviation of 91.7 and a 95% confidence interval of the mean +/- 92.

Table 10 Standard Reference Water Sample M94 Report for V

Code Number	Reported value	Pct. dev. from mean	Methods	References
6	<100.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
8	<100.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
13	16.4	325.4	REJECT EMISSION, IC PLASMA	3,5
15	< 2.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3,5
22	101.0	2519.6	REJECT EMISSION, IC PLASMA	3,5
24	73.0	1793.3	REJECT ATOMIC ABSORPTION, FLAMELESS	3,5
38	< 10.0		IGNORED EMISSION, IC PLASMA	3,5
41	<200.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
47	1.4	-63.7	ATOMIC ABSORPTION, FLAMELESS	3
49	2.0	-48.1	EMISSION, IC PLASMA	3,5
52	7.0	81.6	EMISSION, IC PLASMA	3,5
54	2.7	-30.0	EMISSION, IC PLASMA	3,5
57	1.0	-74.1	COLORIMETRIC, CATALYTIC OXIDATION	4
63	< 40.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
73	6.0	55.6	EMISSION, IC PLASMA	3,5
80	2.0	-48.1	EMISSION, IC PLASMA	3,5
83	<100.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
93	<100.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
98	< 3.0		IGNORED EMISSION, IC PLASMA	3,5
103	< 2.0		IGNORED EMISSION, IC PLASMA	3,5
104	200.0	5087.3	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
110	10.0	159.4	EMISSION, IC PLASMA	3,5
127	< 20.0		IGNORED OTHER	
142	2.6	-32.6	EMISSION, IC PLASMA	3,5

24 Labs had a total range of 1.0 to 200.0 and a mean of 3.86 with a standard deviation of 3.09 and a 95% confidence interval of the mean +/- 2.37.

Table 11. Statistics by method for standard reference sample M94

Determination	Method	Range:		Mean	Standard Deviation	N
		from	to			
ALK(CACO3)	'TITRATION, COLORIMETRIC'	215.000	- 280.000	247.750	12.456	20
	'TITRATION, ELECTROMETRIC'	222.000	- 500.000	243.413	7.658	46
	NOT REPORTED	225.000	- 239.000	233.000	7.211	3
	OTHER	240.000	- 246.000	243.250	2.754	4
	_OVER-ALL_	215.000	- 500.000	244.070	7.782	71
B	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	2000.000	-2000.000	-----	-----	-
	COLORIMETRIC, AZOMETHINE	215.000	- 250.000	231.250	14.361	4
	COLORIMETRIC, CARMINE (GARMINIC ACID)	200.000	- 480.000	366.667	147.422	3
	COLORIMETRIC, CURCUMIN	60.000	- 313.000	220.250	80.688	8
	EMISSION, DC PLASMA	216.000	- 257.000	-----	-----	-
	EMISSION, IC PLASMA	141.000	- 262.000	211.385	29.293	13
	_OVER-ALL_	60.000	-2000.000	223.857	35.345	28
BR	ION CHROMATOGRAPHY	230.000	- 420.000	365.333	75.807	6
	_OVER-ALL_	0.000	-5900.000	393.000	210.514	9
CA	ATOMIC ABSORPTION, DIRECT, AIR	79.000	- 143.000	99.395	7.678	38
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	96.000	- 112.000	101.833	5.382	6
	EMISSION, IC PLASMA	90.000	- 113.000	101.588	5.001	17
	OTHER	10.000	- 110.000	100.500	7.047	4
	TITRATION, EDTA	104.000	- 180.000	106.500	1.915	4
_OVER-ALL_	10.000	- 180.000	100.465	6.897	71	
CL	COLORIMETRIC, FERRIC THIOCYANATE	46.000	- 92.000	65.120	2.743	25
	ION CHROMATOGRAPHY	59.000	- 74.000	65.364	4.081	11
	NOT REPORTED	57.000	- 63.000	59.667	3.055	3
	TITRATION, MERCURIC NITRATE	63.000	- 70.000	65.471	2.183	17
	TITRATION, SILVER NITRATE	50.000	- 76.000	65.733	2.017	15
	_OVER-ALL_	6.000	- 92.000	65.041	2.997	73
DSRD 180	NOT REPORTED	760.000	- 940.000	870.000	96.437	3
	RESIDUE ON EVAPORATION	92.000	-1074.000	968.462	46.053	26
	RESIDUE FILTRABLE	312.000	- 994.000	950.161	19.241	31
	_OVER-ALL_	92.000	-1074.000	948.517	27.370	60
F	COLORIMETRIC, LANTHANUM ALIZARIN "COMPLEXONE"	2.100	- 2.500	2.333	0.208	3
	COLORIMETRIC, SPADNS	2.300	- 2.500	2.467	0.153	3
	ION CHROMATOGRAPHY	1.600	- 2.900	2.344	0.397	9
	ION SELECTIVE ELECTRODE	2.100	- 3.000	2.280	0.118	35
	_OVER-ALL_	1.600	- 3.400	2.304	0.134	53
K	ATOMIC ABSORPTION, DIRECT, AIR	4.700	- 10.500	6.212	0.614	43
	EMISSION, FLAME, PHOTOMETRIC	2.100	- 9.000	6.183	2.347	6
	EMISSION, IC PLASMA	5.000	- 7.000	5.992	0.494	12
	NOT REPORTED	5.200	- 15.600	5.500	0.361	3
	OTHER	6.000	- 7.500	6.567	0.814	3
	_OVER-ALL_	2.100	- 15.600	6.172	0.630	65
MG	ATOMIC ABSORPTION, DIRECT, AIR	34.000	- 97.000	46.868	2.462	38
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	42.000	- 51.000	47.200	3.421	5
	EMISSION, IC PLASMA	30.000	- 51.000	46.750	2.236	16
	OTHER	46.000	- 55.000	49.667	4.726	3
	TITRATION, EDTA	44.000	- 61.000	49.000	8.124	4
_OVER-ALL_	30.000	- 97.000	46.742	2.394	66	
NA	ATOMIC ABSORPTION, DIRECT, AIR	105.000	- 163.000	138.558	6.076	43
	EMISSION, FLAME	133.000	- 201.000	154.714	24.047	7
	EMISSION, IC PLASMA	126.000	- 152.000	136.588	6.746	17
	OTHER	138.000	- 160.000	147.333	11.372	3
	_OVER-ALL_	105.000	- 201.000	138.406	6.146	69
NO2-N	COLORIMETRIC, DIAZOTIZATION	0.001	- 0.480	0.004	0.003	15
	ION CHROMATOGRAPHY	0.006	- 3.400	-----	-----	-
	_OVER-ALL_	0.001	- 3.400	0.006	0.006	18
NO3-N	COLORIMETRIC, BRUCINE	3.600	- 5.200	4.467	0.668	6
	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	0.600	- 5.400	4.288	0.379	50
	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3.900	- 5.200	4.550	0.569	4
	ION CHROMATOGRAPHY	4.300	- 22.800	4.480	0.130	5
	OTHER	4.200	- 5.000	4.583	0.299	6
_OVER-ALL_	0.600	- 22.800	4.359	0.402	73	
P, TOTAL	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	0.340	- 3.490	1.142	0.067	40
	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	0.040	- 7.000	0.962	0.446	12
	EMISSION, IC PLASMA	1.000	- 1.500	1.235	0.206	4
	OTHER	1.000	- 1.700	1.110	0.088	5
	_OVER-ALL_	0.040	- 7.000	1.151	0.111	63
PH	ELECTROMETRIC	8.000	- 8.800	8.571	0.097	75
	NOT REPORTED	7.700	- 8.710	8.570	0.099	4
	_OVER-ALL_	7.700	- 8.800	8.571	0.096	79
SIO2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	12.000	- 23.000	17.444	3.283	9
	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	18.000	- 25.000	18.625	0.744	8
	COLORIMETRIC, MOLYBDOSILICIC ACID	13.000	- 19.000	17.875	1.126	8
	COLORIMETRIC, SODIUM SULFITE REDUCTION TO MOLYBDATE BLUE	16.000	- 19.000	17.200	1.304	5
	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID REDUCE-HETEROPOLY BLUE	18.000	- 21.000	19.000	1.414	4
	EMISSION, IC PLASMA	13.000	- 24.000	18.333	2.741	12
	_OVER-ALL_	12.000	- 25.000	18.122	2.429	49
	---	---	---	---	---	---
SO4	COLORIMETRIC, METHYL THYMOL BLUE	20.000	- 455.000	376.905	29.895	21
	GRAVIMETRIC, BARIUM SULFATE	200.000	- 405.000	377.500	17.928	8
	ION CHROMATOGRAPHY	240.000	- 400.000	351.818	55.959	11
	NOT REPORTED	360.000	- 385.000	385.000	14.434	2
	THORIN TITRATION	360.000	- 390.000	376.667	15.275	3
	TURBIDIMETRIC, BARIUM SULFATE	35.000	- 410.000	377.500	22.311	18
	_OVER-ALL_	20.000	- 455.000	377.623	18.430	62
	---	---	---	---	---	---



Table 11. Statistics by method for standard reference sample M94

Determin- ation	Method	Range:		Mean	Standard Deviation	N
		from	to			
SP. COND.	DIRECT READING INSTRUMENT	720.000	-1440.000	1322.308	75.931	39
	NOT REPORTED	950.000	-1400.000	1400.000	0.000	2
	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1100.000	-1600.000	1342.727	59.491	33
	_OVER-ALL_	720.000	-1600.000	1330.789	73.353	76
SR	ATOMIC ABSORPTION, DIRECT, AIR	32.000	-1200.000	1022.364	141.681	11
	EMISSION, IC PLASMA	965.000	-1220.000	1083.700	68.307	10
	OTHER	1.000	-1300.000	800.333	699.429	3
	_OVER-ALL_	1.000	-1300.000	1075.720	91.662	25
V	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	200.000	- 200.000	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	1.400	- 73.000	-----	-----	-
	EMISSION, IC PLASMA	2.000	- 101.000	6.088	5.065	8
	_OVER-ALL_	1.000	- 200.000	3.856	3.087	9

Table 12 Standard Reference Water Sample T95 Report for ACID@CAC03

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	1170	0.9	'TITRATION, ELECTROMETRIC'	1, 2, 3, 4
19	1150	-0.8	'TITRATION, ELECTROMETRIC'	1, 2, 3, 4
30	1160	0.1	'TITRATION, ELECTROMETRIC'	1, 2, 3, 4
37	1200	3.5	'TITRATION, ELECTROMETRIC'	1, 2, 3, 4
38	1100	-5.1	'TITRATION, ELECTROMETRIC'	1, 2, 3, 4
41	1160	0.1	'TITRATION, ELECTROMETRIC'	1, 2, 3, 4
48	1120	-3.4	'TITRATION, ELECTROMETRIC'	1, 2, 3, 4
54	1160	0.1	'TITRATION, ELECTROMETRIC'	1, 2, 3, 4
80	1150	-0.8	'TITRATION, ELECTROMETRIC'	1, 2, 3, 4
83	1160	0.1	'TITRATION, ELECTROMETRIC'	1, 2, 3, 4
96	1220	5.2	'TITRATION, COLORIMETRIC'	1, 2, 3
102	1130	-2.5	OTHER	1, 2, 3
104	1150	-0.8	'TITRATION, COLORIMETRIC'	1, 2, 3
112	1200	3.5	'TITRATION, ELECTROMETRIC'	1, 2, 3, 4

14 Labs had a total range of 1100 to 1220 and a mean of 1159 with a standard deviation of 32 and a 95% confidence interval of the mean +/- 18.

Table 12 Standard Reference Water Sample T95 Report for AG

Code Number	Reported value	Pct. dev. from mean	Methods	References
2	0.8	-54.4	ATOMIC ABSORPTION, FLAMELESS	3
4	< 25.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
6	< 5.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
7	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
8	< 0.5		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
9	5.0	185.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
10	< 0.3		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
13	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
15	0.1	-94.3	ATOMIC ABSORPTION, FLAMELESS	3
16	2.7	54.1	ATOMIC ABSORPTION, FLAMELESS	3
20	< 0.1		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
22	< 20.0		IGNORED EMISSION, IC PLASMA	3
24	1.4	-20.1	ATOMIC ABSORPTION, FLAMELESS	3
25	2.0	14.1	ATOMIC ABSORPTION, FLAMELESS	3
27	< 2.5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
30	3.0	71.2	ATOMIC ABSORPTION, FLAMELESS	3
32	2.0	14.1	EMISSION, IC PLASMA	3
34	< 1.0		IGNORED OTHER	3
36	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
37	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
38	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
40	1.0	-42.9	ATOMIC ABSORPTION, FLAMELESS	3
41	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
43	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
46	0.2	-88.6	ATOMIC ABSORPTION, FLAMELESS	3
47	1.2	-31.5	ATOMIC ABSORPTION, FLAMELESS	3
48	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
49	1.0	-42.9	EMISSION, IC PLASMA	3
50	< 6.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
54	< 1.0		IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1, 2, 4
57	< 0.7	-60.1	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
58	< 1.0		IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1, 2, 4
59	3.2	82.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
63	2.0	14.1	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
70	10.0	470.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
73	1.0	-42.9	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
74	4.4	151.1	NOT REPORTED	3
76	< 1.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
80	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
81	< 1.0		IGNORED EMISSION, IC PLASMA	3
83	< 20.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
84	< 5.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
90	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
93	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
94	90.0	5035.2	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
96	1.5	-14.4	OTHER	3
98	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
103	< 2.0		IGNORED EMISSION, IC PLASMA	3
104	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3
110	560.0	3E+04	REJECT EMISSION, IC PLASMA	1, 2, 3
118	< 0.5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
122	0.1	-94.3	ATOMIC ABSORPTION, FLAMELESS	3
124	< 20.0		IGNORED NOT REPORTED	3
127	< 1.0		IGNORED OTHER	3
142	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3

56 Labs had a total range of 0.1 to 560.0 and a mean of 1.75 with a standard deviation of 1.38 and a 95% confidence interval of the mean +/- 0.67.

Table 12 Standard Reference Water Sample T95 Report for AL

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	30	-70.7	ATOMIC ABSORPTION, CHELATION EXTRACTION, NITROUS OXIDE	2,4
4	< 1000		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
6	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
7	< 80		IGNORED EMISSION, IC PLASMA	3,5
8	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
9	190	85.7	IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
13	180	75.9	EMISSION, IC PLASMA	3,5
15	30	-70.7	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
18	< 200		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
20	< 0		IGNORED ATOMIC ABSORPTION, CHELATION EXTRACTION, NITROUS OXIDE	2,4
22	< 100		IGNORED EMISSION, IC PLASMA	3,5
24	50	-51.1	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
25	260	154.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
30	110	7.5	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
33	< 100		IGNORED NOT REPORTED	
34	90	-12.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
36	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
37	30	-70.7	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
38	690	574.4	REJECT EMISSION, IC PLASMA	3,5
40	80	-21.8	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
41	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
46	50	-51.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
47	50	-51.1	ATOMIC ABSORPTION, DIRECT, FLAMELESS	1,2,3,4
48	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
49	30	-70.7	EMISSION, IC PLASMA	3,5
51	40	-60.9	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
53	< 500		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
54	20	-80.5	ATOMIC ABSORPTION, CHELATION EXTRACTION, AIR-ACETYLENE	1,7
57	30	-70.7	EMISSION, DC PLASMA	7
58	< 80		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
63	420	310.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
69	10	-90.2	EMISSION, IC PLASMA	3,5
73	410	300.8	EMISSION, IC PLASMA	3,5
76	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
83	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
84	60	-41.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
93	< 200		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
98	40	-60.9	EMISSION, IC PLASMA	3,5
103	< 10		IGNORED EMISSION, IC PLASMA	3,5
104	100	-2.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
110	200	95.5	EMISSION, IC PLASMA	3,5
112	100	-2.3	EMISSION, DC PLASMA	7
127	20	-80.5	OTHER	
142	30	-70.7	EMISSION, IC PLASMA	3,5

44 Labs had a total range of 10 to 690 and a mean of 102 with a standard deviation of 112 and a 95% confidence interval of the mean +/- 45.

Table 12 Standard Reference Water Sample T95 Report for AS

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1.0		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
2	< 0.6	-46.1		
4	< 10.0		IGNORED SPECTROPHOTOMETRIC, SILVER DIETHYL DITHIOCARBAMATE	2,3,4
6	< 2.0		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
7	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
8	< 5.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
9	< 8.0		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
10	< 3.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
13	< 4.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
15	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
16	< 0.5	-55.1	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
20	< 0.0		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
22	< 4.0		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
24	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
25	< 4.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
27	< 5.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
30	< 0.5	-55.1	IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
34	< 3.0		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
36	< 0.6	-46.1	IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
37	< 1.0		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
38	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
40	< 0.1		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
41	< 5.0	349.4	REJECT ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
43	< 2.0	79.8	REJECT ATOMIC ABSORPTION, FLAMELESS	3
46	< 0.2	-82.0	REJECT ATOMIC ABSORPTION, FLAMELESS	3
47	< 1.4	25.8	REJECT ATOMIC ABSORPTION, FLAMELESS	3
48	< 5.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
49	< 0.6	-46.1	IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
53	< 2.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
54	< 7.2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
56	< 7.2	547.2	REJECT ATOMIC ABSORPTION, FLAMELESS	3
57	< 1.0		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
58	< 3.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
63	< 1.4	25.8	REJECT ATOMIC ABSORPTION, FLAMELESS	3
69	< 22.8	1949.4	REJECT EMISSION, IC PLASMA	3
70	< 1.0	169.7	REJECT ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
73	< 14.0	1158.4	REJECT EMISSION, IC PLASMA	3
74	< 1.1	-1.1	REJECT ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
76	< 2.0		IGNORED ATOMIC ABSORPTION, HYDRIDE, (ZINC)	1,2,3,4
80	< 1.2	7.9	IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
81	< 5.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
83	< 5.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
84	< 10.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
90	< 0.1		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
93	< 0.5		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
96	< 16.0		IGNORED OTHER	
98	< 5.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
104	< 1.6	43.8	IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
107	< 2.0	79.8	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
108	< 1.0		IGNORED ATOMIC ABSORPTION, HYDRIDE, (NABH4)	1,4
110	< 90.0	7989.9	REJECT EMISSION, IC PLASMA	3
112	< 2.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
118	< 10.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
121	< 0.6	-46.1	IGNORED OTHER	
122	< 0.5	-55.1	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
127	< 4.0		IGNORED OTHER	
142	< 2.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3

57 Labs had a total range of 0.2 to 90.0 and a mean of 1.11 with a standard deviation of 0.76 and a 95% confidence interval of the mean +/- 0.40.

Table 12 Standard Reference Water Sample T95 Report for B

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	1210	10.9	EMISSION, IC PLASMA	3
2	630	-42.2	COLORIMETRIC, CURCUMIN	1,2,3,4
8	190	-82.6	COLORIMETRIC, CURCUMIN	1,2,3,4
15	2000	83.4	COLORIMETRIC, CURCUMIN	1,2,3,4
22	1050	-3.7	EMISSION, IC PLASMA	3
33	1190	9.1	EMISSION, DC PLASMA	3
38	< 100		IGNORED NOT REPORTED	7
40	1130	3.6	EMISSION, IC PLASMA	3
41	1100	0.9	EMISSION, IC PLASMA	3
47	380	-65.2	COLORIMETRIC, CURCUMIN	1,2,3,4
48	1000	-8.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	7
49	1100	0.9	EMISSION, IC PLASMA	3
49	1120	2.7	EMISSION, IC PLASMA	3
51	1170	7.3	EMISSION, IC PLASMA	3
52	1160	6.4	EMISSION, IC PLASMA	3
54	1300	19.2	COLORIMETRIC, AZOMETHINE	5
63	1600	46.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	7
73	1100	0.9	EMISSION, IC PLASMA	3
80	1100	0.9	EMISSION, IC PLASMA	3
83	< 2000		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	7
85	1300	19.2	COLORIMETRIC, DIANTHRIMIDE	4
93	230	-78.9	COLORIMETRIC, CURCUMIN	1,2,3,4
98	1250	14.6	EMISSION, IC PLASMA	3
103	1060	-2.8	EMISSION, IC PLASMA	3
108	1000	-8.3	COLORIMETRIC, AZOMETHINE	5
110	1180	8.2	EMISSION, IC PLASMA	3
111	1130	3.6	EMISSION, IC PLASMA	3
112	1150	5.4	EMISSION, DC PLASMA	7
124	1600	46.7	NOT REPORTED	
127	1100	0.9	OTHER	
142	1100	0.9	EMISSION, IC PLASMA	3

31 Labs had a total range of 190 to 2000 and a mean of 1091 with a standard deviation of 368 and a 95% confidence interval of the mean +/- 140.

Table 12 Standard Reference Water Sample T95 Report for BA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	50	3.3	EMISSION, IC PLASMA	3,5
4	< 500		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
6	70	44.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
8	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
9	30	-38.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
13	40	-17.3	EMISSION, IC PLASMA	3,5
15	50	3.3	ATOMIC ABSORPTION, FLAMELESS	3
16	40	-17.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
18	< 200		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
20	< 0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
22	40	-17.3	EMISSION, IC PLASMA	3,5
24	90	86.0	ATOMIC ABSORPTION, FLAMELESS	3
25	< 30		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
27	190	292.7	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
30	80	65.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
34	70	44.7	ATOMIC ABSORPTION, FLAMELESS	3
37	470	871.3	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
38	50	3.3	EMISSION, IC PLASMA	3,5
40	40	-17.3	EMISSION, IC PLASMA	3,5
41	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
43	350	623.3	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
46	200	313.3	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
47	30	-38.0	ATOMIC ABSORPTION, FLAMELESS	3
48	< 50		IGNORED EMISSION, IC PLASMA	3,5
51	40	-17.3	EMISSION, IC PLASMA	3,5
52	40	-17.3	EMISSION, IC PLASMA	3,5
53	< 200		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
54	40	-17.3	EMISSION, IC PLASMA	3,5
57	40	-17.3	EMISSION, IC PLASMA	3,5
58	50	3.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
63	70	44.7	ATOMIC ABSORPTION, FLAMELESS	3
69	40	-17.3	EMISSION, IC PLASMA	3,5
73	40	-17.3	EMISSION, IC PLASMA	3,5
76	< 0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
80	40	-17.3	EMISSION, IC PLASMA	3,5
83	< 500		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
84	60	24.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
93	< 300		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
94	240	396.0	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
96	60	24.0	EMISSION, DC PLASMA	2,7
98	40	-17.3	EMISSION, IC PLASMA	3,5
103	40	-17.3	EMISSION, IC PLASMA	3,5
104	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
110	50	3.3	EMISSION, IC PLASMA	3,5
112	50	3.3	EMISSION, DC PLASMA	2,7
118	< 400		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
124	< 10		IGNORED NOT REPORTED	
127	40	-17.3	OTHER	
142	40	-17.3	EMISSION, IC PLASMA	3,5
144	40	-17.3	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

50 Labs had a total range of 30 to 470 and a mean of 48 with a standard deviation of 14 and a 95% confidence interval of the mean +/- 5.

Table 12 Standard Reference Water Sample T95 Report for BE

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 5		IGNORED EMISSION, IC PLASMA	3,5
6	< 5		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
7	< 20		IGNORED EMISSION, IC PLASMA	3,5
8	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
9	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
13	1	0.0	EMISSION, IC PLASMA	3,5
20	< 0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
22	< 10		IGNORED EMISSION, IC PLASMA	3,5
24	1	0.0	ATOMIC ABSORPTION, FLAMELESS	3
34	< 1		IGNORED EMISSION, IC PLASMA	3,5
37	< 1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
38	< 1		IGNORED EMISSION, IC PLASMA	3,5
41	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
47	< 1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
48	< 20		IGNORED EMISSION, IC PLASMA	3,5
50	< 3		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
54	< 1		IGNORED EMISSION, IC PLASMA	3,5
57	< 1		IGNORED EMISSION, IC PLASMA	3,5
58	< 5		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
63	< 2		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
69	5	400.0	REJECT EMISSION, IC PLASMA	3,5
73	< 1		IGNORED EMISSION, IC PLASMA	3,5
76	< 1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
80	< 1		IGNORED EMISSION, IC PLASMA	3,5
83	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
93	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
98	< 2		IGNORED EMISSION, IC PLASMA	3,5
103	< 0		IGNORED EMISSION, IC PLASMA	3,5
104	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
110	1	0.0	EMISSION, IC PLASMA	3,5
112	< 1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
124	< 10		IGNORED NOT REPORTED	
127	< 5		OTHER	
142	< 1		IGNORED EMISSION, IC PLASMA	3,5

34 Labs had a total range of 1 to 5. INSUFFICIENT DATA TO DETERMINE MEAN AND STANDARD OF DEVIATION.

Table 12 Standard Reference Water Sample T95 Report for CA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	71	-1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	37	-48.7	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	78	8.2	OTHER	1,2,3,4
8	72	-0.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
13	72	-0.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	72	-0.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	75	4.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
22	63	-12.6	EMISSION, IC PLASMA	3,5,7
24	73	1.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	75	4.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
27	73	1.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	72	-0.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	69	-4.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
32	73	1.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
33	63	-12.6	NOT REPORTED	1,7
36	76	5.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	59	-18.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	75	4.1	EMISSION, IC PLASMA	3,5,7
40	70	-2.9	EMISSION, IC PLASMA	3,5,7
41	94	30.4	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
43	78	8.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
47	76	5.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	73	1.3	EMISSION, IC PLASMA	3,5,7
49	75	4.1	EMISSION, IC PLASMA	3,5,7
50	22	-69.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	75	4.1	EMISSION, IC PLASMA	3,5,7
52	70	-2.9	EMISSION, IC PLASMA	3,5,7
53	73	1.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	82	13.8	EMISSION, IC PLASMA	3,5,7
56	126	74.8	REJECT TITRATION, EDTA	1,3
57	69	-4.3	EMISSION, IC PLASMA	3,5,7
58	72	-0.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	80	11.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
61	78	8.2	EMISSION, IC PLASMA	3,5,7
70	71	-1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	69	-4.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	69	-4.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	70	-2.9	EMISSION, IC PLASMA	3,5,7
81	75	4.1	EMISSION, IC PLASMA	3,5,7
83	71	-1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	60	-16.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
86	70	-2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	75	4.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	75	4.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
96	87	20.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
98	65	-9.8	OTHER	3,5,7
102	65	-9.8	EMISSION, IC PLASMA	1,2,3,4
103	72	-0.1	ATOMIC ABSORPTION, DIRECT, AIR	3,5,7
104	72	-0.1	EMISSION, IC PLASMA	1,7
108	72	-0.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
110	77	6.8	ATOMIC ABSORPTION, DIRECT, AIR	1,7
111	73	1.3	EMISSION, IC PLASMA	3,5,7
112	7	-90.3	REJECT EMISSION, IC PLASMA	3,5,7
113	69	-4.3	OTHER	1,2,3,4
121	68	-5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
124	61	-15.4	OTHER	
127	77	6.8	NOT REPORTED	
142	75	4.1	OTHER	
			EMISSION, IC PLASMA	3,5,7

58 Labs had a total range of 7 to 126 and a mean of 72.1 with a standard deviation of 5.3 and a 95% confidence interval of the mean +/- 1.4.

Table 12 Standard Reference Water Sample T95 Report for CD

Code Number	Reported value	Pct. dev. from mean	Methods	Reference
1	< 1.0		IGNORED ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1, 4
2	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
4	0.4	-50.6	ATOMIC ABSORPTION, FLAMELESS	3
6	< 5.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
7	0.4	-50.6	ATOMIC ABSORPTION, FLAMELESS	3
8	< 5.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
9	< 0.4		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
10	0.4	-50.6	ATOMIC ABSORPTION, FLAMELESS	3
13	0.4	-50.6	ATOMIC ABSORPTION, FLAMELESS	3
15	0.3	-62.9	ATOMIC ABSORPTION, FLAMELESS	3
16	0.7	-13.5	ATOMIC ABSORPTION, FLAMELESS	3
18	1.0	23.5	ATOMIC ABSORPTION, FLAMELESS	3
20	< 0.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
22	< 10.0		IGNORED EMISSION, IC PLASMA	3, 5
24	0.5	-38.2	ATOMIC ABSORPTION, FLAMELESS	3
25	< 4.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
27	< 0.5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
29	< 3.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
30	0.3	-62.9	ATOMIC ABSORPTION, FLAMELESS	3
32	33.0	3975.6	REJECT ATOMIC ABSORPTION, FLAMELESS	3
33	< 10.0		IGNORED NOT REPORTED	
34	< 2.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
36	0.4	-50.6	ATOMIC ABSORPTION, FLAMELESS	3
37	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
38	0.3	-62.9	ATOMIC ABSORPTION, FLAMELESS	3
40	1.6	97.6	ATOMIC ABSORPTION, FLAMELESS	3
41	< 0.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
43	0.2	-75.3	ATOMIC ABSORPTION, FLAMELESS	3
46	0.1	-87.6	ATOMIC ABSORPTION, FLAMELESS	3
47	0.3	-62.9	ATOMIC ABSORPTION, FLAMELESS	3
48	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
49	4.0	394.0	REJECT EMISSION, IC PLASMA	3, 5
50	< 3.0		IGNORED ANODIC STRIPPING VOLTAMMETRY, DIFFERENTIAL PULSE	2
51	0.9	11.2	ATOMIC ABSORPTION, FLAMELESS	3
53	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
54	< 1.0		IGNORED ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1, 4
56	0.5	-38.2	ATOMIC ABSORPTION, FLAMELESS	3
57	< 1.0		IGNORED EMISSION, IC PLASMA	3, 5
58	< 1.0		IGNORED ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1, 4
59	3.0	270.5	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
63	3.0	270.5	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
69	3.0	270.5	EMISSION, IC PLASMA	3, 5
73	2.0	147.0	EMISSION, IC PLASMA	3, 5
74	0.4	-50.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
76	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
79	0.1	-87.6	ATOMIC ABSORPTION, FLAMELESS	3
80	< 1.0		IGNORED EMISSION, IC PLASMA	3, 5
81	< 2.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
83	< 5.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
84	< 5.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
85	< 1.0		IGNORED ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1, 4
90	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
93	< 5.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
94	6.0	641.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
95	0.2	-75.3	ATOMIC ABSORPTION, FLAMELESS	3
96	2.0	147.0	EMISSION, DC PLASMA	7
98	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
102	5.0	517.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
103	< 3.0		IGNORED EMISSION, IC PLASMA	3, 5
104	1.0	23.5	ATOMIC ABSORPTION, EXTRACTION, (PDCA/CHCL3)	2, 3
107	0.2	-75.3	ATOMIC ABSORPTION, FLAMELESS	3
110	10.0	1135.0	REJECT EMISSION, IC PLASMA	3, 5
112	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
118	0.3	-62.9	ATOMIC ABSORPTION, FLAMELESS	3
121	0.4	-50.6	OTHER	
122	0.5	-38.2	ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1, 4
124	< 10.0		IGNORED NOT REPORTED	
127	< 2.0		IGNORED OTHER	
142	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
144	0.3	-62.9	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

70 Labs had a total range of 0.1 to 33.0 and a mean of 0.81 with a standard deviation of 0.88 and a 95% confidence interval of the mean +/- 0.32.

Table 12 Standard Reference Water Sample T95 Report for CO

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 3		IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
6	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	< 33	126.7	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	< 25		IGNORED EMISSION, IC PLASMA	3,5
15	< 0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
22	< 10		IGNORED EMISSION, IC PLASMA	3,5
24	< 15	3.1	ATOMIC ABSORPTION, FLAMELESS	3
25	< 19	30.5	ATOMIC ABSORPTION, FLAMELESS	3
30	< 23	58.0	ATOMIC ABSORPTION, FLAMELESS	3
34	< 10		IGNORED EMISSION, IC PLASMA	3,5
41	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
47	< 1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
48	< 20		IGNORED EMISSION, IC PLASMA	3,5
54	< 5		IGNORED EMISSION, IC PLASMA	3,5
57	< 3		IGNORED EMISSION, IC PLASMA	3,5
58	< 58	298.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	< 13	-10.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
63	< 10	-31.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	< 3	-79.4	EMISSION, IC PLASMA	3,5
73	< 1		IGNORED EMISSION, IC PLASMA	3,5
76	< 5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
80	< 5		IGNORED EMISSION, IC PLASMA	3,5
81	< 60		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
83	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
85	< 3	-79.4	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
93	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
98	< 2		IGNORED EMISSION, IC PLASMA	3,5
103	< 3		IGNORED EMISSION, IC PLASMA	3,5
104	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	< 12	-17.6	EMISSION, IC PLASMA	3,5
124	< 10		IGNORED NOT REPORTED	
127	< 1		IGNORED OTHER	
142	< 3		IGNORED EMISSION, IC PLASMA	3,5

34 Labs had a total range of 3 to 58 and a mean of 14.6 with a standard deviation of 9.5 and a 95% confidence interval of the mean +/- 7.3.



Table 12 Standard Reference Water Sample T95 for Report CR TOT

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1		IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1, 3, 4
2	< 1		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
4	< 5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
6	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
7	< 3		IGNORED EMISSION, IC PLASMA	3
8	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
9	5	14.0	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
10	2	-54.4	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
13	< 50		IGNORED EMISSION, IC PLASMA	3
15	1	-77.2	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
16	5	14.0	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
18	1	-77.2	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
20	< 0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
22	63	1336.0	REJECT EMISSION, IC PLASMA	3
24	2	-54.4	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
25	< 10		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
27	< 5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
29	< 40		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
30	4	-8.8	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
32	30	583.8	REJECT EMISSION, IC PLASMA	3
33	< 20		IGNORED NOT REPORTED	3
34	< 5		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
37	5	14.0	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
38	1	-77.2	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
40	1	-77.2	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
41	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
43	2	-54.4	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
46	< 1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
47	2	-54.4	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
48	6	36.8	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
49	6	36.8	EMISSION, IC PLASMA	3
53	< 30		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
54	1	-77.2	IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1, 3, 4
56	8	82.4	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
57	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
58	7	59.6	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
59	10	127.9	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
63	5	14.0	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
69	8	82.4	EMISSION, IC PLASMA	3
73	10	127.9	EMISSION, IC PLASMA	3
76	< 5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
79	12	173.5	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
80	< 3		IGNORED EMISSION, IC PLASMA	3
81	< 5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
83	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
84	< 5		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
85	< 1		IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1, 3, 4
93	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1
94	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
95	1	-77.2	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
96	6	36.8	OTHER	3
98	< 3		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
103	< 6		IGNORED EMISSION, IC PLASMA	3
104	5	14.0	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
107	2	-54.4	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
108	< 2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
110	32	629.4	REJECT EMISSION, IC PLASMA	3
112	1	-77.2	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
113	8	82.4	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
118	< 3		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
121	2	-54.4	OTHER	3
122	3	-31.6	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
124	< 20		IGNORED NOT REPORTED	3
127	4	-8.8	OTHER	3
142	< 5		IGNORED EMISSION, IC PLASMA	3

65 Labs had a total range of 1 to 63 and a mean of 4.4 with a standard deviation of 3.1 and a 95% confidence interval of the mean +/- 1.1.

Table 12 Standard Reference Water Sample T95 Report for CU

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	7	-38.5	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1, 4
2	5	-56.0	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
4	15	31.9	ATOMIC ABSORPTION, FLAMELESS	3
6	11	-3.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
7	< 20		IGNORED EMISSION, IC PLASMA	3, 5
8	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
9	10	-12.1	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
13	14	23.1	EMISSION, IC PLASMA	3, 5
15	8	-29.7	ATOMIC ABSORPTION, FLAMELESS	3
16	11	-3.3	ATOMIC ABSORPTION, FLAMELESS	3
18	22	93.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
20	< 0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
22	12	5.5	EMISSION, IC PLASMA	3, 5
24	18	58.2	ATOMIC ABSORPTION, FLAMELESS	3
25	5	-56.0	ATOMIC ABSORPTION, FLAMELESS	3
29	8	-29.7	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
30	26	128.6	ATOMIC ABSORPTION, FLAMELESS	3
33	18	58.2	EMISSION, IC PLASMA	3, 5
34	10	-12.1	NOT REPORTED	
36	18	58.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
37	8	5.5	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
38	< 10		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
40	4	-64.8	IGNORED EMISSION, IC PLASMA	3, 5
41	< 10		IGNORED EMISSION, IC PLASMA	3, 5
43	9	-20.9	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
47	12	5.5	ATOMIC ABSORPTION, FLAMELESS	3
48	< 10		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
49	13	14.3	EMISSION, IC PLASMA	1, 2, 3, 4
51	12	5.5	ATOMIC ABSORPTION, FLAMELESS	3, 5
52	8	-29.7	EMISSION, IC PLASMA	3, 5
53	8	-29.7	ATOMIC ABSORPTION, FLAMELESS	3, 5
54	16	40.7	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1, 4
56	13	14.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
57	< 10		IGNORED EMISSION, IC PLASMA	3, 5
58	< 5		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
59	13	14.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
63	14	23.1	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
69	18	58.2	EMISSION, IC PLASMA	3, 5
70	20	75.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
73	8	-29.7	EMISSION, IC PLASMA	3, 5
74	11	-3.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
76	20	75.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
79	10	-12.1	ATOMIC ABSORPTION, FLAMELESS	3
80	7	-38.5	EMISSION, IC PLASMA	3, 5
81	8	-29.7	ATOMIC ABSORPTION, FLAMELESS	3
83	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
84	10	-12.1	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
85	8	-29.7	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1, 4
90	11	-3.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
93	4	-64.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
94	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
95	8	-29.7	ATOMIC ABSORPTION, FLAMELESS	3
98	7	-38.5	EMISSION, IC PLASMA	3, 5
102	6	-47.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
103	< 3		IGNORED EMISSION, IC PLASMA	3, 5
104	5	-56.0	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2, 3
107	13	14.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
108	8	-29.7	ATOMIC ABSORPTION, FLAMELESS	3
110	50	339.6	REJECT EMISSION, IC PLASMA	3, 5
112	10	-12.1	OTHER	
113	16	40.7	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
118	8	-29.7	ATOMIC ABSORPTION, FLAMELESS	3
121	16	40.7	OTHER	
122	9	-20.9	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2, 3
124	< 10		IGNORED NOT REPORTED	
127	10	-12.1	OTHER	
142	13	14.3	EMISSION, IC PLASMA	3, 5
144	13	14.3	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

69 Labs had a total range of 4 to 50 and a mean of 11.4 with a standard deviation of 4.7 and a 95% confidence interval of the mean +/- 1.3.

Table 12 Standard Reference Water Sample T95 Report for FE

Code Number	Reported value	Pct. dev. from mean	Methods	Reference
1	< 20		IGNORED EMISSION, IC PLASMA	3,5
2	< 47	281.1	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	< 10		IGNORED EMISSION, IC PLASMA	3,5
8	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	< 10		IGNORED EMISSION, IC PLASMA	3,5
15	< 41	232.4	REJECT ATOMIC ABSORPTION, FLAMELESS	3
16	< 20	62.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
18	< 21	70.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	< 0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	< 30		IGNORED EMISSION, IC PLASMA	3,5
24	< 5	-59.5	ATOMIC ABSORPTION, FLAMELESS	3
25	< 6		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
27	< 30		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	< 60		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	< 10	-18.9	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
32	< 60	386.5	REJECT EMISSION, IC PLASMA	3,5
33	< 10		IGNORED NOT REPORTED	3
34	< 10	-18.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
36	< 22	78.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	< 10	-18.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	< 10		IGNORED EMISSION, IC PLASMA	3,5
40	< 7	-43.2	EMISSION, IC PLASMA	3,5
41	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	< 30		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
46	< 15	21.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
47	< 4	-67.6	ATOMIC ABSORPTION, FLAMELESS	3
48	< 20		IGNORED EMISSION, IC PLASMA	3,5
49	< 1	-91.9	EMISSION, IC PLASMA	3,5
50	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	< 9	-27.0	ATOMIC ABSORPTION, FLAMELESS	3
53	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	< 10	-18.9	EMISSION, IC PLASMA	3,5
56	< 23	86.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	< 7	-43.2	EMISSION, IC PLASMA	3,5
58	< 610	4846.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	< 10	-18.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
63	< 10	-18.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	< 9	-27.0	EMISSION, IC PLASMA	3,5
73	< 12	-2.7	EMISSION, IC PLASMA	3,5
76	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
77	< 10		IGNORED OTHER	3
80	< 50		IGNORED EMISSION, IC PLASMA	3,5
81	< 15	21.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	< 10	-18.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
85	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
90	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	< 21	70.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	< 30		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	< 14	13.5	OTHER	3
98	< 12	-2.7	ATOMIC ABSORPTION, FLAMELESS	3,5
103	< 2		IGNORED EMISSION, IC PLASMA	3,5
104	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
108	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	< 21	70.3	EMISSION, IC PLASMA	3,5
112	< 20	62.2	OTHER	3
113	< 107	767.6	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	< 5	-59.5	OTHER	3
122	< 63	410.8	REJECT ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
123	< 12	-2.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
124	< 1		IGNORED NOT REPORTED	3
127	< 1	-91.9	OTHER	3
142	< 24	94.6	EMISSION, IC PLASMA	3,5

68 Labs had a total range of 1 to 610 and a mean of 12.3 with a standard deviation of 6.6 and a 95% confidence interval of the mean +/- 2.5.

Table 12 Standard Reference Water Sample T95 Report for HG

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.60	30.7	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
2	0.39	-15.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
6	0.40	-12.8	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
7	0.40	-12.8	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
8	< 0.50		IGNORED	1,2,3,4
9	0.30	-34.6	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
10	4.00	771.6	REJECT	1,2,3,4
13	0.50	9.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
16	0.40	-12.8	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
20	< 0.00		IGNORED	1,2,3,4
22	0.60	30.7	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
24	0.50	9.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
27	0.44	-4.1	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
30	0.42	-8.5	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
33	0.46	0.2	NOT REPORTED	1,2,3,4
34	0.40	-12.8	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
37	0.70	52.5	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
38	0.50	9.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
40	0.50	9.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
41	0.50	9.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
43	0.30	-34.6	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
46	0.20	-56.4	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
48	0.43	-6.3	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
49	6.08	1224.9	REJECT	1,2,3,4
50	< 0.50		IGNORED	1,2,3,4
54	0.43	-6.3	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
56	1800.	4E+05	REJECT	1,2,3,4
57	0.80	74.3	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
58	0.30	-34.6	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
63	2.80	510.2	REJECT	1,2,3,4
69	2.30	401.2	REJECT	1,2,3,4
74	16.00	3386.6	REJECT	1,2,3,4
76	0.50	9.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
80	0.49	6.8	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
81	< 0.50		IGNORED	1,2,3,4
83	0.45	-1.9	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
84	0.45	-1.9	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
90	0.58	26.4	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
95	0.40	-12.8	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
96	< 4.00		IGNORED	1,2,3,4
98	0.45	-1.9	OTHER	1,2,3,4
104	1.50	226.9	REJECT	1,2,3,4
112	0.50	9.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
118	0.40	-12.8	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
121	0.51	11.1	OTHER	1,2,3,4
122	0.56	22.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
127	< 2.00		IGNORED	1,2,3,4
142	0.30	-34.6	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4

48 Labs had a total range of 0.20 to 1800, and a mean of 0.459 with a standard deviation of 0.115 and a 95% confidence interval of the mean +/- 0.040.

Table 12 Standard Reference Water Sample T95 Report for K

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	4.50	-5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	4.43	-7.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	4.87	2.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	4.50	-5.7	EMISSION, IC PLASMA	3
8	4.60	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	6.83	43.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	5.40	13.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	3.80	-20.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	5.00	4.8	EMISSION, IC PLASMA	3
24	4.70	-1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	4.69	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
27	4.91	2.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	9.25	93.9	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	6.66	39.6	EMISSION, FLAME, PHOTOMETRIC	1,2
32	4.51	-5.5	EMISSION, IC PLASMA	3
33	4.27	-10.5	NOT REPORTED	
36	5.00	4.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	4.86	1.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	5.85	22.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
40	4.30	-9.9	EMISSION, IC PLASMA	3
41	5.10	6.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	4.70	-1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
47	4.70	-1.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	5.50	15.3	EMISSION, FLAME, PHOTOMETRIC	1,2
49	4.50	-5.7	EMISSION, IC PLASMA	3
50	5.15	7.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
52	3.90	-18.3	EMISSION, IC PLASMA	3
53	4.60	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	4.30	-9.9	EMISSION, IC PLASMA	3
56	10.90	128.5	REJECT EMISSION, FLAME, PHOTOMETRIC	1,2
57	4.50	-5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
58	4.60	-3.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	4.61	-3.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
61	5.30	11.1	EMISSION, IC PLASMA	3
63	17.00	256.3	REJECT EMISSION, FLAME, PHOTOMETRIC	1,2
72	4.40	-7.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	4.27	-10.5	EMISSION, IC PLASMA	3
80	5.70	19.5	OTHER	
81	8.50	78.2	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	5.40	13.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	5.40	13.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	5.75	20.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	5.03	5.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	3.56	-25.4	ION SELECTIVE ELECTRODE	1,2,3,4
98	4.50	-5.7	EMISSION, IC PLASMA	3
102	7.40	55.1	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
103	4.61	-3.4	EMISSION, IC PLASMA	3
104	4.50	-5.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
108	4.40	-7.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	5.00	4.8	EMISSION, IC PLASMA	3
111	3.03	-36.5	NOT REPORTED	
112	4.70	-1.5	OTHER	
113	4.32	-9.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	4.44	-6.9	OTHER	
142	4.40	-7.8	EMISSION, IC PLASMA	3

55 Labs had a total range of 3.03 to 17.00 and a mean of 4.771 with a standard deviation of 0.670 and a 95% confidence interval of the mean +/- 0.190.

Table 12 Standard Reference Water Sample T95 Report for LI

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	30	3.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
6	25	-13.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
8	40	38.4	EMISSION, FLAME	1,2,4
13	40	38.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
15	23	-20.4	OTHER	1,2,4
22	< 1000		IGNORED	3,5
36	25	-13.5	EMISSION, IC PLASMA	1,2,4
41	30	3.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
47	27	-6.6	EMISSION, FLAME	1,2,4
49	32	10.7	EMISSION, IC PLASMA	3,5
50	17	-41.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
51	26	-10.0	EMISSION, IC PLASMA	3,5
52	34	17.7	EMISSION, IC PLASMA	3,5
54	30	3.8	EMISSION, IC PLASMA	3,5
57	33	14.2	EMISSION, IC PLASMA	3,5
59	27	-6.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
63	31	7.3	EMISSION, FLAME	1
73	29	0.4	EMISSION, IC PLASMA	3,5
83	< 100		IGNORED	1,2,4
103	20	-30.8	ATOMIC ABSORPTION, DIRECT, AIR	3,5
110	30	3.8	EMISSION, IC PLASMA	3,5
127	< 10		IGNORED	3,5
142	< 50		IGNORED	3,5

23 Labs had a total range of 17 to 40 and a mean of 28.9 with a standard deviation of 5.8 and a 95% confidence interval of the mean +/- 2.8.

Table 12 Standard Reference Water Sample T95 Report for MG

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	33	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	32	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	33	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	33	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	3	-90.8	REJECT	1,7
13	32	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	30	-7.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	10	-69.3	REJECT	1,2,3,4
22	35	7.6	EMISSION, IC PLASMA	3,5
24	30	-7.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	32	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
27	32	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	69	112.1	REJECT	1,2,3,4
30	33	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
32	29	-10.9	EMISSION, IC PLASMA	3,5
33	32	-1.7	NOT REPORTED	
36	32	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	3	-90.8	REJECT	1,2,3,4
38	35	7.6	EMISSION, IC PLASMA	3,5
40	31	-4.7	EMISSION, IC PLASMA	3,5
41	32	-1.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
43	33	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
47	39	19.9	REJECT	1,2,3,4
48	31	-4.7	EMISSION, IC PLASMA	3,5
49	34	4.5	EMISSION, IC PLASMA	3,5
50	34	4.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	35	7.6	EMISSION, IC PLASMA	3,5
52	32	-1.7	EMISSION, IC PLASMA	3,5
53	37	13.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	33	1.4	EMISSION, IC PLASMA	3,5
56	43	32.2	REJECT	2
57	33	-1.7	TITRATION, EDTA	3,5
58	33	-1.7	EMISSION, IC PLASMA	3,5
59	40	22.9	REJECT	1,2,3,4
61	36	10.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
63	46	41.4	EMISSION, IC PLASMA	3,5
69	31	-4.7	REJECT	1,2,3,4
70	33	1.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
72	32	-1.7	EMISSION, IC PLASMA	3,5
73	32	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	32	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
81	31	-4.7	EMISSION, IC PLASMA	3,5
83	35	7.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	34	4.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
86	33	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	31	-4.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	51	56.7	REJECT	1,2,3,4
98	30	-7.8	OTHER	
102	30	-7.8	EMISSION, IC PLASMA	3,5
103	33	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	33	1.4	EMISSION, IC PLASMA	3,5
108	32	-1.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
110	33	1.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
111	33	1.4	EMISSION, IC PLASMA	3,5
112	32	-1.7	EMISSION, IC PLASMA	3,5
113	32	-1.7	OTHER	
118	32	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	32	-1.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
124	29	-10.9	OTHER	
127	37	13.7	NOT REPORTED	
142	34	4.5	OTHER	
			EMISSION, IC PLASMA	3,5

61 Labs had a total range of 3 to 69 and a mean of 32.5 with a standard deviation of 1.7 and a 95% confidence interval of the mean +/- 0.5.

Table 12 Standard Reference Water Sample T95 Report for MN

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 10		IGNORED EMISSION, IC PLASMA	3, 5
2	< 4	-29.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
4	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
6	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
7	< 20		IGNORED EMISSION, IC PLASMA	3, 5
8	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
9	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
13	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
16	< 2	-64.8	IGNORED EMISSION, IC PLASMA	3, 5
18	< 10	76.1	ATOMIC ABSORPTION, FLAMELESS	3
20	< 0		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
22	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
24	< 4	-29.6	IGNORED EMISSION, IC PLASMA	3, 5
25	< 3		ATOMIC ABSORPTION, FLAMELESS	3
27	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	3
29	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
30	< 3	-47.2	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
32	< 5	-12.0	ATOMIC ABSORPTION, FLAMELESS	3
33	< 10		EMISSION, IC PLASMA	3, 5
34	< 5		IGNORED NOT REPORTED	
36	< 5	-12.0	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
37	< 10		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
38	< 5		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
40	< 1	-82.4	IGNORED EMISSION, IC PLASMA	3, 5
41	< 10		EMISSION, IC PLASMA	3, 5
43	< 3	-47.2	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
46	< 4	-29.6	ATOMIC ABSORPTION, FLAMELESS	3
47	< 2	-64.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
48	< 20		ATOMIC ABSORPTION, FLAMELESS	3
49	< 1	-82.4	IGNORED EMISSION, IC PLASMA	3, 5
50	< 10		EMISSION, IC PLASMA	3, 5
51	< 3	-47.2	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
53	< 10		ATOMIC ABSORPTION, FLAMELESS	3
54	< 4	-29.6	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
56	< 1		EMISSION, IC PLASMA	3, 5
57	< 2	-64.8	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
58	< 5		EMISSION, IC PLASMA	3, 5
59	< 3	-47.2	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
63	< 4	-29.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
69	18	217.0	EMISSION, IC PLASMA	3, 5
73	< 3	-47.2	EMISSION, IC PLASMA	3, 5
77	< 10	76.1	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
79	< 3	-47.2	ATOMIC ABSORPTION, FLAMELESS	3
80	< 5		IGNORED EMISSION, IC PLASMA	3, 5
81	< 20		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
83	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
84	< 5		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
85	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
86	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
90	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
93	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
94	< 20	252.2	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
98	< 1		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
103	< 2		IGNORED EMISSION, IC PLASMA	3, 5
104	< 10		IGNORED EMISSION, IC PLASMA	3, 5
108	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
110	< 5	-12.0	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
112	< 10	76.1	EMISSION, IC PLASMA	3, 5
113	< 9	58.5	OTHER	
118	< 40		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
121	< 2	-64.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
122	< 15	164.1	OTHER	
123	< 4	-29.6	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2, 3
124	< 10		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
127	< 1		IGNORED NOT REPORTED	
142	< 5		IGNORED OTHER	
			IGNORED EMISSION, IC PLASMA	3, 5

66 Labs had a total range of 1 to 20 and a mean of 5.7 with a standard deviation of 5.0 and a 95% confidence interval of the mean +/- 1.9.

Table 12 Standard Reference Water Sample T95 Report for MO

Code Number	Reported value	Pct. dev. from mean	Methods	References
4	< 10		IGNORED ATOMIC ABS, EXTRACTION, 8 HYDROXYQUINOLINE/MIBK, NITROUS OXIDE	4
6	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
7	< 30		IGNORED EMISSION, IC PLASMA	3, 5
8	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
13	< 12	-5.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
15	< 8	-36.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
22	< 20		ATOMIC ABSORPTION, FLAMELESS	3
24	< 17	34.5	IGNORED EMISSION, IC PLASMA	3, 5
25	< 31	145.3	ATOMIC ABSORPTION, FLAMELESS	3
34	< 50		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
38	< 20		IGNORED EMISSION, IC PLASMA	3, 5
41	< 100		EMISSION, IC PLASMA	3, 5
46	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
47	< 8	-36.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
54	< 7	-44.6	ATOMIC ABSORPTION, FLAMELESS	3
57	< 10		IGNORED ATOMIC ABS, EXTRACTION, 8 HYDROXYQUINOLINE/MIBK, NITROUS OXIDE	4
63	< 30		EMISSION, IC PLASMA	3, 5
73	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
76	< 50	-20.9	EMISSION, IC PLASMA	3, 5
83	< 100		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
93	< 200		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
98	< 5		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
103	< 6		EMISSION, IC PLASMA	3, 5
104	< 10		IGNORED EMISSION, IC PLASMA	3, 5
108	< 4	-68.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1, 2, 3
110	< 25	97.8	ATOMIC ABSORPTION, FLAMELESS	3
127	< 9	-28.8	EMISSION, IC PLASMA	3, 5
142	< 8	-36.7	OTHER	
			EMISSION, IC PLASMA	3, 5

28 Labs had a total range of 4 to 31 and a mean of 12.6 with a standard deviation of 8.4 and a 95% confidence interval of the mean +/- 5.6.

Table 12 Standard Reference Water Sample T95 Report for NA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	191	1.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	192	1.5	EMISSION, FLAME	1,2
6	189	-0.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	203	7.4	EMISSION, IC PLASMA	3,5
8	204	7.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	151	-20.1	EMISSION, IC PLASMA	3,5
15	200	5.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
18	185	-2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	190	0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	188	-0.6	EMISSION, IC PLASMA	3,5
24	183	-3.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	180	-4.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
29	185	-2.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	186	-1.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
32	182	-3.7	EMISSION, IC PLASMA	1,2,3,4
33	18	-90.5	REJECT NOT REPORTED	3,5
36	195	3.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	167	-11.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	198	4.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
40	185	-2.2	EMISSION, IC PLASMA	3,5
41	179	-5.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	187	-1.1	EMISSION, FLAME	1,2,3,4
47	250	32.2	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
48	194	2.6	EMISSION, IC PLASMA	3,5
49	197	4.2	EMISSION, IC PLASMA	3,5
50	197	4.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	185	-2.2	EMISSION, IC PLASMA	3,5
52	190	0.5	EMISSION, IC PLASMA	3,5
53	191	1.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	182	-3.7	EMISSION, IC PLASMA	3,5
56	193	2.1	EMISSION, FLAME	1,2
57	190	0.5	EMISSION, IC PLASMA	3,5
58	182	-3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	208	10.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
61	215	13.7	EMISSION, IC PLASMA	3,5
63	248	31.2	REJECT EMISSION, FLAME	1,2
69	297	57.1	REJECT EMISSION, IC PLASMA	3,5
70	190	0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	190	0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	199	5.3	EMISSION, IC PLASMA	3,5
80	184	-2.7	EMISSION, IC PLASMA	3,5
81	175	-7.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	190	0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	226	19.5	EMISSION, FLAME	1,2
86	215	13.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	186	-1.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	150	-20.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	187	-1.1	OTHER	
98	175	-7.4	EMISSION, IC PLASMA	3,5
102	196	3.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
103	192	1.5	EMISSION, IC PLASMA	3,5
104	193	2.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
108	190	0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	170	-10.1	EMISSION, IC PLASMA	3,5
111	189	-0.0	EMISSION, IC PLASMA	3,5
112	188	-0.6	OTHER	
118	190	0.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	190	0.5	OTHER	
127	3	-98.4	REJECT OTHER	
142	190	0.5	EMISSION, IC PLASMA	3,5

60 Labs had a total range of 3 to 297 and a mean of 189.1 with a standard deviation of 12.8 and a 95% confidence interval of the mean +/- 3.4.



Table 12 Standard Reference Water Sample T95 Report for NI

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1		IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
2	< 1		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	< 20		IGNORED EMISSION, IC PLASMA	3,5
8	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	60	429.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	4	-64.7	ATOMIC ABSORPTION, FLAMELESS	3
13	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	2	-82.4	ATOMIC ABSORPTION, FLAMELESS	3
20	< 0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	< 40		IGNORED EMISSION, IC PLASMA	3,5
24	7	-38.3	ATOMIC ABSORPTION, FLAMELESS	3
25	26	129.2	ATOMIC ABSORPTION, FLAMELESS	3
29	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	16	41.1	ATOMIC ABSORPTION, FLAMELESS	3
32	5	-55.9	EMISSION, IC PLASMA	3,5
33	< 20		IGNORED NOT REPORTED	
34	14	23.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	2	-82.4	ATOMIC ABSORPTION, FLAMELESS	3
38	9	-20.7	ATOMIC ABSORPTION, FLAMELESS	3
40	9	-20.7	ATOMIC ABSORPTION, FLAMELESS	3
41	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	3	-73.6	ATOMIC ABSORPTION, FLAMELESS	3
47	23	102.8	ATOMIC ABSORPTION, FLAMELESS	3
48	< 50		IGNORED EMISSION, IC PLASMA	3,5
49	6	-47.1	EMISSION, IC PLASMA	3,5
53	< 40		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	2	-82.4	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
57	4	-64.7	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
58	31	173.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	7	-38.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
63	6	-47.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	12	5.8	EMISSION, IC PLASMA	3,5
70	40	252.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
73	8	-29.5	EMISSION, IC PLASMA	3,5
76	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
79	4	-64.7	ATOMIC ABSORPTION, FLAMELESS	3
80	< 5		IGNORED EMISSION, IC PLASMA	3,5
81	40	252.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	20	76.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
85	1	-91.2	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,4
90	11	-3.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	80	605.3	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
96	5	-55.9	OTHER	
98	< 10		IGNORED EMISSION, IC PLASMA	3,5
103	< 6		IGNORED EMISSION, IC PLASMA	3,5
104	3	-73.6	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
108	< 6		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
110	24	111.6	EMISSION, IC PLASMA	3,5
112	1	-91.2	OTHER	
113	25	120.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	< 20		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
121	6	-47.1	OTHER	
122	2	-82.4	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
124	< 30		IGNORED NOT REPORTED	
127	11	-3.0	OTHER	
142	< 4		IGNORED EMISSION, IC PLASMA	3,5
144	8	-29.5	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

60 Labs had a total range of 1 to 80 and a mean of 11.3  
with a standard deviation of 10.8 and a 95% confidence interval of the mean +/- 3.7.

Table 12 Standard Reference Water Sample T95 Report for PB

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1		IGNORED	
2	< 4	-57.4	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1, 4
4	< 5		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
6	< 50		ATOMIC ABSORPTION, FLAMELESS	3
7	< 50		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
8	< 9	-4.2	EMISSION, IC PLASMA	3, 5
9	< 4		ATOMIC ABSORPTION, FLAMELESS	3
10	< 3	-68.1	ATOMIC ABSORPTION, FLAMELESS	3
13	< 4		ATOMIC ABSORPTION, FLAMELESS	3
15	< 0		ATOMIC ABSORPTION, FLAMELESS	3
16	< 2	-78.7	ATOMIC ABSORPTION, FLAMELESS	3
18	< 9	-4.2	ATOMIC ABSORPTION, FLAMELESS	3
20	< 0		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
22	< 50		EMISSION, IC PLASMA	3, 5
24	< 6	-36.1	ATOMIC ABSORPTION, FLAMELESS	3
25	< 8		ATOMIC ABSORPTION, FLAMELESS	3
27	< 5		ATOMIC ABSORPTION, FLAMELESS	3
29	< 20		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
30	< 3	-68.1	ATOMIC ABSORPTION, FLAMELESS	3
32	< 38	304.6	ATOMIC ABSORPTION, FLAMELESS	3
33	< 50		IGNORED	
34	< 1	-89.4	NOT REPORTED	
37	< 1		OTHER	
38	< 1	-89.4	ATOMIC ABSORPTION, FLAMELESS	3
40	< 2	-78.7	ATOMIC ABSORPTION, FLAMELESS	3
41	< 8	-14.8	ATOMIC ABSORPTION, FLAMELESS	3
43	< 1	-89.4	ATOMIC ABSORPTION, FLAMELESS	3
46	< 1		ATOMIC ABSORPTION, FLAMELESS	3
47	< 3	-68.1	ATOMIC ABSORPTION, FLAMELESS	3
48	< 5		ATOMIC ABSORPTION, FLAMELESS	3
50	< 25		ANODIC STRIPPING VOLTAMMETRY	2
51	< 7	-25.5	ATOMIC ABSORPTION, FLAMELESS	3
54	< 114	1113.7	EXTRACTION (APDC/MIBK)	1, 4
57	< 10		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
58	< 15		EMISSION, IC PLASMA	3, 5
59	< 16	70.3	EXTRACTION (APDC/MIBK)	1, 2, 3, 4
63	< 10		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
69	< 31	230.0	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
73	< 32	240.7	EMISSION, IC PLASMA	3, 5
74	< 27	187.5	EMISSION, IC PLASMA	3, 5
76	< 7	-25.5	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
79	< 4	-57.4	ATOMIC ABSORPTION, FLAMELESS	3
80	< 5		ATOMIC ABSORPTION, FLAMELESS	3
81	< 5		ATOMIC ABSORPTION, FLAMELESS	3
83	< 20		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
84	< 50		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
85	< 2	-78.7	EXTRACTION (APDC/MIBK)	1, 4
93	< 20		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
94	< 100		ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
96	< 29	208.7	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7
98	< 5		ATOMIC ABSORPTION, FLAMELESS	3
103	< 37		EMISSION, IC PLASMA	3, 5
104	< 4	-57.4	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2, 3
107	< 1		ATOMIC ABSORPTION, FLAMELESS	3
110	< 1260	1E+04	EMISSION, IC PLASMA	3, 5
112	< 1		ATOMIC ABSORPTION, FLAMELESS	3
118	< 3		ATOMIC ABSORPTION, FLAMELESS	3
121	< 8	-14.8	OTHER	
124	< 50		NOT REPORTED	
127	< 1	-89.4	OTHER	
142	< 4		EMISSION, IC PLASMA	3, 5
144	< 1	-89.4	EMISSION, DC PLASMA	7

63 Labs had a total range of 1 to 1260 and a mean of 9.4 with a standard deviation of 11.1 and a 95% confidence interval of the mean +/- 4.3.

Table 12 Standard Reference Water Sample T95 Report for SB

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1		IGNORED	
7	< 100		EMISSION, IC PLASMA	3
8	< 7	16.7	EMISSION, IC PLASMA	3
13	< 4		ATOMIC ABSORPTION, FLAMELESS	3
22	< 500		ATOMIC ABSORPTION, FLAMELESS	3
38	< 2	-66.7	EMISSION, IC PLASMA	3
41	< 200		ATOMIC ABSORPTION, FLAMELESS	3
43	< 3		ATOMIC ABSORPTION, DIRECT, AIR	1, 3
47	< 34	466.7	ATOMIC ABSORPTION, FLAMELESS	3
48	< 5		NOT REPORTED	
48	< 5		ATOMIC ABSORPTION, FLAMELESS	3
54	< 5		ATOMIC ABSORPTION, FLAMELESS	3
57	< 1		ATOMIC ABSORPTION, FLAMELESS	3
58	< 3		ATOMIC ABSORPTION, HYDRIDE	2, 4
63	< 1	-83.3	ATOMIC ABSORPTION, FLAMELESS	3
69	< 10	66.7	ATOMIC ABSORPTION, FLAMELESS	3
73	< 10	66.7	EMISSION, IC PLASMA	3
76	< 5		EMISSION, IC PLASMA	3
83	< 100		ATOMIC ABSORPTION, FLAMELESS	3
98	< 40		ATOMIC ABSORPTION, DIRECT, AIR	1, 3
104	< 1		EMISSION, IC PLASMA	3
110	< 130	2066.7	ATOMIC ABSORPTION, HYDRIDE	2, 4
112	< 1		EMISSION, IC PLASMA	3
127	< 2		ATOMIC ABSORPTION, FLAMELESS	3
142	< 2		OTHER	
142	< 2		ATOMIC ABSORPTION, FLAMELESS	3

24 Labs had a total range of 1 to 130 and a mean of 6.0 with a standard deviation of 4.3 and a 95% confidence interval of the mean +/- 5.3.

Table 12 Standard Reference Water Sample T95 Report for SE

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	45	-18.6		
6	< 2		IGNORED ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
7	59	6.7	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
8	14	-74.7	ATOMIC ABSORPTION, FLAMELESS	3
9	51	-7.8	ATOMIC ABSORPTION, FLAMELESS	3
13	65	17.6	ATOMIC ABSORPTION, FLAMELESS	3
15	54	-2.3	ATOMIC ABSORPTION, FLAMELESS	3
16	49	-11.4	ATOMIC ABSORPTION, FLAMELESS	3
18	< 1		IGNORED ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
22	< 1		IGNORED ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
24	< 1		IGNORED ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
25	54	-2.3	ATOMIC ABSORPTION, FLAMELESS	3
27	51	-7.8	ATOMIC ABSORPTION, FLAMELESS	3
30	23	-58.4	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
34	60	8.5	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
36	< 1		IGNORED ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
37	35	-36.7	ATOMIC ABSORPTION, FLAMELESS	3
38	62	12.1	ATOMIC ABSORPTION, FLAMELESS	3
40	67	21.2	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
41	60	8.5	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
43	62	12.1	OTHER	3
46	32	-42.1	ATOMIC ABSORPTION, FLAMELESS	3
47	50	-9.6	ATOMIC ABSORPTION, FLAMELESS	3
48	29	-47.5	ATOMIC ABSORPTION, FLAMELESS	3
49	61	10.3	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
50	67	21.2	OTHER	3
52	250	352.2	REJECT EMISSION, IC PLASMA	3
53	69	24.8	ATOMIC ABSORPTION, FLAMELESS	3
54	49	-11.4	ATOMIC ABSORPTION, FLAMELESS	3
56	46	-16.8	ATOMIC ABSORPTION, FLAMELESS	3
57	74	33.8	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
58	45	-18.6	ATOMIC ABSORPTION, FLAMELESS	3
63	62	12.1	ATOMIC ABSORPTION, FLAMELESS	3
68	65	17.6	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
69	67	21.2	EMISSION, IC PLASMA	3
70	4	-92.8	ATOMIC ABSORPTION, FLAMELESS	3
73	60	8.5	EMISSION, IC PLASMA	3
74	78	41.1	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
76	73	32.0	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
80	65	17.6	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
81	59	6.7	ATOMIC ABSORPTION, FLAMELESS	3
83	< 100		IGNORED EMISSION, DC PLASMA	7
84	50	-9.6	ATOMIC ABSORPTION, FLAMELESS	3
90	< 1		IGNORED ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
93	26	-53.0	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
96	121	118.9	REJECT OTHER	3
98	78	41.1	EMISSION, IC PLASMA	3
104	54	-2.3	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
108	72	30.2	ATOMIC ABSORPTION, HYDRIDE	1, 2, 3, 4
110	330	496.9	REJECT EMISSION, IC PLASMA	3
112	90	62.8	ATOMIC ABSORPTION, FLAMELESS	3
121	53	-4.1	OTHER	3
122	44	-20.4	ATOMIC ABSORPTION, FLAMELESS	3
127	100	80.9	OTHER	3
142	55	-0.5	ATOMIC ABSORPTION, FLAMELESS	3

55 Labs had a total range of 4 to 330 and a mean of 55.3 with a standard deviation of 18.3 and a 95% confidence interval of the mean +/- 5.5.

Table 12 Standard Reference Water Sample T95 Report for SIO2

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	7.00	-14.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
13	8.45	3.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
15	8.00	-2.0	COLORIMETRIC, MOLYBDOSILICIC ACID	1, 2, 3
20	8.60	5.3	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
22	5.57	-31.8	EMISSION, IC PLASMA	5
25	7.90	-3.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
30	8.30	1.6	COLORIMETRIC, AMINO-NAPHTHOL SULFONIC ACID	3
32	2.60	-68.2	REJECT EMISSION, IC PLASMA	5
36	35.50	334.7	REJECT COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
38	7.00	-14.3	EMISSION, IC PLASMA	5
40	9.00	10.2	EMISSION, IC PLASMA	5
41	6.63	-18.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
47	8.60	5.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
49	8.30	1.6	EMISSION, IC PLASMA	5
52	7.69	-5.8	EMISSION, IC PLASMA	5
54	9.20	12.7	EMISSION, IC PLASMA	5
57	7.60	-6.9	EMISSION, IC PLASMA	5
73	11.30	38.4	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
83	8.00	-2.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
93	5.72	-30.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
95	8.10	-0.8	COLORIMETRIC, MOLYBDOSILICIC ACID	1, 2, 3
98	7.82	-4.2	EMISSION, IC PLASMA	5
103	7.11	-12.9	EMISSION, IC PLASMA	5
104	11.10	35.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4
110	10.77	31.9	EMISSION, IC PLASMA	5
112	8.00	-2.0	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	4
142	8.40	2.9	EMISSION, IC PLASMA	5

27 Labs had a total range of 2.60 to 35.50 and a mean of 8.166 with a standard deviation of 1.403 and a 95% confidence interval of the mean +/- 0.579.

Table 12 Standard Reference Water Sample T95 Report for SR

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	830	-0.8	EMISSION, IC PLASMA	3,5
6	850	1.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
8	860	2.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
13	850	1.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
15	1000	19.5	OTHER	
22	740	-11.6	EMISSION, IC PLASMA	3,5
24	760	-9.2	ATOMIC ABSORPTION, FLAMELESS	7
34	800	-4.4	EMISSION, IC PLASMA	3,5
36	729	-12.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
41	680	-18.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
47	830	-0.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
49	830	-0.8	EMISSION, IC PLASMA	3,5
51	826	-1.3	EMISSION, IC PLASMA	3,5
52	850	1.6	EMISSION, IC PLASMA	3,5
53	866	3.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
54	920	9.9	EMISSION, IC PLASMA	3,5
57	810	-3.2	EMISSION, IC PLASMA	3,5
63	28	-96.7	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
73	830	-0.8	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7
80	840	0.4	ATOMIC ABSORPTION, FLAMELESS	3
93	854	2.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
98	830	-0.8	EMISSION, IC PLASMA	3,5
103	869	3.8	EMISSION, IC PLASMA	3,5
104	900	7.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
110	900	7.5	EMISSION, IC PLASMA	3,5
112	850	1.6	OTHER	
127	850	1.6	OTHER	
142	820	-2.0	EMISSION, IC PLASMA	3,5
144	859	2.6	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

29 Labs had a total range of 28 to 1000 and a mean of 836.9 with a standard deviation of 60.8 and a 95% confidence interval of the mean +/- 23.6.

Table 12 Standard Reference Water Sample T95 Report for TL

Code Number	Reported value	Pct. dev. from mean	Methods	References
6	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
7	< 60		IGNORED OTHER	
8	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
15	< 1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
22	< 500		IGNORED NOT REPORTED	
38	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
41	< 200		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
43	< 1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
47	< 0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
48	< 1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
54	< 10		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
57	< 0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
63	< 2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
76	< 10		IGNORED NOT REPORTED	
83	< 20		IGNORED ANODIC STRIPPING VOLTAMMETRY	
93	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
98	< 30		IGNORED OTHER	
104	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
110	6	0.0	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	3
112	< 1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
127	< 1		IGNORED OTHER	
142	< 20		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
144	< 1		IGNORED MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

23 Labs had a total range of 0 to 6. INSUFFICIENT DATA TO DETERMINE MEAN AND STANDARD OF DEVIATION.

Table 12 Standard Reference Water Sample T95 Report for V

Code Number	Reported value	Pct. dev. from mean	Methods	References
6	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
8	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
13	< 12	-56.2	EMISSION, IC PLASMA	3,5
15	< 2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
22	70	155.5	EMISSION, IC PLASMA	3,5
24	52	89.8	ATOMIC ABSORPTION, FLAMELESS	3
34	< 10		IGNORED EMISSION, IC PLASMA	3,5
38	< 10		IGNORED EMISSION, IC PLASMA	3,5
41	< 200		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
47	1	-96.4	ATOMIC ABSORPTION, FLAMELESS	3
48	< 100		EMISSION, IC PLASMA	3,5
49	2	-92.7	EMISSION, IC PLASMA	3,5
54	3	-89.1	EMISSION, IC PLASMA	3,5
57	< 6		IGNORED EMISSION, IC PLASMA	3,5
63	< 40		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
69	4	-85.4	EMISSION, IC PLASMA	3,5
73	5	-81.8	EMISSION, IC PLASMA	3,5
80	< 5		IGNORED EMISSION, IC PLASMA	3,5
83	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
93	< 100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
98	< 3		IGNORED EMISSION, IC PLASMA	3,5
103	< 2		IGNORED EMISSION, IC PLASMA	3,5
104	100	265.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2
110	25	-8.8	EMISSION, IC PLASMA	3,5
127	< 30		IGNORED OTHER	
142	< 2		IGNORED EMISSION, IC PLASMA	3,5

26 Labs had a total range of 1 to 100 and a mean of 27.4 with a standard deviation of 34.8 and a 95% confidence interval of the mean +/- 24.9.

Table 12 Standard Reference Water Sample T95 Report for ZN

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	20	10.9	EMISSION, IC PLASMA	3,5
2	8	-55.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
4	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	18	-0.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
7	< 5		IGNORED EMISSION, IC PLASMA	3,5
8	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
9	15	-16.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	20	10.9	EMISSION, IC PLASMA	3,5
16	27	49.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
18	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	< 1		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	< 10		IGNORED EMISSION, IC PLASMA	3,5
24	28	55.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
25	< 5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
29	17	-5.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	22	22.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
32	399	2112.0	REJECT EMISSION, IC PLASMA	3,5
33	< 10		IGNORED NOT REPORTED	
34	22	22.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
36	18	-0.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	23	27.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	10	-44.6	EMISSION, IC PLASMA	3,5
40	20	10.9	EMISSION, IC PLASMA	3,5
41	20	10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	10	-44.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
47	140	676.2	REJECT ATOMIC ABSORPTION, FLAMELESS	3
48	20	10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
49	16	-11.3	EMISSION, IC PLASMA	3,5
51	14	-22.4	ATOMIC ABSORPTION, FLAMELESS	3
52	15	-16.8	EMISSION, IC PLASMA	3,5
53	18	-0.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
54	18	-0.2	EMISSION, IC PLASMA	3,5
56	20	10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	34	88.5	EMISSION, IC PLASMA	3,5
58	22	22.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	18	-0.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
63	19	5.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
69	13	-27.9	EMISSION, IC PLASMA	3,5
70	10	-44.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	18	-0.2	ANODIC STRIPPING VOLTAMMETRY	
73	15	-16.8	EMISSION, IC PLASMA	3,5
74	17	-5.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
76	< 20		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	18	-0.2	EMISSION, IC PLASMA	3,5
81	4	-77.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	20	10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	20	10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
85	20	10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
90	11	-39.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
93	20	10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
94	26	44.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
95	16	-11.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
98	18	-0.2	EMISSION, IC PLASMA	3,5
102	57	216.0	REJECT ANODIC STRIPPING VOLTAMMETRY	
103	16	-11.3	EMISSION, IC PLASMA	3,5
104	12	-33.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
107	32	77.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
108	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	50	177.2	REJECT EMISSION, IC PLASMA	3,5
112	10	-44.6	OTHER	
113	26	44.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	20	10.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
121	17	-5.8	OTHER	
122	6	-66.7	ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2
124	20	10.9	NOT REPORTED	
127	19	5.3	OTHER	
142	20	10.9	EMISSION, IC PLASMA	3,5

67 Labs had a total range of 4 to 399 and a mean of 18.0 with a standard deviation of 5.8 and a 95% confidence interval of the mean +/- 1.6.

Table 13. Statistics by method for standard reference sample T95

Determination	Method	Range:		Mean	Standard Deviation	N
		from	to			
ACID@CACO3	'TITRATION, COLORIMETRIC'	1150.000	- 1220.000	1176.667	37.859	3
	'TITRATION, ELECTROMETRIC'	1100.000	- 1200.000	1157.000	30.930	10
	_OVER-ALL_	1100.000	- 1220.000	1159.286	31.977	14
AG	ATOMIC ABSORPTION, DIRECT, AIR	2.000	- 90.000	4.920	3.065	5
	ATOMIC ABSORPTION, FLAMELESS	0.100	- 3.000	1.200	1.502	11
	EMISSION, IC PLASMA	1.000	- 560.000	-----	-----	-
	OTHER	1.500	- 1.500	-----	-----	-
_OVER-ALL_	0.100	- 560.000	1.753	1.382	19	
AL	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	50.000	- 420.000	167.143	134.625	7
	ATOMIC ABSORPTION, DIRECT, FLAMELESS	30.000	- 110.000	55.714	29.358	7
	EMISSION, IC PLASMA	10.000	- 690.000	198.750	240.264	8
	_OVER-ALL_	10.000	- 690.000	102.308	112.368	26
AS	ATOMIC ABSORPTION, FLAMELESS	0.200	- 7.200	1.143	0.744	7
	ATOMIC ABSORPTION, HYDRIDE, (NABH4)	0.500	- 5.000	1.700	1.561	8
	EMISSION, IC PLASMA	14.000	- 90.000	42.267	41.572	3
	OTHER	0.600	- 0.600	-----	-----	-
_OVER-ALL_	0.200	- 90.000	1.113	0.755	16	
B	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1000.000	- 1600.000	-----	-----	-
	COLORIMETRIC, CURCUMIN	190.000	- 2000.000	357.500	199.228	4
	EMISSION, IC PLASMA	1050.000	- 1250.000	1130.667	54.572	15
	_OVER-ALL_	190.000	- 2000.000	1090.690	368.306	29
BA	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	30.000	- 350.000	130.000	108.321	10
	ATOMIC ABSORPTION, FLAMELESS	30.000	- 470.000	64.000	24.083	5
	EMISSION, IC PLASMA	40.000	- 50.000	41.875	4.031	16
	_OVER-ALL_	30.000	- 470.000	48.387	14.398	31
BE	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	-----	-	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	1.000	- 1.000	-----	-----	-
	EMISSION, IC PLASMA	1.000	- 5.000	-----	-----	-
	_OVER-ALL_	1.000	- 5.000	-----	-----	-
CA	ATOMIC ABSORPTION, DIRECT, AIR	22.000	- 80.000	71.542	4.899	24
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	71.000	- 94.000	73.000	1.673	6
	EMISSION, IC PLASMA	63.000	- 82.000	72.412	4.718	17
	OTHER	7.000	- 87.000	77.500	7.767	4
_OVER-ALL_	7.000	- 126.000	72.076	5.262	53	
CD	ATOMIC ABSORPTION, DIRECT, AIR	0.400	- 6.000	3.480	2.157	5
	ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	0.500	- 0.500	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	0.100	- 33.000	0.395	0.237	20
	EMISSION, IC PLASMA	2.000	- 10.000	4.750	3.594	4
_OVER-ALL_	0.100	- 33.000	0.810	0.877	31	
CO	ATOMIC ABSORPTION, DIRECT, AIR	10.000	- 58.000	28.500	22.159	4
	ATOMIC ABSORPTION, FLAMELESS	15.000	- 23.000	19.000	4.000	3
	EMISSION, IC PLASMA	3.000	- 12.000	-----	-----	-
	_OVER-ALL_	3.000	- 58.000	14.556	9.541	9
CR TOT	ATOMIC ABSORPTION, DIRECT, AIR	5.000	- 10.000	6.667	2.066	6
	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1.000	- 1.000	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	1.000	- 12.000	2.765	2.107	17
	EMISSION, IC PLASMA	6.000	- 63.000	24.833	21.876	6
OTHER	2.000	- 6.000	4.000	2.000	3	
_OVER-ALL_	1.000	- 63.000	4.387	3.127	31	
CU	ATOMIC ABSORPTION, DIRECT, AIR	4.000	- 22.000	12.474	5.048	19
	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	7.000	- 16.000	10.333	4.933	3
	ATOMIC ABSORPTION, FLAMELESS	5.000	- 26.000	9.867	3.292	15
	EMISSION, IC PLASMA	4.000	- 50.000	11.091	4.636	11
OTHER	10.000	- 16.000	-----	-----	-	
_OVER-ALL_	4.000	- 50.000	11.375	4.720	56	
FE	ATOMIC ABSORPTION, DIRECT, AIR	10.000	- 610.000	15.308	5.345	13
	ATOMIC ABSORPTION, FLAMELESS	4.000	- 41.000	8.000	3.391	5
	EMISSION, IC PLASMA	1.000	- 60.000	11.375	7.615	8
	OTHER	1.000	- 20.000	10.000	8.602	4
_OVER-ALL_	1.000	- 610.000	12.333	6.578	30	
HG	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	0.200	- 1800.000	0.459	0.120	32
	OTHER	0.510	- 0.510	-----	-----	-
	_OVER-ALL_	0.200	- 1800.000	0.459	0.115	35
K	ATOMIC ABSORPTION, DIRECT, AIR	3.800	- 9.250	4.813	0.454	28
	EMISSION, FLAME PHOTOMETRIC	5.500	- 17.000	10.015	5.203	4
	EMISSION, IC PLASMA	3.900	- 5.300	4.545	0.368	13
	OTHER	4.440	- 5.700	4.947	0.665	3
_OVER-ALL_	3.030	- 17.000	4.771	0.670	50	
LI	ATOMIC ABSORPTION, DIRECT, AIR	17.000	- 40.000	27.714	6.969	7
	EMISSION, FLAME	27.000	- 40.000	32.667	6.658	3
	EMISSION, IC PLASMA	20.000	- 34.000	29.250	4.496	8
	_OVER-ALL_	17.000	- 40.000	28.895	5.820	19
MG	ATOMIC ABSORPTION, DIRECT, AIR	3.000	- 69.000	32.130	1.140	23
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	3.000	- 46.000	29.800	15.991	5
	EMISSION, IC PLASMA	29.000	- 36.000	32.700	1.809	20
	OTHER	32.000	- 51.000	38.000	8.981	4
_OVER-ALL_	3.000	- 69.000	32.539	1.743	52	

Table 13. Statistics by method for standard reference sample T95

Determination	Method	Range:		Mean	Standard Deviation	N
		from	to			
MN	ATOMIC ABSORPTION, DIRECT, AIR	3.000	20.000	5.889	2.800	9
	ATOMIC ABSORPTION, FLAMELESS	2.000	4.000	2.857	0.690	7
	EMISSION, IC PLASMA	1.000	18.000	3.000	1.732	7
	OTHER	2.000	10.000	-----	-----	-
	_OVER-ALL_	1.000	20.000	5.679	4.982	28
MO	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	12.000	12.000	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	4.000	31.000	13.600	10.831	5
	EMISSION, IC PLASMA	8.000	25.000	14.333	9.292	3
	OTHER	4.000	31.000	12.636	8.370	11
	_OVER-ALL_	4.000	31.000	12.636	8.370	11
NA	ATOMIC ABSORPTION, DIRECT, AIR	150.000	250.000	190.179	9.813	28
	EMISSION, FLAME	187.000	248.000	209.200	26.640	5
	EMISSION, IC PLASMA	151.000	297.000	189.444	10.240	18
	OTHER	3.000	190.000	188.333	1.528	3
	_OVER-ALL_	3.000	297.000	189.073	12.761	55
NI	ATOMIC ABSORPTION, DIRECT, AIR	6.000	80.000	30.364	23.338	11
	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1.000	4.000	2.333	1.528	3
	ATOMIC ABSORPTION, FLAMELESS	2.000	26.000	9.545	8.478	11
	EMISSION, IC PLASMA	5.000	24.000	11.000	7.746	5
	OTHER	1.000	11.000	5.750	4.113	4
_OVER-ALL_	1.000	80.000	11.343	10.754	35	
PB	ATOMIC ABSORPTION, DIRECT, AIR	4.000	114.000	40.250	50.056	4
	ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	2.000	4.000	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	1.000	38.000	4.643	2.925	14
	EMISSION, IC PLASMA	31.000	1260.000	-----	-----	-
	OTHER	1.000	8.000	-----	-----	-
_OVER-ALL_	1.000	1260.000	9.393	11.083	28	
SB	ATOMIC ABSORPTION, FLAMELESS	1.000	7.000	3.333	3.215	3
	EMISSION, IC PLASMA	10.000	130.000	-----	-----	-
	_OVER-ALL_	1.000	130.000	6.000	4.301	5
SE	ATOMIC ABSORPTION, FLAMELESS	4.000	90.000	49.083	17.639	24
	ATOMIC ABSORPTION, HYDRIDE	23.000	78.000	58.786	16.867	14
	EMISSION, IC PLASMA	60.000	330.000	157.000	124.828	5
	OTHER	53.000	121.000	80.600	28.728	5
	_OVER-ALL_	4.000	330.000	55.289	18.347	45
SIO2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	5.720	11.100	7.925	1.613	8
	COLORIMETRIC, ASCORBIC ACID REDUCTION TO MOLYBDATE BLUE	8.000	35.500	9.300	1.758	3
	EMISSION, IC PLASMA	2.600	10.770	8.042	1.354	11
	_OVER-ALL_	2.600	35.500	8.166	1.403	25
SR	ATOMIC ABSORPTION, DIRECT, AIR	28.000	900.000	824.333	71.488	9
	EMISSION, IC PLASMA	740.000	920.000	835.417	46.792	12
	OTHER	850.000	1000.000	-----	-----	-
	_OVER-ALL_	28.000	1000.000	836.893	60.804	28
TL	ATOMIC ABSORPTION, DIRECT, AIR	-----	-----	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	-----	-----	-----	-----	-
	OTHER	-----	-----	-----	-----	-
	_OVER-ALL_	6.000	6.000	-----	-----	-
V	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	100.000	100.000	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	1.000	52.000	-----	-----	-
	EMISSION, IC PLASMA	2.000	70.000	8.500	8.826	6
	_OVER-ALL_	1.000	100.000	27.400	34.834	10
ZN	ATOMIC ABSORPTION, DIRECT, AIR	4.000	32.000	18.719	6.007	32
	ATOMIC ABSORPTION, FLAMELESS	14.000	140.000	-----	-----	-
	EMISSION, IC PLASMA	10.000	399.000	16.846	3.078	13
	OTHER	10.000	19.000	15.333	4.726	3
	_OVER-ALL_	4.000	399.000	18.038	5.804	53

Table 14 Standard Reference Water Sample N16 Report for NH3-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
2	3.31	-2.5	OTHER	
6	3.20	-5.8	COLORIMETRIC, PHENATE	1,2,3
7	4.50	32.5	COLORIMETRIC, PHENATE	1,2,3
8	3.31	-2.5	COLORIMETRIC, PHENATE	1,2,3
10	24.00	606.8	REJECT ION SELECTIVE ELECTRODE	1,2,3,4
12	2.40	-29.3	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,2,3,4
13	2.90	14.9	COLORIMETRIC, PHENATE	1,2,3
14	2.76	-18.7	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
20	3.40	0.1	COLORIMETRIC, INDOPHENOL	4
22	3.90	14.9	ION SELECTIVE ELECTRODE	1,2,3,4
23	3.69	8.7	COLORIMETRIC, PHENATE	1,2,3
25	3.23	-4.9	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
27	4.17	22.8	COLORIMETRIC, INDOPHENOL	4
29	3.49	2.8	OTHER	
30	3.60	6.0	COLORIMETRIC, INDOPHENOL	4
32	3.10	-8.7	ION SELECTIVE ELECTRODE	1,2,3,4
33	1.99	-41.4	NOT REPORTED	
34	3.80	11.9	ION SELECTIVE ELECTRODE	1,2,3,4
35	3.30	-2.8	ION SELECTIVE ELECTRODE	1,2,3,4
36	3.39	-0.2	ION SELECTIVE ELECTRODE	1,2,3,4
37	3.08	-9.3	COLORIMETRIC, PHENATE	1,2,3
38	3.40	0.1	COLORIMETRIC, PHENATE	1,2,3
40	3.50	3.1	ION SELECTIVE ELECTRODE	1,2,3,4
41	3.80	11.9	COLORIMETRIC, PHENATE	1,2,3
43	3.60	6.0	COLORIMETRIC, PHENATE	1,2,3
44	3.50	3.1	ION SELECTIVE ELECTRODE	1,2,3,4
48	3.34	-1.6	OTHER	
53	3.18	-6.3	COLORIMETRIC, PHENATE	1,2,3
55	2.42	-28.7	COLORIMETRIC, INDOPHENOL	4
56	4.29	26.3	ION SELECTIVE ELECTRODE	1,2,3,4
57	3.50	3.1	COLORIMETRIC, PHENATE	1,2,3
58	3.66	7.8	COLORIMETRIC, PHENATE	1,2,3
65	2.92	-14.0	ION SELECTIVE ELECTRODE	1,2,3,4
69	3.46	1.9	NOT REPORTED	
72	3.30	-2.8	COLORIMETRIC, PHENATE	1,2,3
73	3.85	13.4	COLORIMETRIC, PHENATE	1,2,3
74	1.74	-48.8	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
76	3.00	-11.6	ION SELECTIVE ELECTRODE	1,2,3,4
79	0.40	-88.2	REJECT COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
81	3.20	-5.8	COLORIMETRIC, PHENATE	1,2,3
82	3.38	-0.5	COLORIMETRIC, PHENATE	1,2,3
83	3.50	3.1	OTHER	
85	3.42	0.7	COLORIMETRIC, PHENATE	1,2,3
86	3.71	9.3	COLORIMETRIC, PHENATE	1,2,3
91	3.60	6.0	COLORIMETRIC, PHENATE	1,2,3
93	4.16	22.5	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
94	3.37	-0.8	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
95	3.65	7.5	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
96	3.62	6.6	ION SELECTIVE ELECTRODE	1,2,3,4
99	2.49	-26.7	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
104	5.16	52.0	ION SELECTIVE ELECTRODE	1,2,3,4
106	0.34	-90.0	REJECT COLORIMETRIC, INDOPHENOL	4
109	2.90	-14.6	NOT REPORTED	
113	3.71	9.3	ION SELECTIVE ELECTRODE	1,2,3,4
121	3.33	-1.9	COLORIMETRIC, DISTILLATION, NESSLERIZATION	1,4
124	2.17	-36.1	NOT REPORTED	
128	3.80	11.9	COLORIMETRIC, PHENATE	1,2,3
142	3.60	6.0	COLORIMETRIC, PHENATE	1,2,3

58 Labs had a total range of 0.34 to 24.00 and a mean of 3.396 with a standard deviation of 0.587 and a 95% confidence interval of the mean +/- 0.159.



Table 14 Standard Reference Water Sample N16 Report for NO2-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
2	1.10	0.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
6	1.10	0.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
7	1.14	4.4	COLORIMETRIC, DIAZOTIZATION	1,3,4
8	1.12	2.6	COLORIMETRIC, DIAZOTIZATION	1,3,4
12	1.10	0.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
13	1.20	9.9	COLORIMETRIC, DIAZOTIZATION	1,3,4
14	1.21	10.8	COLORIMETRIC, DIAZOTIZATION	1,3,4
15	1.00	-8.4	COLORIMETRIC, DIAZOTIZATION	1,3,4
20	1.10	0.7	ION CHROMATOGRAPHY	2,3,6
22	2.20	101.4	REJECT ION CHROMATOGRAPHY	2,3,6
23	1.12	2.6	COLORIMETRIC, DIAZOTIZATION	1,3,4
24	0.89	-18.5	NOT REPORTED	
27	1.17	7.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
29	1.55	41.9	REJECT COLORIMETRIC, DIAZOTIZATION	1,3,4
30	1.13	3.5	COLORIMETRIC, DIAZOTIZATION	1,3,4
34	1.20	9.9	COLORIMETRIC, DIAZOTIZATION	1,3,4
35	1.10	0.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
36	1.63	49.3	REJECT ION CHROMATOGRAPHY	2,3,6
37	1.14	4.4	OTHER	
38	1.10	0.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
40	1.21	10.8	COLORIMETRIC, DIAZOTIZATION	1,3,4
41	1.61	47.4	REJECT COLORIMETRIC, DIAZOTIZATION	1,3,4
43	1.07	-2.0	COLORIMETRIC, DIAZOTIZATION	1,3,4
44	3.44	215.0	REJECT ION CHROMATOGRAPHY	2,3,6
46	1.02	-6.6	COLORIMETRIC, DIAZOTIZATION	1,3,4
48	1.08	-1.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
53	1.08	-1.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
55	0.93	-14.8	COLORIMETRIC, DIAZOTIZATION	1,3,4
56	0.73	-33.2	REJECT COLORIMETRIC, DIAZOTIZATION	1,3,4
57	1.10	0.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
58	1.09	-0.2	COLORIMETRIC, DIAZOTIZATION	1,3,4
59	1.08	-1.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
65	1.09	-0.2	COLORIMETRIC, DIAZOTIZATION	1,3,4
72	1.00	-8.4	COLORIMETRIC, DIAZOTIZATION	1,3,4
74	1.08	-1.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
76	1.05	-3.9	COLORIMETRIC, DIAZOTIZATION	1,3,4
79	1.40	28.2	REJECT COLORIMETRIC, DIAZOTIZATION	1,3,4
81	1.40	28.2	REJECT ION CHROMATOGRAPHY	2,3,6
83	1.08	-1.1	COLORIMETRIC, DIAZOTIZATION	1,3,4
85	1.11	1.6	COLORIMETRIC, DIAZOTIZATION	1,3,4
86	1.06	-2.9	COLORIMETRIC, DIAZOTIZATION	1,3,4
91	1.10	0.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
93	1.06	-2.9	COLORIMETRIC, DIAZOTIZATION	1,3,4
95	1.20	9.9	COLORIMETRIC, DIAZOTIZATION	1,3,4
99	1.01	-7.5	COLORIMETRIC, DIAZOTIZATION	1,3,4
108	1.00	-8.4	COLORIMETRIC, DIAZOTIZATION	1,3,4
109	1.09	-0.2	COLORIMETRIC, DIAZOTIZATION	1,3,4
113	1.13	3.5	COLORIMETRIC, DIAZOTIZATION	1,3,4
121	1.12	2.6	OTHER	
128	1.10	0.7	COLORIMETRIC, DIAZOTIZATION	1,3,4
142	1.10	0.7	COLORIMETRIC, DIAZOTIZATION	1,3,4

51 Labs had a total range of 0.73 to 3.44 and a mean of 1.092 with a standard deviation of 0.067 and a 95% confidence interval of the mean +/- 0.021.

Table 14 Standard Reference Water Sample N16 Report for NO3-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
2	1.29	-34.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
6	1.68	-14.3	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
7	1.86	-5.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
8	1.68	-14.3	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
10	1.90	-3.0	ION CHROMATOGRAPHY	2, 3, 6, 7
12	1.88	-4.0	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
13	1.80	-8.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
14	2.98	52.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
20	0.03	-98.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
22	1.70	-13.2	REJECT ION CHROMATOGRAPHY	2, 3, 6, 7
23	1.52	-22.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
24	1.95	-0.5	NOT REPORTED	1, 2, 3, 4
25	2.82	43.9	OTHER	
27	2.02	3.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
29	2.79	42.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
30	1.54	-21.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
32	2.60	32.7	ION SELECTIVE ELECTRODE	1, 2, 3, 4
33	2.73	39.3	NOT REPORTED	
34	1.50	-23.4	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
35	1.40	-28.5	COLORIMETRIC, BRUCINE	1, 2, 3, 4
36	1.56	-20.4	ION CHROMATOGRAPHY	2, 3, 6, 7
37	1.61	-17.8	COLORIMETRIC, DEVARDA'S ALLOY REDUCTION, DIAZOTIZATION	1
38	1.80	-8.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
40	1.90	-3.0	COLORIMETRIC, BRUCINE	1, 2, 3, 4
41	8.11	313.9	REJECT COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
43	1.99	1.6	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
45	2.20	12.3	OTHER	
46	2.65	35.3	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
47	2.10	7.2	COLORIMETRIC, BRUCINE	1, 2, 3, 4
48	1.76	-10.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
49	1.59	-18.8	ION CHROMATOGRAPHY	2, 3, 6, 7
53	2.63	34.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
55	3.03	54.6	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
56	2.05	4.6	COLORIMETRIC, BRUCINE	1, 2, 3, 4
57	1.60	-18.3	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
58	2.63	34.2	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
59	2.36	20.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
65	1.70	-13.2	OTHER	
72	1.90	-3.0	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
74	2.65	35.3	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
76	1.85	-5.6	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
79	0.48	-75.5	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
81	2.00	2.1	COLORIMETRIC, BRUCINE	1, 2, 3, 4
82	1.47	26.1	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
83	1.71	-12.7	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
84	2.73	39.3	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
85	1.74	-11.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
86	4.53	131.2	REJECT COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
91	1.30	-33.6	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
93	10.74	448.2	REJECT OTHER	1, 2, 3, 4
95	1.36	-30.6	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
96	2.67	36.3	ION SELECTIVE ELECTRODE	1, 2, 3, 4
99	1.78	-9.2	COLORIMETRIC, BRUCINE	1, 2, 3, 4
104	1.66	-15.3	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
108	1.75	-10.7	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4
121	1.61	-17.8	OTHER	
126	2.00	2.1	COLORIMETRIC, BRUCINE	1, 2, 3, 4
128	1.60	-18.3	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3
142	1.70	-13.2	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1, 2, 3, 4

59 Labs had a total range of 0.03 to 10.74 and a mean of 1.959 with a standard deviation of 0.509 and a 95% confidence interval of the mean +/- 0.138.

Table 14 Standard Reference Water Sample N16 Report for ORG-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
2	1.42	-13.2	COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	3,7
6	1.20	-26.6	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	2,3,4
7	0.30	-81.7	OTHER	
8	1.88	14.9	COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	3,7
13	2.00	22.3	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
20	0.80	-51.1	COLORIMETRIC, DIGESTION, DISTILLATION, INDOPHENOL	4
29	0.86	-47.4	OTHER	
30	0.09	-94.5	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
34	0.62	-62.1	DIGESTION, DISTILLATION, TITRATION	2,3,4
35	2.00	22.3	DIGESTION, DISTILLATION, ION SELECTIVE ELECTRODE	1,2,3
37	2.54	55.3	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
38	0.90	-45.0	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
40	0.71	-56.6	DIGESTION, DISTILLATION, ION SELECTIVE ELECTRODE	1,2,3
43	0.90	-45.0	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
48	3.04	85.8	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
53	4.70	187.3	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
55	0.82	-49.9	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	2,3,4
57	1.30	-20.5	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
58	0.39	-76.2	COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	3,7
65	4.43	170.8	DIGESTION, DISTILLATION, TITRATION	2,3,4
73	0.71	-56.6	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
74	3.06	87.1	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
75	0.27	-83.5	DIGESTION, DISTILLATION, ION SELECTIVE ELECTRODE	1,2,3
79	4.20	156.7	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	2,3,4
82	2.08	27.1	OTHER	
83	4.90	199.5	DIGESTION, DISTILLATION, TITRATION	2,3,4
85	1.33	-18.7	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	3,4
86	1.30	-20.5	COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	3,7
91	1.20	-26.6	OTHER	
95	1.71	4.5	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	2,3,4
99	1.34	-18.1	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	2,3,4
121	1.12	-31.5	OTHER	
128	0.70	-57.2	NOT REPORTED	
142	0.80	-51.1	NOT REPORTED	

34 Labs had a total range of 0.09 to 4.90 and a mean of 1.636 with a standard deviation of 1.301 and a 95% confidence interval of the mean +/- 0.453.

Table 14 Standard Reference Water Sample N16 Report for P, TOTAL

Code Number	Reported value	Pct. dev. from mean	Methods	References
2	2.17	0.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
3	2.23	3.2	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
6	2.39	10.6	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
7	2.27	5.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
8	2.19	1.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
12	2.07	-4.2	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
13	2.40	11.1	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
15	2.28	5.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
22	2.59	19.9	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
23	2.52	16.7	EMISSION, IC PLASMA	3, 5
29	1.57	-27.3	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
30	2.36	9.2	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
33	1.31	-30.1	NOT REPORTED	
34	1.95	-9.7	NOT REPORTED	
35	2.30	6.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
37	1.94	-10.2	OTHER	
38	2.10	-2.8	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
40	2.28	5.5	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
41	2.00	-7.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
43	1.90	-12.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
44	1.25	-42.1	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
47	2.60	20.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
48	2.26	4.6	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
49	2.26	4.6	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
53	2.35	8.8	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
55	4.47	106.9	REJECT	
56	0.88	-59.3	REJECT	
57	2.00	-7.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
58	2.30	6.5	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
59	2.04	-5.6	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
65	2.06	-4.6	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
69	1.45	-32.9	NOT REPORTED	
72	2.20	1.8	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
73	2.29	6.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
74	2.39	10.6	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
76	2.30	6.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
79	2.10	-2.8	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
81	1.90	-12.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
82	2.13	-1.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
83	2.20	1.8	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
84	2.32	7.4	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
85	2.22	2.8	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
86	1.90	-12.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
91	2.30	6.5	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
93	1.90	-12.0	OTHER	
95	2.26	4.6	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
96	2.27	5.1	OTHER	
99	2.58	19.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
104	2.44	13.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
108	2.86	32.4	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
109	1.08	-50.0	REJECT	
113	0.44	-79.6	REJECT	
121	2.16	-0.0	OTHER	
126	1.62	-25.0	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	1, 2, 3, 4
128	2.00	-7.4	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	4
142	2.40	11.1	EMISSION, IC PLASMA	3, 5

56 Labs had a total range of 0.44 to 4.47 and a mean of 2.160 with a standard deviation of 0.302 and a 95% confidence interval of the mean +/- 0.084.

Table 14 Standard Reference Water Sample N16 Report for PO4-P

Code Number	Reported value	Pct. dev. from mean	Methods	References
2	1.42	-3.4	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
3	< 0.05		COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
6	1.47	-0.0	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
7	1.44	-2.1	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
8	1.40	-4.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
10	1.50	2.0	ION CHROMATOGRAPHY	2, 3, 6
12	1.42	-3.4	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
13	1.40	-4.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
14	2.08	41.5	REJECT	1, 2, 3, 4
15	1.42	-3.4	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
20	1.50	2.0	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
22	0.09	-93.9	REJECT	2, 3, 6
23	1.44	-2.1	ION CHROMATOGRAPHY	1, 2, 3, 4
24	1.50	2.0	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
25	1.67	13.6	NOT REPORTED	
27	1.67	13.6	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
29	0.09	-93.9	REJECT	1, 2, 3, 4
30	1.50	2.0	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
34	1.35	-8.2	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
35	1.44	-2.1	OTHER	
36	1.46	-0.7	ION CHROMATOGRAPHY	2, 3, 6
37	1.51	2.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
38	1.30	-11.6	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
40	1.44	-2.1	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
41	1.47	-0.0	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
43	1.48	0.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
46	0.97	-34.0	REJECT	
47	1.70	15.6	OTHER	1, 2, 3, 4
48	1.26	-14.3	OTHER	
49	1.54	4.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
55	1.20	-18.4	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
56	2.71	84.3	REJECT	1, 2, 3, 4
57	1.40	-4.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
58	1.54	4.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
59	1.40	-4.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
65	1.48	0.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
69	1.16	-21.1	NOT REPORTED	
74	1.55	5.4	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
76	1.50	2.0	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
79	1.50	2.0	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
81	1.80	22.4	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
82	1.42	-3.4	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
83	1.40	-4.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
84	1.44	-2.1	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
85	1.48	0.7	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
86	1.62	10.2	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
91	1.40	-4.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
93	1.50	2.0	OTHER	
95	1.45	-1.4	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
99	1.58	7.5	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
104	1.50	2.0	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
108	0.44	-70.1	REJECT	1, 2, 3, 4
109	0.70	-52.4	REJECT	1, 2, 3, 4
113	1.35	-8.2	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
128	1.60	8.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4
142	1.60	8.8	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	1, 2, 3, 4

56 Labs had a total range of 0.09 to 2.71 and a mean of 1.470 with a standard deviation of 0.118 and a 95% confidence interval of the mean +/- 0.034.

Table 15. Statistics by method for standard reference sample N16

Determination	Method	Range:		Mean	Standard Deviation	N
		from	to			
NH3-N	COLORIMETRIC, DISTILLATION, NESSLERIZATION	0.400	4.160	2.753	1.079	10
	COLORIMETRIC, INDOPHENOL	0.340	4.170	2.786	1.506	5
	COLORIMETRIC, PHENATE	3.080	4.500	3.509	0.247	20
	ION SELECTIVE ELECTRODE	2.920	24.000	3.503	0.397	12
	NOT REPORTED	1.990	3.460	2.630	0.679	4
	OTHER	3.340	3.500	-----	-----	-
	_OVER-ALL_	0.340	24.000	3.396	0.587	55
NO2-N	COLORIMETRIC, DIAZOTIZATION	0.730	1.610	1.095	0.062	39
	ION CHROMATOGRAPHY	1.100	3.440	1.954	0.923	5
	_OVER-ALL_	0.730	3.440	1.092	0.067	43
NO3-N	COLORIMETRIC, BRUCINE	1.400	2.100	1.972	0.115	6
	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	0.030	8.110	1.869	0.660	32
	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	1.600	2.630	2.030	0.433	4
	ION CHROMATOGRAPHY	1.560	1.900	1.688	0.154	4
	OTHER	1.610	10.740	2.083	0.556	4
_OVER-ALL_	0.030	10.740	1.959	0.509	55	
ORG-N	COLORIMETRIC, BLOCK DIGESTION, SALICYLATE HYPOCHLORITE	0.090	4.700	1.870	1.352	11
	COLORIMETRIC, DIGESTION, DISTILLATION, NESSLERIZATION	0.820	4.200	1.268	0.368	4
	COLORIMETRIC, DIGESTION, DISTILLATION, PHENATE	0.390	1.880	1.248	0.624	4
	DIGESTION, DISTILLATION, ION SELECTIVE ELECTRODE	0.270	2.000	0.993	0.899	3
	DIGESTION, DISTILLATION, TITRATION	0.620	4.900	3.317	2.347	3
	OTHER	0.300	2.080	1.112	0.646	5
	_OVER-ALL_	0.090	4.900	1.636	1.301	34
P, TOTAL	COLORIMETRIC, H2SO4/PERSULF DIG. ASCORBIC ACID PHOSPHOMOLYBD	0.440	4.470	2.210	0.256	35
	COLORIMETRIC, BLK DIG, H2SO4, K&HG2SO4, PHOSPHOMOLYBDATE	0.880	2.350	2.144	0.181	8
	OTHER	1.900	2.300	2.158	0.182	4
	_OVER-ALL_	0.440	4.470	2.160	0.302	52
PO4-P	COLORIMETRIC, ASCORBIC ACID PHOSPHOMOLYBDATE	0.090	2.710	1.482	0.112	41
	ION CHROMATOGRAPHY	0.090	1.500	-----	-----	-
	OTHER	0.970	1.500	1.293	0.238	4
	_OVER-ALL_	0.090	2.710	1.470	0.118	48

Table 16 Standard Reference Water Sample P8 Report for ACID@CAC03

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 0.1		IGNORED 'TITRATION, ELECTROMETRIC'	1,2,3,4
2	< 1.0	-81.0	IGNORED 'TITRATION, ELECTROMETRIC'	1,2,3,4
8	< 5.0		IGNORED 'TITRATION, ELECTROMETRIC'	1,2,3,4
13	6.0	14.1	'TITRATION, ELECTROMETRIC'	1,2,3,4
19	15.0	185.3	'TITRATION, ELECTROMETRIC'	1,2,3,4
30	0.8	-84.8	'TITRATION, ELECTROMETRIC'	1,2,3,4
37	1.7	-67.7	'TITRATION, ELECTROMETRIC'	1,2,3,4
38	2.0	-62.0	'TITRATION, ELECTROMETRIC'	1,2,3,4
41	44.0	736.8	REJECT 'TITRATION, ELECTROMETRIC'	1,2,3,4
53	5.0	-4.9	'TITRATION, ELECTROMETRIC'	1,2,3,4
57	< 0.1		IGNORED 'TITRATION, ELECTROMETRIC'	1,2,3,4
64	2.0	-62.0	'TITRATION, ELECTROMETRIC'	1,2,3,4
83	2.0	-62.0	'TITRATION, ELECTROMETRIC'	1,2,3,4
102	12.6	139.6	'TITRATION, COLORIMETRIC'	1,2,3
104	2.0	-62.0	'TITRATION, COLORIMETRIC'	1,2,3
112	13.0	147.2	'TITRATION, ELECTROMETRIC'	1,2,3,4

16 Labs had a total range of 0.8 to 44.0 and a mean of 5.26 with a standard deviation of 5.24 and a 95% confidence interval of the mean +/- 3.33.

Table 16 Standard Reference Water Sample P8 Report for AG

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
2	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
6	< 5.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
8	< 0.5		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
10	< 0.3		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
13	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
15	< 0.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
22	< 20.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
30	0.2	-88.9	IGNORED EMISSION, IC PLASMA	3
37	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
40	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
41	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
43	< 0.2		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
47	1.2	-33.3	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
49	4.0	122.2	EMISSION, IC PLASMA	3
50	< 6.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
53	< 0.5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
57	< 0.0		IGNORED ATOMIC ABSORPTION, EXTRACTION (APDC/MIBK)	1,2,4
83	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
104	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3
110	60.0	3233.3	REJECT EMISSION, IC PLASMA	3
124	< 20.0		IGNORED NOT REPORTED	3
142	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3

23 Labs had a total range of 0.2 to 60.0. INSUFFICIENT DATA TO DEFINE MEAN AND STANDARD OF DEVIATION.

Table 16 Standard Reference Water Sample P8 Report for CA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	31.000	1E+04	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	0.220	-6.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	0.280	19.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	0.260	10.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	0.251	6.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	<4.000		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	0.260	10.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	1.000	325.5	REJECT EMISSION, IC PLASMA	3,5,7
26	0.230	-2.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	0.200	-14.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	0.300	27.7	NOT REPORTED	1,2,3,4
37	0.190	-19.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	0.180	-23.4	EMISSION, IC PLASMA	3,5,7
40	0.250	6.4	EMISSION, IC PLASMA	3,5,7
41	<0.010		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
43	0.250	6.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
47	0.250	6.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
49	0.220	-6.4	EMISSION, IC PLASMA	3,5,7
50	<50.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	0.216	-8.1	EMISSION, IC PLASMA	3,5,7
53	<1.000		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	0.200	-14.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
64	0.230	-2.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
67	0.160	-31.9	EMISSION, IC PLASMA	3,5,7
71	0.184	-21.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	0.240	2.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	<0.200		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
86	0.300	27.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
102	0.230	-2.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	0.250	6.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	0.420	78.7	REJECT EMISSION, IC PLASMA	3,5,7
111	0.200	-14.9	EMISSION, IC PLASMA	3,5,7
112	0.300	27.7	OTHER	3,5,7
113	0.600	155.3	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
124	<0.010		IGNORED NOT REPORTED	1,2,3,4
125	0.194	-17.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
142	0.300	27.7	EMISSION, IC PLASMA	3,5,7

37 Labs had a total range of 0.160 to 31.00 and a mean of .2350 with a standard deviation of .0396 and a 95% confidence interval of the mean +/- .0157.

Table 16 Standard Reference Water Sample P8 Report for CD

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 0.10		IGNORED ATOMIC ABSORPTION, EXTRACTION, (APDC/MIBK)	1,4
2	0.05	-95.7	ATOMIC ABSORPTION, FLAMELESS	1,3
6	< 5.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	< 5.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	< 0.10		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
13	< 0.20		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
15	0.08	-93.1	ATOMIC ABSORPTION, FLAMELESS	3
22	<10.00		IGNORED EMISSION, IC PLASMA	3,5
30	0.20	-82.7	ATOMIC ABSORPTION, FLAMELESS	3
37	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
38	0.12	-89.6	ATOMIC ABSORPTION, FLAMELESS	3
40	< 0.01		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
41	< 5.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	< 0.10		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
47	0.05	-95.7	ATOMIC ABSORPTION, FLAMELESS	3
49	1.00	-13.5	EMISSION, IC PLASMA	3,5
50	5.00	332.3	ANODIC STRIPPING VOLTAMMETRY, DIFFERENTIAL PULSE	2
51	0.08	-93.1	ATOMIC ABSORPTION, FLAMELESS	3
53	< 0.20		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
57	< 0.10		IGNORED OTHER	
83	< 5.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
102	2.00	72.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	< 1.00		IGNORED ATOMIC ABSORPTION, EXTRACTION, (PDCA/CHCL3)	2,3,5
110	5.00	332.3	EMISSION, IC PLASMA	3
112	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
124	<10.00		IGNORED NOT REPORTED	
142	0.10	-91.4	ATOMIC ABSORPTION, FLAMELESS	3
144	0.20	-82.7	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

28 Labs had a total range of 0.05 to 5.00.  
INSUFFICIENT DATA TO DEFINE MEAN AND STANDARD OF DEVIATION.

Table 16 Standard Reference Water Sample P8 Report for CL

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	<1.000		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
2	0.500	200.5	ION SELECTIVE ELECTRODE	1,2,3,4
6	0.100	-39.9	ION CHROMATOGRAPHY	2,3,6,7
13	<1.000		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
15	<0.500		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
19	0.050	-70.0	ION CHROMATOGRAPHY	2,3,6,7
22	0.100	-39.9	ION CHROMATOGRAPHY	2,3,6,7
26	0.080	-51.9	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
30	<0.200		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
37	1.540	825.5	REJECT ION SELECTIVE ELECTRODE	1,2,3,4
38	<0.100		IGNORED ION CHROMATOGRAPHY	2,3,6,7
40	<0.001		IGNORED TITRATION, SILVER NITRATE	1,2,4
41	2.000	1101.9	REJECT COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
43	0.100	-39.9	TITRATION, MERCURIC NITRATE	1,2,3,4
47	0.084	-49.5	ION CHROMATOGRAPHY	2,3,6,7
49	0.370	122.4	NOT REPORTED	
51	0.099	-40.5	ION CHROMATOGRAPHY	2,3,6,7
53	<2.000		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
57	0.150	-9.9	ION CHROMATOGRAPHY	2,3,6,7
67	0.096	-42.3	ION CHROMATOGRAPHY	2,3,6,7
71	0.102	-38.7	ION CHROMATOGRAPHY	2,3,6,7
82	0.400	140.4	COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
83	<1.000		IGNORED TITRATION, SILVER NITRATE	1,2,4
86	<0.500		IGNORED COLORIMETRIC, FERRIC THIOCYANATE	1,2,3,4
102	0.370	122.4	TITRATION, MERCURIC NITRATE	1,2,3,4
104	<1.000		IGNORED TITRATION, MERCURIC NITRATE	1,2,3,4
112	<0.500		IGNORED TITRATION, MERCURIC NITRATE	1,2,3,4
118	0.070	-57.9	ION CHROMATOGRAPHY	2,3,6,7
119	0.213	28.0	ION SELECTIVE ELECTRODE	1,2,3,4
124	1.400	741.3	REJECT NOT REPORTED	
125	0.071	-57.3	ION CHROMATOGRAPHY	2,3,6,7
142	0.040	-76.0	ION CHROMATOGRAPHY	2,3,6,7

32 Labs had a total range of 0.040 to 2.000 and a mean of .1664  
with a standard deviation of .1415 and a 95% confidence interval of the mean +/- .0704.

Table 16 Standard Reference Water Sample P8 Report for CO

Code Number	Reported value	Pct. dev. from mean	Methods	References
6	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	25.60	36.2	REJECT EMISSION, IC PLASMA	3,5
15	< 0.10		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
22	<10.00		IGNORED EMISSION, IC PLASMA	3,5
30	< 0.01		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
41	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
47	< 0.10		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
57	< 0.30		IGNORED OTHER	
83	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	12.00	-36.2	REJECT EMISSION, IC PLASMA	3,5
124	<10.00		IGNORED NOT REPORTED	
142	< 3.00		IGNORED EMISSION, IC PLASMA	3,5

14 Labs had a total range of 12.00 to 25.60.  
INSUFFICIENT DATA TO DEFINE MEAN AND STANDARD OF DEVIATION.



Table 16 Standard Reference Water Sample P8 Report for CR TOT

Code Number	Reported value	Pct. dev. from mean	Methods	References
2	< 0.20		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
6	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	< 0.60		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
13	<50.00		IGNORED EMISSION, IC PLASMA	3
15	< 0.50		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
22	<50.00		IGNORED EMISSION, IC PLASMA	3
30	1.00	-9.9	ATOMIC ABSORPTION, FLAMELESS	3
37	1.00	-9.9	ATOMIC ABSORPTION, FLAMELESS	3
38	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
40	0.70	-36.9	ATOMIC ABSORPTION, FLAMELESS	3
41	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	1.80	62.2	ATOMIC ABSORPTION, FLAMELESS	3
47	< 0.05		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
49	2.00	80.2	EMISSION, IC PLASMA	3
53	<30.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	0.16	-85.6	OTHER	3
83	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	< 1.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	8.00	620.7	REJECT EMISSION, IC PLASMA	3
112	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
124	<20.00		IGNORED NOT REPORTED	3
142	< 5.00		IGNORED EMISSION, IC PLASMA	3

23 Labs had a total range of 0.16 to 8.00 and a mean of 1.110 with a standard deviation of 0.688 and a 95% confidence interval of the mean +/- 0.722.

Table 16 Standard Reference Water Sample P8 Report for CU

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
2	< 1.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	< 5.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	<10.00		IGNORED EMISSION, IC PLASMA	3,5
15	0.60	-71.9	ATOMIC ABSORPTION, FLAMELESS	3
22	<10.00		IGNORED EMISSION, IC PLASMA	3,5
30	2.00	-6.3	ATOMIC ABSORPTION, FLAMELESS	3
37	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
38	<10.00		IGNORED EMISSION, IC PLASMA	3,5
40	< 0.01		IGNORED EMISSION, IC PLASMA	3,5
41	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	1.80	-15.7	ATOMIC ABSORPTION, FLAMELESS	3
47	0.73	-65.8	ATOMIC ABSORPTION, FLAMELESS	3
49	5.00	134.1	EMISSION, IC PLASMA	3,5
51	0.76	-64.4	ATOMIC ABSORPTION, FLAMELESS	3
53	0.70	-67.2	ATOMIC ABSORPTION, FLAMELESS	3
57	0.50	-76.6	OTHER	3
83	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
102	35.00	1539.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	< 1.00		IGNORED ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
110	10.00	368.3	REJECT NOT REPORTED	3
112	<10.00		IGNORED OTHER	3
113	5.00	134.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	< 3.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
124	<10.00		IGNORED NOT REPORTED	3
142	5.00	134.1	EMISSION, IC PLASMA	3,5
144	1.40	-34.4	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

28 Labs had a total range of 0.50 to 35.00 and a mean of 2.136 with a standard deviation of 1.904 and a 95% confidence interval of the mean +/- 1.279.

Table 16 Standard Reference Water Sample P8 Report for F

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.290	181.8	ION SELECTIVE ELECTRODE	1,2,3,4
2	0.040	-61.1	OTHER	3
6	<0.050		IGNORED ION SELECTIVE ELECTRODE	1,2,3,4
8	0.100	-2.8	COLORIMETRIC, CEROUS ALIZARIN "COMPLEXONE"	3
13	<0.100		IGNORED ION SELECTIVE ELECTRODE	1,2,3,4
15	<0.100		IGNORED ION SELECTIVE ELECTRODE	1,2,3,4
22	<0.100		IGNORED ION CHROMATOGRAPHY	2,3,6
30	0.008	-92.2	ION SELECTIVE ELECTRODE	1,2,3,4
37	<0.050		IGNORED ION SELECTIVE ELECTRODE	1,2,3,4
38	0.020	-80.6	ION CHROMATOGRAPHY	2,3,6
40	0.030	-70.8	ION SELECTIVE ELECTRODE	1,2,3,4
41	0.190	84.6	COLORIMETRIC, LANTHANUM ALIZARIN "COMPLEXONE"	1
47	0.018	-82.5	ION CHROMATOGRAPHY	2,3,6
49	0.100	-2.8	ION CHROMATOGRAPHY	2,3,6
50	<20.00		IGNORED COLORIMETRIC, CEROUS ALIZARIN "COMPLEXONE"	3
51	0.022	-78.6	ION CHROMATOGRAPHY	2,3,6
57	0.260	152.7	ION CHROMATOGRAPHY	2,3,6
83	<0.200		IGNORED ION SELECTIVE ELECTRODE	1,2,3,4
86	<0.500		IGNORED ION SELECTIVE ELECTRODE	1,2,3,4
104	0.210	104.1	ION SELECTIVE ELECTRODE	1,2,3,4
112	<0.010		IGNORED ION SELECTIVE ELECTRODE	1,2,3,4
142	0.050	-51.4	ION SELECTIVE ELECTRODE	1,2,3,4

22 Labs had a total range of 0.008 to 0.290 and a mean of .1029 with a standard deviation of .1002 and a 95% confidence interval of the mean +/- .0606.

Table 16 Standard Reference Water Sample P8 Report for PE

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 20.0		IGNORED EMISSION, IC PLASMA	3, 5
2	< 15.8	37.7	IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
6	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
8	< 50.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
13	< 10.0		IGNORED EMISSION, IC PLASMA	3, 5
15	< 20.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3, 5
22	< 30.0		IGNORED EMISSION, IC PLASMA	3, 5
30	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3, 5
37	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
38	< 10.0		IGNORED EMISSION, IC PLASMA	3, 5
40	< 0.1		IGNORED EMISSION, IC PLASMA	3, 5
41	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
43	< 30.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
47	< 0.4	-96.5	IGNORED ATOMIC ABSORPTION, FLAMELESS	3
49	1.0	-91.3	EMISSION, IC PLASMA	3, 5
50	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
51	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
53	< 50.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
57	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
67	1.6	-86.1	EMISSION, IC PLASMA	3, 5
83	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
104	< 10.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
110	8.0	-30.3	EMISSION, IC PLASMA	3, 5
112	< 10.0		IGNORED OTHER	
113	29.5	157.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
118	<100.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
124	< 10.0		IGNORED NOT REPORTED	
142	24.0	109.2	EMISSION, IC PLASMA	3, 5

28 Labs had a total range of 0.4 to 29.5.  
INSUFFICIENT DATA TO DEFINE MEAN AND STANDARD OF DEVIATION.

Table 16 Standard Reference Water Sample P8 Report for K

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	<0.100		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
6	0.040	-24.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
8	0.060	12.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
10	0.090	69.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
13	0.081	52.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
15	<0.100		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
19	0.020	-62.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
22	<1.000		IGNORED EMISSION, IC PLASMA	3
26	0.060	12.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
30	0.041	-22.9	EMISSION, FLAME, PHOTOMETRIC	1, 2
37	0.030	-43.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
38	<0.100		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
40	0.020	-62.4	EMISSION, IC PLASMA	3
41	0.090	69.2	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
43	0.040	-24.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
47	0.060	12.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
50	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
53	<0.500		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
57	0.020	-62.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
64	<0.001		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
67	0.064	20.3	EMISSION, FLAME, PHOTOMETRIC	1, 2
71	0.055	3.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
72	0.061	14.7	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
83	<0.100		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
102	0.150	182.0	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
104	0.030	-43.6	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
110	1.900	3471.4	REJECT EMISSION, IC PLASMA	3
112	0.070	31.6	OTHER	
113	0.080	50.4	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
125	0.051	-4.1	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
142	<0.300		IGNORED EMISSION, IC PLASMA	3

31 Labs had a total range of 0.020 to 1.900 and a mean of .0532  
with a standard deviation of .0226 and a 95% confidence interval of the mean +/- .0106.

Table 16 Standard Reference Water Sample P8 Report for MG

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	<0.100		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	<0.500		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	0.020	-38.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	0.040	22.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
13	0.030	-8.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
15	<2.400		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
19	0.030	-8.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	<1.000		IGNORED EMISSION, IC PLASMA	3,5
26	0.030	-8.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	0.024	-26.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	0.100	206.7	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	0.020	-38.7	EMISSION, IC PLASMA	3,5
40	0.030	-8.0	EMISSION, IC PLASMA	3,5
41	0.020	-38.7	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
43	0.028	-14.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
47	0.040	22.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
49	0.030	-8.0	EMISSION, IC PLASMA	3,5
50	<6.000		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	0.025	-23.3	EMISSION, IC PLASMA	3,5
53	<0.500		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	0.040	22.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
64	0.036	10.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
67	0.026	-20.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	0.022	-32.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
72	0.027	-17.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	<0.100		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
102	0.040	22.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	0.030	-8.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	0.110	237.4	REJECT EMISSION, IC PLASMA	3,5
111	0.060	84.0	EMISSION, IC PLASMA	3,5
112	0.050	53.4	OTHER	3,5
113	0.050	53.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
124	<0.010		IGNORED NOT REPORTED	1,2,3,4
125	0.028	-14.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
142	0.040	22.7	EMISSION, IC PLASMA	3,5

35 Labs had a total range of 0.020 to 0.110 and a mean of .0326 with a standard deviation of .0102 and a 95% confidence interval of the mean +/- .0042.

Table 16 Standard Reference Water Sample P8 Report for MN

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	<10.00		IGNORED EMISSION, IC PLASMA	3,5
2	6.70	-2.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	<10.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
13	<10.00		IGNORED EMISSION, IC PLASMA	3,5
22	<10.00		IGNORED EMISSION, IC PLASMA	3,5
30	8.00	16.4	ATOMIC ABSORPTION, FLAMELESS	3
37	10.00	45.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	< 5.00		IGNORED EMISSION, IC PLASMA	3,5
40	6.00	-12.7	EMISSION, IC PLASMA	3,5
41	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	7.80	13.5	ATOMIC ABSORPTION, FLAMELESS	3
47	6.10	-11.2	ATOMIC ABSORPTION, FLAMELESS	3
49	8.00	16.4	EMISSION, IC PLASMA	3,5
50	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
51	7.36	7.1	ATOMIC ABSORPTION, FLAMELESS	3
53	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
57	6.00	-12.7	OTHER	3,5
67	4.00	-41.8	EMISSION, IC PLASMA	3,5
83	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
86	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	5.00	-27.2	EMISSION, IC PLASMA	3,5
112	<10.00		IGNORED OTHER	3,5
113	14.00	103.7	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	<40.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
124	<10.00		IGNORED NOT REPORTED	3,5
142	7.50	9.1	EMISSION, IC PLASMA	3,5

29 Labs had a total range of 4.00 to 14.00 and a mean of 6.872 with a standard deviation of 1.588 and a 95% confidence interval of the mean +/- 1.009.

Table 16 Standard Reference Water Sample P8 Report for NA

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	0.100	0.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
2	0.080	-19.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	0.210	111.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	0.120	21.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
10	0.160	61.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	0.270	172.5	EMISSION, IC PLASMA	3,5
15	<0.100		IGNORED	
19	0.040	-59.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	3.000	2927.2	REJECT	
26	0.040	-59.6	EMISSION, IC PLASMA	3,5
30	0.052	-47.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	0.079	-20.3	EMISSION, FLAME	1,2,3,4
38	<0.100		IGNORED	
40	0.060		ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
41	0.070	-29.4	EMISSION, IC PLASMA	3,5
43	0.020	-79.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
47	0.070	-29.4	EMISSION, FLAME	1,2,3,4
50	<10.00		IGNORED	
51	0.144	45.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
53	<5.000		IGNORED	
57	0.070	-29.4	EMISSION, IC PLASMA	3,5
64	0.020	-79.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
67	0.084	-15.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
71	0.079	-20.3	EMISSION, FLAME	1,2,3,4
72	0.063	-36.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
83	<0.100		IGNORED	
102	0.500	404.5	REJECT	
104	0.130	31.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	0.250	152.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	0.090	-9.2	EMISSION, IC PLASMA	3,5
118	<0.100		OTHER	
125	0.078	-21.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
142	0.400	303.6	REJECT	
142	0.400	303.6	EMISSION, IC PLASMA	3,5

33 Labs had a total range of 0.020 to 3.000 and a mean of .0991 with a standard deviation of .0661 and a 95% confidence interval of the mean +/- .0279.

Table 16 Standard Reference Water Sample P8 Report for NH3-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	<0.020		IGNORED	
6	0.022	-19.4	OTHER	1,2,3
8	0.030	9.9	COLORIMETRIC, PHENATE	1,2,3
10	<0.040		IGNORED	
13	0.040	46.5	COLORIMETRIC, PHENATE	1,2,3,4
15	<0.020		IGNORED	
22	<0.100		IGNORED	
26	0.020	-26.7	ION SELECTIVE ELECTRODE	1,2,3,4
30	0.034	24.5	COLORIMETRIC, INDOPHENOL	4
37	<0.050		IGNORED	
38	0.020	-26.7	COLORIMETRIC, PHENATE	1,2,3
40	<0.100		IGNORED	
41	0.020	-26.7	COLORIMETRIC, PHENATE	1,2,3,4
43	<0.100		IGNORED	
44	0.150	449.5	ION SELECTIVE ELECTRODE	1,2,3,4
50	0.030	9.9	NOT REPORTED	
57	0.044	61.2	COLORIMETRIC, INDOPHENOL	4
64	0.020	-26.7	COLORIMETRIC, INDOPHENOL	4
67	0.025	-8.4	COLORIMETRIC, INDOPHENOL	4
71	0.015	-45.1	COLORIMETRIC, PHENATE	1,2,3
72	0.020	-26.7	COLORIMETRIC, PHENATE	1,2,3
82	0.027	-1.1	COLORIMETRIC, PHENATE	1,2,3
83	<0.100		IGNORED	
86	0.050	83.2	OTHER	1,2,3
104	0.020	-26.7	COLORIMETRIC, PHENATE	1,2,3,4
109	<0.050		IGNORED	
112	<0.010		ION SELECTIVE ELECTRODE	1,2,3,4
112	<0.010		IGNORED	
142	<0.100		IGNORED	
142	<0.100		COLORIMETRIC, PHENATE	1,2,3

28 Labs had a total range of 0.015 to 0.150 and a mean of .0273 with a standard deviation of .0101 and a 95% confidence interval of the mean +/- .0054.

Table 16 Standard Reference Water Sample P8 Report for NO3-N

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	<1.000		IGNORED	
6	0.069	48.1	ION CHROMATOGRAPHY	2,3,6,7
8	0.050	7.3	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
22	0.050	7.3	ION CHROMATOGRAPHY	2,3,6,7
26	0.050	7.3	ION CHROMATOGRAPHY	2,3,6,7
49	0.150	221.9	ION CHROMATOGRAPHY	2,3,6,7
51	0.278	496.6	REJECT	
64	0.062	33.0	ION CHROMATOGRAPHY	2,3,6,7
67	0.060	28.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
71	0.056	20.2	ION CHROMATOGRAPHY	1,2,3,4
118	0.060	28.8	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	1,2,3,4
119	0.009	-80.7	ION CHROMATOGRAPHY	2,3,6,7
119	0.009	-80.7	COLORIMETRIC, HYDRAZINE REDUCTION, DIAZOTIZATION	3

13 Labs had a total range of 0.000 to 0.278 and a mean of .0466 with a standard deviation of .0231 and a 95% confidence interval of the mean +/- .0165.

Table 16 Standard Reference Water Sample P8 Report for PB

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
2	0.15	-72.9	ANODIC STRIPPING VOLTAMMETRY	2
6	<50.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	30.00	5315.2	REJECT ATOMIC ABSORPTION, FLAMELESS	3
10	< 3.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
13	< 4.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
15	< 0.20		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
22	<50.00		IGNORED EMISSION, IC PLASMA	3,5
30	0.20	-63.9	ATOMIC ABSORPTION, FLAMELESS	3
37	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
38	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
40	10.00	1705.1	REJECT ATOMIC ABSORPTION, FLAMELESS	3
41	< 5.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
43	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
47	0.75	35.4	ATOMIC ABSORPTION, FLAMELESS	3
50	<25.00		IGNORED ANODIC STRIPPING VOLTAMMETRY	2
51	0.77	39.0	ATOMIC ABSORPTION, FLAMELESS	3
53	< 2.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
57	< 0.30		IGNORED OTHER	3
83	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	< 2.00		IGNORED ATOMIC ABSORPTION, EXTRACTION (PDCA/CHCL3)	2,3
110	100.	2E+04	REJECT EMISSION, IC PLASMA	3,5
112	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
124	<50.00		IGNORED NOT REPORTED	3
142	< 4.00		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
144	0.90	62.5	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

26 Labs had a total range of 0.15 to 100.  
INSUFFICIENT DATA TO DEFINE MEAN AND STANDARD OF DEVIATION.

Table 16 Standard Reference Water Sample P8 Report for PH

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	5.90	-4.0	ELECTROMETRIC	1,2,3,4
2	5.98	-2.7	ELECTROMETRIC	1,2,3,4
6	6.00	-2.4	ELECTROMETRIC	1,2,3,4
8	5.70	-7.3	ELECTROMETRIC	1,2,3,4
10	6.80	10.6	ELECTROMETRIC	1,2,3,4
13	6.48	5.4	ELECTROMETRIC	1,2,3,4
15	7.10	15.5	ELECTROMETRIC	1,2,3,4
22	7.75	26.0	ELECTROMETRIC	1,2,3,4
26	5.85	-4.9	ELECTROMETRIC	1,2,3,4
30	6.28	2.1	ELECTROMETRIC	1,2,3,4
33	8.19	33.2	NOT REPORTED	1,2,3,4
37	6.09	-1.0	ELECTROMETRIC	1,2,3,4
38	6.19	0.7	ELECTROMETRIC	1,2,3,4
40	6.19	0.7	ELECTROMETRIC	1,2,3,4
41	5.10	-17.1	ELECTROMETRIC	1,2,3,4
43	6.09	-1.0	ELECTROMETRIC	1,2,3,4
44	8.20	33.4	ELECTROMETRIC	1,2,3,4
47	4.60	-25.2	ELECTROMETRIC	1,2,3,4
50	5.08	-17.4	NOT REPORTED	1,2,3,4
53	5.57	-9.4	ELECTROMETRIC	1,2,3,4
57	4.60	-25.2	ELECTROMETRIC	1,2,3,4
64	6.10	-0.8	ELECTROMETRIC	1,2,3,4
67	5.66	-7.9	ELECTROMETRIC	1,2,3,4
71	5.67	-7.8	ELECTROMETRIC	1,2,3,4
83	5.80	-5.7	ELECTROMETRIC	1,2,3,4
86	5.85	-4.9	ELECTROMETRIC	1,2,3,4
102	8.00	30.1	ELECTROMETRIC	1,2,3,4
104	5.49	-10.7	ELECTROMETRIC	1,2,3,4
111	8.30	35.0	ELECTROMETRIC	1,2,3,4
112	5.80	-5.7	ELECTROMETRIC	1,2,3,4
113	6.22	1.2	ELECTROMETRIC	1,2,3,4
118	6.01	-2.3	ELECTROMETRIC	1,2,3,4
142	4.26	-30.7	ELECTROMETRIC	1,2,3,4

33 Labs had a total range of 4.26 to 8.30 and a mean of 6.149  
with a standard deviation of 1.017 and a 95% confidence interval of the mean +/- 0.359.

Table 16 Standard Reference Water Sample P8 Report for SO4

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	<2.000		IGNORED GRAVIMETRIC, BARIUM SULFATE	1,2,3
2	<2.500		IGNORED TURBIDIMETRIC, BARIUM SULFATE	1,2,3
6	0.310	-57.0	ION CHROMATOGRAPHY	2,6,7
8	<5.000		COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
13	<10.00		IGNORED COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
15	<5.000		IGNORED TURBIDIMETRIC, BARIUM SULFATE	1,2,3
19	0.710	-1.5	ION CHROMATOGRAPHY	2,6,7
22	0.390	-45.9	ION CHROMATOGRAPHY	2,6,7
26	0.350	-51.5	ION CHROMATOGRAPHY	2,6,7
30	<1.000		IGNORED COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
33	25.00	3366.9	NOT REPORTED	
37	2.390	231.4	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
38	0.300	-58.4	ION CHROMATOGRAPHY	2,6,7
40	<0.001		IGNORED TURBIDIMETRIC, BARIUM SULFATE	1,2,3
41	7.000	870.7	REJECT COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
43	<1.000		IGNORED TURBIDIMETRIC, BARIUM SULFATE	1,2,3
47	78.90	1E+04	REJECT ION CHROMATOGRAPHY	2,6,7
49	0.880	22.0	ION CHROMATOGRAPHY	2,6,7
50	<1.000		IGNORED NOT REPORTED	
51	0.335	-53.5	ION CHROMATOGRAPHY	2,6,7
53	5.000	593.4	REJECT COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
57	0.330	-54.2	ION CHROMATOGRAPHY	2,6,7
67	0.126	-82.5	ION CHROMATOGRAPHY	2,6,7
71	0.417	-42.2	ION CHROMATOGRAPHY	2,6,7
83	4.000	454.7	REJECT GRAVIMETRIC, BARIUM SULFATE	1,2,3
86	0.800	10.9	COLORIMETRIC, METHYL THYMOL BLUE	1,3,4
104	2.000	177.4	GRAVIMETRIC, BARIUM SULFATE	1,2,3
112	<1.000		IGNORED TURBIDIMETRIC, BARIUM SULFATE	1,2,3
113	2.100	191.2	GRAVIMETRIC, BARIUM SULFATE	1,2,3
118	0.500	-30.7	ION CHROMATOGRAPHY	2,6,7
119	0.301	-58.3	ION CHROMATOGRAPHY	2,6,7
125	0.370	-48.7	ION CHROMATOGRAPHY	2,6,7
142	0.370	-48.7	ION CHROMATOGRAPHY	2,6,7

33 Labs had a total range of 0.126 to 78.90 and a mean of .7211 with a standard deviation of .6932 and a 95% confidence interval of the mean +/- .3447.

Table 16 Standard Reference Water Sample P8 Report for SP. COND.

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	2.00	-55.7	DIRECT READING INSTRUMENT	4
2	3.42	-24.3	DIRECT READING INSTRUMENT	4
6	3.80	-15.9	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
8	4.00	-11.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
10	2.91	-35.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
13	6.90	52.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
15	32.00	608.4	REJECT DIRECT READING INSTRUMENT	4
22	4.11	-9.0	DIRECT READING INSTRUMENT	4
26	4.10	-9.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
30	3.40	-24.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
33	17.00	276.3	REJECT NOT REPORTED	
37	4.00	-11.4	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
38	2.80	-38.0	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
40	5.00	10.7	DIRECT READING INSTRUMENT	4
41	7.00	55.0	DIRECT READING INSTRUMENT	4
43	< 1.00		IGNORED DIRECT READING INSTRUMENT	4
47	8.70	92.6	DIRECT READING INSTRUMENT	4
50	5.70	26.2	NOT REPORTED	
53	481.	1E+04	REJECT WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
57	6.00	32.8	DIRECT READING INSTRUMENT	4
64	3.20	-29.2	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
67	4.33	-4.1	DIRECT READING INSTRUMENT	4
71	3.16	-30.0	DIRECT READING INSTRUMENT	4
72	2.70	-40.2	DIRECT READING INSTRUMENT	4
82	4.00	-11.4	DIRECT READING INSTRUMENT	4
83	3.17	-29.8	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
86	4.70	4.0	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
102	4.80	6.3	DIRECT READING INSTRUMENT	4
104	5.00	10.7	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
109	6.80	50.5	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
111	6.40	41.7	DIRECT READING INSTRUMENT	4
118	3.00	-33.6	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	1,2,3,4
142	5.90	30.6	DIRECT READING INSTRUMENT	4

33 Labs had a total range of 2.00 to 481. and a mean of 4.517 with a standard deviation of 1.595 and a 95% confidence interval of the mean +/- 0.606.

Table 16 Standard Reference Water Sample P8 Report for TL

Code Number	Reported value	Pct. dev. from mean	Methods	References
6	< 50.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
8	<100.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
13	< 4.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
15	< 0.5		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
22	<500.0		IGNORED NOT REPORTED	3
38	< 50.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
41	<200.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
43	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
47	< 0.3	-40.0	REJECT ATOMIC ABSORPTION, FLAMELESS	3
57	< 0.2		IGNORED OTHER	3
83	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
104	< 20.0		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,3
110	600.0	1E+05	REJECT MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7
112	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
142	< 2.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
144	0.7	40.0	REJECT MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

16 Labs had a total range of 0.3 to 600.0 and a mean of 0.50 with a standard deviation of 0.45 and a 95% confidence interval of the mean +/- 0.00.

Table 16 Standard Reference Water Sample P8 Report for ZN

Code Number	Reported value	Pct. dev. from mean	Methods	References
1	<10.00		IGNORED EMISSION, IC PLASMA	3,5
2	4.60	-47.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
6	6.00	-31.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
8	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	13.40	53.6	EMISSION, IC PLASMA	1,2,3,4
22	<10.00		IGNORED EMISSION, IC PLASMA	3,5
30	12.00	37.6	ATOMIC ABSORPTION, DIRECT, AIR	3,5
37	13.00	49.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	<10.00		IGNORED EMISSION, IC PLASMA	1,2,3,4
40	11.00	26.1	EMISSION, IC PLASMA	3,5
41	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	5.00	-42.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
47	110.	1161.0	REJECT ATOMIC ABSORPTION, FLAMELESS	3
49	6.00	-31.2	EMISSION, IC PLASMA	1,2,3,4
51	7.10	-18.6	ATOMIC ABSORPTION, FLAMELESS	3
53	<10.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	3,5
57	3.70	-57.6	OTHER	3
67	4.10	-53.0	EMISSION, IC PLASMA	1,2,3,4
83	20.00	129.3	ATOMIC ABSORPTION, DIRECT, AIR	3,5
102	4.00	-54.1	ANODIC STRIPPING VOLTAMMETRY	1,2,3,4
104	4.00	-54.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
110	11.00	26.1	EMISSION, IC PLASMA	3,5
112	<10.00		IGNORED OTHER	3,5
113	14.70	68.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
118	<20.00		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
124	<10.00		IGNORED NOT REPORTED	1,2,3,4
142	8.70	-0.3	EMISSION, IC PLASMA	3,5

27 Labs had a total range of 3.70 to 110. and a mean of 8.724 with a standard deviation of 4.768 and a 95% confidence interval of the mean +/- 2.452.

Table 17. Statistics by method for standard reference sample P8

Determination	Method	Range:		Mean	Standard Deviation	N
		from	to			
ACID@CACO3	'TITRATION, COLORIMETRIC'	2.000	- 2.000	2.000	0.000	3
	'TITRATION, ELECTROMETRIC'	0.800	- 44.000	5.563	5.555	8
	_OVER-ALL_	0.800	- 44.000	5.258	5.244	12
AG	ATOMIC ABSORPTION, DIRECT, AIR	-----	-----	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	0.200	- 1.200	-----	-----	-
	EMISSION, IC PLASMA	4.000	- 60.000	-----	-----	-
_OVER-ALL_	0.200	- 60.000	1.800	1.970	3	
CA	ATOMIC ABSORPTION, DIRECT, AIR	0.184	- 31.000	0.233	0.032	17
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	0.260	- 0.260	-----	-----	-
	EMISSION, IC PLASMA	0.160	- 1.000	0.243	0.083	8
_OVER-ALL_	0.160	- 31.000	0.235	0.040	27	
CD	ATOMIC ABSORPTION, DIRECT, AIR	2.000	- 2.000	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	0.050	- 0.200	0.097	0.052	7
	EMISSION, IC PLASMA	1.000	- 5.000	-----	-----	-
_OVER-ALL_	0.050	- 5.000	1.157	1.885	12	
CL	COLORIMETRIC, FERRIC THIOCYANATE	0.080	- 2.000	0.827	1.029	3
	ION CHROMATOGRAPHY	0.040	- 0.150	0.087	0.030	11
	ION SELECTIVE ELECTRODE	0.213	- 1.540	0.751	0.698	3
	TITRATION, MERCURIC NITRATE	0.100	- 0.370	-----	-----	-
_OVER-ALL_	0.040	- 2.000	0.166	0.142	18	
CO	ATOMIC ABSORPTION, DIRECT, AIR	-----	-----	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	-----	-----	-----	-----	-
	EMISSION, IC PLASMA	12.000	- 25.600	-----	-----	-
_OVER-ALL_	12.000	- 25.600	-----	-----	-	
CR TOT	ATOMIC ABSORPTION, DIRECT, AIR	-----	-----	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	0.700	- 1.800	1.125	0.472	4
	EMISSION, IC PLASMA	2.000	- 8.000	-----	-----	-
_OVER-ALL_	0.160	- 8.000	1.110	0.688	6	
CU	ATOMIC ABSORPTION, DIRECT, AIR	5.000	- 35.000	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	0.600	- 2.000	1.098	0.626	6
	EMISSION, IC PLASMA	5.000	- 5.000	-----	-----	-
_OVER-ALL_	0.500	- 35.000	2.136	1.904	11	
F	ION CHROMATOGRAPHY	0.018	- 0.260	0.084	0.104	5
	ION SELECTIVE ELECTRODE	0.008	- 0.290	0.118	0.125	5
	_OVER-ALL_	0.008	- 0.290	0.103	0.100	13
FE	ATOMIC ABSORPTION, DIRECT, AIR	15.800	- 29.500	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	0.400	- 0.400	-----	-----	-
	EMISSION, IC PLASMA	1.000	- 24.000	8.650	10.712	4
_OVER-ALL_	0.400	- 29.500	11.471	11.849	7	
K	ATOMIC ABSORPTION, DIRECT, AIR	0.020	- 0.150	0.054	0.023	16
	EMISSION, IC PLASMA	0.020	- 1.900	-----	-----	-
	_OVER-ALL_	0.020	- 1.900	0.053	0.023	20
MG	ATOMIC ABSORPTION, DIRECT, AIR	0.020	- 0.100	0.031	0.008	16
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	0.020	- 0.040	-----	-----	-
	EMISSION, IC PLASMA	0.020	- 0.110	0.034	0.014	6
_OVER-ALL_	0.020	- 0.110	0.033	0.010	25	
MN	ATOMIC ABSORPTION, DIRECT, AIR	6.700	- 14.000	10.233	3.656	3
	ATOMIC ABSORPTION, FLAMELESS	6.100	- 8.000	7.315	0.853	4
	EMISSION, IC PLASMA	4.000	- 8.000	6.100	1.673	5
_OVER-ALL_	4.000	- 14.000	6.872	1.588	12	
NA	ATOMIC ABSORPTION, DIRECT, AIR	0.020	- 0.500	0.088	0.048	16
	EMISSION, FLAME	0.020	- 0.084	0.052	0.032	3
	EMISSION, IC PLASMA	0.060	- 3.000	0.225	0.129	5
_OVER-ALL_	0.020	- 3.000	0.099	0.066	24	
NH3-N	COLORIMETRIC, INDOPHENOL	0.020	- 0.044	0.029	0.010	5
	COLORIMETRIC, PHENATE	0.015	- 0.050	0.027	0.011	9
	ION SELECTIVE ELECTRODE	0.020	- 0.150	-----	-----	-
OTHER	-----	-----	-----	-----	-	
_OVER-ALL_	0.015	- 0.150	0.027	0.010	16	
NO3-N	COLORIMETRIC, CADMIUM REDUCTION, DIAZOTIZATION	0.056	- 0.069	0.062	0.007	3
	ION CHROMATOGRAPHY	0.050	- 0.278	0.108	0.092	6
	_OVER-ALL_	0.000	- 0.278	0.047	0.023	10
PB	ATOMIC ABSORPTION, FLAMELESS	0.200	- 30.000	8.344	12.778	5
	_OVER-ALL_	0.150	- 100.000	0.554	0.351	5
PH	ELECTROMETRIC	4.260	- 8.300	6.117	0.961	31
	_OVER-ALL_	4.260	- 8.300	6.149	1.017	33
SO4	COLORIMETRIC, METHYL THYMOL BLUE	0.800	- 7.000	3.798	2.749	4
	GRAVIMETRIC, BARIUM SULFATE	2.000	- 4.000	-----	-----	-
	ION CHROMATOGRAPHY	0.126	- 78.900	0.361	0.059	11
	TURBIDIMETRIC, BARIUM SULFATE	-----	-----	-----	-----	-
_OVER-ALL_	0.126	- 78.900	0.721	0.693	18	
SP. COND.	DIRECT READING INSTRUMENT	2.000	- 32.000	4.823	1.829	14
	WHEATSTONE BRIDGE-TYPE CONDUCTIVITY METER	2.800	- 481.000	4.127	1.327	14
	_OVER-ALL_	2.000	- 481.000	4.517	1.595	29
TL	ATOMIC ABSORPTION, DIRECT, AIR	-----	-----	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	0.300	- 0.300	-----	-----	-
_OVER-ALL_	0.300	- 600.000	-----	-----	-	
ZN	ATOMIC ABSORPTION, DIRECT, AIR	4.000	- 20.000	9.913	5.870	8
	EMISSION, IC PLASMA	4.100	- 3.400	9.033	3.477	6
	_OVER-ALL_	3.700	- 110.000	8.724	4.768	7



Table 18 Standard Reference Water Sample SED3 Report for AL

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	15	-20.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
13	21	11.6	EMISSION, IC PLASMA	3,5
20	31	64.8	ATOMIC ABSORPTION, CHELATION EXTRACTION, NITROUS OXIDE	2,4
22	16	-15.0	EMISSION, IC PLASMA	3,5
30	8	-57.5	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
37	8	-57.5	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
38	22	16.9	EMISSION, IC PLASMA	3,5
40	19	1.0	ATOMIC ABSORPTION, DIRECT, FLAMELESS	3
41	75	298.7	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
43	25	32.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
80	18	-4.3	EMISSION, IC PLASMA	3,5
84	25	32.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
104	26	38.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
112	21	11.6	EMISSION, DC PLASMA	7
135	16	-15.0	EMISSION, IC PLASMA	3,5
142	11	-41.5	EMISSION, IC PLASMA	3,5
144	19	1.0	EMISSION, IC PLASMA	3,5

17 Labs had a total range of 8 to 75 and a mean of 18.8 with a standard deviation of 6.4 and a 95% confidence interval of the mean +/- 3.4.

Table 18 Standard Reference Water Sample SED3 Report for B

Code Number	Reported value	Pct. dev. from mean	Methods	References
22	< 10		IGNORED EMISSION, IC PLASMA	3
38	51	98.1	EMISSION, IC PLASMA	3
112	5	-80.6	EMISSION, DC PLASMA	7
127	< 30		IGNORED OTHER	
135	22	-14.6	EMISSION, IC PLASMA	3
144	25	-2.9	EMISSION, IC PLASMA	3

6 Labs had a total range of 5 to 51. INSUFFICIENT DATA TO DEFINE MEAN AND STANDARD OF DEVIATION.

Table 18 Standard Reference Water Sample SED3 Report for BA

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	190	4.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
13	210	15.0	EMISSION, IC PLASMA	3,5
20	180	-1.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
22	200	9.5	EMISSION, IC PLASMA	3,5
30	130	-28.8	ATOMIC ABSORPTION, FLAMELESS	3
37	150	-17.9	ATOMIC ABSORPTION, FLAMELESS	3
38	250	36.9	EMISSION, IC PLASMA	3,5
40	210	15.0	EMISSION, IC PLASMA	3,5
41	790	332.5	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
43	450	146.4	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
80	190	4.0	EMISSION, IC PLASMA	3,5
84	150	-17.9	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
104	390	113.5	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
112	160	-12.4	EMISSION, DC PLASMA	2,7
127	200	9.5	OTHER	
135	160	-12.4	EMISSION, IC PLASMA	3,5
142	170	-6.9	EMISSION, IC PLASMA	3,5
144	190	4.0	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

18 Labs had a total range of 130 to 790 and a mean of 183 with a standard deviation of 30 and a 95% confidence interval of the mean +/- 17.

Table 18 Standard Reference Water Sample SED3 Report for BE

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	< 1.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
13	1.2	7.3	EMISSION, IC PLASMA	3,5
20	0.8	-28.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
22	< 1.0		IGNORED EMISSION, IC PLASMA	3,5
37	0.7	-37.4	ATOMIC ABSORPTION, FLAMELESS	3
38	< 1.0		IGNORED EMISSION, IC PLASMA	3,5
41	2.5	123.6	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
43	0.8	-28.5	ATOMIC ABSORPTION, FLAMELESS	3
50	366.0	3E+04	REJECT ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
80	1.0	-10.6	EMISSION, IC PLASMA	3,5
104	< 1.0		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3,4
112	2.0	78.9	OTHER	
127	1.1	-1.6	OTHER	
135	0.4	-64.2	EMISSION, IC PLASMA	3,5
142	0.9	-19.5	EMISSION, IC PLASMA	3,5
144	0.9	-19.5	EMISSION, IC PLASMA	3,5

16 Labs had a total range of 0.4 to 366.0 and a mean of 1.12 with a standard deviation of 0.61 and a 95% confidence interval of the mean +/- 0.41.

Table 18 Standard Reference Water Sample SED3 Report for C, INORG

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	0.1	-89.5	REJECT	
20	1.8	89.5	REJECT	

2 Labs had a total range of 0.1 to 1.8. INSUFFICIENT DATA TO DEFINE MEAN AND STANDARD OF DEVIATION.

Table 18 Standard Reference Water Sample SED3 Report for C, TOTAL

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	6	9.1	REJECT	
38	5	-9.1	REJECT	

2 Labs had a total range of 5 to 6.  
INSUFFICIENT DATA TO DEFINE MEAN AND STANDARD OF DEVIATION.

Table 18 Standard Reference Water Sample SED3 Report for CA

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	4	-11.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
13	5	11.1	EMISSION, IC PLASMA	3,5,7
20	< 1		IGNORED	3,5,7
22	4	-11.1	EMISSION, IC PLASMA	3,5,7
30	4	-11.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
37	3	-33.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	8	77.8	EMISSION, IC PLASMA	3,5,7
40	4	-11.1	EMISSION, IC PLASMA	3,5,7
41	9	100.0	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
43	4	-11.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	1196	3E+04	REJECT	1,2,3,4
59	1	-77.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	4	-11.1	EMISSION, IC PLASMA	3,5,7
84	5	11.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	7	55.6	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
112	2	-55.6	EMISSION, DC PLASMA	7
127	5	11.1	OTHER	
135	4	-11.1	EMISSION, IC PLASMA	3,5,7
142	4	-11.1	EMISSION, IC PLASMA	3,5,7
144	4	-11.1	EMISSION, IC PLASMA	3,5,7

20 Labs had a total range of 1 to 1196 and a mean of 4.5  
with a standard deviation of 1.9 and a 95% confidence interval of the mean +/- 1.0.

Table 18 Standard Reference Water Sample SED3 Report for CD

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	< 0.5		IGNORED	1,2,3,4
13	< 1.0		IGNORED	3,5
22	< 1.0		IGNORED	3,5
30	0.2	-74.3	ATOMIC ABSORPTION, FLAMELESS	3
33	1.6	105.7	NOT REPORTED	
37	< 0.1		IGNORED	3
38	< 1.0		IGNORED	3
40	< 0.2		IGNORED	3
41	0.6	-22.9	ATOMIC ABSORPTION, FLAMELESS	3
43	0.6	-22.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	< 3.0		IGNORED	2
59	1.4	80.0	ANODIC STRIPPING VOLTAMMETRY, DIFFERENTIAL PULSE	1,2,3,4
80	0.2	-74.3	ATOMIC ABSORPTION, FLAMELESS	3
84	< 5.0		IGNORED	1,2,3,4
104	< 1.0		IGNORED	1,2,3,4
112	< 0.2		IGNORED	1,2,3,4
127	< 0.2		IGNORED	
135	1.8	131.4	EMISSION, IC PLASMA	3,5
142	0.1	-87.1	ATOMIC ABSORPTION, FLAMELESS	3
144	0.5	-35.7	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

20 Labs had a total range of 0.1 to 1.8 and a mean of 0.78  
with a standard deviation of 0.65 and a 95% confidence interval of the mean +/- 0.50.

Table 18 Standard Reference Water Sample SED3 Report for CO

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	4	0.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	< 3		IGNORED	3,5
22	3	-25.0	EMISSION, IC PLASMA	3,5
30	2	-50.0	ATOMIC ABSORPTION, FLAMELESS	3
41	25	525.0	REJECT	1,2,3,4
59	3	-25.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	6	50.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	5	25.0	OTHER	
127	5	25.0	OTHER	
135	4	0.0	EMISSION, IC PLASMA	3,5
142	4	0.0	EMISSION, IC PLASMA	3,5
144	65	1525.0	REJECT	3,5

12 Labs had a total range of 2 to 65 and a mean of 4.0  
with a standard deviation of 1.2 and a 95% confidence interval of the mean +/- 0.9.

Table 18 Standard Reference Water Sample SED3 Report for CR TOT

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	14	-24.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	8	-56.8	EMISSION, IC PLASMA	3
20	26	40.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	39	110.8	EMISSION, IC PLASMA	3
30	4	-78.4	ATOMIC ABSORPTION, FLAMELESS	3
33	4	-78.4	NOT REPORTED	
37	9	-51.4	ATOMIC ABSORPTION, FLAMELESS	3
38	20	8.1	ATOMIC ABSORPTION, FLAMELESS	3
40	35	89.2	ATOMIC ABSORPTION, FLAMELESS	3
41	37	100.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	27	45.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	12	-35.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	14	-24.3	EMISSION, IC PLASMA	3
84	22	18.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	17	-8.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	15	-18.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	22	18.9	OTHER	1,2,3,4
135	15	-18.9	EMISSION, IC PLASMA	3
142	12	-35.1	EMISSION, IC PLASMA	3
144	18	-2.7	EMISSION, IC PLASMA	3

20 Labs had a total range of 4 to 39 and a mean of 18.5 with a standard deviation of 10.2 and a 95% confidence interval of the mean +/- 4.8.

Table 18 Standard Reference Water Sample SED3 Report for CU

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	< 1		IGNORED	
13	9	-35.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	13	-6.4	EMISSION, IC PLASMA	3,5
22	13	-6.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
30	13	-6.4	EMISSION, IC PLASMA	3,5
33	11	-20.8	ATOMIC ABSORPTION, FLAMELESS	3
37	10	-28.0	NOT REPORTED	
38	20	43.9	ATOMIC ABSORPTION, FLAMELESS	3
40	19	36.7	EMISSION, IC PLASMA	3,5
41	20	43.9	EMISSION, IC PLASMA	3,5
43	16	15.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	11	-20.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	13	-6.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
84	18	29.5	EMISSION, IC PLASMA	3,5
104	17	22.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	13	-6.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	13	-6.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
135	12	-13.6	OTHER	
142	10	-28.0	EMISSION, IC PLASMA	3,5
144	13	-6.4	EMISSION, IC PLASMA	3,5
			MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

20 Labs had a total range of 9 to 20 and a mean of 13.9 with a standard deviation of 3.4 and a 95% confidence interval of the mean +/- 1.7.

Table 18 Standard Reference Water Sample SED3 Report for FE

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	14	-1.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	18	26.6	EMISSION, IC PLASMA	3,5
20	5	-64.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	14	-1.6	EMISSION, IC PLASMA	3,5
30	10	-29.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	9	-36.7	NOT REPORTED	
37	9	-36.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	22	54.7	EMISSION, IC PLASMA	3,5
40	18	26.6	EMISSION, IC PLASMA	3,5
41	3	-78.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	21	47.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	17901	1E+05	REJECT	
59	12	-15.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	19	33.6	EMISSION, IC PLASMA	3,5
84	22	54.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	16	12.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	200	1306.3	REJECT	
135	15	5.5	OTHER	
142	13	-8.6	EMISSION, IC PLASMA	3,5
144	16	12.5	EMISSION, IC PLASMA	3,5

20 Labs had a total range of 3 to 17901 and a mean of 14.2 with a standard deviation of 5.5 and a 95% confidence interval of the mean +/- 2.7.

Table 18 Standard Reference Water Sample SED3 Report for HG

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	< 0.01		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
13	< 1.00		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
20	< 0.00		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
22	< 0.20		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
30	0.02	-90.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
33	21.40	1E+04	REJECT NOT REPORTED	
37	0.04	-80.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
38	< 0.10		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
40	0.34	70.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
50	886.50	4E+05	REJECT ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
80	0.07	-65.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
84	0.53	165.0	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
104	< 0.10		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
112	< 0.16		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
127	< 0.20		IGNORED OTHER	
135	< 0.03		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4
142	< 0.05		IGNORED ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	1,2,3,4

17 Labs had a total range of 0.02 to 886.5 and a mean of 0.200 with a standard deviation of 0.226 and a 95% confidence interval of the mean +/- 0.280.

Table 18 Standard Reference Water Sample SED3 Report for K

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	2.0	-33.0	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	4.1	37.3	EMISSION, FLAME, PHOTOMETRIC	1,2
20	2.6	-12.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	2.6	-12.9	EMISSION, IC PLASMA	3
30	1.8	-39.7	EMISSION, FLAME, PHOTOMETRIC	1,2
33	1.3	-56.5	NOT REPORTED	
37	1.9	-36.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
40	2.9	-2.9	EMISSION, IC PLASMA	3
41	48.0	1507.1	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	4.7	57.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	2866.0		REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	2.8	-6.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	3.1	3.8	EMISSION, IC PLASMA	3
84	5.7	90.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	24.8	730.3	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	3.5	17.2	OTHER	
135	3.2	7.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
142	2.6	-12.9	EMISSION, IC PLASMA	3

18 Labs had a total range of 1.3 to 2866.0 and a mean of 2.99 with a standard deviation of 1.16 and a 95% confidence interval of the mean +/- 0.64.

Table 18 Standard Reference Water Sample SED3 Report for LI

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	13	-20.4	EMISSION, FLAME	1
13	21	28.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
22	13	-20.4	EMISSION, IC PLASMA	3,5
41	27	65.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
50	16600	1E+05	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
59	13	-20.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
112	17	4.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
127	< 1		IGNORED OTHER	
135	18	10.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
142	8	-51.0	EMISSION, IC PLASMA	3,5
144	17	4.1	EMISSION, IC PLASMA	3,5

11 Labs had a total range of 8 to 16600 and a mean of 16.3 with a standard deviation of 5.5 and a 95% confidence interval of the mean +/- 4.2.

Table 18 Standard Reference Water Sample SED3 Report for MG

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	3.1	-14.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
13	4.2	15.8	EMISSION, IC PLASMA	3,5
20	3.5	-3.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	3.8	4.8	EMISSION, IC PLASMA	3,5
30	2.5	-31.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
33	0.7	-80.7	NOT REPORTED	
37	2.8	-22.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	10.0	175.8	REJECT EMISSION, IC PLASMA	1,2,3,4
40	3.6	-0.7	EMISSION, IC PLASMA	3,5
41	6.5	79.2	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
43	4.7	29.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	2766.0	8E+04	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	2.7	-25.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	3.5	-3.5	EMISSION, IC PLASMA	3,5
84	3.4	48.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	3.6	54.4	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,7
112	3.1	-14.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	3.8	4.8	OTHER	
135	3.2	-11.8	EMISSION, IC PLASMA	3,5
142	3.2	-11.8	EMISSION, IC PLASMA	3,5
144	3.0	-17.3	EMISSION, IC PLASMA	3,5

21 Labs had a total range of 0.7 to 2766.0 and a mean of 3.63 with a standard deviation of 1.28 and a 95% confidence interval of the mean +/- 0.62.

Table 18 Standard Reference Water Sample SED3 Report for MN

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	120	1.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	140	18.8	EMISSION, IC PLASMA	3,5
22	100	-15.2	EMISSION, IC PLASMA	3,5
30	90	-23.7	ATOMIC ABSORPTION, FLAMELESS	3
33	60	-49.1	NOT REPORTED	
37	860	629.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
38	150	27.2	EMISSION, IC PLASMA	3,5
40	120	1.8	EMISSION, IC PLASMA	3,5
41	120	1.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	150	27.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	110	-6.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	90	-23.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	120	1.8	EMISSION, IC PLASMA	3,5
84	130	10.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	130	10.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	140	18.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	160	35.7	OTHER	1,2,3,4
135	80	-32.1	EMISSION, IC PLASMA	3,5
142	110	-6.7	EMISSION, IC PLASMA	3,5
144	120	1.8	EMISSION, IC PLASMA	3,5

20 Labs had a total range of 60 to 860 and a mean of 118 with a standard deviation of 26 and a 95% confidence interval of the mean +/- 12.

Table 18 Standard Reference Water Sample SED3 Report for MO

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3
13	< 10		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3
22	< 2		IGNORED EMISSION, IC PLASMA	3,5
38	< 2		IGNORED EMISSION, IC PLASMA	3,5
41	< 5		IGNORED ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3
104	13	77.3	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3
112	8	9.1	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,2,3
127	1	-86.4	OTHER	1,2,3
135	< 22		IGNORED EMISSION, IC PLASMA	3,5

9 Labs had a total range of 1 to 13. INSUFFICIENT DATA TO DEFINE MEAN AND STANDARD OF DEVIATION.

Table 18 Standard Reference Water Sample SED3 Report for NA

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	169	16.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	127	-12.2	EMISSION, IC PLASMA	3,5
20	86	-40.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	263	81.9	EMISSION, IC PLASMA	3,5
30	158	9.3	EMISSION, FLAME	1,2
33	169	16.9	NOT REPORTED	
37	4	-97.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
40	275	90.2	EMISSION, IC PLASMA	3,5
41	18	-87.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	182	25.9	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	165	14.1	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	1	-99.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	249	72.2	EMISSION, IC PLASMA	3,5
84	< 1		IGNORED EMISSION, FLAME	3,5
104	9	-93.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	157	8.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	190	31.4	OTHER	
135	160	10.7	EMISSION, IC PLASMA	3,5
142	210	45.2	EMISSION, IC PLASMA	3,5
144	155	7.2	EMISSION, IC PLASMA	3,5

20 Labs had a total range of 1 to 275 and a mean of 144.6 with a standard deviation of 85.1 and a 95% confidence interval of the mean +/- 41.0.

Table 18 Standard Reference Water Sample SED3 Report for NI

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	12	-27.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
13	21	26.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
20	17	2.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	10	-39.8	EMISSION, IC PLASMA	3,5
30	25	50.6	ATOMIC ABSORPTION, FLAMELESS	3
33	17	-2.4	NOT REPORTED	
37	7	-57.8	ATOMIC ABSORPTION, FLAMELESS	3
38	23	38.6	ATOMIC ABSORPTION, FLAMELESS	3
40	19	14.5	ATOMIC ABSORPTION, FLAMELESS	3
41	34	104.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	18	8.4	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
59	12	-27.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	13	-21.7	EMISSION, IC PLASMA	3,5
84	28	68.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	10	-39.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	12	-27.7	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
127	14	-15.7	OTHER	1,2,3,4
135	11	-33.7	EMISSION, IC PLASMA	3,5
142	12	-27.7	EMISSION, IC PLASMA	3,5
144	17	2.4	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

20 Labs had a total range of 7 to 34 and a mean of 16.6 with a standard deviation of 6.8 and a 95% confidence interval of the mean +/- 3.2.

Table 18 Standard Reference Water Sample SED3 Report for PB

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	1090	8407.3	REJECT ATOMIC ABSORPTION, FLAMELESS	3
13	< 5		IGNORED EMISSION, IC PLASMA	3,5
20	8	-37.6	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
22	19	48.3	EMISSION, IC PLASMA	3,5
30	9	-29.8	ATOMIC ABSORPTION, FLAMELESS	3
33	8	-37.6	NOT REPORTED	
37	5	-61.0	ATOMIC ABSORPTION, FLAMELESS	3
38	13	-24.9	ATOMIC ABSORPTION, FLAMELESS	3
40	3	-76.6	ATOMIC ABSORPTION, FLAMELESS	3
41	43	251.2	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
43	19	48.3	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
50	8660	7E+04	REJECT ANODIC STRIPPING VOLTAMMETRY	2
59	28	118.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
80	12	-6.3	ATOMIC ABSORPTION, FLAMELESS	3
84	< 50		IGNORED ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
104	6	-53.2	ATOMIC ABSORPTION, DIRECT, AIR	1,2,3,4
112	24	87.3	EMISSION, DC PLASMA	7
127	13	1.5	OTHER	
135	26	102.9	EMISSION, IC PLASMA	3,5
142	8	-37.6	EMISSION, IC PLASMA	3,5
144	1	-92.2	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

21 Labs had a total range of 1 to 8660 and a mean of 12.8 with a standard deviation of 8.3 and a 95% confidence interval of the mean +/- 4.4.

Table 18 Standard Reference Water Sample SED3 Report for SE

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	< 1.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
13	< 0.6		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
22	27.3	3312.5	REJECT ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
37	1.6	100.0	ATOMIC ABSORPTION, FLAMELESS	3
38	< 0.1		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
40	0.2	-75.0	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
41	< 0.3	-62.5	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
43	< 0.9		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
50	1260.0	2E+05	REJECT OTHER	
80	0.2	-75.0	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
84	< 10.0		IGNORED ATOMIC ABSORPTION, FLAMELESS	3
104	0.2	-75.0	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
112	< 2.7	237.5	EMISSION, DC PLASMA	7
127	< 2.0		IGNORED OTHER	
135	0.4	-50.0	ATOMIC ABSORPTION, HYDRIDE	1,2,3,4
142	< 0.4		IGNORED ATOMIC ABSORPTION, FLAMELESS	3

16 Labs had a total range of 0.2 to 1260.0 and a mean of 0.80 with a standard deviation of 0.98 and a 95% confidence interval of the mean +/- 0.90.

Table 18 Standard Reference Water Sample SED3 Report for SR

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	49	14.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
13	52	21.8	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
22	41	-4.0	EMISSION, IC PLASMA	3,5
41	19	-55.5	ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
80	52	21.8	EMISSION, IC PLASMA	3,5
104	98	129.5	REJECT ATOMIC ABSORPTION, DIRECT, AIR	1,2,4
112	41	-4.0	EMISSION, DC PLASMA	7
127	56	31.1	OTHER	
135	20	-53.2	OTHER	
142	47	10.1	EMISSION, IC PLASMA	3,5
144	50	17.1	MASS SPECTROMETRY, IC PLASMA, ISOTOPE DILUTION	7

11 Labs had a total range of 19 to 98 and a mean of 42.7 with a standard deviation of 13.1 and a 95% confidence interval of the mean +/- 9.4.

Table 18 Standard Reference Water Sample SED3 Report for V

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	30	-30.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
13	28	-35.2	EMISSION, IC PLASMA	3,5
22	43	-0.4	EMISSION, IC PLASMA	3,5
38	81	87.6	EMISSION, IC PLASMA	3,5
41	85	96.8	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
80	27	-37.5	EMISSION, IC PLASMA	3,5
104	46	6.5	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	1,3
112	18	-58.3	EMISSION, DC PLASMA	7
127	40	-7.4	OTHER	
135	< 22		IGNORED OTHER	
142	25	-42.1	EMISSION, IC PLASMA	3,5
144	52	20.4	EMISSION, IC PLASMA	3,5

12 Labs had a total range of 18 to 85 and a mean of 43.2 with a standard deviation of 22.1 and a 95% confidence interval of the mean +/- 14.9.

Table 18 Standard Reference Water Sample SED3 Report for ZN

Code Number	Reported value	Pct. dev. from mean	Methods	References
8	46	-17.5	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
13	56	0.4	EMISSION, IC PLASMA	3, 5
20	55	-1.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
22	65	16.6	EMISSION, IC PLASMA	3, 5
30	33	-40.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
33	34	-39.0	NOT REPORTED	
37	33	-40.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
38	70	25.6	EMISSION, IC PLASMA	3, 5
40	62	11.2	EMISSION, IC PLASMA	3, 5
41	80	43.5	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
43	64	14.8	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
59	43	-22.9	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
80	71	27.4	EMISSION, IC PLASMA	3, 5
84	80	43.5	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
104	58	4.0	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
112	55	-1.3	ATOMIC ABSORPTION, DIRECT, AIR	1, 2, 3, 4
127	66	18.4	OTHER	1, 2, 3, 4
135	43	-22.9	EMISSION, IC PLASMA	3, 5
142	47	-15.7	EMISSION, IC PLASMA	3, 5
144	54	-3.1	EMISSION, IC PLASMA	3, 5

20 Labs had a total range of 33 to 80 and a mean of 55.8 with a standard deviation of 14.4 and a 95% confidence interval of the mean +/- 6.7.

Table 19. Statistics by method for standard reference sample SED3

Determination	Method	Range:		Mean	Standard Deviation	N
		from	to			
AL	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	15.000	75.000	25.000	-----	2
	ATOMIC ABSORPTION, DIRECT, FLAMELESS	8.000	19.000	8.000	-----	2
	EMISSION, IC PLASMA	11.000	22.000	17.571	3.690	7
	OVER-ALL	8.000	75.000	18.813	6.442	16
B	EMISSION, IC PLASMA	22.000	51.000	32.667	15.948	3
	OVER-ALL	5.000	51.000	25.750	18.998	4
BA	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	150.000	790.000	358.333	244.656	6
	EMISSION, IC PLASMA	160.000	250.000	198.571	29.681	7
	OVER-ALL	130.000	790.000	182.667	30.347	15
BE	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	0.800	366.000	-----	-----	-
	EMISSION, IC PLASMA	0.400	1.200	0.880	0.295	5
	OVER-ALL	0.400	366.000	1.118	0.608	11
CA	ATOMIC ABSORPTION, DIRECT, AIR	1.000	1196.000	3.400	1.517	5
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	4.000	9.000	6.667	2.517	3
	EMISSION, IC PLASMA	4.000	8.000	4.000	0.000	6
	OVER-ALL	1.000	1196.000	4.500	1.917	18
CD	ATOMIC ABSORPTION, DIRECT, AIR	0.600	1.400	-----	-----	-
	ATOMIC ABSORPTION, FLAMELESS	0.100	0.600	0.275	0.222	4
	EMISSION, IC PLASMA	1.800	1.800	-----	-----	-
	OVER-ALL	0.100	1.800	0.778	0.650	9
CO	ATOMIC ABSORPTION, DIRECT, AIR	3.000	25.000	4.333	1.528	3
	EMISSION, IC PLASMA	3.000	65.000	4.000	-----	2
	OVER-ALL	2.000	65.000	4.000	1.225	9
CR TOT	ATOMIC ABSORPTION, DIRECT, AIR	12.000	37.000	21.250	8.447	8
	ATOMIC ABSORPTION, FLAMELESS	4.000	35.000	17.000	13.736	4
	EMISSION, IC PLASMA	8.000	39.000	13.400	3.715	5
	OVER-ALL	4.000	39.000	18.500	10.164	20
CU	ATOMIC ABSORPTION, DIRECT, AIR	11.000	20.000	15.429	3.207	7
	EMISSION, IC PLASMA	9.000	20.000	13.714	4.231	7
	OVER-ALL	9.000	20.000	13.895	3.430	19
FE	ATOMIC ABSORPTION, DIRECT, AIR	3.000	-17901.000	12.444	6.540	9
	EMISSION, IC PLASMA	13.000	22.000	16.875	2.949	8
	OVER-ALL	3.000	-17901.000	14.222	5.505	18
HG	ATOMIC ABSORPTION, FLAMELESS, COLD VAPOR	0.020	886.500	0.200	0.239	5
	OVER-ALL	0.020	886.500	0.200	0.226	5
K	ATOMIC ABSORPTION, DIRECT, AIR	1.900	2866.000	3.271	1.421	7
	EMISSION, IC PLASMA	2.600	3.100	2.800	0.245	4
K	OVER-ALL	1.300	2866.000	2.987	1.158	15
LI	ATOMIC ABSORPTION, DIRECT, AIR	13.000	-16600.000	19.200	5.215	5
	EMISSION, IC PLASMA	8.000	17.000	12.667	4.509	3
	OVER-ALL	8.000	-16600.000	16.333	5.500	9
MG	ATOMIC ABSORPTION, DIRECT, AIR	2.500	2766.000	3.529	1.106	7
	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	3.100	6.500	5.067	1.762	3
	EMISSION, IC PLASMA	3.000	10.000	3.500	0.412	7
	OVER-ALL	0.700	2766.000	3.626	1.281	19
MN	ATOMIC ABSORPTION, DIRECT, AIR	90.000	860.000	123.750	18.468	8
	EMISSION, IC PLASMA	80.000	150.000	117.500	21.876	8
	OVER-ALL	60.000	860.000	117.895	25.729	19
MO	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	8.000	13.000	-----	-----	-
	EMISSION, IC PLASMA	-----	-----	-----	-----	-
	OVER-ALL	1.000	13.000	7.333	6.028	3
NA	ATOMIC ABSORPTION, DIRECT, AIR	1.000	182.000	87.889	80.512	9
	EMISSION, IC PLASMA	127.000	275.000	205.571	58.923	7
	OVER-ALL	1.000	275.000	144.579	85.145	19
NI	ATOMIC ABSORPTION, DIRECT, AIR	10.000	34.000	18.222	8.197	9
	ATOMIC ABSORPTION, FLAMELESS	7.000	25.000	18.500	8.062	4
	EMISSION, IC PLASMA	10.000	13.000	11.500	1.291	4
	OVER-ALL	7.000	34.000	16.600	6.832	20
PB	ATOMIC ABSORPTION, DIRECT, AIR	6.000	45.000	21.200	15.991	5
	ATOMIC ABSORPTION, FLAMELESS	3.000	1090.000	9.000	5.244	5
	EMISSION, IC PLASMA	8.000	26.000	17.667	9.074	3
	OVER-ALL	1.000	8660.000	12.813	8.344	16
SE	ATOMIC ABSORPTION, FLAMELESS	1.600	1.600	-----	-----	-
	ATOMIC ABSORPTION, HYDRIDE	0.200	27.300	0.260	0.089	5
	OVER-ALL	0.200	1260.000	0.800	0.978	7
SR	ATOMIC ABSORPTION, DIRECT, AIR	19.000	98.000	54.500	32.604	4
	EMISSION, IC PLASMA	41.000	52.000	46.667	5.508	3
	OVER-ALL	19.000	98.000	42.700	13.098	10
V	ATOMIC ABSORPTION, DIRECT, NITROUS OXIDE	30.000	85.000	53.667	28.290	3
	EMISSION, IC PLASMA	25.000	81.000	42.667	21.584	6
	OVER-ALL	18.000	85.000	43.182	22.149	11
ZN	ATOMIC ABSORPTION, DIRECT, AIR	33.000	80.000	54.700	16.773	10
	EMISSION, IC PLASMA	43.000	71.000	58.500	10.296	8
	OVER-ALL	33.000	80.000	55.750	14.371	20